

The results set forth herein are provided by SGS North America Inc.

*e-Hardcopy 2.0*  
*Automated Report*

## Technical Report for

**AECOM, INC.**

**N6274223F0104 RH Fire Suppression System**

**60697810**

**SGS Job Number: FC8392**

**Sampling Date: 08/01/23**



### Report to:

**AECOM, Inc**  
**7595 Technology Way**  
**Denver, CO 80237**  
**katie.abbott@aecom.com; mark.kromis@aecom.com;**  
**watson.tanji@aecom.com; kristin.rutherford@aecom.com**  
**ATTN: Katie Abbott**

**Total number of pages in report: 32**



Test results contained within this data package meet the requirements of the National Environmental Laboratory Accreditation Program and/or state specific certification programs as applicable unless noted in the narrative, comments or footnotes.

**Norm Farmer**  
**Technical Director**

**Client Service contact: Elvin Kumar 407-425-6700**

Certifications: FL(E83510), LA(03051), KS(E-10327), NC(573), NJ(FL002), NY(12022), SC(96038001)

DoD ELAP(ANAB L2229), AZ(AZ0806), CA(2937), TX(T104704404), PA(68-03573), VA(460177),

AL, AK, AR, CT, IA, KY, MA, MI, MS, ND, NH, NV, OK, OR, IL, UT, VT, WA, WI, WV

This report shall not be reproduced, except in its entirety, without the written approval of SGS.

Test results relate only to samples analyzed.

# Table of Contents

-1-

<b>Section 1: Sample Summary</b> .....	<b>3</b>
<b>Section 2: Case Narrative/Conformance Summary</b> .....	<b>4</b>
<b>Section 3: Summary of Hits</b> .....	<b>5</b>
<b>Section 4: Sample Results</b> .....	<b>6</b>
<b>4.1: FC8392-1: AF-RHMW10-WGN01LF-2308</b> .....	<b>7</b>
<b>Section 5: Misc. Forms</b> .....	<b>10</b>
<b>5.1: Chain of Custody</b> .....	<b>11</b>
<b>5.2: QC Evaluation: DOD QSM5.x Limits</b> .....	<b>13</b>
<b>Section 6: MS Semi-volatiles - QC Data Summaries</b> .....	<b>14</b>
<b>6.1: Method Blank Summary</b> .....	<b>15</b>
<b>6.2: Blank Spike Summary</b> .....	<b>25</b>
<b>6.3: Matrix Spike Summary</b> .....	<b>29</b>
<b>6.4: Duplicate Summary</b> .....	<b>31</b>

1

2

3

4

5

6



## Sample Summary

AECOM, INC.

Job No: FC8392

N6274223F0104 RH Fire Suppression System  
Project No: 60697810

Sample Number	Collected Date	Time By	Received	Matrix Code	Type	Client Sample ID
FC8392-1	08/01/23	10:40	CPAY 08/03/23	AQ	Ground Water	AF-RHMW10-WGN01LF-2308

## SAMPLE DELIVERY GROUP CASE NARRATIVE

**Client:** AECOM, INC.

**Job No:** FC8392

**Site:** N6274223F0104 RH Fire Suppression System

**Report Date:** 8/14/2023 5:11:44 PM

On 08/03/2023, 1 Sample(s), 0 Trip Blank(s) and 0 Field Blank(s) were received at SGS North America Inc - Orlando. at a maximum corrected temperature of 4 C. Samples were intact and chemically preserved, unless noted below. A SGS North America Inc. - Orlando Job Number of FC8392 was assigned to the project.

Laboratory sample ID, client sample ID and dates of sample collection are detailed in the report's Results Summary Section. Specified quality control criteria were achieved for this job except as noted below. For more information, please refer to the analytical results and QC summary pages.

### MS Semi-volatiles By Method EPA DRAFT 1633

**Matrix:** AQ

**Batch ID:** OP98297

Sample(s) FC8439-3MS, FC8439-4DUP were used as the QC samples indicated.

Matrix Spike Recovery(s) for 8:2 Fluorotelomer sulfonate are outside control limits. Probable cause is due to matrix interference.

RPD(s) for Duplicate for Perfluorooctanoic acid are outside control limits for sample OP98297-DUP. Probable cause is due to sample non-homogeneity.

OP98297-MB: All results and recoveries corrected for double EIS.

FC8392-1 for 8:2 Fluorotelomer sulfonate: Associated CCV outside of control limits high, sample was ND.

FC8392-1 for 9Cl-PF3ONS (F-53B Major): Associated CCV outside control limits high. Sample was ND.

OP98297-LLBS for 13C2-8:2FTS: Outside control limits.

OP98297-LLBS for 13C2-6:2FTS: Outside control limits.

SGS North America Inc. - Orlando certifies that data reported for samples received, listed on the associated custody chain or analytical task order, were produced to specifications meeting the Quality System precision, accuracy and completeness objectives except as noted. Estimated non-standard method measurement uncertainty data is available on request, based on quality control bias and implicit for standard methods. Acceptable uncertainty requires tested parameter quality control data to meet method criteria. SGS North America Inc.- Orlando is not responsible for data quality assumptions if partial reports are used and recommends that this report be used in its entirety.

Narrative prepared by:

---

Kim Benham, Client Services (*Signature on File*)

## Summary of Hits

**Job Number:** FC8392  
**Account:** AECOM, INC.  
**Project:** N6274223F0104 RH Fire Suppression System  
**Collected:** 08/01/23



Lab Sample ID	Client Sample ID	Result/ Qual	LOQ	LOD	Units	Method
---------------	------------------	-----------------	-----	-----	-------	--------

FC8392-1      AF-RHMW10-WGN01LF-2308

No hits reported in this sample.

**Sample Results**

---

**Report of Analysis**

---

# Report of Analysis

Client Sample ID:	AF-RHMW10-WGN01LF-2308		
Lab Sample ID:	FC8392-1	Date Sampled:	08/01/23
Matrix:	AQ - Ground Water	Date Received:	08/03/23
Method:	EPA DRAFT 1633 EPA 1633 DRAFT	Percent Solids:	n/a
Project:	N6274223F0104 RH Fire Suppression System		

Run #	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
Run #1	4Q48783.D	1	08/09/23 20:02	AL	08/08/23 10:50	OP98297	S4Q713
Run #2							

Run #	Initial Volume	Final Volume
Run #1	570 ml	5.0 ml
Run #2		

CAS No.	Compound	Result	LOQ	LOD	DL	Units	Q
---------	----------	--------	-----	-----	----	-------	---

**PERFLUOROALKYL CARBOXYLIC ACIDS**

375-22-4	Perfluorobutanoic acid	3.5 U	14	3.5	1.7	ng/l	
2706-90-3	Perfluoropentanoic acid	1.8 U	7.0	1.8	0.82	ng/l	
307-24-4	Perfluorohexanoic acid	1.8 U	3.5	1.8	0.44	ng/l	
375-85-9	Perfluoroheptanoic acid	1.8 U	3.5	1.8	0.44	ng/l	
335-67-1	Perfluorooctanoic acid	0.88 U	3.5	0.88	0.44	ng/l	
375-95-1	Perfluorononanoic acid	1.8 U	3.5	1.8	0.54	ng/l	
335-76-2	Perfluorodecanoic acid	1.8 U	3.5	1.8	0.44	ng/l	
2058-94-8	Perfluoroundecanoic acid	1.8 U	3.5	1.8	0.53	ng/l	
307-55-1	Perfluorododecanoic acid	1.8 U	3.5	1.8	0.53	ng/l	
72629-94-8	Perfluorotridecanoic acid	1.8 U	3.5	1.8	0.74	ng/l	
376-06-7	Perfluorotetradecanoic acid	1.8 U	3.5	1.8	0.44	ng/l	

**PERFLUOROALKYL SULFONIC ACIDS**

375-73-5	Perfluorobutanesulfonic acid	1.8 U	3.5	1.8	0.44	ng/l	
2706-91-4	Perfluoropentanesulfonic acid	3.5 U	4.4	3.5	0.98	ng/l	
355-46-4	Perfluorohexanesulfonic acid	1.8 U	3.5	1.8	0.61	ng/l	
375-92-8	Perfluoroheptanesulfonic acid	1.8 U	3.5	1.8	0.44	ng/l	
1763-23-1	Perfluorooctanesulfonic acid	1.8 U	3.5	1.8	0.47	ng/l	
68259-12-1	Perfluorononanesulfonic acid	1.8 U	3.5	1.8	0.50	ng/l	
335-77-3	Perfluorodecanesulfonic acid	1.8 U	3.5	1.8	0.56	ng/l	
79780-39-5	Perfluorododecanesulfonic aci	3.5 U	4.4	3.5	1.0	ng/l	

**FLUOROTELOMER SULFONIC ACIDS**

757124-72-4	4:2 Fluorotelomer sulfonate	7.0 U	18	7.0	2.8	ng/l	
27619-97-2	6:2 Fluorotelomer sulfonate	7.0 U	18	7.0	3.0	ng/l	
39108-34-4	8:2 Fluorotelomer sulfonate <sup>a</sup>	7.0 U	18	7.0	3.6	ng/l	

**PERFLUOROOCCTANE SULFONAMIDES**

754-91-6	PFOSA	1.8 U	3.5	1.8	0.59	ng/l	
31506-32-8	MeFOSA	3.5 U	7.0	3.5	0.88	ng/l	
4151-50-2	EtFOSA	3.5 U	7.0	3.5	0.88	ng/l	

U = Not detected      LOD = Limit of Detection      J = Indicates an estimated value  
 LOQ = Limit of Quantitation      DL = Detection Limit      B = Indicates analyte found in associated method blank  
 E = Indicates value exceeds calibration range      N = Indicates presumptive evidence of a compound

4.1  
4

# Report of Analysis

Client Sample ID:	AF-RHMW10-WGN01LF-2308		
Lab Sample ID:	FC8392-1	Date Sampled:	08/01/23
Matrix:	AQ - Ground Water	Date Received:	08/03/23
Method:	EPA DRAFT 1633 EPA 1633 DRAFT	Percent Solids:	n/a
Project:	N6274223F0104 RH Fire Suppression System		

CAS No.	Compound	Result	LOQ	LOD	DL	Units	Q
---------	----------	--------	-----	-----	----	-------	---

**PERFLUOROOCCTANE SULFONAMIDOACETIC ACIDS**

2355-31-9	MeFOSAA	3.5 U	4.4	3.5	0.88	ng/l	
2991-50-6	EtFOSAA	3.5 U	4.4	3.5	1.2	ng/l	

**PERFLUOROOCCTANE SULFONAMIDO ETHANOLS**

24448-09-7	MeFOSE	18 U	35	18	3.8	ng/l	
1691-99-2	EtFOSE	18 U	35	18	6.5	ng/l	

**PER and POLYFLUOROETHER CARBOXYLIC ACIDS**

13252-13-6	HFPO-DA (GenX)	1.8 U	3.5	1.8	0.88	ng/l	
919005-14-4	ADONA	3.5 U	7.0	3.5	1.6	ng/l	
377-73-1	PFMPA	1.8 U	7.0	1.8	0.88	ng/l	
863090-89-5	PFMBA	3.5 U	7.0	3.5	1.0	ng/l	
151772-58-6	NFDHA	3.5 U	7.0	3.5	1.1	ng/l	

**PER and POLYFLUOROETHER SULFONIC ACIDS**

756426-58-1	9Cl-PF3ONS (F-53B Major) <sup>b</sup>	3.5 U	7.0	3.5	1.2	ng/l	
763051-92-9	11Cl-PF3OUdS (F-53B Minor)	3.5 U	7.0	3.5	1.5	ng/l	
113507-82-7	PFEESA	1.8 U	7.0	1.8	0.68	ng/l	

**FLUOROTELOMER CARBOXYLIC ACIDS**

356-02-5	3:3 Fluorotelomer carboxylate	8.8 U	18	8.8	4.0	ng/l	
914637-49-3	5:3 Fluorotelomer carboxylate	18 U	88	18	7.7	ng/l	
812-70-4	7:3 Fluorotelomer carboxylate	18 U	88	18	6.9	ng/l	

CAS No.	ID Standard Recoveries	Run# 1	Run# 2	Limits
---------	------------------------	--------	--------	--------

	13C4-PFBA	110%		20-150%
	13C5-PFPeA	100%		20-150%
	13C5-PFHxA	108%		20-150%
	13C4-PFHpA	112%		20-150%
	13C8-PFOA	105%		20-150%
	13C9-PFNA	93%		20-150%
	13C6-PFDA	100%		20-150%
	13C7-PFUnDA	111%		20-150%
	13C2-PFDoDA	101%		20-150%
	13C2-PFTeDA	87%		20-150%
	13C3-PFBS	106%		20-150%
	13C3-PFHxS	115%		20-150%

U = Not detected      LOD = Limit of Detection      J = Indicates an estimated value  
 LOQ = Limit of Quantitation      DL = Detection Limit      B = Indicates analyte found in associated method blank  
 E = Indicates value exceeds calibration range      N = Indicates presumptive evidence of a compound



## Report of Analysis

Client Sample ID: AF-RHMW10-WGN01LF-2308		Date Sampled: 08/01/23
Lab Sample ID: FC8392-1		Date Received: 08/03/23
Matrix: AQ - Ground Water		Percent Solids: n/a
Method: EPA DRAFT 1633 EPA 1633 DRAFT		
Project: N6274223F0104 RH Fire Suppression System		

CAS No.	ID Standard Recoveries	Run# 1	Run# 2	Limits
	13C8-PFOS	102%		20-150%
	13C8-FOSA	75%		20-150%
	d3-MeFOSA	78%		20-150%
	d5-EtFOSA	76%		20-150%
	d3-MeFOSAA	104%		20-150%
	d5-EtFOSAA	105%		20-150%
	d7-MeFOSE	66%		20-150%
	d9-EtFOSE	74%		20-150%
	13C2-4:2FTS	144%		20-180%
	13C2-6:2FTS	149%		20-180%
	13C2-8:2FTS	141%		20-180%
	13C3-HFPO-DA	95%		20-150%

- (a) Associated CCV outside of control limits high, sample was ND.
- (b) Associated CCV outside control limits high. Sample was ND.

---

U = Not detected      LOD = Limit of Detection      J = Indicates an estimated value  
 LOQ = Limit of Quantitation      DL = Detection Limit      B = Indicates analyte found in associated method blank  
 E = Indicates value exceeds calibration range      N = Indicates presumptive evidence of a compound

## Misc. Forms

### Custody Documents and Other Forms

---

**Includes the following where applicable:**

- Chain of Custody
- QC Evaluation: DOD QSM5.x Limits



SGS North America Inc - Orlando  
Chain of Custody

4405 Vineland Road, Suite C-15 Orlando, FL 32811  
TEL: 407-425-6700 FAX: 407-425-0707  
www.sgs.com

**FC8392**

SGS - ORLANDO JOB #:

COC #: 2308AFSG03

PAGE **1** OF **1**

Client / Reporting Information		Project Information				Analytical Information														Matrix Codes	
Company Name: AECOM		Project Name: N6274223F0104 RH Fire Suppression System				<div style="text-align: right; font-size: 2em; font-weight: bold; opacity: 0.5;"> <del>2308 8/3/23</del> </div>														DW - Drinking Water GW - Ground Water WW - Water SW - Surface Water SO - Soil SL - Sludge LIQ - Other Liquid AIR - Air SOL - Other Solid WP - Wipe	
Address: 1001 Bishop St. ste 1600		Street																			
City: Honolulu State: HI Zip: 96813		City Honolulu State Hawaii																			
Project Contact: Katie Abbott Email: katie.abbott@aecom.com		Project # 60697810																			
Project Manager: Watson Tanji Email: watson.tanji@aecom.com		Fax #																			
Phone #: 303-796-4624 / 808-954-4512		Client Purchase Order #																			
Sampler(s) Name(s) (Printed) Sampler 1: <u>Christian Perez</u> Sampler 2: <u>Andy Young</u>																					
SGS Orlando Sample #		COLLECTION			CONTAINER INFORMATION											PFAS EPA Draft 1633		LAB USE ONLY			
Field ID / Point of Collection		DATE	TIME	SAMPLED BY	MATRIX	TOTAL # OF BOTTLES	OTHER	NONE	HCl	NACH	IN03	H2SO4	NACH-ZINC	DI WATER	MCOH						
1 AF-RHMW10-WGN01LF-2308		8/1/23	1040	CP,AY	GW	3		X													
																		INITIAL ASSESSMENT <u>ZB</u>			
																		LABEL VERIFICATION <u>ZB</u>			
																		4.2 ITR			
Turnaround Time ( Business days)				Data Deliverable Information							Comments / Remarks										
10 Day (Business) Approved By: / Date: 7 Day <input checked="" type="checkbox"/> 5 Day 3 Day RUSH 2 Day RUSH 1 Day RUSH Other				<input type="checkbox"/> COMMERCIAL "A" (RESULTS ONLY) <input type="checkbox"/> COMMERCIAL "B" (RESULTS PLUS QC) <input type="checkbox"/> REDT1 (EPA LEVEL 3) <input checked="" type="checkbox"/> FULLT1 (EPA LEVEL 4) <input checked="" type="checkbox"/> EDD'S							EDMS upload database: JBPHE EDMS Coverage: AFFF Assessment Sampling GW United AWB - 010-06268791										
Rush T/A Data Available VIA Email or Lablink				Sample Custody must be documented below each time samples change possession, including courier delivery.																	
Relinquished by Sampler/Affiliation		Date Time:		Received By/Affiliation				Relinquished By/Affiliation				Date Time:		Received By/Affiliation							
1 Andy Young / AECOM		8/1/23 1330		2 Brittany Tominez / AECOM				3 Brittany Tominez / AECOM				8/3/23 1400		4 [Signature] / AECOM							
Relinquished by/Affiliation		Date Time:		Received By/Affiliation				Relinquished By/Affiliation				Date Time:		Received By/Affiliation							
5				6				7						8							

PFAS\_COCs\_ALL\_082023.xls Rev 031318

FC8392: Chain of Custody

Page 1 of 2



5.1 5

## SGS Sample Receipt Summary

Job Number: FC8392

Client: AECOM

Project: N6274223F0104 RH Fire Suppression System

Date / Time Received: 8/3/2023 10:00:00 AM

Delivery Method: United Cargo/Airspace

Airbill #'s: United Cargo AWB #: 016-06268791

Therm ID: IR 1;

Therm CF: -0.2;

# of Coolers: 1

Cooler Temps (Raw Measured) °C: Cooler 1: (4.2);

Cooler Temps (Corrected) °C: Cooler 1: (4.0);

**Cooler Information**

Y or N

- 1. Custody Seals Present
- 2. Custody Seals Intact
- 3. Temp criteria achieved
- 4. Cooler temp verification IR Gun
- 5. Cooler media Ice (Bag)

**Trip Blank Information**

Y or N N/A

- 1. Trip Blank present / cooler
  - 2. Trip Blank listed on COC
- W or S N/A
- 3. Type Of TB Received

**Sample Information**

Y or N N/A

- 1. Sample labels present on bottles
- 2. Samples preserved properly
- 3. Sufficient volume/containers recvd for analysis:
- 4. Condition of sample Intact
- 5. Sample recvd within HT
- 6. Dates/Times/IDs on COC match Sample Label
- 7. VOCs have headspace
- 8. Bottles received for unspecified tests
- 9. Compositing instructions clear
- 10. Voa Soil Kits/Jars received past 48hrs?
- 11. % Solids Jar received?
- 12. Residual Chlorine Present?

**Misc. Information**

Number of Encores: 25-Gram \_\_\_\_\_ 5-Gram \_\_\_\_\_ Number of 5035 Field Kits: \_\_\_\_\_ Number of Lab Filtered Metals: \_\_\_\_\_  
 Test Strip Lot #'s: pH 0-3 230320 pH 10-12 \_\_\_\_\_ Other: (Specify) pH 1.0 - 12.0 222221  
 Residual Chlorine Test Strip Lot #: \_\_\_\_\_

Comments

SM001  
Rev. Date 05/24/17

Technician: ZANEB

Date: 8/3/2023 10:00:00 AM

Reviewer: ZB

Date: 8/9/2023

**FC8392: Chain of Custody**

Page 2 of 2

5.1  
5

# QC Evaluation: DOD QSM5.x Limits

**Job Number:** FC8392  
**Account:** AECOM, INC.  
**Project:** N6274223F0104 RH Fire Suppression System  
**Collected:** 08/01/23

QC Sample ID	CAS#	Analyte	Sample Result Type	Result Type	Units	Limits
--------------	------	---------	--------------------	-------------	-------	--------

No DOD QSM5.x Limits found for methods in this job.

---

\* Sample used for QC is not from job FC8392

5.2  
5

## MS Semi-volatiles

---

### QC Data Summaries

---

#### Includes the following where applicable:

- Method Blank Summaries
- Blank Spike Summaries
- Matrix Spike and Duplicate Summaries

# Instrument Blank

Job Number: FC8392  
 Account: AECOMCOD AECOM, INC.  
 Project: N6274223F0104 RH Fire Suppression System

Sample	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
S4Q713-IBLK	4Q48763.D	1	08/09/23	AL	n/a	n/a	S4Q713

The QC reported here applies to the following samples:

Method: EPA DRAFT 1633

FC8392-1

CAS No.	Compound	Result	RL	MDL	Units	Q
375-22-4	Perfluorobutanoic acid	ND	0.016	0.0019	ug/l	
2706-90-3	Perfluoropentanoic acid	ND	0.0080	0.00094	ug/l	
307-24-4	Perfluorohexanoic acid	ND	0.0040	0.00050	ug/l	
375-85-9	Perfluoroheptanoic acid	ND	0.0040	0.00050	ug/l	
335-67-1	Perfluorooctanoic acid	ND	0.0040	0.00050	ug/l	
375-95-1	Perfluorononanoic acid	ND	0.0040	0.00061	ug/l	
335-76-2	Perfluorodecanoic acid	ND	0.0040	0.00050	ug/l	
2058-94-8	Perfluoroundecanoic acid	ND	0.0040	0.00060	ug/l	
307-55-1	Perfluorododecanoic acid	ND	0.0040	0.00060	ug/l	
72629-94-8	Perfluorotridecanoic acid	ND	0.0040	0.00084	ug/l	
376-06-7	Perfluorotetradecanoic acid	ND	0.0040	0.00050	ug/l	
375-73-5	Perfluorobutanesulfonic acid	ND	0.0040	0.00050	ug/l	
2706-91-4	Perfluoropentanesulfonic acid	ND	0.0050	0.0011	ug/l	
355-46-4	Perfluorohexanesulfonic acid	ND	0.0040	0.00070	ug/l	
375-92-8	Perfluoroheptanesulfonic acid	ND	0.0040	0.00050	ug/l	
1763-23-1	Perfluorooctanesulfonic acid	ND	0.0040	0.00054	ug/l	
68259-12-1	Perfluorononanesulfonic acid	ND	0.0040	0.00057	ug/l	
335-77-3	Perfluorodecanesulfonic acid	ND	0.0040	0.00064	ug/l	
79780-39-5	Perfluorododecanesulfonic aci	ND	0.0050	0.0011	ug/l	
757124-72-44:2	Fluorotelomer sulfonate	ND	0.020	0.0032	ug/l	
27619-97-2	6:2 Fluorotelomer sulfonate	ND	0.020	0.0035	ug/l	
39108-34-4	8:2 Fluorotelomer sulfonate	ND	0.020	0.0041	ug/l	
754-91-6	PFOSA	ND	0.0040	0.00067	ug/l	
31506-32-8	MeFOSA	ND	0.0080	0.0010	ug/l	
4151-50-2	EtFOSA	0.0020	0.0080	0.0010	ug/l	J
2355-31-9	MeFOSAA	ND	0.0050	0.0010	ug/l	
2991-50-6	EtFOSAA	ND	0.0050	0.0013	ug/l	
24448-09-7	MeFOSE	ND	0.040	0.0044	ug/l	
1691-99-2	EtFOSE	ND	0.040	0.0074	ug/l	
13252-13-6	HFPO-DA (GenX)	ND	0.0040	0.0010	ug/l	
919005-14-4	ADONA	ND	0.0080	0.0019	ug/l	
377-73-1	PFMPA	ND	0.0080	0.0010	ug/l	
863090-89-5	PFMBA	ND	0.0080	0.0011	ug/l	
151772-58-6	NFDHA	ND	0.0080	0.0012	ug/l	
756426-58-19	Cl-PF3ONS (F-53B Major)	ND	0.0080	0.0014	ug/l	
763051-92-91	Cl-PF3OUdS (F-53B Minor)	ND	0.0080	0.0018	ug/l	

# Instrument Blank

Job Number: FC8392  
 Account: AECOMCOD AECOM, INC.  
 Project: N6274223F0104 RH Fire Suppression System

Sample	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
S4Q713-IBLK	4Q48763.D	1	08/09/23	AL	n/a	n/a	S4Q713

The QC reported here applies to the following samples:

Method: EPA DRAFT 1633

FC8392-1

CAS No.	Compound	Result	RL	MDL	Units	Q
113507-82-7	PFEESA	ND	0.0080	0.00078	ug/l	
356-02-5	3:3 Fluorotelomer carboxylate	ND	0.020	0.0045	ug/l	
914637-49-35:3	Fluorotelomer carboxylate	ND	0.10	0.0087	ug/l	
812-70-4	7:3 Fluorotelomer carboxylate	ND	0.10	0.0079	ug/l	

CAS No.	ID Standard Recoveries	Limits
	13C4-PFBA	106% 20-150%
	13C5-PFPeA	95% 20-150%
	13C5-PFHxA	100% 20-150%
	13C4-PFHpA	101% 20-150%
	13C8-PFOA	99% 20-150%
	13C9-PFNA	96% 20-150%
	13C6-PFDA	102% 20-150%
	13C7-PFUnDA	111% 20-150%
	13C2-PFDoDA	101% 20-150%
	13C2-PFTeDA	100% 20-150%
	13C3-PFBS	90% 20-150%
	13C3-PFHxS	94% 20-150%
	13C8-PFOS	91% 20-150%
	13C8-FOSA	92% 20-150%
	d3-MeFOSAA	105% 20-150%
	d5-EtFOSAA	103% 20-150%
	13C2-4:2FTS	148% 20-180%
	13C2-6:2FTS	130% 20-180%
	13C2-8:2FTS	138% 20-180%

6.1.1

6



# Continuing Calibration Blank

Job Number: FC8392  
 Account: AECOMCOD AECOM, INC.  
 Project: N6274223F0104 RH Fire Suppression System

Sample	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
S4Q713-ICCB	4Q48777.D	1	08/09/23	AL	n/a	n/a	S4Q713

The QC reported here applies to the following samples:

Method: EPA DRAFT 1633

FC8392-1

CAS No.	Compound	Result	RL	MDL	Units	Q
375-22-4	Perfluorobutanoic acid	ND	0.016	0.0019	ug/l	
2706-90-3	Perfluoropentanoic acid	ND	0.0080	0.00094	ug/l	
307-24-4	Perfluorohexanoic acid	ND	0.0040	0.00050	ug/l	
375-85-9	Perfluoroheptanoic acid	ND	0.0040	0.00050	ug/l	
335-67-1	Perfluorooctanoic acid	ND	0.0040	0.00050	ug/l	
375-95-1	Perfluorononanoic acid	ND	0.0040	0.00061	ug/l	
335-76-2	Perfluorodecanoic acid	ND	0.0040	0.00050	ug/l	
2058-94-8	Perfluoroundecanoic acid	ND	0.0040	0.00060	ug/l	
307-55-1	Perfluorododecanoic acid	ND	0.0040	0.00060	ug/l	
72629-94-8	Perfluorotridecanoic acid	ND	0.0040	0.00084	ug/l	
376-06-7	Perfluorotetradecanoic acid	ND	0.0040	0.00050	ug/l	
375-73-5	Perfluorobutanesulfonic acid	ND	0.0040	0.00050	ug/l	
2706-91-4	Perfluoropentanesulfonic acid	ND	0.0050	0.0011	ug/l	
355-46-4	Perfluorohexanesulfonic acid	ND	0.0040	0.00070	ug/l	
375-92-8	Perfluoroheptanesulfonic acid	ND	0.0040	0.00050	ug/l	
1763-23-1	Perfluorooctanesulfonic acid	ND	0.0040	0.00054	ug/l	
68259-12-1	Perfluorononanesulfonic acid	ND	0.0040	0.00057	ug/l	
335-77-3	Perfluorodecanesulfonic acid	ND	0.0040	0.00064	ug/l	
79780-39-5	Perfluorododecanesulfonic aci	ND	0.0050	0.0011	ug/l	
757124-72-44:2	Fluorotelomer sulfonate	ND	0.020	0.0032	ug/l	
27619-97-2	6:2 Fluorotelomer sulfonate	ND	0.020	0.0035	ug/l	
39108-34-4	8:2 Fluorotelomer sulfonate	ND	0.020	0.0041	ug/l	
754-91-6	PFOSA	ND	0.0040	0.00067	ug/l	
31506-32-8	MeFOSA	ND	0.0080	0.0010	ug/l	
4151-50-2	EtFOSA	ND	0.0080	0.0010	ug/l	
2355-31-9	MeFOSAA	ND	0.0050	0.0010	ug/l	
2991-50-6	EtFOSAA	ND	0.0050	0.0013	ug/l	
24448-09-7	MeFOSE	ND	0.040	0.0044	ug/l	
1691-99-2	EtFOSE	ND	0.040	0.0074	ug/l	
13252-13-6	HFPO-DA (GenX)	ND	0.0040	0.0010	ug/l	
919005-14-4	ADONA	ND	0.0080	0.0019	ug/l	
377-73-1	PFMPA	ND	0.0080	0.0010	ug/l	
863090-89-5	PFMBA	ND	0.0080	0.0011	ug/l	
151772-58-6	NFDHA	ND	0.0080	0.0012	ug/l	
756426-58-19	Cl-PF3ONS (F-53B Major)	ND	0.0080	0.0014	ug/l	
763051-92-91	Cl-PF3OUdS (F-53B Minor)	ND	0.0080	0.0018	ug/l	

# Continuing Calibration Blank

Job Number: FC8392  
 Account: AECOMCOD AECOM, INC.  
 Project: N6274223F0104 RH Fire Suppression System

Sample	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
S4Q713-ICCB	4Q48777.D	1	08/09/23	AL	n/a	n/a	S4Q713

The QC reported here applies to the following samples:

Method: EPA DRAFT 1633

FC8392-1

CAS No.	Compound	Result	RL	MDL	Units	Q
113507-82-7	PFEESA	ND	0.0080	0.00078	ug/l	
356-02-5	3:3 Fluorotelomer carboxylate	ND	0.020	0.0045	ug/l	
914637-49-35:3	Fluorotelomer carboxylate	ND	0.10	0.0087	ug/l	
812-70-4	7:3 Fluorotelomer carboxylate	ND	0.10	0.0079	ug/l	

CAS No.	ID Standard Recoveries	Limits
	13C4-PFBA	110% 20-150%
	13C5-PFPeA	89% 20-150%
	13C5-PFHxA	102% 20-150%
	13C4-PFHpA	103% 20-150%
	13C8-PFOA	101% 20-150%
	13C9-PFNA	93% 20-150%
	13C6-PFDA	107% 20-150%
	13C7-PFUnDA	110% 20-150%
	13C2-PFDoDA	108% 20-150%
	13C2-PFTeDA	103% 20-150%
	13C3-PFBS	90% 20-150%
	13C3-PFHxS	97% 20-150%
	13C8-PFOS	93% 20-150%
	13C8-FOSA	93% 20-150%
	d3-MeFOSAA	106% 20-150%
	d5-EtFOSAA	108% 20-150%
	13C2-4:2FTS	125% 20-180%
	13C2-6:2FTS	135% 20-180%
	13C2-8:2FTS	133% 20-180%

# Method Blank Summary

Job Number: FC8392  
 Account: AECOMCOD AECOM, INC.  
 Project: N6274223F0104 RH Fire Suppression System

Sample	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
OP98297-MB <sup>a</sup>	4Q48781.D	1	08/09/23	AL	08/08/23	OP98297	S4Q713

The QC reported here applies to the following samples:

Method: EPA DRAFT 1633

FC8392-1

CAS No.	Compound	Result	RL	MDL	Units	Q
375-22-4	Perfluorobutanoic acid	ND	0.016	0.0019	ug/l	
2706-90-3	Perfluoropentanoic acid	ND	0.0080	0.00094	ug/l	
307-24-4	Perfluorohexanoic acid	ND	0.0040	0.00050	ug/l	
375-85-9	Perfluoroheptanoic acid	ND	0.0040	0.00050	ug/l	
335-67-1	Perfluorooctanoic acid	ND	0.0040	0.00050	ug/l	
375-95-1	Perfluorononanoic acid	ND	0.0040	0.00061	ug/l	
335-76-2	Perfluorodecanoic acid	ND	0.0040	0.00050	ug/l	
2058-94-8	Perfluoroundecanoic acid	ND	0.0040	0.00060	ug/l	
307-55-1	Perfluorododecanoic acid	ND	0.0040	0.00060	ug/l	
72629-94-8	Perfluorotridecanoic acid	ND	0.0040	0.00084	ug/l	
376-06-7	Perfluorotetradecanoic acid	ND	0.0040	0.00050	ug/l	
375-73-5	Perfluorobutanesulfonic acid	ND	0.0040	0.00050	ug/l	
2706-91-4	Perfluoropentanesulfonic acid	ND	0.0050	0.0011	ug/l	
355-46-4	Perfluorohexanesulfonic acid	ND	0.0040	0.00070	ug/l	
375-92-8	Perfluoroheptanesulfonic acid	ND	0.0040	0.00050	ug/l	
1763-23-1	Perfluorooctanesulfonic acid	ND	0.0040	0.00054	ug/l	
68259-12-1	Perfluorononanesulfonic acid	ND	0.0040	0.00057	ug/l	
335-77-3	Perfluorodecanesulfonic acid	ND	0.0040	0.00064	ug/l	
79780-39-5	Perfluorododecanesulfonic aci	ND	0.0050	0.0011	ug/l	
757124-72-44:2	Fluorotelomer sulfonate	ND	0.020	0.0032	ug/l	
27619-97-2	6:2 Fluorotelomer sulfonate	ND	0.020	0.0035	ug/l	
39108-34-4	8:2 Fluorotelomer sulfonate	ND	0.020	0.0041	ug/l	
754-91-6	PFOSA	ND	0.0040	0.00067	ug/l	
31506-32-8	MeFOSA	ND	0.0080	0.0010	ug/l	
4151-50-2	EtFOSA	ND	0.0080	0.0010	ug/l	
2355-31-9	MeFOSAA	ND	0.0050	0.0010	ug/l	
2991-50-6	EtFOSAA	ND	0.0050	0.0013	ug/l	
24448-09-7	MeFOSE	ND	0.040	0.0044	ug/l	
1691-99-2	EtFOSE	ND	0.040	0.0074	ug/l	
13252-13-6	HFPO-DA (GenX)	ND	0.0040	0.0010	ug/l	
919005-14-4	ADONA	ND	0.0080	0.0019	ug/l	
377-73-1	PFMPA	ND	0.0080	0.0010	ug/l	
863090-89-5	PFMBA	ND	0.0080	0.0011	ug/l	
151772-58-6	NFDHA	ND	0.0080	0.0012	ug/l	
756426-58-19	Cl-PF3ONS (F-53B Major)	ND	0.0080	0.0014	ug/l	
763051-92-91	Cl-PF3OUdS (F-53B Minor)	ND	0.0080	0.0018	ug/l	

# Method Blank Summary

Job Number: FC8392  
 Account: AECOMCOD AECOM, INC.  
 Project: N6274223F0104 RH Fire Suppression System

Sample	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
OP98297-MB <sup>a</sup>	4Q48781.D	1	08/09/23	AL	08/08/23	OP98297	S4Q713

The QC reported here applies to the following samples:

Method: EPA DRAFT 1633

FC8392-1

CAS No.	Compound	Result	RL	MDL	Units	Q
113507-82-7	PFEESA	ND	0.0080	0.00078	ug/l	
356-02-5	3:3 Fluorotelomer carboxylate	ND	0.020	0.0045	ug/l	
914637-49-35:3	Fluorotelomer carboxylate	ND	0.10	0.0087	ug/l	
812-70-4	7:3 Fluorotelomer carboxylate	ND	0.10	0.0079	ug/l	

CAS No.	ID Standard Recoveries	Limits
	13C4-PFBA	85% 20-150%
	13C5-PFPeA	74% 20-150%
	13C5-PFHxA	81% 20-150%
	13C4-PFHpA	84% 20-150%
	13C8-PFOA	77% 20-150%
	13C9-PFNA	72% 20-150%
	13C6-PFDA	79% 20-150%
	13C7-PFUnDA	80% 20-150%
	13C2-PFDoDA	77% 20-150%
	13C2-PFTeDA	71% 20-150%
	13C3-PFBS	76% 20-150%
	13C3-PFHxS	79% 20-150%
	13C8-PFOS	81% 20-150%
	13C8-FOSA	44% 20-150%
	d3-MeFOSA	49% 20-150%
	d5-EtFOSA	51% 20-150%
	d3-MeFOSAA	77% 20-150%
	d5-EtFOSAA	82% 20-150%
	d7-MeFOSE	37% 20-150%
	d9-EtFOSE	45% 20-150%
	13C2-4:2FTS	112% 20-180%
	13C2-6:2FTS	111% 20-180%
	13C2-8:2FTS	104% 20-180%
	13C3-HFPO-DA	73% 20-150%

(a) All results and recoveries corrected for double EIS.

6.1.3  
6

# Continuing Calibration Blank

Job Number: FC8392  
 Account: AECOMCOD AECOM, INC.  
 Project: N6274223F0104 RH Fire Suppression System

Sample	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
S4Q713-ICCB	4Q48788.D	1	08/09/23	AL	n/a	n/a	S4Q713

The QC reported here applies to the following samples:

Method: EPA DRAFT 1633

OP98297-DUP, OP98297-MS

CAS No.	Compound	Result	RL	MDL	Units	Q
375-22-4	Perfluorobutanoic acid	ND	0.016	0.0019	ug/l	
2706-90-3	Perfluoropentanoic acid	ND	0.0080	0.00094	ug/l	
307-24-4	Perfluorohexanoic acid	ND	0.0040	0.00050	ug/l	
375-85-9	Perfluoroheptanoic acid	ND	0.0040	0.00050	ug/l	
335-67-1	Perfluorooctanoic acid	ND	0.0040	0.00050	ug/l	
375-95-1	Perfluorononanoic acid	ND	0.0040	0.00061	ug/l	
335-76-2	Perfluorodecanoic acid	ND	0.0040	0.00050	ug/l	
2058-94-8	Perfluoroundecanoic acid	ND	0.0040	0.00060	ug/l	
307-55-1	Perfluorododecanoic acid	ND	0.0040	0.00060	ug/l	
72629-94-8	Perfluorotridecanoic acid	ND	0.0040	0.00084	ug/l	
376-06-7	Perfluorotetradecanoic acid	ND	0.0040	0.00050	ug/l	
375-73-5	Perfluorobutanesulfonic acid	ND	0.0040	0.00050	ug/l	
2706-91-4	Perfluoropentanesulfonic acid	ND	0.0050	0.0011	ug/l	
355-46-4	Perfluorohexanesulfonic acid	ND	0.0040	0.00070	ug/l	
375-92-8	Perfluoroheptanesulfonic acid	ND	0.0040	0.00050	ug/l	
1763-23-1	Perfluorooctanesulfonic acid	ND	0.0040	0.00054	ug/l	
68259-12-1	Perfluorononanesulfonic acid	ND	0.0040	0.00057	ug/l	
335-77-3	Perfluorodecanesulfonic acid	ND	0.0040	0.00064	ug/l	
79780-39-5	Perfluorododecanesulfonic aci	ND	0.0050	0.0011	ug/l	
757124-72-44:2	Fluorotelomer sulfonate	ND	0.020	0.0032	ug/l	
27619-97-2	6:2 Fluorotelomer sulfonate	ND	0.020	0.0035	ug/l	
39108-34-4	8:2 Fluorotelomer sulfonate	ND	0.020	0.0041	ug/l	
754-91-6	PFOSA	ND	0.0040	0.00067	ug/l	
31506-32-8	MeFOSA	ND	0.0080	0.0010	ug/l	
4151-50-2	EtFOSA	ND	0.0080	0.0010	ug/l	
2355-31-9	MeFOSAA	ND	0.0050	0.0010	ug/l	
2991-50-6	EtFOSAA	ND	0.0050	0.0013	ug/l	
24448-09-7	MeFOSE	ND	0.040	0.0044	ug/l	
1691-99-2	EtFOSE	ND	0.040	0.0074	ug/l	
13252-13-6	HFPO-DA (GenX)	ND	0.0040	0.0010	ug/l	
919005-14-4	ADONA	ND	0.0080	0.0019	ug/l	
377-73-1	PFMPA	ND	0.0080	0.0010	ug/l	
863090-89-5	PFMBA	ND	0.0080	0.0011	ug/l	
151772-58-6	NFDHA	ND	0.0080	0.0012	ug/l	
756426-58-19	Cl-PF3ONS (F-53B Major)	ND	0.0080	0.0014	ug/l	
763051-92-91	Cl-PF3OUdS (F-53B Minor)	ND	0.0080	0.0018	ug/l	

# Continuing Calibration Blank

Job Number: FC8392  
 Account: AECOMCOD AECOM, INC.  
 Project: N6274223F0104 RH Fire Suppression System

Sample	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
S4Q713-ICCB	4Q48788.D	1	08/09/23	AL	n/a	n/a	S4Q713

The QC reported here applies to the following samples:

Method: EPA DRAFT 1633

OP98297-DUP, OP98297-MS

CAS No.	Compound	Result	RL	MDL	Units	Q
113507-82-7	PFEESA	ND	0.0080	0.00078	ug/l	
356-02-5	3:3 Fluorotelomer carboxylate	ND	0.020	0.0045	ug/l	
914637-49-35:3	Fluorotelomer carboxylate	ND	0.10	0.0087	ug/l	
812-70-4	7:3 Fluorotelomer carboxylate	ND	0.10	0.0079	ug/l	

CAS No.	ID Standard Recoveries	Limits
	13C4-PFBA	111% 20-150%
	13C5-PFPeA	88% 20-150%
	13C5-PFHxA	99% 20-150%
	13C4-PFHpA	103% 20-150%
	13C8-PFOA	102% 20-150%
	13C9-PFNA	84% 20-150%
	13C6-PFDA	101% 20-150%
	13C7-PFUnDA	116% 20-150%
	13C2-PFDoDA	106% 20-150%
	13C2-PFTeDA	99% 20-150%
	13C3-PFBS	91% 20-150%
	13C3-PFHxS	99% 20-150%
	13C8-PFOS	93% 20-150%
	13C8-FOSA	92% 20-150%
	d3-MeFOSAA	104% 20-150%
	d5-EtFOSAA	104% 20-150%
	13C2-4:2FTS	141% 20-180%
	13C2-6:2FTS	119% 20-180%
	13C2-8:2FTS	139% 20-180%

# Continuing Calibration Blank

Job Number: FC8392  
 Account: AECOMCOD AECOM, INC.  
 Project: N6274223F0104 RH Fire Suppression System

Sample	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
S4Q713-ICCB	4Q48800.D	1	08/10/23	AL	n/a	n/a	S4Q713

The QC reported here applies to the following samples:

Method: EPA DRAFT 1633

OP98297-DUP, OP98297-MS

CAS No.	Compound	Result	RL	MDL	Units	Q
375-22-4	Perfluorobutanoic acid	ND	0.016	0.0019	ug/l	
2706-90-3	Perfluoropentanoic acid	ND	0.0080	0.00094	ug/l	
307-24-4	Perfluorohexanoic acid	ND	0.0040	0.00050	ug/l	
375-85-9	Perfluoroheptanoic acid	ND	0.0040	0.00050	ug/l	
335-67-1	Perfluorooctanoic acid	ND	0.0040	0.00050	ug/l	
375-95-1	Perfluorononanoic acid	ND	0.0040	0.00061	ug/l	
335-76-2	Perfluorodecanoic acid	ND	0.0040	0.00050	ug/l	
2058-94-8	Perfluoroundecanoic acid	ND	0.0040	0.00060	ug/l	
307-55-1	Perfluorododecanoic acid	ND	0.0040	0.00060	ug/l	
72629-94-8	Perfluorotridecanoic acid	ND	0.0040	0.00084	ug/l	
376-06-7	Perfluorotetradecanoic acid	ND	0.0040	0.00050	ug/l	
375-73-5	Perfluorobutanesulfonic acid	ND	0.0040	0.00050	ug/l	
2706-91-4	Perfluoropentanesulfonic acid	ND	0.0050	0.0011	ug/l	
355-46-4	Perfluorohexanesulfonic acid	ND	0.0040	0.00070	ug/l	
375-92-8	Perfluoroheptanesulfonic acid	ND	0.0040	0.00050	ug/l	
1763-23-1	Perfluorooctanesulfonic acid	ND	0.0040	0.00054	ug/l	
68259-12-1	Perfluorononanesulfonic acid	ND	0.0040	0.00057	ug/l	
335-77-3	Perfluorodecanesulfonic acid	ND	0.0040	0.00064	ug/l	
79780-39-5	Perfluorododecanesulfonic aci	ND	0.0050	0.0011	ug/l	
757124-72-44:2	Fluorotelomer sulfonate	ND	0.020	0.0032	ug/l	
27619-97-2	6:2 Fluorotelomer sulfonate	ND	0.020	0.0035	ug/l	
39108-34-4	8:2 Fluorotelomer sulfonate	ND	0.020	0.0041	ug/l	
754-91-6	PFOSA	ND	0.0040	0.00067	ug/l	
31506-32-8	MeFOSA	ND	0.0080	0.0010	ug/l	
4151-50-2	EtFOSA	ND	0.0080	0.0010	ug/l	
2355-31-9	MeFOSAA	ND	0.0050	0.0010	ug/l	
2991-50-6	EtFOSAA	ND	0.0050	0.0013	ug/l	
24448-09-7	MeFOSE	ND	0.040	0.0044	ug/l	
1691-99-2	EtFOSE	ND	0.040	0.0074	ug/l	
13252-13-6	HFPO-DA (GenX)	ND	0.0040	0.0010	ug/l	
919005-14-4	ADONA	ND	0.0080	0.0019	ug/l	
377-73-1	PFMPA	ND	0.0080	0.0010	ug/l	
863090-89-5	PFMBA	ND	0.0080	0.0011	ug/l	
151772-58-6	NFDHA	ND	0.0080	0.0012	ug/l	
756426-58-19	Cl-PF3ONS (F-53B Major)	ND	0.0080	0.0014	ug/l	
763051-92-91	Cl-PF3OUdS (F-53B Minor)	ND	0.0080	0.0018	ug/l	

# Continuing Calibration Blank

Job Number: FC8392  
 Account: AECOMCOD AECOM, INC.  
 Project: N6274223F0104 RH Fire Suppression System

Sample	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
S4Q713-ICCB	4Q48800.D	1	08/10/23	AL	n/a	n/a	S4Q713

The QC reported here applies to the following samples:

Method: EPA DRAFT 1633

OP98297-DUP, OP98297-MS

CAS No.	Compound	Result	RL	MDL	Units	Q
113507-82-7	PFEESA	ND	0.0080	0.00078	ug/l	
356-02-5	3:3 Fluorotelomer carboxylate	ND	0.020	0.0045	ug/l	
914637-49-35:3	Fluorotelomer carboxylate	ND	0.10	0.0087	ug/l	
812-70-4	7:3 Fluorotelomer carboxylate	ND	0.10	0.0079	ug/l	

CAS No.	ID Standard Recoveries	Limits
	13C4-PFBA	111% 20-150%
	13C5-PFPeA	88% 20-150%
	13C5-PFHxA	97% 20-150%
	13C4-PFHpA	99% 20-150%
	13C8-PFOA	100% 20-150%
	13C9-PFNA	93% 20-150%
	13C6-PFDA	100% 20-150%
	13C7-PFUnDA	105% 20-150%
	13C2-PFDoDA	103% 20-150%
	13C2-PFTeDA	94% 20-150%
	13C3-PFBS	90% 20-150%
	13C3-PFHxS	102% 20-150%
	13C8-PFOS	91% 20-150%
	13C8-FOSA	89% 20-150%
	d3-MeFOSAA	96% 20-150%
	d5-EtFOSAA	100% 20-150%
	13C2-4:2FTS	131% 20-180%
	13C2-6:2FTS	137% 20-180%
	13C2-8:2FTS	142% 20-180%



# Blank Spike Summary

Job Number: FC8392  
 Account: AECOMCOD AECOM, INC.  
 Project: N6274223F0104 RH Fire Suppression System

Sample	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
OP98297-LLBS	4Q48780.D	1	08/09/23	AL	08/08/23	OP98297	S4Q713

The QC reported here applies to the following samples:

Method: EPA DRAFT 1633

FC8392-1

CAS No.	Compound	Spike ug/l	BSP ug/l	BSP %	Limits
375-22-4	Perfluorobutanoic acid	0.03	0.0319	106	40-150
2706-90-3	Perfluoropentanoic acid	0.015	0.0180	120	40-150
307-24-4	Perfluorohexanoic acid	0.0075	0.0090	120	40-150
375-85-9	Perfluoroheptanoic acid	0.0075	0.0078	104	40-150
335-67-1	Perfluorooctanoic acid	0.0075	0.0090	120	40-150
375-95-1	Perfluorononanoic acid	0.0075	0.0088	117	40-150
335-76-2	Perfluorodecanoic acid	0.0075	0.0084	112	40-150
2058-94-8	Perfluoroundecanoic acid	0.0075	0.0079	105	40-150
307-55-1	Perfluorododecanoic acid	0.0075	0.0080	107	40-150
72629-94-8	Perfluorotridecanoic acid	0.0075	0.0089	119	40-150
376-06-7	Perfluorotetradecanoic acid	0.0075	0.0084	112	40-150
375-73-5	Perfluorobutanesulfonic acid	0.00665	0.0084	126	40-150
2706-91-4	Perfluoropentanesulfonic acid	0.00706	0.0079	112	40-150
355-46-4	Perfluorohexanesulfonic acid	0.00686	0.0082	120	40-150
375-92-8	Perfluoroheptanesulfonic acid	0.00715	0.0072	101	40-150
1763-23-1	Perfluorooctanesulfonic acid	0.00696	0.0090	129	40-150
68259-12-1	Perfluorononanesulfonic acid	0.00722	0.0082	114	40-150
335-77-3	Perfluorodecanesulfonic acid	0.00724	0.0077	106	40-150
79780-39-5	Perfluorododecanesulfonic aci	0.00728	0.0079	109	40-150
757124-72-44:2	Fluorotelomer sulfonate	0.0281	0.0321	114	40-150
27619-97-2	6:2 Fluorotelomer sulfonate	0.0285	0.0316	111	40-150
39108-34-4	8:2 Fluorotelomer sulfonate	0.0288	0.0284	99	40-150
754-91-6	PFOSA	0.0075	0.0083	111	40-150
31506-32-8	MeFOSA	0.015	0.0138	92	40-150
4151-50-2	EtFOSA	0.015	0.0158	105	40-150
2355-31-9	MeFOSAA	0.0075	0.0090	120	40-150
2991-50-6	EtFOSAA	0.0075	0.0079	105	40-150
24448-09-7	MeFOSE	0.0375	0.0381	102	40-150
1691-99-2	EtFOSE	0.0375	0.0363	97	40-150
13252-13-6	HFPO-DA (GenX)	0.015	0.0166	111	40-150
919005-14-4	ADONA	0.0142	0.0174	123	40-150
377-73-1	PFMPA	0.015	0.0175	117	40-150
863090-89-5	PFMBA	0.015	0.0177	118	40-150
151772-58-6	NFDHA	0.015	0.0154	103	40-150
756426-58-19	Cl-PF3ONS (F-53B Major)	0.014	0.0175	125	40-150
763051-92-91	Cl-PF3OUdS (F-53B Minor)	0.0142	0.0157	111	40-150

\* = Outside of Control Limits.

# Blank Spike Summary

Job Number: FC8392  
 Account: AECOMCOD AECOM, INC.  
 Project: N6274223F0104 RH Fire Suppression System

Sample	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
OP98297-LLBS	4Q48780.D	1	08/09/23	AL	08/08/23	OP98297	S4Q713

The QC reported here applies to the following samples:

Method: EPA DRAFT 1633

FC8392-1

CAS No.	Compound	Spike ug/l	BSP ug/l	BSP %	Limits
113507-82-7	PFEESA	0.0134	0.0141	106	40-150
356-02-5	3:3 Fluorotelomer carboxylate	0.0375	0.0248	66	40-150
914637-49-35:3	Fluorotelomer carboxylate	0.188	0.172	92	40-150
812-70-4	7:3 Fluorotelomer carboxylate	0.188	0.192	102	40-150

CAS No.	ID Standard Recoveries	BSP	Limits
	13C4-PFBA	142%	20-150%
	13C5-PFPeA	114%	20-150%
	13C5-PFHxA	132%	20-150%
	13C4-PFHpA	137%	20-150%
	13C8-PFOA	127%	20-150%
	13C9-PFNA	115%	20-150%
	13C6-PFDA	127%	20-150%
	13C7-PFUnDA	133%	20-150%
	13C2-PFDoDA	118%	20-150%
	13C2-PFTeDA	106%	20-150%
	13C3-PFBS	121%	20-150%
	13C3-PFHxS	127%	20-150%
	13C8-PFOS	111%	20-150%
	13C8-FOSA	60%	20-150%
	d3-MeFOSA	73%	20-150%
	d5-EtFOSA	74%	20-150%
	d3-MeFOSAA	118%	20-150%
	d5-EtFOSAA	120%	20-150%
	d7-MeFOSE	58%	20-150%
	d9-EtFOSE	70%	20-150%
	13C2-4:2FTS	168%	20-180%
	13C2-6:2FTS	181%* a	20-180%
	13C2-8:2FTS	200%* a	20-180%
	13C3-HFPO-DA	114%	20-150%

(a) Outside control limits.

\* = Outside of Control Limits.

# Blank Spike Summary

Job Number: FC8392  
 Account: AECOMCOD AECOM, INC.  
 Project: N6274223F0104 RH Fire Suppression System

Sample	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
OP98297-BS	4Q48779.D	1	08/09/23	AL	08/08/23	OP98297	S4Q713

The QC reported here applies to the following samples:

Method: EPA DRAFT 1633

FC8392-1

CAS No.	Compound	Spike ug/l	BSP ug/l	BSP %	Limits
375-22-4	Perfluorobutanoic acid	0.1	0.111	111	40-150
2706-90-3	Perfluoropentanoic acid	0.05	0.0643	129	40-150
307-24-4	Perfluorohexanoic acid	0.025	0.0284	114	40-150
375-85-9	Perfluoroheptanoic acid	0.025	0.0282	113	40-150
335-67-1	Perfluorooctanoic acid	0.025	0.0300	120	40-150
375-95-1	Perfluorononanoic acid	0.025	0.0319	128	40-150
335-76-2	Perfluorodecanoic acid	0.025	0.0318	127	40-150
2058-94-8	Perfluoroundecanoic acid	0.025	0.0292	117	40-150
307-55-1	Perfluorododecanoic acid	0.025	0.0274	110	40-150
72629-94-8	Perfluorotridecanoic acid	0.025	0.0304	122	40-150
376-06-7	Perfluorotetradecanoic acid	0.025	0.0302	121	40-150
375-73-5	Perfluorobutanesulfonic acid	0.0222	0.0315	142	40-150
2706-91-4	Perfluoropentanesulfonic acid	0.0235	0.0257	109	40-150
355-46-4	Perfluorohexanesulfonic acid	0.0229	0.0299	131	40-150
375-92-8	Perfluoroheptanesulfonic acid	0.0238	0.0272	114	40-150
1763-23-1	Perfluorooctanesulfonic acid	0.0232	0.0299	129	40-150
68259-12-1	Perfluorononanesulfonic acid	0.0241	0.0308	128	40-150
335-77-3	Perfluorodecanesulfonic acid	0.0241	0.0274	114	40-150
79780-39-5	Perfluorododecanesulfonic aci	0.0243	0.0280	115	40-150
757124-72-44:2	Fluorotelomer sulfonate	0.0938	0.108	115	40-150
27619-97-2	6:2 Fluorotelomer sulfonate	0.095	0.108	114	40-150
39108-34-4	8:2 Fluorotelomer sulfonate	0.096	0.119	124	40-150
754-91-6	PFOSA	0.025	0.0293	117	40-150
31506-32-8	MeFOSA	0.05	0.0533	107	40-150
4151-50-2	EtFOSA	0.05	0.0562	112	40-150
2355-31-9	MeFOSAA	0.025	0.0310	124	40-150
2991-50-6	EtFOSAA	0.025	0.0360	144	40-150
24448-09-7	MeFOSE	0.125	0.140	112	40-150
1691-99-2	EtFOSE	0.125	0.138	110	40-150
13252-13-6	HFPO-DA (GenX)	0.05	0.0545	109	40-150
919005-14-4	ADONA	0.0473	0.0604	128	40-150
377-73-1	PFMPA	0.05	0.0621	124	40-150
863090-89-5	PFMBA	0.05	0.0632	126	40-150
151772-58-6	NFDHA	0.05	0.0577	115	40-150
756426-58-19	Cl-PF3ONS (F-53B Major)	0.0468	0.0629	135	40-150
763051-92-91	Cl-PF3OUdS (F-53B Minor)	0.0473	0.0608	129	40-150

\* = Outside of Control Limits.

# Blank Spike Summary

Job Number: FC8392  
 Account: AECOMCOD AECOM, INC.  
 Project: N6274223F0104 RH Fire Suppression System

Sample	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
OP98297-BS	4Q48779.D	1	08/09/23	AL	08/08/23	OP98297	S4Q713

The QC reported here applies to the following samples:

Method: EPA DRAFT 1633

FC8392-1

CAS No.	Compound	Spike ug/l	BSP ug/l	BSP %	Limits
113507-82-7	PFEESA	0.0445	0.0489	110	40-150
356-02-5	3:3 Fluorotelomer carboxylate	0.125	0.0953	76	40-150
914637-49-35:3	Fluorotelomer carboxylate	0.625	0.610	98	40-150
812-70-4	7:3 Fluorotelomer carboxylate	0.625	0.672	108	40-150

CAS No.	ID Standard Recoveries	BSP	Limits
	13C4-PFBA	140%	20-150%
	13C5-PFPeA	111%	20-150%
	13C5-PFHxA	130%	20-150%
	13C4-PFHpA	135%	20-150%
	13C8-PFOA	127%	20-150%
	13C9-PFNA	108%	20-150%
	13C6-PFDA	130%	20-150%
	13C7-PFUnDA	137%	20-150%
	13C2-PFDoDA	135%	20-150%
	13C2-PFTeDA	116%	20-150%
	13C3-PFBS	114%	20-150%
	13C3-PFHxS	125%	20-150%
	13C8-PFOS	112%	20-150%
	13C8-FOSA	68%	20-150%
	d3-MeFOSA	74%	20-150%
	d5-EtFOSA	74%	20-150%
	d3-MeFOSAA	123%	20-150%
	d5-EtFOSAA	101%	20-150%
	d7-MeFOSE	58%	20-150%
	d9-EtFOSE	69%	20-150%
	13C2-4:2FTS	159%	20-180%
	13C2-6:2FTS	168%	20-180%
	13C2-8:2FTS	173%	20-180%
	13C3-HFPO-DA	114%	20-150%

\* = Outside of Control Limits.

# Matrix Spike Summary

Job Number: FC8392  
 Account: AECOMCOD AECOM, INC.  
 Project: N6274223F0104 RH Fire Suppression System

Sample	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
OP98297-MS	4Q48790.D	1	08/09/23	AL	08/08/23	OP98297	S4Q713
FC8439-3	4Q48789.D	1	08/09/23	AL	08/08/23	OP98297	S4Q713

The QC reported here applies to the following samples:

Method: EPA DRAFT 1633

FC8392-1

CAS No.	Compound	FC8439-3 ug/l	Spike Q	MS ug/l	MS %	Limits
375-22-4	Perfluorobutanoic acid	0.015 U	0.0909	0.106	117	40-150
2706-90-3	Perfluoropentanoic acid	0.0074 U	0.0455	0.0603	133	40-150
307-24-4	Perfluorohexanoic acid	0.0037 U	0.0227	0.0260	114	40-150
375-85-9	Perfluoroheptanoic acid	0.0037 U	0.0227	0.0268	118	40-150
335-67-1	Perfluorooctanoic acid	0.0037 U	0.0227	0.0290	128	40-150
375-95-1	Perfluorononanoic acid	0.0037 U	0.0227	0.0279	123	40-150
335-76-2	Perfluorodecanoic acid	0.0037 U	0.0227	0.0292	128	40-150
2058-94-8	Perfluoroundecanoic acid	0.0037 U	0.0227	0.0270	119	40-150
307-55-1	Perfluorododecanoic acid	0.0037 U	0.0227	0.0269	118	40-150
72629-94-8	Perfluorotridecanoic acid	0.0037 U	0.0227	0.0276	121	40-150
376-06-7	Perfluorotetradecanoic acid	0.0037 U	0.0227	0.0284	125	40-150
375-73-5	Perfluorobutanesulfonic acid	0.0037 U	0.0202	0.0295	146	40-150
2706-91-4	Perfluoropentanesulfonic acid	0.0046 U	0.0214	0.0260	122	40-150
355-46-4	Perfluorohexanesulfonic acid	0.0037 U	0.0208	0.0282	136	40-150
375-92-8	Perfluoroheptanesulfonic acid	0.0037 U	0.0217	0.0250	115	40-150
1763-23-1	Perfluorooctanesulfonic acid	0.0037 U	0.0211	0.0252	119	40-150
68259-12-1	Perfluorononanesulfonic acid	0.0037 U	0.0219	0.0282	129	40-150
335-77-3	Perfluorodecanesulfonic acid	0.0037 U	0.0219	0.0247	113	40-150
79780-39-5	Perfluorododecanesulfonic aci	0.0046 U	0.022	0.0250	113	40-150
757124-72-44:2	Fluorotelomer sulfonate	0.019 U	0.0852	0.0991	116	40-150
27619-97-2	6:2 Fluorotelomer sulfonate	0.019 U	0.0864	0.0981	114	40-150
39108-34-4	8:2 Fluorotelomer sulfonate	0.019 U	0.0873	0.133	152*	40-150
754-91-6	PFOSA	0.0037 U	0.0227	0.0279	123	40-150
31506-32-8	MeFOSA	0.0074 U	0.0455	0.0473	104	40-150
4151-50-2	EtFOSA	0.0074 U	0.0455	0.0545	120	40-150
2355-31-9	MeFOSAA	0.0046 U	0.0227	0.0328	144	40-150
2991-50-6	EtFOSAA	0.0046 U	0.0227	0.0254	112	40-150
24448-09-7	MeFOSE	0.037 U	0.114	0.121	106	40-150
1691-99-2	EtFOSE	0.037 U	0.114	0.126	111	40-150
13252-13-6	HFPO-DA (GenX)	0.0037 U	0.0455	0.0512	113	40-150
919005-14-4	ADONA	0.0074 U	0.043	0.0574	134	40-150
377-73-1	PFMPA	0.0074 U	0.0455	0.0548	121	40-150
863090-89-5	PFMBA	0.0074 U	0.0455	0.0587	129	40-150
151772-58-6	NFDHA	0.0074 U	0.0455	0.0524	115	40-150
756426-58-19	Cl-PF3ONS (F-53B Major)	0.0074 U	0.0425	0.0587	138	40-150
763051-92-91	Cl-PF3OUdS (F-53B Minor)	0.0074 U	0.043	0.0549	128	40-150

\* = Outside of Control Limits.

# Matrix Spike Summary

Job Number: FC8392  
 Account: AECOMCOD AECOM, INC.  
 Project: N6274223F0104 RH Fire Suppression System

Sample	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
OP98297-MS	4Q48790.D	1	08/09/23	AL	08/08/23	OP98297	S4Q713
FC8439-3	4Q48789.D	1	08/09/23	AL	08/08/23	OP98297	S4Q713

The QC reported here applies to the following samples:

Method: EPA DRAFT 1633

FC8392-1

CAS No.	Compound	FC8439-3 ug/l	Spike Q	MS ug/l	MS %	Limits
113507-82-7	PFEESA	0.0074 U	0.0405	0.0449	111	40-150
356-02-5	3:3 Fluorotelomer carboxylate	0.019 U	0.114	0.111	98	40-150
914637-49-35:3	Fluorotelomer carboxylate	0.093 U	0.568	0.593	104	40-150
812-70-4	7:3 Fluorotelomer carboxylate	0.093 U	0.568	0.664	117	40-150

CAS No.	ID Standard Recoveries	MS	FC8439-3	Limits
	13C4-PFBA	107%	108%	20-150%
	13C5-PFPeA	107%	108%	20-150%
	13C5-PFHxA	124%	128%	20-150%
	13C4-PFHpA	126%	129%	20-150%
	13C8-PFOA	120%	120%	20-150%
	13C9-PFNA	101%	102%	20-150%
	13C6-PFDA	127%	122%	20-150%
	13C7-PFUnDA	130%	128%	20-150%
	13C2-PFDoDA	123%	115%	20-150%
	13C2-PFTeDA	103%	98%	20-150%
	13C3-PFBS	120%	108%	20-150%
	13C3-PFHxS	129%	114%	20-150%
	13C8-PFOS	113%	114%	20-150%
	13C8-FOSA	83%	90%	20-150%
	d3-MeFOSA	102%	101%	20-150%
	d5-EtFOSA	91%	91%	20-150%
	d3-MeFOSAA	118%	109%	20-150%
	d5-EtFOSAA	127%	120%	20-150%
	d7-MeFOSE	75%	82%	20-150%
	d9-EtFOSE	85%	87%	20-150%
	13C2-4:2FTS	170%	156%	20-180%
	13C2-6:2FTS	179%	162%	20-180%
	13C2-8:2FTS	155%	165%	20-180%
	13C3-HFPO-DA	107%	106%	20-150%

\* = Outside of Control Limits.

# Duplicate Summary

Job Number: FC8392  
 Account: AECOMCOD AECOM, INC.  
 Project: N6274223F0104 RH Fire Suppression System

Sample	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
OP98297-DUP	4Q48792.D	1	08/09/23	AL	08/08/23	OP98297	S4Q713
FC8439-4	4Q48791.D	1	08/09/23	AL	08/08/23	OP98297	S4Q713
FC8439-4 <sup>a</sup>	6Q22667.D	10	08/10/23	MV	08/08/23	OP98297	S6Q330

The QC reported here applies to the following samples:

Method: EPA DRAFT 1633

FC8392-1

CAS No.	Compound	FC8439-4 ug/l	DUP Q ug/l	Q	RPD	Limits
375-22-4	Perfluorobutanoic acid	0.015 U	ND		nc	30
2706-90-3	Perfluoropentanoic acid	0.0074 U	ND		nc	30
307-24-4	Perfluorohexanoic acid	0.0037 U	ND		nc	30
375-85-9	Perfluoroheptanoic acid	0.0037 U	ND		nc	30
335-67-1	Perfluorooctanoic acid	0.00095 J	0.00056 J		52*	30
375-95-1	Perfluorononanoic acid	0.0037 U	ND		nc	30
335-76-2	Perfluorodecanoic acid	0.0037 U	ND		nc	30
2058-94-8	Perfluoroundecanoic acid	0.0037 U	ND		nc	30
307-55-1	Perfluorododecanoic acid	0.0037 U	ND		nc	30
72629-94-8	Perfluorotridecanoic acid	0.0037 U	ND		nc	30
376-06-7	Perfluorotetradecanoic acid	0.0037 U	ND		nc	30
375-73-5	Perfluorobutanesulfonic acid	0.0037 U	ND		nc	30
2706-91-4	Perfluoropentanesulfonic acid	0.0046 U	ND		nc	30
355-46-4	Perfluorohexanesulfonic acid	0.0037 U	ND		nc	30
375-92-8	Perfluoroheptanesulfonic acid	0.0037 U	ND		nc	30
1763-23-1	Perfluorooctanesulfonic acid	0.0037 U	ND		nc	30
68259-12-1	Perfluorononanesulfonic acid	0.0037 U	ND		nc	30
335-77-3	Perfluorodecanesulfonic acid	0.0037 U	ND		nc	30
79780-39-5	Perfluorododecanesulfonic aci	0.0046 U	ND		nc	30
757124-72-44:2	Fluorotelomer sulfonate	0.19 U <sup>b</sup>	ND		nc	30
27619-97-2	6:2 Fluorotelomer sulfonate	0.19 U <sup>b</sup>	ND		nc	30
39108-34-4	8:2 Fluorotelomer sulfonate	0.19 U <sup>b</sup>	ND		nc	30
754-91-6	PFOSA	0.0037 U	ND		nc	30
31506-32-8	MeFOSA	0.0074 U	ND		nc	30
4151-50-2	EtFOSA	0.0074 U	ND		nc	30
2355-31-9	MeFOSAA	0.0046 U	ND		nc	30
2991-50-6	EtFOSAA	0.0046 U	ND		nc	30
24448-09-7	MeFOSE	0.037 U	ND		nc	30
1691-99-2	EtFOSE	0.037 U	ND		nc	30
13252-13-6	HFPO-DA (GenX)	0.0037 U	ND		nc	30
919005-14-4	ADONA	0.0074 U	ND		nc	30
377-73-1	PFMPA	0.0074 U	ND		nc	30
863090-89-5	PFMBA	0.0074 U	ND		nc	30
151772-58-6	NFDHA	0.0074 U	ND		nc	30
756426-58-19	Cl-PF3ONS (F-53B Major)	0.0074 U	ND		nc	30
763051-92-91	Cl-PF3OUdS (F-53B Minor)	0.0074 U	ND		nc	30

\* = Outside of Control Limits.

# Duplicate Summary

Job Number: FC8392  
 Account: AECOMCOD AECOM, INC.  
 Project: N6274223F0104 RH Fire Suppression System

Sample	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
OP98297-DUP	4Q48792.D	1	08/09/23	AL	08/08/23	OP98297	S4Q713
FC8439-4	4Q48791.D	1	08/09/23	AL	08/08/23	OP98297	S4Q713
FC8439-4 <sup>a</sup>	6Q22667.D	10	08/10/23	MV	08/08/23	OP98297	S6Q330

The QC reported here applies to the following samples:

Method: EPA DRAFT 1633

FC8392-1

CAS No.	Compound	FC8439-4 ug/l	DUP Q	ug/l	Q	RPD	Limits
113507-82-7	PFEESA	0.0074 U	ND	ND	nc	30	
356-02-5	3:3 Fluorotelomer carboxylate	0.019 U	ND	ND	nc	30	
914637-49-35:3	Fluorotelomer carboxylate	0.093 U	ND	ND	nc	30	
812-70-4	7:3 Fluorotelomer carboxylate	0.093 U	ND	ND	nc	30	

CAS No.	ID Standard Recoveries	DUP	FC8439-4	FC8439-4	Limits
	13C4-PFBA	88%	96%	91%	20-150%
	13C5-PFPeA	77%	80%	121%	20-150%
	13C5-PFHxA	125%	131%	127%	20-150%
	13C4-PFHpA	133%	139%	131%	20-150%
	13C8-PFOA	121%	118%	117%	20-150%
	13C9-PFNA	107%	106%	115%	20-150%
	13C6-PFDA	120%	126%	121%	20-150%
	13C7-PFUnDA	124%	135%	121%	20-150%
	13C2-PFDoDA	106%	114%	107%	20-150%
	13C2-PFTeDA	82%	88%	91%	20-150%
	13C3-PFBS	113%	114%	116%	20-150%
	13C3-PFHxS	128%	116%	113%	20-150%
	13C8-PFOS	110%	120%	113%	20-150%
	13C8-FOSA	101%	109%	106%	20-150%
	d3-MeFOSA	98%	101%	89%	20-150%
	d5-EtFOSA	94%	92%	94%	20-150%
	d3-MeFOSAA	138%	138%	113%	20-150%
	d5-EtFOSAA	145%	146%	106%	20-150%
	d7-MeFOSE	83%	86%	92%	20-150%
	d9-EtFOSE	85%	92%	94%	20-150%
	13C2-4:2FTS	265%* c	258%* c	109%	20-180%
	13C2-6:2FTS	254%* c	227%* c	137%	20-180%
	13C2-8:2FTS	180%	184%* c	142%	20-180%
	13C3-HFPO-DA	88%	92%	120%	20-150%

- (a) Dilution required (ID recovery standard failure).
- (b) Result is from Run #2.
- (c) Outside control limits.

\* = Outside of Control Limits.