

 **ANALYTICAL REPORT****PREPARED FOR**

Attn: Terri Choy  
AECOM

1001 Bishop Street  
Honolulu HI 96813

Generated 12/13/2023 2:14 PM

**JOB DESCRIPTION**

Red Hill - AFFF Assessment Sampling

**JOB NUMBER**

580-134720-1

# Eurofins Seattle

## Job Notes

This report may not be reproduced except in full, and with written approval from the laboratory. The results relate only to the samples tested. For questions please contact the Project Manager at the e-mail address or telephone number listed on this page.

The test results in this report relate only to the samples as received by the laboratory and will meet all requirements of the methodology, with any exceptions noted. This report shall not be reproduced except in full, without the express written approval of the laboratory. All questions should be directed to the Eurofins Environment Testing Northwest, LLC Project Manager.

## Authorization



Generated  
12/13/2023 2:14 PM

---

Authorized for release by  
Tracy L Dutton, Client Relations Manager  
[Tracy.Dutton@et.eurofinsus.com](mailto:Tracy.Dutton@et.eurofinsus.com)  
253 248-4970

# Table of Contents

Cover Title Page . . . . .	1
Data Summaries . . . . .	5
Definitions . . . . .	5
Case Narrative . . . . .	6
Detection Summary . . . . .	7
Client Sample Results . . . . .	8
Default Detection Limits . . . . .	9
QC Sample Results . . . . .	10
QC Association . . . . .	11
Chronicle . . . . .	12
Certification Summary . . . . .	13
Method Summary . . . . .	14
Sample Summary . . . . .	15
Reagent Traceability . . . . .	16
COAs . . . . .	17
Organic Sample Data . . . . .	31
GC Semi VOA . . . . .	31
Method 8015C - DAI Glycols . . . . .	31
Method 8015C - DAI Glycols QC Summary . . . . .	32
Method 8015C - DAI Glycols Sample Data . . . . .	39
Standards Data . . . . .	60
Method 8015C - DAI Glycols ICAL Data . . . . .	60
Method 8015C - DAI Glycols CCAL Data . . . . .	90
Raw QC Data . . . . .	103
Method 8015C - DAI Glycols Blank Data . . . . .	103
Method 8015C - DAI Glycols LCS/LCSD Data . . . . .	106

# Table of Contents

Method 8015C - DAI Glycols MS/MSD Data .....	112
Method 8015C - DAI Glycols Run Logs .....	118
Method 8015C - DAI Glycols Prep Data .....	120
Subcontracted Data .....	122
Shipping and Receiving Documents .....	123
Client Chain of Custody .....	124
Sample Receipt Checklist .....	128

# Definitions/Glossary

Client: AECOM  
Project/Site: Red Hill - AFFF Assessment Sampling

Job ID: 580-134720-1

## Qualifiers

### GC Semi VOA

Qualifier	Qualifier Description
Q	One or more quality control criteria failed.
U	Undetected at the Limit of Detection.

## Glossary

Abbreviation	These commonly used abbreviations may or may not be present in this report.
α	Listed under the "D" column to designate that the result is reported on a dry weight basis
%R	Percent Recovery
CFL	Contains Free Liquid
CFU	Colony Forming Unit
CNF	Contains No Free Liquid
DER	Duplicate Error Ratio (normalized absolute difference)
Dil Fac	Dilution Factor
DL	Detection Limit (DoD/DOE)
DL, RA, RE, IN	Indicates a Dilution, Re-analysis, Re-extraction, or additional Initial metals/anion analysis of the sample
DLC	Decision Level Concentration (Radiochemistry)
EDL	Estimated Detection Limit (Dioxin)
LOD	Limit of Detection (DoD/DOE)
LOQ	Limit of Quantitation (DoD/DOE)
MCL	EPA recommended "Maximum Contaminant Level"
MDA	Minimum Detectable Activity (Radiochemistry)
MDC	Minimum Detectable Concentration (Radiochemistry)
MDL	Method Detection Limit
ML	Minimum Level (Dioxin)
MPN	Most Probable Number
MQL	Method Quantitation Limit
NC	Not Calculated
ND	Not Detected at the reporting limit (or MDL or EDL if shown)
NEG	Negative / Absent
POS	Positive / Present
PQL	Practical Quantitation Limit
PRES	Presumptive
QC	Quality Control
RER	Relative Error Ratio (Radiochemistry)
RL	Reporting Limit or Requested Limit (Radiochemistry)
RPD	Relative Percent Difference, a measure of the relative difference between two points
TEF	Toxicity Equivalent Factor (Dioxin)
TEQ	Toxicity Equivalent Quotient (Dioxin)
TNTC	Too Numerous To Count

**Job Narrative  
580-134720-1**

Analytical test results meet all requirements of the associated regulatory program listed on the Accreditation/Certification Summary Page unless otherwise noted under the individual analysis. Data qualifiers are applied to indicate exceptions. Noncompliant quality control (QC) is further explained in narrative comments.

Matrix QC may not be reported if insufficient sample or site-specific QC samples were not submitted. In these situations, to demonstrate precision and accuracy at a batch level, a LCS/LCSD may be performed, unless otherwise specified in the method. Surrogate and/or isotope dilution analyte recoveries (if applicable) which are outside of the QC window are confirmed unless attributed to a dilution or otherwise noted in the narrative.

**Receipt**

The samples were received on 12/8/2023 4:00 PM. Unless otherwise noted below, the samples arrived in good condition, and, where required, properly preserved and on ice. The temperature of the cooler at receipt time was 5.0°C

Note: All samples which require thermal preservation are considered acceptable if the arrival temperature is within 2C of the required temperature or method specified range. For samples with a specified temperature of 4C, samples with a temperature ranging from just above freezing temperature of water to 6C shall be acceptable. Samples that are hand delivered immediately following collection may not meet these criteria, however they will be deemed acceptable according to NELAC standards, if there is evidence that the chilling process has begun, such as arrival on ice, etc.

**Method 8015C GLY - Glycols- Direct Injection (GC/FID)**

Samples AF-RHMW12A-WGFD01LF-2312 (134720-7), AF-RHMW12A-WGN01LF-2312 (134720-6), AF-RHMW17D-WGN01LF-2312 (134720-2), AF-RHMW17D-WQFB01-2312 (134720-3), AF-RHMW17S-WGN01LF-2312 (134720-4), AF-RHMW17S-WQEB01-2312 (134720-5) and AF-RHMW17-WGN01LF-2312 (134720-1) were analyzed for Glycols- Direct Injection (GC/FID). The samples were analyzed on 12/12/2023 and 12/13/2023.

The continuing calibration verification (CCV) associated with batch 680-812978 recovered above the upper control limit for 2-(2-Butoxyethoxy)ethanol. The samples associated with this CCV were non-detects for the affected analyte; therefore, the data has been reported.

# Detection Summary

Client: AECOM  
Project/Site: Red Hill - AFFF Assessment Sampling

Job ID: 580-134720-1

**Client Sample ID: AF-RHMW17-WGN01LF-2312**

**Lab Sample ID: 580-134720-1**

No Detections.

**Client Sample ID: AF-RHMW17D-WGN01LF-2312**

**Lab Sample ID: 580-134720-2**

No Detections.

**Client Sample ID: AF-RHMW17D-WQFB01-2312**

**Lab Sample ID: 580-134720-3**

No Detections.

**Client Sample ID: AF-RHMW17S-WGN01LF-2312**

**Lab Sample ID: 580-134720-4**

No Detections.

**Client Sample ID: AF-RHMW17S-WQEB01-2312**

**Lab Sample ID: 580-134720-5**

No Detections.

**Client Sample ID: AF-RHMW12A-WGN01LF-2312**

**Lab Sample ID: 580-134720-6**

No Detections.

**Client Sample ID: AF-RHMW12A-WGFD01LF-2312**

**Lab Sample ID: 580-134720-7**

No Detections.

This Detection Summary does not include radiochemical test results.

# Client Sample Results

Client: AECOM  
Project/Site: Red Hill - AFFF Assessment Sampling

Job ID: 580-134720-1

**Client Sample ID: AF-RHMW17-WGN01LF-2312**

**Lab Sample ID: 580-134720-1**

Date Collected: 12/06/23 11:50

Matrix: Water

Date Received: 12/08/23 16:00

**Method: SW846 8015C GLY - Glycols- Direct Injection (GC/FID)**

Analyte	Result	Qualifier	LOQ	DL	Unit	D	Prepared	Analyzed	Dil Fac
2-(2-Butoxyethoxy)ethanol	3.0	U Q	5.0	1.1	mg/L			12/12/23 23:48	1

**Client Sample ID: AF-RHMW17D-WGN01LF-2312**

**Lab Sample ID: 580-134720-2**

Date Collected: 12/06/23 10:40

Matrix: Water

Date Received: 12/08/23 16:00

**Method: SW846 8015C GLY - Glycols- Direct Injection (GC/FID)**

Analyte	Result	Qualifier	LOQ	DL	Unit	D	Prepared	Analyzed	Dil Fac
2-(2-Butoxyethoxy)ethanol	3.0	U Q	5.0	1.1	mg/L			12/13/23 00:11	1

**Client Sample ID: AF-RHMW17D-WQFB01-2312**

**Lab Sample ID: 580-134720-3**

Date Collected: 12/06/23 10:10

Matrix: Water

Date Received: 12/08/23 16:00

**Method: SW846 8015C GLY - Glycols- Direct Injection (GC/FID)**

Analyte	Result	Qualifier	LOQ	DL	Unit	D	Prepared	Analyzed	Dil Fac
2-(2-Butoxyethoxy)ethanol	3.0	U Q	5.0	1.1	mg/L			12/13/23 00:35	1

**Client Sample ID: AF-RHMW17S-WGN01LF-2312**

**Lab Sample ID: 580-134720-4**

Date Collected: 12/06/23 09:05

Matrix: Water

Date Received: 12/08/23 16:00

**Method: SW846 8015C GLY - Glycols- Direct Injection (GC/FID)**

Analyte	Result	Qualifier	LOQ	DL	Unit	D	Prepared	Analyzed	Dil Fac
2-(2-Butoxyethoxy)ethanol	3.0	U Q	5.0	1.1	mg/L			12/13/23 00:58	1

**Client Sample ID: AF-RHMW17S-WQEB01-2312**

**Lab Sample ID: 580-134720-5**

Date Collected: 12/06/23 09:40

Matrix: Water

Date Received: 12/08/23 16:00

**Method: SW846 8015C GLY - Glycols- Direct Injection (GC/FID)**

Analyte	Result	Qualifier	LOQ	DL	Unit	D	Prepared	Analyzed	Dil Fac
2-(2-Butoxyethoxy)ethanol	3.0	U Q	5.0	1.1	mg/L			12/13/23 01:21	1

**Client Sample ID: AF-RHMW12A-WGN01LF-2312**

**Lab Sample ID: 580-134720-6**

Date Collected: 12/07/23 09:15

Matrix: Water

Date Received: 12/08/23 16:00

**Method: SW846 8015C GLY - Glycols- Direct Injection (GC/FID)**

Analyte	Result	Qualifier	LOQ	DL	Unit	D	Prepared	Analyzed	Dil Fac
2-(2-Butoxyethoxy)ethanol	3.0	U Q	5.0	1.1	mg/L			12/13/23 01:45	1

**Client Sample ID: AF-RHMW12A-WGFD01LF-2312**

**Lab Sample ID: 580-134720-7**

Date Collected: 12/07/23 09:15

Matrix: Water

Date Received: 12/08/23 16:00

**Method: SW846 8015C GLY - Glycols- Direct Injection (GC/FID)**

Analyte	Result	Qualifier	LOQ	DL	Unit	D	Prepared	Analyzed	Dil Fac
2-(2-Butoxyethoxy)ethanol	3.0	U Q	5.0	1.1	mg/L			12/13/23 02:08	1



# Default Detection Limits

Client: AECOM

Job ID: 580-134720-1

Project/Site: Red Hill - AFFF Assessment Sampling

---

## Method: 8015C GLY - Glycols- Direct Injection (GC/FID)

---

Analyte	LOQ	DL	Units
2-(2-Butoxyethoxy)ethanol	5.0	1.1	mg/L

# QC Sample Results

Client: AECOM  
 Project/Site: Red Hill - AFFF Assessment Sampling

Job ID: 580-134720-1

## Method: 8015C GLY - Glycols- Direct Injection (GC/FID)

**Lab Sample ID: MB 680-812978/11**  
**Matrix: Water**  
**Analysis Batch: 812978**

**Client Sample ID: Method Blank**  
**Prep Type: Total/NA**

Analyte	MB Result	MB Qualifier	LOQ	DL	Unit	D	Prepared	Analyzed	Dil Fac
2-(2-Butoxyethoxy)ethanol	3.0	U	5.0	1.1	mg/L			12/12/23 19:07	1

**Lab Sample ID: LCS 680-812978/7**  
**Matrix: Water**  
**Analysis Batch: 812978**

**Client Sample ID: Lab Control Sample**  
**Prep Type: Total/NA**

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec Limits
2-(2-Butoxyethoxy)ethanol	20.0	19.9		mg/L		100	50 - 150

**Lab Sample ID: LCSD 680-812978/8**  
**Matrix: Water**  
**Analysis Batch: 812978**

**Client Sample ID: Lab Control Sample Dup**  
**Prep Type: Total/NA**

Analyte	Spike Added	LCSD Result	LCSD Qualifier	Unit	D	%Rec	%Rec Limits	RPD	RPD Limit
2-(2-Butoxyethoxy)ethanol	20.0	20.8		mg/L		104	50 - 150	4	50

**Lab Sample ID: 580-134720-1 MS**  
**Matrix: Water**  
**Analysis Batch: 812978**

**Client Sample ID: AF-RHMW17-WGN01LF-2312**  
**Prep Type: Total/NA**

Analyte	Sample Result	Sample Qualifier	Spike Added	MS Result	MS Qualifier	Unit	D	%Rec	%Rec Limits
2-(2-Butoxyethoxy)ethanol	3.0	U Q	20.0	17.7		mg/L		89	50 - 150

**Lab Sample ID: 580-134720-1 MSD**  
**Matrix: Water**  
**Analysis Batch: 812978**

**Client Sample ID: AF-RHMW17-WGN01LF-2312**  
**Prep Type: Total/NA**

Analyte	Sample Result	Sample Qualifier	Spike Added	MSD Result	MSD Qualifier	Unit	D	%Rec	%Rec Limits	RPD	RPD Limit
2-(2-Butoxyethoxy)ethanol	3.0	U Q	20.0	15.3		mg/L		77	50 - 150	15	50

# QC Association Summary

Client: AECOM  
Project/Site: Red Hill - AFFF Assessment Sampling

Job ID: 580-134720-1

## GC Semi VOA

### Analysis Batch: 812978

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
580-134720-1	AF-RHMW17-WGN01LF-2312	Total/NA	Water	8015C GLY	
580-134720-2	AF-RHMW17D-WGN01LF-2312	Total/NA	Water	8015C GLY	
580-134720-3	AF-RHMW17D-WQFB01-2312	Total/NA	Water	8015C GLY	
580-134720-4	AF-RHMW17S-WGN01LF-2312	Total/NA	Water	8015C GLY	
580-134720-5	AF-RHMW17S-WQEB01-2312	Total/NA	Water	8015C GLY	
580-134720-6	AF-RHMW12A-WGN01LF-2312	Total/NA	Water	8015C GLY	
580-134720-7	AF-RHMW12A-WGFD01LF-2312	Total/NA	Water	8015C GLY	
MB 680-812978/11	Method Blank	Total/NA	Water	8015C GLY	
LCS 680-812978/7	Lab Control Sample	Total/NA	Water	8015C GLY	
LCSD 680-812978/8	Lab Control Sample Dup	Total/NA	Water	8015C GLY	
580-134720-1 MS	AF-RHMW17-WGN01LF-2312	Total/NA	Water	8015C GLY	
580-134720-1 MSD	AF-RHMW17-WGN01LF-2312	Total/NA	Water	8015C GLY	

# Lab Chronicle

Client: AECOM  
Project/Site: Red Hill - AFFF Assessment Sampling

Job ID: 580-134720-1

**Client Sample ID: AF-RHMW17-WGN01LF-2312**

**Lab Sample ID: 580-134720-1**

Date Collected: 12/06/23 11:50

Matrix: Water

Date Received: 12/08/23 16:00

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Analysis	8015C GLY		1	812978	DBM	EET SAV	12/12/23 23:48

**Client Sample ID: AF-RHMW17D-WGN01LF-2312**

**Lab Sample ID: 580-134720-2**

Date Collected: 12/06/23 10:40

Matrix: Water

Date Received: 12/08/23 16:00

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Analysis	8015C GLY		1	812978	DBM	EET SAV	12/13/23 00:11

**Client Sample ID: AF-RHMW17D-WQFB01-2312**

**Lab Sample ID: 580-134720-3**

Date Collected: 12/06/23 10:10

Matrix: Water

Date Received: 12/08/23 16:00

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Analysis	8015C GLY		1	812978	DBM	EET SAV	12/13/23 00:35

**Client Sample ID: AF-RHMW17S-WGN01LF-2312**

**Lab Sample ID: 580-134720-4**

Date Collected: 12/06/23 09:05

Matrix: Water

Date Received: 12/08/23 16:00

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Analysis	8015C GLY		1	812978	DBM	EET SAV	12/13/23 00:58

**Client Sample ID: AF-RHMW17S-WQEB01-2312**

**Lab Sample ID: 580-134720-5**

Date Collected: 12/06/23 09:40

Matrix: Water

Date Received: 12/08/23 16:00

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Analysis	8015C GLY		1	812978	DBM	EET SAV	12/13/23 01:21

**Client Sample ID: AF-RHMW12A-WGN01LF-2312**

**Lab Sample ID: 580-134720-6**

Date Collected: 12/07/23 09:15

Matrix: Water

Date Received: 12/08/23 16:00

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Analysis	8015C GLY		1	812978	DBM	EET SAV	12/13/23 01:45

**Client Sample ID: AF-RHMW12A-WGFD01LF-2312**

**Lab Sample ID: 580-134720-7**

Date Collected: 12/07/23 09:15

Matrix: Water

Date Received: 12/08/23 16:00

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Analysis	8015C GLY		1	812978	DBM	EET SAV	12/13/23 02:08

## Laboratory References:

EET SAV = Eurofins Savannah, 5102 LaRoche Avenue, Savannah, GA 31404, TEL (912)354-7858

# Accreditation/Certification Summary

Client: AECOM  
Project/Site: Red Hill - AFFF Assessment Sampling

Job ID: 580-134720-1

## Laboratory: Eurofins Savannah

Unless otherwise noted, all analytes for this laboratory were covered under each accreditation/certification below.

Authority	Program	Identification Number	Expiration Date
ANAB	Dept. of Defense ELAP	L2463	09-22-24

The following analytes are included in this report, but the laboratory is not certified by the governing authority. This list may include analytes for which the agency does not offer certification.

Analysis Method	Prep Method	Matrix	Analyte
8015C GLY		Water	2-(2-Butoxyethoxy)ethanol

# Method Summary

Client: AECOM

Job ID: 580-134720-1

Project/Site: Red Hill - AFFF Assessment Sampling

---

---

<b>Method</b>	<b>Method Description</b>	<b>Protocol</b>	<b>Laboratory</b>
8015C GLY	Glycols- Direct Injection (GC/FID)	SW846	EET SAV

**Protocol References:**

SW846 = "Test Methods For Evaluating Solid Waste, Physical/Chemical Methods", Third Edition, November 1986 And Its Updates.

**Laboratory References:**

EET SAV = Eurofins Savannah, 5102 LaRoche Avenue, Savannah, GA 31404, TEL (912)354-7858

# Sample Summary

Client: AECOM

Job ID: 580-134720-1

Project/Site: Red Hill - AFFF Assessment Sampling

---

<b>Lab Sample ID</b>	<b>Client Sample ID</b>	<b>Matrix</b>	<b>Collected</b>	<b>Received</b>
580-134720-1	AF-RHMW17-WGN01LF-2312	Water	12/06/23 11:50	12/08/23 16:00
580-134720-2	AF-RHMW17D-WGN01LF-2312	Water	12/06/23 10:40	12/08/23 16:00
580-134720-3	AF-RHMW17D-WQFB01-2312	Water	12/06/23 10:10	12/08/23 16:00
580-134720-4	AF-RHMW17S-WGN01LF-2312	Water	12/06/23 09:05	12/08/23 16:00
580-134720-5	AF-RHMW17S-WQEB01-2312	Water	12/06/23 09:40	12/08/23 16:00
580-134720-6	AF-RHMW12A-WGN01LF-2312	Water	12/07/23 09:15	12/08/23 16:00
580-134720-7	AF-RHMW12A-WGFD01LF-2312	Water	12/07/23 09:15	12/08/23 16:00

REAGENT TRACEABILITY SUMMARY

Lab Name: Eurofins Savannah

Job No.: 580-134720-1

SDG No.:

Reagent ID	Exp Date	Prep Date	Dilutant Used	Reagent Final Volume	Parent Reagent		Analyte	Concentration
					Reagent ID	Volume Added		
SG_Gly_CAL_00053	12/27/23		o2si, Lot 480919			(Purchased Reagent)	2,2'-Oxybisethanol	2000 ug/mL
							2-(2-Butoxyethoxy)ethanol	2000 ug/mL
							2-Butoxyethanol	2000 ug/mL
							4-Hydroxy-4-methyl-2-pentanone	2000 ug/mL
							Dipropylene Glycol Methyl Ether	2000 ug/mL
							Ethanol, 2-propoxy	2000 ug/mL
							Ethylene glycol	2000 ug/mL
							Propylene glycol	2000 ug/mL
SG_Gly_CAL_00056	01/17/24		o2si, Lot 500821			(Purchased Reagent)	2-(2-Butoxyethoxy)ethanol	2000 ug/mL
SG_GLY_ISTD_00129	02/09/24		Agilent, Lot 0006738806			(Purchased Reagent)	n-Heptyl Alcohol	5000 ug/mL
SG_GlyICV_00062	05/02/24		o2si, Lot 454407			(Purchased Reagent)	2-(2-Butoxyethoxy)ethanol	2000 ug/mL



Reagent

---

**SG\_Gly\_CAL\_00053**



ISO/IEC 17025 Accredited  
Chemical Testing Lab  
Cert. No. 3031.01



ISO 17034 Accredited  
Reference Material Producer  
Cert. No. 3031.02

## Certificate of Analysis

Rev 0

Page 1 of 3

Catalog No.	Lot No.	Storage	Solvent	Date Received	Exp. Date
G34-120070-04	480919	≤ -10 °C	P/T Methanol		2-May-2024

**Description:**

ISO 17034 -Custom Volatiles Mix,105-12, 2000 & 4,000 mg/L, 1 mL

**Container:**

1 ml Ampule, Amber Glass

### Certified Values:

The certified value is based on gravimetric and volumetric preparation of this Certified Reference Material (CRM). This CRM has been confirmed by GC/MS, GC, HPLC, UPLC/HRAM-MS, UV/VIS, Enzymatic, and/or wet chemistry techniques using internally developed method(s) against independent source(s). The uncertainty value is calculated for a 95% confidence interval with a *k* value of 2. The purity of neat materials not traceable to an ISO 17034:2016 accredited Reference Material Provider is traceable to internal analysis by GC, GC/MS, HPLC, Enzymatic, or wet chemistry techniques and compared to a National Metrological Institute such as NIST where feasible.

Compound	CAS No.	Purity (%)	Neat Material Lot No.	Concentration
2-butoxyethanol	111-76-2	99.6	311.9.2P	1986 ± 100 mg/L
diethylene glycol butyl ether	112-34-5	99.8	2323.7.2P	2008 ± 100 mg/L
propyl cellosolve	2807-30-9	99.9	1570.7.2P	1980 ± 100 mg/L
dipropylene glycol monomethyl ether	34590-94-8	99.7	2333.7.2P	2014 ± 100 mg/L
ethylene glycol	107-21-1	100	307.201.1P	1968 ± 99 mg/L
di(ethylene glycol)	111-46-6	99.5	309.7.2P	1994 ± 100 mg/L
tri(ethylene glycol)	112-27-6	99.9	310.7.2.1.1P	1974 ± 110 mg/L
4-Hydroxy-4-methyl-2-pentanone	123-42-2	98	2334.286.1P	1991 ± 110 mg/L
1,2-propanediol	57-55-6	99.5	306.9.3P	1998 ± 100 mg/L
tetraethylene glycol	112-60-7	98	3754.7.1P	3959 ± 200 mg/L

### Intended Uses:

This CRM is intended for use as a calibration standard or a quality control standard for chromatography equipment such as GC, GC/MS, HPLC, and HPLC/MS. It may also be used for various USEPA, NIOSH and ASTM methods.

Recommended storage container for ampuled products after opening is a 12 mm x 32 mm amber vial with screw cap Teflon lined silicon septum. The modeled % change per day can be calculated using the following:

# Certificate of Analysis

Page 2 of 3

Catalog No. G34-120070-04

Lot No. 480919

Expiration Date 2 -May-2024

$$\% \text{ Change} = 116192x^{-2.578} + 40.383e^{-0.03y}$$

where x = boiling point of the most volatile analyte in the mix (in degrees K)

y = boiling point of the solvent (in degrees K)

This model assumes the container is stored at -10 °C and is unopened during storage. The user should determine what the acceptable error for their process is and calculate the maximum number of days the opened ampule should be stored.

## Method of Preparation:

This standard was prepared gravimetrically using balances calibrated with National Institute of Standards and Technology (NIST) traceable weights (NIST Test Numbers 822/273070-06, 822/275141-07, 822/278993-10). Only calibrated Class A volumetric glassware and/or calibrated syringes were used to prepare this standard. Raw materials may have been checked for stoichiometry and purity prior to use. This standard has been analyzed against an independent source.

## Packaging and Storage:

The solution should be stored according to the following storage requirements:  $\leq -10$  °C

Once the product is opened, it should be transferred to a vial with minimum head space if the product was received in a sealed ampule.

## Glassware Calibration:

Only Class A glassware and/or calibrated syringes are used in the manufacture and quality control of standards. All glassware is calibrated using NIST traceable weights.

## Weights and Balance Calibration:

Weights used to perform daily checks on balances are calibrated annually by the State of South Carolina Department of Agriculture Metrology Laboratory and are traceable to NIST. Balances are checked daily in accordance to procedure O2-LB-G-002. Balances are calibrated annually by an ISO/IEC 17025:2017 accredited metrology service.

## Homogeneity:

Homogeneity has been established in accordance with internal procedure O2-QS-011 and has a maximum uncertainty of 0.1%. This is consistent with the intended use of this CRM. The homogeneity of this product has been confirmed by procedures consistent with ISO/IEC 17025:2017 and ISO 17034:2016. The homogeneity of this CRM is valid for sample sub-sizes that the end user can quantitatively reproduce.

## Hazardous Information:

Refer to MSDS.

## Calculation of Uncertainty:

The following equations are used to calculate the value of the expanded uncertainty:

$u = ku_c$  u = Expanded Uncertainty, k = the coverage factor at the 95% confidence level, k = 2,  $u_c$  = the combined uncertainty

$u_c = (u_{\text{char}}^2 + u_{\text{tran}}^2 + u_{\text{homo}}^2 + u_{\text{ls}}^2)^{1/2}$  where  $u_i$  are the individual uncertainty components for manufacturing, transportation, homogeneity, and shelf life. While no significant uncertainty was detected in the replicates, a minimum contribution to

Manufactured By:




Brian Stokes

3 -May-2022

Production Chemist I

Certified By:



Tyler Sherman

14 -Jun-2022

Quality Control Chemist I

Released By:



Susan Mathews

14 -Jun-2022

Quality Control Team Lead

7290B Investment Drive • North Charleston, SC 29418  
Phone: 866.272.0932 • Fax: 866.509.5146 www.o2si.com

# Certificate of Analysis

Catalog No. G34-120070-04

Lot No. 480919

Expiration Date 2 -May-2024

uncertainty was added for homogeneity and long term stability as described in ISO Guide 35:2017.


## Expiration Information:

The stability of this product is based upon rigorous short term and long term testing of the solution for the certified value. These tests include the effect of temperature and packaging on the product. Studies on the short term instability have determined no contribution to instability as observed on the concentration under controlled transportation conditions. This standard is guaranteed until 2-May-2024

## Quality Standard Documentation:

- ISO/IEC 17025:2017 "General Requirements for the Competence of Testing and Calibration" - Chemical Testing - Accredited A2LA Certificate Number 3031.01
- ISO 17034:2016 "General Requirements for the Competence of Reference Material Producers" - Reference Material Production - Accredited A2LA Certificate Number 3031.02

Manufactured By:



Brian Stokes

3 -May-2022

Production Chemist I

Certified By:



Tyler Sherman

14 -Jun-2022

Quality Control Chemist I

7290B Investment Drive • North Charleston, SC 29418  
Phone: 866.272.0932 • Fax: 866.509.5146 www.o2si.com

Released By:



Susan Mathews

14 -Jun-2022

Quality Control Team Lead

Reagent

---

**SG\_Gly\_CAL\_00056**



ISO/IEC 17025 Accredited  
Chemical Testing Lab  
Cert. No. 3031.01



ISO 17034 Accredited  
Reference Material Producer  
Cert. No. 3031.02

Rev 0

## Certificate of Analysis

Page 1 of 3

Catalog No.	Lot No.	Storage	Solvent	Date Received	Exp. Date
G34-120070-04-SS	507036	≤ -10 °C	P/T Methanol		31-Jul-2025

### Description:

ISO 17034 -Custom Volatiles Mix,105-12, Second Source, 2000 & 4,000 mg/L, 1 mL

### Container:

1 ml Ampule, Amber Glass

### Certified Values:

The certified value is based on gravimetric and volumetric preparation of this Certified Reference Material (CRM). This CRM has been confirmed by GC/MS, GC, HPLC, UPLC/HRAM-MS, UV/VIS, Enzymatic, and/or wet chemistry techniques using internally developed method(s) against independent source(s). The uncertainty value is calculated for a 95% confidence interval with a *k* value of 2. The purity of neat materials not traceable to an ISO 17034:2016 accredited Reference Material Provider is traceable to internal analysis by GC, GC/MS, HPLC, Enzymatic, or wet chemistry techniques and compared to a National Metrological Institute such as NIST where feasible.

Compound	CAS No.	Purity (%)	Neat Material Lot No.	Concentration
2-butoxyethanol	111-76-2	99.5	311.7.1.1S	2006 ± 25 mg/L
diethylene glycol butyl ether	112-34-5	99.8	2323.7.2.1S	2000 ± 25 mg/L
propyl cellosolve	2807-30-9	99.9	1570.7.2.1S	1994 ± 25 mg/L
dipropylene glycol monomethyl ether	34590-94-8	99.7	2333.7.2.1S	2018 ± 25 mg/L
ethylene glycol	107-21-1	99.9	307.1.5.1S	2002 ± 24 mg/L
di(ethylene glycol)	111-46-6	99.5	309.7.2.1S	2006 ± 25 mg/L
tri(ethylene glycol)	112-27-6	100	310.7.4.1S	2004 ± 25 mg/L
4-Hydroxy-4-methyl-2-pentanone	123-42-2	98	2334.286.1.1S	2003 ± 47 mg/L
1,2-propanediol	57-55-6	99.9	306.9.4.1S	2010 ± 25 mg/L
tetraethylene glycol	112-60-7	98	3754.7.1.1S	3998 ± 47 mg/L

### Intended Uses:

This CRM is intended for use as a calibration standard or a quality control standard for chromatography equipment such as GC, GC/MS, HPLC, and HPLC/MS. It may also be used for various USEPA, NIOSH and ASTM methods.

Recommended storage container for ampuled products after opening is a 12 mm x 32 mm amber vial with screw cap Teflon lined silicon septum. The modeled % change per day can be calculated using the following:

# Certificate of Analysis

Page 2 of 3

Catalog No. G34-120070-04-SS

Lot No. 507036

Expiration Date 31-Jul-2025

$$\% \text{ Change} = 116192x^{-2.578} + 40.383e^{-0.03y}$$

where  $x$  = boiling point of the most volatile analyte in the mix (in degrees K)  
 $y$  = boiling point of the solvent (in degrees K)

This model assumes the container is stored at  $-10\text{ }^{\circ}\text{C}$  and is unopened during storage. The user should determine what the acceptable error for their process is and calculate the maximum number of days the opened ampule should be stored.

## Method of Preparation:

This standard was prepared gravimetrically using balances calibrated with National Institute of Standards and Technology (NIST) traceable weights (NIST Test Numbers 822/273070-06, 822/275141-07, 822/278993-10). Only calibrated Class A volumetric glassware and/or calibrated syringes were used to prepare this standard. Raw materials may have been checked for stoichiometry and purity prior to use. This standard has been analyzed against an independent source.

## Packaging and Storage:

The solution should be stored according to the following storage requirements:  $\leq -10\text{ }^{\circ}\text{C}$   
Once the product is opened, it should be transferred to a vial with minimum head space if the product was received in a sealed ampule.

## Glassware Calibration:

Only Class A glassware and/or calibrated syringes are used in the manufacture and quality control of standards. All glassware is calibrated using NIST traceable weights.

## Weights and Balance Calibration:

Weights used to perform daily checks on balances are calibrated annually and are traceable to NIST. Balances are checked daily in accordance to procedure O2-LB-G-002. Balances are calibrated annually by an ISO/IEC 17025:2017 accredited metrology service.

## Homogeneity:

Homogeneity has been established in accordance with internal procedure O2-QS-011 and has a maximum uncertainty of 0.1%. This is consistent with the intended use of this CRM. The homogeneity of this product has been confirmed by procedures consistent with ISO/IEC 17025:2017 and ISO 17034:2016. The homogeneity of this CRM is valid for sample sub-sizes that the end user can quantitatively reproduce.

## Hazardous Information:

Refer to MSDS.

## Calculation of Uncertainty:

The following equations are used to calculate the value of the expanded uncertainty:  
 $u = ku_c$   $u$  = Expanded Uncertainty,  $k$  = the coverage factor at the 95% confidence level,  $k = 2$ ,  $u_c$  = the combined uncertainty  
 $u_c = (u_{\text{char}}^2 + u_{\text{tran}}^2 + u_{\text{homo}}^2 + u_{\text{ls}}^2)^{1/2}$  where  $u_i$  are the individual uncertainty components for manufacturing, transportation, homogeneity, and shelf life. While no significant uncertainty was detected in the replicates, a minimum contribution to

Manufactured By:




Andrea Schaible  
1-Aug-2023  
Production Chemist I

Certified By:



Jared Ball  
14-Aug-2023  
Quality Control Chemist II

Released By:



Susan Mathews  
14-Aug-2023  
Quality Control Team Lead

7290B Investment Drive • North Charleston, SC 29418  
Phone: 866.272.0932 • Fax: 866.509.5146 www.o2si.com

# Certificate of Analysis

Page 3 of 3

Catalog No. G34-120070-04-SS

Lot No. 507036

Expiration Date 31-Jul-2025

uncertainty was added for homogeneity and long term stability as described in ISO Guide 35:2017.

## Expiration Information:

The stability of this product is based upon rigorous short term and long term testing of the solution for the certified value. These tests include the effect of temperature and packaging on the product. Studies on the short term instability have determined no contribution to instability as observed on the concentration under controlled transportation conditions. This standard is guaranteed until 31-Jul-2025

## Quality Standard Documentation:

- ISO/IEC 17025:2017 "General Requirements for the Competence of Testing and Calibration" - Chemical Testing - Accredited A2LA Certificate Number 3031.01
- ISO 17034:2016 "General Requirements for the Competence of Reference Material Producers" - Reference Material Production - Accredited A2LA Certificate Number 3031.02

Manufactured By:



Andrea Schaible

1-Aug-2023

Production Chemist I

Certified By:



Jared Ball

14-Aug-2023

Quality Control Chemist II

7290B Investment Drive • North Charleston, SC 29418  
Phone: 866.272.0932 • Fax: 866.509.5146 www.o2si.com

Released By:



Susan Mathews

14-Aug-2023

Quality Control Team Lead



Reagent

---

**SG\_GLY\_ISTD\_00129**

**Reference Material Certificate**  
**Product Information Sheet**

**Product Name:** Custom Standard

**Lot Number:** 0006759266

**Product Number:** CUS-6046

**Lot Issue Date:** 24-Aug-2023

**Storage Conditions:** Store at Room Temperature (15° to 30°C).

**Expiration Date:** 30-Sep-2025

Component Name	Concentration	Uncertainty	CAS#	Analyte Lot
n-heptanol	5018	± 25 µg/mL	000111-70-6	RM04540

**Matrix:** methanol (methyl alcohol)

**Description:**

This document is prepared in accordance with ISO 17034 and Guide 31. This analytical reference material standard was manufactured and verified in accordance with an ISO 9001 registered quality system and analyte concentrations were verified by an ISO 17025 accredited laboratory. The concentration and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed above.

**Traceability:**

The balances used for these measurements are calibrated with weights traceable to NIST in compliance with ANSI/NCCL Z540.3, ISO 9001, ISO 17025, and ISO 17034. Calibrated Class A glassware is used for volumetric measurements. Thermometers are calibrated against a NIST traceable thermometer in accordance with NIST Special Publication 1088.

**Homogeneity:**

This analytical reference standard was unitized according to an in-house procedure and is guaranteed to be homogeneous. There is no minimum sub-sample size required.

**Instructions for Use:**

Sample aliquots for analysis should be withdrawn at 20°C to 25°C immediately after opening the container and should be processed without delay for the certified values to be valid within the stated uncertainties.

**Safety:**

Refer to the Safety Data Sheet on [www.agilent.com](http://www.agilent.com) for information regarding this analytical reference material.

**Intended Use:**

This analytical reference standard is intended for the preparation of working reference samples for use in routine laboratory analyses, calibration of instruments, validation of analytical methods, assessments of measurement methods, and continuing calibration verification.

**Expiration of Certification:**

The certification of this analytical reference standard is valid until the expiration date specified above, provided the material is handled and stored in accordance with the instructions given in this certificate. This certification is nullified if the material is damaged, contaminated, or otherwise modified.

Reagent

---

**SG\_GlyICV\_00062**



ISO/IEC 17025 Accredited  
Chemical Testing Lab  
Cert. No. 3031.01



ISO 17034 Accredited  
Reference Material Producer  
Cert. No. 3031.02

Rev 0

## Certificate of Analysis

Page 1 of 3

Catalog No.	Lot No.	Storage	Solvent	Date Received	Exp. Date
G34-120070-04	480919	≤ -10 °C	P/T Methanol		2-May-2024

### Description:

ISO 17034 -Custom Volatiles Mix,105-12, 2000 & 4,000 mg/L, 1 mL

### Container:

1 ml Ampule, Amber Glass

### Certified Values:

The certified value is based on gravimetric and volumetric preparation of this Certified Reference Material (CRM). This CRM has been confirmed by GC/MS, GC, HPLC, UPLC/HRAM-MS, UV/VIS, Enzymatic, and/or wet chemistry techniques using internally developed method(s) against independent source(s). The uncertainty value is calculated for a 95% confidence interval with a *k* value of 2. The purity of neat materials not traceable to an ISO 17034:2016 accredited Reference Material Provider is traceable to internal analysis by GC, GC/MS, HPLC, Enzymatic, or wet chemistry techniques and compared to a National Metrological Institute such as NIST where feasible.

Compound	CAS No.	Purity (%)	Neat Material Lot No.	Concentration
2-butoxyethanol	111-76-2	99.6	311.9.2P	1986 ± 100 mg/L
diethylene glycol butyl ether	112-34-5	99.8	2323.7.2P	2008 ± 100 mg/L
propyl cellosolve	2807-30-9	99.9	1570.7.2P	1980 ± 100 mg/L
dipropylene glycol monomethyl ether	34590-94-8	99.7	2333.7.2P	2014 ± 100 mg/L
ethylene glycol	107-21-1	100	307.201.1P	1968 ± 100 mg/L
di(ethylene glycol)	111-46-6	99.5	309.7.2P	1994 ± 100 mg/L
tri(ethylene glycol)	112-27-6	99.9	310.7.2.1.1P	1974 ± 110 mg/L
4-Hydroxy-4-methyl-2-pentanone	123-42-2	98	2334.286.1P	1991 ± 110 mg/L
1,2-propanediol	57-55-6	99.5	306.9.3P	1998 ± 100 mg/L
tetraethylene glycol	112-60-7	98	3754.7.1P	3959 ± 200 mg/L

### Intended Uses:

This CRM is intended for use as a calibration standard or a quality control standard for chromatography equipment such as GC, GC/MS, HPLC, and HPLC/MS. It may also be used for various USEPA, NIOSH and ASTM methods.

Recommended storage container for ampuled products after opening is a 12 mm x 32 mm amber vial with screw cap Teflon lined silicon septum. The modeled % change per day can be calculated using the following:

# Certificate of Analysis

Page 2 of 3

Catalog No. G34-120070-04

Lot No. 480919

Expiration Date 2 -May-2024

$$\% \text{ Change} = 116192x^{-2.578} + 40.383e^{-0.03y}$$

where x = boiling point of the most volatile analyte in the mix (in degrees K)

y = boiling point of the solvent (in degrees K)

This model assumes the container is stored at -10 °C and is unopened during storage. The user should determine what the acceptable error for their process is and calculate the maximum number of days the opened ampule should be stored.

## Method of Preparation:

This standard was prepared gravimetrically using balances calibrated with National Institute of Standards and Technology (NIST) traceable weights (NIST Test Numbers 822/273070-06, 822/275141-07, 822/278993-10). Only calibrated Class A volumetric glassware and/or calibrated syringes were used to prepare this standard. Raw materials may have been checked for stoichiometry and purity prior to use. This standard has been analyzed against an independent source.

## Packaging and Storage:

The solution should be stored according to the following storage requirements:  $\leq -10$  °C

Once the product is opened, it should be transferred to a vial with minimum head space if the product was received in a sealed ampule.

## Glassware Calibration:

Only Class A glassware and/or calibrated syringes are used in the manufacture and quality control of standards. All glassware is calibrated using NIST traceable weights.

## Weights and Balance Calibration:

Weights used to perform daily checks on balances are calibrated annually and are traceable to NIST. Balances are checked daily in accordance to procedure O2-LB-G-002. Balances are calibrated annually by an ISO/IEC 17025:2017 accredited metrology service.

## Homogeneity:

Homogeneity has been established in accordance with internal procedure O2-QS-011 and has a maximum uncertainty of 0.1%. This is consistent with the intended use of this CRM. The homogeneity of this product has been confirmed by procedures consistent with ISO/IEC 17025:2017 and ISO 17034:2016. The homogeneity of this CRM is valid for sample sub-sizes that the end user can quantitatively reproduce.

## Hazardous Information:

Refer to MSDS.

## Calculation of Uncertainty:

The following equations are used to calculate the value of the expanded uncertainty:

$u = k u_c$  u = Expanded Uncertainty, k = the coverage factor at the 95% confidence level, k = 2,  $u_c$  = the combined uncertainty

$u_c = (u_{\text{char}}^2 + u_{\text{tran}}^2 + u_{\text{homo}}^2 + u_{\text{lis}}^2)^{1/2}$  where  $u_i$  are the individual uncertainty components for manufacturing, transportation, homogeneity, and shelf life. While no significant uncertainty was detected in the replicates, a minimum contribution to

Manufactured By:




Brian Stokes

3 -May-2022

Production Chemist I

Certified By:



Tyler Sherman

14 -Jun-2022

Quality Control Chemist I

Released By:



Susan Mathews

14 -Jun-2022

Quality Control Team Lead

7290B Investment Drive • North Charleston, SC 29418  
Phone: 866.272.0932 • Fax: 866.509.5146 www.o2si.com

# Certificate of Analysis

Catalog No. G34-120070-04

Lot No. 480919

Expiration Date 2 -May-2024

uncertainty was added for homogeneity and long term stability as described in ISO Guide 35:2017.

## Expiration Information:

The stability of this product is based upon rigorous short term and long term testing of the solution for the certified value. These tests include the effect of temperature and packaging on the product. Studies on the short term instability have determined no contribution to instability as observed on the concentration under controlled transportation conditions. This standard is guaranteed until 2-May-2024

## Quality Standard Documentation:

- ISO/IEC 17025:2017 "General Requirements for the Competence of Testing and Calibration" - Chemical Testing - Accredited A2LA Certificate Number 3031.01
- ISO 17034:2016 "General Requirements for the Competence of Reference Material Producers" - Reference Material Production - Accredited A2LA Certificate Number 3031.02

### Manufactured By:

*Brian Stokes*

---

Brian Stokes  
3 -May-2022

**Production Chemist I**

### Certified By:

*Tyler Sherman*

---

Tyler Sherman  
14 -Jun-2022

**Quality Control Chemist I**

7290B Investment Drive • North Charleston, SC 29418  
Phone: 866.272.0932 • Fax: 866.509.5146 www.o2si.com

### Released By:

*Susan Mathews*

---

Susan Mathews  
14 -Jun-2022

**Quality Control Team Lead**

# Method 8015C - DAI Glycols

---

Glycols -Direct Injection (GC/FID) -  
Method 8015C

FORM III  
GC SEMI VOA LAB CONTROL SAMPLE RECOVERY

Lab Name: Eurofins Savannah Job No.: 580-134720-1  
 SDG No.: \_\_\_\_\_  
 Matrix: Water Level: Low Lab File ID: 1IL12007.D  
 Lab ID: LCS 680-812978/7 Client ID: \_\_\_\_\_

COMPOUND	SPIKE ADDED (mg/L)	LCS CONCENTRATION (mg/L)	LCS % REC	QC LIMITS REC	#
2-(2-Butoxyethoxy) ethanol	20.0	19.9	100	50-150	

# Column to be used to flag recovery and RPD values  
 FORM III 8015C GLY



FORM III  
GC SEMI VOA LAB CONTROL SAMPLE DUPLICATE RECOVERY

Lab Name: Eurofins Savannah Job No.: 580-134720-1  
 SDG No.: \_\_\_\_\_  
 Matrix: Water Level: Low Lab File ID: 1IL12008.D  
 Lab ID: LCSD 680-812978/8 Client ID: \_\_\_\_\_

COMPOUND	SPIKE ADDED (mg/L)	LCSD CONCENTRATION (mg/L)	LCSD % REC	% RPD	QC LIMITS		#
					RPD	REC	
2-(2-Butoxyethoxy) ethanol	20.0	20.8	104	4	50	50-150	

# Column to be used to flag recovery and RPD values  
 FORM III 8015C GLY

FORM III  
GC SEMI VOA MATRIX SPIKE RECOVERY

Lab Name: Eurofins Savannah Job No.: 580-134720-1  
 SDG No.: \_\_\_\_\_  
 Matrix: Water Level: Low Lab File ID: 1IL12030.D  
 Lab ID: 580-134720-1 MS Client ID: AF-RHMW17-WGN01LF-2312 MS

COMPOUND	SPIKE ADDED (mg/L)	SAMPLE CONCENTRATION (mg/L)	MS CONCENTRATION (mg/L)	MS % REC	QC LIMITS REC	#
2-(2-Butoxyethoxy) ethanol	20.0	3.0 U	17.7	89	50-150	

# Column to be used to flag recovery and RPD values  
 FORM III 8015C GLY

FORM III  
GC SEMI VOA MATRIX SPIKE DUPLICATE RECOVERY

Lab Name: Eurofins Savannah Job No.: 580-134720-1

SDG No.: \_\_\_\_\_

Matrix: Water Level: Low Lab File ID: 11L12031.D

Lab ID: 580-134720-1 MSD Client ID: AF-RHWW17-WGN01LF-2312 MSD

COMPOUND	SPIKE ADDED (mg/L)	MSD CONCENTRATION (mg/L)	MSD % REC	% RPD	QC LIMITS		#
					RPD	REC	
2-(2-Butoxyethoxy) ethanol	20.0	15.3	77	15	50	50-150	

# Column to be used to flag recovery and RPD values

FORM III 8015C GLY

FORM IV  
GC SEMI VOA METHOD BLANK SUMMARY

Lab Name: Eurofins Savannah Job No.: 580-134720-1  
 SDG No.: \_\_\_\_\_  
 Lab Sample ID: MB 680-812978/11  
 Matrix: Water Date Extracted: \_\_\_\_\_  
 Lab File ID: (1) 1IL12011.D Lab File ID: (2) \_\_\_\_\_  
 Date Analyzed: (1) 12/12/2023 19:07 Date Analyzed: (2) \_\_\_\_\_  
 Instrument ID: (1) CVGG2 Instrument ID: (2) \_\_\_\_\_  
 GC Column: (1) J&W DB WAX ID: 0.45 (mm) GC Column: (2) \_\_\_\_\_ ID: \_\_\_\_\_

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES:

CLIENT SAMPLE ID	LAB SAMPLE ID	DATE ANALYZED 1	DATE ANALYZED 2
	LCS 680-812978/7	12/12/2023 17:34	
	LCSD 680-812978/8	12/12/2023 17:57	
AF-RHMMW17-WGN01LF-2312	580-134720-1	12/12/2023 23:48	
AF-RHMMW17D-WGN01LF-2312	580-134720-2	12/13/2023 00:11	
AF-RHMMW17D-WQFB01-2312	580-134720-3	12/13/2023 00:35	
AF-RHMMW17S-WGN01LF-2312	580-134720-4	12/13/2023 00:58	
AF-RHMMW17S-WQEB01-2312	580-134720-5	12/13/2023 01:21	
AF-RHMMW12A-WGN01LF-2312	580-134720-6	12/13/2023 01:45	
AF-RHMMW12A-WGFD01LF-2312	580-134720-7	12/13/2023 02:08	
AF-RHMMW17-WGN01LF-2312 MS	580-134720-1 MS	12/13/2023 02:31	
AF-RHMMW17-WGN01LF-2312 MSD	580-134720-1 MSD	12/13/2023 02:55	

FORM VIII  
GC SEMI VOA INTERNAL STANDARD AREA AND RETENTION TIME SUMMARY

Lab Name: Eurofins Savannah Job No.: 580-134720-1  
 SDG No.: \_\_\_\_\_  
 Sample No.: ICIS 680-808404/9 Date Analyzed: 11/15/2023 16:34  
 Instrument ID: CVGG2 GC Column: J&W DB WAX ID: 0.45 (mm)  
 Lab File ID (Standard): 1IK15009.D Heated Purge: (Y/N) N  
 Calibration ID: 92674

	nHPA		#	RT #	#	RT #
	AREA #	RT #				
INITIAL CALIBRATION MID-POINT	2506763	6.28				
UPPER LIMIT	5013526	6.78				
LOWER LIMIT	1253382	5.78				
LAB SAMPLE ID	CLIENT SAMPLE ID					
ICV 680-808404/14		2248055	6.28			

nHPA = n-Heptyl Alcohol

Area Limit = 50%-200% of internal standard area  
 RT Limit = ± 0.5 minutes of internal standard RT

# Column used to flag values outside QC limits

FORM VIII  
GC SEMI VOA INTERNAL STANDARD AREA AND RETENTION TIME SUMMARY

Lab Name: Eurofins Savannah Job No.: 580-134720-1  
 SDG No.: \_\_\_\_\_  
 Sample No.: CCVIS 680-812978/6 Date Analyzed: 12/12/2023 17:10  
 Instrument ID: CVGG2 GC Column: J&W DB WAX ID: 0.45 (mm)  
 Lab File ID (Standard): 1IL12006.D Heated Purge: (Y/N) N  
 Calibration ID: 92674

		nHPA					
		AREA #	RT #	#	RT #	#	RT #
12/24 HOUR STD		1981601	6.25				
UPPER LIMIT		3963202	6.75				
LOWER LIMIT		990801	5.75				
LAB SAMPLE ID	CLIENT SAMPLE ID						
LCS 680-812978/7		3064834	6.25				
LCSD 680-812978/8		3251792	6.25				
MB 680-812978/11		2765692	6.25				
580-134720-1	AF-RHMW17-WGN01LF-2 312	2787818	6.25				
580-134720-2	AF-RHMW17D-WGN01LF- 2312	1546301	6.25				
580-134720-3	AF-RHMW17D-WQFB01-2 312	3323054	6.25				
580-134720-4	AF-RHMW17S-WGN01LF- 2312	2739650	6.25				
580-134720-5	AF-RHMW17S-WQEB01-2 312	3229024	6.25				
580-134720-6	AF-RHMW12A-WGN01LF- 2312	2010591	6.25				
580-134720-7	AF-RHMW12A-WGFD01LF -2312	2277287	6.26				
580-134720-1 MS	AF-RHMW17-WGN01LF-2 312 MS	2678030	6.25				
580-134720-1 MSD	AF-RHMW17-WGN01LF-2 312 MSD	2998400	6.25				
CCV 680-812978/33		2796242	6.25				

nHPA = n-Heptyl Alcohol

Area Limit = 50%-200% of internal standard area  
 RT Limit = ± 0.5 minutes of internal standard RT

# Column used to flag values outside QC limits

FORM I  
GC SEMI VOA ORGANICS ANALYSIS DATA SHEET

Lab Name: Eurofins Savannah Job No.: 580-134720-1  
 SDG No.: \_\_\_\_\_  
 Client Sample ID: AF-RHMW17-WGN01LF-2312 Lab Sample ID: 580-134720-1  
 Matrix: Water Lab File ID: 1IL12023.D  
 Analysis Method: 8015C GLY Date Collected: 12/06/2023 11:50  
 Extraction Method: \_\_\_\_\_ Date Extracted: \_\_\_\_\_  
 Sample wt/vol: 1(mL) Date Analyzed: 12/12/2023 23:48  
 Con. Extract Vol.: 1(mL) Dilution Factor: 1  
 Injection Volume: 1(uL) GC Column: J&W DB WAX ID: 0.45(mm)  
 % Moisture: \_\_\_\_\_ % Solids: \_\_\_\_\_ GPC Cleanup: (Y/N) N  
 Cleanup Factor: \_\_\_\_\_  
 Analysis Batch No.: 812978 Units: mg/L

CAS NO.	COMPOUND NAME	RESULT	Q	LOQ	LOD	DL
112-34-5	2-(2-Butoxyethoxy)ethanol	3.0	U Q	5.0	3.0	1.1

Eurofins Savannah  
Target Compound Quantitation Report

Data File: \\chromfs\Savannah\ChromData\CVGG2\20231212-91533.b\1IL12023.D  
 Lims ID: 580-134720-A-1  
 Client ID: AF-RHMW17-WGN01LF-2312  
 Sample Type: Client  
 Inject. Date: 12-Dec-2023 23:48:22      ALS Bottle#: 0      Worklist Smp#: 23  
 Injection Vol: 1.0 ul      Dil. Factor: 1.0000  
 Sample Info: 680-0091533-023  
 Operator ID:      Instrument ID: CVGG2  
 Method: \\chromfs\Savannah\ChromData\CVGG2\20231212-91533.b\8015\_GLY\_VGG.m  
 Limit Group: 8015C\_DAI  
 Last Update: 13-Dec-2023 15:05:29      Calib Date: 15-Nov-2023 17:44:20  
 Integrator: Falcon  
 Quant Method: Internal Standard      Quant By: Initial Calibration  
 Last ICal File: \\chromfs\Savannah\ChromData\CVGG2\20231115-90783.b\1IK15012.D  
 Column 1 : J&W DB WAX ( 0.45 mm)      Det: GC FID2B  
 Process Host: CTX1647

First Level Reviewer: AR8P      Date: 13-Dec-2023 15:04:56

RT (min.)	Exp RT (min.)	Dlt RT (min.)	Response	OnCol Amt ug/ml	Flags
--------------	------------------	------------------	----------	--------------------	-------

\* 4 n-Heptyl Alcohol  
 6.248    6.254    -0.006    2787818    50.0

**Reagents:**

SG\_GLY\_ISTD\_00129      Amount Added: 10.00      Units: uL      Run Reagent



Eurofins Savannah

Data File: \\chromfs\Savannah\ChromData\CVGG2\20231212-91533.b\1IL12023.D

Injection Date: 12-Dec-2023 23:48:22

Instrument ID: CVGG2

Operator ID:

Lims ID: 580-134720-A-1

Lab Sample ID: 680-134720-1

Worklist Smp#: 23

Client ID: AF-RHMW17-WGN01LF-2312

Injection Vol: 1.0 ul


Dil. Factor: 1.0000

ALS Bottle#: 0

Method: 8015\_GLY\_VGG

Limit Group: 8015C\_DAI

Column: J&W DB WAX (0.45 mm)



FORM I  
GC SEMI VOA ORGANICS ANALYSIS DATA SHEET

Lab Name: Eurofins Savannah Job No.: 580-134720-1  
 SDG No.: \_\_\_\_\_  
 Client Sample ID: AF-RHMW17D-WGN01LF-2312 Lab Sample ID: 580-134720-2  
 Matrix: Water Lab File ID: 1IL12024.D  
 Analysis Method: 8015C GLY Date Collected: 12/06/2023 10:40  
 Extraction Method: \_\_\_\_\_ Date Extracted: \_\_\_\_\_  
 Sample wt/vol: 1(mL) Date Analyzed: 12/13/2023 00:11  
 Con. Extract Vol.: 1(mL) Dilution Factor: 1  
 Injection Volume: 1(uL) GC Column: J&W DB WAX ID: 0.45(mm)  
 % Moisture: \_\_\_\_\_ % Solids: \_\_\_\_\_ GPC Cleanup: (Y/N) N  
 Cleanup Factor: \_\_\_\_\_  
 Analysis Batch No.: 812978 Units: mg/L

CAS NO.	COMPOUND NAME	RESULT	Q	LOQ	LOD	DL
112-34-5	2-(2-Butoxyethoxy)ethanol	3.0	U Q	5.0	3.0	1.1

Eurofins Savannah  
Target Compound Quantitation Report

Data File: \\chromfs\Savannah\ChromData\CVGG2\20231212-91533.b\1IL12024.D  
 Lims ID: 580-134720-B-2  
 Client ID: AF-RHMW17D-WGN01LF-2312  
 Sample Type: Client  
 Inject. Date: 13-Dec-2023 00:11:42      ALS Bottle#: 0      Worklist Smp#: 24  
 Injection Vol: 1.0 ul      Dil. Factor: 1.0000  
 Sample Info: 680-0091533-024  
 Operator ID:      Instrument ID: CVGG2  
 Method: \\chromfs\Savannah\ChromData\CVGG2\20231212-91533.b\8015\_GLY\_VGG.m  
 Limit Group: 8015C\_DAI  
 Last Update: 13-Dec-2023 15:05:29      Calib Date: 15-Nov-2023 17:44:20  
 Integrator: Falcon  
 Quant Method: Internal Standard      Quant By: Initial Calibration  
 Last ICal File: \\chromfs\Savannah\ChromData\CVGG2\20231115-90783.b\1IK15012.D  
 Column 1 : J&W DB WAX ( 0.45 mm)      Det: GC FID2B  
 Process Host: CTX1647

First Level Reviewer: AR8P      Date: 13-Dec-2023 15:05:00

RT (min.)	Exp RT (min.)	Dlt RT (min.)	Response	OnCol Amt ug/ml	Flags
--------------	------------------	------------------	----------	--------------------	-------

\* 4 n-Heptyl Alcohol  
 6.253    6.254    -0.001    1546301    50.0

**Reagents:**

SG\_GLY\_ISTD\_00129      Amount Added: 10.00      Units: uL      Run Reagent

Eurofins Savannah

Data File: \\chromfs\Savannah\ChromData\CVGG2\20231212-91533.b\1IL12024.D

Injection Date: 13-Dec-2023 00:11:42

Instrument ID: CVGG2

Operator ID:

Lims ID: 580-134720-B-2

Lab Sample ID: 680-134720-2

Worklist Smp#: 24

Client ID: AF-RHMW17D-WGN01LF-2312

Injection Vol: 1.0 ul


Dil. Factor: 1.0000

ALS Bottle#: 0

Method: 8015\_GLY\_VGG

Limit Group: 8015C\_DAI

Column: J&W DB WAX (0.45 mm)



FORM I  
GC SEMI VOA ORGANICS ANALYSIS DATA SHEET

Lab Name: Eurofins Savannah Job No.: 580-134720-1  
 SDG No.: \_\_\_\_\_  
 Client Sample ID: AF-RHWW17D-WQFB01-2312 Lab Sample ID: 580-134720-3  
 Matrix: Water Lab File ID: 1IL12025.D  
 Analysis Method: 8015C GLY Date Collected: 12/06/2023 10:10  
 Extraction Method: \_\_\_\_\_ Date Extracted: \_\_\_\_\_  
 Sample wt/vol: 1(mL) Date Analyzed: 12/13/2023 00:35  
 Con. Extract Vol.: 1(mL) Dilution Factor: 1  
 Injection Volume: 1(uL) GC Column: J&W DB WAX ID: 0.45(mm)  
 % Moisture: \_\_\_\_\_ % Solids: \_\_\_\_\_ GPC Cleanup: (Y/N) N  
 Cleanup Factor: \_\_\_\_\_  
 Analysis Batch No.: 812978 Units: mg/L

CAS NO.	COMPOUND NAME	RESULT	Q	LOQ	LOD	DL
112-34-5	2-(2-Butoxyethoxy)ethanol	3.0	U Q	5.0	3.0	1.1

Eurofins Savannah  
Target Compound Quantitation Report

Data File: \\chromfs\Savannah\ChromData\CVGG2\20231212-91533.b\1IL12025.D  
 Lims ID: 580-134720-A-3  
 Client ID: AF-RHMW17D-WQFB01-2312  
 Sample Type: Client  
 Inject. Date: 13-Dec-2023 00:35:11      ALS Bottle#: 0      Worklist Smp#: 25  
 Injection Vol: 1.0 ul      Dil. Factor: 1.0000  
 Sample Info: 680-0091533-025  
 Operator ID:      Instrument ID: CVGG2

Method: \\chromfs\Savannah\ChromData\CVGG2\20231212-91533.b\8015\_GLY\_VGG.m  
 Limit Group: 8015C\_DAI  
 Last Update: 13-Dec-2023 15:05:29      Calib Date: 15-Nov-2023 17:44:20  
 Integrator: Falcon  
 Quant Method: Internal Standard      Quant By: Initial Calibration  
 Last ICal File: \\chromfs\Savannah\ChromData\CVGG2\20231115-90783.b\1IK15012.D  
 Column 1 : J&W DB WAX ( 0.45 mm)      Det: GC FID2B  
 Process Host: CTX1647

RT (min.)	Exp RT (min.)	Dlt RT (min.)	Response	OnCol Amt ug/ml	Flags
--------------	------------------	------------------	----------	--------------------	-------

\* 4 n-Heptyl Alcohol  
 6.251    6.254    -0.003    3323054    50.0

**Reagents:**

SG\_GLY\_ISTD\_00129      Amount Added: 10.00      Units: uL      Run Reagent

Eurofins Savannah

Data File: \\chromfs\Savannah\ChromData\CVGG2\20231212-91533.b\1L12025.D

Injection Date: 13-Dec-2023 00:35:11

Instrument ID: CVGG2

Operator ID:

Lims ID: 580-134720-A-3

Lab Sample ID: 680-134720-3

Worklist Smp#: 25

Client ID: AF-RHMW17D-WQFB01-2312

Injection Vol: 1.0 ul


Dil. Factor: 1.0000

ALS Bottle#: 0

Method: 8015\_GLY\_VGG

Limit Group: 8015C\_DAI

Column: J&W DB WAX (0.45 mm)



FORM I  
GC SEMI VOA ORGANICS ANALYSIS DATA SHEET

Lab Name: Eurofins Savannah Job No.: 580-134720-1  
 SDG No.: \_\_\_\_\_  
 Client Sample ID: AF-RHWW17S-WGN01LF-2312 Lab Sample ID: 580-134720-4  
 Matrix: Water Lab File ID: 1IL12026.D  
 Analysis Method: 8015C GLY Date Collected: 12/06/2023 09:05  
 Extraction Method: \_\_\_\_\_ Date Extracted: \_\_\_\_\_  
 Sample wt/vol: 1(mL) Date Analyzed: 12/13/2023 00:58  
 Con. Extract Vol.: 1(mL) Dilution Factor: 1  
 Injection Volume: 1(uL) GC Column: J&W DB WAX ID: 0.45(mm)  
 % Moisture: \_\_\_\_\_ % Solids: \_\_\_\_\_ GPC Cleanup: (Y/N) N  
 Cleanup Factor: \_\_\_\_\_  
 Analysis Batch No.: 812978 Units: mg/L

CAS NO.	COMPOUND NAME	RESULT	Q	LOQ	LOD	DL
112-34-5	2-(2-Butoxyethoxy)ethanol	3.0	U Q	5.0	3.0	1.1



Eurofins Savannah  
Target Compound Quantitation Report

Data File: \\chromfs\Savannah\ChromData\CVGG2\20231212-91533.b\1IL12026.D  
 Lims ID: 580-134720-A-4  
 Client ID: AF-RHMW17S-WGN01LF-2312  
 Sample Type: Client  
 Inject. Date: 13-Dec-2023 00:58:32      ALS Bottle#: 0      Worklist Smp#: 26  
 Injection Vol: 1.0 ul      Dil. Factor: 1.0000  
 Sample Info: 680-0091533-026  
 Operator ID:      Instrument ID: CVGG2

Method: \\chromfs\Savannah\ChromData\CVGG2\20231212-91533.b\8015\_GLY\_VGG.m  
 Limit Group: 8015C\_DAI  
 Last Update: 13-Dec-2023 15:05:29      Calib Date: 15-Nov-2023 17:44:20  
 Integrator: Falcon  
 Quant Method: Internal Standard      Quant By: Initial Calibration  
 Last ICal File: \\chromfs\Savannah\ChromData\CVGG2\20231115-90783.b\1IK15012.D  
 Column 1 : J&W DB WAX ( 0.45 mm)      Det: GC FID2B  
 Process Host: CTX1647

RT (min.)	Exp RT (min.)	Dlt RT (min.)	Response	OnCol Amt ug/ml	Flags
--------------	------------------	------------------	----------	--------------------	-------

\* 4 n-Heptyl Alcohol  
 6.251    6.254    -0.003    2739650    50.0

**Reagents:**

SG\_GLY\_ISTD\_00129      Amount Added: 10.00      Units: uL      Run Reagent

Eurofins Savannah

Data File: \\chromfs\Savannah\ChromData\CVGG2\20231212-91533.b\1IL12026.D

Injection Date: 13-Dec-2023 00:58:32

Instrument ID: CVGG2

Operator ID:

Lims ID: 580-134720-A-4

Lab Sample ID: 680-134720-4

Worklist Smp#: 26

Client ID: AF-RHMW17S-WGN01LF-2312

Injection Vol: 1.0 ul


Dil. Factor: 1.0000

ALS Bottle#: 0

Method: 8015\_GLY\_VGG

Limit Group: 8015C\_DAI

Column: J&W DB WAX (0.45 mm)



FORM I  
GC SEMI VOA ORGANICS ANALYSIS DATA SHEET

Lab Name: Eurofins Savannah Job No.: 580-134720-1  
 SDG No.: \_\_\_\_\_  
 Client Sample ID: AF-RHWW17S-WQEB01-2312 Lab Sample ID: 580-134720-5  
 Matrix: Water Lab File ID: 1IL12027.D  
 Analysis Method: 8015C GLY Date Collected: 12/06/2023 09:40  
 Extraction Method: \_\_\_\_\_ Date Extracted: \_\_\_\_\_  
 Sample wt/vol: 1(mL) Date Analyzed: 12/13/2023 01:21  
 Con. Extract Vol.: 1(mL) Dilution Factor: 1  
 Injection Volume: 1(uL) GC Column: J&W DB WAX ID: 0.45(mm)  
 % Moisture: \_\_\_\_\_ % Solids: \_\_\_\_\_ GPC Cleanup: (Y/N) N  
 Cleanup Factor: \_\_\_\_\_  
 Analysis Batch No.: 812978 Units: mg/L

CAS NO.	COMPOUND NAME	RESULT	Q	LOQ	LOD	DL
112-34-5	2-(2-Butoxyethoxy)ethanol	3.0	U Q	5.0	3.0	1.1

Eurofins Savannah  
Target Compound Quantitation Report

Data File: \\chromfs\Savannah\ChromData\CVGG2\20231212-91533.b\1IL12027.D  
 Lims ID: 580-134720-A-5  
 Client ID: AF-RHMW17S-WQEB01-2312  
 Sample Type: Client  
 Inject. Date: 13-Dec-2023 01:21:51      ALS Bottle#: 0      Worklist Smp#: 27  
 Injection Vol: 1.0 ul      Dil. Factor: 1.0000  
 Sample Info: 680-0091533-027  
 Operator ID:      Instrument ID: CVGG2

Method: \\chromfs\Savannah\ChromData\CVGG2\20231212-91533.b\8015\_GLY\_VGG.m  
 Limit Group: 8015C\_DAI  
 Last Update: 13-Dec-2023 15:05:29      Calib Date: 15-Nov-2023 17:44:20  
 Integrator: Falcon  
 Quant Method: Internal Standard      Quant By: Initial Calibration  
 Last ICal File: \\chromfs\Savannah\ChromData\CVGG2\20231115-90783.b\1IK15012.D  
 Column 1 : J&W DB WAX ( 0.45 mm)      Det: GC FID2B  
 Process Host: CTX1647

RT (min.)	Exp RT (min.)	Dlt RT (min.)	Response	OnCol Amt ug/ml	Flags
--------------	------------------	------------------	----------	--------------------	-------

\* 4 n-Heptyl Alcohol  
 6.245    6.254    -0.009    3229024    50.0

**Reagents:**

SG\_GLY\_ISTD\_00129      Amount Added: 10.00      Units: uL      Run Reagent

Eurofins Savannah

Data File: \\chromfs\Savannah\ChromData\CVGG2\20231212-91533.b\1IL12027.D

Injection Date: 13-Dec-2023 01:21:51

Instrument ID: CVGG2

Operator ID:

Lims ID: 580-134720-A-5

Lab Sample ID: 680-134720-5

Worklist Smp#: 27

Client ID: AF-RHMW17S-WQEB01-2312

Injection Vol: 1.0 ul


Dil. Factor: 1.0000

ALS Bottle#: 0

Method: 8015\_GLY\_VGG

Limit Group: 8015C\_DAI

Column: J&W DB WAX (0.45 mm)



FORM I  
GC SEMI VOA ORGANICS ANALYSIS DATA SHEET

Lab Name: Eurofins Savannah Job No.: 580-134720-1  
 SDG No.: \_\_\_\_\_  
 Client Sample ID: AF-RHMW12A-WGN01LF-2312 Lab Sample ID: 580-134720-6  
 Matrix: Water Lab File ID: 1IL12028.D  
 Analysis Method: 8015C GLY Date Collected: 12/07/2023 09:15  
 Extraction Method: \_\_\_\_\_ Date Extracted: \_\_\_\_\_  
 Sample wt/vol: 1(mL) Date Analyzed: 12/13/2023 01:45  
 Con. Extract Vol.: 1(mL) Dilution Factor: 1  
 Injection Volume: 1(uL) GC Column: J&W DB WAX ID: 0.45(mm)  
 % Moisture: \_\_\_\_\_ % Solids: \_\_\_\_\_ GPC Cleanup: (Y/N) N  
 Cleanup Factor: \_\_\_\_\_  
 Analysis Batch No.: 812978 Units: mg/L

CAS NO.	COMPOUND NAME	RESULT	Q	LOQ	LOD	DL
112-34-5	2-(2-Butoxyethoxy)ethanol	3.0	U Q	5.0	3.0	1.1

Eurofins Savannah  
Target Compound Quantitation Report

Data File: \\chromfs\Savannah\ChromData\CVGG2\20231212-91533.b\1IL12028.D  
 Lims ID: 580-134720-A-6  
 Client ID: AF-RHMW12A-WGN01LF-2312  
 Sample Type: Client  
 Inject. Date: 13-Dec-2023 01:45:16      ALS Bottle#: 0      Worklist Smp#: 28  
 Injection Vol: 1.0 ul      Dil. Factor: 1.0000  
 Sample Info: 680-0091533-028  
 Operator ID:      Instrument ID: CVGG2  
  
 Method: \\chromfs\Savannah\ChromData\CVGG2\20231212-91533.b\8015\_GLY\_VGG.m  
 Limit Group: 8015C\_DAI  
 Last Update: 13-Dec-2023 15:05:29      Calib Date: 15-Nov-2023 17:44:20  
 Integrator: Falcon  
 Quant Method: Internal Standard      Quant By: Initial Calibration  
 Last ICal File: \\chromfs\Savannah\ChromData\CVGG2\20231115-90783.b\1IK15012.D  
  
 Column 1 : J&W DB WAX ( 0.45 mm)      Det: GC FID2B  
 Process Host: CTX1647

RT (min.)	Exp RT (min.)	Dlt RT (min.)	Response	OnCol Amt ug/ml	Flags
--------------	------------------	------------------	----------	--------------------	-------

\* 4 n-Heptyl Alcohol  
 6.253    6.254    -0.001    2010591    50.0

**Reagents:**

SG\_GLY\_ISTD\_00129      Amount Added: 10.00      Units: uL      Run Reagent

Eurofins Savannah

Data File: \\chromfs\Savannah\ChromData\CVGG2\20231212-91533.b\1IL12028.D

Injection Date: 13-Dec-2023 01:45:16

Instrument ID: CVGG2

Operator ID:

Lims ID: 580-134720-A-6

Lab Sample ID: 680-134720-6

Worklist Smp#: 28

Client ID: AF-RHMW12A-WGN01LF-2312

Injection Vol: 1.0 ul


Dil. Factor: 1.0000

ALS Bottle#: 0

Method: 8015\_GLY\_VGG

Limit Group: 8015C\_DAI

Column: J&W DB WAX (0.45 mm)





FORM I  
GC SEMI VOA ORGANICS ANALYSIS DATA SHEET

Lab Name: Eurofins Savannah Job No.: 580-134720-1  
 SDG No.: \_\_\_\_\_  
 Client Sample ID: AF-RHWW12A-WGFD01LF-2312 Lab Sample ID: 580-134720-7  
 Matrix: Water Lab File ID: 1IL12029.D  
 Analysis Method: 8015C GLY Date Collected: 12/07/2023 09:15  
 Extraction Method: \_\_\_\_\_ Date Extracted: \_\_\_\_\_  
 Sample wt/vol: 1(mL) Date Analyzed: 12/13/2023 02:08  
 Con. Extract Vol.: 1(mL) Dilution Factor: 1  
 Injection Volume: 1(uL) GC Column: J&W DB WAX ID: 0.45(mm)  
 % Moisture: \_\_\_\_\_ % Solids: \_\_\_\_\_ GPC Cleanup: (Y/N) N  
 Cleanup Factor: \_\_\_\_\_  
 Analysis Batch No.: 812978 Units: mg/L

CAS NO.	COMPOUND NAME	RESULT	Q	LOQ	LOD	DL
112-34-5	2-(2-Butoxyethoxy)ethanol	3.0	U Q	5.0	3.0	1.1

Eurofins Savannah  
Target Compound Quantitation Report

Data File: \\chromfs\Savannah\ChromData\CVGG2\20231212-91533.b\1IL12029.D  
 Lims ID: 580-134720-B-7  
 Client ID: AF-RHMW12A-WGFD01LF-2312  
 Sample Type: Client  
 Inject. Date: 13-Dec-2023 02:08:36 ALS Bottle#: 0 Worklist Smp#: 29  
 Injection Vol: 1.0 ul Dil. Factor: 1.0000  
 Sample Info: 680-0091533-029  
 Operator ID: Instrument ID: CVGG2

Method: \\chromfs\Savannah\ChromData\CVGG2\20231212-91533.b\8015\_GLY\_VGG.m  
 Limit Group: 8015C\_DAI  
 Last Update: 13-Dec-2023 15:05:29 Calib Date: 15-Nov-2023 17:44:20  
 Integrator: Falcon  
 Quant Method: Internal Standard Quant By: Initial Calibration  
 Last ICal File: \\chromfs\Savannah\ChromData\CVGG2\20231115-90783.b\1IK15012.D  
 Column 1 : J&W DB WAX ( 0.45 mm) Det: GC FID2B  
 Process Host: CTX1647

RT (min.)	Exp RT (min.)	Dlt RT (min.)	Response	OnCol Amt ug/ml	Flags
--------------	------------------	------------------	----------	--------------------	-------

\* 4 n-Heptyl Alcohol  
 6.255 6.254 0.001 2277287 50.0

Reagents:

SG\_GLY\_ISTD\_00129 Amount Added: 10.00 Units: uL Run Reagent

Eurofins Savannah

Data File: \\chromfs\Savannah\ChromData\CVGG2\20231212-91533.b\1L12029.D

Injection Date: 13-Dec-2023 02:08:36

Instrument ID: CVGG2

Operator ID:

Lims ID: 580-134720-B-7

Lab Sample ID: 680-134720-7

Worklist Smp#: 29

Client ID: AF-RHMW12A-WGFD01LF-2312

Injection Vol: 1.0 ul


Dil. Factor: 1.0000

ALS Bottle#: 0

Method: 8015\_GLY\_VGG

Limit Group: 8015C\_DAI

Column: J&W DB WAX (0.45 mm)



FORM VI  
GC SEMI VOA BY INTERNAL STANDARD - INITIAL CALIBRATION DATA  
CURVE EVALUATION

Lab Name: Eurofins Savannah Job No.: 580-134720-1 Analy Batch No.: 808404

SDG No.: \_\_\_\_\_

Instrument ID: CVGG2 GC Column: J&W DB WAX ID: 0.45 (mm) Heated Purge: (Y/N) N

Calibration Start Date: 11/15/2023 15:23 Calibration End Date: 11/15/2023 17:44 Calibration ID: 92674

Calibration Files

LEVEL:	LAB SAMPLE ID:	LAB FILE ID:
Level 1	IC 680-808404/12	1IK15012.D
Level 2	IC 680-808404/11	1IK15011.D
Level 3	IC 680-808404/10	1IK15010.D
Level 4	ICIS 680-808404/9	1IK15009.D
Level 5	IC 680-808404/8	1IK15008.D
Level 6	IC 680-808404/7	1IK15007.D
Level 7	IC 680-808404/6	1IK15006.D

ANALYTE	RRF					CURVE TYPE	COEFFICIENT			#	MIN RRF	%RSD /RSE	#	MAX %RSD /RSE	R^2 OR COD	#	MIN R^2 OR COD
	LVL 1	LVL 2	LVL 3	LVL 4	LVL 5		B	M1	M2								
	LVL 6	LVL 7															
Ethanol, 2-propoxy	1.1302 ++++	0.7908 0.5926	++++	0.6956	0.6816	Lin2	0.995 9	0.622 8						0.9960		0.9900	
4-Hydroxy-4-methyl-2-pentanone	1.1112 ++++	0.7643 0.5781	++++	0.6600	0.6461	Lin2	1.013 6	0.593 8						0.9970		0.9900	
2-Butoxyethanol	1.2800 ++++	0.8884 0.6678	++++	0.7869	0.7698	Lin2	1.131 2	0.702 1						0.9950		0.9900	
Dipropylene Glycol Methyl Ether	0.1210 ++++	0.0739 0.0492	++++	0.0570	0.0529	Lin2	0.142 8	0.048 5						0.9980		0.9900	
Propylene glycol	0.8828 ++++	0.5811 0.4192	++++	0.4821	0.4062	Lin2	0.943 9	0.407 0						0.9970		0.9900	
Ethylene glycol	0.6653 ++++	0.4183 0.3134	++++	0.3652	0.3000	Lin2	0.712 0	0.301 4						0.9930		0.9900	
2-(2-Butoxyethoxy)ethanol	1.1471 ++++	0.7679 0.5480	++++	0.6196	0.5767	Lin2	1.192 4	0.545 9						0.9990		0.9900	
2,2'-Oxybisethanol	0.7807 ++++	0.4897 0.3378	++++	0.4078	0.3226	Lin2	0.897 1	0.327 8						0.9930		0.9900	
Triethylene Glycol	1.2266 ++++	0.6725 0.3849	++++	0.4728	0.3668	Lin2	1.723 9	0.355 8						0.9930		0.9900	
Tetraethylene Glycol	0.7840 ++++	0.4933 0.3409	++++	0.4186	0.3306	Lin2	1.778 0	0.334 8						0.9930		0.9900	

Note: The M1 coefficient is the same as Ave RRF for an Ave curve type. RSD is calculated for Ave curve types. RSE is used for all other types.

FORM VI  
GC SEMI VOA BY INTERNAL STANDARD - INITIAL CALIBRATION DATA  
RESPONSE AND CONCENTRATION

Lab Name: Eurofins Savannah Job No.: 580-134720-1 Analy Batch No.: 808404

SDG No.: \_\_\_\_\_

Instrument ID: CVGG2 GC Column: J&W DB WAX ID: 0.45(mm) Heated Purge: (Y/N) N

Calibration Start Date: 11/15/2023 15:23 Calibration End Date: 11/15/2023 17:44 Calibration ID: 92674

Calibration Files:

LEVEL:	LAB SAMPLE ID:	LAB FILE ID:
Level 1	IC 680-808404/12	1IK15012.D
Level 2	IC 680-808404/11	1IK15011.D
Level 3	IC 680-808404/10	1IK15010.D
Level 4	ICIS 680-808404/9	1IK15009.D
Level 5	IC 680-808404/8	1IK15008.D
Level 6	IC 680-808404/7	1IK15007.D
Level 7	IC 680-808404/6	1IK15006.D

ANALYTE	IS REF	CURVE TYPE	RESPONSE					CONCENTRATION (UG/ML)				
			LVL 1 LVL 6	LVL 2 LVL 7	LVL 3	LVL 4	LVL 5	LVL 1 LVL 6	LVL 2 LVL 7	LVL 3	LVL 4	LVL 5
Ethanol, 2-propoxy	nHPA	Lin2	89414 ++++	175351 3100269	++++	697507	1795383	2.00 ++++	5.00 100	++++	20.0	50.0
4-Hydroxy-4-methyl-2-pentanone	nHPA	Lin2	87911 ++++	169482 3024238	++++	661829	1701776	2.00 ++++	5.00 100	++++	20.0	50.0
2-Butoxyethanol	nHPA	Lin2	101265 ++++	197005 3493941	++++	789027	2027645	2.00 ++++	5.00 100	++++	20.0	50.0
Dipropylene Glycol Methyl Ether	nHPA	Lin2	9574 ++++	16390 257340	++++	57169	139230	2.00 ++++	5.00 100	++++	20.0	50.0
Propylene glycol	nHPA	Lin2	69847 ++++	128858 2193222	++++	483381	1070046	2.00 ++++	5.00 100	++++	20.0	50.0
Ethylene glycol	nHPA	Lin2	52632 ++++	92757 1639426	++++	366220	790234	2.00 ++++	5.00 100	++++	20.0	50.0
2-(2-Butoxyethoxy)ethanol	nHPA	Lin2	90755 ++++	170281 2866855	++++	621252	1519207	2.00 ++++	5.00 100	++++	20.0	50.0
2,2'-Oxybisethanol	nHPA	Lin2	61764 ++++	108584 1767042	++++	408918	849792	2.00 ++++	5.00 100	++++	20.0	50.0
Triethylene Glycol	nHPA	Lin2	97042 ++++	149118 2013907	++++	474118	966093	2.00 ++++	5.00 100	++++	20.0	50.0
Tetraethylene Glycol	nHPA	Lin2	124057 ++++	218792 3567157	++++	839423	1741870	4.00 ++++	10.0 200	++++	40.0	100

Curve Type Legend:

Lin2 = Linear 1/conc^2 ISTD

FORM VI  
GC SEMI VOA BY INTERNAL STANDARD - INITIAL CALIBRATION DATA  
READBACK PERCENT ERROR

Lab Name: Eurofins Savannah Job No.: 580-134720-1 Analy Batch No.: 808404

SDG No.: \_\_\_\_\_

Instrument ID: CVGG2 GC Column: J&W DB WAX ID: 0.45 (mm) Heated Purge: (Y/N) N

Calibration Start Date: 11/15/2023 15:23 Calibration End Date: 11/15/2023 17:44 Calibration ID: 92674

Calibration Files:

LEVEL:	LAB SAMPLE ID:	LAB FILE ID:
Level 1	IC 680-808404/12	1IK15012.D
Level 2	IC 680-808404/11	1IK15011.D
Level 3	IC 680-808404/10	1IK15010.D
Level 4	ICIS 680-808404/9	1IK15009.D
Level 5	IC 680-808404/8	1IK15008.D
Level 6	IC 680-808404/7	1IK15007.D
Level 7	IC 680-808404/6	1IK15006.D

ANALYTE	PERCENT ERROR						PERCENT ERROR LIMIT					
	LVL 1 # LVL 7 #	LVL 2 #	LVL 3 #	LVL 4 #	LVL 5 #	LVL 6 #	LVL 1 LVL 7	LVL 2	LVL 3	LVL 4	LVL 5	LVL 6
Ethanol, 2-propoxy	1.5 -6.4	-5.0	++++	3.7	6.2	++++	20 20	20		20	20	
4-Hydroxy-4-methyl-2-pentanone	1.8 -4.4	-5.4	++++	2.6	5.4	++++	20 20	20		20	20	
2-Butoxyethanol	1.7 -6.5	-5.7	++++	4.0	6.4	++++	20 20	20		20	20	
Dipropylene Glycol Methyl Ether	2.2 -1.6	-6.5	++++	2.8	3.1	++++	20 20	20		20	20	
Propylene glycol	0.9 0.7	-3.6	++++	6.8	-4.8	++++	20 20	20		20	20	
Ethylene glycol	2.6 1.6	-8.4	++++	9.4	-5.2	++++	20 20	20		20	20	
2-(2-Butoxyethoxy)ethanol	0.9 -1.8	-3.0	++++	2.6	1.3	++++	20 20	20		20	20	
2,2'-Oxybisethanol	1.3 0.3	-5.3	++++	10.7	-7.0	++++	20 20	20		20	20	
Triethylene Glycol	2.5 3.3	-7.9	++++	8.7	-6.6	++++	20 20	20		20	20	
Tetraethylene Glycol	1.4 -0.8	-5.8	++++	11.7	-6.6	++++	20 20	20		20	20	

Eurofins Savannah  
Target Compound Quantitation Report

Data File: \\chromfs\Savannah\ChromData\CVGG2\20231115-90783.b\1IK15006.D  
 Lims ID: ic g7  
 Client ID:  
 Sample Type: IC Calib Level: 7  
 Inject. Date: 15-Nov-2023 15:23:43 ALS Bottle#: 0 Worklist Smp#: 6  
 Injection Vol: 1.0 ul Dil. Factor: 1.0000  
 Sample Info: 680-0090783-006  
 Operator ID: Instrument ID: CVGG2  
 Sublist: chrom-8015\_GLY\_VGG\*sub2  
 Method: \\chromfs\Savannah\ChromData\CVGG2\20231115-90783.b\8015\_GLY\_VGG.m  
 Limit Group: 8015C\_DAI  
 Last Update: 16-Nov-2023 11:39:04 Calib Date: 15-Nov-2023 17:44:20  
 Integrator: Falcon  
 Quant Method: Internal Standard Quant By: Initial Calibration  
 Last ICal File: \\chromfs\Savannah\ChromData\CVGG2\20231115-90783.b\1IK15012.D  
 Column 1 : J&W DB WAX ( 0.45 mm) Det: GC FID2B  
 Process Host: CTX1640

First Level Reviewer: AR8P Date: 15-Nov-2023 17:21:23

RT (min.)	Exp RT (min.)	Dlt RT (min.)	Response	Cal Amt ug/ml	OnCol Amt ug/ml	Flags
--------------	------------------	------------------	----------	------------------	--------------------	-------

1 Ethanol, 2-propoxy	4.471	4.471	0.000	3100269	100.0	93.6
2 4-Hydroxy-4-methyl-2-pentanone	5.258	5.258	0.000	3024238	100.0	95.6
3 2-Butoxyethanol	5.689	5.689	0.000	3493941	100.0	93.5
* 4 n-Heptyl Alcohol	6.278	6.278	0.000	2615890	50.0	50.0
5 Dipropylene Glycol Methyl Ether	7.326	7.326	0.000	257340	100.0	98.4
6 Propylene glycol	8.250	8.250	0.000	2193222	100.0	100.7
7 Ethylene glycol	8.700	8.700	0.000	1639426	100.0	101.6
8 2-(2-Butoxyethoxy)ethanol	9.808	9.808	0.000	2866855	100.0	98.2
9 2,2'-Oxybisethanol	10.504	10.504	0.000	1767042	100.0	100.3
10 Triethylene Glycol	11.697	11.697	0.000	2013907	100.0	103.3
11 Tetraethylene Glycol	13.989	13.989	0.000	3567157	200.0	198.3

QC Flag Legend  
Processing Flags

**Reagents:**

SG\_Gly\_CAL\_00053

Amount Added: 50.00

Units: uL

SG\_GLY\_ISTD\_00129

Amount Added: 10.00

Units: uL

Run Reagent



Eurofins Savannah

Data File: \\chromfs\Savannah\ChromData\CVGG2\20231115-90783.b\1IK15006.D

Injection Date: 15-Nov-2023 15:23:43

Instrument ID: CVGG2

Operator ID:

Lims ID: ic g7

Worklist Smp#: 6

Client ID:

Injection Vol: 1.0 ul


Dil. Factor: 1.0000

ALS Bottle#: 0

Method: 8015\_GLY\_VGG

Limit Group: 8015C\_DAI

Column: J&W DB WAX (0.45 mm)



Eurofins Savannah  
Target Compound Quantitation Report

Data File: \\chromfs\Savannah\ChromData\CVGG2\20231115-90783.b\1IK15007.D  
 Lims ID: ic g6  
 Client ID:  
 Sample Type: IC Calib Level: 6  
 Inject. Date: 15-Nov-2023 15:47:13 ALS Bottle#: 0 Worklist Smp#: 7  
 Injection Vol: 1.0 ul Dil. Factor: 1.0000  
 Sample Info: 680-0090783-007  
 Operator ID: Instrument ID: CVGG2  
 Sublist: chrom-8015\_GLY\_VGG\*sub2  
 Method: \\chromfs\Savannah\ChromData\CVGG2\20231115-90783.b\8015\_GLY\_VGG.m  
 Limit Group: 8015C\_DAI  
 Last Update: 16-Nov-2023 11:39:05 Calib Date: 15-Nov-2023 17:44:20  
 Integrator: Falcon  
 Quant Method: Internal Standard Quant By: Initial Calibration  
 Last ICal File: \\chromfs\Savannah\ChromData\CVGG2\20231115-90783.b\1IK15012.D  
 Column 1 : J&W DB WAX ( 0.45 mm) Det: GC FID2B  
 Process Host: CTX1640

First Level Reviewer: AR8P Date: 16-Nov-2023 11:09:45

RT (min.)	Exp RT (min.)	Dlt RT (min.)	Response	Cal Amt ug/ml	OnCol Amt ug/ml	Flags
--------------	------------------	------------------	----------	------------------	--------------------	-------

1 Ethanol, 2-propoxy	4.472	4.471	0.001	2511726	80.0	106.2
2 4-Hydroxy-4-methyl-2-pentanone	5.259	5.258	0.001	2406799	80.0	106.6
3 2-Butoxyethanol	5.689	5.689	0.000	2829474	80.0	106.1
* 4 n-Heptyl Alcohol	6.277	6.278	-0.001	1870888	50.0	50.0
5 Dipropylene Glycol Methyl Ether	7.327	7.326	0.001	199119	80.0	106.7
6 Propylene glycol	8.254	8.250	0.004	1598739	80.0	102.7
7 Ethylene glycol	8.701	8.700	0.001	1187415	80.0	102.9
8 2-(2-Butoxyethoxy)ethanol	9.808	9.808	0.000	2188973	80.0	105.0
9 2,2'-Oxybisethanol	10.504	10.504	0.000	1241858	80.0	98.5
10 Triethylene Glycol	11.698	11.697	0.001	1362169	80.0	97.5
11 Tetraethylene Glycol	13.988	13.989	-0.001	2497176	160.0	194.0

### QC Flag Legend

Processing Flags

**Reagents:**

SG\_Gly\_CAL\_00053

Amount Added: 40.00

Units: uL

SG\_GLY\_ISTD\_00129

Amount Added: 10.00

Units: uL

Run Reagent

Eurofins Savannah

Data File: \\chromfs\Savannah\ChromData\CVGG2\20231115-90783.b\1IK15007.D

Injection Date: 15-Nov-2023 15:47:13

Instrument ID: CVGG2

Operator ID:

Lims ID: ic g6

Worklist Smp#: 7

Client ID:

Injection Vol: 1.0 ul


Dil. Factor: 1.0000

ALS Bottle#: 0

Method: 8015\_GLY\_VGG

Limit Group: 8015C\_DAI

Column: J&W DB WAX (0.45 mm)



Eurofins Savannah  
Target Compound Quantitation Report

Data File: \\chromfs\Savannah\ChromData\CVGG2\20231115-90783.b\1IK15008.D  
 Lims ID: ic g5  
 Client ID:  
 Sample Type: IC Calib Level: 5  
 Inject. Date: 15-Nov-2023 16:10:37 ALS Bottle#: 0 Worklist Smp#: 8  
 Injection Vol: 1.0 ul Dil. Factor: 1.0000  
 Sample Info: 680-0090783-008  
 Operator ID: Instrument ID: CVGG2  
 Sublist: chrom-8015\_GLY\_VGG\*sub2  
 Method: \\chromfs\Savannah\ChromData\CVGG2\20231115-90783.b\8015\_GLY\_VGG.m  
 Limit Group: 8015C\_DAI  
 Last Update: 16-Nov-2023 11:39:07 Calib Date: 15-Nov-2023 17:44:20  
 Integrator: Falcon  
 Quant Method: Internal Standard Quant By: Initial Calibration  
 Last ICal File: \\chromfs\Savannah\ChromData\CVGG2\20231115-90783.b\1IK15012.D  
 Column 1 : J&W DB WAX ( 0.45 mm) Det: GC FID2B  
 Process Host: CTX1640

RT (min.)	Exp RT (min.)	Dlt RT (min.)	Response	Cal Amt ug/ml	OnCol Amt ug/ml	Flags
-----------	---------------	---------------	----------	---------------	-----------------	-------

1 Ethanol, 2-propoxy	4.475	4.471	0.004	1795383	50.0	53.1
2 4-Hydroxy-4-methyl-2-pentanone	5.262	5.258	0.004	1701776	50.0	52.7
3 2-Butoxyethanol	5.689	5.689	0.000	2027645	50.0	53.2
* 4 n-Heptyl Alcohol	6.275	6.278	-0.003	2634100	50.0	50.0
5 Dipropylene Glycol Methyl Ether	7.329	7.326	0.003	139230	50.0	51.5
6 Propylene glycol	8.256	8.250	0.006	1070046	50.0	47.6
7 Ethylene glycol	8.703	8.700	0.003	790234	50.0	47.4
8 2-(2-Butoxyethoxy)ethanol	9.808	9.808	0.000	1519207	50.0	50.6
9 2,2'-Oxybisethanol	10.505	10.504	0.001	849792	50.0	46.5
10 Triethylene Glycol	11.698	11.697	0.001	966093	50.0	46.7
11 Tetraethylene Glycol	13.990	13.989	0.001	1741870	100.0	93.4

Reagents:

SG\_Gly\_CAL\_00053 Amount Added: 25.00 Units: uL  
 SG\_GLY\_ISTD\_00129 Amount Added: 10.00 Units: uL Run Reagent

Eurofins Savannah

Data File: \\chromfs\Savannah\ChromData\CVGG2\20231115-90783.b\1IK15008.D

Injection Date: 15-Nov-2023 16:10:37

Instrument ID: CVGG2

Operator ID:

Lims ID: ic g5

Worklist Smp#: 8

Client ID:

Injection Vol: 1.0 ul


Dil. Factor: 1.0000

ALS Bottle#: 0

Method: 8015\_GLY\_VGG

Limit Group: 8015C\_DAI

Column: J&W DB WAX (0.45 mm)



Eurofins Savannah  
Target Compound Quantitation Report

Data File: \\chromfs\Savannah\ChromData\CVGG2\20231115-90783.b\1IK15009.D  
 Lims ID: icis g4  
 Client ID:  
 Sample Type: ICIS Calib Level: 4  
 Inject. Date: 15-Nov-2023 16:34:03 ALS Bottle#: 0 Worklist Smp#: 9  
 Injection Vol: 1.0 ul Dil. Factor: 1.0000  
 Sample Info: 680-0090783-009  
 Operator ID: Instrument ID: CVGG2  
 Sublist: chrom-8015\_GLY\_VGG\*sub2  
 Method: \\chromfs\Savannah\ChromData\CVGG2\20231115-90783.b\8015\_GLY\_VGG.m  
 Limit Group: 8015C\_DAI  
 Last Update: 16-Nov-2023 11:39:08 Calib Date: 15-Nov-2023 17:44:20  
 Integrator: Falcon  
 Quant Method: Internal Standard Quant By: Initial Calibration  
 Last ICal File: \\chromfs\Savannah\ChromData\CVGG2\20231115-90783.b\1IK15012.D  
 Column 1 : J&W DB WAX ( 0.45 mm) Det: GC FID2B  
 Process Host: CTX1640

RT (min.)	Exp RT (min.)	Dlt RT (min.)	Response	Cal Amt ug/ml	OnCol Amt ug/ml	Flags
-----------	---------------	---------------	----------	---------------	-----------------	-------

1 Ethanol, 2-propoxy	4.474	4.474	0.000	697507	20.0	20.7
2 4-Hydroxy-4-methyl-2-pentanone	5.260	5.260	0.000	661829	20.0	20.5
3 2-Butoxyethanol	5.690	5.690	0.000	789027	20.0	20.8
* 4 n-Heptyl Alcohol	6.278	6.278	0.000	2506763	50.0	50.0
5 Dipropylene Glycol Methyl Ether	7.328	7.328	0.000	57169	20.0	20.6
6 Propylene glycol	8.258	8.258	0.000	483381	20.0	21.4
7 Ethylene glycol	8.703	8.703	0.000	366220	20.0	21.9
8 2-(2-Butoxyethoxy)ethanol	9.807	9.807	0.000	621252	20.0	20.5
9 2,2'-Oxybisethanol	10.504	10.504	0.000	408918	20.0	22.1
10 Triethylene Glycol	11.698	11.698	0.000	474118	20.0	21.7
11 Tetraethylene Glycol	13.988	13.988	0.000	839423	40.0	44.7

Reagents:

SG\_Gly\_CAL\_00053 Amount Added: 10.00 Units: uL  
 SG\_GLY\_ISTD\_00129 Amount Added: 10.00 Units: uL Run Reagent

Eurofins Savannah

Data File: \\chromfs\Savannah\ChromData\CVGG2\20231115-90783.b\1IK15009.D

Injection Date: 15-Nov-2023 16:34:03

Instrument ID: CVGG2

Operator ID:

Lims ID: icis g4

Worklist Smp#: 9

Client ID:

Injection Vol: 1.0 ul


Dil. Factor: 1.0000

ALS Bottle#: 0

Method: 8015\_GLY\_VGG

Limit Group: 8015C\_DAI

Column: J&W DB WAX (0.45 mm)





Eurofins Savannah  
Target Compound Quantitation Report

Data File: \\chromfs\Savannah\ChromData\CVGG2\20231115-90783.b\1IK15010.D  
 Lims ID: ic g3  
 Client ID:  
 Sample Type: IC Calib Level: 3  
 Inject. Date: 15-Nov-2023 16:57:34 ALS Bottle#: 0 Worklist Smp#: 10  
 Injection Vol: 1.0 ul Dil. Factor: 1.0000  
 Sample Info: 680-0090783-010  
 Operator ID: Instrument ID: CVGG2  
 Sublist: chrom-8015\_GLY\_VGG\*sub2  
 Method: \\chromfs\Savannah\ChromData\CVGG2\20231115-90783.b\8015\_GLY\_VGG.m  
 Limit Group: 8015C\_DAI  
 Last Update: 16-Nov-2023 11:39:09 Calib Date: 15-Nov-2023 17:44:20  
 Integrator: Falcon  
 Quant Method: Internal Standard Quant By: Initial Calibration  
 Last ICal File: \\chromfs\Savannah\ChromData\CVGG2\20231115-90783.b\1IK15012.D  
 Column 1 : J&W DB WAX ( 0.45 mm) Det: GC FID2B  
 Process Host: CTX1640

RT (min.)	Exp RT (min.)	Dlt RT (min.)	Response	Cal Amt ug/ml	OnCol Amt ug/ml	Flags
-----------	---------------	---------------	----------	---------------	-----------------	-------

1 Ethanol, 2-propoxy	4.473	4.474	-0.001	316326	10.0	11.3
2 4-Hydroxy-4-methyl-2-pentanone	5.259	5.260	-0.001	315601	10.0	11.8
3 2-Butoxyethanol	5.691	5.690	0.001	352337	10.0	11.1
* 4 n-Heptyl Alcohol	6.279	6.278	0.001	1966905	50.0	50.0
5 Dipropylene Glycol Methyl Ether	7.326	7.328	-0.002	30938	10.0	13.3
6 Propylene glycol	8.256	8.258	-0.002	284510	10.0	15.4
7 Ethylene glycol	8.703	8.703	0.000	213013	10.0	15.6
8 2-(2-Butoxyethoxy)ethanol	9.807	9.807	0.000	333376	10.0	13.3
9 2,2'-Oxybisethanol	10.504	10.504	0.000	250607	10.0	16.7
10 Triethylene Glycol	11.697	11.698	-0.001	302489	10.0	16.8
11 Tetraethylene Glycol	13.986	13.988	-0.002	526705	20.0	34.7

Reagents:

SG\_Gly\_CAL\_00053 Amount Added: 5.00 Units: uL  
 SG\_GLY\_ISTD\_00129 Amount Added: 10.00 Units: uL Run Reagent

Eurofins Savannah

Data File: \\chromfs\Savannah\ChromData\CVGG2\20231115-90783.b\1IK15010.D

Injection Date: 15-Nov-2023 16:57:34

Instrument ID: CVGG2

Operator ID:

Lims ID: ic g3

Worklist Smp#: 10

Client ID:

Injection Vol: 1.0 ul


Dil. Factor: 1.0000

ALS Bottle#: 0

Method: 8015\_GLY\_VGG

Limit Group: 8015C\_DAI

Column: J&W DB WAX (0.45 mm)



Eurofins Savannah  
Target Compound Quantitation Report

Data File: \\chromfs\Savannah\ChromData\CVGG2\20231115-90783.b\1IK15011.D  
 Lims ID: ic g2  
 Client ID:  
 Sample Type: IC Calib Level: 2  
 Inject. Date: 15-Nov-2023 17:20:58 ALS Bottle#: 0 Worklist Smp#: 11  
 Injection Vol: 1.0 ul Dil. Factor: 1.0000  
 Sample Info: 680-0090783-011  
 Operator ID: Instrument ID: CVGG2  
 Sublist: chrom-8015\_GLY\_VGG\*sub2  
 Method: \\chromfs\Savannah\ChromData\CVGG2\20231115-90783.b\8015\_GLY\_VGG.m  
 Limit Group: 8015C\_DAI  
 Last Update: 16-Nov-2023 11:39:11 Calib Date: 15-Nov-2023 17:44:20  
 Integrator: Falcon  
 Quant Method: Internal Standard Quant By: Initial Calibration  
 Last ICal File: \\chromfs\Savannah\ChromData\CVGG2\20231115-90783.b\1IK15012.D  
 Column 1 : J&W DB WAX ( 0.45 mm) Det: GC FID2B  
 Process Host: CTX1640

RT (min.)	Exp RT (min.)	Dlt RT (min.)	Response	Cal Amt ug/ml	OnCol Amt ug/ml	Flags
-----------	---------------	---------------	----------	---------------	-----------------	-------

1 Ethanol, 2-propoxy	4.472	4.474	-0.002	175351	5.00	4.75
2 4-Hydroxy-4-methyl-2-pentanone	5.257	5.260	-0.003	169482	5.00	4.73
3 2-Butoxyethanol	5.688	5.690	-0.002	197005	5.00	4.72
* 4 n-Heptyl Alcohol	6.277	6.278	-0.001	2217440	50.0	50.0
5 Dipropylene Glycol Methyl Ether	7.325	7.328	-0.003	16390	5.00	4.67
6 Propylene glycol	8.257	8.258	-0.001	128858	5.00	4.82
7 Ethylene glycol	8.704	8.703	0.001	92757	5.00	4.58
8 2-(2-Butoxyethoxy)ethanol	9.807	9.807	0.000	170281	5.00	4.85
9 2,2'-Oxybisethanol	10.503	10.504	-0.001	108584	5.00	4.73
10 Triethylene Glycol	11.697	11.698	-0.001	149118	5.00	4.61
11 Tetraethylene Glycol	13.987	13.988	-0.001	218792	10.0	9.42

Reagents:

SG\_Gly\_CAL\_00053 Amount Added: 2.50 Units: uL  
 SG\_GLY\_ISTD\_00129 Amount Added: 10.00 Units: uL Run Reagent

Eurofins Savannah

Data File: \\chromfs\Savannah\ChromData\CVGG2\20231115-90783.b\1IK15011.D

Injection Date: 15-Nov-2023 17:20:58

Instrument ID: CVGG2

Operator ID:

Lims ID: ic g2

Worklist Smp#: 11

Client ID:

Injection Vol: 1.0 ul


Dil. Factor: 1.0000

ALS Bottle#: 0

Method: 8015\_GLY\_VGG

Limit Group: 8015C\_DAI

Column: J&W DB WAX (0.45 mm)



Eurofins Savannah  
Target Compound Quantitation Report

Data File: \\chromfs\Savannah\ChromData\CVGG2\20231115-90783.b\1IK15012.D  
 Lims ID: ic g1  
 Client ID:  
 Sample Type: IC Calib Level: 1  
 Inject. Date: 15-Nov-2023 17:44:20 ALS Bottle#: 0 Worklist Smp#: 12  
 Injection Vol: 1.0 ul Dil. Factor: 1.0000  
 Sample Info: 680-0090783-012  
 Operator ID: Instrument ID: CVGG2  
 Sublist: chrom-8015\_GLY\_VGG\*sub2  
 Method: \\chromfs\Savannah\ChromData\CVGG2\20231115-90783.b\8015\_GLY\_VGG.m  
 Limit Group: 8015C\_DAI  
 Last Update: 16-Nov-2023 11:39:12 Calib Date: 15-Nov-2023 17:44:20  
 Integrator: Falcon  
 Quant Method: Internal Standard Quant By: Initial Calibration  
 Last ICal File: \\chromfs\Savannah\ChromData\CVGG2\20231115-90783.b\1IK15012.D  
 Column 1 : J&W DB WAX ( 0.45 mm) Det: GC FID2B  
 Process Host: CTX1640

First Level Reviewer: AR8P Date: 15-Nov-2023 19:05:07

RT (min.)	Exp RT (min.)	Dlt RT (min.)	Response	Cal Amt ug/ml	OnCol Amt ug/ml	Flags
--------------	------------------	------------------	----------	------------------	--------------------	-------

1 Ethanol, 2-propoxy	4.474	4.474	0.000	89414	2.00	2.03
2 4-Hydroxy-4-methyl-2-pentanone	5.261	5.260	0.001	87911	2.00	2.04
3 2-Butoxyethanol	5.691	5.690	0.001	101265	2.00	2.03
* 4 n-Heptyl Alcohol	6.279	6.278	0.001	1977887	50.0	50.0
5 Dipropylene Glycol Methyl Ether	7.329	7.328	0.001	9574	2.00	2.04
6 Propylene glycol	8.258	8.258	0.000	69847	2.00	2.02
7 Ethylene glycol	8.708	8.703	0.005	52632	2.00	2.05
8 2-(2-Butoxyethoxy)ethanol	9.809	9.807	0.002	90755	2.00	2.02
9 2,2'-Oxybisethanol	10.505	10.504	0.001	61764	2.00	2.03
10 Triethylene Glycol	11.699	11.698	0.001	97042	2.00	2.05
11 Tetraethylene Glycol	13.989	13.988	0.001	124057	4.00	4.06

**QC Flag Legend**  
Processing Flags

Reagents:

SG\_Gly\_CAL\_00053

Amount Added: 1.00

Units: uL

SG\_GLY\_ISTD\_00129

Amount Added: 10.00

Units: uL

Run Reagent

Eurofins Savannah

Data File: \\chromfs\Savannah\ChromData\CVGG2\20231115-90783.b\1IK15012.D

Injection Date: 15-Nov-2023 17:44:20

Instrument ID: CVGG2

Operator ID:

Lims ID: ic g1

Worklist Smp#: 12

Client ID:

Injection Vol: 1.0 ul


Dil. Factor: 1.0000

ALS Bottle#: 0

Method: 8015\_GLY\_VGG

Limit Group: 8015C\_DAI

Column: J&W DB WAX (0.45 mm)



Calibration


/ Ethanol, 2-propoxy

Curve Type: Linear  
 Weighting: Conc\_Sq  
 Origin: None  
 Dependency: Response  
 Calib Mode: ISTD  
 Response Base: AREA  
 RF Rounding: 0

Curve Coefficients	
Intercept:	0.9959
Slope:	0.6228

Error Coefficients	
Standard Error:	2110000
Relative Standard Error:	6.4
Correlation Coefficient:	0.994
Coefficient of Determination (Adjusted):	0.996

ID	Level	Concentration	Rel. Resp.	IS Amount	IS Response	RRF	Used
1	IC 680-808404/12	2.0	2.260341	50.0	1977887.0	1.130171	Y
2	IC 680-808404/11	5.0	3.953906	50.0	2217440.0	0.790781	Y
3	IC 680-808404/10	10.0	8.041212	50.0	1966905.0	0.804121	N
4	ICIS 680-808404/9	20.0	13.912504	50.0	2506763.0	0.695625	Y
5	IC 680-808404/8	50.0	34.079629	50.0	2634100.0	0.681593	Y
6	IC 680-808404/7	80.0	67.126573	50.0	1870888.0	0.839082	N
7	IC 680-808404/6	100.0	59.258398	50.0	2615890.0	0.592584	Y





Calibration


/ 4-Hydroxy-4-methyl-2-pentanone

Curve Type: Linear  
 Weighting: Conc\_Sq  
 Origin: None  
 Dependency: Response  
 Calib Mode: ISTD  
 Response Base: AREA  
 RF Rounding: 0

Curve Coefficients	
Intercept:	1.014
Slope:	0.5938

Error Coefficients	
Standard Error:	2040000
Relative Standard Error:	5.4
Correlation Coefficient:	0.996
Coefficient of Determination (Adjusted):	0.997

ID	Level	Concentration	Rel. Resp.	IS Amount	IS Response	RRF	Used
1	IC 680-808404/12	2.0	2.222346	50.0	1977887.0	1.111173	Y
2	IC 680-808404/11	5.0	3.821569	50.0	2217440.0	0.764314	Y
3	IC 680-808404/10	10.0	8.022782	50.0	1966905.0	0.802278	N
4	ICIS 680-808404/9	20.0	13.200869	50.0	2506763.0	0.660043	Y
5	IC 680-808404/8	50.0	32.302798	50.0	2634100.0	0.646056	Y
6	IC 680-808404/7	80.0	64.32237	50.0	1870888.0	0.80403	N
7	IC 680-808404/6	100.0	57.805145	50.0	2615890.0	0.578051	Y



Calibration


/ 2-Butoxyethanol

Curve Type: Linear  
 Weighting: Conc\_Sq  
 Origin: None  
 Dependency: Response  
 Calib Mode: ISTD  
 Response Base: AREA  
 RF Rounding: 0

Curve Coefficients	
Intercept:	1.131
Slope:	0.7021

Error Coefficients	
Standard Error:	2380000
Relative Standard Error:	6.7
Correlation Coefficient:	0.994
Coefficient of Determination (Adjusted):	0.995

ID	Level	Concentration	Rel. Resp.	IS Amount	IS Response	RRF	Used
1	IC 680-808404/12	2.0	2.559929	50.0	1977887.0	1.279964	Y
2	IC 680-808404/11	5.0	4.442172	50.0	2217440.0	0.888434	Y
3	IC 680-808404/10	10.0	8.956635	50.0	1966905.0	0.895663	N
4	ICIS 680-808404/9	20.0	15.737966	50.0	2506763.0	0.786898	Y
5	IC 680-808404/8	50.0	38.488383	50.0	2634100.0	0.769768	Y
6	IC 680-808404/7	80.0	75.618476	50.0	1870888.0	0.945231	N
7	IC 680-808404/6	100.0	66.783026	50.0	2615890.0	0.66783	Y



Calibration


/ Dipropylene Glycol Methyl Ether

Curve Type: Linear  
 Weighting: Conc\_Sq  
 Origin: None  
 Dependency: Response  
 Calib Mode: ISTD  
 Response Base: AREA  
 RF Rounding: 0

Curve Coefficients	
Intercept:	0.1428
Slope:	0.04852

Error Coefficients	
Standard Error:	172000
Relative Standard Error:	4.7
Correlation Coefficient:	0.999
Coefficient of Determination (Adjusted):	0.998

ID	Level	Concentration	Rel. Resp.	IS Amount	IS Response	RRF	Used
1	IC 680-808404/12	2.0	0.242026	50.0	1977887.0	0.121013	Y
2	IC 680-808404/11	5.0	0.36957	50.0	2217440.0	0.073914	Y
3	IC 680-808404/10	10.0	0.786464	50.0	1966905.0	0.078646	N
4	ICIS 680-808404/9	20.0	1.140295	50.0	2506763.0	0.057015	Y
5	IC 680-808404/8	50.0	2.642838	50.0	2634100.0	0.052857	Y
6	IC 680-808404/7	80.0	5.32151	50.0	1870888.0	0.066519	N
7	IC 680-808404/6	100.0	4.918785	50.0	2615890.0	0.049188	Y



**Calibration**


/ Propylene glycol

Curve Type: Linear  
 Weighting: Conc\_Sq  
 Origin: None  
 Dependency: Response  
 Calib Mode: ISTD  
 Response Base: AREA  
 RF Rounding: 0

Curve Coefficients	
Intercept:	0.9439
Slope:	0.407

Error Coefficients	
Standard Error:	1440000
Relative Standard Error:	5.3
Correlation Coefficient:	0.999
Coefficient of Determination (Adjusted):	0.997

ID	Level	Concentration	Rel. Resp.	IS Amount	IS Response	RRF	Used
1	IC 680-808404/12	2.0	1.765697	50.0	1977887.0	0.882849	Y
2	IC 680-808404/11	5.0	2.905558	50.0	2217440.0	0.581112	Y
3	IC 680-808404/10	10.0	7.232429	50.0	1966905.0	0.723243	N
4	ICIS 680-808404/9	20.0	9.641538	50.0	2506763.0	0.482077	Y
5	IC 680-808404/8	50.0	20.311416	50.0	2634100.0	0.406228	Y
6	IC 680-808404/7	80.0	42.726743	50.0	1870888.0	0.534084	N
7	IC 680-808404/6	100.0	41.921143	50.0	2615890.0	0.419211	Y



**Calibration**


/ Ethylene glycol

Curve Type: Linear  
 Weighting: Conc\_Sq  
 Origin: None  
 Dependency: Response  
 Calib Mode: ISTD  
 Response Base: AREA  
 RF Rounding: 0

Curve Coefficients	
Intercept:	0.712
Slope:	0.3014

Error Coefficients	
Standard Error:	1070000
Relative Standard Error:	8.1
Correlation Coefficient:	0.999
Coefficient of Determination (Adjusted):	0.993

ID	Level	Concentration	Rel. Resp.	IS Amount	IS Response	RRF	Used
1	IC 680-808404/12	2.0	1.330511	50.0	1977887.0	0.665255	Y
2	IC 680-808404/11	5.0	2.091533	50.0	2217440.0	0.418307	Y
3	IC 680-808404/10	10.0	5.414929	50.0	1966905.0	0.541493	N
4	ICIS 680-808404/9	20.0	7.304639	50.0	2506763.0	0.365232	Y
5	IC 680-808404/8	50.0	15.000076	50.0	2634100.0	0.300002	Y
6	IC 680-808404/7	80.0	31.733995	50.0	1870888.0	0.396675	N
7	IC 680-808404/6	100.0	31.335912	50.0	2615890.0	0.313359	Y



Calibration


/ 2-(2-Butoxyethoxy)ethanol

Curve Type: Linear  
 Weighting: Conc\_Sq  
 Origin: None  
 Dependency: Response  
 Calib Mode: ISTD  
 Response Base: AREA  
 RF Rounding: 0

Curve Coefficients	
Intercept:	1.192
Slope:	0.5459

Error Coefficients	
Standard Error:	1910000
Relative Standard Error:	2.7
Correlation Coefficient:	0.999
Coefficient of Determination (Adjusted):	0.999

ID	Level	Concentration	Rel. Resp.	IS Amount	IS Response	RRF	Used
1	IC 680-808404/12	2.0	2.294241	50.0	1977887.0	1.147121	Y
2	IC 680-808404/11	5.0	3.839585	50.0	2217440.0	0.767917	Y
3	IC 680-808404/10	10.0	8.474634	50.0	1966905.0	0.847463	N
4	ICIS 680-808404/9	20.0	12.391518	50.0	2506763.0	0.619576	Y
5	IC 680-808404/8	50.0	28.837307	50.0	2634100.0	0.576746	Y
6	IC 680-808404/7	80.0	58.50091	50.0	1870888.0	0.731261	N
7	IC 680-808404/6	100.0	54.796933	50.0	2615890.0	0.547969	Y



**Calibration**


**/ 2,2'-Oxybisethanol**

Curve Type: Linear  
 Weighting: Conc\_Sq  
 Origin: None  
 Dependency: Response  
 Calib Mode: ISTD  
 Response Base: AREA  
 RF Rounding: 0

Curve Coefficients	
Intercept:	0.8971
Slope:	0.3278

Error Coefficients	
Standard Error:	1160000
Relative Standard Error:	8.1
Correlation Coefficient:	0.998
Coefficient of Determination (Adjusted):	0.993

ID	Level	Concentration	Rel. Resp.	IS Amount	IS Response	RRF	Used
1	IC 680-808404/12	2.0	1.561363	50.0	1977887.0	0.780682	Y
2	IC 680-808404/11	5.0	2.448409	50.0	2217440.0	0.489682	Y
3	IC 680-808404/10	10.0	6.370592	50.0	1966905.0	0.637059	N
4	ICIS 680-808404/9	20.0	8.156296	50.0	2506763.0	0.407815	Y
5	IC 680-808404/8	50.0	16.130595	50.0	2634100.0	0.322612	Y
6	IC 680-808404/7	80.0	33.188999	50.0	1870888.0	0.414862	N
7	IC 680-808404/6	100.0	33.775159	50.0	2615890.0	0.337752	Y



**Calibration**


/ Triethylene Glycol

Curve Type: Linear  
 Weighting: Conc\_Sq  
 Origin: None  
 Dependency: Response  
 Calib Mode: ISTD  
 Response Base: AREA  
 RF Rounding: 0

Curve Coefficients	
Intercept:	1.724
Slope:	0.3558

Error Coefficients	
Standard Error:	1320000
Relative Standard Error:	8.1
Correlation Coefficient:	0.998
Coefficient of Determination (Adjusted):	0.993

ID	Level	Concentration	Rel. Resp.	IS Amount	IS Response	RRF	Used
1	IC 680-808404/12	2.0	2.453174	50.0	1977887.0	1.226587	Y
2	IC 680-808404/11	5.0	3.362391	50.0	2217440.0	0.672478	Y
3	IC 680-808404/10	10.0	7.689466	50.0	1966905.0	0.768947	N
4	ICIS 680-808404/9	20.0	9.456778	50.0	2506763.0	0.472839	Y
5	IC 680-808404/8	50.0	18.338199	50.0	2634100.0	0.366764	Y
6	IC 680-808404/7	80.0	36.404344	50.0	1870888.0	0.455054	N
7	IC 680-808404/6	100.0	38.493725	50.0	2615890.0	0.384937	Y





**Calibration**


/ Tetraethylene Glycol

Curve Type: Linear  
 Weighting: Conc\_Sq  
 Origin: None  
 Dependency: Response  
 Calib Mode: ISTD  
 Response Base: AREA  
 RF Rounding: 0

Curve Coefficients	
Intercept:	1.778
Slope:	0.3348

Error Coefficients	
Standard Error:	2350000
Relative Standard Error:	8.5
Correlation Coefficient:	0.999
Coefficient of Determination (Adjusted):	0.993

ID	Level	Concentration	Rel. Resp.	IS Amount	IS Response	RRF	Used
1	IC 680-808404/12	4.0	3.136099	50.0	1977887.0	0.784025	Y
2	IC 680-808404/11	10.0	4.933437	50.0	2217440.0	0.493344	Y
3	IC 680-808404/10	20.0	13.389182	50.0	1966905.0	0.669459	N
4	ICIS 680-808404/9	40.0	16.743166	50.0	2506763.0	0.418579	Y
5	IC 680-808404/8	100.0	33.063855	50.0	2634100.0	0.330639	Y
6	IC 680-808404/7	160.0	66.73772	50.0	1870888.0	0.417111	N
7	IC 680-808404/6	200.0	68.182473	50.0	2615890.0	0.340912	Y



FORM VII  
GC SEMI VOA CONTINUING CALIBRATION DATA

Lab Name: Eurofins Savannah Job No.: 580-134720-1  
 SDG No.: \_\_\_\_\_  
 Lab Sample ID: ICV 680-808404/14 Calibration Date: 11/15/2023 18:31  
 Instrument ID: CVGG2 Calib Start Date: 11/15/2023 15:23  
 GC Column: J&W DB WAX ID: 0.45 (mm) Calib End Date: 11/15/2023 17:44  
 Lab File ID: 1IK15014.D Conc. Units: mg/L

ANALYTE	CURVE TYPE	AVE RRF	RRF	MIN RRF	CALC AMOUNT	SPIKE AMOUNT	%D	MAX %D
Ethanol, 2-propoxy	Lin2		0.7102		21.2	20.0	6.0	20.0
4-Hydroxy-4-methyl-2-pentano ne	Lin2		0.6926		21.6	20.0	8.1	20.0
2-Butoxyethanol	Lin2		0.7889		20.9	20.0	4.3	20.0
Dipropylene Glycol Methyl Ether	Lin2		0.0630		23.0	20.0	15.0	20.0
Propylene glycol	Lin2		0.5517		24.8	20.0	24.0*	20.0
Ethylene glycol	Lin2		0.4160		25.2	20.0	26.2*	20.0
2-(2-Butoxyethoxy)ethanol	Lin2		0.6852		22.9	20.0	14.6	20.0
2,2'-Oxybisethanol	Lin2		0.4579		25.2	20.0	26.0*	20.0
Triethylene Glycol	Lin2		0.5216		24.5	20.0	22.4*	20.0
Tetraethylene Glycol	Lin2		0.4596		49.6	40.0	24.0*	20.0

FORM VII  
GC SEMI VOA CONTINUING CALIBRATION RETENTION TIME SUMMARY

Lab Name: Eurofins Savannah Job No.: 580-134720-1  
 SDG No.: \_\_\_\_\_  
 Lab Sample ID: ICV 680-808404/14 Calibration Date: 11/15/2023 18:31  
 Instrument ID: CVGG2 Calib Start Date: 11/15/2023 15:23  
 GC Column: J&W DB WAX ID: 0.45 (mm) Calib End Date: 11/15/2023 17:44  
 Lab File ID: 1IK15014.D

Analyte	RT	RT WINDOW	
		FROM	TO
Ethanol, 2-propoxy	4.48	4.39	4.56
4-Hydroxy-4-methyl-2-pentanone	5.27	5.16	5.37
2-Butoxyethanol	5.70	5.58	5.80
Dipropylene Glycol Methyl Ether	7.33	7.18	7.48
Propylene glycol	8.26	8.09	8.42
Ethylene glycol	8.71	8.53	8.88
2-(2-Butoxyethoxy)ethanol	9.81	9.61	10.00
2,2'-Oxybisethanol	10.51	10.29	10.71
Triethylene Glycol	11.70	11.46	11.93
Tetraethylene Glycol	13.99	13.71	14.27

Eurofins Savannah  
Target Compound Quantitation Report

Data File: \\chromfs\Savannah\ChromData\CVGG2\20231115-90783.b\1IK15014.D  
 Lims ID: icv gly  
 Client ID:  
 Sample Type: ICV  
 Inject. Date: 15-Nov-2023 18:31:10 ALS Bottle#: 0 Worklist Smp#: 14  
 Injection Vol: 1.0 ul Dil. Factor: 1.0000  
 Sample Info: 680-0090783-014  
 Operator ID: Instrument ID: CVGG2  
 Sublist:

Method: \\chromfs\Savannah\ChromData\CVGG2\20231115-90783.b\8015\_GLY\_VGG.m  
 Limit Group: 8015C\_DAI  
 Last Update: 16-Nov-2023 11:39:12 Calib Date: 15-Nov-2023 17:44:20  
 Integrator: Falcon  
 Quant Method: Internal Standard Quant By: Initial Calibration  
 Last ICal File: \\chromfs\Savannah\ChromData\CVGG2\20231115-90783.b\1IK15012.D  
 Column 1 : J&W DB WAX ( 0.45 mm) Det: GC FID2B  
 Process Host: CTX1640

First Level Reviewer: AR8P Date: 16-Nov-2023 11:15:03

RT (min.)	Exp RT (min.)	Dlt RT (min.)	Response	Cal Amt ug/ml	OnCol Amt ug/ml	Flags
--------------	------------------	------------------	----------	------------------	--------------------	-------

1 Ethanol, 2-propoxy	4.479	4.474	0.005	638653	20.0	21.2
2 4-Hydroxy-4-methyl-2-pentanone	5.265	5.260	0.005	622810	20.0	21.6
3 2-Butoxyethanol	5.695	5.690	0.005	709392	20.0	20.9
* 4 n-Heptyl Alcohol	6.283	6.278	0.005	2248055	50.0	50.0
5 Dipropylene Glycol Methyl Ether	7.333	7.328	0.005	56613	20.0	23.0
6 Propylene glycol	8.262	8.258	0.004	496126	20.0	24.8
7 Ethylene glycol	8.705	8.703	0.002	374101	20.0	25.2
8 2-(2-Butoxyethoxy)ethanol	9.809	9.807	0.002	616120	20.0	22.9
9 2,2'-Oxybisethanol	10.505	10.504	0.001	411713	20.0	25.2
10 Triethylene Glycol	11.700	11.698	0.002	469063	20.0	24.5
11 Tetraethylene Glycol	13.992	13.988	0.004	826585	40.0	49.6

QC Flag Legend  
Processing Flags

Reagents:

SG\_GlyICV\_00062

Amount Added: 10.00

Units: uL

SG\_GLY\_ISTD\_00129

Amount Added: 10.00

Units: uL

Run Reagent

Eurofins Savannah

Data File: \\chromfs\Savannah\ChromData\CVGG2\20231115-90783.b\1IK15014.D

Injection Date: 15-Nov-2023 18:31:10

Instrument ID: CVGG2

Operator ID:

Lims ID: icv gly

Worklist Smp#: 14

Client ID:

Injection Vol: 1.0 ul


Dil. Factor: 1.0000

ALS Bottle#: 0

Method: 8015\_GLY\_VGG

Limit Group: 8015C\_DAI

Column: J&W DB WAX (0.45 mm)



FORM VII  
GC SEMI VOA CONTINUING CALIBRATION DATA

Lab Name: Eurofins Savannah Job No.: 580-134720-1  
 SDG No.: \_\_\_\_\_  
 Lab Sample ID: CCVIS 680-812978/6 Calibration Date: 12/12/2023 17:10  
 Instrument ID: CVGG2 Calib Start Date: 11/15/2023 15:23  
 GC Column: J&W DB WAX ID: 0.45 (mm) Calib End Date: 11/15/2023 17:44  
 Lab File ID: 1IL12006.D Conc. Units: mg/L

ANALYTE	CURVE TYPE	AVE RRF	RRF	MIN RRF	CALC AMOUNT	SPIKE AMOUNT	%D	MAX %D
Ethanol, 2-propoxy	Lin2		0.9004		27.3	20.0	36.6*	20.0
4-Hydroxy-4-methyl-2-pentane	Lin2		0.8540		27.1	20.0	35.3*	20.0
2-Butoxyethanol	Lin2		1.018		27.4	20.0	36.9*	20.0
Dipropylene Glycol Methyl Ether	Lin2		0.0764		28.6	20.0	42.8*	20.0
Propylene glycol	Lin2		0.6317		28.7	20.0	43.6*	20.0
Ethylene glycol	Lin2		0.4117		25.0	20.0	24.8*	20.0
2-(2-Butoxyethoxy)ethanol	Lin2		0.8448		28.8	20.0	43.9*	20.0
2,2'-Oxybisethanol	Lin2		0.4370		23.9	20.0	19.6	20.0
Triethylene Glycol	Lin2		0.5372		25.4	20.0	26.8*	20.0
Tetraethylene Glycol	Lin2		0.4980		54.2	40.0	35.5*	20.0

FORM VII  
GC SEMI VOA CONTINUING CALIBRATION RETENTION TIME SUMMARY

Lab Name: Eurofins Savannah Job No.: 580-134720-1  
 SDG No.: \_\_\_\_\_  
 Lab Sample ID: CCVIS 680-812978/6 Calibration Date: 12/12/2023 17:10  
 Instrument ID: CVGG2 Calib Start Date: 11/15/2023 15:23  
 GC Column: J&W DB WAX ID: 0.45 (mm) Calib End Date: 11/15/2023 17:44  
 Lab File ID: 1IL12006.D

Analyte	RT	RT WINDOW	
		FROM	TO
Ethanol, 2-propoxy	4.45	4.36	4.54
4-Hydroxy-4-methyl-2-pentanone	5.23	5.13	5.34
2-Butoxyethanol	5.67	5.55	5.78
Dipropylene Glycol Methyl Ether	7.30	7.15	7.45
Propylene glycol	8.28	8.11	8.44
Ethylene glycol	8.70	8.52	8.87
2-(2-Butoxyethoxy)ethanol	9.80	9.60	9.99
2,2'-Oxybisethanol	10.50	10.29	10.71
Triethylene Glycol	11.68	11.45	11.92
Tetraethylene Glycol	13.95	13.67	14.23



Eurofins Savannah  
Target Compound Quantitation Report

Data File: \\chromfs\Savannah\ChromData\CVGG2\20231212-91533.b\1IL12006.D  
 Lims ID: ccvis  
 Client ID:  
 Sample Type: CCVIS  
 Inject. Date: 12-Dec-2023 17:10:38 ALS Bottle#: 0 Worklist Smp#: 6  
 Injection Vol: 1.0 ul Dil. Factor: 1.0000  
 Sample Info: 680-0091533-006  
 Operator ID: Instrument ID: CVGG2  
 Sublist: chrom-8015\_GLY\_VGG\*sub2  
 Method: \\chromfs\Savannah\ChromData\CVGG2\20231212-91533.b\8015\_GLY\_VGG.m  
 Limit Group: 8015C\_DAI  
 Last Update: 13-Dec-2023 15:05:29 Calib Date: 15-Nov-2023 17:44:20  
 Integrator: Falcon  
 Quant Method: Internal Standard Quant By: Initial Calibration  
 Last ICal File: \\chromfs\Savannah\ChromData\CVGG2\20231115-90783.b\1IK15012.D  
 Column 1 : J&W DB WAX ( 0.45 mm) Det: GC FID2B  
 Process Host: CTX1647

RT (min.)	Exp RT (min.)	Dlt RT (min.)	Response	Cal Amt ug/ml	OnCol Amt ug/ml	Flags
-----------	---------------	---------------	----------	---------------	-----------------	-------

1 Ethanol, 2-propoxy	4.452	4.452	0.000	713692	20.0	27.3
2 4-Hydroxy-4-methyl-2-pentanone	5.234	5.234	0.000	676904	20.0	27.1
3 2-Butoxyethanol	5.666	5.666	0.000	806622	20.0	27.4
* 4 n-Heptyl Alcohol	6.254	6.254	0.000	1981601	50.0	50.0
5 Dipropylene Glycol Methyl Ether	7.300	7.300	0.000	60562	20.0	28.6
6 Propylene glycol	8.276	8.276	0.000	500725	20.0	28.7
7 Ethylene glycol	8.696	8.696	0.000	326362	20.0	25.0
8 2-(2-Butoxyethoxy)ethanol	9.796	9.796	0.000	669662	20.0	28.8
9 2,2'-Oxybisethanol	10.496	10.496	0.000	346371	20.0	23.9
10 Triethylene Glycol	11.682	11.682	0.000	425828	20.0	25.4
11 Tetraethylene Glycol	13.951	13.951	0.000	789448	40.0	54.2

Reagents:

SG\_Gly\_CAL\_00053 Amount Added: 10.00 Units: uL  
 SG\_GLY\_ISTD\_00129 Amount Added: 10.00 Units: uL Run Reagent

Eurofins Savannah

Data File: \\chromfs\Savannah\ChromData\CVGG2\20231212-91533.b\1IL12006.D

Injection Date: 12-Dec-2023 17:10:38

Instrument ID: CVGG2

Operator ID:

Lims ID: ccvis

Worklist Smp#: 6

Client ID:

Injection Vol: 1.0 ul


Dil. Factor: 1.0000

ALS Bottle#: 0

Method: 8015\_GLY\_VGG

Limit Group: 8015C\_DAI

Column: J&W DB WAX (0.45 mm)



FORM VII  
GC SEMI VOA CONTINUING CALIBRATION DATA

Lab Name: Eurofins Savannah Job No.: 580-134720-1  
 SDG No.: \_\_\_\_\_  
 Lab Sample ID: CCV 680-812978/33 Calibration Date: 12/13/2023 03:42  
 Instrument ID: CVGG2 Calib Start Date: 11/15/2023 15:23  
 GC Column: J&W DB WAX ID: 0.45 (mm) Calib End Date: 11/15/2023 17:44  
 Lab File ID: 1IL12033.D Conc. Units: mg/L

ANALYTE	CURVE TYPE	AVE RRF	RRF	MIN RRF	CALC AMOUNT	SPIKE AMOUNT	%D	MAX %D
Ethanol, 2-propoxy	Lin2		0.5925		17.4	20.0	-12.9	20.0
4-Hydroxy-4-methyl-2-pentane	Lin2		0.5147		15.6	20.0	-21.9*	20.0
2-Butoxyethanol	Lin2		0.6696		17.5	20.0	-12.7	20.0
Dipropylene Glycol Methyl Ether	Lin2		0.0475		16.6	20.0	-16.8	20.0
Propylene glycol	Lin2		0.3371		14.2	20.0	-28.8*	20.0
Ethylene glycol	Lin2		0.2230		12.4	20.0	-37.8*	20.0
2-(2-Butoxyethoxy)ethanol	Lin2		0.5149		16.7	20.0	-16.6	20.0
2,2'-Oxybisethanol	Lin2		0.1948		9.15	20.0	-54.2*	20.0
Triethylene Glycol	Lin2		0.2167		7.33	20.0	-63.3*	20.0
Tetraethylene Glycol	Lin2				10.0	40.0	-100.0*	20.0

FORM VII  
GC SEMI VOA CONTINUING CALIBRATION RETENTION TIME SUMMARY

Lab Name: Eurofins Savannah Job No.: 580-134720-1  
 SDG No.: \_\_\_\_\_  
 Lab Sample ID: CCV 680-812978/33 Calibration Date: 12/13/2023 03:42  
 Instrument ID: CVGG2 Calib Start Date: 11/15/2023 15:23  
 GC Column: J&W DB WAX ID: 0.45 (mm) Calib End Date: 11/15/2023 17:44  
 Lab File ID: 1IL12033.D

Analyte	RT	RT WINDOW	
		FROM	TO
Ethanol, 2-propoxy	4.45	4.36	4.54
4-Hydroxy-4-methyl-2-pentanone	5.23	5.13	5.34
2-Butoxyethanol	5.66	5.55	5.78
Dipropylene Glycol Methyl Ether	7.30	7.15	7.44
Propylene glycol	8.29	8.13	8.46
Ethylene glycol	8.71	8.53	8.88
2-(2-Butoxyethoxy)ethanol	9.79	9.60	9.99
2,2'-Oxybisethanol	10.50	10.29	10.71
Triethylene Glycol	11.69	11.46	11.92
Tetraethylene Glycol			

Eurofins Savannah  
Target Compound Quantitation Report

Data File: \\chromfs\Savannah\ChromData\CVGG2\20231212-91533.b\1IL12033.D  
 Lims ID: ccv G4  
 Client ID:  
 Sample Type: CCV  
 Inject. Date: 13-Dec-2023 03:42:01 ALS Bottle#: 0 Worklist Smp#: 33  
 Injection Vol: 1.0 ul Dil. Factor: 1.0000  
 Sample Info: 680-0091533-033  
 Operator ID: Instrument ID: CVGG2  
 Sublist: chrom-8015\_GLY\_VGG\*sub2  
 Method: \\chromfs\Savannah\ChromData\CVGG2\20231212-91533.b\8015\_GLY\_VGG.m  
 Limit Group: 8015C\_DAI  
 Last Update: 13-Dec-2023 15:05:42 Calib Date: 15-Nov-2023 17:44:20  
 Integrator: Falcon  
 Quant Method: Internal Standard Quant By: Initial Calibration  
 Last ICal File: \\chromfs\Savannah\ChromData\CVGG2\20231115-90783.b\1IK15012.D  
 Column 1 : J&W DB WAX ( 0.45 mm) Det: GC FID2B  
 Process Host: CTX1647

RT (min.)	Exp RT (min.)	Dlt RT (min.)	Response	Cal Amt ug/ml	OnCol Amt ug/ml	Flags
-----------	---------------	---------------	----------	---------------	-----------------	-------

1 Ethanol, 2-propoxy	4.451	4.451	0.000	662740	20.0	17.4
2 4-Hydroxy-4-methyl-2-pentanone	5.231	5.231	0.000	575689	20.0	15.6
3 2-Butoxyethanol	5.662	5.662	0.000	748995	20.0	17.5
* 4 n-Heptyl Alcohol	6.251	6.251	0.000	2796242	50.0	50.0
5 Dipropylene Glycol Methyl Ether	7.296	7.296	0.000	53124	20.0	16.6
6 Propylene glycol	8.294	8.294	0.000	377046	20.0	14.2
7 Ethylene glycol	8.706	8.706	0.000	249401	20.0	12.4
8 2-(2-Butoxyethoxy)ethanol	9.794	9.794	0.000	575909	20.0	16.7
9 2,2'-Oxybisethanol	10.497	10.497	0.000	217930	20.0	9.15
10 Triethylene Glycol	11.691	11.691	0.000	242350	20.0	7.33

Reagents:

SG\_Gly\_CAL\_00053 Amount Added: 10.00 Units: uL  
 SG\_GLY\_ISTD\_00129 Amount Added: 10.00 Units: uL Run Reagent

Eurofins Savannah

Data File: \\chromfs\Savannah\ChromData\CVGG2\20231212-91533.b\1IL12033.D

Injection Date: 13-Dec-2023 03:42:01

Instrument ID: CVGG2

Operator ID:

Lims ID: ccv G4

Worklist Smp#: 33

Client ID:

Injection Vol: 1.0 ul


Dil. Factor: 1.0000

ALS Bottle#: 0

Method: 8015\_GLY\_VGG

Limit Group: 8015C\_DAI

Column: J&W DB WAX (0.45 mm)



FORM I  
GC SEMI VOA ORGANICS ANALYSIS DATA SHEET

Lab Name: Eurofins Savannah Job No.: 580-134720-1  
 SDG No.: \_\_\_\_\_  
 Client Sample ID: \_\_\_\_\_ Lab Sample ID: MB 680-812978/11  
 Matrix: Water Lab File ID: 1IL12011.D  
 Analysis Method: 8015C GLY Date Collected: \_\_\_\_\_  
 Extraction Method: \_\_\_\_\_ Date Extracted: \_\_\_\_\_  
 Sample wt/vol: 1(mL) Date Analyzed: 12/12/2023 19:07  
 Con. Extract Vol.: 1(mL) Dilution Factor: 1  
 Injection Volume: 1(uL) GC Column: J&W DB WAX ID: 0.45(mm)  
 % Moisture: \_\_\_\_\_ % Solids: \_\_\_\_\_ GPC Cleanup: (Y/N) N  
 Cleanup Factor: \_\_\_\_\_  
 Analysis Batch No.: 812978 Units: mg/L

CAS NO.	COMPOUND NAME	RESULT	Q	LOQ	LOD	DL
112-34-5	2-(2-Butoxyethoxy)ethanol	3.0	U	5.0	3.0	1.1

Eurofins Savannah  
Target Compound Quantitation Report

Data File: \\chromfs\Savannah\ChromData\CVGG2\20231212-91533.b\1IL12011.D  
 Lims ID: mb  
 Client ID:  
 Sample Type: MB  
 Inject. Date: 12-Dec-2023 19:07:46 ALS Bottle#: 0 Worklist Smp#: 11  
 Injection Vol: 1.0 ul Dil. Factor: 1.0000  
 Sample Info: 680-0091533-011  
 Operator ID: Instrument ID: CVGG2  
 Method: \\chromfs\Savannah\ChromData\CVGG2\20231212-91533.b\8015\_GLY\_VGG.m  
 Limit Group: 8015C\_DAI  
 Last Update: 13-Dec-2023 15:05:42 Calib Date: 15-Nov-2023 17:44:20  
 Integrator: Falcon  
 Quant Method: Internal Standard Quant By: Initial Calibration  
 Last ICal File: \\chromfs\Savannah\ChromData\CVGG2\20231115-90783.b\1IK15012.D  
 Column 1 : J&W DB WAX ( 0.45 mm) Det: GC FID2B  
 Process Host: CTX1647

First Level Reviewer: AR8P Date: 13-Dec-2023 15:04:05

RT (min.)	Exp RT (min.)	Dlt RT (min.)	Response	Cal Amt ug/ml	OnCol Amt ug/ml	Flags
--------------	------------------	------------------	----------	------------------	--------------------	-------

\* 4 n-Heptyl Alcohol  
 6.251 6.251 0.000 2765692 50.0 50.0

Reagents:

SG\_GLY\_ISTD\_00129 Amount Added: 10.00 Units: uL Run Reagent



Eurofins Savannah

Data File: \\chromfs\Savannah\ChromData\CVGG2\20231212-91533.b\1IL12011.D

Injection Date: 12-Dec-2023 19:07:46

Instrument ID: CVGG2

Operator ID:

Lims ID: mb

Worklist Smp#: 11

Client ID:

Injection Vol: 1.0 ul


Dil. Factor: 1.0000

ALS Bottle#: 0

Method: 8015\_GLY\_VGG

Limit Group: 8015C\_DAI

Column: J&W DB WAX (0.45 mm)



FORM I  
GC SEMI VOA ORGANICS ANALYSIS DATA SHEET

Lab Name: Eurofins Savannah Job No.: 580-134720-1  
 SDG No.: \_\_\_\_\_  
 Client Sample ID: \_\_\_\_\_ Lab Sample ID: LCS 680-812978/7  
 Matrix: Water Lab File ID: 1IL12007.D  
 Analysis Method: 8015C GLY Date Collected: \_\_\_\_\_  
 Extraction Method: \_\_\_\_\_ Date Extracted: \_\_\_\_\_  
 Sample wt/vol: 1(mL) Date Analyzed: 12/12/2023 17:34  
 Con. Extract Vol.: 1(mL) Dilution Factor: 1  
 Injection Volume: 1(uL) GC Column: J&W DB WAX ID: 0.45(mm)  
 % Moisture: \_\_\_\_\_ % Solids: \_\_\_\_\_ GPC Cleanup: (Y/N) N  
 Cleanup Factor: \_\_\_\_\_  
 Analysis Batch No.: 812978 Units: mg/L

CAS NO.	COMPOUND NAME	RESULT	Q	LOQ	LOD	DL
112-34-5	2-(2-Butoxyethoxy)ethanol	19.9		5.0	3.0	1.1

Eurofins Savannah  
Target Compound Quantitation Report

Data File: \\chromfs\Savannah\ChromData\CVGG2\20231212-91533.b\1IL12007.D  
 Lims ID: lcs  
 Client ID:  
 Sample Type: LCS  
 Inject. Date: 12-Dec-2023 17:34:07 ALS Bottle#: 0 Worklist Smp#: 7  
 Injection Vol: 1.0 ul Dil. Factor: 1.0000  
 Sample Info: 680-0091533-007  
 Operator ID: Instrument ID: CVGG2  
 Method: \\chromfs\Savannah\ChromData\CVGG2\20231212-91533.b\8015\_GLY\_VGG.m  
 Limit Group: 8015C\_DAI  
 Last Update: 13-Dec-2023 15:05:28 Calib Date: 15-Nov-2023 17:44:20  
 Integrator: Falcon  
 Quant Method: Internal Standard Quant By: Initial Calibration  
 Last ICal File: \\chromfs\Savannah\ChromData\CVGG2\20231115-90783.b\1IK15012.D  
 Column 1 : J&W DB WAX ( 0.45 mm) Det: GC FID2B  
 Process Host: CTX1647

RT (min.)	Exp RT (min.)	Dlt RT (min.)	Response	Cal Amt ug/ml	OnCol Amt ug/ml	Flags
--------------	------------------	------------------	----------	------------------	--------------------	-------

1 Ethanol, 2-propoxy	4.460	4.460	0.000	1010293	20.0	24.9
2 4-Hydroxy-4-methyl-2-pentanone	5.245	5.245	0.000	921501	20.0	23.6
3 2-Butoxyethanol	5.667	5.667	0.000	1142067	20.0	24.9
* 4 n-Heptyl Alcohol	6.248	6.248	0.000	3064834	50.0	50.0
5 Dipropylene Glycol Methyl Ether	7.304	7.304	0.000	70009	20.0	20.6
6 Propylene glycol	8.272	8.272	0.000	421172	20.0	14.6
7 Ethylene glycol	8.697	8.697	0.000	263962	20.0	11.9
8 2-(2-Butoxyethoxy)ethanol	9.795	9.795	0.000	739497	20.0	19.9
9 2,2'-Oxybisethanol	10.495	10.495	0.000	270054	20.0	10.7
10 Triethylene Glycol	11.682	11.682	0.000	343214	20.0	10.9
11 Tetraethylene Glycol	13.950	13.950	0.000	618973	40.0	24.8

**Reagents:**

SG\_GlylCV\_00062 Amount Added: 10.00 Units: uL  
 SG\_GLY\_ISTD\_00129 Amount Added: 10.00 Units: uL Run Reagent

Eurofins Savannah

Data File: \\chromfs\Savannah\ChromData\CVGG2\20231212-91533.b\1IL12007.D

Injection Date: 12-Dec-2023 17:34:07

Instrument ID: CVGG2

Operator ID:

Lims ID: lcs

Worklist Smp#: 7

Client ID:

Injection Vol: 1.0 ul


Dil. Factor: 1.0000

ALS Bottle#: 0

Method: 8015\_GLY\_VGG

Limit Group: 8015C\_DAI

Column: J&W DB WAX (0.45 mm)



FORM I  
GC SEMI VOA ORGANICS ANALYSIS DATA SHEET

Lab Name: Eurofins Savannah Job No.: 580-134720-1  
 SDG No.: \_\_\_\_\_  
 Client Sample ID: \_\_\_\_\_ Lab Sample ID: LCSD 680-812978/8  
 Matrix: Water Lab File ID: 1IL12008.D  
 Analysis Method: 8015C GLY Date Collected: \_\_\_\_\_  
 Extraction Method: \_\_\_\_\_ Date Extracted: \_\_\_\_\_  
 Sample wt/vol: 1(mL) Date Analyzed: 12/12/2023 17:57  
 Con. Extract Vol.: 1(mL) Dilution Factor: 1  
 Injection Volume: 1(uL) GC Column: J&W DB WAX ID: 0.45(mm)  
 % Moisture: \_\_\_\_\_ % Solids: \_\_\_\_\_ GPC Cleanup: (Y/N) N  
 Cleanup Factor: \_\_\_\_\_  
 Analysis Batch No.: 812978 Units: mg/L

CAS NO.	COMPOUND NAME	RESULT	Q	LOQ	LOD	DL
112-34-5	2-(2-Butoxyethoxy)ethanol	20.8		5.0	3.0	1.1

Eurofins Savannah  
Target Compound Quantitation Report

Data File: \\chromfs\Savannah\ChromData\CVGG2\20231212-91533.b\1IL12008.D  
 Lims ID: lcsd  
 Client ID:  
 Sample Type: LCSD  
 Inject. Date: 12-Dec-2023 17:57:29 ALS Bottle#: 0 Worklist Smp#: 8  
 Injection Vol: 1.0 ul Dil. Factor: 1.0000  
 Sample Info: 680-0091533-008  
 Operator ID: Instrument ID: CVGG2  
 Method: \\chromfs\Savannah\ChromData\CVGG2\20231212-91533.b\8015\_GLY\_VGG.m  
 Limit Group: 8015C\_DAI  
 Last Update: 13-Dec-2023 15:04:56 Calib Date: 15-Nov-2023 17:44:20  
 Integrator: Falcon  
 Quant Method: Internal Standard Quant By: Initial Calibration  
 Last ICal File: \\chromfs\Savannah\ChromData\CVGG2\20231115-90783.b\1IK15012.D  
 Column 1 : J&W DB WAX ( 0.45 mm) Det: GC FID2B  
 Process Host: CTX1647

RT (min.)	Exp RT (min.)	Dlt RT (min.)	Response	Cal Amt ug/ml	OnCol Amt ug/ml	Flags
--------------	------------------	------------------	----------	------------------	--------------------	-------

1 Ethanol, 2-propoxy	4.462	4.454	0.008	1096695	20.0	25.5
2 4-Hydroxy-4-methyl-2-pentanone	5.246	5.236	0.010	994692	20.0	24.0
3 2-Butoxyethanol	5.667	5.663	0.004	1247162	20.0	25.7
* 4 n-Heptyl Alcohol	6.249	6.249	0.000	3251792	50.0	50.0
5 Dipropylene Glycol Methyl Ether	7.305	7.300	0.005	77127	20.0	21.5
6 Propylene glycol	8.272	8.296	-0.024	502658	20.0	16.7
7 Ethylene glycol	8.696	8.707	-0.011	324671	20.0	14.2
8 2-(2-Butoxyethoxy)ethanol	9.795	9.795	0.000	815143	20.0	20.8
9 2,2'-Oxybisethanol	10.495	10.499	-0.004	343713	20.0	13.4
10 Triethylene Glycol	11.682	11.689	-0.007	388136	20.0	11.9
11 Tetraethylene Glycol	13.947	13.950	-0.003	726176	40.0	28.0

Reagents:

SG\_GlyICV\_00062 Amount Added: 10.00 Units: uL  
 SG\_GLY\_ISTD\_00129 Amount Added: 10.00 Units: uL Run Reagent

Eurofins Savannah

Data File: \\chromfs\Savannah\ChromData\CVGG2\20231212-91533.b\1IL12008.D

Injection Date: 12-Dec-2023 17:57:29

Instrument ID: CVGG2

Operator ID:

Lims ID: lcsd

Worklist Smp#: 8

Client ID:

Injection Vol: 1.0 ul


Dil. Factor: 1.0000

ALS Bottle#: 0

Method: 8015\_GLY\_VGG

Limit Group: 8015C\_DAI

Column: J&W DB WAX (0.45 mm)



FORM I  
GC SEMI VOA ORGANICS ANALYSIS DATA SHEET

Lab Name: Eurofins Savannah Job No.: 580-134720-1  
 SDG No.: \_\_\_\_\_  
 Client Sample ID: AF-RHMW17-WGN01LF-2312 MS Lab Sample ID: 580-134720-1 MS  
 Matrix: Water Lab File ID: 1IL12030.D  
 Analysis Method: 8015C GLY Date Collected: 12/06/2023 11:50  
 Extraction Method: \_\_\_\_\_ Date Extracted: \_\_\_\_\_  
 Sample wt/vol: 1(mL) Date Analyzed: 12/13/2023 02:31  
 Con. Extract Vol.: 1(mL) Dilution Factor: 1  
 Injection Volume: 1(uL) GC Column: J&W DB WAX ID: 0.45(mm)  
 % Moisture: \_\_\_\_\_ % Solids: \_\_\_\_\_ GPC Cleanup: (Y/N) N  
 Cleanup Factor: \_\_\_\_\_  
 Analysis Batch No.: 812978 Units: mg/L

CAS NO.	COMPOUND NAME	RESULT	Q	LOQ	LOD	DL
112-34-5	2-(2-Butoxyethoxy)ethanol	17.7		5.0	3.0	1.1



Eurofins Savannah  
Target Compound Quantitation Report

Data File: \\chromfs\Savannah\ChromData\CVGG2\20231212-91533.b\1IL12030.D  
 Lims ID: 580-134720-A-1 MS  
 Client ID:  
 Sample Type: MS  
 Inject. Date: 13-Dec-2023 02:31:58 ALS Bottle#: 0 Worklist Smp#: 30  
 Injection Vol: 1.0 ul Dil. Factor: 1.0000  
 Sample Info: 680-0091533-030  
 Operator ID: Instrument ID: CVGG2  
 Method: \\chromfs\Savannah\ChromData\CVGG2\20231212-91533.b\8015\_GLY\_VGG.m  
 Limit Group: 8015C\_DAI  
 Last Update: 13-Dec-2023 15:05:29 Calib Date: 15-Nov-2023 17:44:20  
 Integrator: Falcon  
 Quant Method: Internal Standard Quant By: Initial Calibration  
 Last ICal File: \\chromfs\Savannah\ChromData\CVGG2\20231115-90783.b\1IK15012.D  
 Column 1 : J&W DB WAX ( 0.45 mm) Det: GC FID2B  
 Process Host: CTX1647

RT (min.)	Exp RT (min.)	Dlt RT (min.)	Response	Cal Amt ug/ml	OnCol Amt ug/ml	Flags
-----------	---------------	---------------	----------	---------------	-----------------	-------

1 Ethanol, 2-propoxy	4.453	4.452	0.001	714448	20.0	19.8
2 4-Hydroxy-4-methyl-2-pentanone	5.236	5.234	0.002	661671	20.0	19.1
3 2-Butoxyethanol	5.663	5.666	-0.003	786467	20.0	19.3
* 4 n-Heptyl Alcohol	6.249	6.254	-0.005	2678030	50.0	50.0
5 Dipropylene Glycol Methyl Ether	7.301	7.300	0.001	53773	20.0	17.7
6 Propylene glycol	8.299	8.276	0.023	397806	20.0	15.9
7 Ethylene glycol	8.706	8.696	0.010	358900	20.0	19.9
8 2-(2-Butoxyethoxy)ethanol	9.796	9.796	0.000	581819	20.0	17.7
9 2,2'-Oxybisethanol	10.501	10.496	0.005	146079	20.0	5.58

Reagents:

SG\_Gly\_CAL\_00056 Amount Added: 10.00 Units: uL  
 SG\_GLY\_ISTD\_00129 Amount Added: 10.00 Units: uL Run Reagent

Eurofins Savannah

Data File: \\chromfs\Savannah\ChromData\CVGG2\20231212-91533.b\1IL12030.D

Injection Date: 13-Dec-2023 02:31:58

Instrument ID: CVGG2

Operator ID:

Lims ID: 580-134720-A-1 MS

Worklist Smp#: 30

Client ID:

Injection Vol: 1.0 ul


Dil. Factor: 1.0000

ALS Bottle#: 0

Method: 8015\_GLY\_VGG

Limit Group: 8015C\_DAI

Column: J&W DB WAX (0.45 mm)



FORM I  
GC SEMI VOA ORGANICS ANALYSIS DATA SHEET

Lab Name: <u>Eurofins Savannah</u>	Job No.: <u>580-134720-1</u>
SDG No.: _____	
Client Sample ID: <u>AF-RHMW17-WGN01LF-2312</u> MSD	Lab Sample ID: <u>580-134720-1 MSD</u>
Matrix: <u>Water</u>	Lab File ID: <u>1IL12031.D</u>
Analysis Method: <u>8015C GLY</u>	Date Collected: <u>12/06/2023 11:50</u>
Extraction Method: _____	Date Extracted: _____
Sample wt/vol: <u>1(mL)</u>	Date Analyzed: <u>12/13/2023 02:55</u>
Con. Extract Vol.: <u>1(mL)</u>	Dilution Factor: <u>1</u>
Injection Volume: <u>1(uL)</u>	GC Column: <u>J&amp;W DB WAX</u> ID: <u>0.45(mm)</u>
% Moisture: _____ % Solids: _____	GPC Cleanup: (Y/N) <u>N</u>
Cleanup Factor: _____	
Analysis Batch No.: <u>812978</u>	Units: <u>mg/L</u>

CAS NO.	COMPOUND NAME	RESULT	Q	LOQ	LOD	DL
112-34-5	2-(2-Butoxyethoxy)ethanol	15.3		5.0	3.0	1.1

Eurofins Savannah  
Target Compound Quantitation Report

Data File: \\chromfs\Savannah\ChromData\CVGG2\20231212-91533.b\1IL12031.D  
 Lims ID: 580-134720-A-1 MSD  
 Client ID:  
 Sample Type: MSD  
 Inject. Date: 13-Dec-2023 02:55:24 ALS Bottle#: 0 Worklist Smp#: 31  
 Injection Vol: 1.0 ul Dil. Factor: 1.0000  
 Sample Info: 680-0091533-031  
 Operator ID: Instrument ID: CVGG2  
 Method: \\chromfs\Savannah\ChromData\CVGG2\20231212-91533.b\8015\_GLY\_VGG.m  
 Limit Group: 8015C\_DAI  
 Last Update: 13-Dec-2023 15:05:29 Calib Date: 15-Nov-2023 17:44:20  
 Integrator: Falcon  
 Quant Method: Internal Standard Quant By: Initial Calibration  
 Last ICal File: \\chromfs\Savannah\ChromData\CVGG2\20231115-90783.b\1IK15012.D  
 Column 1 : J&W DB WAX ( 0.45 mm) Det: GC FID2B  
 Process Host: CTX1647

RT (min.)	Exp RT (min.)	Dlt RT (min.)	Response	Cal Amt ug/ml	OnCol Amt ug/ml	Flags
--------------	------------------	------------------	----------	------------------	--------------------	-------

1 Ethanol, 2-propoxy	4.449	4.452	-0.003	698217	20.0	17.1
2 4-Hydroxy-4-methyl-2-pentanone	5.231	5.234	-0.003	626506	20.0	15.9
3 2-Butoxyethanol	5.663	5.666	-0.003	791536	20.0	17.2
* 4 n-Heptyl Alcohol	6.253	6.254	-0.001	2998400	50.0	50.0
5 Dipropylene Glycol Methyl Ether	7.297	7.300	-0.003	52846	20.0	15.2
6 Propylene glycol	8.327	8.276	0.051	392686	20.0	13.8
7 Ethylene glycol	8.708	8.696	0.012	372609	20.0	18.3
8 2-(2-Butoxyethoxy)ethanol	9.796	9.796	0.000	572551	20.0	15.3
9 2,2'-Oxybisethanol	10.500	10.496	0.004	183631	20.0	6.61
10 Triethylene Glycol	11.693	11.682	0.011	141301	20.0	1.78

Reagents:

SG\_Gly\_CAL\_00056 Amount Added: 10.00 Units: uL  
 SG\_GLY\_ISTD\_00129 Amount Added: 10.00 Units: uL Run Reagent

Eurofins Savannah

Data File: \\chromfs\Savannah\ChromData\CVGG2\20231212-91533.b\1IL12031.D

Injection Date: 13-Dec-2023 02:55:24

Instrument ID: CVGG2

Operator ID:

Lims ID: 580-134720-A-1 MSD

Worklist Smp#: 31

Client ID:

Injection Vol: 1.0 ul


Dil. Factor: 1.0000

ALS Bottle#: 0

Method: 8015\_GLY\_VGG

Limit Group: 8015C\_DAI

Column: J&W DB WAX (0.45 mm)



GC SEMI VOA ANALYSIS RUN LOG

Lab Name: Eurofins Savannah Job No.: 580-134720-1

SDG No.: \_\_\_\_\_

Instrument ID: CVGG2 Start Date: 11/15/2023 15:23

Analysis Batch Number: 808404 End Date: 11/15/2023 18:31

LAB SAMPLE ID	CLIENT SAMPLE ID	DATE ANALYZED	DILUTION FACTOR	LAB FILE ID	COLUMN ID
IC 680-808404/6		11/15/2023 15:23	1	1IK15006.D	J&W DB WAX 0.45 (mm)
IC 680-808404/7		11/15/2023 15:47	1	1IK15007.D	J&W DB WAX 0.45 (mm)
IC 680-808404/8		11/15/2023 16:10	1	1IK15008.D	J&W DB WAX 0.45 (mm)
ICIS 680-808404/9		11/15/2023 16:34	1	1IK15009.D	J&W DB WAX 0.45 (mm)
IC 680-808404/10		11/15/2023 16:57	1	1IK15010.D	J&W DB WAX 0.45 (mm)
IC 680-808404/11		11/15/2023 17:20	1	1IK15011.D	J&W DB WAX 0.45 (mm)
IC 680-808404/12		11/15/2023 17:44	1	1IK15012.D	J&W DB WAX 0.45 (mm)
ICV 680-808404/14		11/15/2023 18:31	1	1IK15014.D	J&W DB WAX 0.45 (mm)

GC SEMI VOA ANALYSIS RUN LOG

Lab Name: Eurofins Savannah Job No.: 580-134720-1

SDG No.: \_\_\_\_\_

Instrument ID: CVGG2 Start Date: 12/12/2023 17:10

Analysis Batch Number: 812978 End Date: 12/13/2023 03:42

LAB SAMPLE ID	CLIENT SAMPLE ID	DATE ANALYZED	DILUTION FACTOR	LAB FILE ID	COLUMN ID
CCVIS 680-812978/6		12/12/2023 17:10	1	1IL12006.D	J&W DB WAX 0.45 (mm)
LCS 680-812978/7		12/12/2023 17:34	1	1IL12007.D	J&W DB WAX 0.45 (mm)
LCSD 680-812978/8		12/12/2023 17:57	1	1IL12008.D	J&W DB WAX 0.45 (mm)
MB 680-812978/11		12/12/2023 19:07	1	1IL12011.D	J&W DB WAX 0.45 (mm)
ZZZZZ		12/12/2023 21:04	1		J&W DB WAX 0.45 (mm)
ZZZZZ		12/12/2023 21:28	1		J&W DB WAX 0.45 (mm)
ZZZZZ		12/12/2023 21:51	1		J&W DB WAX 0.45 (mm)
ZZZZZ		12/12/2023 22:14	1		J&W DB WAX 0.45 (mm)
ZZZZZ		12/12/2023 22:38	1		J&W DB WAX 0.45 (mm)
ZZZZZ		12/12/2023 23:01	1		J&W DB WAX 0.45 (mm)
ZZZZZ		12/12/2023 23:25	1		J&W DB WAX 0.45 (mm)
580-134720-1	AF-RHMW17-WGN01LF-2312	12/12/2023 23:48	1	1IL12023.D	J&W DB WAX 0.45 (mm)
580-134720-2	AF-RHMW17D-WGN01LF-2312	12/13/2023 00:11	1	1IL12024.D	J&W DB WAX 0.45 (mm)
580-134720-3	AF-RHMW17D-WQFB01-2312	12/13/2023 00:35	1	1IL12025.D	J&W DB WAX 0.45 (mm)
580-134720-4	AF-RHMW17S-WGN01LF-2312	12/13/2023 00:58	1	1IL12026.D	J&W DB WAX 0.45 (mm)
580-134720-5	AF-RHMW17S-WQEB01-2312	12/13/2023 01:21	1	1IL12027.D	J&W DB WAX 0.45 (mm)
580-134720-6	AF-RHMW12A-WGN01LF-2312	12/13/2023 01:45	1	1IL12028.D	J&W DB WAX 0.45 (mm)
580-134720-7	AF-RHMW12A-WGFD01LF-2312	12/13/2023 02:08	1	1IL12029.D	J&W DB WAX 0.45 (mm)
580-134720-1 MS	AF-RHMW17-WGN01LF-2312 MS	12/13/2023 02:31	1	1IL12030.D	J&W DB WAX 0.45 (mm)
580-134720-1 MSD	AF-RHMW17-WGN01LF-2312 MSD	12/13/2023 02:55	1	1IL12031.D	J&W DB WAX 0.45 (mm)
CCV 680-812978/33		12/13/2023 03:42	1	1IL12033.D	J&W DB WAX 0.45 (mm)
ZZZZZ		12/13/2023 03:42	1		J&W DB WAX 0.45 (mm)

GC SEMI VOA BATCH WORKSHEET

Lab Name: Eurofins Savannah Job No.: 580-134720-1

SDG No.: \_\_\_\_\_

Batch Number: 808404 Batch Start Date: 11/15/23 15:23 Batch Analyst: Mullis, David B

Batch Method: 8015C GLY Batch End Date: \_\_\_\_\_

Lab Sample ID	Client Sample ID	Method Chain	Basis	FinalAmount	SG_Gly_CAL 00053	SG_GLY_ISTD 00129	SG_GlyICV 00062		
IC 680-808404/6		8015C GLY		1 mL	50 uL	10 uL			
IC 680-808404/7		8015C GLY		1 mL	40 uL	10 uL			
IC 680-808404/8		8015C GLY		1 mL	25 uL	10 uL			
ICIS 680-808404/9		8015C GLY		1 mL	10 uL	10 uL			
IC 680-808404/10		8015C GLY		1 mL	5 uL	10 uL			
IC 680-808404/11		8015C GLY		1 mL	2.5 uL	10 uL			
IC 680-808404/12		8015C GLY		1 mL	1 uL	10 uL			
ICV 680-808404/14		8015C GLY		1 mL		10 uL	10 uL		

Batch Notes	

Basis	Basis Description

The pound sign (#) in the amount added field denotes that the reagent was used undiluted. All calculations are performed using the stated concentration for this reagent.



GC SEMI VOA BATCH WORKSHEET

Lab Name: Eurofins Savannah Job No.: 580-134720-1

SDG No.: \_\_\_\_\_

Batch Number: 812978 Batch Start Date: 12/12/23 17:10 Batch Analyst: Mullis, David B

Batch Method: 8015C GLY Batch End Date: \_\_\_\_\_

Lab Sample ID	Client Sample ID	Method Chain	Basis	FinalAmount	SG_Gly_CAL 00053	SG_Gly_CAL 00056	SG_GLY_ISTD 00129	SG_GlyICV 00062	
CCVIS 680-812978/6		8015C GLY		1 mL	10 uL		10 uL		
LCS 680-812978/7		8015C GLY		1 mL			10 uL	10 uL	
LCSD 680-812978/8		8015C GLY		1 mL			10 uL	10 uL	
MB 680-812978/11		8015C GLY		1 mL			10 uL		
580-134720-A-1	AF-RHMW17-WGN01L F-2312	8015C GLY	T	1 mL			10 uL		
580-134720-B-2	AF-RHMW17D-WGN01 LF-2312	8015C GLY	T	1 mL			10 uL		
580-134720-A-3	AF-RHMW17D-WQFBO 1-2312	8015C GLY	T	1 mL			10 uL		
580-134720-A-4	AF-RHMW17S-WGN01 LF-2312	8015C GLY	T	1 mL			10 uL		
580-134720-A-5	AF-RHMW17S-WQEBO 1-2312	8015C GLY	T	1 mL			10 uL		
580-134720-A-6	AF-RHMW12A-WGN01 LF-2312	8015C GLY	T	1 mL			10 uL		
580-134720-B-7	AF-RHMW12A-WGFD0 1LF-2312	8015C GLY	T	1 mL			10 uL		
580-134720-A-1 MS	AF-RHMW17-WGN01L F-2312	8015C GLY	T	1 mL		10 uL	10 uL		
580-134720-A-1 MSD	AF-RHMW17-WGN01L F-2312	8015C GLY	T	1 mL		10 uL	10 uL		
CCV 680-812978/33		8015C GLY		1 mL	10 uL		10 uL		

Batch Notes	

Basis	Basis Description
T	Total/NA


The pound sign (#) in the amount added field denotes that the reagent was used undiluted. All calculations are performed using the stated concentration for this reagent.

# Subcontract Data

# Shipping and Receiving Documents

**Chain of Custody Record**

AWB: 016-97613596

<b>Client Information</b>		Sampler: <i>Max Galtin</i>		Lab PM: Elaine Walker		Carrier Tracking No(s): <i>07713 FedEx United Cargo</i>		COC No: 2312AFE10			
Client Contact:		Phone: <i>808-621-3051</i>		E-Mail: <i>M.Elaine.Walker@EurofinsET.com</i>		State of Origin: Hawaii		Page: Page 1 of 1			
Company: AECOM				PWSID:		<b>Analysis Requested</b>					
Address: 1001 Bishop St. Suite 1600		Due Date Requested: See subcontract		Field Filtered Sample (Yes or No) <input checked="" type="checkbox"/> Perform MSMSD (Yes or No) <input checked="" type="checkbox"/> 8015C_DAL_GL_DS/2-(p-butoxyethoxy)ethanol		Total Number of containers 3		Preservation Codes: A - HCL                      M - Hexane B - NaOH                    N - None C - Zn Acetate              O - AsNaO2 D - Nitric Acid              P - Na2O4S E - NaHSO4                 Q - Na2SO3 F - MeOH                    R - Na2S2O3 G - Amchlor                 S - H2SO4 H - Ascorbic Acid          T - TSP Dodecahydrate I - Ice                         U - Acetone J - DI Water                 V - MCAA K - EDTA                    W - pH 4-5 L - EDA                      Z - other (specify)			
City: Honolulu		TAT Requested (days): <b>Rush - 5 Day</b>									
State, Zip: Hawaii 96813		Compliance Project: <input type="checkbox"/> Yes <input type="checkbox"/> No									
Phone: 808-954-4512 / 770-331-0794		PO #:									
Email: Watson Tanji (watson.tanji@aecom.com) / Mark Kromis (mark.kromis@aecom.com)		WO #:									
Project Name: CTO N6274223F0104		Project #: 60697810		Project #: 60697810		Project #: 60697810		Project #: 60697810			
Site: RHSF		SSOW#:		SSOW#:		SSOW#:		SSOW#:			
<b>Sample Identification</b>		Sample Date		Sample Time		Sample Type (C=Comp, G=grab)		Matrix (W=water, S=solid, O=soil, BT=Tissue, A=Air)			
AF-RHMW17-WGN01LF-2312		<i>12/6/23</i>		<i>1150</i>		G      W		N      N      X			
 580-134720 Chain of Custody											
<b>Possible Hazard Identification</b> <input checked="" type="checkbox"/> Non-Hazard <input type="checkbox"/> Flammable <input type="checkbox"/> Skin Irritant <input type="checkbox"/> Poison B <input type="checkbox"/> Unknown <input type="checkbox"/> Radiological					<b>Sample Disposal (A fee may be assessed if samples are retained longer than 1 month)</b> <input type="checkbox"/> Return To Client <input checked="" type="checkbox"/> Disposal By Lab <input type="checkbox"/> Archive For _____ Months						
Deliverable Requested: I, II, III, IV, Other (specify)					Prelim data (Level 1or2)=see TAT above. DoD Stage 4 report standard TAT. AECOM EQUS FDD.						
Special Instructions/QC Requirements: DOD QSM project.											
Empty Kit Relinquished by:		Date:		Time:		Method of Shipment:					
Relinquished by: <i>Max Galtin</i>		Date/Time: <i>12/6/23 1335</i>		Company: AECOM		Received by: <i>Alex Edwards</i>		Date/Time: <i>12/6/23 1335</i>		Company: AECOM	
Relinquished by: <i>Alex Edwards</i>		Date/Time: <i>12/7/23 1350</i>		Company: AECOM		Received by: <i>[Signature]</i>		Date/Time: <i>12/8/23 1600</i>		Company:	
Custody Seals Intact: <input checked="" type="checkbox"/> Yes <input type="checkbox"/> No		Custody Seal No.:		Cooler Temperature (°C) and Other Remarks: <i>4.7/5.0</i>							

### Chain of Custody Record


**Environment Testing America**  
*United AWD 4/7/23*  
*016-97613580*

<b>Client Information</b>			Sampler: <i>Max Gattlin</i>			Lab PM: Elaine Walker			Carrier Tracking No(s): FedEx			COC No: 2312AFEA11		
Client Contact:			Phone: <i>808-521-3051</i>			E-Mail: <i>M.Elaine.Walker@EurofinsET.com</i>			State of Origin: Hawaii			Page: Page 1 of 1		
Company: AECOM			PWSID:			<b>Analysis Requested</b>						Job #:		
Address: 1001 Bishop St. Suite 1600			Due Date Requested: see subcontract			<div style="display: flex; justify-content: space-between;"> <div style="writing-mode: vertical-rl; transform: rotate(180deg);">Field Filtered Sample (Yes or No)</div> <div style="writing-mode: vertical-rl; transform: rotate(180deg);">Perform MS/MSD (Yes or No)</div> <div style="writing-mode: vertical-rl; transform: rotate(180deg);">8015C_DAL_GL_DS/2-(2-butoxyethoxy)-ethanol</div> </div> <p style="font-size: 2em; opacity: 0.5; transform: rotate(45deg); position: absolute; top: 50%; left: 50%;"><i>Max Gattlin 12/16/23</i></p>						<b>Preservation Codes:</b> A - HCL                    M - Hexane B - NaOH                N - None C - Zn Acetate        O - AsNaO2 D - Nitric Acid        P - Na2O4S E - NaHSO4            Q - Na2SO3 F - MeOH                R - Na2S2O3 G - Amchlor            S - H2SO4 H - Ascorbic Acid    T - TSP Dodecahydrate I - Ice                     U - Acetone J - DI Water            V - MCAA K - EDTA                W - pH 4-5 L - EDA                  Z - other (specify)		
City: Honolulu			TAT Requested (days): <b>Rush - 5 Day</b>											
State, Zip: Hawaii 96813			Compliance Project: <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No											
Phone: 808-954-4512 / 770-331-0794			PO #:											
Email: Watson Tanji (watson.tanji@aecom.com)/ Mark Kromis (mark.kromis@aecom.com)			WO #:											
Project Name: N6274223F0104			Project #: 60697810			Total Number of Containers			Special Instructions/Note:					
Site: R:MSF			SSOW#:											

Sample Identification	Sample Date	Sample Time	Sample Type (C=comp, G=grab)	Matrix (W=water, Solid, Overwater/oil, BT=Tissue, A=Air)	Field Filtered Sample (Yes or No)	Perform MS/MSD (Yes or No)	8015C_DAL_GL_DS/2-(2-butoxyethoxy)-ethanol	Total Number of Containers	Special Instructions/Note:
AF-RHMW17D-WGN01LF-2312	12/6/23	1040	G	W	N	N	X	3	
AF-RHMW17D-WQFB01-2312	12/6/23	1010	G	W	N	N	X	3	
<div style="display: flex; justify-content: space-around; align-items: center;"> <p><i>Max Gattlin</i></p> <p><i>12/16/23</i></p> </div>									

<b>Possible Hazard Identification</b> <input checked="" type="checkbox"/> Non-Hazard <input type="checkbox"/> Flammable <input type="checkbox"/> Skin Irritant <input type="checkbox"/> Poison B <input type="checkbox"/> Unknown <input type="checkbox"/> Radiological					<b>Sample Disposal (A fee may be assessed if samples are retained longer than 1 month)</b> <input type="checkbox"/> Return To Client <input checked="" type="checkbox"/> Disposal By Lab <input type="checkbox"/> Archive For _____ Months						
Deliverable Requested: I, II, III, IV, Other (specify)			Prelim data (Level 1or2)=see TAT above. DoD Stage 4 report standard TAT_AECOM/EQ/US EDD.			Special Instructions/QC Requirements: DOD QSM project.					
Empty Kit Relinquished by:					Date:		Time:		Method of Shipment:		
Relinquished by: <i>Max Gattlin</i>		Date/Time: <i>12/16/23 1335</i>		Company: AECOM		Received by: <i>Alex Edmonds</i>		Date/Time: <i>12/16/23/1335</i>		Company: AECOM	
Relinquished by: <i>Alex Edmonds</i>		Date/Time: <i>12/16/23 1350</i>		Company: AECOM		Received by: <i>[Signature]</i>		Date/Time: <i>12/16/23 1600</i>		Company:	
Custody Seals Intact: <input checked="" type="checkbox"/> Yes <input type="checkbox"/> No			Custody Seal No.:			Cooler Temperature(s) °C and Other Remarks: <i>U. 7/5.0</i>					

### Chain of Custody Record

016-976 13596

<b>Client Information</b>			Sampler: <i>Max Gattin</i>	Lab PM: Elaine Walker		Carrier Tracking No(s): <i>FeDEX</i>	COC No: 2312AFE12																																																																																																																																																
Client Contact:			Phone: 800-521-3051	E-Mail: M.Elaine.Walker@EurofinsET.com	State of Origin: Hawaii	Page: Page 1 of 1	Job #:																																																																																																																																																
Company: AECOM		PWSID:		<b>Analysis Requested</b>				<b>Preservation Codes:</b> A - HCL M - Hexane B - NaOH N - None C - Zn Acetate O - AsNaO2 D - Nitric Acid P - Na2O4S E - NaHSO4 Q - Na2SO3 F - MeOH R - Na2S2O3 G - Amchlor S - H2SO4 H - Ascorbic Acid T - TSP Dodecahydrate I - Ice U - Acetone J - DI Water V - MCAA K - EDTA W - pH 4-5 L - EDA Z - other (specify)  Other:																																																																																																																																															
Address: 1001 Bishop St. Suite 1600		Due Date Requested: see subcontract		Field Filtered Sample (Yes or No) <i>Max Gattin 12/6/23</i> Perform MSMSD (Yes or No) 8015C_DAL_GL_DS/2-(2-butoxyethoxy)ethanol																																																																																																																																																			
City: Honolulu		TAT Requested (days): <b>Rush - 5 Day</b>																																																																																																																																																					
State, Zip: Hawaii 96813		Compliance Project: <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No																																																																																																																																																					
Phone: 808-954-4512 / 770-331-0794		PO #:		<table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Sample Identification</th> <th>Sample Date</th> <th>Sample Time</th> <th>Sample Type (C=comp, G=grab)</th> <th>Matrix (W=water, S=solid, O=organic, BT=leach, A=Air)</th> <th>Field Filtered Sample (Yes or No)</th> <th>Perform MSMSD (Yes or No)</th> <th>Field Filtered Sample (Yes or No)</th> <th>Field Filtered Sample (Yes or No)</th> <th>Field Filtered Sample (Yes or No)</th> <th>Field Filtered Sample (Yes or No)</th> <th>Field Filtered Sample (Yes or No)</th> <th>Field Filtered Sample (Yes or No)</th> <th>Field Filtered Sample (Yes or No)</th> <th>Total Number of Containers</th> <th>Special Instructions/Note:</th> </tr> </thead> <tbody> <tr> <td colspan="16" style="text-align: center;">Preservation Code: A</td> </tr> <tr> <td>AF-RHMW17S-WGN01LF-2312</td> <td>12/6/23</td> <td>0905</td> <td>G</td> <td>W</td> <td>N</td> <td>N</td> <td>X</td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td>3</td> <td></td> </tr> <tr> <td>AF-RHMW17S-WQEB01-2312</td> <td>12/6/23</td> <td>0940</td> <td>G</td> <td>W</td> <td>N</td> <td>N</td> <td>X</td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td>3</td> <td></td> </tr> <tr><td> </td><td> </td><td> </td><td> </td><td> </td><td> </td><td> </td><td> </td><td> </td><td> </td><td> </td><td> </td><td> </td><td> </td><td> </td><td> </td></tr> <tr><td> </td><td> </td><td> </td><td> </td><td> </td><td> </td><td> </td><td> </td><td> </td><td> </td><td> </td><td> </td><td> </td><td> </td><td> </td><td> </td></tr> <tr><td> </td><td> </td><td> </td><td> </td><td> </td><td> </td><td> </td><td> </td><td> </td><td> </td><td> </td><td> </td><td> </td><td> </td><td> </td><td> </td></tr> <tr><td> </td><td> </td><td> </td><td> </td><td> </td><td> </td><td> </td><td> </td><td> </td><td> </td><td> </td><td> </td><td> </td><td> </td><td> </td><td> </td></tr> <tr><td> </td><td> </td><td> </td><td> </td><td> </td><td> </td><td> </td><td> </td><td> </td><td> </td><td> </td><td> </td><td> </td><td> </td><td> </td><td> </td></tr> </tbody> </table>				Sample Identification	Sample Date	Sample Time	Sample Type (C=comp, G=grab)	Matrix (W=water, S=solid, O=organic, BT=leach, A=Air)	Field Filtered Sample (Yes or No)	Perform MSMSD (Yes or No)	Field Filtered Sample (Yes or No)	Field Filtered Sample (Yes or No)	Field Filtered Sample (Yes or No)	Field Filtered Sample (Yes or No)	Field Filtered Sample (Yes or No)	Field Filtered Sample (Yes or No)	Field Filtered Sample (Yes or No)	Total Number of Containers	Special Instructions/Note:	Preservation Code: A																AF-RHMW17S-WGN01LF-2312	12/6/23	0905	G	W	N	N	X							3		AF-RHMW17S-WQEB01-2312	12/6/23	0940	G	W	N	N	X							3																																																																																	
Sample Identification	Sample Date	Sample Time	Sample Type (C=comp, G=grab)					Matrix (W=water, S=solid, O=organic, BT=leach, A=Air)	Field Filtered Sample (Yes or No)	Perform MSMSD (Yes or No)	Field Filtered Sample (Yes or No)	Field Filtered Sample (Yes or No)	Field Filtered Sample (Yes or No)	Field Filtered Sample (Yes or No)	Field Filtered Sample (Yes or No)	Field Filtered Sample (Yes or No)	Field Filtered Sample (Yes or No)	Total Number of Containers	Special Instructions/Note:																																																																																																																																				
Preservation Code: A																																																																																																																																																							
AF-RHMW17S-WGN01LF-2312	12/6/23	0905	G					W	N	N	X							3																																																																																																																																					
AF-RHMW17S-WQEB01-2312	12/6/23	0940	G	W	N	N	X							3																																																																																																																																									
Email: Watson Tanji (watson.tanji@aecom.com) / Mark Kromis (mark.kromis@aecom.com)		WO #:		8015C_DAL_GL_DS/2-(2-butoxyethoxy)ethanol <i>Max Gattin 12/6/23</i>																																																																																																																																																			
Project Name: CTO N6274223F0104		Project #: 60697810																																																																																																																																																					
Site: RHSF		SSOW#:		<table border="1" style="width:100%; border-collapse: collapse;"> <tr> <td colspan="2">Possible Hazard Identification</td> <td colspan="2">Sample Disposal (A fee may be assessed if samples are retained longer than 1 month)</td> </tr> <tr> <td colspan="2"> <input checked="" type="checkbox"/> Non-Hazard <input type="checkbox"/> Flammable <input type="checkbox"/> Skin Irritant <input type="checkbox"/> Poison B <input type="checkbox"/> Unknown <input type="checkbox"/> Radiological         </td> <td colspan="2"> <input type="checkbox"/> Return To Client <input checked="" type="checkbox"/> Disposal By Lab <input type="checkbox"/> Archive For _____ Months         </td> </tr> <tr> <td colspan="2">Deliverable Requested: I, II, III, IV, Other (specify)</td> <td colspan="2">Special Instructions/QC Requirements: DOD QSM project.</td> </tr> </table>				Possible Hazard Identification		Sample Disposal (A fee may be assessed if samples are retained longer than 1 month)		<input checked="" type="checkbox"/> Non-Hazard <input type="checkbox"/> Flammable <input type="checkbox"/> Skin Irritant <input type="checkbox"/> Poison B <input type="checkbox"/> Unknown <input type="checkbox"/> Radiological		<input type="checkbox"/> Return To Client <input checked="" type="checkbox"/> Disposal By Lab <input type="checkbox"/> Archive For _____ Months		Deliverable Requested: I, II, III, IV, Other (specify)		Special Instructions/QC Requirements: DOD QSM project.																																																																																																																																					
Possible Hazard Identification		Sample Disposal (A fee may be assessed if samples are retained longer than 1 month)																																																																																																																																																					
<input checked="" type="checkbox"/> Non-Hazard <input type="checkbox"/> Flammable <input type="checkbox"/> Skin Irritant <input type="checkbox"/> Poison B <input type="checkbox"/> Unknown <input type="checkbox"/> Radiological		<input type="checkbox"/> Return To Client <input checked="" type="checkbox"/> Disposal By Lab <input type="checkbox"/> Archive For _____ Months																																																																																																																																																					
Deliverable Requested: I, II, III, IV, Other (specify)		Special Instructions/QC Requirements: DOD QSM project.																																																																																																																																																					
Empty Kit Relinquished by:		Date:		Time:		Method of Shipment:																																																																																																																																																	
Relinquished by: <i>Max Gattin</i>		Date/Time: 12/6/23 1325		Company: AECOM		Received by: <i>Alex Edmonds</i>																																																																																																																																																	
Relinquished by: <i>Alex Edmonds</i>		Date/Time: 12/7/23 1350		Company: AECOM		Received by: <i>Alex Edmonds</i>																																																																																																																																																	
Relinquished by:		Date/Time:		Company:		Received by:																																																																																																																																																	
Custody Seals Intact: <input checked="" type="checkbox"/> Yes <input type="checkbox"/> No		Custody Seal No.:		Cooler Temperature(s) C and Other Remarks: <i>4.7/5.0</i>																																																																																																																																																			

**Eurofins FGS, Seattle**

5755 8th Street East  
Tacoma, WA 98424

**Chain of Custody Record**

UNITED AWB  
12/13/23  
D10-9752984

<b>Client Information</b>				Sampler: <i>Josie Vilangos</i>	Lab PM: Elaine Walker	Carrier Tracking No(s): FedEx <i>MM 12/7</i>	COC No: 2312AFEA05																																																																																																					
Client Contact:				Phone: <i>224-645-7852</i>	E-Mail: <i>M.Elaine.Walker@EurofinsET.com</i>	State of Origin: Hawaii	Page: Page 1 of 1																																																																																																					
Company: <b>AECOM</b>			PWSID:	<b>Analysis Requested</b>			Job #:																																																																																																					
Address: 1001 Bishop St. Suite 1600 City: Honolulu State, Zip: Hawaii 96813		Due Date Requested: see subcontract TAT Requested (days): <b>Rush - 5 Day</b>					<div style="text-align: center; font-size: 2em; font-family: cursive;"> <i>12/7/23</i> </div> <p style="writing-mode: vertical-rl; transform: rotate(180deg);">8016C_DAI_GL_DS1/2-(2-butyloxyethoxy)-ethanol</p> <p style="writing-mode: vertical-rl; transform: rotate(180deg);">Field Filtered Sample (Yes or No)</p> <p style="writing-mode: vertical-rl; transform: rotate(180deg);">Perform MS/MSD (Yes or No)</p> <p style="writing-mode: vertical-rl; transform: rotate(180deg);">Total Number of containers</p>																																																																																																					
Project Name: CTO N6274223F0104			Compliance Project: <input type="checkbox"/> Yes <input type="checkbox"/> No		Preservation Codes: A - HCL                      M - Hexane B - NaOH                      N - None C - Zn Acetate                O - AsNaO2 D - Nitric Acid                P - Na2O4S E - NaHSO4                    Q - Na2SO3 F - MeOH                      R - Na2S2O3 G - Amchlor                    S - H2SO4 H - Ascorbic Acid            T - TSP Dodecahydrate I - Ice                            U - Acetone J - DI Water                    V - MCAA K - EDTA                      W - pH 4-5 L - EDTA                      Z - other (specify) Other:																																																																																																							
Site: RHSF			Project #: 60697810							Special Instructions/Note:																																																																																																		
<table border="1"> <thead> <tr> <th>Sample Identification</th> <th>Sample Date</th> <th>Sample Time</th> <th>Sample Type (C=Comp, G=grab)</th> <th>Matrix (Water, Sewage, Wastewater, BT-Tissue, Air)</th> <th>Field Filtered Sample (Yes or No)</th> <th>Perform MS/MSD (Yes or No)</th> <th>8016C_DAI_GL_DS1/2-(2-butyloxyethoxy)-ethanol</th> <th>Total Number of containers</th> </tr> </thead> <tbody> <tr> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> </tr> <tr> <td><i>AF-RHMW12A-WGN01LF-2312</i></td> <td><i>12/7/23</i></td> <td><i>0915</i></td> <td><i>G</i></td> <td><i>W</i></td> <td><i>N</i></td> <td><i>N</i></td> <td><i>X</i></td> <td><i>3</i></td> </tr> <tr> <td><i>AF-RHMW12A-WGFD01LF-2312</i></td> <td><i>12/7/23</i></td> <td><i>0915</i></td> <td><i>G</i></td> <td><i>W</i></td> <td><i>N</i></td> <td><i>N</i></td> <td><i>X</i></td> <td><i>3</i></td> </tr> <tr> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> </tr> <tr> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> </tr> <tr> <td colspan="4">                 Possible Hazard Identification  <input checked="" type="checkbox"/> Non-Hazard    <input type="checkbox"/> Flammable    <input type="checkbox"/> Skin Irritant    <input type="checkbox"/> Poison B    <input type="checkbox"/> Unknown    <input type="checkbox"/> Radiological             </td> <td colspan="4">                 Sample Disposal (A fee may be assessed if samples are retained longer than 1 month)  <input type="checkbox"/> Return To Client    <input checked="" type="checkbox"/> Disposal By Lab    <input type="checkbox"/> Archive For _____ Months             </td> </tr> <tr> <td colspan="4">Deliverable Requested: I, II, III, IV, Other (specify)</td> <td colspan="4">Prelim data (Level 1or2)=see TAT above. DoD Stage 4 report standard TAT. AECOM EQUIS EDD.</td> </tr> <tr> <td colspan="2">Empty Kit Relinquished by:</td> <td>Date:</td> <td>Time:</td> <td colspan="4">Method of Shipment:</td> </tr> <tr> <td colspan="2">Relinquished by: <i>Michaela Nain</i></td> <td>Date/Time: <i>12/7/23 1330</i></td> <td>Company: <i>AECOM</i></td> <td colspan="2">Received by: <i>Michaela Nain</i></td> <td>Date/Time: <i>12/7/23 1330</i></td> <td>Company: <i>AECOM</i></td> </tr> <tr> <td colspan="2">Relinquished by: <i>Michaela Nain</i></td> <td>Date/Time: <i>12/7/23 1350</i></td> <td>Company: <i>AECOM</i></td> <td colspan="2">Received by: <i>[Signature]</i></td> <td>Date/Time: <i>12/2/23 1600</i></td> <td>Company: <i>[Blank]</i></td> </tr> <tr> <td colspan="4">Custody Seals Intact: <input checked="" type="checkbox"/></td> <td colspan="4">Cooler Temperature(s)/C and Other Remarks:</td> </tr> </tbody> </table>				Sample Identification	Sample Date	Sample Time						Sample Type (C=Comp, G=grab)	Matrix (Water, Sewage, Wastewater, BT-Tissue, Air)	Field Filtered Sample (Yes or No)	Perform MS/MSD (Yes or No)	8016C_DAI_GL_DS1/2-(2-butyloxyethoxy)-ethanol	Total Number of containers										<i>AF-RHMW12A-WGN01LF-2312</i>	<i>12/7/23</i>	<i>0915</i>	<i>G</i>	<i>W</i>	<i>N</i>	<i>N</i>	<i>X</i>	<i>3</i>	<i>AF-RHMW12A-WGFD01LF-2312</i>	<i>12/7/23</i>	<i>0915</i>	<i>G</i>	<i>W</i>	<i>N</i>	<i>N</i>	<i>X</i>	<i>3</i>																			Possible Hazard Identification <input checked="" type="checkbox"/> Non-Hazard <input type="checkbox"/> Flammable <input type="checkbox"/> Skin Irritant <input type="checkbox"/> Poison B <input type="checkbox"/> Unknown <input type="checkbox"/> Radiological				Sample Disposal (A fee may be assessed if samples are retained longer than 1 month) <input type="checkbox"/> Return To Client <input checked="" type="checkbox"/> Disposal By Lab <input type="checkbox"/> Archive For _____ Months				Deliverable Requested: I, II, III, IV, Other (specify)				Prelim data (Level 1or2)=see TAT above. DoD Stage 4 report standard TAT. AECOM EQUIS EDD.				Empty Kit Relinquished by:		Date:	Time:	Method of Shipment:				Relinquished by: <i>Michaela Nain</i>		Date/Time: <i>12/7/23 1330</i>	Company: <i>AECOM</i>	Received by: <i>Michaela Nain</i>		Date/Time: <i>12/7/23 1330</i>	Company: <i>AECOM</i>	Relinquished by: <i>Michaela Nain</i>		Date/Time: <i>12/7/23 1350</i>	Company: <i>AECOM</i>	Received by: <i>[Signature]</i>		Date/Time: <i>12/2/23 1600</i>	Company: <i>[Blank]</i>	Custody Seals Intact: <input checked="" type="checkbox"/>				Cooler Temperature(s)/C and Other Remarks:	
Sample Identification	Sample Date	Sample Time	Sample Type (C=Comp, G=grab)	Matrix (Water, Sewage, Wastewater, BT-Tissue, Air)	Field Filtered Sample (Yes or No)	Perform MS/MSD (Yes or No)				8016C_DAI_GL_DS1/2-(2-butyloxyethoxy)-ethanol	Total Number of containers																																																																																																	
<i>AF-RHMW12A-WGN01LF-2312</i>	<i>12/7/23</i>	<i>0915</i>	<i>G</i>	<i>W</i>	<i>N</i>	<i>N</i>				<i>X</i>	<i>3</i>																																																																																																	
<i>AF-RHMW12A-WGFD01LF-2312</i>	<i>12/7/23</i>	<i>0915</i>	<i>G</i>	<i>W</i>	<i>N</i>	<i>N</i>				<i>X</i>	<i>3</i>																																																																																																	
Possible Hazard Identification <input checked="" type="checkbox"/> Non-Hazard <input type="checkbox"/> Flammable <input type="checkbox"/> Skin Irritant <input type="checkbox"/> Poison B <input type="checkbox"/> Unknown <input type="checkbox"/> Radiological				Sample Disposal (A fee may be assessed if samples are retained longer than 1 month) <input type="checkbox"/> Return To Client <input checked="" type="checkbox"/> Disposal By Lab <input type="checkbox"/> Archive For _____ Months																																																																																																								
Deliverable Requested: I, II, III, IV, Other (specify)				Prelim data (Level 1or2)=see TAT above. DoD Stage 4 report standard TAT. AECOM EQUIS EDD.																																																																																																								
Empty Kit Relinquished by:		Date:	Time:	Method of Shipment:																																																																																																								
Relinquished by: <i>Michaela Nain</i>		Date/Time: <i>12/7/23 1330</i>	Company: <i>AECOM</i>	Received by: <i>Michaela Nain</i>		Date/Time: <i>12/7/23 1330</i>	Company: <i>AECOM</i>																																																																																																					
Relinquished by: <i>Michaela Nain</i>		Date/Time: <i>12/7/23 1350</i>	Company: <i>AECOM</i>	Received by: <i>[Signature]</i>		Date/Time: <i>12/2/23 1600</i>	Company: <i>[Blank]</i>																																																																																																					
Custody Seals Intact: <input checked="" type="checkbox"/>				Cooler Temperature(s)/C and Other Remarks:																																																																																																								

*4.7/5.0*

# Login Sample Receipt Checklist

Client: AECOM

Job Number: 580-134720-1

**Login Number: 134720**  
**List Number: 1**  
**Creator: Johnson, Corey M**

**List Source: Eurofins Seattle**

Question	Answer	Comment
Radioactivity wasn't checked or is $\leq$ background as measured by a survey meter.	N/A	
The cooler's custody seal, if present, is intact.	True	
Sample custody seals, if present, are intact.	True	
The cooler or samples do not appear to have been compromised or tampered with.	True	
Samples were received on ice.	True	
Cooler Temperature is acceptable.	True	
Cooler Temperature is recorded.	True	
COC is present.	True	
COC is filled out in ink and legible.	True	
COC is filled out with all pertinent information.	True	
Is the Field Sampler's name present on COC?	True	
There are no discrepancies between the containers received and the COC.	True	
Samples are received within Holding Time (excluding tests with immediate HTs)	True	
Sample containers have legible labels.	True	
Containers are not broken or leaking.	True	
Sample collection date/times are provided.	True	
Appropriate sample containers are used.	True	
Sample bottles are completely filled.	True	
Sample Preservation Verified.	N/A	
There is sufficient vol. for all requested analyses, incl. any requested MS/MSDs	True	
Containers requiring zero headspace have no headspace or bubble is $<6\text{mm}$ (1/4").	N/A	
Multiphasic samples are not present.	True	
Samples do not require splitting or compositing.	True	
Residual Chlorine Checked.	N/A	