



# ANALYTICAL SUMMARY REPORT

January 20, 2022

AECOM - Honolulu  
1001 Bishop Street, Suite 1600  
Honolulu HI, 96813-3698

Work Order: B21121020 Quote ID: 5912

Project Name: CV18F0126/60571032.02.20.01

Energy Laboratories Inc Billings MT received the following 6 samples from AECOM - Honolulu on 12/11/2021 for analysis.

Lab ID	Client Sample ID	Collect Date	Received Date	Matrix	Test
B21121020-001	ERH2213 (RHMW01R)	12/08/21 15:10	12/11/2021	Ground Water	DRO-Liquid-Liquid Extraction SW3520C Separatory Funnel SW3510C Liquid-Liquid Ext. Carbon, Total Organic SW9060A 8260-Volatile Organic Compounds-BTEX SW8260B Gasoline Range Organics SW8015C Diesel Range Organics SW8015C Low Level PAH SW8270C
B21121020-002	ERH2215 (RHMW02)	12/08/21 16:35	12/11/2021	Ground Water	Same As Above
B21121020-003	ERH2212 (Trip Blank)-14525	12/08/21 15:05	12/11/2021	Trip Blank	Gasoline Range Organics SW8015C
B21121020-004	ERH2212 (Trip Blank)-14525	12/08/21 15:05	12/11/2021	Trip Blank	8260-Volatile Organic Compounds-BTEX SW8260B
B21121020-005	ERH2214 (Trip Blank)-14525	12/08/21 16:35	12/11/2021	Trip Blank	Gasoline Range Organics SW8015C
B21121020-006	ERH2214 (Trip Blank)-14525	12/08/21 16:35	12/11/2021	Trip Blank	8260-Volatile Organic Compounds-BTEX SW8260B

The analyses presented in this report were performed by Energy Laboratories, Inc., 1120 S 27th St., Billings, MT 59101, unless otherwise noted. Any exceptions or problems with the analyses are noted in the report package. Any issues encountered during sample receipt are documented in the Work Order Receipt Checklist.

The results as reported relate only to the item(s) submitted for testing. This report shall be used or copied only in its entirety. Energy Laboratories, Inc. is not responsible for the consequences arising from the use of a partial report.

If you have any questions regarding these test results, please contact your Project Manager.

Report Approved By:



**CLIENT:** AECOM - Honolulu  
**Project:** CV18F0126/60571032.02.20.01  
**Work Order:** B21121020

**Report Date:** 1/20/2022

## CASE NARRATIVE

### General Comments:

For any question please contact your Project Manager at (406) 252-6325 or [billingspm@energylab.com](mailto:billingspm@energylab.com).

All analyses have been performed in accordance with DOD QSM Version 5.3 unless otherwise noted below. The specific methodologies used in obtaining the enclosed analytical results are indicated on the Analytical Summary Report and the Laboratory Analytical Report. The samples were received at the appropriate temperature and in accordance with the chain of custody unless otherwise noted in the Work Order Receipt Checklist.

The tests listed below are accredited and meet the requirements of DoD QSM Version 5.3 as verified by ANSI-ASQ National Accreditation Board (ANAB) certificate number ADE-2588. Exceptions to this require client authorization and records documenting this approval are attached in the Sample Management Records. Accreditation may not be offered or required for all methods and analytes reported in this package. Refer to the certificate and scope of accreditation located at <https://www.energylab.com/whyus/certifications-quality-control/> or contact your project manager.

Tests for Total Organic Carbon by SW0060A associated with analyst identified as ELI-CA were subcontracted to Energy Laboratories, PO Box 247, Casper, WY, EPA Number WY00002.

Project specific matrix quality control samples may not be reported if site specific samples were not submitted. Matrix quality control samples were performed on project samples where adequate volume was available. All quality control measures met criteria unless otherwise noted in the Analytical QC Exceptions report and in the Analysis Specific Comments below. Where available, sample management records are attached.

The Level IV Validation Package includes data reports for all analyses associated with the instrument calibration, quality control (QC) sample analysis, and sample analysis. All analytical data is within method specifications except as noted in the Analytical QC Exceptions report or the Analysis Specific Comments below. The analytical report identifies preparation batch and analytical run IDs associated with each result for a sample. Only the raw data associated with the parameters listed on this report should be validated.

### Analysis Specific Comments:

An Analytical QC Exceptions Report has been attached, summarizing all qualified QC results. Corrective actions regarding Semi-Volatile Organic Compounds by EPA8270C analysis are summarized below.

#### Method SW8270C:

MB-162126, LLCS-162126, B21121001-001ALMS

A documented laboratory error occurred during extraction, which contributed to low surrogate and analyte recoveries in these samples. The surrogate Nitrobenzene-d5 and analyte Naphthalene were not recovered in the MB-162126 and LLCS-162126. The associated LCSDuplicate sample, LLCSD-162126, had normal recoveries. Target analytes were not detected in sample ERH2215 (RHMW02), B21121020-001.

B21121020-002ALMS

The recovery for Naphthalene was above the quality control limit in the matrix spike sample, which may indicate a non-homogenous sample matrix between the two bottles submitted.



Trust our People. Trust our Data.

# Chain of Custody & Analytical Request Record

www.energylab.com

COC # 202112-05-NOI

Page 1 of 2

### Account Information (Billing information)

Company/Name	AECOM	
Contact	Alethea Ramos / Margie Pascua	
Phone	808-529-7283 / 808-356-5373	
Mailing Address	1001 Bishop St, Suite 1600	
City, State, Zip	Honolulu, Hawaii 96813	
Email	alethea.ramos / margie.pascua@aecom.com	
Receive Invoice	<input type="checkbox"/> Hard Copy <input checked="" type="checkbox"/> Email	Receive Report <input type="checkbox"/> Hard Copy <input checked="" type="checkbox"/> Email
Purchase Order	Quote	Bottle Order
N/A	N/A	N/A

### Report Information (if different than Account Information)

Company/Name	AECOM	
Contact	see Account information	
Phone		
Mailing Address		
City, State, Zip		
Email	USAPimaging@aecom.com	
Receive Report	<input type="checkbox"/> Hard Copy <input checked="" type="checkbox"/> Email	
Special Report/Formats:	<input checked="" type="checkbox"/> LEVEL IV <input checked="" type="checkbox"/> INELAC <input checked="" type="checkbox"/> EDD/EDT (contact laboratory) <input type="checkbox"/> Other _____	

### Comments

1. Project performed under DoD QSM
2. TPH-DRO/RRO needs 3520 extraction
3. Preliminary data (or Level 1) in 1-2 business days; Level IV report in 10 working days.
4. Note: NOI log is separate from other COCs.
5. \*PAH 8270D SIM: 1-methylnaphthalene, 2-methylnaphthalene, naphthalene

### Project Information

Project Name, PWSID, Permit, etc	CV18F0126/60571032.02.20.01	
Sampler Name GM, MY, NL, AM, DM, CF	Sampler Phone 808-987-3201	
Sample Origin State Hawaii	EPA/State Compliance <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No	
URANIUM MINING CLIENTS MUST indicate sample type		
<input type="checkbox"/> Unprocessed Ore		
<input type="checkbox"/> Processed Ore (Ground or Refined) **CALL BEFORE SENDING		
<input type="checkbox"/> 11(e)2 @product Material (Can ONLY be Submitted to ELI Casper Location)		

### Matrix Codes

- A - Air
- W - Water
- S - Soils/Solids
- V - Vegetation
- B - Bioassay
- O - Oil
- DW - Drinking Water

### Analysis Requested

	BTEX by 8260	TPH-g by 8015	PAHs by 8270D SIM*	EPA 8015 TPH-DRO/RRO	EPA 8015 TPH-DRO/RRO w/SGC	TOC by 9060	8270D SVOC Full List											
1	✓	✓																
2	✓	✓	✓	✓	✓	✓												
3	✓	✓																
4	✓	✓	✓	✓	✓	✓												
5																		
6																		
7																		
8																		
9																		

All turnaround times are standard unless marked as RUSH.  
Energy Laboratories MUST be contacted prior to RUSH sample submittal for charges and scheduling - See Instructions Page

Sample Identification (Name, Location, Interval, etc.)	Collection		Number of Containers	Matrix (See Codes Above)	BTEX by 8260	TPH-g by 8015	PAHs by 8270D SIM*	EPA 8015 TPH-DRO/RRO	EPA 8015 TPH-DRO/RRO w/SGC	TOC by 9060	8270D SVOC Full List	See Attached	RUSH TAT	ELI LAB ID Laboratory Use Only
	Date	Time												
1 ERH2212 (Trip Blank)	12/08/2021	11:05 am	0	WQ	✓	✓							✗	BA1121220
2 ERH2213 (RHMW01R)	12/08/2021	11:10 am	4	GW	✓	✓	✓	✓	✓	✓			✗	
3 ERH2214 (Trip Blank)	12/08/2021	12:30 pm	0	WQ	✓	✓							✗	
4 ERH2215 (RHMW02)	12/08/2021	12:35 pm	2	GW	✓	✓	✓	✓	✓	✓			✗	
5														
6														
7														
8														
9														

ELI is REQUIRED to provide preservative traceability. If the preservatives supplied with the bottle order were NOT used, please attach your preservative information with this COC.

Custody Record MUST be signed	Relinquished by (print) Alexander Edmonds	Date/Time 12/9/21 12:30 pm	Signature <i>Alexander Edmonds</i>	Received by (print) M. Edmonds	Date/Time 12/21 1125	Signature <i>M. Edmonds</i>			
	Relinquished by (print)	Date/Time	Signature	Received by (print)	Date/Time	Signature			
LABORATORY USE ONLY									
Shipped By	Cooler ID(s)	Custody Seals Y N C B	Intact Y N	Receipt Temp °C	Temp Blank Y N	On Ice Y N	Payment Type CC Cash Check	Amount \$	Receipt Number (cash/check only)

In certain circumstances, samples submitted to Energy Laboratories, Inc. may be subcontracted to other certified laboratories in order to complete the analysis requested. This serves as notice of this possibility. All subcontracted data will be clearly notated on your analytical report.



Trust our People. Trust our Data.

# Chain of Custody & Analytical Request Record

www.energylab.com

COC # 202112-06-NOI

Page 2 of 2

### Account Information (Billing information)

Company/Name	AECOM	
Contact	Alethea Ramos / Margie Pascua	
Phone	808-529-7283 / 808-356-5373	
Mailing Address	1001 Bishop St., Suite 1600	
City, State, Zip	Honolulu, Hawaii 96813	
Email	alethea.ramos / margie.pascua@aecom.com	
Receive Invoice	<input type="checkbox"/> Hard Copy <input checked="" type="checkbox"/> Email	Receive Report <input type="checkbox"/> Hard Copy <input checked="" type="checkbox"/> Email
Purchase Order	Quote	Bottle Order
N/A	N/A	N/A

### Report Information (If different than Account Information)

Company/Name	AECOM	
Contact	see Account information	
Phone		
Mailing Address		
City, State, Zip		
Email	USAPimaging@aecom.com	
Receive Report	<input type="checkbox"/> Hard Copy <input checked="" type="checkbox"/> Email	
Special Report/Format:	<input checked="" type="checkbox"/> LEVEL IV <input type="checkbox"/> NELAC <input checked="" type="checkbox"/> EDD/EDT (contact laboratory) <input type="checkbox"/> Other	

### Comments

1. Project performed under DoD QSM.  
 2. TPH-DRO/RRO needs 3520 extraction.  
 3. Preliminary data (or Level 1) in 1-2 business days; Level IV report in 10 working days.  
 4. Note: NOI log is separate from other COCs.  
 5. \*PAH 8270D SIM: 1-methylnaphthalene, 2-methylnaphthalene, naphthalene

### Project Information

Project Name, PWSID, Permit, etc.	CV18F0126/60571032.02.20.01		
Sampler Name	GM, MY, NL, AM, DM, CF	Sampler Phone	808-987-3201
Sample Origin State	Hawaii	EPA/State Compliance	<input type="checkbox"/> Yes <input checked="" type="checkbox"/> No
URANIUM MINING CLIENTS MUST indicate sample type			
<input type="checkbox"/> Unprocessed Ore			
<input type="checkbox"/> Processed Ore (Ground or Refined) **CALL BEFORE SENDING			
<input type="checkbox"/> 11(a)2 Byproduct Material (Can ONLY be Submitted to ELI Casper Location)			

### Matrix Codes

- A - Air
- W - Water
- S - Soils/Solids
- V - Vegetation
- B - Bioscience
- O - Oil
- DW - Drinking Water

### Analysis Requested

	BTEX by 8260	TPH-g by 8015	PAHs by 8270D SIM*	EPA8015 TPH-DRO/RRO	EPA 8015 TPH-DRO/RRO w/SGC	TOC by 9060	8270D SVOC Full List
✓	✓						
✓	✓	✓	✓	✓	✓	✓	
✓	✓						
✓	✓	✓	✓	✓	✓	✓	

See Attached

All turnaround times are standard unless marked as RUSH.  
 Energy Laboratories MUST be contacted prior to RUSH sample submittal for charges and scheduling - See Instructions Page

Sample Identification (Name, Location, Interval, etc.)	Collection		Number of Containers	Matrix (See Codes Above)	BTEX by 8260	TPH-g by 8015	PAHs by 8270D SIM*	EPA8015 TPH-DRO/RRO	EPA 8015 TPH-DRO/RRO w/SGC	TOC by 9060	8270D SVOC Full List	RUSH TAT	ELI LAB ID Laboratory Use Only
	Date	Time											
1 ERH2212 (Trip Blank)	12/08/2021	11:05 am	6	WQ	✓	✓						✓	B2112102D
2 ERH2213 (RHMW01R)	12/08/2021	11:10 am	8	GW	✓	✓	✓	✓	✓	✓		✓	
3 ERH2214 (Trip Blank)	12/08/2021	12:30 pm	6	WQ	✓	✓						✓	
4 ERH2215 (RHMW02)	12/08/2021	12:35 pm	10	GW	✓	✓	✓	✓	✓	✓		✓	
5 TB TPA 14525 2012													
6 TB BTEX 14525 2012													
7 TB TPA 14525 2014													
8 TB BTEX 14525 2014													
9													

ELI is REQUIRED to provide preservative traceability. If the preservatives supplied with the bottle order were NOT used, please attach your preservative information with this COC.

Custody Record MUST be signed	Relinquished by (print) Alexander Edmonds	Date/Time 12/9/21 12:30PM	Signature <i>Alex Edmonds</i>	Received by (print)	Date/Time	Signature
	Relinquished by (print)	Date/Time	Signature	Received by Laboratory (print) Sara Gutrich	Date/Time 12/31/21 09:50	Signature <i>Sara Gutrich</i>
LABORATORY USE ONLY						
Shipped By	Cooler ID(s)	Custody Seals Y N C B	Intact Y N	Receipt Temp 0.7 °C	Temp Blank Y N	On Ice Y N
						Payment Type CC Cash Check
					Amount \$	Receipt Number (cash/check only)

In certain circumstances, samples submitted to Energy Laboratories, Inc. may be subcontracted to other certified laboratories in order to complete the analysis requested. This serves as notice of this possibility. All subcontracted data will be clearly notated on your analytical report.



Work Order Receipt Checklist

AECOM - Honolulu

B21121020

Login completed by: Tabitha Edwards
Reviewed by: BL2000\darcy
Reviewed Date: 12/16/2021

Date Received: 12/11/2021
Received by: leb
Carrier name: FedEx

- Shipping container/cooler in good condition? Yes [x] No [ ] Not Present [ ]
Custody seals intact on all shipping container(s)/cooler(s)? Yes [x] No [ ] Not Present [ ]
Custody seals intact on all sample bottles? Yes [x] No [ ] Not Present [ ]
Chain of custody present? Yes [x] No [ ]
Chain of custody signed when relinquished and received? Yes [x] No [ ]
Chain of custody agrees with sample labels? Yes [x] No [ ]
Samples in proper container/bottle? Yes [x] No [ ]
Sample containers intact? Yes [x] No [ ]
Sufficient sample volume for indicated test? Yes [x] No [ ]
All samples received within holding time? (Exclude analyses that are considered field parameters such as pH, DO, Res Cl, Sulfite, Ferrous Iron, etc.) Yes [x] No [ ]
Temp Blank received in all shipping container(s)/cooler(s)? Yes [x] No [ ] Not Applicable [ ]
Container/Temp Blank temperature: 0.4°C On Ice
Water - VOA vials have zero headspace? Yes [x] No [ ] Not Applicable [ ]
Water - pH acceptable upon receipt? Yes [x] No [ ] Not Applicable [ ]

Standard Reporting Procedures:

Lab measurement of analytes considered field parameters that require analysis within 15 minutes of sampling such as pH, Dissolved Oxygen and Residual Chlorine, are qualified as being analyzed outside of recommended holding time.

Solid/soil samples are reported on a wet weight basis (as received) unless specifically indicated. If moisture corrected, data units are typically noted as -dry. For agricultural and mining soil parameters/characteristics, all samples are dried and ground prior to sample analysis.

Contact and Corrective Action Comments:

The collection time indicated on the Chain of Custody for all samples is in Hawaii-Aleutian Standard Time. The collection time has been converted (+4 Hours) to Mountain Standard Time.

The containers and preservative for Total Organic Carbon (TOC) by method 9060 were not provided by Energy Laboratories Inc. The VOA vials were received without the preservative traceability bottle label provided by Energy Laboratories. Proceed with analysis without preservative traceability per email from Alethea Ramos on 12/11/2021.

The following containers (Chain of Custody 2 of 2) were received on 12/13/21 by Staci Gottlob at 0.7°C on ice from Fedex Next Day Air:
ERH2212 (Trip Blank) - All containers
ERH2213 (RHMW01R) - All VOA vials
ERH2214 (Trip Blank) - All containers
ERH2215 (RHMW02) - 2 One-Liter Amber Glass Sulfuric Preserved and all VOA vials

## Qualifiers and Abbreviations

Qualifier	Qualifier Description
##	Limit of Quantitation (LOQ) for this analyte exceeds the Maximum Contaminant Level (MCL)
*	Result exceeds the Maximum Contaminant Level (MCL)
A	The analyte level was greater than four times the spike level - in accordance with the method, percent recovery is not calculated
B	Analyte detected in the method blank
C	Continuing calibration verification was outside of the quality control advisory limits
D	Limit of Quantitation (LOQ) increased due to sample matrix
E	Estimated value - result exceeds the instrument upper quantitation limit
H	Analysis performed past the method holding time
J	The reported result is an estimated value
L	Lowest Limit of Quantitation (LOQ) available for the analytical method used
N	Analyte concentration was not sufficiently high to calculate a Relative Percent Difference (RPD) for the serial dilution test
O	Diluted out
P	Poor method performance - method validations have shown no recoveries at low concentrations or method performance was erratic
Q	Values reported below the Limit of Quantitation (LOQ) are statistically invalid
R	Relative Percent Difference (RPD) exceeds advisory limit
S	Spike recovery outside of advisory limits
T	Analyte detected in the associated trip blank
U	Not detected at the Limit of Detection (LOD)
V	The RPD value for this duplicate represents the RER value and the RPD limit of 2 is the RER upper limit.

## Qualifiers and Abbreviations

### Abbreviation

Reporting	Explanation of Abbreviation
DF	Dilution Factor
DL	Detection Limit
LOD	Limit of Detection
LOQ	Limit of Quantitation
MCL	Maximum Contaminant Level
MDC	Minimum Detectable Concentration
ND	Not detected at the Limit of Quantitation (LOQ)
RBSL	Risk-Based Screening Levels
REC	Recovery
RER	Relative Error Ratio
RPD	Relative Percent Difference
SPK	Spike

Sample Types	Explanation of Abbreviation
CCB	Continuing Calibration Blank
CCV	Continuing Calibration Verification Standard
DUP	Sample Duplicate
ICSA	Interference Check Sample A
ICSAB	Interference Check Sample AB
ICV	Initial Calibration Verification Standard
LCS	Laboratory Control Sample
LCSD	Laboratory Control Sample Duplicate
LFB	Laboratory Fortified Blank
LRB	Laboratory Reagent Blank
MBLK	Method Blank
MS	Sample Matrix Spike
MSD	Sample Matrix Spike Duplicate
PDS	Post Digestion/Distillation Spike
QCS	Quality Control Sample
SD	Serial Dilution
SRM	Standard Reference Material



### LABORATORY ANALYTICAL REPORT

Prepared by Billings, MT Branch

Lab ID: B21121020-001  
Collection Date: 12/08/2021 15:10  
Date Received: 12/11/2021  
Report Date: 01/20/2022

Client: AECOM - Honolulu  
Client Sample ID: ERH2213 (RHMW01R)  
Project: CV18F0126/60571032.02.20.01  
Matrix: Ground Water

Analyses	Result	Units	DF	Qual	LOQ	LOD	DL	MCL	Method	Analysis Date / By	RunID : Run Order	BatchID
<b>AGGREGATE ORGANICS</b>												
Organic Carbon, Total (TOC) - TOC Range is 0.8 to 0.8	0.78	mg/L	1		0.50	0.50	0.17		SW9060A	12/14/2021 23:04/eli-ca	SUB-C277842 : 10	C_R277842
<b>VOLATILE ORGANIC COMPOUNDS</b>												
Benzene	ND	ug/L	1	U	1.0	0.20	0.05		SW8260B	12/14/2021 15:28/sbd	SV5972.I_211214A : 12	R371802
Ethylbenzene	ND	ug/L	1	U	1.0	0.20	0.05		SW8260B	12/14/2021 15:28/sbd	SV5972.I_211214A : 12	R371802
Toluene	ND	ug/L	1	U	1.0	0.20	0.06		SW8260B	12/14/2021 15:28/sbd	SV5972.I_211214A : 12	R371802
m+p-Xylenes	ND	ug/L	1	U	1.0	0.20	0.07		SW8260B	12/14/2021 15:28/sbd	SV5972.I_211214A : 12	R371802
o-Xylene	ND	ug/L	1	U	1.0	0.20	0.04		SW8260B	12/14/2021 15:28/sbd	SV5972.I_211214A : 12	R371802
Xylenes, Total	ND	ug/L	1	U	1.0	0.20	0.04		SW8260B	12/14/2021 15:28/sbd	SV5972.I_211214A : 12	R371802
Surr: 1,2-Dichloroethane-d4	98.0	%REC	1		81-118				SW8260B	12/14/2021 15:28/sbd	SV5972.I_211214A : 12	R371802
Surr: Dibromofluoromethane	100.0	%REC	1		80-119				SW8260B	12/14/2021 15:28/sbd	SV5972.I_211214A : 12	R371802
Surr: p-Bromofluorobenzene	100.0	%REC	1		85-114				SW8260B	12/14/2021 15:28/sbd	SV5972.I_211214A : 12	R371802
Surr: Toluene-d8	93.0	%REC	1		89-112				SW8260B	12/14/2021 15:28/sbd	SV5972.I_211214A : 12	R371802
<b>PETROLEUM HYDROCARBONS-VOLATILE</b>												
C6 to C10	ND	ug/L	1	U	20	8.7	2.3		SW8015C	12/15/2021 05:57/jp	PE 1_211214A : 27	R371801
Total Purgeable Hydrocarbons	36	ug/L	1		20	10	3.6		SW8015C	12/15/2021 05:57/jp	PE 1_211214A : 27	R371801
Surr: Trifluorotoluene	75.0	%REC	1		70-130				SW8015C	12/15/2021 05:57/jp	PE 1_211214A : 27	R371801
- Note 1: C6 to C10 is defined as all hydrocarbons eluting between 2-Methylpentane and 1,2,4-Trimethylbenzene.												
- Note 2: Total Purgeable Hydrocarbons are defined as the total hydrocarbon response regardless of elution time.												
<b>PETROLEUM HYDROCARBONS-SEMI-VOLATILE</b>												
Diesel Range Organics (C10 to C24)	0.29	mg/L	1	J	0.30	0.14	0.037		SW8015C	12/14/2021 23:35/amn	GCFID-HP5-B_211214A : 13	162151
Diesel Range Organics (SGT-C10 to C24)	0.055	mg/L	1	J	0.30	0.11	0.037		SW8015C	12/16/2021 07:16/amn	GCFID-HP5-B_211214B : 14	162151
Oil Range Hydrocarbons (C24 to C40)	0.23	mg/L	1	J	0.30	0.14	0.084		SW8015C	12/14/2021 23:35/amn	GCFID-HP5-B_211214A : 13	162151
Oil Range Hydrocarbons (SGT-C24 to C40)	ND	mg/L	1	U	0.30	0.14	0.084		SW8015C	12/16/2021 07:16/amn	GCFID-HP5-B_211214B : 14	162151
Total Extractable Hydrocarbons	0.59	mg/L	1		0.30	0.14	0.071		SW8015C	12/14/2021 23:35/amn	GCFID-HP5-B_211214A : 13	162151
Total Extractable Hydrocarbons (SGT)	0.068	mg/L	1	J	0.30	0.11	0.031		SW8015C	12/16/2021 07:16/amn	GCFID-HP5-B_211214B : 14	162151
Surr: o-Terphenyl	96.0	%REC	1		56-125				SW8015C	12/14/2021 23:35/amn	GCFID-HP5-B_211214A : 13	162151
Surr: o-Terphenyl (SGT)	89.0	%REC	1		56-125				SW8015C	12/16/2021 07:16/amn	GCFID-HP5-B_211214B : 14	162151
Surr: n-Triacontane	101.0	%REC	1		50-150				SW8015C	12/14/2021 23:35/amn	GCFID-HP5-B_211214A : 13	162151
Surr: n-Triacontane (SGT)	91.0	%REC	1		50-150				SW8015C	12/16/2021 07:16/amn	GCFID-HP5-B_211214B : 14	162151
- Note: Total Extractable Hydrocarbons are defined as the total hydrocarbon responses regardless of elution time.												
<b>SEMI-VOLATILE ORGANIC COMPOUNDS (LOW LEVEL) BY SIM</b>												
1-Methylnaphthalene	ND	ug/L	1	U	0.10	0.10	0.021		SW8270C	12/15/2021 07:28/jph	SV5975.I_211214B : 4	162126
2-Methylnaphthalene	ND	ug/L	1	U	0.10	0.10	0.018		SW8270C	12/15/2021 07:28/jph	SV5975.I_211214B : 4	162126
Naphthalene	ND	ug/L	1	U	0.10	0.10	0.029		SW8270C	12/15/2021 07:28/jph	SV5975.I_211214B : 4	162126
Surr: 2-Fluorobiphenyl	78.0	%REC	1		53-106				SW8270C	12/15/2021 07:28/jph	SV5975.I_211214B : 4	162126
Surr: Nitrobenzene-d5	69.0	%REC	1		55-111				SW8270C	12/15/2021 07:28/jph	SV5975.I_211214B : 4	162126





### LABORATORY ANALYTICAL REPORT

Prepared by Billings, MT Branch

**Client:** AECOM - Honolulu  
**Client Sample ID:** ERH2213 (RHMW01R)  
**Project:** CV18F0126/60571032.02.20.01  
**Matrix:** Ground Water

**Lab ID:** B21121020-001  
**Collection Date:** 12/08/2021 15:10  
**Date Received:** 12/11/2021  
**Report Date:** 01/20/2022

Analyses	Result	Units	DF	Qual	LOQ	LOD	DL	MCL	Method	Analysis Date / By	RunID : Run Order	BatchID
<b>SEMI-VOLATILE ORGANIC COMPOUNDS (LOW LEVEL) BY SIM</b>												
Surr: Terphenyl-d14	87.0	%REC	1		58-132				SW8270C	12/15/2021 07:28/jph	SV5975.I_211214B : 4	162126



### LABORATORY ANALYTICAL REPORT

Prepared by Billings, MT Branch

**Lab ID:** B21121020-002  
**Collection Date:** 12/08/2021 16:35  
**Date Received:** 12/11/2021  
**Report Date:** 01/20/2022

**Client:** AECOM - Honolulu  
**Client Sample ID:** ERH2215 (RHMW02)  
**Project:** CV18F0126/60571032.02.20.01  
**Matrix:** Ground Water

Analyses	Result	Units	DF	Qual	LOQ	LOD	DL	MCL	Method	Analysis Date / By	RunID : Run Order	BatchID
<b>AGGREGATE ORGANICS</b>												
Organic Carbon, Total (TOC) - TOC Range is 3.4 to 3.4	3.4	mg/L	1		0.50	0.50	0.17		SW9060A	12/14/2021 23:45/eli-ca	SUB-C277842 : 11	C_R277842
<b>VOLATILE ORGANIC COMPOUNDS</b>												
Benzene	ND	ug/L	1	U	1.0	0.20	0.05		SW8260B	12/14/2021 15:53/sbd	SV5972.I_211214A : 13	R371802
Ethylbenzene	ND	ug/L	1	U	1.0	0.20	0.05		SW8260B	12/14/2021 15:53/sbd	SV5972.I_211214A : 13	R371802
Toluene	ND	ug/L	1	U	1.0	0.20	0.06		SW8260B	12/14/2021 15:53/sbd	SV5972.I_211214A : 13	R371802
m+p-Xylenes	ND	ug/L	1	U	1.0	0.20	0.07		SW8260B	12/14/2021 15:53/sbd	SV5972.I_211214A : 13	R371802
o-Xylene	ND	ug/L	1	U	1.0	0.20	0.04		SW8260B	12/14/2021 15:53/sbd	SV5972.I_211214A : 13	R371802
Xylenes, Total	ND	ug/L	1	U	1.0	0.20	0.04		SW8260B	12/14/2021 15:53/sbd	SV5972.I_211214A : 13	R371802
Surr: 1,2-Dichloroethane-d4	99.0	%REC	1		81-118				SW8260B	12/14/2021 15:53/sbd	SV5972.I_211214A : 13	R371802
Surr: Dibromofluoromethane	100.0	%REC	1		80-119				SW8260B	12/14/2021 15:53/sbd	SV5972.I_211214A : 13	R371802
Surr: p-Bromofluorobenzene	103.0	%REC	1		85-114				SW8260B	12/14/2021 15:53/sbd	SV5972.I_211214A : 13	R371802
Surr: Toluene-d8	90.0	%REC	1		89-112				SW8260B	12/14/2021 15:53/sbd	SV5972.I_211214A : 13	R371802
<b>PETROLEUM HYDROCARBONS-VOLATILE</b>												
C6 to C10	12	ug/L	1	J	20	8.7	2.3		SW8015C	12/15/2021 07:06/jp	PE 1_211214A : 28	R371801
Total Purgeable Hydrocarbons	237	ug/L	1		20	10	3.6		SW8015C	12/15/2021 07:06/jp	PE 1_211214A : 28	R371801
Surr: Trifluorotoluene	79.0	%REC	1		70-130				SW8015C	12/15/2021 07:06/jp	PE 1_211214A : 28	R371801
- Note 1: C6 to C10 is defined as all hydrocarbons eluting between 2-Methylpentane and 1,2,4-Trimethylbenzene.												
- Note 2: Total Purgeable Hydrocarbons are defined as the total hydrocarbon response regardless of elution time.												
<b>PETROLEUM HYDROCARBONS-SEMI-VOLATILE</b>												
Diesel Range Organics (C10 to C24)	1.7	mg/L	1		0.30	0.15	0.038		SW8015C	12/15/2021 01:02/amn	GCFID-HP5-B_211214A : 14	162151
Diesel Range Organics (SGT-C10 to C24)	0.41	mg/L	1		0.30	0.12	0.038		SW8015C	12/16/2021 07:59/amn	GCFID-HP5-B_211214B : 15	162151
Oil Range Hydrocarbons (C24 to C40)	0.34	mg/L	1		0.30	0.15	0.085		SW8015C	12/15/2021 01:02/amn	GCFID-HP5-B_211214A : 14	162151
Oil Range Hydrocarbons (SGT-C24 to C40)	ND	mg/L	1	U	0.30	0.15	0.085		SW8015C	12/16/2021 07:59/amn	GCFID-HP5-B_211214B : 15	162151
Total Extractable Hydrocarbons	2.1	mg/L	1		0.30	0.15	0.073		SW8015C	12/15/2021 01:02/amn	GCFID-HP5-B_211214A : 14	162151
Total Extractable Hydrocarbons (SGT)	0.43	mg/L	1		0.30	0.12	0.032		SW8015C	12/16/2021 07:59/amn	GCFID-HP5-B_211214B : 15	162151
Surr: o-Terphenyl	87.0	%REC	1		56-125				SW8015C	12/15/2021 01:02/amn	GCFID-HP5-B_211214A : 14	162151
Surr: o-Terphenyl (SGT)	86.0	%REC	1		56-125				SW8015C	12/16/2021 07:59/amn	GCFID-HP5-B_211214B : 15	162151
Surr: n-Triacontane	92.0	%REC	1		50-150				SW8015C	12/15/2021 01:02/amn	GCFID-HP5-B_211214A : 14	162151
Surr: n-Triacontane (SGT)	89.0	%REC	1		50-150				SW8015C	12/16/2021 07:59/amn	GCFID-HP5-B_211214B : 15	162151
- Note: Total Extractable Hydrocarbons are defined as the total hydrocarbon responses regardless of elution time.												
<b>SEMI-VOLATILE ORGANIC COMPOUNDS (LOW LEVEL) BY SIM</b>												
1-Methylnaphthalene	14	ug/L	5		0.51	0.51	0.11		SW8270C	12/15/2021 11:49/jph	SV5975.I_211214B : 12	162189
2-Methylnaphthalene	11	ug/L	5		0.51	0.51	0.090		SW8270C	12/15/2021 11:49/jph	SV5975.I_211214B : 12	162189
Naphthalene	30	ug/L	5		0.51	0.51	0.15		SW8270C	12/15/2021 11:49/jph	SV5975.I_211214B : 12	162189
Surr: 2-Fluorobiphenyl	78.0	%REC	1		53-106				SW8270C	12/15/2021 08:01/jph	SV5975.I_211214B : 5	162126
Surr: Nitrobenzene-d5	88.0	%REC	1		55-111				SW8270C	12/15/2021 08:01/jph	SV5975.I_211214B : 5	162126



### LABORATORY ANALYTICAL REPORT

Prepared by Billings, MT Branch

**Client:** AECOM - Honolulu  
**Client Sample ID:** ERH2215 (RHMW02)  
**Project:** CV18F0126/60571032.02.20.01  
**Matrix:** Ground Water

**Lab ID:** B21121020-002  
**Collection Date:** 12/08/2021 16:35  
**Date Received:** 12/11/2021  
**Report Date:** 01/20/2022

Analyses	Result	Units	DF	Qual	LOQ	LOD	DL	MCL	Method	Analysis Date / By	RunID : Run Order	BatchID
<b>SEMI-VOLATILE ORGANIC COMPOUNDS (LOW LEVEL) BY SIM</b>												
Surr: Terphenyl-d14	107.0	%REC	1		58-132				SW8270C	12/15/2021 08:01/jph	SV5975.I_211214B : 5	162126



**LABORATORY ANALYTICAL REPORT**

Prepared by Billings, MT Branch

**Lab ID:** B21121020-003

**Collection Date:** 12/08/2021 15:05

**Date Received:** 12/11/2021

**Report Date:** 01/20/2022

**Client:** AECOM - Honolulu  
**Client Sample ID:** ERH2212 (Trip Blank)-14525  
**Project:** CV18F0126/60571032.02.20.01  
**Matrix:** Trip Blank

Analyses	Result	Units	DF	Qual	LOQ	LOD	DL	MCL	Method	Analysis Date / By	RunID : Run Order	BatchID
<b>PETROLEUM HYDROCARBONS-VOLATILE</b>												
C6 to C10	ND	ug/L	1	U	20	8.7	2.3		SW8015C	12/14/2021 17:56/jp	PE 1_211214A : 13	R371801
Total Purgeable Hydrocarbons	ND	ug/L	1	U	20	10	3.6		SW8015C	12/14/2021 17:56/jp	PE 1_211214A : 13	R371801
Surr: Trifluorotoluene	84.0	%REC	1		70-130				SW8015C	12/14/2021 17:56/jp	PE 1_211214A : 13	R371801

- Note 1: C6 to C10 is defined as all hydrocarbons eluting between 2-Methylpentane and 1,2,4-Trimethylbenzene.  
- Note 2: Total Purgeable Hydrocarbons are defined as the total hydrocarbon response regardless of elution time.



**LABORATORY ANALYTICAL REPORT**

Prepared by Billings, MT Branch

**Lab ID:** B21121020-004  
**Collection Date:** 12/08/2021 15:05  
**Date Received:** 12/11/2021  
**Report Date:** 01/20/2022

**Client:** AECOM - Honolulu  
**Client Sample ID:** ERH2212 (Trip Blank)-14525  
**Project:** CV18F0126/60571032.02.20.01  
**Matrix:** Trip Blank

Analyses	Result	Units	DF	Qual	LOQ	LOD	DL	MCL	Method	Analysis Date / By	RunID : Run Order	BatchID
<b>VOLATILE ORGANIC COMPOUNDS</b>												
Benzene	ND	ug/L	1	U	1.0	0.20	0.05		SW8260B	12/14/2021 19:15/sbd	SV5972.I_211214A : 20	R371802
Ethylbenzene	ND	ug/L	1	U	1.0	0.20	0.05		SW8260B	12/14/2021 19:15/sbd	SV5972.I_211214A : 20	R371802
Toluene	ND	ug/L	1	U	1.0	0.20	0.06		SW8260B	12/14/2021 19:15/sbd	SV5972.I_211214A : 20	R371802
m+p-Xylenes	ND	ug/L	1	U	1.0	0.20	0.07		SW8260B	12/14/2021 19:15/sbd	SV5972.I_211214A : 20	R371802
o-Xylene	ND	ug/L	1	U	1.0	0.20	0.04		SW8260B	12/14/2021 19:15/sbd	SV5972.I_211214A : 20	R371802
Xylenes, Total	ND	ug/L	1	U	1.0	0.20	0.04		SW8260B	12/14/2021 19:15/sbd	SV5972.I_211214A : 20	R371802
Surr: 1,2-Dichloroethane-d4	98.0	%REC	1		81-118				SW8260B	12/14/2021 19:15/sbd	SV5972.I_211214A : 20	R371802
Surr: Dibromofluoromethane	101.0	%REC	1		80-119				SW8260B	12/14/2021 19:15/sbd	SV5972.I_211214A : 20	R371802
Surr: p-Bromofluorobenzene	102.0	%REC	1		85-114				SW8260B	12/14/2021 19:15/sbd	SV5972.I_211214A : 20	R371802
Surr: Toluene-d8	93.0	%REC	1		89-112				SW8260B	12/14/2021 19:15/sbd	SV5972.I_211214A : 20	R371802



**LABORATORY ANALYTICAL REPORT**

Prepared by Billings, MT Branch

**Lab ID:** B21121020-005

**Collection Date:** 12/08/2021 16:35

**Date Received:** 12/11/2021

**Report Date:** 01/20/2022

**Client:** AECOM - Honolulu  
**Client Sample ID:** ERH2214 (Trip Blank)-14525  
**Project:** CV18F0126/60571032.02.20.01  
**Matrix:** Trip Blank

Analyses	Result	Units	DF	Qual	LOQ	LOD	DL	MCL	Method	Analysis Date / By	RunID : Run Order	BatchID
<b>PETROLEUM HYDROCARBONS-VOLATILE</b>												
C6 to C10	ND	ug/L	1	U	20	8.7	2.3		SW8015C	12/14/2021 18:31/jp	PE 1_211214A : 14	R371801
Total Purgeable Hydrocarbons	ND	ug/L	1	U	20	10	3.6		SW8015C	12/14/2021 18:31/jp	PE 1_211214A : 14	R371801
Surr: Trifluorotoluene	85.0	%REC	1		70-130				SW8015C	12/14/2021 18:31/jp	PE 1_211214A : 14	R371801
- Note 1: C6 to C10 is defined as all hydrocarbons eluting between 2-Methylpentane and 1,2,4-Trimethylbenzene.												
- Note 2: Total Purgeable Hydrocarbons are defined as the total hydrocarbon response regardless of elution time.												



### LABORATORY ANALYTICAL REPORT

Prepared by Billings, MT Branch

**Client:** AECOM - Honolulu  
**Client Sample ID:** ERH2214 (Trip Blank)-14525  
**Project:** CV18F0126/60571032.02.20.01  
**Matrix:** Trip Blank

**Lab ID:** B21121020-006  
**Collection Date:** 12/08/2021 16:35  
**Date Received:** 12/11/2021  
**Report Date:** 01/20/2022

Analyses	Result	Units	DF	Qual	LOQ	LOD	DL	MCL	Method	Analysis Date / By	RunID : Run Order	BatchID
<b>VOLATILE ORGANIC COMPOUNDS</b>												
Benzene	ND	ug/L	1	U	1.0	0.20	0.05		SW8260B	12/14/2021 19:40/sbd	SV5972.I_211214A : 21	R371802
Ethylbenzene	ND	ug/L	1	U	1.0	0.20	0.05		SW8260B	12/14/2021 19:40/sbd	SV5972.I_211214A : 21	R371802
Toluene	ND	ug/L	1	U	1.0	0.20	0.06		SW8260B	12/14/2021 19:40/sbd	SV5972.I_211214A : 21	R371802
m+p-Xylenes	ND	ug/L	1	U	1.0	0.20	0.07		SW8260B	12/14/2021 19:40/sbd	SV5972.I_211214A : 21	R371802
o-Xylene	ND	ug/L	1	U	1.0	0.20	0.04		SW8260B	12/14/2021 19:40/sbd	SV5972.I_211214A : 21	R371802
Xylenes, Total	ND	ug/L	1	U	1.0	0.20	0.04		SW8260B	12/14/2021 19:40/sbd	SV5972.I_211214A : 21	R371802
Surr: 1,2-Dichloroethane-d4	96.0	%REC	1			81-118			SW8260B	12/14/2021 19:40/sbd	SV5972.I_211214A : 21	R371802
Surr: Dibromofluoromethane	101.0	%REC	1			80-119			SW8260B	12/14/2021 19:40/sbd	SV5972.I_211214A : 21	R371802
Surr: p-Bromofluorobenzene	100.0	%REC	1			85-114			SW8260B	12/14/2021 19:40/sbd	SV5972.I_211214A : 21	R371802
Surr: Toluene-d8	94.0	%REC	1			89-112			SW8260B	12/14/2021 19:40/sbd	SV5972.I_211214A : 21	R371802



### Analytical QC Summary Report

Prepared by Billings, MT Branch

**Client:** AECOM - Honolulu  
**Workorder:** B21121020  
**Project:** CV18F0126/60571032.02.20.01

**Report Date:** 01/20/2022

**Run ID: Run Order:** SUB-C277842: 2      **SampType:** Method Blank      **Batch ID:** C\_R277842  
**Method:** SW9060A      **Analysis Date:** 12/14/2021 17:32      **Prep Date:**  
**Lab ID:** MBLK      **Units:** mg/L      **Prep Method:**

Analytes	Result	LOQ	Spk value	Spk RefVal	%REC	LowLimit	HighLimit	RPD RefVal	%RPD	RPDLimit	Qual
Organic Carbon, Total (TOC)	ND	0.25									

Associated Samples: **B21121020-001E, B21121020-002E**  
- TOC Range is 0.1 to 0.2

**Run ID: Run Order:** SUB-C277842: 1      **SampType:** Laboratory Control Sample      **Batch ID:** C\_R277842  
**Method:** SW9060A      **Analysis Date:** 12/14/2021 16:11      **Prep Date:**  
**Lab ID:** LCS      **Units:** mg/L      **Prep Method:**

Analytes	Result	LOQ	Spk value	Spk RefVal	%REC	LowLimit	HighLimit	RPD RefVal	%RPD	RPDLimit	Qual
Organic Carbon, Total (TOC)	5.0	0.50	5.0		101.0	91	111				

Associated Samples: **B21121020-001E, B21121020-002E**  
- TOC Range is 5.0 to 5.1

**Run ID: Run Order:** SUB-C277842: 3      **SampType:** Continuing Calibration Verification Standard      **Batch ID:** C\_R277842  
**Method:** SW9060A      **Analysis Date:** 12/14/2021 18:11      **Prep Date:**  
**Lab ID:** CCV      **Units:** mg/L      **Prep Method:**

Analytes	Result	LOQ	Spk value	Spk RefVal	%REC	LowLimit	HighLimit	RPD RefVal	%RPD	RPDLimit	Qual
Organic Carbon, Total (TOC)	5.0	0.50	5.0		100.0	90	110				

Associated Samples: **B21121020-001E, B21121020-002E**  
- TOC Range is 5.0 to 5.0





### Analytical QC Summary Report

Prepared by Billings, MT Branch

**Client:** AECOM - Honolulu  
**Workorder:** B21121020  
**Project:** CV18F0126/60571032.02.20.01

**Report Date:** 01/20/2022

**Run ID: Run Order:** SV5972.I\_211214A: 4  
**Method:** SW8260B  
**Lab ID:** MBLK121421

**SampType:** Method Blank  
**Analysis Date:** 12/14/2021 12:04  
**Units:** ug/L

**Batch ID:** R371802  
**Prep Date:**  
**Prep Method:**

Analytes	Result	LOQ	Spk value	Spk RefVal	%REC	LowLimit	HighLimit	RPD RefVal	%RPD	RPDLimit	Qual
Benzene	ND	0.50									
Ethylbenzene	ND	0.50									
Toluene	ND	0.50									
m+p-Xylenes	ND	0.50									
o-Xylene	ND	0.50									
Xylenes, Total	ND	0.50									
Surr: 1,2-Dichloroethane-d4	9.7	0.50	10		97.0	81	118				
Surr: Dibromofluoromethane	9.9	0.50	10		99.0	80	119				
Surr: p-Bromofluorobenzene	9.9	0.50	10		99.0	85	114				
Surr: Toluene-d8	9.2	0.50	10		92.0	89	112				

Associated Samples: B21121020-001D, B21121020-002D, B21121020-004A, B21121020-006A



### Analytical QC Summary Report

Prepared by Billings, MT Branch

**Client:** AECOM - Honolulu  
**Workorder:** B21121020  
**Project:** CV18F0126/60571032.02.20.01

**Report Date:** 01/20/2022

**Run ID: Run Order:** SV5972.I\_211214A: 3      **SampType:** Laboratory Control Sample      **Batch ID:** R371802  
**Method:** SW8260B      **Analysis Date:** 12/14/2021 11:14      **Prep Date:**  
**Lab ID:** LCS121421      **Units:** ug/L      **Prep Method:**

Analytes	Result	LOQ	Spk value	Spk RefVal	%REC	LowLimit	HighLimit	RPD RefVal	%RPD	RPDLimit	Qual
Benzene	4.6	0.50	5.0		93.0	79	120				
Ethylbenzene	4.5	0.50	5.0		90.0	79	121				
Toluene	4.5	0.50	5.0		90.0	80	121				
m+p-Xylenes	8.7	0.50	10		87.0	80	121				
o-Xylene	4.5	0.50	5.0		89.0	78	122				
Xylenes, Total	13	0.50	15		88.0	79	121				
Surr: 1,2-Dichloroethane-d4	9.7	0.50	10		97.0	81	118				
Surr: Dibromofluoromethane	9.9	0.50	10		99.0	80	119				
Surr: p-Bromofluorobenzene	10	0.50	10		101.0	85	114				
Surr: Toluene-d8	9.1	0.50	10		91.0	89	112				

Associated Samples: **B21121020-001D, B21121020-002D, B21121020-004A, B21121020-006A**

**Run ID: Run Order:** SV5972.I\_211214A: 23      **SampType:** Sample Matrix Spike      **Batch ID:** R371802  
**Method:** SW8260B      **Analysis Date:** 12/14/2021 20:05      **Prep Date:**  
**Lab ID:** B21121012-001DMS      **Units:** ug/L      **Prep Method:**

Analytes	Result	LOQ	Spk value	Spk RefVal	%REC	LowLimit	HighLimit	RPD RefVal	%RPD	RPDLimit	Qual
Benzene	4.6	0.50	5.0	0.0	92.0	79	120				
Ethylbenzene	4.5	0.50	5.0	0.0	91.0	79	121				
Toluene	4.6	0.50	5.0	0.0	92.0	80	121				
m+p-Xylenes	8.7	0.50	10	0.0	87.0	80	121				
o-Xylene	4.5	0.50	5.0	0.0	89.0	78	122				
Xylenes, Total	13	0.50	15	0.0	88.0	79	121				
Surr: 1,2-Dichloroethane-d4	9.7	0.50	10	0.0	97.0	81	118				
Surr: Dibromofluoromethane	10	0.50	10	0.0	100.0	80	119				
Surr: p-Bromofluorobenzene	9.9	0.50	10	0.0	99.0	85	114				



### Analytical QC Summary Report

Prepared by Billings, MT Branch

**Client:** AECOM - Honolulu  
**Workorder:** B21121020  
**Project:** CV18F0126/60571032.02.20.01

**Report Date:** 01/20/2022

**Run ID: Run Order:** SV5972.I\_211214A: 23      **SampType:** Sample Matrix Spike      **Batch ID:** R371802  
**Method:** SW8260B      **Analysis Date:** 12/14/2021 20:05      **Prep Date:**  
**Lab ID:** B21121012-001DMS      **Units:** ug/L      **Prep Method:**

Analytes	Result	LOQ	Spk value	Spk RefVal	%REC	LowLimit	HighLimit	RPD RefVal	%RPD	RPDLimit	Qual
Surr: Toluene-d8	9.3	0.50	10	0.0	93.0	89	112				

Associated Samples: B21121020-001D, B21121020-002D, B21121020-004A, B21121020-006A

**Run ID: Run Order:** SV5972.I\_211214A: 24      **SampType:** Sample Matrix Spike Duplicate      **Batch ID:** R371802  
**Method:** SW8260B      **Analysis Date:** 12/14/2021 20:31      **Prep Date:**  
**Lab ID:** B21121012-001DMSD      **Units:** ug/L      **Prep Method:**

Analytes	Result	LOQ	Spk value	Spk RefVal	%REC	LowLimit	HighLimit	RPD RefVal	%RPD	RPDLimit	Qual
Benzene	4.3	0.50	5.0	0.0	86.0	79	120	4.6	6.9	20.0	
Ethylbenzene	4.2	0.50	5.0	0.0	84.0	79	121	4.5	7.3	20.0	
Toluene	4.3	0.50	5.0	0.0	86.0	80	121	4.6	6.6	20.0	
m+p-Xylenes	8.1	0.50	10	0.0	81.0	80	121	8.7	6.7	20.0	
o-Xylene	4.2	0.50	5.0	0.0	83.0	78	122	4.5	6.9	20.0	
Xylenes, Total	12	0.50	15	0.0	82.0	79	121	13	6.8	20.0	
Surr: 1,2-Dichloroethane-d4	9.8	0.50	10	0.0	98.0	81	118	0.0			
Surr: Dibromofluoromethane	9.9	0.50	10	0.0	99.0	80	119	0.0			
Surr: p-Bromofluorobenzene	9.7	0.50	10	0.0	97.0	85	114	0.0			
Surr: Toluene-d8	9.5	0.50	10	0.0	95.0	89	112	0.0			

Associated Samples: B21121020-001D, B21121020-002D, B21121020-004A, B21121020-006A



### Analytical QC Summary Report

Prepared by Billings, MT Branch

**Client:** AECOM - Honolulu  
**Workorder:** B21121020  
**Project:** CV18F0126/60571032.02.20.01

**Report Date:** 01/20/2022

**Run ID: Run Order:** SV5972.I\_211214A: 2      **SampType:** Continuing Calibration Verification Standard      **Batch ID:** R371802  
**Method:** SW8260B      **Analysis Date:** 12/14/2021 10:49      **Prep Date:**  
**Lab ID:** CCV121421      **Units:** ug/L      **Prep Method:**

Analytes	Result	LOQ	Spk value	Spk RefVal	%REC	LowLimit	HighLimit	RPD RefVal	%RPD	RPDLimit	Qual
Benzene	4.9	0.50	5.0		98.0	80	120				
Ethylbenzene	4.6	0.50	5.0		92.0	80	120				
Toluene	4.6	0.50	5.0		92.0	80	120				
m+p-Xylenes	9.1	0.50	10		91.0	80	120				
o-Xylene	4.5	0.50	5.0		89.0	80	120				
Xylenes, Total	14	0.50	15		90.0	80	120				
Surr: 1,2-Dichloroethane-d4	9.4	0.50	10		94.0	80	120				
Surr: Dibromofluoromethane	9.9	0.50	10		99.0	80	120				
Surr: p-Bromofluorobenzene	9.8	0.50	10		98.0	80	120				
Surr: Toluene-d8	9.2	0.50	10		92.0	80	120				

Associated Samples: **B21121020-001D, B21121020-002D, B21121020-004A, B21121020-006A**



### Analytical QC Summary Report

Prepared by Billings, MT Branch

**Client:** AECOM - Honolulu  
**Workorder:** B21121020  
**Project:** CV18F0126/60571032.02.20.01

**Report Date:** 01/20/2022

**Run ID: Run Order:** SV5972.I\_211214A: 25      **SampType:** Continuing Calibration Verification Standard      **Batch ID:** R371802  
**Method:** SW8260B      **Analysis Date:** 12/14/2021 20:56      **Prep Date:**  
**Lab ID:** CCV\_CLOSING\_121421      **Units:** ug/L      **Prep Method:**

Analytes	Result	LOQ	Spk value	Spk RefVal	%REC	LowLimit	HighLimit	RPD RefVal	%RPD	RPDLimit	Qual
Benzene	5.1	0.50	5.0		102.0	50	150				
Ethylbenzene	4.9	0.50	5.0		99.0	50	150				
Toluene	5.1	0.50	5.0		101.0	50	150				
m+p-Xylenes	9.7	0.50	10		97.0	50	150				
o-Xylene	4.9	0.50	5.0		97.0	50	150				
Xylenes, Total	15	0.50	15		97.0	50	150				
Surr: 1,2-Dichloroethane-d4	9.6	0.50	10		96.0	50	150				
Surr: Dibromofluoromethane	9.8	0.50	10		98.0	50	150				
Surr: p-Bromofluorobenzene	9.8	0.50	10		98.0	50	150				
Surr: Toluene-d8	9.5	0.50	10		95.0	50	150				

Associated Samples: **B21121020-001D, B21121020-002D, B21121020-004A, B21121020-006A**



### Analytical QC Summary Report

Prepared by Billings, MT Branch

**Client:** AECOM - Honolulu  
**Workorder:** B21121020  
**Project:** CV18F0126/60571032.02.20.01

**Report Date:** 01/20/2022

**Run ID: Run Order:** PE 1\_211214A: 4      **SampType:** Method Blank      **Batch ID:** R371801  
**Method:** SW8015C      **Analysis Date:** 12/14/2021 12:14      **Prep Date:**  
**Lab ID:** MBLK\_1214PE105r      **Units:** ug/L      **Prep Method:**

Analytes	Result	LOQ	Spk value	Spk RefVal	%REC	LowLimit	HighLimit	RPD RefVal	%RPD	RPDLimit	Qual
C6 to C10	ND	10									
Total Purgeable Hydrocarbons	ND	10									
Surr: Trifluorotoluene	21	1.0	25		85.0	70	130				

Associated Samples: B21121020-001C, B21121020-002C, B21121020-003A, B21121020-005A

**Run ID: Run Order:** PE 1\_211214A: 20      **SampType:** Method Blank      **Batch ID:** R371801  
**Method:** SW8015C      **Analysis Date:** 12/14/2021 22:31      **Prep Date:**  
**Lab ID:** MBLK\_1214PE123r      **Units:** ug/L      **Prep Method:**

Analytes	Result	LOQ	Spk value	Spk RefVal	%REC	LowLimit	HighLimit	RPD RefVal	%RPD	RPDLimit	Qual
C6 to C10	ND	10									
Total Purgeable Hydrocarbons	ND	10									
Surr: Trifluorotoluene	20	1.0	25		81.0	70	130				

Associated Samples: B21121020-001C, B21121020-002C, B21121020-003A, B21121020-005A

**Run ID: Run Order:** PE 1\_211214A: 3      **SampType:** Laboratory Control Sample      **Batch ID:** R371801  
**Method:** SW8015C      **Analysis Date:** 12/14/2021 11:39      **Prep Date:**  
**Lab ID:** LCS\_1214PE104r      **Units:** ug/L      **Prep Method:**

Analytes	Result	LOQ	Spk value	Spk RefVal	%REC	LowLimit	HighLimit	RPD RefVal	%RPD	RPDLimit	Qual
C6 to C10	178	20	170		105.0	78	122				
Total Purgeable Hydrocarbons	216	20	200		108.0	70	130				
Surr: Trifluorotoluene	24	1.0	25		95.0	70	130				

Associated Samples: B21121020-001C, B21121020-002C, B21121020-003A, B21121020-005A



### Analytical QC Summary Report

Prepared by Billings, MT Branch

**Client:** AECOM - Honolulu  
**Workorder:** B21121020  
**Project:** CV18F0126/60571032.02.20.01

**Report Date:** 01/20/2022

**Run ID: Run Order:** PE 1\_211214A: 19      **SampType:** Laboratory Control Sample      **Batch ID:** R371801  
**Method:** SW8015C      **Analysis Date:** 12/14/2021 21:57      **Prep Date:**  
**Lab ID:** LCS\_1214PE122r      **Units:** ug/L      **Prep Method:**

Analytes	Result	LOQ	Spk value	Spk RefVal	%REC	LowLimit	HighLimit	RPD RefVal	%RPD	RPDLimit	Qual
C6 to C10	171	20	170		100.0	78	122				
Total Purgeable Hydrocarbons	207	20	200		103.0	70	130				
Surr: Trifluorotoluene	23	1.0	25		92.0	70	130				

Associated Samples: B21121020-001C, B21121020-002C, B21121020-003A, B21121020-005A

**Run ID: Run Order:** PE 1\_211214A: 15      **SampType:** Sample Matrix Spike      **Batch ID:** R371801  
**Method:** SW8015C      **Analysis Date:** 12/14/2021 19:05      **Prep Date:**  
**Lab ID:** B21121001-001CMS      **Units:** ug/L      **Prep Method:**

Analytes	Result	LOQ	Spk value	Spk RefVal	%REC	LowLimit	HighLimit	RPD RefVal	%RPD	RPDLimit	Qual
C6 to C10	162	20	170	0.0	95.0	78	122				
Total Purgeable Hydrocarbons	198	20	200	0.0	99.0	70	130				
Surr: Trifluorotoluene	23	1.0	25	0.0	92.0	70	130				

Associated Samples: B21121020-001C, B21121020-002C, B21121020-003A, B21121020-005A

**Run ID: Run Order:** PE 1\_211214A: 16      **SampType:** Sample Matrix Spike Duplicate      **Batch ID:** R371801  
**Method:** SW8015C      **Analysis Date:** 12/14/2021 19:39      **Prep Date:**  
**Lab ID:** B21121001-001CMSD      **Units:** ug/L      **Prep Method:**

Analytes	Result	LOQ	Spk value	Spk RefVal	%REC	LowLimit	HighLimit	RPD RefVal	%RPD	RPDLimit	Qual
C6 to C10	171	20	170	0.0	101.0	78	122	162	5.6	20.0	
Total Purgeable Hydrocarbons	210	20	200	0.0	105.0	70	130	198	5.8	20.0	
Surr: Trifluorotoluene	24	1.0	25	0.0	94.0	70	130	0.0			

Associated Samples: B21121020-001C, B21121020-002C, B21121020-003A, B21121020-005A



### Analytical QC Summary Report

Prepared by Billings, MT Branch

**Client:** AECOM - Honolulu  
**Workorder:** B21121020  
**Project:** CV18F0126/60571032.02.20.01

**Report Date:** 01/20/2022

**Run ID: Run Order:** GCFID-HP5-B\_211214A: 5      **SampType:** Method Blank      **Batch ID:** 162151  
**Method:** SW8015C      **Analysis Date:** 12/14/2021 13:25      **Prep Date:** 12/13/2021 13:18  
**Lab ID:** MB-162151      **Units:** mg/L      **Prep Method:** SW3520C

Analytes	Result	LOQ	Spk value	Spk RefVal	%REC	LowLimit	HighLimit	RPD RefVal	%RPD	RPDLimit	Qual
Diesel Range Organics (C10 to C24)	ND	0.15									
Oil Range Hydrocarbons (C24 to C40)	ND	0.15									
Total Extractable Hydrocarbons	ND	0.15									
Surr: o-Terphenyl	0.19	0.0020	0.20		93.0	56	150				
Surr: n-Triacontane	0.098	0.0020	0.10		98.0	50	125				

Associated Samples: **B21121020-001B, B21121020-002B**

**Run ID: Run Order:** GCFID-HP5-B\_211214B: 5      **SampType:** Method Blank      **Batch ID:** 162151  
**Method:** SW8015C      **Analysis Date:** 12/15/2021 21:57      **Prep Date:** 12/13/2021 13:18  
**Lab ID:** MB-162151      **Units:** mg/L      **Prep Method:** SW3520C

Analytes	Result	LOQ	Spk value	Spk RefVal	%REC	LowLimit	HighLimit	RPD RefVal	%RPD	RPDLimit	Qual
Diesel Range Organics (SGT-C10 to C24)	ND	0.15									
Oil Range Hydrocarbons (SGT-C24 to C40)	ND	0.15									
Total Extractable Hydrocarbons (SGT)	ND	0.15									
Surr: o-Terphenyl (SGT)	0.17	0.0020	0.20		84.0	56	125				
Surr: n-Triacontane (SGT)	0.082	0.0020	0.10		82.0	50	150				

Associated Samples: **B21121020-001B, B21121020-002B**





### Analytical QC Summary Report

Prepared by Billings, MT Branch

**Client:** AECOM - Honolulu  
**Workorder:** B21121020  
**Project:** CV18F0126/60571032.02.20.01

**Report Date:** 01/20/2022

**Run ID: Run Order:** GCFID-HP5-B\_211214A: 3      **SampType:** Laboratory Control Sample      **Batch ID:** 162151  
**Method:** SW8015C      **Analysis Date:** 12/14/2021 11:59      **Prep Date:** 12/13/2021 13:18  
**Lab ID:** LCS-162151      **Units:** mg/L      **Prep Method:** SW3520C

Analytes	Result	LOQ	Spk value	Spk RefVal	%REC	LowLimit	HighLimit	RPD RefVal	%RPD	RPDLimit	Qual
Diesel Range Organics (C10 to C24)	13	0.30	15		87.0	36	132				
Total Extractable Hydrocarbons	14	0.30	15		93.0	60	132				
Surr: o-Terphenyl	0.18	0.0020	0.20		92.0	56	125				

Associated Samples: **B21121020-001B, B21121020-002B**

**Run ID: Run Order:** GCFID-HP5-B\_211214A: 4      **SampType:** Laboratory Control Sample Duplicate      **Batch ID:** 162151  
**Method:** SW8015C      **Analysis Date:** 12/14/2021 12:42      **Prep Date:** 12/13/2021 13:18  
**Lab ID:** LCSD-162151      **Units:** mg/L      **Prep Method:** SW3520C

Analytes	Result	LOQ	Spk value	Spk RefVal	%REC	LowLimit	HighLimit	RPD RefVal	%RPD	RPDLimit	Qual
Diesel Range Organics (C10 to C24)	14	0.30	15		91.0	36	132	13	3.9	20.0	
Total Extractable Hydrocarbons	15	0.30	15		97.0	60	132	14	3.8	20.0	
Surr: o-Terphenyl	0.19	0.0020	0.20		96.0	56	125	0.0			

Associated Samples: **B21121020-001B, B21121020-002B**

**Run ID: Run Order:** GCFID-HP5-B\_211214A: 22      **SampType:** Laboratory Control Sample      **Batch ID:** 162151  
**Method:** SW8015C      **Analysis Date:** 12/15/2021 12:32      **Prep Date:** 12/13/2021 13:18  
**Lab ID:** LCS-162151-RRO      **Units:** mg/L      **Prep Method:** SW3520C

Analytes	Result	LOQ	Spk value	Spk RefVal	%REC	LowLimit	HighLimit	RPD RefVal	%RPD	RPDLimit	Qual
TEH(Oil Range)	4.7	0.30	5.0		94.0	41	113				
Surr: n-Triacontane	0.095	0.0020	0.10		95.0	50	150				

Associated Samples: **B21121020-001B, B21121020-002B**



### Analytical QC Summary Report

Prepared by Billings, MT Branch

**Client:** AECOM - Honolulu  
**Workorder:** B21121020  
**Project:** CV18F0126/60571032.02.20.01

**Report Date:** 01/20/2022

**Run ID: Run Order:** GCFID-HP5-B\_211214A: 23      **SampType:** Laboratory Control Sample Duplicate      **Batch ID:** 162151  
**Method:** SW8015C      **Analysis Date:** 12/15/2021 13:15      **Prep Date:** 12/13/2021 13:18  
**Lab ID:** LCSD-162151-RRO      **Units:** mg/L      **Prep Method:** SW3520C

Analytes	Result	LOQ	Spk value	Spk RefVal	%REC	LowLimit	HighLimit	RPD RefVal	%RPD	RPDLimit	Qual
TEH(Oil Range)	5.2	0.30	5.0		104.0	41	113	4.7	10.0	20.0	
Surr: n-Triacontane	0.098	0.0020	0.10		98.0	50	150	0.0			

Associated Samples: **B21121020-001B, B21121020-002B**

**Run ID: Run Order:** GCFID-HP5-B\_211214B: 3      **SampType:** Laboratory Control Sample      **Batch ID:** 162151  
**Method:** SW8015C      **Analysis Date:** 12/15/2021 20:31      **Prep Date:** 12/13/2021 13:18  
**Lab ID:** LCS-162151      **Units:** mg/L      **Prep Method:** SW3520C

Analytes	Result	LOQ	Spk value	Spk RefVal	%REC	LowLimit	HighLimit	RPD RefVal	%RPD	RPDLimit	Qual
Diesel Range Organics (SGT-C10 to C24)	12	0.30	15		79.0	36	132				
Total Extractable Hydrocarbons (SGT)	13	0.30	15		85.0	60	132				
Surr: o-Terphenyl (SGT)	0.17	0.0020	0.20		86.0	56	125				

Associated Samples: **B21121020-001B, B21121020-002B**

**Run ID: Run Order:** GCFID-HP5-B\_211214B: 4      **SampType:** Laboratory Control Sample Duplicate      **Batch ID:** 162151  
**Method:** SW8015C      **Analysis Date:** 12/15/2021 21:14      **Prep Date:** 12/13/2021 13:18  
**Lab ID:** LCSD-162151      **Units:** mg/L      **Prep Method:** SW3520C

Analytes	Result	LOQ	Spk value	Spk RefVal	%REC	LowLimit	HighLimit	RPD RefVal	%RPD	RPDLimit	Qual
Diesel Range Organics (SGT-C10 to C24)	13	0.30	15		88.0	36	132	12	10.0	20.0	
Total Extractable Hydrocarbons (SGT)	14	0.30	15		93.0	60	132	13	9.9	20.0	
Surr: o-Terphenyl (SGT)	0.19	0.0020	0.20		94.0	56	125	0.0			

Associated Samples: **B21121020-001B, B21121020-002B**



### Analytical QC Summary Report

Prepared by Billings, MT Branch

**Client:** AECOM - Honolulu  
**Workorder:** B21121020  
**Project:** CV18F0126/60571032.02.20.01

**Report Date:** 01/20/2022

**Run ID: Run Order:** GCFID-HP5-B\_211214B: 18      **SampType:** Laboratory Control Sample      **Batch ID:** 162151  
**Method:** SW8015C      **Analysis Date:** 12/16/2021 10:52      **Prep Date:** 12/13/2021 13:18  
**Lab ID:** LCS-162151-RRO      **Units:** mg/L      **Prep Method:** SW3520C

Analytes	Result	LOQ	Spk value	Spk RefVal	%REC	LowLimit	HighLimit	RPD RefVal	%RPD	RPDLimit	Qual
TEH (SGT-Oil Range)	4.7	0.30	5.0		94.0	41	113				
Surr: n-Triacontane (SGT)	0.091	0.0020	0.10		91.0	50	150				

Associated Samples: **B21121020-001B, B21121020-002B**

**Run ID: Run Order:** GCFID-HP5-B\_211214B: 19      **SampType:** Laboratory Control Sample Duplicate      **Batch ID:** 162151  
**Method:** SW8015C      **Analysis Date:** 12/16/2021 12:17      **Prep Date:** 12/13/2021 13:18  
**Lab ID:** LCSD-162151-RRO      **Units:** mg/L      **Prep Method:** SW3520C

Analytes	Result	LOQ	Spk value	Spk RefVal	%REC	LowLimit	HighLimit	RPD RefVal	%RPD	RPDLimit	Qual
TEH (SGT-Oil Range)	4.7	0.30	5.0		94.0	41	113	4.7	0.9	20.0	
Surr: n-Triacontane (SGT)	0.089	0.0020	0.10		89.0	50	150	0.0			

Associated Samples: **B21121020-001B, B21121020-002B**

**Run ID: Run Order:** GCFID-HP5-B\_211214A: 6      **SampType:** Sample Matrix Spike      **Batch ID:** 162151  
**Method:** SW8015C      **Analysis Date:** 12/14/2021 14:51      **Prep Date:** 12/13/2021 13:19  
**Lab ID:** B21121019-001BMS      **Units:** mg/L      **Prep Method:** SW3520C

Analytes	Result	LOQ	Spk value	Spk RefVal	%REC	LowLimit	HighLimit	RPD RefVal	%RPD	RPDLimit	Qual
Diesel Range Organics (C10 to C24)	14	0.30	14	0.12	96.0	36	132				
Total Extractable Hydrocarbons	15	0.30	14	0.41	102.0	60	132				
Surr: o-Terphenyl	0.18	0.0020	0.19	0.0	96.0	56	125				

Associated Samples: **B21121020-001B, B21121020-002B**



### Analytical QC Summary Report

Prepared by Billings, MT Branch

**Client:** AECOM - Honolulu  
**Workorder:** B21121020  
**Project:** CV18F0126/60571032.02.20.01

**Report Date:** 01/20/2022

**Run ID: Run Order:** GCFID-HP5-B\_211214A: 9      **SampType:** Sample Matrix Spike      **Batch ID:** 162151  
**Method:** SW8015C      **Analysis Date:** 12/14/2021 18:32      **Prep Date:** 12/13/2021 13:19  
**Lab ID:** B21121019-002BMS-RRO      **Units:** mg/L      **Prep Method:** SW3520C

Analytes	Result	LOQ	Spk value	Spk RefVal	%REC	LowLimit	HighLimit	RPD RefVal	%RPD	RPDLimit	Qual
TEH(Oil Range)	5.0	0.30	4.8	0.23	99.0	41	113				
Surr: n-Triacontane	0.094	0.0020	0.096	0.0	98.0	50	150				

Associated Samples: **B21121020-001B, B21121020-002B**

**Run ID: Run Order:** GCFID-HP5-B\_211214B: 7      **SampType:** Sample Matrix Spike      **Batch ID:** 162151  
**Method:** SW8015C      **Analysis Date:** 12/15/2021 23:23      **Prep Date:** 12/13/2021 13:19  
**Lab ID:** B21121019-001BMS      **Units:** mg/L      **Prep Method:** SW3520C

Analytes	Result	LOQ	Spk value	Spk RefVal	%REC	LowLimit	HighLimit	RPD RefVal	%RPD	RPDLimit	Qual
Diesel Range Organics (SGT-C10 to C24)	13	0.30	14	0.0	89.0	36	132				
Total Extractable Hydrocarbons (SGT)	13	0.30	14	0.0	94.0	60	132				
Surr: o-Terphenyl (SGT)	0.18	0.0020	0.19	0.0	93.0	56	125				

Associated Samples: **B21121020-001B, B21121020-002B**

**Run ID: Run Order:** GCFID-HP5-B\_211214B: 11      **SampType:** Sample Matrix Spike      **Batch ID:** 162151  
**Method:** SW8015C      **Analysis Date:** 12/16/2021 03:41      **Prep Date:** 12/13/2021 13:19  
**Lab ID:** B21121019-002BMS-RRO      **Units:** mg/L      **Prep Method:** SW3520C

Analytes	Result	LOQ	Spk value	Spk RefVal	%REC	LowLimit	HighLimit	RPD RefVal	%RPD	RPDLimit	Qual
TEH (SGT-Oil Range)	4.6	0.30	4.8	0.0	95.0	41	113				
Surr: n-Triacontane (SGT)	0.083	0.0020	0.096	0.0	86.0	50	150				

Associated Samples: **B21121020-001B, B21121020-002B**



### Analytical QC Summary Report

Prepared by Billings, MT Branch

**Client:** AECOM - Honolulu  
**Workorder:** B21121020  
**Project:** CV18F0126/60571032.02.20.01

**Report Date:** 01/20/2022

**Run ID: Run Order:** PE 1\_211214A: 2      **SampType:** Continuing Calibration Verification Standard      **Batch ID:** R371801  
**Method:** SW8015C      **Analysis Date:** 12/14/2021 11:05      **Prep Date:**  
**Lab ID:** CCV\_1214PE103r      **Units:** ug/L      **Prep Method:**

Analytes	Result	LOQ	Spk value	Spk RefVal	%REC	LowLimit	HighLimit	RPD RefVal	%RPD	RPDLimit	Qual
C6 to C10	174	20	168		104.0	80	120				
Total Purgeable Hydrocarbons	208	20	200		104.0	80	120				
Surr: Trifluorotoluene	25	1.0	25		99.0	80	120				

Associated Samples: B21121020-001C, B21121020-002C, B21121020-003A, B21121020-005A

**Run ID: Run Order:** PE 1\_211214A: 18      **SampType:** Continuing Calibration Verification Standard      **Batch ID:** R371801  
**Method:** SW8015C      **Analysis Date:** 12/14/2021 21:22      **Prep Date:**  
**Lab ID:** CCV\_1214PE121r      **Units:** ug/L      **Prep Method:**

Analytes	Result	LOQ	Spk value	Spk RefVal	%REC	LowLimit	HighLimit	RPD RefVal	%RPD	RPDLimit	Qual
C6 to C10	165	20	168		98.0	80	120				
Total Purgeable Hydrocarbons	196	20	200		98.0	80	120				
Surr: Trifluorotoluene	23	1.0	25		92.0	80	120				

Associated Samples: B21121020-001C, B21121020-002C, B21121020-003A, B21121020-005A

**Run ID: Run Order:** PE 1\_211214A: 30      **SampType:** Continuing Calibration Verification Standard      **Batch ID:** R371801  
**Method:** SW8015C      **Analysis Date:** 12/15/2021 08:49      **Prep Date:**  
**Lab ID:** CCV\_1214PE141r      **Units:** ug/L      **Prep Method:**

Analytes	Result	LOQ	Spk value	Spk RefVal	%REC	LowLimit	HighLimit	RPD RefVal	%RPD	RPDLimit	Qual
C6 to C10	160	20	168		95.0	80	120				
Total Purgeable Hydrocarbons	193	20	200		96.0	80	120				
Surr: Trifluorotoluene	23	1.0	25		91.0	80	120				

Associated Samples: B21121020-001C, B21121020-002C, B21121020-003A, B21121020-005A



### Analytical QC Summary Report

Prepared by Billings, MT Branch

**Client:** AECOM - Honolulu  
**Workorder:** B21121020  
**Project:** CV18F0126/60571032.02.20.01

**Report Date:** 01/20/2022

**Run ID: Run Order:** GCFID-HP5-B\_211214A: 10      **SampType:** Continuing Calibration Verification Standard      **Batch ID:** R371804  
**Method:** SW8015C      **Analysis Date:** 12/14/2021 19:59      **Prep Date:**  
**Lab ID:** CCV\_1214HP518r-W      **Units:** mg/L      **Prep Method:**

Analytes	Result	LOQ	Spk value	Spk RefVal	%REC	LowLimit	HighLimit	RPD RefVal	%RPD	RPDLimit	Qual
TEH(Oil Range)	4.9	0.30	5.0		98.0	80	120				
Surr: n-Triacontane	0.21	0.0020	0.20		104.0	80	120				

Associated Samples: **B21121020-001B, B21121020-002B**

**Run ID: Run Order:** GCFID-HP5-B\_211214A: 11      **SampType:** Continuing Calibration Verification Standard      **Batch ID:** R371804  
**Method:** SW8015C      **Analysis Date:** 12/14/2021 20:42      **Prep Date:**  
**Lab ID:** CCV\_1214HP519r      **Units:** mg/L      **Prep Method:**

Analytes	Result	LOQ	Spk value	Spk RefVal	%REC	LowLimit	HighLimit	RPD RefVal	%RPD	RPDLimit	Qual
Diesel Range Organics (C10 to C24)	15	0.30	15		100.0	80	120				
Total Extractable Hydrocarbons	16	0.30	15		104.0	80	120				
Surr: o-Terphenyl	0.20	0.0020	0.20		101.0	80	120				

Associated Samples: **B21121020-001B, B21121020-002B**

**Run ID: Run Order:** GCFID-HP5-B\_211214A: 17      **SampType:** Continuing Calibration Verification Standard      **Batch ID:** R371804  
**Method:** SW8015C      **Analysis Date:** 12/15/2021 06:47      **Prep Date:**  
**Lab ID:** CCV\_1214HP533r-W      **Units:** mg/L      **Prep Method:**

Analytes	Result	LOQ	Spk value	Spk RefVal	%REC	LowLimit	HighLimit	RPD RefVal	%RPD	RPDLimit	Qual
TEH(Oil Range)	5.1	0.30	5.0		101.0	80	120				
Surr: n-Triacontane	0.21	0.0020	0.20		106.0	80	120				

Associated Samples: **B21121020-001B, B21121020-002B**



### Analytical QC Summary Report

Prepared by Billings, MT Branch

**Client:** AECOM - Honolulu  
**Workorder:** B21121020  
**Project:** CV18F0126/60571032.02.20.01

**Report Date:** 01/20/2022

**Run ID: Run Order:** GCFID-HP5-B\_211214A: 18      **SampType:** Continuing Calibration Verification Standard      **Batch ID:** R371804  
**Method:** SW8015C      **Analysis Date:** 12/15/2021 07:32      **Prep Date:**  
**Lab ID:** CCV\_1214HP534r      **Units:** mg/L      **Prep Method:**

Analytes	Result	LOQ	Spk value	Spk RefVal	%REC	LowLimit	HighLimit	RPD RefVal	%RPD	RPDLimit	Qual
Diesel Range Organics (C10 to C24)	15	0.30	15		102.0	80	120				
Total Extractable Hydrocarbons	16	0.30	15		106.0	80	120				
Surr: o-Terphenyl	0.20	0.0020	0.20		102.0	80	120				

Associated Samples: **B21121020-001B, B21121020-002B**

**Run ID: Run Order:** GCFID-HP5-B\_211214B: 12      **SampType:** Continuing Calibration Verification Standard      **Batch ID:** R371919  
**Method:** SW8015C      **Analysis Date:** 12/16/2021 05:07      **Prep Date:**  
**Lab ID:** CCV\_1214HP564r-W      **Units:** mg/L      **Prep Method:**

Analytes	Result	LOQ	Spk value	Spk RefVal	%REC	LowLimit	HighLimit	RPD RefVal	%RPD	RPDLimit	Qual
TEH(Oil Range)	4.9	0.30	5.0		97.0	80	120				
Surr: n-Triacontane	0.21	0.0020	0.20		106.0	80	120				

Associated Samples: **B21121020-001B, B21121020-002B**

**Run ID: Run Order:** GCFID-HP5-B\_211214B: 13      **SampType:** Continuing Calibration Verification Standard      **Batch ID:** R371919  
**Method:** SW8015C      **Analysis Date:** 12/16/2021 05:50      **Prep Date:**  
**Lab ID:** CCV\_1214HP565r      **Units:** mg/L      **Prep Method:**

Analytes	Result	LOQ	Spk value	Spk RefVal	%REC	LowLimit	HighLimit	RPD RefVal	%RPD	RPDLimit	Qual
Diesel Range Organics (C10 to C24)	15	0.30	15		102.0	80	120				
Total Extractable Hydrocarbons	16	0.30	15		106.0	80	120				
Surr: o-Terphenyl	0.20	0.0020	0.20		102.0	80	120				

Associated Samples: **B21121020-001B, B21121020-002B**



### Analytical QC Summary Report

Prepared by Billings, MT Branch

**Client:** AECOM - Honolulu  
**Workorder:** B21121020  
**Project:** CV18F0126/60571032.02.20.01

**Report Date:** 01/20/2022

**Run ID: Run Order:** GCFID-HP5-B\_211214B: 20      **SampType:** Continuing Calibration Verification Standard      **Batch ID:** R371919  
**Method:** SW8015C      **Analysis Date:** 12/16/2021 13:43      **Prep Date:**  
**Lab ID:** CCV\_1214HP576r-W      **Units:** mg/L      **Prep Method:**

Analytes	Result	LOQ	Spk value	Spk RefVal	%REC	LowLimit	HighLimit	RPD RefVal	%RPD	RPDLimit	Qual
TEH(Oil Range)	4.9	0.30	5.0		98.0	80	120				
Surr: n-Triacontane	0.21	0.0020	0.20		106.0	80	120				

Associated Samples: **B21121020-001B, B21121020-002B**

**Run ID: Run Order:** GCFID-HP5-B\_211214B: 21      **SampType:** Continuing Calibration Verification Standard      **Batch ID:** R371919  
**Method:** SW8015C      **Analysis Date:** 12/16/2021 14:25      **Prep Date:**  
**Lab ID:** CCV\_1214HP577r      **Units:** mg/L      **Prep Method:**

Analytes	Result	LOQ	Spk value	Spk RefVal	%REC	LowLimit	HighLimit	RPD RefVal	%RPD	RPDLimit	Qual
Diesel Range Organics (C10 to C24)	15	0.30	15		101.0	80	120				
Total Extractable Hydrocarbons	16	0.30	15		105.0	80	120				
Surr: o-Terphenyl	0.20	0.0020	0.20		100.0	80	120				

Associated Samples: **B21121020-001B, B21121020-002B**





### Analytical QC Summary Report

Prepared by Billings, MT Branch

**Client:** AECOM - Honolulu  
**Workorder:** B21121020  
**Project:** CV18F0126/60571032.02.20.01

**Report Date:** 01/20/2022

**Run ID: Run Order:** SV5975.I\_211214A: 11      **SampType:** Method Blank      **Batch ID:** 162126  
**Method:** SW8270C      **Analysis Date:** 12/14/2021 22:55      **Prep Date:** 12/13/2021 10:12  
**Lab ID:** MB-162126      **Units:** ug/L      **Prep Method:** SW3510C

Analytes	Result	LOQ	Spk value	Spk RefVal	%REC	LowLimit	HighLimit	RPD RefVal	%RPD	RPDLimit	Qual
1-Methylnaphthalene	ND	0.10									
2-Methylnaphthalene	ND	0.10									
Naphthalene	ND	0.10									

Associated Samples: **B21121020-001A, B21121020-002A**

- A documented laboratory error occurred during extraction, which contributed to low surrogate recoveries in this sample. The surrogate Nitrobenzene-d5 was not recovered.

**Run ID: Run Order:** SV5975.I\_211214A: 12      **SampType:** Method Blank      **Batch ID:** 162126  
**Method:** SW8270C      **Analysis Date:** 12/14/2021 23:28      **Prep Date:** 12/13/2021 10:12  
**Lab ID:** MB-162126      **Units:** ug/L      **Prep Method:** SW3510C

Analytes	Result	LOQ	Spk value	Spk RefVal	%REC	LowLimit	HighLimit	RPD RefVal	%RPD	RPDLimit	Qual
Surr: 2-Fluorobiphenyl	27	2.0	100		27.0	53	106				S
Surr: Nitrobenzene-d5	ND	2.0	5.0		0.0	55	111				S
Surr: Terphenyl-d14	104	2.0	100		104.0	58	132				

Associated Samples: **B21121020-001A, B21121020-002A**



### Analytical QC Summary Report

Prepared by Billings, MT Branch

**Client:** AECOM - Honolulu  
**Workorder:** B21121020  
**Project:** CV18F0126/60571032.02.20.01

**Report Date:** 01/20/2022

**Run ID: Run Order:** SV5975.I\_211214A: 13      **SampType:** Laboratory Control Sample      **Batch ID:** 162126  
**Method:** SW8270C      **Analysis Date:** 12/15/2021 00:01      **Prep Date:** 12/13/2021 10:13  
**Lab ID:** LLCS-162126      **Units:** ug/L      **Prep Method:** SW3510C

Analytes	Result	LOQ	Spk value	Spk RefVal	%REC	LowLimit	HighLimit	RPD RefVal	%RPD	RPDLimit	Qual
1-Methylnaphthalene	0.28	0.10	5.0		6.0	41	115				S
2-Methylnaphthalene	0.21	0.10	5.0		4.0	39	114				S
Naphthalene	ND	0.10	5.0		0.0	43	114				S
Surr: 2-Fluorobiphenyl	0.79	0.10	5.0		16.0	53	106				S
Surr: Nitrobenzene-d5	ND	0.10	5.0		0.0	55	111				S
Surr: Terphenyl-d14	5.2	0.10	5.0		104.0	58	132				

Associated Samples: **B21121020-001A, B21121020-002A**

- A documented laboratory error occurred during extraction, which contributed to low surrogate and analyte recoveries in this sample. The surrogate Nitrobenzene-d5 and analyte Naphthalene were not recovered. The associated LCSDuplicate sample, LLCSD-162126, had normal recoveries.

**Run ID: Run Order:** SV5975.I\_211214A: 14      **SampType:** Laboratory Control Sample Duplicate      **Batch ID:** 162126  
**Method:** SW8270C      **Analysis Date:** 12/15/2021 00:33      **Prep Date:** 12/13/2021 10:13  
**Lab ID:** LLCSD-162126      **Units:** ug/L      **Prep Method:** SW3510C

Analytes	Result	LOQ	Spk value	Spk RefVal	%REC	LowLimit	HighLimit	RPD RefVal	%RPD	RPDLimit	Qual
1-Methylnaphthalene	3.1	0.10	5.0		61.0	41	115	0.28	166.0	40.0	R
2-Methylnaphthalene	3.5	0.10	5.0		69.0	39	114	0.21	177.0	40.0	R
Naphthalene	3.3	0.10	5.0		65.0	43	114	0.0		40.0	
Surr: 2-Fluorobiphenyl	3.4	0.10	5.0		68.0	53	106	0.0	0.0		
Surr: Nitrobenzene-d5	3.8	0.10	5.0		75.0	55	111	0.0	0.0		
Surr: Terphenyl-d14	5.3	0.10	5.0		105.0	58	132	0.0	0.0		

Associated Samples: **B21121020-001A, B21121020-002A**



### Analytical QC Summary Report

Prepared by Billings, MT Branch

**Client:** AECOM - Honolulu  
**Workorder:** B21121020  
**Project:** CV18F0126/60571032.02.20.01

**Report Date:** 01/20/2022

**Run ID: Run Order:** SV5975.I\_211214A: 16      **SampType:** Sample Matrix Spike      **Batch ID:** 162126  
**Method:** SW8270C      **Analysis Date:** 12/15/2021 01:39      **Prep Date:** 12/13/2021 10:16  
**Lab ID:** B21121001-001ALMS      **Units:** ug/L      **Prep Method:** SW3510C

Analytes	Result	LOQ	Spk value	Spk RefVal	%REC	LowLimit	HighLimit	RPD RefVal	%RPD	RPDLimit	Qual
1-Methylnaphthalene	2.4	0.10	5.2	0.078	45.0	41	115				
2-Methylnaphthalene	2.4	0.10	5.2	0.084	45.0	39	114				
Naphthalene	2.0	0.10	5.2	0.072	37.0	43	114				S
Surr: 2-Fluorobiphenyl	3.0	0.10	5.2	0.0	57.0	53	106				
Surr: Nitrobenzene-d5	2.1	0.10	5.2	0.0	41.0	55	111				S
Surr: Terphenyl-d14	5.1	0.10	5.2	0.0	100.0	58	132				

Associated Samples: **B21121020-001A, B21121020-002A**

- A documented laboratory error occurred during extraction, which contributed to low surrogate and analyte recoveries in this sample.

**Run ID: Run Order:** SV5975.I\_211214B: 6      **SampType:** Method Blank      **Batch ID:** 162189  
**Method:** SW8270C      **Analysis Date:** 12/15/2021 08:33      **Prep Date:** 12/14/2021 15:17  
**Lab ID:** MB-162189      **Units:** ug/L      **Prep Method:** SW3510C

Analytes	Result	LOQ	Spk value	Spk RefVal	%REC	LowLimit	HighLimit	RPD RefVal	%RPD	RPDLimit	Qual
1-Methylnaphthalene	ND	0.10									
2-Methylnaphthalene	ND	0.10									
Naphthalene	ND	0.10									

Associated Samples: **B21121020-002A**



### Analytical QC Summary Report

Prepared by Billings, MT Branch

**Client:** AECOM - Honolulu  
**Workorder:** B21121020  
**Project:** CV18F0126/60571032.02.20.01

**Report Date:** 01/20/2022

**Run ID: Run Order:** SV5975.I\_211214B: 13      **SampType:** Method Blank      **Batch ID:** 162189  
**Method:** SW8270C      **Analysis Date:** 12/15/2021 12:21      **Prep Date:** 12/14/2021 15:17  
**Lab ID:** MB-162189      **Units:** ug/L      **Prep Method:** SW3510C

Analytes	Result	LOQ	Spk value	Spk RefVal	%REC	LowLimit	HighLimit	RPD RefVal	%RPD	RPDLimit	Qual
Surr: 2-Fluorobiphenyl	67	2.0	100		67.0	53	106				
Surr: Nitrobenzene-d5	66	2.0	100		66.0	55	111				
Surr: Terphenyl-d14	101	2.0	100		101.0	58	132				

Associated Samples: **B21121020-002A**

**Run ID: Run Order:** SV5975.I\_211214B: 8      **SampType:** Laboratory Control Sample      **Batch ID:** 162189  
**Method:** SW8270C      **Analysis Date:** 12/15/2021 09:38      **Prep Date:** 12/14/2021 15:17  
**Lab ID:** LLCS-162189      **Units:** ug/L      **Prep Method:** SW3510C

Analytes	Result	LOQ	Spk value	Spk RefVal	%REC	LowLimit	HighLimit	RPD RefVal	%RPD	RPDLimit	Qual
1-Methylnaphthalene	3.1	0.10	5.0		62.0	41	115				
2-Methylnaphthalene	3.6	0.10	5.0		72.0	39	114				
Naphthalene	3.3	0.10	5.0		66.0	43	114				
Surr: 2-Fluorobiphenyl	4.2	0.10	5.0		84.0	53	106				
Surr: Nitrobenzene-d5	4.6	0.10	5.0		92.0	55	111				
Surr: Terphenyl-d14	4.9	0.10	5.0		98.0	58	132				

Associated Samples: **B21121020-002A**

**Run ID: Run Order:** SV5975.I\_211214B: 9      **SampType:** Laboratory Control Sample Duplicate      **Batch ID:** 162189  
**Method:** SW8270C      **Analysis Date:** 12/15/2021 10:11      **Prep Date:** 12/14/2021 15:17  
**Lab ID:** LLCSD-162189      **Units:** ug/L      **Prep Method:** SW3510C

Analytes	Result	LOQ	Spk value	Spk RefVal	%REC	LowLimit	HighLimit	RPD RefVal	%RPD	RPDLimit	Qual
1-Methylnaphthalene	3.7	0.10	5.0		74.0	41	115	3.1	16.0	40.0	
2-Methylnaphthalene	4.2	0.10	5.0		83.0	39	114	3.6	14.0	40.0	
Naphthalene	3.8	0.10	5.0		75.0	43	114	3.3	13.0	40.0	



### Analytical QC Summary Report

Prepared by Billings, MT Branch

**Client:** AECOM - Honolulu  
**Workorder:** B21121020  
**Project:** CV18F0126/60571032.02.20.01

**Report Date:** 01/20/2022

**Run ID: Run Order:** SV5975.I\_211214B: 9      **SampType:** Laboratory Control Sample Duplicate      **Batch ID:** 162189  
**Method:** SW8270C      **Analysis Date:** 12/15/2021 10:11      **Prep Date:** 12/14/2021 15:17  
**Lab ID:** LLCSD-162189      **Units:** ug/L      **Prep Method:** SW3510C

Analytes	Result	LOQ	Spk value	Spk RefVal	%REC	LowLimit	HighLimit	RPD RefVal	%RPD	RPDLimit	Qual
Surr: 2-Fluorobiphenyl	4.9	0.10	5.0		97.0	53	106	0.0	0.0		
Surr: Nitrobenzene-d5	4.9	0.10	5.0		99.0	55	111	0.0	0.0		
Surr: Terphenyl-d14	5.4	0.10	5.0		108.0	58	132	0.0	0.0		

Associated Samples: **B21121020-002A**

**Run ID: Run Order:** SV5975.I\_211214B: 14      **SampType:** Sample Matrix Spike      **Batch ID:** 162189  
**Method:** SW8270C      **Analysis Date:** 12/15/2021 12:54      **Prep Date:** 12/14/2021 14:45  
**Lab ID:** B21121020-002ALMS      **Units:** ug/L      **Prep Method:** SW3510C

Analytes	Result	LOQ	Spk value	Spk RefVal	%REC	LowLimit	HighLimit	RPD RefVal	%RPD	RPDLimit	Qual
1-Methylnaphthalene	18	0.96	4.8	14	84.0	41	115				
2-Methylnaphthalene	15	0.96	4.8	11	79.0	39	114				
Naphthalene	37	0.96	4.8	30	155.0	43	114				S
Surr: 2-Fluorobiphenyl	3.5	0.96	4.8	0.0	73.0	53	106				
Surr: Nitrobenzene-d5	4.2	0.96	4.8	0.0	87.0	55	111				
Surr: Terphenyl-d14	4.7	0.96	4.8	0.0	98.0	58	132				

Associated Samples: **B21121020-002A**



### Analytical QC Summary Report

Prepared by Billings, MT Branch

**Client:** AECOM - Honolulu  
**Workorder:** B21121020  
**Project:** CV18F0126/60571032.02.20.01

**Report Date:** 01/20/2022

**Run ID: Run Order:** SV5975.I\_211214B: 2      **SampType:** Continuing Calibration Verification Standard      **Batch ID:** R371813  
**Method:** SW8270C      **Analysis Date:** 12/15/2021 06:23      **Prep Date:**  
**Lab ID:** 14-Dec-21\_CCV\_25      **Units:** ug/L      **Prep Method:**

Analytes	Result	LOQ	Spk value	Spk RefVal	%REC	LowLimit	HighLimit	RPD RefVal	%RPD	RPDLimit	Qual
1-Methylnaphthalene	2.0	0.10	2.0		98.0	80	120				
2-Methylnaphthalene	2.0	0.10	2.0		100.0	80	120				
Naphthalene	1.9	0.10	2.0		93.0	80	120				
Surr: 2-Fluorobiphenyl	2.2	0.10	2.0		108.0	80	120				
Surr: Nitrobenzene-d5	2.1	0.10	2.0		103.0	80	120				
Surr: Terphenyl-d14	1.8	0.10	2.0		92.0	80	120				

Associated Samples: **B21121020-001A, B21121020-002A**

**Run ID: Run Order:** SV5975.I\_211214B: 15      **SampType:** Continuing Calibration Verification Standard      **Batch ID:** R371813  
**Method:** SW8270C      **Analysis Date:** 12/15/2021 13:26      **Prep Date:**  
**Lab ID:** 14-Dec-21\_CCV\_38      **Units:** ug/L      **Prep Method:**

Analytes	Result	LOQ	Spk value	Spk RefVal	%REC	LowLimit	HighLimit	RPD RefVal	%RPD	RPDLimit	Qual
1-Methylnaphthalene	2.0	0.10	2.0		99.0	80	120				
2-Methylnaphthalene	2.0	0.10	2.0		99.0	80	120				
Naphthalene	1.8	0.10	2.0		91.0	80	120				
Surr: 2-Fluorobiphenyl	2.1	0.10	2.0		107.0	80	120				
Surr: Nitrobenzene-d5	2.3	0.10	2.0		113.0	80	120				
Surr: Terphenyl-d14	1.9	0.10	2.0		95.0	80	120				

Associated Samples: **B21121020-001A, B21121020-002A**

### Analytical QC Exceptions Report

Prepared by Billings, MT Branch

**Client:** AECOM - Honolulu  
**Workorder:** B21121020  
**Project:** CV18F0126/60571032.02.20.01

**Report Date:** 01/20/2022

Analysis Method	Analysis	Batch ID	Associated Samples	Sample Type	Lab ID	Analysis Date	Analysis Time	Analyte	%REC	Low Limit	High Limit	% RPD	RPD Limit	Qual
SW8270C	Low Level PAH	162126	001A, 002A	MBLK	MB-162126	12/14/2021	23:28	Surr: 2-Fluorobiphenyl	27.0	53	106			S
								Surr: Nitrobenzene-d5	0.0	55	111			S
				LCS-DOD	LLCS-162126	12/15/2021	00:01	Surr: 2-Fluorobiphenyl	16.0	53	106			S
								Surr: Nitrobenzene-d5	0.0	55	111			S
								1-Methylnaphthalene	6.0	41	115			S
								2-Methylnaphthalene	4.0	39	114			S
		LCSD-DOD	LLCSD-162126	12/15/2021	00:33	1-Methylnaphthalene	61.0	41	115	166	40.0	R		
						2-Methylnaphthalene	69.0	39	114	177	40.0	R		
						Surr: Nitrobenzene-d5	41.0	55	111			S		
		MS-DOD	B21121001-001ALMS	12/15/2021	01:39	Naphthalene	37.0	43	114			S		
						MS-DOD	B21121020-002ALMS	12/15/2021	12:54	Naphthalene	155.0	43	114	



## Preparation and Analysis Dates Report

Work Order: B21121020

Client: AECOM - Honolulu

Project Name: CV18F0126/60571032.02.20.01

Report Date: 1/20/2022

Lab ID	Client Sample ID	Collection Date	Matrix	Test Name	TCLP Date	Prep Method	Prep Date	Prep Batch	Analysis Method	Analysis Date
001A	ERH2213 (RHMW01R)	12/08/2021 15:10	Ground Water	Low Level PAH		SW3510C	12/13/2021 16:54	162126	SW8270C	12/15/2021 07:28
001B	ERH2213 (RHMW01R)	12/08/2021 15:10	Ground Water	Diesel Range Organics		SW3520C	12/13/2021 13:19	162151	SW8015C	12/14/2021 23:35
						SW3520C	12/13/2021 13:19	162151	SW8015C	12/16/2021 07:16
002A	ERH2215 (RHMW02)	12/08/2021 16:35	Ground Water	Low Level PAH		SW3510C	12/13/2021 16:54	162126	SW8270C	12/15/2021 08:01
						SW3510C	12/13/2021 16:54	162189	SW8270C	12/15/2021 11:49
002B	ERH2215 (RHMW02)	12/08/2021 16:35	Ground Water	Diesel Range Organics		SW3520C	12/13/2021 13:19	162151	SW8015C	12/15/2021 01:02
						SW3520C	12/13/2021 13:19	162151	SW8015C	12/16/2021 07:59





## Chemical Abstracts Service (CAS) Registry Numbers

Prepared by Billings, MT Branch

**Client:** AECOM - Honolulu

**Workorder:** B21121020

**Project:** CV18F0126/60571032.02.20.01


**Report Date:** 01/20/2022

Analyses	CAS No
<b>AGGREGATE ORGANICS</b>	
Organic Carbon, Total (TOC)	7440-44-0
<b>VOLATILE ORGANIC COMPOUNDS</b>	
Benzene	71-43-2
Ethylbenzene	100-41-4
Toluene	108-88-3
m+p-Xylenes	179601-23-1
o-Xylene	95-47-6
Xylenes, Total	1330-20-7
<b>PETROLEUM HYDROCARBONS-VOLATILE</b>	
C6 to C10	
Total Purgeable Hydrocarbons	
<b>PETROLEUM HYDROCARBONS-SEMI-VOLATILE</b>	
Diesel Range Organics (C10 to C24)	
Diesel Range Organics (SGT-C10 to C24)	
Oil Range Hydrocarbons (C24 to C40)	
Oil Range Hydrocarbons (SGT-C24 to C40)	
Total Extractable Hydrocarbons	
Total Extractable Hydrocarbons (SGT)	
<b>SEMI-VOLATILE ORGANIC COMPOUNDS (LOW LEVEL) BY SIM</b>	
1-Methylnaphthalene	90-12-0
2-Methylnaphthalene	91-57-6
Naphthalene	91-20-3

ERH2213 (RHMW01R)

G:\Org\PE1\DAT\PE1121421\_b\1214PE1B.0036.RAW

B21121020-001C ;1214PE1 , \$HC-8015-GRO-W,



**GASOLINE RANGE ORGANICS CHROMATOGRAM REPORT**

Sample Name: B21121020-001C ;1214PE1 , \$HC-8015-GRO-W,  
Raw File: G:\Org\PE1\DAT\PE1121421\_b\1214PE1B.0036.RAW  
Date & Time Acquired: 12/15/2021 5:57:53 AM  
Method File: G:\Org\PE1\Methods\211208G1020-1B%.MET  
Calibration File: G:\Org\PE1\Cals\211208GRO8015CB.CAL  
Sample Weight: 5 Dilution: 1 S.A.: 1

Mean RF for GRO: 945.9678  
Mean RF for TPH: 909.3915  
Rt range for Gasoline Range Organics: 4.45 to 17.93


SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
**Trifluorotoluene	8.69	25.	18.844	75.37

GRO Area:10248.96 GRO Amount: 2.166874  
TPH Area:165066.4 TPH Amount: 36.3026

ERH2215 (RHMW02)

G:\Org\PE1\DAT\PE1121421\_b\1214PE1B.0038.RAW

B21121020-002C ;1214PE1 , \$HC-8015-GRO-W,



**GASOLINE RANGE ORGANICS CHROMATOGRAM REPORT**

Sample Name: B21121020-002C ;1214PE1 , \$HC-8015-GRO-W,  
Raw File: G:\Org\PE1\DAT\PE1121421\_b\1214PE1B.0038.RAW  
Date & Time Acquired: 12/15/2021 7:06:23 AM  
Method File: G:\Org\PE1\Methods\211208G1020-2B%.MET  
Calibration File: G:\Org\PE1\Cals\211208GRO8015CB.CAL  
Sample Weight: 5 Dilution: 1 S.A.: 1

Mean RF for GRO: 945.9678  
Mean RF for TPH: 909.3915  
Rt range for Gasoline Range Organics: 4.45 to 17.93


SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
**Trifluorotoluene	8.692	25.	19.684	78.74

GRO Area:59099.18 GRO Amount: 12.49497  
TPH Area:1078927 TPH Amount: 237.2854

ERH2212 (Trip Blank)-14525

G:\Org\PE1\DAT\PE1121421\_b\1214PE1B.0015.RAW

B21121020-003A ;1214PE1 , \$HC-8015-GRO-W,



**GASOLINE RANGE ORGANICS CHROMATOGRAM REPORT**

Sample Name: B21121020-003A ;1214PE1 , \$HC-8015-GRO-W,  
Raw File: G:\Org\PE1\DAT\PE1121421\_b\1214PE1B.0015.RAW  
Date & Time Acquired: 12/14/2021 5:56:57 PM  
Method File: G:\Org\PE1\Methods\211208GROB%.MET  
Calibration File: G:\Org\PE1\Cals\211208GRO8015CB.CAL  
Sample Weight: 5 Dilution: 1 S.A.: 1

Mean RF for GRO: 945.9678  
Mean RF for TPH: 909.3915  
Rt range for Gasoline Range Organics: 4.45 to 17.93


SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
**Trifluorotoluene	8.695	25.	20.999	84.

GRO Area:2998.399 GRO Amount: 0.6339325  
TPH Area:7467.492 TPH Amount: 1.642305

ERH2214 (Trip Blank)-14525

G:\Org\PE1\DAT\PE1121421\_b\1214PE1B.0016.RAW

B21121020-005A ;1214PE1 , \$HC-8015-GRO-W,



**GASOLINE RANGE ORGANICS CHROMATOGRAM REPORT**

Sample Name: B21121020-005A ;1214PE1 , \$HC-8015-GRO-W,  
Raw File: G:\Org\PE1\DAT\PE1121421\_b\1214PE1B.0016.RAW  
Date & Time Acquired: 12/14/2021 6:31:23 PM  
Method File: G:\Org\PE1\Methods\211208GROB%.MET  
Calibration File: G:\Org\PE1\Cals\211208GRO8015CB.CAL  
Sample Weight: 5 Dilution: 1 S.A.: 1

Mean RF for GRO: 945.9678  
Mean RF for TPH: 909.3915  
Rt range for Gasoline Range Organics: 4.45 to 17.93

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
**Trifluorotoluene	8.694	25.	21.211	84.85


GRO Area:3960.696 GRO Amount: 0.837385  
TPH Area:6675.809 TPH Amount: 1.468192

ERH2213 (RHMW01R)

G:\org\HP5\DAT\HP5121421\_b\1214HP5.0023.RAW

Batch ID: 162151

B21121020-001B ;1214HP5 , \$HC-8015-DRO-W,



**DIESEL RANGE ORGANICS CHROMATOGRAM REPORT**

Sample Name: B21121020-001B ;1214HP5 , \$HC-8015-DRO-W,  
 Raw File: G:\org\HP5\DAT\HP5121421\_b\1214HP5.0023.RAW  
 Date & Time Acquired: 12/14/2021 11:35:45 PM  
 Method File: G:\Org\HP5\Methods\D3\_8015-121423-IH-L%.met  
 Calibration File: G:\Org\HP5\Cals\SW8015C\_DRO211102IH-24-Tri.CAL  
 Sample Weight: 1050 Dilution: 1 S.A.: 1

Mean RF for TEH: 31353.19

Rt range for Diesel Range Organics: 6.46 to 14.35

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*o-Terphenyl	12.173	.19	.187	98.38	-
*1-Chlorooctadecane	12.982	.19	.003	1.4	-
*#Triacontane	16.247	.19	.102	53.6	-


DRO Area: 9412167 DRO Amount: 0.2859029  
 TEH Area: 1.95192E+07 TEH Amount: 0.5929129

ERH2213 (RHMW01R)

G:\org\HP5\DAT\HP5121421\_b\1214HP5.0023.RAW

Batch ID: 162151

B21121020-001B ;1214HP5 , \$HC-8015-DRO-W,



**RESIDUAL RANGE ORGANICS CHROMATOGRAM**

Sample Name: B21121020-001B ;1214HP5 , \$HC-8015-DRO-W,  
Raw File: G:\org\HP5\DAT\HP5121421\_b\1214HP5.0023.RAW  
Date & Time Acquired: 12/14/2021 11:35:45 PM  
Method File: G:\Org\HP5\Methods\D3\_OROS-121423-AI-L%.MET  
Calibration File: G:\Org\HP5\Cals\SW8015C\_ORO211017AI-SAMP.CAL  
Sample Weight: 1050 Dilution: 1 S.A.: 1

Mean RF for for Residual Range Organics Calculations: 28542.41  
Rt range for Residual Range Organics: 14.25 to 21.5

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*#Triacontane	16.247	.476	.102	21.44


RRO Area:6850202 RRO AMOUNT: 0.2285722

ERH2215 (RHMW02)

Batch ID: 162151

G:\Org\HP5\DAT\HP5121421\_b\1214HP5.0025.RAW

B21121020-002B ;1214HP5 , \$HC-8015-DRO-W,



**DIESEL RANGE ORGANICS CHROMATOGRAM REPORT**

Sample Name: B21121020-002B ;1214HP5 , \$HC-8015-DRO-W,  
Raw File: G:\Org\HP5\DAT\HP5121421\_b\1214HP5.0025.RAW  
Date & Time Acquired: 12/15/2021 1:02:11 AM  
Method File: G:\Org\HP5\Methods\D3\_8015-121423-IH-L%.met  
Calibration File: G:\Org\HP5\Cals\SW8015C\_DRO211102IH-24-Tri.CAL  
Sample Weight: 1030 Dilution: 1 S.A.: 1

Mean RF for TEH: 31353.19

Rt range for Diesel Range Organics: 6.46 to 14.35

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*o-Terphenyl	12.173	.194	.198	102.22	-
*1-Chlorooctadecane	29.936	.194	.	.	-
*#Triacontane	16.246	.194	.104	53.79	-

DRO Area: 5.334132E+07 DRO Amount: 1.651752  
TEH Area: 6.706593E+07 TEH Amount: 2.076744




ERH2215 (RHMW02)

G:\org\HP5\DAT\HP5121421\_b\1214HP5.0025.RAW

Batch ID: 162151

B21121020-002B ;1214HP5 , \$HC-8015-DRO-W,



**RESIDUAL RANGE ORGANICS CHROMATOGRAM**

Sample Name: B21121020-002B ;1214HP5 , \$HC-8015-DRO-W,  
Raw File: G:\org\HP5\DAT\HP5121421\_b\1214HP5.0025.RAW  
Date & Time Acquired: 12/15/2021 1:02:11 AM  
Method File: G:\Org\HP5\Methods\D3\_OROS-121423-AI-L%.MET  
Calibration File: G:\Org\HP5\Cals\SW8015C\_ORO211017AI-SAMP.CAL  
Sample Weight: 1030 Dilution: 1 S.A.: 1

Mean RF for for Residual Range Organics Calculations: 28542.41  
Rt range for Residual Range Organics: 14.25 to 21.5

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*#Triacontane	16.246	.485	.104	21.52


RRO Area:1.012333E+07 RRO AMOUNT: 0.3443465

ERH2213 (RHMW01R)

Batch ID: 162151

G:\org\HP5\DAT\HP5121421\_b\1214HP5.0067.RAW

B21121020-001B ;1214HP5 , \$HC-8015-DRO-W, SGT



**DIESEL RANGE ORGANICS CHROMATOGRAM REPORT**

Sample Name: B21121020-001B ;1214HP5 , \$HC-8015-DRO-W, SGT  
 Raw File: G:\org\HP5\DAT\HP5121421\_b\1214HP5.0067.RAW  
 Date & Time Acquired: 12/16/2021 7:16:33 AM  
 Method File: G:\Org\HP5\Methods\DR\_8015-C24T-II-L%.met  
 Calibration File: G:\Org\HP5\Cals\SW8015C\_DRO211102II-24-Tri.CAL  
 Sample Weight: 1050 Dilution: 1 S.A.: 1

Mean RF for TEH: 31353.19

Rt range for Diesel Range Organics: 6.45 to 14.32

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*o-Terphenyl	12.159	.19	.17	89.33	-
*1-Chlorooctadecane	12.999	.19	.	.12	-
*#Triacontane	16.231	.19	.087	45.9	-


DRO Area:1825767 DRO Amount: 5.545928E-02  
 TEH Area:2243469 TEH Amount: 6.814735E-02

ERH2213 (RHMW01R)

G:\org\HP5\DAT\HP5121421\_b\1214HP5.0067.RAW

Batch ID: 162151

B21121020-001B ;1214HP5 , \$HC-8015-DRO-W, SGT



**RESIDUAL RANGE ORGANICS CHROMATOGRAM**

Sample Name: B21121020-001B ;1214HP5 , \$HC-8015-DRO-W, SGT  
 Raw File: G:\org\HP5\DAT\HP5121421\_b\1214HP5.0067.RAW  
 Date & Time Acquired: 12/16/2021 7:16:33 AM  
 Method File: G:\Org\HP5\Methods\DR\_OROS-AJ-L%.MET  
 Calibration File: G:\Org\HP5\Cals\SW8015C\_ORO211017AJ-SAMP.CAL  
 Sample Weight: 1050 Dilution: 1 S.A.: 1

Mean RF for for Residual Range Organics Calculations: 28542.41  
 Rt range for Residual Range Organics: 14.22 to 21.4

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*#Triacontane	16.231	.476	.087	18.36


RRO Area:249006.2 RRO AMOUNT: 8.308644E-03

ERH2215 (RHMW02)

G:\org\HP5\DAT\HP5121421\_b\1214HP5.0068.RAW

Batch ID: 162151

B21121020-002B ;1214HP5 , \$HC-8015-DRO-W, SGT



**DIESEL RANGE ORGANICS CHROMATOGRAM REPORT**

Sample Name: B21121020-002B ;1214HP5 , \$HC-8015-DRO-W, SGT  
 Raw File: G:\org\HP5\DAT\HP5121421\_b\1214HP5.0068.RAW  
 Date & Time Acquired: 12/16/2021 7:59:47 AM  
 Method File: G:\Org\HP5\Methods\DR\_8015-C24T-II-L%.met  
 Calibration File: G:\Org\HP5\Cals\SW8015C\_DRO211102II-24-Tri.CAL  
 Sample Weight: 1030 Dilution: 1 S.A.: 1

Mean RF for TEH: 31353.19

Rt range for Diesel Range Organics: 6.45 to 14.32

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*o-Terphenyl	12.159	.194	.168	86.64	-
*1-Chlorooctadecane	13.003	.194	.	.08	-
*#Triacontane	16.233	.194	.087	44.85	-


DRO Area:1.31982E+07 DRO Amount: 0.4086917  
 TEH Area:1.373437E+07 TEH Amount: 0.4252945

ERH2215 (RHMW02)

G:\org\HP5\DAT\HP5121421\_b\1214HP5.0068.RAW

Batch ID: 162151

B21121020-002B ;1214HP5 , \$HC-8015-DRO-W, SGT



**RESIDUAL RANGE ORGANICS CHROMATOGRAM**

Sample Name: B21121020-002B ;1214HP5 , \$HC-8015-DRO-W, SGT  
 Raw File: G:\org\HP5\DAT\HP5121421\_b\1214HP5.0068.RAW  
 Date & Time Acquired: 12/16/2021 7:59:47 AM  
 Method File: G:\Org\HP5\Methods\DR\_OROS-AJ-L%.MET  
 Calibration File: G:\Org\HP5\Cals\SW8015C\_ORO211017AJ-SAMP.CAL  
 Sample Weight: 1030 Dilution: 1 S.A.: 1

Mean RF for for Residual Range Organics Calculations: 28542.41  
 Rt range for Residual Range Organics: 14.22 to 21.4

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*#Triacontane	16.233	.485	.087	17.94

RRO Area:231133.6 RRO AMOUNT: 7.862039E-03

---

**From:** Ramos, Alethea <alethea.ramos@aecom.com>  
**Sent:** Monday, December 13, 2021 3:11 PM  
**To:** Tabitha Edwards  
**Cc:** Pascua, Margie; billingsPM@energylab.com  
**Subject:** RE: [EXTERNAL] FW: CV18F0126: Expedited NOI Groundwater Samples, Saturday 12/12 Submission

**Categories:** Must Attend

Hi Tabitha,

I believe Casper WY is DoD ELAP accredited in the TOC 9060 method. I spoke to Shari and she indicated there is a daily courier between Billings and Casper, and would be appx. a day delay. Under those stipulations, please subcontract these samples and inform on expedited TAT.

Thank you,

**Alethea Ramos, CIH**  
Environmental Scientist, Environmental Health & Science, Environment  
D +1-808-529-7283  
M +1-808-389-5383  
[alethea.ramos@aecom.com](mailto:alethea.ramos@aecom.com)

**AECOM**  
1001 Bishop Street  
Suite 1600  
Honolulu, HI 96813, United States of America  
aecom.com

Imagine it. Delivered.

[LinkedIn](#) | [Twitter](#) | [Facebook](#) | [Instagram](#)



[Fortune World's Most Admired Companies 2020](#)

---

**From:** Tabitha Edwards <tedwards@energylab.com>  
**Sent:** Monday, December 13, 2021 7:05 AM  
**To:** Ramos, Alethea <alethea.ramos@aecom.com>  
**Cc:** Pascua, Margie <Margie.Pascua@aecom.com>; billingsPM@energylab.com  
**Subject:** [EXTERNAL] FW: CV18F0126: Expedited NOI Groundwater Samples, Saturday 12/12 Submission  
**Importance:** High

Alethea,

The TOC by 9060 must be subcontracted to our office in Casper, WY. I need authorization from you to subcontract these. Once that has been received we will discuss the TAT with them and let you know what is achievable.

Thank you,

**Energy Laboratories, Inc.**

Trust our People. Trust our Data.

**Tabitha Edwards** | Office Manager | Billings, MT

O: 406-869-6286 | [tedwards@energylab.com](mailto:tedwards@energylab.com) | [www.energylab.com](http://www.energylab.com)

This transmission may contain confidential information and is for the use of the intended recipient(s). If you received this in error, please contact the sender and delete this email and all copies.

***We want to help you ship successfully!** Please plan ahead and allow extra time to receive supplies from the lab and for the lab to receive your samples. All carriers are in full-swing holiday peak season operating with double the volume and limited capacity. We appreciate your business so please contact your local branch or Project Manager to discuss adjustments to your shipping schedule or to ask questions.*

---

**From:** Ramos, Alethea [<mailto:alethea.ramos@aecom.com>]

**Sent:** Saturday, December 11, 2021 3:20 AM

**To:** Shari Endy; [billingsPM@energylab.com](mailto:billingsPM@energylab.com)

**Cc:** Jillian Miller; Pascua, Margie; KaaihiliChoy, Terri Ann

**Subject:** CV18F0126: Expedited NOI Groundwater Samples, Saturday 12/12 Submission

**Importance:** High

Hi Shari and Billings PM,

You will be receiving a Saturday shipment (12/12) of groundwater samples indicated in the attached COCs. We will need results by **Wednesday, December 15<sup>th</sup>**, and will pay any fees incurred for an expedited TAT. Please proceed with analysis without preservation traceability. Please see below tracking information links:

<https://www.fedex.com/fedextrack/?trknbr=287337969629&trkqual=2459558000~287337969629~FX>

<https://www.fedex.com/fedextrack/?trknbr=287343101019&trkqual=2459559000~287343101019~FX>

Thank you,

**Alethea Ramos, CIH**

Environmental Scientist, Environmental Health & Science, Environment

D +1-808-529-7283

M +1-808-389-5383

[alethea.ramos@aecom.com](mailto:alethea.ramos@aecom.com)

**AECOM**

1001 Bishop Street  
Suite 1600  
Honolulu, HI 96813, United States of America  
aecom.com

**Imagine it. Delivered.**

[LinkedIn](#) | [Twitter](#) | [Facebook](#) | [Instagram](#)



[Fortune World's Most Admired Companies 2020](#)