

Energy Laboratories Inc

ANALYTICAL RUN Summary

03-Nov-21

Run ID GCFID-HP5-B_211102A

Run Start Date: 11/2/2021
Analyst: Ann Nebel
Ical:
Column ID:
Comments: ICAL for 8015C_DRO211002IA

Std ID	Std Name	Std Amount	Std Units	Samp Amount	Samp Units	SampType	Expiration Date
DRO211012A	Diesel Fuel #2 50,000 ug/mL in DCM					CAL-DIESE	4/30/2023
DRO211012B	#2 Diesel in Acetone 150,000 ug/mL					SECOND S	11/5/2023
DRO211025A	ALI CCV Mix-200ug/mL					MARKER	5/31/2022
DRO211101A	OTP-4000 ug/mL DCM					CAL-SURR	9/30/2024

Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist					
14822192	CCV_1102HP50	HC-8015-DRO-	CCV		11/2/2021 8:31:3	1	R369667		0	0						
Analyte	T	Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD	Q
Total Extractable Hydrocarbons	A	mg/L		3.665751		15	0	0	0.0749	0.3	50	24%	80	120	0%	S
o-Terphenyl	S	mg/L		0.2015393		0.2	0	0	0.000429	0.002	0	101%	80	120	0%	

Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist					
14822193	CCV_1102HP50	HC-8015-DRO-	CAL1		11/2/2021 9:57:0	1	R369667		0	0						
Analyte	T	Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD	Q
o-Terphenyl	S	mg/L		0.00195342		0.002	0	0	0.000429	0.002	0	98%	80	120	0%	

Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist					
14822194	CCV_1102HP50	HC-8015-DRO-	CAL2		11/2/2021 10:39:	1	R369667		0	0						
Analyte	T	Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD	Q
o-Terphenyl	S	mg/L		0.04992196		0.05	0	0	0.000429	0.002	0	100%	80	120	0%	

Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist					
14822195	CCV_1102HP50	HC-8015-DRO-	CAL3		11/2/2021 11:22:	1	R369667		0	0						
Analyte	T	Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD	Q
o-Terphenyl	S	mg/L		0.1939565		0.2	0	0	0.000429	0.002	0	97%	80	120	0%	
Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist					
14822196	CCV_1102HP50	HC-8015-DRO-	CAL4		11/2/2021 12:05:	1	R369667		0	0						
Analyte	T	Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD	Q
o-Terphenyl	S	mg/L		0.5113316		0.5	0	0	0.000429	0.002	0	102%	80	120	0%	
Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist					
14822197	CCV_1102HP50	HC-8015-DRO-	CAL5		11/2/2021 12:49:	1	R369667		0	0						
Analyte	T	Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD	Q
o-Terphenyl	S	mg/L		1.032406		1	0	0	0.000429	0.002	0	103%	80	120	0%	
Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist					
14822198	CCV_1102HP51	HC-8015-DRO-	CAL1		11/2/2021 1:32:0	1	R369667		0	0						
Analyte	T	Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD	Q
Total Extractable Hydrocarbons	A	mg/L		0.1551854		0.15	0	0	0.0749	0.3	50	103%	80	120	0%	
Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist					
14822199	CCV_1102HP51	HC-8015-DRO-	CAL2		11/2/2021 2:15:0	1	R369667		0	0						
Analyte	T	Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD	Q
Total Extractable Hydrocarbons	A	mg/L		3.698073		3.75	0	0	0.0749	0.3	50	99%	80	120	0%	
Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist					
14822200	CCV_1102HP51	HC-8015-DRO-	CAL3		11/2/2021 2:58:2	1	R369667		0	0						
Analyte	T	Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD	Q
Total Extractable Hydrocarbons	A	mg/L		15.26249		15	0	0	0.0749	0.3	50	102%	80	120	0%	

Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist					
14822201	CCV_1102HP51	HC-8015-DRO-	CAL4		11/2/2021 3:41:3	1	R369667		0	0						
Analyte	T	Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD	Q
Total Extractable Hydrocarbons	A	mg/L		36.59341		37.5	0	0	0.0749	0.3	50	98%	80	120	0%	
Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist					
14822202	CCV_1102HP51	HC-8015-DRO-	CAL5		11/2/2021 4:24:5	1	R369667		0	0						
Analyte	T	Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD	Q
Total Extractable Hydrocarbons	A	mg/L		49.2977		50	0	0	0.0749	0.3	50	99%	80	120	0%	
Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist					
14822203	CCV_1102HP51	HC-8015-DRO-	ICV		11/2/2021 5:51:3	1	R369667		0	0						
Analyte	T	Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD	Q
Total Extractable Hydrocarbons	A	mg/L		14.98994		15	0	0	0.0749	0.3	50	100%	80	120	0%	

Write Sequence	Data File	Sample Name	Method	Weight	Dil Factor	Amt Inj.	IS	Cal ID
		CCV_1102HP508r, DRO ;1102HP5 , DRO211025A	G:\Org\HP5\Methods\DC_8015-IA-L%.met	1	1	1	1	0
		DCM-Baseline Check-V04	G:\Org\HP5\Methods\DR_8015-HP-LEXP.met	1	1	1	1	0
		CCV_1102HP505r, CAL1 ;1102HP5 , 2 ug per mL OTP (10 uL of Cal3 + 990 uL DCM(14408)	G:\Org\HP5\Methods\DS_8015-IA-L#.met	1	1	1	1	0
		CCV_1102HP506r, CAL2 ;1102HP5 , 50 ug per mL OTP (100 uL Cal4 + 900 uL of DCM(14408)	G:\Org\HP5\Methods\DS_8015-IA-L#.met	1	1	1	1	0
		CCV_1102HP507r, CAL3 ;1102HP5 , 200 ug per mL OTP (100uL of Cal5 + 400 uL DCM(14408)	G:\Org\HP5\Methods\DS_8015-IA-L#.met	1	1	1	1	0
		CCV_1102HP508r, CAL4 ;1102HP5 , 500 ug per mL OTP (250uL of Cal5 + 250 uL DCM(14408)	G:\Org\HP5\Methods\DS_8015-IA-L#.met	1	1	1	1	0
		CCV_1102HP509r, CAL5 ;1102HP5 , 1000 ug per mL OTP (250 uL 4000 ug/mL OTP DRO21101A + 750 DCM(14408)	G:\Org\HP5\Methods\DS_8015-IA-L#.met	1	1	1	1	0
		CCV_1102HP510r, CAL1 ;1102HP5 , 150 ug per mL Diesel (10 uL of Cal3 + 990 uL DCM(14408),	G:\Org\HP5\Methods\DC_8015-IA-L%.met	1	1	1	1	0
		CCV_1102HP511r, CAL2 ;1102HP5 , 3750 ug per mL Diesel (100 uL Cal4 + 900 uL of DCM(14408)	G:\Org\HP5\Methods\DC_8015-IA-L%.met	1	1	1	1	0
		CCV_1102HP512r, CAL3 ;1102HP5 , 15000 ug per mL Diesel (300 uL of DRO211012A + 700 uL DCM(14408)	G:\Org\HP5\Methods\DC_8015-IA-L%.met	1	1	1	1	0
		CCV_1102HP513r, CAL4 ;1102HP5 , 37500ug per mL Diesel (750 uL of DRO211012A + 250 uL DCM(14408)	G:\Org\HP5\Methods\DC_8015-IA-L%.met	1	1	1	1	0
		CCV_1102HP514r, CAL5 ;1102HP5 , 50000 ug per mL Diesel (200 uL of DRO211012A)	G:\Org\HP5\Methods\DC_8015-IA-L%.met	1	1	1	1	0
		DCM-Baseline Check-V15	G:\Org\HP5\Methods\DR_8015-HP-LEXP.met	1	1	1	1	0
		CCV_1102HP516r, Second Source ;1102HP5 , 15000 ug per mL (100uL of DRO211012B + 900uL DCM(14408)	G:\Org\HP5\Methods\DC_8015-IA-L%.met	1	1	1	1	0

File Name: G:\Org\HP5\Cals\SW8015C_DRO211102IA.CAL

Version: 14

Creator: AMN 11/02/2021

Description: 8015C-DRO. New ICal Per 1102HP5 (2021)-2 uL Inj.; COD added using OTP RFs

Reason for change:

External standard calibration

Standard injection volume: 1

Standard sample weight: 1

Area reject threshold: 500

Reference peak area reject threshold: 500

Amount units: nanograms


No default component

Method of calculating data point averages: Equal weight for all updates

No calibration update report

All levels are normal data points.


1 DRO Range Start



Expected retention time: 6.64 minutes
 Search window: 0.05 minutes
 No retention time reference component
 Group number: 0
 High alarm limit: 0
 Low alarm limit: 0
 Component constant: 0
 Single peak quantification by area
 $Y = 31353.19 X + 0$
 Average CF fit with equal weighting, forced to origin
 Coefficient of determination: 0.9992341
 Average error: 2.083%
 Average CF: 31353.19
 RSD: 2.487%

Level	Amount	Response	Cal Factor	Error, %	Source	Date and time
1	150	4865557	32437.05	3.457	Manual	11/3/2021 6:44:55 AM
2	3750	1.159464E+08	30919.04	-1.385	Manual	11/3/2021 6:45:31 AM
3	15000	4.785279E+08	31901.86	1.750	Manual	11/3/2021 6:46:08 AM
4	37500	1.14732E+09	30595.2	-2.418	Manual	11/3/2021 6:45:45 AM
5	50000	1.54564E+09	30912.8	-1.405	Manual	11/3/2021 6:45:18 AM


2 *o-Terphenyl



Expected retention time: 12.29 minutes
 Search window: 0.05 minutes
 No retention time reference component
 Group number: 0
 High alarm limit: 0
 Low alarm limit: 0
 Component constant: 0
 Single peak quantification by area
 Y = 35509.21 X + 0
 Average CF fit with equal weighting, forced to origin
 Coefficient of determination: 0.9983284
 Average error: 2.203%
 Average CF: 35509.21
 RSD: 2.749%

Level	Amount	Response	Cal Factor	Error, %	Source	Date and time
1	2	69364.34	34682.17	-2.329	G:\Org\HP5\DAT\HP5110221_b\1102HP5.0005.BND	11/2/2021 2:21:29 PM
2	50	1772689	35453.78	-0.156	G:\Org\HP5\DAT\HP5110221_b\1102HP5.0006.BND	11/2/2021 2:21:35 PM
3	200	6887244	34436.22	-3.022	G:\Org\HP5\DAT\HP5110221_b\1102HP5.0007.BND	11/2/2021 2:21:41 PM
4	500	1.815698E+07	36313.96	2.266	G:\Org\HP5\DAT\HP5110221_b\1102HP5.0008.BND	11/2/2021 2:21:47 PM
5	1000	3.665993E+07	36659.93	3.241	G:\Org\HP5\DAT\HP5110221_b\1102HP5.0009.BND	11/2/2021 2:21:52 PM

3 *1-Chlorooctadecane




Expected retention time: 13.1 minutes
 Search window: 0.05 minutes
 No retention time reference component
 Group number: 0
 High alarm limit: 0
 Low alarm limit: 0
 Component constant: 0
 Single peak quantification by area
 Y = 35509.21 X + 0
 Average CF fit with equal weighting, forced to origin
 Coefficient of determination: 0.9983284
 Average error: 2.203%
 Average CF: 35509.21
 RSD: 2.749%

Level	Amount	Response	Cal Factor	Error, %	Source	Date and time
1	2	69364.34	34682.17	-2.329	Manual	11/2/2021 2:21:57 PM
2	50	1772689	35453.78	-0.156	Manual	11/2/2021 2:21:58 PM
3	200	6887244	34436.22	-3.022	Manual	11/2/2021 2:22:00 PM
4	500	1.815698E+07	36313.96	2.266	Manual	11/2/2021 2:22:02 PM
5	1000	3.665993E+07	36659.93	3.241	Manual	11/2/2021 2:22:04 PM

G:\org\HP5\DAT\HP5110221_b\1102HP5.0003.RAW

CCV_1102HP508r, DRO ;1102HP5 , DRO211025A



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: CCV_1102HP508r, DRO ;1102HP5 , DRO211025A
 Raw File: G:\org\HP5\DAT\HP5110221_b\1102HP5.0003.RAW
 Date & Time Acquired: 11/2/2021 8:31:35 AM
 Method File: G:\Org\HP5\Methods\DC_8015-IA-L%.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO21102IA.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for TEH: 31353.19

Rt range for Diesel Range Organics: 6.59 to 15.91


SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*o-Terphenyl	12.292	200.	201.539	100.77
*1-Chlorooctadecane	13.095	200.	163.573	81.79

DRO Area: 9.38791E+07 DRO Amount: 2994.244
 TEH Area: 1.14933E+08 TEH Amount: 3665.75

CONTINUING CALIBRATION REPORT: G:\org\HP5\DAT\HP5110221_b\1102HP5.0003.RAW

COMPOUND	ACTUAL (NG)	MEASURED (NG)	%RECOVERY	LIMITS
TOTAL DRO	15000.	3665.75	24.44	85-115

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	LIMITS
*o-Terphenyl	12.292	200.	201.539	100.77	85-115
*1-Chlorooctadecane	13.095	200.	163.573	81.79	85-115



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: DCM-Baseline Check-V04
 Raw File: G:\org\HP5\DAT\HP5110221_b\1102HP5.0004.RAW
 Date & Time Acquired: 11/2/2021 9:14:27 AM
 Method File: G:\Org\HP5\Methods\DR_8015-HP-LEXP.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO210108HP.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1


Mean RF for TEH: 29457.33
 Rt range for Diesel Range Organics: 7.125 to 16.28

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*o-Terphenyl	12.746	200.	.026	.01
*1-Chlorooctadecane	29.929	200.	.	.

DRO Area: 35138 DRO Amount: 1.192844
 TEH Area: 110269.4 TEH Amount: 3.74336

G:\org\HP5\DAT\HP5110221_b\1102HP5.0005.RAW

CCV_1102HP505r, CAL1 ;1102HP5 , 2 ug per mL OTP



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: CCV_1102HP505r, CAL1 ;1102HP5 , 2 ug per mL OTP
 Raw File: G:\org\HP5\DAT\HP5110221_b\1102HP5.0005.RAW
 Date & Time Acquired: 11/2/2021 9:57:01 AM
 Method File: G:\Org\HP5\Methods\DS_8015-IA-L#.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO211102IA.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for TEH: 31353.19

Rt range for Diesel Range Organics: 6.59 to 15.91

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*o-Terphenyl	12.299	200.	1.953	.98	-
*1-Chlorooctadecane	15.561	200.	.	.	-


DRO Area:37026.36 DRO Amount: 1.180944
 TEH Area:88429.33 TEH Amount: 2.820425

CONTINUING CALIBRATION REPORT: G:\org\HP5\DAT\HP5110221_b\1102HP5.0005.RAW
 COMPOUND ACTUAL (NG) MEASURED (NG) %RECOVERY LIMITS
 TOTAL DRO 15000. . . 85-115

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	LIMITS
*o-Terphenyl	12.299	200.	1.953	.98	85-115
*1-Chlorooctadecane	15.561	200.	.	.	85-115

G:\org\HP5\DAT\HP5110221_b\1102HP5.0006.RAW

CCV_1102HP506r, CAL2 ;1102HP5 , 50 ug per mL OTP



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: CCV_1102HP506r, CAL2 ;1102HP5 , 50 ug per mL OTP
 Raw File: G:\org\HP5\DAT\HP5110221_b\1102HP5.0006.RAW
 Date & Time Acquired: 11/2/2021 10:39:43 AM
 Method File: G:\Org\HP5\Methods\DS_8015-IA-L#.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO211102IA.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for TEH: 31353.19

Rt range for Diesel Range Organics: 6.59 to 15.91

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*o-Terphenyl	12.29	200.	49.922	24.96	-
*1-Chlorooctadecane	15.697	200.	.	.	-

DRO Area:141449.5 DRO Amount: 4.511485
 TEH Area:310030.7 TEH Amount: 9.88833


CONTINUING CALIBRATION REPORT: G:\org\HP5\DAT\HP5110221_b\1102HP5.0006.RAW

COMPOUND	ACTUAL (NG)	MEASURED (NG)	%RECOVERY	LIMITS
TOTAL DRO	15000.	.	.	85-115

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	LIMITS
*o-Terphenyl	12.29	200.	49.922	24.96	85-115
*1-Chlorooctadecane	15.697	200.	.	.	85-115

G:\org\HP5\DAT\HP5110221_b\1102HP5.0007.RAW

CCV_1102HP507r, CAL3 ;1102HP5 , 200 ug per mL OTP



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: CCV_1102HP507r, CAL3 ;1102HP5 , 200 ug per mL OTP
 Raw File: G:\org\HP5\DAT\HP5110221_b\1102HP5.0007.RAW
 Date & Time Acquired: 11/2/2021 11:22:37 AM
 Method File: G:\Org\HP5\Methods\DS_8015-IA-L#.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO211102IA.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for TEH: 31353.19

Rt range for Diesel Range Organics: 6.59 to 15.91


SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*o-Terphenyl	12.29	200.	193.957	96.98
*1-Chlorooctadecane	15.779	200.	.	-

DRO Area:338082.7 DRO Amount: 10.78304
 TEH Area:638415.8 TEH Amount: 20.36207

CONTINUING CALIBRATION REPORT: G:\org\HP5\DAT\HP5110221_b\1102HP5.0007.RAW

COMPOUND	ACTUAL (NG)	MEASURED (NG)	%RECOVERY	LIMITS
TOTAL DRO	15000.	20.36	.14	85-115

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	LIMITS
*o-Terphenyl	12.29	200.	193.957	96.98	85-115
*1-Chlorooctadecane	15.779	200.	.	.	85-115



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: CCV_1102HP508r, CAL4 ;1102HP5 , 500 ug per mL OTP
 Raw File: G:\org\HP5\DAT\HP5110221_b\1102HP5.0008.RAW
 Date & Time Acquired: 11/2/2021 12:05:44 PM
 Method File: G:\Org\HP5\Methods\DS_8015-IA-L#.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO211102IA.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for TEH: 31353.19

Rt range for Diesel Range Organics: 6.59 to 15.91

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*o-Terphenyl	12.3	200.	511.332	255.67	-
*1-Chlorooctadecane	15.798	200.	.	.	-

DRO Area:456992 DRO Amount: 14.57561
 TEH Area:825752.1 TEH Amount: 26.3371


CONTINUING CALIBRATION REPORT: G:\org\HP5\DAT\HP5110221_b\1102HP5.0008.RAW

COMPOUND	ACTUAL (NG)	MEASURED (NG)	%RECOVERY	LIMITS
TOTAL DRO	15000.	26.34	.18	85-115

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	LIMITS
*o-Terphenyl	12.3	200.	511.332	255.67	85-115
*1-Chlorooctadecane	15.798	200.	.	.	85-115

G:\org\HP5\DAT\HP5110221_b\1102HP5.0009.RAW

CCV_1102HP509r, CAL5 ;1102HP5 , 1000 ug per mL OTP



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: CCV_1102HP509r, CAL5 ;1102HP5 , 1000 ug per mL OTP
 Raw File: G:\org\HP5\DAT\HP5110221_b\1102HP5.0009.RAW
 Date & Time Acquired: 11/2/2021 12:49:02 PM
 Method File: G:\Org\HP5\Methods\DS_8015-IA-L#.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO211102IA.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for TEH: 31353.19

Rt range for Diesel Range Organics: 6.59 to 15.91

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*o-Terphenyl	12.31	200.	1032.406	516.2	-
*1-Chlorooctadecane	15.803	200.	.	.	-

DRO Area:461032.4 DRO Amount: 14.70448
 TEH Area:724020.4 TEH Amount: 23.0924


CONTINUING CALIBRATION REPORT: G:\org\HP5\DAT\HP5110221_b\1102HP5.0009.RAW

COMPOUND	ACTUAL (NG)	MEASURED (NG)	%RECOVERY	LIMITS
TOTAL DRO	15000.	23.09	.15	85-115

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	LIMITS
*o-Terphenyl	12.31	200.	1032.406	516.2	85-115
*1-Chlorooctadecane	15.803	200.	.	.	85-115

G:\org\HP5\DAT\HP5110221_b\1102HP5.0010.RAW

CCV_1102HP510r, CAL1 ;1102HP5 , 150 ug per mL Diesel



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: CCV_1102HP510r, CAL1 ;1102HP5 , 150 ug per mL Diesel
 Raw File: G:\org\HP5\DAT\HP5110221_b\1102HP5.0010.RAW
 Date & Time Acquired: 11/2/2021 1:32:06 PM
 Method File: G:\Org\HP5\Methods\DC_8015-IA-L%.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO211102IA.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for TEH: 31353.19

Rt range for Diesel Range Organics: 6.59 to 15.91

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*o-Terphenyl	12.285	200.	1.416	.71	-
*1-Chlorooctadecane	13.13	200.	1.44	.72	-

DRO Area:4571415 DRO Amount: 145.8038
 TEH Area:4865557 TEH Amount: 155.1854


CONTINUING CALIBRATION REPORT: G:\org\HP5\DAT\HP5110221_b\1102HP5.0010.RAW

COMPOUND	ACTUAL (NG)	MEASURED (NG)	%RECOVERY	LIMITS
TOTAL DRO	15000.	155.19	1.03	85-115

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	LIMITS
*o-Terphenyl	12.285	200.	1.416	.71	85-115
*1-Chlorooctadecane	13.13	200.	1.44	.72	85-115

G:\org\HP5\DAT\HP5110221_b\1102HP5.0011.RAW

CCV_1102HP511r, CAL2 ;1102HP5 , 3750 ug per mL Diesel



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: CCV_1102HP511r, CAL2 ;1102HP5 , 3750 ug per mL Diesel
 Raw File: G:\org\HP5\DAT\HP5110221_b\1102HP5.0011.RAW
 Date & Time Acquired: 11/2/2021 2:15:08 PM
 Method File: G:\Org\HP5\Methods\DC_8015-IA-L%.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO211102IA.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for TEH: 31353.19

Rt range for Diesel Range Organics: 6.59 to 15.91

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*o-Terphenyl	12.252	200.	28.988	14.49	-
*1-Chlorooctadecane	13.122	200.	39.19	19.59	-

DRO Area: 1.131291E+08 DRO Amount: 3608.216
 TEH Area: 1.159464E+08 TEH Amount: 3698.073


CONTINUING CALIBRATION REPORT: G:\org\HP5\DAT\HP5110221_b\1102HP5.0011.RAW

COMPOUND	ACTUAL (NG)	MEASURED (NG)	%RECOVERY	LIMITS
TOTAL DRO	15000.	3698.07	24.65	85-115

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	LIMITS
*o-Terphenyl	12.252	200.	28.988	14.49	85-115
*1-Chlorooctadecane	13.122	200.	39.19	19.59	85-115

G:\org\HP5\DAT\HP5110221_b\1102HP5.0012.RAW

CCV_1102HP512r, CAL3 ;1102HP5 , 15000 ug per mL Diesel



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: CCV_1102HP512r, CAL3 ;1102HP5 , 15000 ug per mL Diesel
 Raw File: G:\org\HP5\DAT\HP5110221_b\1102HP5.0012.RAW
 Date & Time Acquired: 11/2/2021 2:58:26 PM
 Method File: G:\Org\HP5\Methods\DC_8015-IA-L%.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO211102IA.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for TEH: 31353.19

Rt range for Diesel Range Organics: 6.59 to 15.91

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*o-Terphenyl	12.257	200.	92.35	46.18
*1-Chlorooctadecane	13.125	200.	158.994	79.5

DRO Area: 4.667999E+08 DRO Amount: 14888.43
 TEH Area: 4.785279E+08 TEH Amount: 15262.49


CONTINUING CALIBRATION REPORT: G:\org\HP5\DAT\HP5110221_b\1102HP5.0012.RAW

COMPOUND	ACTUAL (NG)	MEASURED (NG)	%RECOVERY	LIMITS
TOTAL DRO	15000.	15262.49	101.75	85-115

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	LIMITS
*o-Terphenyl	12.257	200.	92.35	46.18	85-115
*1-Chlorooctadecane	13.125	200.	158.994	79.5	85-115

G:\org\HP5\DAT\HP5110221_b\1102HP5.0013.RAW

CCV_1102HP513r, CAL4 ;1102HP5 , 37500ug per mL Diesel



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: CCV_1102HP513r, CAL4 ;1102HP5 , 37500ug per mL Diesel
 Raw File: G:\org\HP5\DAT\HP5110221_b\1102HP5.0013.RAW
 Date & Time Acquired: 11/2/2021 3:41:37 PM
 Method File: G:\Org\HP5\Methods\DC_8015-IA-L%.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO211102IA.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for TEH: 31353.19

Rt range for Diesel Range Organics: 6.59 to 15.91

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*o-Terphenyl	12.268	200.	238.956	119.48
*1-Chlorooctadecane	13.133	200.	386.008	193.

DRO Area:1.118993E+09 DRO Amount: 35689.91
 TEH Area:1.14732E+09 TEH Amount: 36593.41


CONTINUING CALIBRATION REPORT: G:\org\HP5\DAT\HP5110221_b\1102HP5.0013.RAW

COMPOUND	ACTUAL (NG)	MEASURED (NG)	%RECOVERY	LIMITS
TOTAL DRO	15000.	36593.41	243.96	85-115

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	LIMITS
*o-Terphenyl	12.268	200.	238.956	119.48	85-115
*1-Chlorooctadecane	13.133	200.	386.008	193.	85-115

G:\org\HP5\DAT\HP5110221_b\1102HP5.0014.RAW

CCV_1102HP514r, CAL5 ;1102HP5 , 50000 ug per mL Diesel



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: CCV_1102HP514r, CAL5 ;1102HP5 , 50000 ug per mL Diesel
 Raw File: G:\org\HP5\DAT\HP5110221_b\1102HP5.0014.RAW
 Date & Time Acquired: 11/2/2021 4:24:53 PM
 Method File: G:\Org\HP5\Methods\DC_8015-IA-L%.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO211102IA.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for TEH: 31353.19

Rt range for Diesel Range Organics: 6.59 to 15.91


SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*o-Terphenyl	29.921	200.	.	-
*1-Chlorooctadecane	13.135	200.	512.63	256.31 -

DRO Area: 1.507978E+09 DRO Amount: 48096.49
 TEH Area: 1.54564E+09 TEH Amount: 49297.7

CONTINUING CALIBRATION REPORT: G:\org\HP5\DAT\HP5110221_b\1102HP5.0014.RAW

COMPOUND	ACTUAL (NG)	MEASURED (NG)	%RECOVERY	LIMITS
TOTAL DRO	15000.	49297.7	328.65	85-115

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	LIMITS
*o-Terphenyl	29.921	200.	.	.	85-115
*1-Chlorooctadecane	13.135	200.	512.63	256.31	85-115




DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: DCM-Baseline Check-V15
 Raw File: G:\Org\HP5\DAT\HP5110221_b\1102HP5.0015.RAW
 Date & Time Acquired: 11/2/2021 5:08:11 PM
 Method File: G:\Org\HP5\Methods\DR_8015-HP-LEXP.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO210108HP.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for TEH: 29457.33
 Rt range for Diesel Range Organics: 7.125 to 16.28

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*o-Terphenyl	12.744	200.	.041	.02	-
*1-Chlorooctadecane	13.606	200.	.017	.01	-

DRO Area:305831.5 DRO Amount: 10.38219
 TEH Area:517467.1 TEH Amount: 17.56667



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: CCV_1102HP516r, Second Source ;1102HP5 , 15000 ug per mL
 Raw File: G:\org\HP5\DAT\HP5110221_b\1102HP5.0016.RAW
 Date & Time Acquired: 11/2/2021 5:51:31 PM
 Method File: G:\Org\HP5\Methods\DC_8015-IA-L%.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO211102IA.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for TEH: 31353.19

Rt range for Diesel Range Organics: 6.59 to 15.91

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*o-Terphenyl	12.261	200.	146.398	73.2
*1-Chlorooctadecane	13.127	200.	209.876	104.94

DRO Area: 4.432555E+08 DRO Amount: 14137.49
 TEH Area: 4.699825E+08 TEH Amount: 14989.94

CONTINUING CALIBRATION REPORT: G:\org\HP5\DAT\HP5110221_b\1102HP5.0016.RAW

COMPOUND	ACTUAL (NG)	MEASURED (NG)	%RECOVERY	LIMITS
TOTAL DRO	15000.	14989.94	99.93	85-115

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	LIMITS
*o-Terphenyl	12.261	200.	146.398	73.2	85-115
*1-Chlorooctadecane	13.127	200.	209.876	104.94	85-115

Write Sequence	Data File	Sample Name	Method	Weight	Dil Factor	Amt Inj.	IS	Cal ID	Manul Integrations
	G:\org\HP5\DAT\HP5110221_b1102HP5.03r	CCV_1102HP508r, DRO ;1102HP5 ,DRO211025A	G:\Org\HP5\Methods\DC_8015-IA-L%.met	1	1	1	1	0	No integrations
	G:\org\HP5\DAT\HP5110221_b1102HP5.04r	DCM-Baseline Check-V04	G:\Org\HP5\Methods\DR_8015-HP-LEXP.met	1	1	1	1	0	No integrations
	G:\org\HP5\DAT\HP5110221_b1102HP5.05r	CCV_1102HP505r, CAL1 ;1102HP5 , 2 ug per mL OTP (10 uL of Cal3 + 990 uL DCM(14408))	G:\Org\HP5\Methods\DS_8015-IA-L%.met	1	1	1	1	0	Surrogates are integrated using a valley to valley integration using Set Baseline Now placed slight before at 12.21 and slightly after the surrogate peak at 12.63 and scaling showing surrogate peak from 11.5-14.
	G:\org\HP5\DAT\HP5110221_b1102HP5.06r	CCV_1102HP506r, CAL2 ;1102HP5 , 50 ug per mL OTP (100 uL Cal4 + 900 uL of DCM(14408))	G:\Org\HP5\Methods\DS_8015-IA-L%.met	1	1	1	1	0	Surrogates are integrated using a valley to valley integration using Set Baseline Now placed slight before at 12.21 and slightly after the surrogate peak at 12.63 and scaling showing surrogate peak from 11.5-14.
	G:\org\HP5\DAT\HP5110221_b1102HP5.07r	CCV_1102HP507r, CAL3 ;1102HP5 , 200 ug per mL OTP (100uL of Cal5 + 400 uL DCM(14408))	G:\Org\HP5\Methods\DS_8015-IA-L%.met	1	1	1	1	0	Surrogates are integrated using a valley to valley integration using Set Baseline Now placed slight before at 12.21 and slightly after the surrogate peak at 12.63 and scaling showing surrogate peak from 11.5-14.
	G:\org\HP5\DAT\HP5110221_b1102HP5.08r	CCV_1102HP508r, CAL4 ;1102HP5 , 500 ug per mL OTP (250uL of Cal5 + 250 uL DCM(14408))	G:\Org\HP5\Methods\DS_8015-IA-L%.met	1	1	1	1	0	Surrogates are integrated using a valley to valley integration using Set Baseline Now placed slight before at 12.21 and slightly after the surrogate peak at 12.63 and scaling showing surrogate peak from 11.5-14.
	G:\org\HP5\DAT\HP5110221_b1102HP5.09r	CCV_1102HP509r, CAL5 ;1102HP5 , 1000 ug per mL OTP (250 uL 4000 ug/mL OTP DRO211011A + 750 DCM(14408))	G:\Org\HP5\Methods\DS_8015-IA-L%.met	1	1	1	1	0	Surrogates are integrated using a valley to valley integration using Set Baseline Now placed slight before at 12.21 and slightly after the surrogate peak at 12.63 and scaling showing surrogate peak from 11.5-14.
	G:\org\HP5\DAT\HP5110221_b1102HP5.10r	CCV_1102HP510r, CAL1 ;1102HP5 , 150 ug per mL Diesel (10 uL of Cal3 + 990 uL DCM(14408)),	G:\Org\HP5\Methods\DC_8015-IA-L%.met	1	1	1	1	0	The integration of Diesel Range Organics and Total Extractable Hydrocarbons is the hydrocarbon response with reference to the baseline. Assigned Set Baseline All Valley on at 16.83
	G:\org\HP5\DAT\HP5110221_b1102HP5.11r	CCV_1102HP511r, CAL2 ;1102HP5 , 3750 ug per mL Diesel (100 uL Cal4 + 900 uL of DCM(14408))	G:\Org\HP5\Methods\DC_8015-IA-L%.met	1	1	1	1	0	The integration of Diesel Range Organics and Total Extractable Hydrocarbons is the hydrocarbon response with reference to the baseline. Assigned Set Baseline All Valley on at 17.5
	G:\org\HP5\DAT\HP5110221_b1102HP5.12r	CCV_1102HP512r, CAL3 ;1102HP5 , 15000 ug per mL Diesel (300 uL of DRO211012A + 700 uL DCM(14408))	G:\Org\HP5\Methods\DC_8015-IA-L%.met	1	1	1	1	0	The integration of Diesel Range Organics and Total Extractable Hydrocarbons is the hydrocarbon response with reference to the baseline. Assigned Set Baseline All Valley on at 17.5
	G:\org\HP5\DAT\HP5110221_b1102HP5.13r	CCV_1102HP513r, CAL4 ;1102HP5 , 37500ug per mL Diesel (750 uL of DRO211012A + 250 uL DCM(14408))	G:\Org\HP5\Methods\DC_8015-IA-L%.met	1	1	1	1	0	The integration of Diesel Range Organics and Total Extractable Hydrocarbons is the hydrocarbon response with reference to the baseline. Assigned Set Baseline All Valley on at 17.5
	G:\org\HP5\DAT\HP5110221_b1102HP5.14r	CCV_1102HP514r, CAL5 ;1102HP5 , 50000 ug per mL Diesel (200 uL of DRO211012A)	G:\Org\HP5\Methods\DC_8015-IA-L%.met	1	1	1	1	0	The integration of Diesel Range Organics and Total Extractable Hydrocarbons is the hydrocarbon response with reference to the baseline. Assigned Set Baseline All Valley on at 17.5
	G:\org\HP5\DAT\HP5110221_b1102HP5.15r	DCM-Baseline Check-V15	G:\Org\HP5\Methods\DR_8015-HP-LEXP.met	1	1	1	1	0	No integrations
	G:\org\HP5\DAT\HP5110221_b1102HP5.16r	CCV_1102HP516r, Second Source ;1102HP5 , 15000 ug per mL (100uL of DRO211012B + 900uL DCM(14408))	G:\Org\HP5\Methods\DC_8015-IA-L%.met	1	1	1	1	0	The integration of Diesel Range Organics and Total Extractable Hydrocarbons is the hydrocarbon response with reference to the baseline. Assigned Set Baseline All Valley on at 17.5



Digitally signed by
Ann Nebel
Date: 2021.11.09 12:32:44 -07:00

Energy Laboratories Inc

ANALYTICAL RUN Summary

31-Mar-21

Run ID GCFID-HP5-B_210218B

Run Start Date: 2/18/2021
Analyst: Ann Nebel
Ical:
Column ID:
Comments: 8015 OIL range calibration SW8015_OIL210218

Std ID	Std Name	Std Amount	Std Units	Samp Amount	Samp Units	SampType	Expiration Date
DRO180918C	50,000 ug/mL Oil Std For AK103 RRO-In DCM					CAL	8/31/2025
DRO210204A	Carbon Scan STD					MARKER	3/5/2028
DRO210217A	20,000 ug/mL Oil Std For AK103 RRO-In DCM					ICV	8/23/2021

Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist				
14282665	CCV_0218HP50	HC-8015-DRO-	CAL1		2/18/2021 12:03:	1	R356533		0	0					
Analyte		T Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD Q
TEH(Oil Range)		A mg/L		0.1468323		0.15	0	0	0	0.3	0	98%	80	120	0%

Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist				
14282666	CCV_0218HP50	HC-8015-DRO-	CAL2		2/18/2021 1:27:3	1	R356533		0	0					
Analyte		T Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD Q
TEH(Oil Range)		A mg/L		1.062811		1	0	0	0.15	0.3	0	106%	80	120	0%

Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist				
14282667	CCV_0218HP50	HC-8015-DRO-	CAL3		2/18/2021 2:51:0	1	R356533		0	0					
Analyte		T Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD Q
TEH(Oil Range)		A mg/L		5.035713		5	0	0	0.15	0.3	0	101%	80	120	0%

Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist					
14282668	CCV_0218HP50	HC-8015-DRO-	CAL4		2/18/2021 4:14:3	1	R356533		0	0						
Analyte	T	Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD	Q
TEH(Oil Range)	A	mg/L		14.69295		15	0	0	0.15	0.3	0	98%	80	120	0%	
Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist					
14282669	CCV_0218HP51	HC-8015-DRO-	CAL5		2/18/2021 5:38:3	1	R356533		0	0						
Analyte	T	Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD	Q
TEH(Oil Range)	A	mg/L		28.20769		30	0	0	0.15	0.3	0	94%	80	120	0%	
Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist					
14282670	CCV_0218HP51	HC-8015-DRO-	ICV		2/18/2021 8:27:3	1	R356533		0	0						
Analyte	T	Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD	Q
TEH(Oil Range)	A	mg/L		5.474465		5	0	0	0.15	0.3	0	109%	80	120	0%	

Write Sequence	Data File	Sample Name	Method	Weight	Dil Factor	Amt Inj.	IS	Cal ID
		Marker_0218HP501r, DRO C40 ;0218HP5 , DRO210204A	G:\Org\HP5\Methods\CSC210212.met	1	1	1	1	0
		DCM-Baseline Check-V02	G:\Org\HP5\Methods\DR_8015-HE-LEXP.met	1	1	1	1	0
		CCV_0218HP503r, CAL1 ;0218HP5 , 150 ug per mL Oil (10 uL of Cal4 + 990 uL DCM(13510)	G:\Org\HP5\Methods\DR_OIL-021803-AA-L0.MET	1	1	1	1	0
		DCM-Baseline Check-V04	G:\Org\HP5\Methods\DR_8015-HE-LEXP.met	1	1	1	1	0
		CCV_0218HP505r, CAL2 ;0218HP5 , 1000 ug per mL Oil (200 uL of Cal 3 +800 uL DCM(13510)	G:\Org\HP5\Methods\DR_OIL-021805-AA-L0.MET	1	1	1	1	0
		DCM-Baseline Check-V06	G:\Org\HP5\Methods\DR_8015-HE-LEXP.met	1	1	1	1	0
		CCV_0218HP507r, CAL3 ;0218HP5 , 5000 ug per mL Oil (100 uL of DRO180918C + 900 uL DCM(13510)	G:\Org\HP5\Methods\DR_OIL-021807-AA-L0.MET	1	1	1	1	0
		DCM-Baseline Check-V08	G:\Org\HP5\Methods\DR_8015-HE-LEXP.met	1	1	1	1	0
		CCV_0218HP509r, CAL4 ;0218HP5 , 15000 ug per mL Oil (200 uL of CAL5 + 200 uL DCM(13510)	G:\Org\HP5\Methods\DR_OIL-021807-AA-L0.MET	1	1	1	1	0
		DCM-Baseline Check-V10	G:\Org\HP5\Methods\DR_8015-HE-LEXP.met	1	1	1	1	0
		CCV_0218HP511r, CAL5 ;0218HP5 , 30000 ug per mL Oil (600 uL of DRO180918C + 400 uL of DCM)	G:\Org\HP5\Methods\DR_OIL-021811-AA-L0.MET	1	1	1	1	0
		DCM-Baseline Check-V12	G:\Org\HP5\Methods\DR_8015-HE-LEXP.met	1	1	1	1	0
		DCM-Baseline Check-V13	G:\Org\HP5\Methods\DR_8015-HE-LEXP.met	1	1	1	1	0
		DCM-Baseline Check-V14	G:\Org\HP5\Methods\DR_8015-HE-LEXP.met	1	1	1	1	0
		CCV_0218HP515r, Second Source ;0218HP5 , 5000 ug per mL (100uL of DRO210217A + 300uL DCM(13510)	G:\Org\HP5\Methods\DR_OIL-021811-AA-L0.MET	1	1	1	1	0

File Name: G:\Org\HP5\Cals\SW8015C_OIL210218AA.CAL

Version: 4

Creator: AMN 3/31/2021

Description: DRO-8015-Oil range. New ICal Per 0218HP5 (2021)-2 uL Inj.; COD added using OTP RFs

Reason for change:

External standard calibration

Standard injection volume: 1

Standard sample weight: 1

Area reject threshold: 500

Reference peak area reject threshold: 500

Amount units: nanograms


No default component

Method of calculating data point averages: Equal weight for all updates

No calibration update report

All levels are normal data points.

1 DRO Range Start



Expected retention time: 6.54 minutes
 Search window: 0.05 minutes
 No retention time reference component
 Group number: 0
 High alarm limit: 0
 Low alarm limit: 0
 Component constant: 0


Single peak quantification by area

$Y = 28542.41 X + 0$

Average CF fit with equal weighting, forced to origin
 Coefficient of determination: 0.9940317
 Average error: 3.209%
 Average CF: 28542.41
 RSD: 4.497%

Level	Amount	Response	Cal Factor	Error, %	Source	Date and time
1	150	4325287	28835.25	1.026	Manual	3/30/2021 11:50:57 AM
2	1000	3.03352E+07	30335.2	6.281	Manual	3/30/2021 11:51:41 AM
3	5000	1.437314E+08	28746.28	0.714	Manual	3/30/2021 11:52:00 AM
4	15000	4.193721E+08	27958.14	-2.047	Manual	3/30/2021 11:52:55 AM
5	30000	8.051155E+08	26837.18	-5.974	Manual	3/30/2021 11:52:32 AM


2 *o-Terphenyl



Expected retention time: 12.14 minutes
 Search window: 0.05 minutes
 No retention time reference component
 Group number: 0
 High alarm limit: 0
 Low alarm limit: 0
 Component constant: 0
 Single peak quantification by area
 $Y = 35071.26 X + 0$
 Average CF fit with equal weighting, forced to origin
 Coefficient of determination: 0.9996535
 Average error: 0.838%
 Average CF: 35071.26
 RSD: 1.097%


Level	Amount	Response	Cal Factor	Error, %	Source	Date and time
1	2	70648.41	35324.2	0.721	G:\Org\HP5\DAT\HP5010821_b\0108HP5.0011.BND	1/11/2021 9:44:43 AM
2	50	1746406	34928.12	-0.408	G:\Org\HP5\DAT\HP5010821_b\0108HP5.0012.BND	1/11/2021 9:44:50 AM
3	200	7110604	35553.02	1.374	G:\Org\HP5\DAT\HP5010821_b\0108HP5.0013.BND	1/11/2021 9:44:58 AM
4	500	1.749965E+07	34999.3	-0.205	G:\Org\HP5\DAT\HP5010821_b\0108HP5.0014.BND	1/11/2021 9:45:02 AM
5	1000	3.455164E+07	34551.64	-1.482	G:\Org\HP5\DAT\HP5010821_b\0108HP5.0015.BND	1/11/2021 9:45:07 AM


3 *1-Chlorooctadecane



Expected retention time: 12.95 minutes
 Search window: 0.05 minutes
 No retention time reference component
 Group number: 0
 High alarm limit: 0
 Low alarm limit: 0
 Component constant: 0
 Single peak quantification by area
 $Y = 35071.26 X + 0$
 Average CF fit with equal weighting, forced to origin
 Coefficient of determination: 0.9996535
 Average error: 0.838%
 Average CF: 35071.26
 RSD: 1.097%

Level	Amount	Response	Cal Factor	Error, %	Source	Date and time
1	2	70648.41	35324.2	0.721	Manual	1/11/2021 9:45:13 AM
2	50	1746406	34928.12	-0.408	Manual	1/11/2021 9:45:15 AM
3	200	7110604	35553.02	1.374	Manual	1/11/2021 9:45:17 AM
4	500	1.749965E+07	34999.3	-0.205	Manual	1/11/2021 9:45:19 AM
5	1000	3.455164E+07	34551.64	-1.482	Manual	1/11/2021 9:45:21 AM





DIESEL RANGE ORGANICS CHROMATOGRAM REPORT


Sample Name: DCM-Baseline Check-V02
 Raw File: G:\org\HP5\DAT\HP5021821_b\0218HP5.0002.RAW
 Date & Time Acquired: 2/18/2021 11:21:40 AM
 Method File: G:\Org\HP5\Methods\DR_8015-HE-LEXP.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO210108HE.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for TEH: 29457.33

Rt range for Diesel Range Organics: 6.49 to 15.75

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*o-Terphenyl	12.158	200.	.034	.02
*1-Chlorooctadecane	29.927	200.	.	.

DRO Area:29553.31 DRO Amount: 1.003258
 TEH Area:144057.3 TEH Amount: 4.890373



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: CCV_0218HP503r, CAL1 ;0218HP5 , 150 ug per mL Oil
 Raw File: G:\org\HP5\DAT\HP5021821_b\0218HP5.0003.RAW
 Date & Time Acquired: 2/18/2021 12:03:33 PM
 Method File: G:\ORG\HP5\METHODS\DR_OIL-021803-AA-L0.MET
 Calibration File: G:\Org\HP5\Cals\SW8015C_OIL210218AA.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for TEH: 28542.41

Rt range for Diesel Range Organics: 6.49 to 15.75


SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*o-Terphenyl	12.146	200.	.02	.01	-
*1-Chlorooctadecane	29.979	200.	.	.	-

DRO Area: 141843.8 DRO Amount: 4.969579
 TEH Area: 3766485 TEH Amount: 131.961

CONTINUING CALIBRATION REPORT: G:\org\HP5\DAT\HP5021821_b\0218HP5.0003.RAW

COMPOUND	ACTUAL (NG)	MEASURED (NG)	%RECOVERY	LIMITS
TOTAL DRO	5000.	131.96	2.64	85-115

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	LIMITS
*o-Terphenyl	12.146	200.	.02	.01	85-115
*1-Chlorooctadecane	29.979	200.	.	.	85-115



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT


Sample Name: DCM-Baseline Check-V04
 Raw File: G:\org\HP5\DAT\HP5021821_b\0218HP5.0004.RAW
 Date & Time Acquired: 2/18/2021 12:45:36 PM
 Method File: G:\Org\HP5\Methods\DR_8015-HE-LEXP.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO210108HE.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for TEH: 29457.33

Rt range for Diesel Range Organics: 6.49 to 15.75

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*o-Terphenyl	12.171	200.	.017	.01
*1-Chlorooctadecane	29.958	200.	.	.

DRO Area: 27245.38 DRO Amount: 0.9249101
 TEH Area: 63574.01 TEH Amount: 2.158173



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: CCV_0218HP505r, CAL2 ;0218HP5 , 1000 ug per mL Oil
 Raw File: G:\org\HP5\DAT\HP5021821_b\0218HP5.0005.RAW
 Date & Time Acquired: 2/18/2021 1:27:30 PM
 Method File: G:\ORG\HP5\METHODS\DR_OIL-021805-AA-L0.MET
 Calibration File: G:\Org\HP5\Cals\SW8015C_OIL210218AA.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for TEH: 28542.41

Rt range for Diesel Range Organics: 6.49 to 15.75


SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*o-Terphenyl	12.136	200.	.034	.02
*1-Chlorooctadecane	12.983	200.	.496	.25

DRO Area: 3721460 DRO Amount: 130.3835
 TEH Area: 3.03352E+07 TEH Amount: 1062.811

CONTINUING CALIBRATION REPORT: G:\org\HP5\DAT\HP5021821_b\0218HP5.0005.RAW

COMPOUND	ACTUAL (NG)	MEASURED (NG)	%RECOVERY	LIMITS
TOTAL DRO	5000.	1062.81	21.26	85-115

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	LIMITS
*o-Terphenyl	12.136	200.	.034	.02	85-115
*1-Chlorooctadecane	12.983	200.	.496	.25	85-115




DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: DCM-Baseline Check-V06
 Raw File: G:\org\HP5\DAT\HP5021821_b\0218HP5.0006.RAW
 Date & Time Acquired: 2/18/2021 2:09:12 PM
 Method File: G:\Org\HP5\Methods\DR_8015-HE-LEXP.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO210108HE.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for TEH: 29457.33
 Rt range for Diesel Range Organics: 6.49 to 15.75

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*o-Terphenyl	12.17	200.	.017	.01
*1-Chlorooctadecane	29.977	200.	.	.

DRO Area:29249.25 DRO Amount: 0.9929362
 TEH Area:123949.1 TEH Amount: 4.20775



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: CCV_0218HP507r, CAL3 ;0218HP5 , 5000 ug per mL Oil
 Raw File: G:\org\HP5\DAT\HP5021821_b\0218HP5.0007.RAW
 Date & Time Acquired: 2/18/2021 2:51:00 PM
 Method File: G:\ORG\HP5\METHODS\DR_OIL-021807-AA-L0.MET
 Calibration File: G:\Org\HP5\Cals\SW8015C_OIL210218AA.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for TEH: 28542.41

Rt range for Diesel Range Organics: 6.49 to 15.75


SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*o-Terphenyl	12.129	200.	.538	.27	-
*1-Chlorooctadecane	12.981	200.	4.244	2.12	-

DRO Area: 2.19239E+07 DRO Amount: 768.1166
 TEH Area: 1.437314E+08 TEH Amount: 5035.713

CONTINUING CALIBRATION REPORT: G:\org\HP5\DAT\HP5021821_b\0218HP5.0007.RAW

COMPOUND	ACTUAL (NG)	MEASURED (NG)	%RECOVERY	LIMITS
TOTAL DRO	5000.	5035.71	100.71	85-115

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	LIMITS
*o-Terphenyl	12.129	200.	.538	.27	85-115
*1-Chlorooctadecane	12.981	200.	4.244	2.12	85-115



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT


Sample Name: DCM-Baseline Check-V08
 Raw File: G:\org\HP5\DAT\HP5021821_b\0218HP5.0008.RAW
 Date & Time Acquired: 2/18/2021 3:32:46 PM
 Method File: G:\Org\HP5\Methods\DR_8015-HE-LEXP.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO210108HE.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for TEH: 29457.33

Rt range for Diesel Range Organics: 6.49 to 15.75

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*o-Terphenyl	12.176	200.	.026	.01
*1-Chlorooctadecane	29.982	200.	.	.

DRO Area:30717.07 DRO Amount: 1.042765
 TEH Area:223672.8 TEH Amount: 7.593112



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: CCV_0218HP509r, CAL4 ;0218HP5 , 15000 ug per mL Oil
 Raw File: G:\org\HP5\DAT\HP5021821_b\0218HP5.0009.RAW
 Date & Time Acquired: 2/18/2021 4:14:34 PM
 Method File: G:\ORG\HP5\METHODS\DR_OIL-021807-AA-L0.MET
 Calibration File: G:\Org\HP5\Cals\SW8015C_OIL210218AA.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for TEH: 28542.41

Rt range for Diesel Range Organics: 6.49 to 15.75


SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*o-Terphenyl	12.124	200.	1.888	.94	-
*1-Chlorooctadecane	12.984	200.	13.129	6.56	-

DRO Area: 6.73131E+07 DRO Amount: 2358.354
 TEH Area: 4.193721E+08 TEH Amount: 14692.95

CONTINUING CALIBRATION REPORT: G:\org\HP5\DAT\HP5021821_b\0218HP5.0009.RAW

COMPOUND	ACTUAL (NG)	MEASURED (NG)	%RECOVERY	LIMITS
TOTAL DRO	5000.	14692.95	293.86	85-115

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	LIMITS
*o-Terphenyl	12.124	200.	1.888	.94	85-115
*1-Chlorooctadecane	12.984	200.	13.129	6.56	85-115




DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: DCM-Baseline Check-V10
 Raw File: G:\org\HP5\DAT\HP5021821_b\0218HP5.0010.RAW
 Date & Time Acquired: 2/18/2021 4:56:16 PM
 Method File: G:\Org\HP5\Methods\DR_8015-HE-LEXP.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO210108HE.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for TEH: 29457.33
 Rt range for Diesel Range Organics: 6.49 to 15.75

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*o-Terphenyl	12.153	200.	.023	.01
*1-Chlorooctadecane	29.957	200.	.	.

DRO Area: 108629.1 DRO Amount: 3.687675
 TEH Area: 543425.7 TEH Amount: 18.44789



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: CCV_0218HP511r, CAL5 ;0218HP5 , 30000 ug per mL Oil
 Raw File: G:\org\HP5\DAT\HP5021821_b\0218HP5.0011.RAW
 Date & Time Acquired: 2/18/2021 5:38:33 PM
 Method File: G:\ORG\HP5\METHODS\DR_OIL-021811-AA-L0.MET
 Calibration File: G:\Org\HP5\Cals\SW8015C_OIL210218AA.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for TEH: 28542.41
 Rt range for Diesel Range Organics: 6.49 to 15.75


SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*o-Terphenyl	12.128	200.	3.828	1.91	-
*1-Chlorooctadecane	12.985	200.	28.22	14.11	-

DRO Area: 1.319327E+08 DRO Amount: 4622.338
 TEH Area: 8.051155E+08 TEH Amount: 28207.69

CONTINUING CALIBRATION REPORT: G:\org\HP5\DAT\HP5021821_b\0218HP5.0011.RAW

COMPOUND	ACTUAL (NG)	MEASURED (NG)	%RECOVERY	LIMITS
TOTAL DRO	5000.	28207.69	564.15	85-115

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	LIMITS
*o-Terphenyl	12.128	200.	3.828	1.91	85-115
*1-Chlorooctadecane	12.985	200.	28.22	14.11	85-115




DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: DCM-Baseline Check-V12
 Raw File: G:\org\HP5\DAT\HP5021821_b\0218HP5.0012.RAW
 Date & Time Acquired: 2/18/2021 6:20:48 PM
 Method File: G:\Org\HP5\Methods\DR_8015-HE-LEXP.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO210108HE.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for TEH: 29457.33
 Rt range for Diesel Range Organics: 6.49 to 15.75

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*o-Terphenyl	12.146	200.	.024	.01
*1-Chlorooctadecane	29.981	200.	.	.

DRO Area:186033.1 DRO Amount: 6.31534
 TEH Area:888262.2 TEH Amount: 30.1542




DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: DCM-Baseline Check-V13
 Raw File: G:\org\HP5\DAT\HP5021821_b\0218HP5.0013.RAW
 Date & Time Acquired: 2/18/2021 7:03:01 PM
 Method File: G:\Org\HP5\Methods\DR_8015-HE-LEXP.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO210108HE.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for TEH: 29457.33
 Rt range for Diesel Range Organics: 6.49 to 15.75

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*o-Terphenyl	12.152	200.	.033	.02
*1-Chlorooctadecane	29.967	200.	.	.

DRO Area: 40824.55 DRO Amount: 1.385888
 TEH Area: 476705.3 TEH Amount: 16.18291




DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: DCM-Baseline Check-V14
 Raw File: G:\org\HP5\DAT\HP5021821_b\0218HP5.0014.RAW
 Date & Time Acquired: 2/18/2021 7:45:21 PM
 Method File: G:\Org\HP5\Methods\DR_8015-HE-LEXP.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO210108HE.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for TEH: 29457.33
 Rt range for Diesel Range Organics: 6.49 to 15.75

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*o-Terphenyl	12.148	200.	.03	.01
*1-Chlorooctadecane	29.949	200.	.	.

DRO Area:30445.24 DRO Amount: 1.033537
 TEH Area:297634.4 TEH Amount: 10.10392



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: CCV_0218HP515r, Second Source ;0218HP5 , 5000 ug per mL
 Raw File: G:\Org\HP5\DAT\HP5021821_b\0218HP5.0015.RAW
 Date & Time Acquired: 2/18/2021 8:27:37 PM
 Method File: G:\ORG\HP5\METHODS\DR_OIL-021811-AA-L0.MET
 Calibration File: G:\Org\HP5\Cals\SW8015C_OIL210218AA.CAL
 Sample Weight: 1000 Dilution: 1 S.A.: 1

Mean RF for TEH: 28542.41
 Rt range for Diesel Range Organics: 6.49 to 15.75

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*o-Terphenyl	12.137	.2	.19	-
*1-Chlorooctadecane	12.984	.2	.003	1.62

DRO Area: 2.558549E+07 DRO Amount: 0.8964027
 TEH Area: 1.562544E+08 TEH Amount: 5.474465

CONTINUING CALIBRATION REPORT: G:\Org\HP5\DAT\HP5021821_b\0218HP5.0015.RAW

COMPOUND	ACTUAL (NG)	MEASURED (NG)	%RECOVERY	LIMITS
TOTAL DRO	5000.	5.47	.11	85-115

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	LIMITS
*o-Terphenyl	12.137	.2	.19	1.62	85-115
*1-Chlorooctadecane	12.984	.2	.003	1.62	85-115

Write Sequence	Data File	Sample Name	Method	Weight	Dil Factor	Amt Inj	IS	Cal ID	Manual Integration
		Marker_0218HP501r, DRO C40 ;0218HP5_ , DRO210204A	G:\Org\HP5\Methods\CSC210212.met	1	1	1	1	0	No Integration
		DCM-Baseline Check-V02	G:\Org\HP5\Methods\DR_8015-HE-LEXP.met	1	1	1	1	0	No Integration
		CCV_0218HP503r, CAL1 ;0218HP5 , 150 ug per mL Oil (10 uL of Cal4 + 990 uL DCM(13510))	G:\Org\HP5\Methods\DR_OIL-021803-AA-L0.MET						The integration of Total Extractable Hydrocarbons (OIL) is the hydrocarbon response with reference to the baseline. Assigned Set Baseline All Valley on at 20.91 changed peak width and scale was set at 150 for the Yaxis.
		DCM-Baseline Check-V04	G:\Org\HP5\Methods\DR_8015-HE-LEXP.met	1	1	1	1	0	No Integration
		CCV_0218HP505r, CAL2 ;0218HP5 , 1000 ug per mL Oil (200 uL of Cal 3 +800 uL DCM(13510))	G:\Org\HP5\Methods\DR_OIL-021805-AA-L0.MET						The integration of Total Extractable Hydrocarbons (OIL) is the hydrocarbon response with reference to the baseline. Assigned Set Baseline All Valley on at 26.79 changed peak width and scale was set at 300 for the Yaxis.
		DCM-Baseline Check-V06	G:\Org\HP5\Methods\DR_8015-HE-LEXP.met	1	1	1	1	0	No Integration
		CCV_0218HP507r, CAL3 ;0218HP5 , 5000 ug per mL Oil (100 uL of DRO180918C + 900 uL DCM(13510))	G:\Org\HP5\Methods\DR_OIL-021807-AA-L0.MET						The integration of Total Extractable Hydrocarbons (OIL) is the hydrocarbon response with reference to the baseline. Scale was set at 3000 for the Yaxis.
		DCM-Baseline Check-V08	G:\Org\HP5\Methods\DR_8015-HE-LEXP.met	1	1	1	1	0	No Integration
		CCV_0218HP509r, CAL4 ;0218HP5 , 15000 ug per mL Oil (200 uL of CAL5 + 200 uL DCM(13510))	G:\Org\HP5\Methods\DR_OIL-021807-AA-L0.MET						The integration of Total Extractable Hydrocarbons (OIL) is the hydrocarbon response with reference to the baseline. Scale was set at 3000 for the Yaxis.
		DCM-Baseline Check-V10	G:\Org\HP5\Methods\DR_8015-HE-LEXP.met	1	1	1	1	0	No Integration
		CCV_0218HP511r, CAL5 ;0218HP5 , 30000 ug per mL Oil (600 uL of DRO180918C + 400 uL of DCM)	G:\Org\HP5\Methods\DR_OIL-021811-AA-L0.MET						The integration of Total Extractable Hydrocarbons (OIL) is the hydrocarbon response with reference to the baseline. Scale was set at 5000 for the Yaxis.
		DCM-Baseline Check-V12	G:\Org\HP5\Methods\DR_8015-HE-LEXP.met	1	1	1	1	0	No Integration
		DCM-Baseline Check-V13	G:\Org\HP5\Methods\DR_8015-HE-LEXP.met	1	1	1	1	0	No Integration
		DCM-Baseline Check-V14	G:\Org\HP5\Methods\DR_8015-HE-LEXP.met	1	1	1	1	0	No Integration
		CCV_0218HP515r, Second Source ;0218HP5 , 5000 ug per mL (100uL of DRO210217A + 300uL DCM(13510))	G:\Org\HP5\Methods\DR_OIL-021811-AA-L0.MET						The integration of Total Extractable Hydrocarbons (OIL) is the hydrocarbon response with reference to the baseline. Scale was set at 5000 for the Yaxis.

Ann Nebel

Digitally signed by
Ann Nebel
Date: 2021.10.29 12:03:40 -06:00

Energy Laboratories Inc

ANALYTICAL RUN Summary

25-Oct-21

Run ID GCFID-HP5-B_211017A

Run Start Date: 10/17/2021
Analyst: Ann Nebel
Ical:
Column ID:
Comments: Triacontane ICAL

Std ID	Std Name	Std Amount	Std Units	Samp Amount	Samp Units	SampType	Expiration Date
DRO211006A	Triacontane SURR 2000 ug/mL					SURR	4/6/2026

Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist
--------	--------	-----------	------------	---------	---------------	----	----------	-----------	--------	--------	--------

14777567	CCV_1017HP50	HC-8015-DRO-	CAL1		10/17/2021 3:30:	1	R368813		0	0	
----------	--------------	--------------	------	--	------------------	---	---------	--	---	---	--

Analyte	T	Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD	Q
n-Triacontane	S	mg/L		0.00202757		0.002	0	0	0.002	0.002	0	101%	80	120	0%	

Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist
--------	--------	-----------	------------	---------	---------------	----	----------	-----------	--------	--------	--------

14777568	CCV_1017HP50	HC-8015-DRO-	CAL2		10/17/2021 4:12:	1	R368813		0	0	
----------	--------------	--------------	------	--	------------------	---	---------	--	---	---	--

Analyte	T	Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD	Q
n-Triacontane	S	mg/L		0.04817772		0.05	0	0	0.002	0.002	0	96%	80	120	0%	

Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist
--------	--------	-----------	------------	---------	---------------	----	----------	-----------	--------	--------	--------

14777569	CCV_1017HP50	HC-8015-DRO-	CAL3		10/17/2021 4:55:	1	R368813		0	0	
----------	--------------	--------------	------	--	------------------	---	---------	--	---	---	--

Analyte	T	Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD	Q
n-Triacontane	S	mg/L		0.2231112		0.2	0	0	0.002	0.002	0	112%	80	120	0%	

Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist
--------	--------	-----------	------------	---------	---------------	----	----------	-----------	--------	--------	--------

14777570	CCV_1017HP50	HC-8015-DRO-	CAL4		10/17/2021 5:38:	1	R368813		0	0	
----------	--------------	--------------	------	--	------------------	---	---------	--	---	---	--

Analyte	T	Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD	Q
n-Triacontane	S	mg/L		0.4700634		0.5	0	0	0.002	0.002	0	94%	80	120	0%	

Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist					
Analyte	T	Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD	Q
14777571	CCV_1017HP50	HC-8015-DRO-	CAL5		10/17/2021 6:20:	1	R368813		0	0						
n-Triacontane	S	mg/L		0.9372648		1	0	0	0.002	0.002	0	94%	80	120	0%	

Write Sequence	Data File	Sample Name	Method	Weight	Dil Factor	Amt Inj.	IS	Cal ID
	G:\org\HP5\DAT\HP5101721_b\1017HP5.01r	DCM-Baseline Check-V01	G:\Org\HP5\Methods\DR_8015-HS-LEXP.met	1	1	1	1	0
	G:\org\HP5\DAT\HP5101721_b\1017HP5.02r	DCM-Baseline Check-V02	G:\Org\HP5\Methods\DR_8015-HS-LEXP.met	1	1	1	1	0
	G:\org\HP5\DAT\HP5101721_b\1017HP5.03r	CCV_1017HP503r, DRO ;1017HP5 , DRO210708A	G:\Org\HP5\Methods\DR_8015-HS-LEXP.met	1	1	1	1	0
	G:\org\HP5\DAT\HP5101721_b\1017HP5.04r	DCM-Baseline Check-V04	G:\Org\HP5\Methods\DR_8015-HS-LEXP.met	1	1	1	1	0
	G:\org\HP5\DAT\HP5101721_b\1017HP5.05r	CCV_1017HP505r, CAL1 ;1017HP5 , 2 ug per mL Triacotane (10 uL of Cal3 + 990 uL DCM(14354)	G:\Org\HP5\Methods\DS_ORO-AA-L0.met	1	1	1	1	0
	G:\org\HP5\DAT\HP5101721_b\1017HP5.06r	CCV_1017HP506r, CAL2 ;1017HP5 , 50 ug per mL Triacotane (100 uL Cal4 + 900 uL of DCM(14354)	G:\Org\HP5\Methods\DS_ORO-AA-L0.met	1	1	1	1	0
	G:\org\HP5\DAT\HP5101721_b\1017HP5.07r	CCV_1017HP507r, CAL3 ;1017HP5 , 200 ug per mL Triacotane (100uL of Cal5 + 400 uL DCM(14354)	G:\Org\HP5\Methods\DS_ORO-AA-L0.met	1	1	1	1	0
	G:\org\HP5\DAT\HP5101721_b\1017HP5.08r	CCV_1017HP508r, CAL4 ;1017HP5 , 500 ug per mL Triacotane (250uL of Cal5 + 250 uL DCM(14354)	G:\Org\HP5\Methods\DS_ORO-AA-L0.met	1	1	1	1	0
	G:\org\HP5\DAT\HP5101721_b\1017HP5.09r	CCV_1017HP509r, CAL5 ;1017HP5 , 1000 ug per mL Triacotane (500 uL 2000 ug/mL Triacotane DRO211006A + 500 DCM(14354)	G:\Org\HP5\Methods\DS_ORO-AA-L0.met	1	1	1	1	0

File Name: G:\Org\HP5\Cals\SW8015C_ORO211017AA.CAL

Version: 11

Creator: AMN

Description: 8015C-Oil Range. New ICal Per 1017HP5 (2021)-2 uL Inj.; RRO copied from 8015 cal for Oil

Reason for change:

External standard calibration

Standard injection volume: 1

Standard sample weight: 1

Area reject threshold: 500

Reference peak area reject threshold: 500

Amount units: nanograms


No default component

Method of calculating data point averages: Equal weight for all updates

No calibration update report

All levels are normal data points.

1 *30-40 Motor Oil



Expected retention time: 6.4 minutes
 Search window: 0.05 minutes
 No retention time reference component
 Group number: 0
 High alarm limit: 0
 Low alarm limit: 0
 Component constant: 0

Single peak quantification by area


$Y = 28542.41 X + 0$

Average CF fit with equal weighting, forced to origin
 Coefficient of determination: 0.9940317
 Average error: 3.209%
 Average CF: 28542.41
 RSD: 4.497%

Level	Amount	Response	Cal Factor	Error, %	Source	Date and time
1	150	4325287	28835.25	1.026	Manual	10/18/2021 2:26:35 PM
2	1000	3.03352E+07	30335.2	6.281	Manual	10/18/2021 2:26:39 PM
3	5000	1.437314E+08	28746.28	0.714	Manual	10/18/2021 2:26:42 PM
4	15000	4.193721E+08	27958.14	-2.047	Manual	10/18/2021 2:26:45 PM
5	30000	8.051155E+08	26837.18	-5.974	Manual	10/18/2021 2:26:47 PM


3

*#Triacontane



Expected retention time: 16.26 minutes
 Search window: 0.05 minutes
 No retention time reference component
 Group number: 0
 High alarm limit: 0
 Low alarm limit: 0
 Component constant: 0
 Single peak quantification by area
 $Y = 28930.14 X + 0$
 Average CF fit with equal weighting, forced to origin
 Coefficient of determination: 0.9919451
 Average error: 5.737%
 Average CF: 28930.14
 RSD: 7.577%

Level	Amount	Response	Cal Factor	Error, %	Source	Date and time
1	2	59020.1	29510.05	2.005	Manual	10/18/2021 2:42:38 PM
2	50	1403134	28062.68	-2.998	G:\Org\HP4\DAT\HP4100621_b\1006HP4.0015.BND	10/7/2021 12:47:26 PM
3	200	6499949	32499.74	12.339	G:\Org\HP4\DAT\HP4100621_b\1006HP4.0017.BND	10/7/2021 12:47:56 PM
4	500	1.366713E+07	27334.26	-5.516	Manual	10/18/2021 2:44:43 PM
5	1000	2.724398E+07	27243.98	-5.828	Manual	10/18/2021 2:43:45 PM



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT


Sample Name: DCM-Baseline Check-V01
 Raw File: G:\org\HP5\DAT\HP5101721_b\1017HP5.0001.RAW
 Date & Time Acquired: 10/17/2021 12:40:02 PM
 Method File: G:\Org\HP5\Methods\DR_8015-HS-LEXP.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO210108Hs.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for TEH: 29457.33

Rt range for Diesel Range Organics: 6.51 to 15.85

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*o-Terphenyl	12.178	200.	.522	.26	-
*1-Chlorooctadecane	13.04	200.	1.235	.62	-

DRO Area: 571771.5 DRO Amount: 19.41016
 TEH Area: 639555.1 TEH Amount: 21.71124




DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: DCM-Baseline Check-V02
 Raw File: G:\org\HP5\DAT\HP5101721_b\1017HP5.0002.RAW
 Date & Time Acquired: 10/17/2021 1:22:21 PM
 Method File: G:\Org\HP5\Methods\DR_8015-HS-LEXP.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO210108Hs.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for TEH: 29457.33
 Rt range for Diesel Range Organics: 6.51 to 15.85

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*o-Terphenyl	29.973	200.	.	-
*1-Chlorooctadecane	13.043	200.	.018	.01 -

DRO Area:58862.56 DRO Amount: 1.998231
 TEH Area:105899.4 TEH Amount: 3.595009



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: CCV_1017HP503r, DRO ;1017HP5 , DRO210708A
 Raw File: G:\org\HP5\DAT\HP5101721_b\1017HP5.0003.RAW
 Date & Time Acquired: 10/17/2021 2:04:53 PM
 Method File: G:\Org\HP5\Methods\DR_8015-HS-LEXP.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO210108Hs.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for TEH: 29457.33
 Rt range for Diesel Range Organics: 6.51 to 15.85


SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*o-Terphenyl	29.982	200.	.	-
*1-Chlorooctadecane	13.037	200.	255.634	127.82 -

DRO Area: 1.592713E+08 DRO Amount: 5406.847
 TEH Area: 2.370861E+08 TEH Amount: 8048.458

CONTINUING CALIBRATION REPORT: G:\org\HP5\DAT\HP5101721_b\1017HP5.0003.RAW

COMPOUND	ACTUAL (NG)	MEASURED (NG)	%RECOVERY	LIMITS
TOTAL DRO	15000.	8048.46	53.66	85-115

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	LIMITS
*o-Terphenyl	29.982	200.	.	.	85-115
*1-Chlorooctadecane	13.037	200.	255.634	127.82	85-115




DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: DCM-Baseline Check-V04
 Raw File: G:\org\HP5\DAT\HP5101721_b\1017HP5.0004.RAW
 Date & Time Acquired: 10/17/2021 2:47:29 PM
 Method File: G:\Org\HP5\Methods\DR_8015-HS-LEXP.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO210108Hs.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for TEH: 29457.33
 Rt range for Diesel Range Organics: 6.51 to 15.85

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*o-Terphenyl	29.987	200.	.	-
*1-Chlorooctadecane	29.987	200.	.	-

DRO Area:32637.46 DRO Amount: 1.107957
 TEH Area:75218.02 TEH Amount: 2.553457



RESIDUAL RANGE ORGANICS CHROMATOGRAM

Sample Name: CCV_1017HP505r, CAL1 ;1017HP5 , 2 ug per mL Triacontane
 Raw File: G:\org\HP5\DAT\HP5101721_b\1017HP5.0005.RAW
 Date & Time Acquired: 10/17/2021 3:30:16 PM
 Method File: G:\Org\HP5\Methods\DS_ORO-AA-L0.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_ORO211017AA.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for for Residual Range Organics Calculations: 28542.41
 Rt range for Residual Range Organics: 12.53 to 30.05

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*#Triacontane	16.258	500.	2.028	.41	-

RRO Area:2747.039 RRO AMOUNT: 9.624412E-02


CONTINUING CALIBRATION REPORT: G:\org\HP5\DAT\HP5101721_b\1017HP5.0005.RAW

COMPOUND	ACTUAL (NG)	MEASURED (NG)	%RECOVERY	LIMITS
*30-40 Motor Oil	5000.	.	75-125	

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	LIMITS
*#Triacontane	16.258	200.	2.028	1.01	75-125

G:\org\HP5\DAT\HP5101721_b\1017HP5.0006.RAW

CCV_1017HP506r, CAL2 ;1017HP5 , 50 ug per mL Triacontane



RESIDUAL RANGE ORGANICS CHROMATOGRAM

Sample Name: CCV_1017HP506r, CAL2 ;1017HP5 , 50 ug per mL Triacontane
 Raw File: G:\org\HP5\DAT\HP5101721_b\1017HP5.0006.RAW
 Date & Time Acquired: 10/17/2021 4:12:57 PM
 Method File: G:\Org\HP5\Methods\DS_ORO-AA-L0.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_ORO211017AA.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for for Residual Range Organics Calculations: 28542.41
 Rt range for Residual Range Organics: 12.53 to 30.05


SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*#Triacontane	16.257	500.	48.178	9.64	-

RRO Area:45902.25 RRO AMOUNT: 1.608212

CONTINUING CALIBRATION REPORT: G:\org\HP5\DAT\HP5101721_b\1017HP5.0006.RAW

COMPOUND	ACTUAL (NG)	MEASURED (NG)	%RECOVERY	LIMITS
*30-40 Motor Oil	5000.	.	75-125	

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	LIMITS
*#Triacontane	16.257	200.	48.178	24.09	75-125



RESIDUAL RANGE ORGANICS CHROMATOGRAM

Sample Name: CCV_1017HP507r, CAL3 ;1017HP5 , 200 ug per mL Triacontane
 Raw File: G:\org\HP5\DAT\HP5101721_b\1017HP5.0007.RAW
 Date & Time Acquired: 10/17/2021 4:55:33 PM
 Method File: G:\Org\HP5\Methods\DS_ORO-AA-L0.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_ORO211017AA.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for for Residual Range Organics Calculations: 28542.41
 Rt range for Residual Range Organics: 12.53 to 30.05


SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*#Triacontane	16.262	500.	223.111	44.62	-

RRO Area:219754.5 RRO AMOUNT: 7.699227

CONTINUING CALIBRATION REPORT: G:\org\HP5\DAT\HP5101721_b\1017HP5.0007.RAW

COMPOUND	ACTUAL (NG)	MEASURED (NG)	%RECOVERY	LIMITS
*30-40 Motor Oil	5000.	.	75-125	

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	LIMITS
*#Triacontane	16.262	200.	223.111	111.56	75-125



RESIDUAL RANGE ORGANICS CHROMATOGRAM

Sample Name: CCV_1017HP508r, CAL4 ;1017HP5 , 500 ug per mL Triacontane
 Raw File: G:\org\HP5\DAT\HP5101721_b\1017HP5.0008.RAW
 Date & Time Acquired: 10/17/2021 5:38:10 PM
 Method File: G:\Org\HP5\Methods\DS_ORO-AA-L0.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_ORO211017AA.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for for Residual Range Organics Calculations: 28542.41
 Rt range for Residual Range Organics: 12.53 to 30.05


SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*#Triacontane	16.267	500.	470.063	94.01

RRO Area:496538.4 RRO AMOUNT: 17.39651

CONTINUING CALIBRATION REPORT: G:\org\HP5\DAT\HP5101721_b\1017HP5.0008.RAW

COMPOUND	ACTUAL (NG)	MEASURED (NG)	%RECOVERY	LIMITS
*30-40 Motor Oil	5000.	.	75-125	

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	LIMITS
*#Triacontane	16.267	200.	470.063	235.03	75-125



RESIDUAL RANGE ORGANICS CHROMATOGRAM

Sample Name: CCV_1017HP509r, CAL5 ;1017HP5 , 1000 ug per mL Triacontane
 Raw File: G:\org\HP5\DAT\HP5101721_b\1017HP5.0009.RAW
 Date & Time Acquired: 10/17/2021 6:20:57 PM
 Method File: G:\Org\HP5\Methods\DS_ORO-AA-L0.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_ORO211017AA.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for for Residual Range Organics Calculations: 28542.41
 Rt range for Residual Range Organics: 12.53 to 30.05

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*#Triacontane	16.275	500.	937.265	187.45	-

RRO Area:979213.9 RRO AMOUNT: 34.30733

CONTINUING CALIBRATION REPORT: G:\org\HP5\DAT\HP5101721_b\1017HP5.0009.RAW

COMPOUND	ACTUAL (NG)	MEASURED (NG)	%RECOVERY	LIMITS
*30-40 Motor Oil	5000.	.	75-125	

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	LIMITS
*#Triacontane	16.275	200.	937.265	468.63	75-125

Write Sequence	Data File	Sample Name	Method	Weight	Dil Factor	Amt Inj.	IS	Cal ID	Manual Integration
	G:\org\HP5\DAT\HP5101721_b\1017HP5.01r	DCM-Baseline Check-V01	G:\Org\HP5\Methods\DR_8015-HS-LEXP.met	1	1	1	1	0	No integration
	G:\org\HP5\DAT\HP5101721_b\1017HP5.02r	DCM-Baseline Check-V02	G:\Org\HP5\Methods\DR_8015-HS-LEXP.met	1	1	1	1	0	No integration
	G:\org\HP5\DAT\HP5101721_b\1017HP5.03r	CCV_1017HP503r, DRO :1017HP5 , DRO210708A	G:\Org\HP5\Methods\DR_8015-HS-LEXP.met	1	1	1	1	0	No integration
	G:\org\HP5\DAT\HP5101721_b\1017HP5.04r	DCM-Baseline Check-V04	G:\Org\HP5\Methods\DR_8015-HS-LEXP.met	1	1	1	1	0	No integration
	G:\org\HP5\DAT\HP5101721_b\1017HP5.05r	CCV_1017HP505r, CAL1 ;1017HP5 , 2 ug per mL Triacotane (10 uL of Cal3 + 990 uL DCM(14354)	G:\Org\HP5\Methods\DS_ORO-AA-L0.met	1	1	1	1	0	Surrogates are integrated using a valley to valley integration using Set Baseline Now placed slight before at 16.18 and slightly after the surrogate peak at 16.45 and scaling showing surrogate peak from 15.0-17.0
	G:\org\HP5\DAT\HP5101721_b\1017HP5.06r	CCV_1017HP506r, CAL2 ;1017HP5 , 50 ug per mL Triacotane (100 uL Cal4 + 900 uL of DCM(14354)	G:\Org\HP5\Methods\DS_ORO-AA-L0.met	1	1	1	1	0	Surrogates are integrated using a valley to valley integration using Set Baseline Now placed slight before at 16.18 and slightly after the surrogate peak at 16.45 and scaling showing surrogate peak from 15.0-17.0
	G:\org\HP5\DAT\HP5101721_b\1017HP5.07r	CCV_1017HP507r, CAL3 ;1017HP5 , 200 ug per mL Triacotane (100uL of Cal5 + 400 uL DCM(14354)	G:\Org\HP5\Methods\DS_ORO-AA-L0.met	1	1	1	1	0	Surrogates are integrated using a valley to valley integration using Set Baseline Now placed slight before at 16.18 and slightly after the surrogate peak at 16.45 and scaling showing surrogate peak from 15.0-17.0
	G:\org\HP5\DAT\HP5101721_b\1017HP5.08r	CCV_1017HP508r, CAL4 ;1017HP5 , 500 ug per mL Triacotane (250uL of Cal5 + 250 uL DCM(14354)	G:\Org\HP5\Methods\DS_ORO-AA-L0.met	1	1	1	1	0	Surrogates are integrated using a valley to valley integration using Set Baseline Now placed slight before at 16.18 and slightly after the surrogate peak at 16.45 and scaling showing surrogate peak from 15.0-17.0
	G:\org\HP5\DAT\HP5101721_b\1017HP5.09r	CCV_1017HP509r, CAL5 ;1017HP5 , 1000 ug per mL Triacotane (500 uL 2000 ug/mL Triacotane DRO211006A + 500 DCM(14354)	G:\Org\HP5\Methods\DS_ORO-AA-L0.met	1	1	1	1	0	Surrogates are integrated using a valley to valley integration using Set Baseline Now placed slight before at 16.18 and slightly after the surrogate peak at 16.45 and scaling showing surrogate peak from 15.0-17.0



Digitally signed by
Ann Nebel
Date: 2021.10.25 17:42:58 -06:00

PREP BATCH REPORT

Prep Code: **HC-3520-DRO**
 Prep Batch **162352** Prep Temp **NA °C**

Technician: **Jillian L Bostwick**
 Batch Units: **ML**

Prep Start Date: **12/20/2021 12:09:03 P**
 Prep End Date: **12/21/2021 1:43:00 P**

Sample ID	Matrix	pH	Initial Samp Amt	Sol Added	Sol Recovered	Final Vol (mL)	Factor	Balance	Prep Start Date	Prep End Date
MB-162352			1000	0	0	1.00	0.001		12/20/2021	12/21/2021
Lines 1-13 Start time: 12:05 PM, 12/20/2021. End time: 12/21/2021 at 6:05 AM.SGT was performed on remaining samples. amn 12/23/2021										
LCS-162352			1000	0	0	1.00	0.001		12/20/2021	12/21/2021
All bottles were completely used, defaced and disposed of on 12/16/2021. SGT was performed on remaining samples. amn 12/23/2021										
LCS-162352-RRO			1000	0	0	1.00	0.001		12/20/2021	12/21/2021
SGT was performed on remaining samples. amn 12/23/2021										
B21121613-001B	Ground Water	2	990	0	0	1.00	0.00101		12/20/2021	12/21/2021
Bottle 1/6. Clear, orange sediment. SGT was performed on remaining samples. amn 12/23/2021										
B21121613-001BMS	Ground Water	2	990	0	0	1.00	0.00101		12/20/2021	12/21/2021
Bottle 2/6. Clear, orange sediment. SGT was performed on remaining samples. amn 12/23/2021										
B21121613-001BMSD	Ground Water	2	1040	0	0	1.00	0.000962		12/20/2021	12/21/2021
Bottle 3/6. Clear, orange sediment. SGT was performed on remaining samples. amn 12/23/2021										
B21121613-001BMS-RRO	Ground Water	2	1040	0	0	1.00	0.000957		12/20/2021	12/21/2021
Bottle 4/6. Clear, orange sediment. SGT was performed on remaining samples. amn 12/23/2021										
B21121613-001BMSD-RRO	Ground Water	2	1060	0	0	1.00	0.000948		12/20/2021	12/21/2021
Bottle 5/6. Clear, orange sediment. SGT was performed on remaining samples. amn 12/23/2021										
B21121605-001A	Ground Water	2	1050	0	0	1.00	0.000952		12/20/2021	12/21/2021
Bottle 1/2. Clear.SGT was performed on remaining samples. amn 12/23/2021										
B21121605-002A	Ground Water	2	1050	0	0	1.00	0.000952		12/20/2021	12/21/2021
Bottle 1/2. Clear.SGT was performed on remaining samples. amn 12/23/2021										
B21121605-003A	Ground Water	2	1010	0	0	1.00	0.00099		12/20/2021	12/21/2021
Bottle 1/2. Clear.SGT was performed on remaining samples. amn 12/23/2021										
B21121609-001C	Ground Water	2	1060	0	0	1.00	0.000943		12/20/2021	12/21/2021
Bottle 1/1. Clear, light sediment.SGT was performed on remaining samples. amn 12/23/2021										
B21121611-001B	Ground Water	2	1030	0	0	1.00	0.000971		12/20/2021	12/21/2021
Bottle 1/2. Clear.SGT was performed on remaining samples. amn 12/23/2021										
B21121616-001C	Ground Water	2	990	0	0	1.00	0.00101		12/20/2021	12/21/2021
Bottle 1/2. Clear. Lines 14-18 Start time: 3:55 PM, 12/20/2021. End time: 12/21/2021 at 9:55 AM.SGT was performed on remaining samples. amn 12/23/2021										

Number	Reagent Name	Exp Date
11	Carbon Filter Water	1/1/2023
13379	PTFE Boiling Stones 27463755	12/30/2025
14206	pH-indicator Strips 0-14 HC160347	8/26/2026
14563	4ML, Amber Vial, 171001407106	11/30/2022
14596	Dichloromethane EC757	10/20/2023

Spk ID	Spike Name	SampType	AmtAdd	Exp Date
FP211207 14244	DCM RINSED FILTER PAPER	all	1	4/6/2026
Sulfate 12/10/21 (Baked Sodium Sulfate	all	Varies	11/29/2026
DRO211129A	Triacotane SURR 1000 ug/mL	all except LCS/D,	100 uL	4/6/2026
DRO211213A	OTP only SURR 2000 ug/mL	All except RRO-L	100 uL	9/30/2024
SG211217(13376)	Baked Silica Gel	ALL-SGT	5 g	2/28/2030
DRO211012B	#2 Diesel in Acetone 150,000 ug/mL	LCS, LCSD, MS	100 uL	11/5/2023
DRO210902A	50,000 ug/mL Oil Std for RRO-In D	LCS/D-RRO, MS-	100 uL	9/1/2026

PREP BATCH REPORT

Prep Code: **HC-3520-DRO**
 Prep Batch **162352** Prep Temp **NA °C**

Technician: **Jillian L Bostwick**
 Batch Units: **ML**

Prep Start Date: **12/20/2021 12:09:03 P**
 Prep End Date: **12/21/2021 1:43:00 P**

Sample ID	Matrix	pH	Initial Samp Amt	Sol Added	Sol Recovered	Final Vol (mL)	Factor	Balance	Prep Start Date	Prep End Date
B21121622-001B	Ground Water	2	960	0	0	1.00	0.00104		12/20/2021	12/21/2021
Bottle 1/2. Clear.SGT was performed on remaining samples. amn 12/23/2021										
B21121622-002B	Ground Water	2	1050	0	0	1.00	0.000952		12/20/2021	12/21/2021
Bottle 1/2. Clear.SGT was performed on remaining samples. amn 12/23/2021										
B21121622-003B	Ground Water	2	900	0	0	1.00	0.00111		12/20/2021	12/21/2021
Bottle 1/2. Clear.										
B21121623-001C	Ground Water	2	1020	0	0	1.00	0.00098		12/20/2021	12/21/2021
Bottle 1/2. Clear.SGT was performed on remaining samples. amn 12/23/2021										
B21121613-002B	Ground Water	2	1020	0	0	1.00	0.00098		12/20/2021	12/21/2021
Bottle 1/2. Clear. Lines 19 Start time: 4:55 PM, 12/20/2021. End time: 12/21/2021 at AM.										

Number	Reagent Name	Exp Date
11	Carbon Filter Water	1/1/2023
13379	PTFE Boiling Stones 27463755	12/30/2025
14206	pH-indicator Strips 0-14 HC160347	8/26/2026
14563	4ML, Amber Vial, 171001407106	11/30/2022
14596	Dichloromethane EC757	10/20/2023

Spk ID	Spike Name	SampType	AmtAdd	Exp Date
FP211207 14244	DCM RINSED FILTER PAPER	all	1	4/6/2026
Sulfate 12/10/21 (Baked Sodium Sulfate	all	Varies	11/29/2026
DRO211129A	Triacotane SURR 1000 ug/mL	all except LCS/D,	100 uL	4/6/2026
DRO211213A	OTP only SURR 2000 ug/mL	All except RRO-L	100 uL	9/30/2024
SG211217(13376)	Baked Silica Gel	ALL-SGT	5 g	2/28/2030
DRO211012B	#2 Diesel in Acetone 150,000 ug/mL	LCS, LCSD, MS	100 uL	11/5/2023
DRO210902A	50,000 ug/mL Oil Std for RRO-In D	LCS/D-RRO, MS-	100 uL	9/1/2026

Energy Laboratories Inc

ANALYTICAL RUN Summary

23-Dec-21

Run ID GCFID-HP5-B_211221A

Run Start Date: 12/21/2021
Analyst: Ann Nebel
Ical:
Column ID:
Comments: DRO-8015-ICAL information is in Index GCFID-HP5-B_211102A 8015C OIL range calibration GCFID-HP5-B_210218B

Std ID	Std Name	Std Amount	Std Units	Samp Amount	Samp Units	SampType	Expiration Date
DRO211207A	Carbon Scan STD-Marker					MARKER	3/5/2028
DRO211220A	8015 CCV-15,000ug/mL + 200 OTP					CCV	4/30/2023
DRO211220C	5,000 ug/mL RRO CCV 200 ug/mL Triacontane					CCV	4/6/2026

Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist					
14943612	CCV_1221HP50	HC-8015-DRO-	CCV		12/21/2021 4:25:	1	R372230		0	0						
Analyte	T	Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD	Q
TEH(Oil Range)	A	mg/L		4.73827734		5	0	0	0.0879	0.3	0	95%	80	120	0%	
n-Triacontane	S	mg/L		0.2138702		0.2	0	0	0.000336	0.002	0	107%	80	120	0%	

Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist					
14943613	CCV_1221HP50	HC-8015-DRO-	CCV		12/21/2021 5:08:	1	R372230		0	0						
Analyte	T	Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD	Q
Diesel Range Organics (C10 to C24)	A	mg/L		14.67219		15	0	0	0.0389	0.3	0	98%	80	120	0%	
Total Extractable Hydrocarbons	A	mg/L		15.2203		15	0	0	0.0749	0.3	50	101%	80	120	0%	
o-Terphenyl	S	mg/L		0.1961226		0.2	0	0	0.000429	0.002	0	98%	80	120	0%	

Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist					
14943614	LCS-162352	HC-8015-DRO-	LCS-DOD		12/21/2021 6:35:	1	162352	12/20/2021	0	0						
Analyte	T	Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD	Q

Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist					
14943614	LCS-162352	HC-8015-DRO-	LCS-DOD		12/21/2021 6:35:	1	162352	12/20/2021	0	0						
Analyte	T	Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD	Q
Diesel Range Organics (C10 to C24)	A	mg/L		13.31435		15	0	0	0.0389	0.3	0	89%	36	132	0%	
Total Extractable Hydrocarbons	A	mg/L		14.26188		15	0	0	0.0749	0.3	50	95%	60	132	0%	
o-Terphenyl	S	mg/L		0.1873054		0.2	0	0	0.000429	0.002	0	94%	56	125	0%	

Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist					
14943615	MB-162352	HC-8015-DRO-	MBLK		12/21/2021 7:18:	1	162352	12/20/2021	0	0						
Analyte	T	Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD	Q
Diesel Range Organics (C10 to C24)	A	mg/L		0		0	0	0	0.0389	0.15	0	0%	0	0	0%	
Oil Range Hydrocarbons (C24 to C40)	A	mg/L		0		0	0	0	0.0879	0.15	0	0%	0	0	0%	
Total Extractable Hydrocarbons	A	mg/L		0		0	0	0	0.0749	0.15	50	0%	0	0	0%	
n-Triacontane	S	mg/L		0.0944		0.1	0	0	0.000336	0.002	0	94%	50	150	0%	
o-Terphenyl	S	mg/L		0.1830072		0.2	0	0	0.000429	0.002	0	92%	56	125	0%	

Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist					
14943616	B21121613-001	HC-8015-DRO-	SAMP		12/21/2021 8:01:	1	162352	12/20/2021	0	0						
Analyte	T	Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD	Q
Diesel Range Organics (C10 to C24)	A	mg/L		0		0	0	0	0.039289	0.303	0	0%	0	0	0%	U
Oil Range Hydrocarbons (C24 to C40)	A	mg/L		0		0	0	0	0.088779	0.303	0	0%	0	0	0%	U
Total Extractable Hydrocarbons	A	mg/L		0		0	0	0	0.075649	0.303	50	0%	0	0	0%	U
n-Triacontane	S	mg/L		0.0923		0.101	0	0	0.0003394	0.00202	0	91%	50	150	0%	
o-Terphenyl	S	mg/L		0.1715049		0.202	0	0	0.0004333	0.00202	0	85%	56	125	0%	

Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist					
14943617	B21121613-001	HC-8015-DRO-	MS-DOD		12/21/2021 8:44:	1	162352	12/20/2021	1E+07	0						
Analyte	T	Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD	Q
Diesel Range Organics (C10 to C24)	A	mg/L		13.80946		15.15	0	0	0.039289	0.303	0	91%	36	132	0%	
Total Extractable Hydrocarbons	A	mg/L		14.83788		15.15	0	0	0.075649	0.303	50	98%	60	132	0%	
o-Terphenyl	S	mg/L		0.1836434		0.202	0	0	0.0004333	0.00202	0	91%	56	125	0%	

Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist					
14943618	B21121613-001	HC-8015-DRO-	MSD-DOD		12/21/2021 9:28:	1	162352	12/20/2021	1E+07	1E+07						
Analyte	T	Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD	Q
Diesel Range Organics (C10 to C24)	A	mg/L		12.21048		14.43	0	13.80946	0.0374218	0.3	0	85%	36	132	12%	
Total Extractable Hydrocarbons	A	mg/L		13.10213		14.43	0	14.83788	0.0720538	0.3	50	91%	60	132	12%	
o-Terphenyl	S	mg/L		0.1690977		0.1924	0	0	0.0004127	0.002	0	88%	56	125	0%	

Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist					
14943619	B21121613-002	HC-8015-DRO-	SAMP		12/21/2021 10:5	1	162352	12/20/2021	0	0						
Analyte	T	Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD	Q
Diesel Range Organics (C10 to C24)	A	mg/L		0		0	0	0	0.038122	0.3	0	0%	0	0	0%	U
Oil Range Hydrocarbons (C24 to C40)	A	mg/L		0		0	0	0	0.086142	0.3	0	0%	0	0	0%	U
Total Extractable Hydrocarbons	A	mg/L		0		0	0	0	0.073402	0.3	50	0%	0	0	0%	U
n-Triacontane	S	mg/L		0.0898		0.098	0	0	0.0003293	0.00196	0	92%	50	150	0%	
o-Terphenyl	S	mg/L		0.1661896		0.196	0	0	0.0004204	0.002	0	85%	56	125	0%	

Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist					
14943620	B21121609-001	HC-8015-DRO-	SAMP		12/22/2021 12:2	1	162352	12/20/2021	0	0						
Analyte	T	Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD	Q
Diesel Range Organics (C10 to C24)	A	mg/L		4.656794		0	0	0	0.0366827	0.3	0	0%	0	0	0%	
Oil Range Hydrocarbons (C24 to C40)	A	mg/L		0.37599784		0	0	0	0.0828897	0.3	0	0%	0	0	0%	
Total Extractable Hydrocarbons	A	mg/L		5.139189		0	0	0	0.0706307	0.3	50	0%	0	0	0%	
n-Triacontane	S	mg/L		0.0843		0.0943	0	0	0.0003168	0.001886	0	89%	50	150	0%	
o-Terphenyl	S	mg/L		0.1517489		0.1886	0	0	0.0004045	0.002	0	80%	56	125	0%	

Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist					
14943621	B21121611-001	HC-8015-DRO-	SAMP		12/22/2021 1:46:	1	162352	12/20/2021	0	0						
Analyte	T	Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD	Q
Diesel Range Organics (C10 to C24)	A	mg/L		0.1643919		0	0	0	0.0377719	0.3	0	0%	0	0	0%	J
Oil Range Hydrocarbons (C24 to C40)	A	mg/L		0.29179561		0	0	0	0.0853509	0.3	0	0%	0	0	0%	J
Total Extractable Hydrocarbons	A	mg/L		0.6027129		0	0	0	0.0727279	0.3	50	0%	0	0	0%	
n-Triacontane	S	mg/L		0.0938		0.0971	0	0	0.0003263	0.001942	0	97%	50	150	0%	
o-Terphenyl	S	mg/L		0.1730123		0.1942	0	0	0.0004166	0.002	0	89%	56	125	0%	

Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist					
14943622	B21121616-001	HC-8015-DRO-	SAMP		12/22/2021 2:30:	1	162352	12/20/2021	0	0						
Analyte	T	Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD	Q
Diesel Range Organics (C10 to C24)	A	mg/L		1.215285		0	0	0	0.039289	0.303	0	0%	0	0	0%	
Oil Range Hydrocarbons (C24 to C40)	A	mg/L		0.26310992		0	0	0	0.088779	0.303	0	0%	0	0	0%	J
Total Extractable Hydrocarbons	A	mg/L		1.655497		0	0	0	0.075649	0.303	50	0%	0	0	0%	
n-Triacontane	S	mg/L		0.1		0.101	0	0	0.0003394	0.00202	0	99%	50	150	0%	
o-Terphenyl	S	mg/L		0.1523631		0.202	0	0	0.0004333	0.00202	0	75%	56	125	0%	
Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist					
14943623	CCV_1221HP52	HC-8015-DRO-	CCV		12/22/2021 3:56:	1	R372230				0	0				
Analyte	T	Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD	Q
TEH(Oil Range)	A	mg/L		4.80260889		5	0	0	0.0879	0.3	0	96%	80	120	0%	
n-Triacontane	S	mg/L		0.2118513		0.2	0	0	0.000336	0.002	0	106%	80	120	0%	
Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist					
14943624	CCV_1221HP52	HC-8015-DRO-	CCV		12/22/2021 4:39:	1	R372230				0	0				
Analyte	T	Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD	Q
Diesel Range Organics (C10 to C24)	A	mg/L		14.86623		15	0	0	0.0389	0.3	0	99%	80	120	0%	
Total Extractable Hydrocarbons	A	mg/L		15.39986		15	0	0	0.0749	0.3	50	103%	80	120	0%	
o-Terphenyl	S	mg/L		0.198526		0.2	0	0	0.000429	0.002	0	99%	80	120	0%	
Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist					
14943625	B21121623-001	HC-8015-DRO-	SAMP		12/22/2021 7:44:	1	162352	12/20/2021	0	0						
Analyte	T	Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD	Q
Diesel Range Organics (C10 to C24)	A	mg/L		0.1946039		0	0	0	0.038122	0.3	0	0%	0	0	0%	J
Oil Range Hydrocarbons (C24 to C40)	A	mg/L		0		0	0	0	0.086142	0.3	0	0%	0	0	0%	U
Total Extractable Hydrocarbons	A	mg/L		0.3643004		0	0	0	0.073402	0.3	50	0%	0	0	0%	
n-Triacontane	S	mg/L		0.0964		0.098	0	0	0.0003293	0.00196	0	98%	50	150	0%	
o-Terphenyl	S	mg/L		0.1681939		0.196	0	0	0.0004204	0.002	0	86%	56	125	0%	

Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist					
14943626	B21121605-001	HC-8015-DRO-	SAMP		12/22/2021 9:10:	1	162352	12/20/2021	0	0						
Analyte	T	Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD	Q
Diesel Range Organics (C10 to C24)	A	mg/L		0.1463492		0	0	0	0.0370328	0.3	0	0%	0	0	0%	J
Oil Range Hydrocarbons (C24 to C40)	A	mg/L		0.10364144		0	0	0	0.0836808	0.3	0	0%	0	0	0%	J
Total Extractable Hydrocarbons	A	mg/L		0.2733932		0	0	0	0.0713048	0.3	50	0%	0	0	0%	J
n-Triacontane	S	mg/L		0.0856		0.0952	0	0	0.0003199	0.001904	0	90%	50	150	0%	
o-Terphenyl	S	mg/L		0.1648605		0.1904	0	0	0.0004084	0.002	0	87%	56	125	0%	
Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist					
14943627	B21121605-002	HC-8015-DRO-	SAMP		12/22/2021 10:3	1	162352	12/20/2021	0	0						
Analyte	T	Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD	Q
Diesel Range Organics (C10 to C24)	A	mg/L		1.537647		0	0	0	0.0370328	0.3	0	0%	0	0	0%	
Oil Range Hydrocarbons (C24 to C40)	A	mg/L		0.17544827		0	0	0	0.0836808	0.3	0	0%	0	0	0%	J
Total Extractable Hydrocarbons	A	mg/L		1.808759		0	0	0	0.0713048	0.3	50	0%	0	0	0%	
n-Triacontane	S	mg/L		0.0944		0.0952	0	0	0.0003199	0.001904	0	99%	50	150	0%	
o-Terphenyl	S	mg/L		0.1767315		0.1904	0	0	0.0004084	0.002	0	93%	56	125	0%	
Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist					
14943628	B21121605-003	HC-8015-DRO-	SAMP		12/22/2021 11:1	1	162352	12/20/2021	0	0						
Analyte	T	Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD	Q
Diesel Range Organics (C10 to C24)	A	mg/L		1.759276		0	0	0	0.038511	0.3	0	0%	0	0	0%	
Oil Range Hydrocarbons (C24 to C40)	A	mg/L		0.17849281		0	0	0	0.087021	0.3	0	0%	0	0	0%	J
Total Extractable Hydrocarbons	A	mg/L		2.009793		0	0	0	0.074151	0.3	50	0%	0	0	0%	
n-Triacontane	S	mg/L		0.0965		0.099	0	0	0.0003326	0.00198	0	97%	50	150	0%	
o-Terphenyl	S	mg/L		0.184375		0.198	0	0	0.0004247	0.002	0	93%	56	125	0%	
Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist					
14943629	B21121622-002	HC-8015-DRO-	SAMP		12/22/2021 1:27:	1	162352	12/20/2021	0	0						
Analyte	T	Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD	Q
Diesel Range Organics (C10 to C24)	A	mg/L		0.3060482		0	0	0	0.0370328	0.3	0	0%	0	0	0%	
Oil Range Hydrocarbons (C24 to C40)	A	mg/L		0.38883811		0	0	0	0.0836808	0.3	0	0%	0	0	0%	
Total Extractable Hydrocarbons	A	mg/L		0.7950797		0	0	0	0.0713048	0.3	50	0%	0	0	0%	
n-Triacontane	S	mg/L		0.0955		0.0952	0	0	0.0003199	0.001904	0	100%	50	150	0%	

Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist					
14943629	B21121622-002	HC-8015-DRO-	SAMP		12/22/2021 1:27:	1	162352	12/20/2021	0	0						
Analyte	T	Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD	Q
o-Terphenyl	S	mg/L		0.1789114		0.1904	0	0	0.0004084	0.002	0	94%	56	125	0%	
Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist					
14943770	CCV_1221HP53	HC-8015-DRO-	CCV		12/22/2021 3:35:	1	R372230				0	0				
Analyte	T	Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD	Q
TEH(Oil Range)	A	mg/L		4.68422070		5	0	0	0.0879	0.3	0	94%	80	120	0%	
n-Triacontane	S	mg/L		0.2093559		0.2	0	0	0.000336	0.002	0	105%	80	120	0%	
Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist					
14943771	CCV_1221HP53	HC-8015-DRO-	CCV		12/22/2021 4:18:	1	R372230				0	0				
Analyte	T	Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD	Q
Diesel Range Organics (C10 to C24)	A	mg/L		14.01734		15	0	0	0.0389	0.3	0	93%	80	120	0%	
Total Extractable Hydrocarbons	A	mg/L		14.5345		15	0	0	0.0749	0.3	50	97%	80	120	0%	
o-Terphenyl	S	mg/L		0.1871337		0.2	0	0	0.000429	0.002	0	94%	80	120	0%	
Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist					
14944673	B21121622-001	HC-8015-DRO-	SAMP		12/22/2021 6:29:	1	162352	12/20/2021	0	0						
Analyte	T	Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD	Q
Diesel Range Organics (C10 to C24)	A	mg/L		0		0	0	0	0.040456	0.312	0	0%	0	0	0%	U
Oil Range Hydrocarbons (C24 to C40)	A	mg/L		0.10430708		0	0	0	0.091416	0.312	0	0%	0	0	0%	J
Total Extractable Hydrocarbons	A	mg/L		0.1345782		0	0	0	0.077896	0.312	50	0%	0	0	0%	J
n-Triacontane	S	mg/L		0.1016		0.104	0	0	0.0003494	0.00208	0	98%	50	150	0%	
o-Terphenyl	S	mg/L		0.1943547		0.208	0	0	0.0004462	0.00208	0	93%	56	125	0%	
Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist					
14944674	B21121622-003	HC-8015-DRO-	SAMP		12/22/2021 7:13:	1	162352	12/20/2021	0	0						
Analyte	T	Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD	Q
Diesel Range Organics (C10 to C24)	A	mg/L		0		0	0	0	0.043179	0.333	0	0%	0	0	0%	U
Oil Range Hydrocarbons (C24 to C40)	A	mg/L		0		0	0	0	0.097569	0.333	0	0%	0	0	0%	U
Total Extractable Hydrocarbons	A	mg/L		0		0	0	0	0.083139	0.333	50	0%	0	0	0%	U

Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist					
14944674	B21121622-003	HC-8015-DRO-	SAMP		12/22/2021 7:13:	1	162352	12/20/2021	0	0						
Analyte	T	Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD	Q
n-Triacontane	S	mg/L		0.1055		0.111	0	0	0.000373	0.00222	0	95%	50	150	0%	
o-Terphenyl	S	mg/L		0.1998307		0.222	0	0	0.0004762	0.00222	0	90%	50	150	0%	
Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist					
14944675	B21121613-001	HC-8015-DRO-	MS-DOD		12/22/2021 8:39:	1	162352	12/20/2021	1E+07	0						
Analyte	T	Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD	Q
TEH(Oil Range)	A	mg/L		4.48855734		4.785	0	0	0.0841203	0.3	0	94%	41	113	0%	
n-Triacontane	S	mg/L		0.0913		0.0957	0	0	0.0003216	0.002	0	95%	50	150	0%	
Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist					
14944676	B21121613-001	HC-8015-DRO-	MSD-DOD		12/22/2021 10:0	1	162352	12/20/2021	1E+07	1E+07						
Analyte	T	Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD	Q
TEH(Oil Range)	A	mg/L		4.49660873		4.74	0	4.4885573	0.0833292	0.3	0	95%	41	113	0%	
n-Triacontane	S	mg/L		0.0895		0.0948	0	0	0.0003185	0.002	0	94%	50	150	0%	
Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist					
14944677	LCS-162352-RR	HC-8015-DRO-	LCS-DOD		12/22/2021 10:4	1	162352	12/20/2021	0	0						
Analyte	T	Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD	Q
TEH(Oil Range)	A	mg/L		4.64328671		5	0	0	0.0879	0.3	0	93%	41	113	0%	
n-Triacontane	S	mg/L		0.09		0.1	0	0	0.000336	0.002	0	90%	50	150	0%	
Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist					
14944678	CCV_1221HP54	HC-8015-DRO-	CCV		12/23/2021 12:1	1	R372230		0	0						
Analyte	T	Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD	Q
TEH(Oil Range)	A	mg/L		4.67174316		5	0	0	0.0879	0.3	0	93%	80	120	0%	
n-Triacontane	S	mg/L		0.207587		0.2	0	0	0.000336	0.002	0	104%	80	120	0%	

Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist					
14944679	CCV_1221HP54	HC-8015-DRO-	CCV		12/23/2021 12:5	1	R372230		0	0						
Analyte	T	Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD	Q
Diesel Range Organics (C10 to C24)	A	mg/L		14.33879		15	0	0	0.0389	0.3	0	96%	80	120	0%	
Total Extractable Hydrocarbons	A	mg/L		14.83493		15	0	0	0.0749	0.3	50	99%	80	120	0%	
o-Terphenyl	S	mg/L		0.1912967		0.2	0	0	0.000429	0.002	0	96%	80	120	0%	

Energy Laboratories Inc

ANALYTICAL RUN Summary

28-Dec-21

Run ID GCFID-HP5-B_211226A

Run Start Date: 12/26/2021
Analyst: Ann Nebel
Ical:
Column ID:
Comments: DRO-8015-ICAL information is in Index GCFID-HP5-B_211102A 8015C OIL range calibration GCFID-HP5-B_210218B

Std ID	Std Name	Std Amount	Std Units	Samp Amount	Samp Units	SampType	Expiration Date
DRO211207A	Carbon Scan STD-Marker					MARKER	3/5/2028
DRO211220A	8015 CCV-15,000ug/mL + 200 OTP					CCV	4/30/2023
DRO211220C	5,000 ug/mL RRO CCV 200 ug/mL Triacontane					CCV	4/6/2026

Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist						
14948157	CCV_1226HP50	HC-8015-DRO-	CCV		12/26/2021 1:24:	1	R372351			0	0						
Analyte		T	Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD	Q
TEH(Oil Range)		A	mg/L		5.07259375		5	0	0	0.0879	0.3	0	101%	80	120	0%	
n-Triacontane		S	mg/L		0.2313521		0.2	0	0	0.000336	0.002	0	116%	80	120	0%	

Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist						
14948158	CCV_1226HP50	HC-8015-DRO-	CCV		12/26/2021 2:07:	1	R372351			0	0						
Analyte		T	Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD	Q
Diesel Range Organics (C10 to C24)		A	mg/L		15.84015		15	0	0	0.0389	0.3	0	106%	80	120	0%	
Total Extractable Hydrocarbons		A	mg/L		16.4279		15	0	0	0.0749	0.3	50	110%	80	120	0%	
o-Terphenyl		S	mg/L		0.2132998		0.2	0	0	0.000429	0.002	0	107%	80	120	0%	

Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist						
14948159	MB-162352	HC-8015-DRO-	MBLK		12/26/2021 4:16:	1	162352	12/20/2021		0	0						
Analyte		T	Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD	Q

Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist					
14948159	MB-162352	HC-8015-DRO-	MBLK		12/26/2021 4:16:	1	162352	12/20/2021	0	0						
Analyte	T	Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD	Q
Diesel Range Organics (SGT-C10 to	A	mg/L		0		0	0	0	0.0389	0.15	0	0%	0	0	0%	
Oil Range Hydrocarbons (SGT-C24 t	A	mg/L		0		0	0	0	0.0879	0.15	0	0%	0	0	0%	
Total Extractable Hydrocarbons (SGT	A	mg/L		0		0	0	0	0.0329	0.15	0	0%	0	0	0%	
n-Triacontane (SGT)	S	mg/L		0.1176		0.1	0	0	0.000336	0.002	0	118%	50	150	0%	
o-Terphenyl (SGT)	S	mg/L		0.2176667		0.2	0	0	0.000429	0.002	0	109%	56	125	0%	
Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist					
14948160	B21121613-001	HC-8015-DRO-	SAMP		12/26/2021 4:59:	1	162352	12/20/2021	0	0						
Analyte	T	Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD	Q
Diesel Range Organics (SGT-C10 to	A	mg/L		0		0	0	0	0.039289	0.303	0	0%	0	0	0%	U
Oil Range Hydrocarbons (SGT-C24 t	A	mg/L		0		0	0	0	0.088779	0.303	0	0%	0	0	0%	U
Total Extractable Hydrocarbons (SGT	A	mg/L		0		0	0	0	0.033229	0.303	0	0%	0	0	0%	U
n-Triacontane (SGT)	S	mg/L		0.0829		0.101	0	0	0.0003394	0.00202	0	82%	50	150	0%	
o-Terphenyl (SGT)	S	mg/L		0.153843		0.202	0	0	0.0004333	0.00202	0	76%	56	125	0%	
Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist					
14948161	B21121613-001	HC-8015-DRO-	MS-DOD		12/26/2021 5:42:	1	162352	12/20/2021	1E+07	0						
Analyte	T	Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD	Q
Diesel Range Organics (SGT-C10 to	A	mg/L		15.84011		15.15	0	0	0.039289	0.303	0	105%	36	132	0%	
Total Extractable Hydrocarbons (SGT	A	mg/L		16.83423		15.15	0	0	0.033229	0.303	0	111%	60	132	0%	
o-Terphenyl (SGT)	S	mg/L		0.2156087		0.202	0	0	0.0004333	0.00202	0	107%	56	125	0%	
Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist					
14948162	B21121613-001	HC-8015-DRO-	MSD-DOD		12/26/2021 6:25:	1	162352	12/20/2021	1E+07	1E+07						
Analyte	T	Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD	Q
Diesel Range Organics (SGT-C10 to	A	mg/L		14.5995		14.43	0	15.84011	0.0374218	0.3	0	101%	36	132	8%	
Total Extractable Hydrocarbons (SGT	A	mg/L		15.5177		14.43	0	16.83423	0.0316498	0.3	0	108%	60	132	8%	
o-Terphenyl (SGT)	S	mg/L		0.2096413		0.1924	0	0	0.0004127	0.002	0	109%	56	125	0%	

Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist					
14948163	B21121622-001	HC-8015-DRO-	SAMP		12/26/2021 7:51:	1	162352	12/20/2021	0	0						
Analyte	T	Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD	Q
Diesel Range Organics (SGT-C10 to A	mg/L			0		0	0	0	0.040456	0.312	0	0%	0	0	0%	U
Oil Range Hydrocarbons (SGT-C24 t A	mg/L			0		0	0	0	0.091416	0.312	0	0%	0	0	0%	U
Total Extractable Hydrocarbons (SGT A	mg/L			0		0	0	0	0.034216	0.312	0	0%	0	0	0%	U
n-Triacontane (SGT)	S	mg/L		0.1008		0.104	0	0	0.0003494	0.00208	0	97%	50	150	0%	
o-Terphenyl (SGT)	S	mg/L		0.1898372		0.208	0	0	0.0004462	0.00208	0	91%	56	125	0%	
Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist					
14948164	B21121605-001	HC-8015-DRO-	SAMP		12/26/2021 8:34:	1	162352	12/20/2021	0	0						
Analyte	T	Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD	Q
Diesel Range Organics (SGT-C10 to A	mg/L			0.0597298		0	0	0	0.0370328	0.3	0	0%	0	0	0%	J
Oil Range Hydrocarbons (SGT-C24 t A	mg/L			0		0	0	0	0.0836808	0.3	0	0%	0	0	0%	U
Total Extractable Hydrocarbons (SGT A	mg/L			0.06545574		0	0	0	0.0313208	0.3	0	0%	0	0	0%	J
n-Triacontane (SGT)	S	mg/L		0.0815		0.0952	0	0	0.0003199	0.001904	0	86%	50	150	0%	
o-Terphenyl (SGT)	S	mg/L		0.1577361		0.1904	0	0	0.0004084	0.001904	0	83%	56	125	0%	
Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist					
14948165	B21121623-001	HC-8015-DRO-	SAMP		12/26/2021 9:17:	1	162352	12/20/2021	0	0						
Analyte	T	Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD	Q
Diesel Range Organics (SGT-C10 to A	mg/L			0.05812604		0	0	0	0.038122	0.3	0	0%	0	0	0%	J
Oil Range Hydrocarbons (SGT-C24 t A	mg/L			0		0	0	0	0.086142	0.3	0	0%	0	0	0%	U
Total Extractable Hydrocarbons (SGT A	mg/L			0.06626495		0	0	0	0.032242	0.3	0	0%	0	0	0%	J
n-Triacontane (SGT)	S	mg/L		0.1113		0.098	0	0	0.0003293	0.00196	0	114%	50	150	0%	
o-Terphenyl (SGT)	S	mg/L		0.1920631		0.196	0	0	0.0004204	0.00196	0	98%	56	125	0%	
Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist					
14948166	B21121611-001	HC-8015-DRO-	SAMP		12/26/2021 10:4	1	162352	12/20/2021	0	0						
Analyte	T	Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD	Q
Diesel Range Organics (SGT-C10 to A	mg/L			0		0	0	0	0.0377719	0.3	0	0%	0	0	0%	U
Oil Range Hydrocarbons (SGT-C24 t A	mg/L			0		0	0	0	0.0853509	0.3	0	0%	0	0	0%	U
Total Extractable Hydrocarbons (SGT A	mg/L			0.04004378		0	0	0	0.0319459	0.3	0	0%	0	0	0%	J
n-Triacontane (SGT)	S	mg/L		0.1055		0.0971	0	0	0.0003263	0.001942	0	109%	50	150	0%	

Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist					
14948166	B21121611-001	HC-8015-DRO-	SAMP		12/26/2021 10:4	1	162352	12/20/2021	0	0						
Analyte	T	Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD	Q
o-Terphenyl (SGT)	S	mg/L		0.1996854		0.1942	0	0	0.0004166	0.001942	0	103%	56	125	0%	
14948167	B21121622-002	HC-8015-DRO-	SAMP		12/26/2021 11:2	1	162352	12/20/2021	0	0						
Analyte	T	Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD	Q
Diesel Range Organics (SGT-C10 to A	mg/L			0		0	0	0	0.0370328	0.3	0	0%	0	0	0%	U
Oil Range Hydrocarbons (SGT-C24 t A	mg/L			0		0	0	0	0.0836808	0.3	0	0%	0	0	0%	U
Total Extractable Hydrocarbons (SGT A	mg/L			0		0	0	0	0.0313208	0.3	0	0%	0	0	0%	U
n-Triacontane (SGT)	S	mg/L		0.0829		0.0952	0	0	0.0003199	0.001904	0	87%	50	150	0%	
o-Terphenyl (SGT)	S	mg/L		0.1551555		0.1904	0	0	0.0004084	0.001904	0	81%	56	125	0%	
14948168	CCV_1226HP50	HC-8015-DRO-	CCV		12/27/2021 12:5	1	R372351				0	0				
Analyte	T	Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD	Q
TEH(Oil Range)	A	mg/L		4.88013965		5	0	0	0.0879	0.3	0	98%	80	120	0%	
n-Triacontane	S	mg/L		0.2220504		0.2	0	0	0.000336	0.002	0	111%	80	120	0%	
14948169	CCV_1226HP50	HC-8015-DRO-	CCV		12/27/2021 1:34:	1	R372351				0	0				
Analyte	T	Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD	Q
Diesel Range Organics (C10 to C24) A	mg/L			15.86322		15	0	0	0.0389	0.3	0	106%	80	120	0%	
Total Extractable Hydrocarbons	A	mg/L		16.45564		15	0	0	0.0749	0.3	50	110%	80	120	0%	
o-Terphenyl	S	mg/L		0.2124378		0.2	0	0	0.000429	0.002	0	106%	80	120	0%	
14948170	B21121616-001	HC-8015-DRO-	SAMP		12/27/2021 3:43:	1	162352	12/20/2021	0	0						
Analyte	T	Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD	Q
Diesel Range Organics (SGT-C10 to A	mg/L			0.2567575		0	0	0	0.039289	0.303	0	0%	0	0	0%	J
Oil Range Hydrocarbons (SGT-C24 t A	mg/L			0		0	0	0	0.088779	0.303	0	0%	0	0	0%	U
Total Extractable Hydrocarbons (SGT A	mg/L			0.2700363		0	0	0	0.033229	0.303	0	0%	0	0	0%	J


Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist					
14948170	B21121616-001	HC-8015-DRO-	SAMP		12/27/2021 3:43:	1	162352	12/20/2021	0	0						
Analyte	T	Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD	Q
n-Triacontane (SGT)	S	mg/L		0.0962		0.101	0	0	0.0003394	0.00202	0	95%	50	150	0%	
o-Terphenyl (SGT)	S	mg/L		0.1451255		0.202	0	0	0.0004333	0.00202	0	72%	56	125	0%	
Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist					
14948171	B21121605-002	HC-8015-DRO-	SAMP		12/27/2021 4:26:	1	162352	12/20/2021	0	0						
Analyte	T	Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD	Q
Diesel Range Organics (SGT-C10 to A	mg/L			1.257842		0	0	0	0.0370328	0.3	0	0%	0	0	0%	
Oil Range Hydrocarbons (SGT-C24 t A	mg/L			0.10345144		0	0	0	0.0836808	0.3	0	0%	0	0	0%	J
Total Extractable Hydrocarbons (SGT A	mg/L			1.385268		0	0	0	0.0313208	0.3	0	0%	0	0	0%	
n-Triacontane (SGT)	S	mg/L		0.0986		0.0952	0	0	0.0003199	0.001904	0	104%	50	150	0%	
o-Terphenyl (SGT)	S	mg/L		0.1885758		0.1904	0	0	0.0004084	0.001904	0	99%	56	125	0%	
Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist					
14948172	B21121605-003	HC-8015-DRO-	SAMP		12/27/2021 5:09:	1	162352	12/20/2021	0	0						
Analyte	T	Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD	Q
Diesel Range Organics (SGT-C10 to A	mg/L			1.397764		0	0	0	0.038511	0.3	0	0%	0	0	0%	
Oil Range Hydrocarbons (SGT-C24 t A	mg/L			0		0	0	0	0.087021	0.3	0	0%	0	0	0%	U
Total Extractable Hydrocarbons (SGT A	mg/L			1.502644		0	0	0	0.032571	0.3	0	0%	0	0	0%	
n-Triacontane (SGT)	S	mg/L		0.0929		0.099	0	0	0.0003326	0.00198	0	94%	50	150	0%	
o-Terphenyl (SGT)	S	mg/L		0.1816916		0.198	0	0	0.0004247	0.00198	0	92%	56	125	0%	
Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist					
14948173	B21121609-001	HC-8015-DRO-	SAMP		12/27/2021 5:51:	1	162352	12/20/2021	0	0						
Analyte	T	Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD	Q
Diesel Range Organics (SGT-C10 to A	mg/L			3.977403		0	0	0	0.0366827	0.3	0	0%	0	0	0%	
Oil Range Hydrocarbons (SGT-C24 t A	mg/L			0.17698956		0	0	0	0.0828897	0.3	0	0%	0	0	0%	J
Total Extractable Hydrocarbons (SGT A	mg/L			4.198934		0	0	0	0.0310247	0.3	0	0%	0	0	0%	
n-Triacontane (SGT)	S	mg/L		0.0955		0.0943	0	0	0.0003168	0.001886	0	101%	50	150	0%	
o-Terphenyl (SGT)	S	mg/L		0.1784275		0.1886	0	0	0.0004045	0.001886	0	95%	56	125	0%	

Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist					
14948174	B21121613-001	HC-8015-DRO-	MS-DOD		12/27/2021 6:34:	1	162352	12/20/2021	1E+07	0						
Analyte	T	Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD	Q
TEH (SGT-Oil Range)	A	mg/L		4.7792449		4.785	0	0	0.0841203	0.3	0	100%	41	113	0%	
n-Triacontane (SGT)	S	mg/L		0.0924		0.0957	0	0	0.0003216	0.002	0	97%	50	150	0%	
Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist					
14948175	LCS-162352	HC-8015-DRO-	LCS-DOD		12/27/2021 8:00:	1	162352	12/20/2021		0	0					
Analyte	T	Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD	Q
Diesel Range Organics (SGT-C10 to	A	mg/L		14.89724		15	0	0	0.0389	0.3	0	99%	36	132	0%	
Total Extractable Hydrocarbons (SGT	A	mg/L		15.84407		15	0	0	0.0329	0.3	0	106%	60	132	0%	
o-Terphenyl (SGT)	S	mg/L		0.2019174		0.2	0	0	0.000429	0.002	0	101%	56	125	0%	
Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist					
14948176	B21121613-001	HC-8015-DRO-	MSD-DOD		12/27/2021 8:43:	1	162352	12/20/2021	1E+07	1E+07						
Analyte	T	Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD	Q
TEH (SGT-Oil Range)	A	mg/L		4.87512207		4.74	0	4.7792449	0.0833292	0.3	0	103%	41	113	2%	
n-Triacontane (SGT)	S	mg/L		0.0938		0.0948	0	0	0.0003185	0.002	0	99%	50	150	0%	
Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist					
14948178	CCV_1226HP53	HC-8015-DRO-	CCV		12/27/2021 10:5	1	R372351			0	0					
Analyte	T	Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD	Q
TEH(Oil Range)	A	mg/L		4.77245410		5	0	0	0.0879	0.3	0	95%	80	120	0%	
n-Triacontane	S	mg/L		0.2152825		0.2	0	0	0.000336	0.002	0	108%	80	120	0%	
Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist					
14948179	CCV_1226HP53	HC-8015-DRO-	CCV		12/27/2021 11:3	1	R372351			0	0					
Analyte	T	Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD	Q
Diesel Range Organics (C10 to C24)	A	mg/L		14.891		15	0	0	0.0389	0.3	0	99%	80	120	0%	
Total Extractable Hydrocarbons	A	mg/L		15.44637		15	0	0	0.0749	0.3	50	103%	80	120	0%	
o-Terphenyl	S	mg/L		0.1995018		0.2	0	0	0.000429	0.002	0	100%	80	120	0%	

Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist					
14950319	CCV_1226HP55	HC-8015-DRO-	CCV		12/27/2021 10:2	1	R372351		0	0						
Analyte	T	Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD	Q
TEH(Oil Range)	A	mg/L		5.13239502		5	0	0	0.0879	0.3	0	103%	80	120	0%	
n-Triacontane	S	mg/L		0.2173307		0.2	0	0	0.000336	0.002	0	109%	80	120	0%	
Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist					
14950320	CCV_1226HP55	HC-8015-DRO-	CCV		12/27/2021 11:0	1	R372351		0	0						
Analyte	T	Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD	Q
Diesel Range Organics (C10 to C24)	A	mg/L		15.44586		15	0	0	0.0389	0.3	0	103%	80	120	0%	
Total Extractable Hydrocarbons	A	mg/L		15.95357		15	0	0	0.0749	0.3	50	106%	80	120	0%	
o-Terphenyl	S	mg/L		0.212485		0.2	0	0	0.000429	0.002	0	106%	80	120	0%	
Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist					
14950321	LCS-162352-RR	HC-8015-DRO-	LCS-DOD		12/28/2021 3:25:	1	162352	12/20/2021	0	0						
Analyte	T	Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD	Q
TEH (SGT-Oil Range)	A	mg/L		5.67233229		5	0	0	0.0879	0.3	0	113%	41	113	0%	
n-Triacontane (SGT)	S	mg/L		0.1133		0.1	0	0	0.000336	0.002	0	113%	50	150	0%	
Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist					
14950322	CCV_1226HP56	HC-8015-DRO-	CCV		12/28/2021 6:17:	1	R372351		0	0						
Analyte	T	Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD	Q
TEH(Oil Range)	A	mg/L		5.55641016		5	0	0	0.0879	0.3	0	111%	80	120	0%	
n-Triacontane	S	mg/L		0.2231189		0.2	0	0	0.000336	0.002	0	112%	80	120	0%	
Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist					
14950323	CCV_1226HP56	HC-8015-DRO-	CCV		12/28/2021 7:00:	1	R372351		0	0						
Analyte	T	Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD	Q
Diesel Range Organics (C10 to C24)	A	mg/L		15.843		15	0	0	0.0389	0.3	0	106%	80	120	0%	
Total Extractable Hydrocarbons	A	mg/L		16.35379		15	0	0	0.0749	0.3	50	109%	80	120	0%	
o-Terphenyl	S	mg/L		0.2145115		0.2	0	0	0.000429	0.002	0	107%	80	120	0%	

Write Sequence	Data File	Sample Name	Method	Weight	Dil Factor	Amt Inj.	IS	Cal ID
		Insert Entries/Have the first cell for entries selected						
		Write Sequence						
		Data File						
10rgHP5.DAT	HP5122121_b11221HP5.011	DCM-Baseline Check-V01	G:\Org\HP5-Methods\DR_8015-IBB-LEXP.met	1	1	1	1	0
10rgHP5.DAT	HP5122121_b11221HP5.022	DCM-Baseline Check-V02	G:\Org\HP5-Methods\DR_8015-IBB-LEXP.met	1	1	1	1	0
10rgHP5.DAT	HP5122121_b11221HP5.033	MARKER_1220HP503r, C40 ;1220HP5 , DRO211207A	G:\Org\HP5-Methods\CSC211221.met	1	1	1	1	0
10rgHP5.DAT	HP5122121_b11221HP5.044	CCV_1221HP504r, RRO ;1221HP5 , DRO211220C	G:\Org\HP5-Methods\DC_ORO-AK-L%.MET G:\Org\HP5-Methods\DS_ORO-AK-L%.MET	1	1	1	1	0
10rgHP5.DAT	HP5122121_b11221HP5.055	CCV_1221HP505r, DRO ;1221HP5 , DRO211220A	G:\Org\HP5-Methods\DC_8015-24-IK-L%.met G:\Org\HP5-Methods\DS_8015-24-IK-L%.met	1	1	1	1	0
10rgHP5.DAT	HP5122121_b11221HP5.066	DCM-Baseline Check-V06	G:\Org\HP5-Methods\DR_8015-IBB-LEXP.met	1	1	1	1	0
10rgHP5.DAT	HP5122121_b11221HP5.077	LCS-162352 ;1221HP5 ,	G:\Org\HP5-Methods\DR_8015-IBB-LEXP.met G:\Org\HP5-Methods\DS_8015-24-IK-L%.met	1000	1	1	1	0
10rgHP5.DAT	HP5122121_b11221HP5.088	MB-162352 ;1221HP5 ,	G:\Org\HP5-Methods\DR_8015-C24T-IK-L%.met G:\Org\HP5-Methods\DR_OROS-AK-L%.MET G:\Org\HP5-Methods\DS_8015-C24T-IK-L%.met	1000	1	1	1	0
10rgHP5.DAT	HP5122121_b11221HP5.099	B21121613-001B ;1221HP5 , \$HC-8015-DRO-W,	G:\Org\HP5-Methods\DR_8015-C24T-IK-L%.met G:\Org\HP5-Methods\DR_OROS-AK-L%.MET G:\Org\HP5-Methods\DS_8015-C24T-IK-L%.met	990	1	1	1	0
10rgHP5.DAT	HP5122121_b11221HP5.100	B21121613-001BMS ;1221HP5 ,	G:\Org\HP5-Methods\DR_8015-IBB-LEXP.met G:\Org\HP5-Methods\DS_8015-24-IK-L%.met	990	1	1	1	0
10rgHP5.DAT	HP5122121_b11221HP5.111	B21121613-001BMSD ;1221HP5 ,	G:\Org\HP5-Methods\DR_8015-IBB-LEXP.met G:\Org\HP5-Methods\DS_8015-24-IK-L%.met	1040	1	1	1	0
10rgHP5.DAT	HP5122121_b11221HP5.122	DCM-Baseline Check-V12	G:\Org\HP5-Methods\DR_8015-IBB-LEXP.met	1	1	1	1	0
10rgHP5.DAT	HP5122121_b11221HP5.133	B21121613-002B ;1221HP5 , \$HC-8015-DRO-W,	G:\Org\HP5-Methods\DR_8015-C24T-IK-L%.met G:\Org\HP5-Methods\DR_OROS-AK-L%.MET G:\Org\HP5-Methods\DS_8015-C24T-IK-L%.met	1020	1	1	1	0
10rgHP5.DAT	HP5122121_b11221HP5.144	DCM-Baseline Check-V14	G:\Org\HP5-Methods\DR_8015-IBB-LEXP.met	1	1	1	1	0
10rgHP5.DAT	HP5122121_b11221HP5.155	B21121609-001C ;1221HP5 , \$HC-8015-DRO-W,	G:\Org\HP5-Methods\DR_8015-C24T-IK-L%.met G:\Org\HP5-Methods\DR_OROS-AK-L%.MET G:\Org\HP5-Methods\DS_8015-122115-IK-L%.met	1060	1	1	1	0
10rgHP5.DAT	HP5122121_b11221HP5.166	DCM-Baseline Check-V16	G:\Org\HP5-Methods\DR_8015-IBB-LEXP.met	1	1	1	1	0
10rgHP5.DAT	HP5122121_b11221HP5.177	B21121611-001B ;1221HP5 , \$HC-8015-DRO-W,	G:\Org\HP5-Methods\DR_8015-C24T-IK-L%.met G:\Org\HP5-Methods\DR_OROS-AK-L%.MET G:\Org\HP5-Methods\DS_8015-122117-IK-L%.met	1030	1	1	1	0
10rgHP5.DAT	HP5122121_b11221HP5.188	B21121616-001C ;1221HP5 , \$HC-8015-DRO-W,	G:\Org\HP5-Methods\DR_8015-IBB-LEXP.met G:\Org\HP5-Methods\DS_8015-24-IK-L%.met	990	1	1	1	0
10rgHP5.DAT	HP5122121_b11221HP5.199	MARKER_1220HP519r, C40 ;1220HP5 , DRO211207A	G:\Org\HP5-Methods\CSC211221.met	1	1	1	1	0
10rgHP5.DAT	HP5122121_b11221HP5.200	CCV_1221HP520r, RRO ;1221HP5 , DRO211220C	G:\Org\HP5-Methods\DC_ORO-AK-L%.MET G:\Org\HP5-Methods\DS_ORO-AK-L%.MET	1	1	1	1	0
10rgHP5.DAT	HP5122121_b11221HP5.211	CCV_1221HP521r, DRO ;1221HP5 , DRO211220A	G:\Org\HP5-Methods\DC_8015-24-IK-L%.met G:\Org\HP5-Methods\DS_8015-24-IK-L%.met	1	1	1	1	0
10rgHP5.DAT	HP5122121_b11221HP5.222	DCM-Baseline Check-V22	G:\Org\HP5-Methods\DR_8015-IBB-LEXP.met	1	1	1	1	0
10rgHP5.DAT	HP5122121_b11221HP5.233	B21121623-001C ;1221HP5 , \$HC-8015-DRO-W,	G:\Org\HP5-Methods\DR_8015-122123-IK-L%.met G:\Org\HP5-Methods\DR_OROS-122123-AK-L%.MET G:\Org\HP5-Methods\DS_8015-C24T-IK-L%.met	1020	1	1	1	0
10rgHP5.DAT	HP5122121_b11221HP5.244	DCM-Baseline Check-V24	G:\Org\HP5-Methods\DR_8015-IBB-LEXP.met	1	1	1	1	0
10rgHP5.DAT	HP5122121_b11221HP5.255	B21121605-001A ;1221HP5 , \$HC-8015-DRO-W,	G:\Org\HP5-Methods\DR_8015-122123-IK-L%.met G:\Org\HP5-Methods\DR_OROS-122123-AK-L%.MET G:\Org\HP5-Methods\DS_8015-C24T-IK-L%.met	1050	1	1	1	0
10rgHP5.DAT	HP5122121_b11221HP5.266	DCM-Baseline Check-V26	G:\Org\HP5-Methods\DR_8015-IBB-LEXP.met	1	1	1	1	0
10rgHP5.DAT	HP5122121_b11221HP5.277	B21121605-002A ;1221HP5 , \$HC-8015-DRO-W,	G:\Org\HP5-Methods\DR_8015-C24T-IK-L%.met G:\Org\HP5-Methods\DR_OROS-AK-L%.MET G:\Org\HP5-Methods\DS_8015-122127-IK-L%.met	1050	1	1	1	0
10rgHP5.DAT	HP5122121_b11221HP5.288	B21121605-003A ;1221HP5 , \$HC-8015-DRO-W,	G:\Org\HP5-Methods\DR_8015-IBB-LEXP.met G:\Org\HP5-Methods\DS_8015-24-IK-L%.met	1010	1	1	1	0
10rgHP5.DAT	HP5122121_b11221HP5.299	DCM-Baseline Check-V29	G:\Org\HP5-Methods\DR_8015-IBB-LEXP.met	1	1	1	1	0
10rgHP5.DAT	HP5122121_b11221HP5.300	B21121622-001B ;1221HP5 , \$HC-8015-DRO-W, due to baseline	G:\Org\HP5-Methods\DR_8015-IBB-LEXP.met	960	1	1	1	0
10rgHP5.DAT	HP5122121_b11221HP5.311	B21121622-002B ;1221HP5 , \$HC-8015-DRO-W,	G:\Org\HP5-Methods\DR_8015-IBB-LEXP.met G:\Org\HP5-Methods\DS_8015-122131-IK-L%.met	1050	1	1	1	0
10rgHP5.DAT	HP5122121_b11221HP5.322	B21121622-003B ;1221HP5 , \$HC-8015-DRO-W, need rerun to verify no carryover	G:\Org\HP5-Methods\DR_8015-C24T-IJ-L0.met	900	1	1	1	0
10rgHP5.DAT	HP5122121_b11221HP5.333	MARKER_1220HP535r, C40 ;1220HP5 , DRO211207A	G:\Org\HP5-Methods\CSC211221.met	1	1	1	1	0
10rgHP5.DAT	HP5122121_b11221HP5.344	CCV_1221HP536r, RRO ;1221HP5 , DRO211220C	G:\Org\HP5-Methods\DC_ORO-AK-L%.MET G:\Org\HP5-Methods\DS_ORO-AK-L%.MET	1	1	1	1	0
10rgHP5.DAT	HP5122121_b11221HP5.355	CCV_1221HP537r, DRO ;1221HP5 , DRO211220A	G:\Org\HP5-Methods\DC_8015-24-IK-L%.met G:\Org\HP5-Methods\DS_8015-24-IK-L%.met	1	1	1	1	0
10rgHP5.DAT	HP5122121_b11221HP5.366	DCM-Baseline Check-V36	G:\Org\HP5-Methods\DR_8015-IBB-LEXP.met	1	1	1	1	0
10rgHP5.DAT	HP5122121_b11221HP5.377	DCM-Baseline Check-V37	G:\Org\HP5-Methods\DR_8015-IBB-LEXP.met	1	1	1	1	0
10rgHP5.DAT	HP5122121_b11221HP5.388	B21121622-001B ;1221HP5 , \$HC-8015-DRO-W, rr	G:\Org\HP5-Methods\DR_8015-122138-IK-L%.met G:\Org\HP5-Methods\DR_OROS-122138-AK-L%.MET G:\Org\HP5-Methods\DS_8015-C24T-IK-L%.met	960	1	1	1	0
10rgHP5.DAT	HP5122121_b11221HP5.399	B21121622-003B ;1221HP5 , \$HC-8015-DRO-W, rr	G:\Org\HP5-Methods\DR_8015-C24T-IK-L%.met G:\Org\HP5-Methods\DR_OROS-AK-L%.MET G:\Org\HP5-Methods\DS_8015-C24T-IK-L%.met	900	1	1	1	0
10rgHP5.DAT	HP5122121_b11221HP5.400	DCM-Baseline Check-V40	G:\Org\HP5-Methods\DR_8015-IBB-LEXP.met	1	1	1	1	0
10rgHP5.DAT	HP5122121_b11221HP5.411	B21121613-001BMS-RRO ;1221HP5 ,	G:\Org\HP5-Methods\DR_8015-C24T-IK-L%.met G:\Org\HP5-Methods\DR_ORO-AK-L%.MET	1045	1	1	1	0
10rgHP5.DAT	HP5122121_b11221HP5.422	DCM-Baseline Check-V42	G:\Org\HP5-Methods\DR_8015-IBB-LEXP.met	1	1	1	1	0
10rgHP5.DAT	HP5122121_b11221HP5.433	B21121613-001BMSD-RRO ;1221HP5 ,	G:\Org\HP5-Methods\DR_8015-IBB-LEXP.met G:\Org\HP5-Methods\DS_ORO-AK-L%.MET	1055	1	1	1	0
10rgHP5.DAT	HP5122121_b11221HP5.444	LCS-162352-RRO ;1221HP5 ,	G:\Org\HP5-Methods\DR_8015-IBB-LEXP.met G:\Org\HP5-Methods\DS_ORO-AK-L%.MET	1000	1	1	1	0
10rgHP5.DAT	HP5122121_b11221HP5.455	MARKER_1220HP546r, C40 ;1220HP5 , DRO211207A	G:\Org\HP5-Methods\CSC211221.met	1	1	1	1	0
10rgHP5.DAT	HP5122121_b11221HP5.466	CCV_1221HP547r, RRO ;1221HP5 , DRO211220C	G:\Org\HP5-Methods\DC_ORO-AK-L%.MET G:\Org\HP5-Methods\DS_ORO-AK-L%.MET	1	1	1	1	0
10rgHP5.DAT	HP5122121_b11221HP5.477	CCV_1221HP548r, DRO ;1221HP5 , DRO211220A	G:\Org\HP5-Methods\DC_8015-24-IK-L%.met G:\Org\HP5-Methods\DS_8015-24-IK-L%.met	1	1	1	1	0

Write Sequence	Data File	Sample Name	Method	Weight	Dil Factor	Amt Inj.	IS	Cal ID
		DCM-Baseline Check-V01	G:\Org\HP5\Methods\DR_8015-IBB-L-EXP.met	1	1	1	1	0
		DCM-Baseline Check-V02	G:\Org\HP5\Methods\DR_8015-IBB-L-EXP.met	1	1	1	1	0
		MARKER_1226HP503r, C40 ;1226HP5 , DRO211207A	G:\Org\HP5\Methods\CSC211226.met	1	1	1	1	0
		CCV_1226HP504r, RRO ;1226HP5 , DRO211220C	G:\Org\HP5\Methods\DC_ORO-AL-L%.MET G:\Org\HP5\Methods\DS_ORO-AL-L%.MET	1	1	1	1	0
		CCV_1226HP505r, DRO ;1226HP5 , DRO211220A	G:\Org\HP5\Methods\DC_8015-24-IL-L%.met G:\Org\HP5\Methods\DS_8015-24-IL-L%.met	1	1	1	1	0
		DCM-Baseline Check-V06	G:\Org\HP5\Methods\DR_8015-IBB-L-EXP.met	1	1	1	1	0
		LCS-162352 ;1226HP5 , Needs RR GC vial cap was loose and sample may have concentrated.	G:\Org\HP5\Methods\D3_8015-24-IL-L0.met	1000	1	1	1	0
		MB-162352 ;1226HP5 , SGT	G:\Org\HP5\Methods\DR_8015-C24T-IL-L%.met G:\Org\HP5\Methods\DR_OROS-AL-L%.MET G:\Org\HP5\Methods\DS_8015-C24T-IL-L%.met	1000	1	1	1	0
		B21121613-001B ;1226HP5 , \$HC-8015-DRO-W, SGT	G:\Org\HP5\Methods\DR_8015-C24T-IL-L%.met G:\Org\HP5\Methods\DR_OROS-AL-L%.MET G:\Org\HP5\Methods\DS_8015-C24T-IL-L%.met	990	1	1	1	0
		B21121613-001BMS ;1226HP5 , SGT	G:\Org\HP5\Methods\D3_8015-24-IL-L%.met G:\Org\HP5\Methods\DS_8015-24-IL-L%.met	990	1	1	1	0
		B21121613-001BMSD ;1226HP5 , SGT	G:\Org\HP5\Methods\D3_8015-24-IL-L%.met G:\Org\HP5\Methods\DS_8015-24-IL-L%.met	1040	1	1	1	0
		DCM-Baseline Check-V12	G:\Org\HP5\Methods\DR_8015-IBB-L-EXP.met	1	1	1	1	0
		B21121622-001B ;1226HP5 , \$HC-8015-DRO-W, SGT	G:\Org\HP5\Methods\DR_8015-C24T-IL-L%.met G:\Org\HP5\Methods\DR_OROS-AL-L%.MET G:\Org\HP5\Methods\DS_8015-C24T-IL-L%.met	960	1	1	1	0
		B21121605-001A ;1226HP5 , \$HC-8015-DRO-W, SGT	G:\Org\HP5\Methods\DR_8015-C24T-IL-L%.met G:\Org\HP5\Methods\DR_OROS-AL-L%.MET G:\Org\HP5\Methods\DS_8015-C24T-IL-L%.met	1050	1	1	1	0
		B21121623-001C ;1226HP5 , \$HC-8015-DRO-W, SGT	G:\Org\HP5\Methods\DR_8015-C24T-IL-L%.met G:\Org\HP5\Methods\DR_OROS-AL-L%.MET G:\Org\HP5\Methods\DS_8015-C24T-IL-L%.met	1020	1	1	1	0
		DCM-Baseline Check-V16	G:\Org\HP5\Methods\DR_8015-IBB-L-EXP.met	1	1	1	1	0
		B21121611-001B ;1226HP5 , \$HC-8015-DRO-W, SGT	G:\Org\HP5\Methods\DR_8015-C24T-IL-L%.met G:\Org\HP5\Methods\DR_OROS-AL-L%.MET G:\Org\HP5\Methods\DS_8015-C24T-IL-L%.met	1030	1	1	1	0
		B21121622-002B ;1226HP5 , \$HC-8015-DRO-W, SGT	G:\Org\HP5\Methods\DR_8015-C24T-IL-L%.met G:\Org\HP5\Methods\DR_OROS-AL-L%.MET G:\Org\HP5\Methods\DS_8015-C24T-IL-L%.met	1050	1	1	1	0
		MARKER_1226HP503r, C40 ;1226HP5 , DRO211207A	G:\Org\HP5\Methods\CSC211226.met	1	1	1	1	0
		CCV_1226HP504r, RRO ;1226HP5 , DRO211220C	G:\Org\HP5\Methods\DC_ORO-AL-L%.MET G:\Org\HP5\Methods\DS_ORO-AL-L%.MET	1	1	1	1	0
		CCV_1226HP505r, DRO ;1226HP5 , DRO211220A	G:\Org\HP5\Methods\DC_8015-24-IL-L%.met G:\Org\HP5\Methods\DS_8015-24-IL-L%.met	1	1	1	1	0
		DCM-Baseline Check-V22	G:\Org\HP5\Methods\DR_8015-IBB-L-EXP.met	1	1	1	1	0
		DCM-Baseline Check-V23	G:\Org\HP5\Methods\DR_8015-IBB-L-EXP.met	1	1	1	1	0
		B21121616-001C ;1226HP5 , \$HC-8015-DRO-W, SGT	G:\Org\HP5\Methods\DR_8015-C24T-IL-L%.met G:\Org\HP5\Methods\DR_OROS-AL-L%.MET G:\Org\HP5\Methods\DS_8015-C24T-IL-L%.met	990	1	1	1	0
		B21121605-002A ;1226HP5 , \$HC-8015-DRO-W, SGT	G:\Org\HP5\Methods\D3_8015-C24T-IL-L%.met G:\Org\HP5\Methods\D3_OROS-AL-L%.MET G:\Org\HP5\Methods\DS_8015-C24T-IL-L%.met	1050	1	1	1	0
		B21121605-003A ;1226HP5 , \$HC-8015-DRO-W, SGT	G:\Org\HP5\Methods\D3_8015-122626-IL-L%.met G:\Org\HP5\Methods\D3_OROS-122626-AL-L%.MET G:\Org\HP5\Methods\DS_8015-C24T-IL-L%.met	1010	1	1	1	0
		B21121609-001C ;1226HP5 , \$HC-8015-DRO-W, SGT	G:\Org\HP5\Methods\D3_8015-122627-IL-L%.met G:\Org\HP5\Methods\D3_OROS-122627-AL-L%.MET G:\Org\HP5\Methods\DS_8015-C24T-IL-L%.met	1060	1	1	1	0
		B21121613-001BMS-RRO ;1226HP5 , SGT	G:\Org\HP5\Methods\D3_ORO-AL-L%.MET G:\Org\HP5\Methods\DS_ORO-AL-L%.MET	1045	1	1	1	0
		DCM-Baseline Check-V29	G:\Org\HP5\Methods\DR_8015-IBB-L-EXP.met	1	1	1	1	0
		LCS-162352 ;1226HP5 , RR-SGT	G:\Org\HP5\Methods\D3_8015-122630-IL-L%.met G:\Org\HP5\Methods\DS_8015-24-IL-L%.met	1000	1	1	1	0
		B21121613-001BMSD-RRO ;1226HP5 , SGT	G:\Org\HP5\Methods\D3_ORO-AL-L%.MET G:\Org\HP5\Methods\DS_ORO-AL-L%.MET	1055	1	1	1	0
		LCS-162352-RRO ;1226HP5 , NEEDS RR	G:\Org\HP5\Methods\DR_8015-IBB-L-EXP.met	1000	1	1	1	0
		MARKER_1226HP533r, C40 ;1226HP5 , DRO211207A	G:\Org\HP5\Methods\CSC211226.met	1	1	1	1	0
		CCV_1226HP534r, RRO ;1226HP5 , DRO211220C	G:\Org\HP5\Methods\DC_ORO-AL-L%.MET G:\Org\HP5\Methods\DS_ORO-AL-L%.MET	1	1	1	1	0
		CCV_1226HP535r, DRO ;1226HP5 , DRO211220A	G:\Org\HP5\Methods\DC_8015-24-IL-L%.met G:\Org\HP5\Methods\DS_8015-24-IL-L%.met	1	1	1	1	0
		DCM-Baseline Check-V36	G:\Org\HP5\Methods\DR_8015-IBB-L-EXP.met	1	1	1	1	0
		MARKER_1226HP549r, C40 ;1226HP5 , DRO211207A	G:\Org\HP5\Methods\CSC211226.met	1	1	1	1	0
		CCV_1226HP550r, RRO ;1226HP5 , DRO211220C	G:\Org\HP5\Methods\DC_ORO-AL-L%.MET G:\Org\HP5\Methods\DS_ORO-AL-L%.MET	1	1	1	1	0
		CCV_1226HP551r, DRO ;1226HP5 , DRO211220A	G:\Org\HP5\Methods\DC_8015-24-IL-L%.met G:\Org\HP5\Methods\DS_8015-24-IL-L%.met	1	1	1	1	0
		DCM-Baseline Check-V52	G:\Org\HP5\Methods\DR_8015-IBB-L-EXP.met	1	1	1	1	0
		DCM-Baseline Check-V56	G:\Org\HP5\Methods\DR_8015-IBB-L-EXP.met	1	1	1	1	0
		LCS-162352-RRO ;1226HP5 , RR-SGT	G:\Org\HP5\Methods\D3_ORO-AL-L%.MET G:\Org\HP5\Methods\DS_ORO-AL-L%.MET	1000	1	1	1	0
		MARKER_1226HP560r, C40 ;1226HP5 , DRO211207A	G:\Org\HP5\Methods\CSC211226.met	1	1	1	1	0
		CCV_1226HP561r, RRO ;1226HP5 , DRO211220C	G:\Org\HP5\Methods\DC_ORO-AL-L%.MET G:\Org\HP5\Methods\DS_ORO-AL-L%.MET	1	1	1	1	0




DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: DCM-Baseline Check-V01
 Raw File: G:\org\HP5\DAT\HP5122121_b\1221HP5.0001.RAW
 Date & Time Acquired: 12/21/2021 2:17:37 PM
 Method File: G:\Org\HP5\Methods\DR_8015-IBb-LEXP.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO211102IB.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for TEH: 31353.19
 Rt range for Diesel Range Organics: 6.54 to 15.86

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*o-Terphenyl	12.259	200.	.055	.03	-
*1-Chlorooctadecane	13.076	200.	.134	.07	-

DRO Area:1009653 DRO Amount: 32.20255
 TEH Area:1709678 TEH Amount: 54.52963




DIESEL RANGE ORGANICS CHROMATOGRAM REPORT


Sample Name: DCM-Baseline Check-V02
 Raw File: G:\org\HP5\DAT\HP5122121_b\1221HP5.0002.RAW
 Date & Time Acquired: 12/21/2021 3:00:29 PM
 Method File: G:\Org\HP5\Methods\DR_8015-IBb-LEXP.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO211102IB.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for TEH: 31353.19
 Rt range for Diesel Range Organics: 6.54 to 15.86

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*o-Terphenyl	29.964	200.	.	-
*1-Chlorooctadecane	13.074	200.	.027	.01

DRO Area:124973.6 DRO Amount: 3.985994
 TEH Area:294331.7 TEH Amount: 9.387614





RESIDUAL RANGE ORGANICS CHROMATOGRAM

Sample Name: CCV_1221HP504r, RRO ;1221HP5 , DRO211220C
 Raw File: G:\org\HP5\DAT\HP5122121_b\1221HP5.0004.RAW
 Date & Time Acquired: 12/21/2021 4:25:45 PM
 Method File: G:\Org\HP5\Methods\DC_ORO-AK-L%.MET
 Calibration File: G:\Org\HP5\Cals\SW8015C_ORO211017AK.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for for ~~Residual~~ TEH (Oil Range) Organics Calculations: 28542.41
 Rt range for ~~Residual~~ TEH (Oil Range) Organics: 12.58 to 30.05

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*#Triacontane	16.304	500.	348.356	69.67	-

RRO TEH (Oil Range) Area:1.352419E+08 RRO TEH (Oil Range) AMOUNT: 4738.277

CONTINUING CALIBRATION REPORT: G:\org\HP5\DAT\HP5122121_b\1221HP5.0004.RAW


COMPOUND	ACTUAL (NG)	MEASURED (NG)	%RECOVERY	LIMITS
*30-40 Motor Oil	5000.	.065	.	75-125

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	LIMITS
*#Triacontane	16.304	200.	348.356	174.18	75-125

AMN 01/11/2022

G:\org\HP5\DAT\HP5122121_b\1221HP5.0004.RAW

CCV_1221HP504r, RRO ;1221HP5 , DRO211220C



RESIDUAL RANGE ORGANICS CHROMATOGRAM

Sample Name: CCV_1221HP504r, RRO ;1221HP5 , DRO211220C
Raw File: G:\org\HP5\DAT\HP5122121_b\1221HP5.0004.RAW
Date & Time Acquired: 12/21/2021 4:25:45 PM
Method File: G:\Org\HP5\Methods\DS_ORO-AK-L%.MET
Calibration File: G:\Org\HP5\Cals\SW8015C_ORO211017AK.CAL
Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for for Residual Range Organics Calculations: 28542.41
Rt range for Residual Range Organics: 12.58 to 30.05


SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*#Triacontane	16.304	500.	213.87	42.77	-

RRO Area:6404931 RRO AMOUNT: 224.4005

CONTINUING CALIBRATION REPORT: G:\org\HP5\DAT\HP5122121_b\1221HP5.0004.RAW

COMPOUND	ACTUAL (NG)	MEASURED (NG)	%RECOVERY	LIMITS
*30-40 Motor Oil	5000.	.065	.	75-125

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	LIMITS
*#Triacontane	16.304	200.	213.87	106.94	75-125



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: CCV_1221HP505r, DRO ;1221HP5 , DRO211220A
 Raw File: G:\org\HP5\DAT\HP5122121_b\1221HP5.0005.RAW
 Date & Time Acquired: 12/21/2021 5:08:55 PM
 Method File: G:\Org\HP5\Methods\DC_8015-24-IK-L%.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO211102Ik-24.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for TEH: 31353.19

Rt range for Diesel Range Organics: 6.51 to 14.38

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*o-Terphenyl	12.222	200.	325.217	162.61
*1-Chlorooctadecane	13.066	200.	155.538	77.77

DRO Area: 4.600201E+08 DRO Amount: 14672.19
 TEH Area: 4.772049E+08 TEH Amount: 15220.3


CONTINUING CALIBRATION REPORT: G:\org\HP5\DAT\HP5122121_b\1221HP5.0005.RAW

COMPOUND	ACTUAL (NG)	MEASURED (NG)	%RECOVERY	LIMITS
TOTAL DRO	15000.	15220.3	101.47	85-115

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	LIMITS
*o-Terphenyl	12.222	200.	325.217	162.61	85-115
*1-Chlorooctadecane	13.066	200.	155.538	77.77	85-115

G:\org\HP5\DAT\HP5122121_b\1221HP5.0005.RAW

CCV_1221HP505r, DRO ;1221HP5 , DRO211220A



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: CCV_1221HP505r, DRO ;1221HP5 , DRO211220A
 Raw File: G:\org\HP5\DAT\HP5122121_b\1221HP5.0005.RAW
 Date & Time Acquired: 12/21/2021 5:08:55 PM
 Method File: G:\Org\HP5\Methods\DS_8015-24-IK-L#.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO211102Ik-24.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for TEH: 31353.19

Rt range for Diesel Range Organics: 6.51 to 14.38


SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*o-Terphenyl	12.222	200.	196.123	98.06
*1-Chlorooctadecane	13.066	200.	36.45	18.23

DRO Area: 2.553877E+08 DRO Amount: 8145.508
 TEH Area: 2.661105E+08 TEH Amount: 8487.508

CONTINUING CALIBRATION REPORT: G:\org\HP5\DAT\HP5122121_b\1221HP5.0005.RAW

COMPOUND	ACTUAL (NG)	MEASURED (NG)	%RECOVERY	LIMITS
TOTAL DRO	15000.	8487.51	56.58	85-115

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	LIMITS
*o-Terphenyl	12.222	200.	196.123	98.06	85-115
*1-Chlorooctadecane	13.066	200.	36.45	18.23	85-115



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: DCM-Baseline Check-V06
 Raw File: G:\org\HP5\DAT\HP5122121_b\1221HP5.0006.RAW
 Date & Time Acquired: 12/21/2021 5:52:10 PM
 Method File: G:\Org\HP5\Methods\DR_8015-IBb-LEXP.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO211102IB.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for TEH: 31353.19
 Rt range for Diesel Range Organics: 6.54 to 15.86


SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*o-Terphenyl	29.974	200.	.	-
*1-Chlorooctadecane	13.064	200.	.019	.01 -

DRO Area:145099.7 DRO Amount: 4.627907
 TEH Area:284114.4 TEH Amount: 9.061737

Batch ID: 162352

LCS-162352 ;1221HP5 ,

G:\org\HP5\DAT\HP5122121_b\1221HP5.0007.RAW



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: LCS-162352 ;1221HP5 ,
 Raw File: G:\org\HP5\DAT\HP5122121_b\1221HP5.0007.RAW
 Date & Time Acquired: 12/21/2021 6:35:14 PM
 Method File: G:\Org\HP5\Methods\D3_8015-24-IK-L%.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO211102IK-24.CAL
 Sample Weight: 1000 Dilution: 1 S.A.: 1

Mean RF for TEH: 31353.19

Rt range for Diesel Range Organics: 6.51 to 14.38


SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*o-Terphenyl	12.222	.2	.324	161.91	-
*1-Chlorooctadecane	13.067	.2	.152	76.05	-

DRO Area: 4.174475E+08 DRO Amount: 13.31435
 TEH Area: 4.471555E+08 TEH Amount: 14.26188

Batch ID: 162352

G:\Org\HP5\DAT\HP5122121_b\1221HP5.0007.RAW

LCS-162352 ;1221HP5 ,



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT


Sample Name: LCS-162352 ;1221HP5 ,
 Raw File: G:\Org\HP5\DAT\HP5122121_b\1221HP5.0007.RAW
 Date & Time Acquired: 12/21/2021 6:35:14 PM
 Method File: G:\Org\HP5\Methods\DS_8015-24-IK-L#.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO211102Ik-24.CAL
 Sample Weight: 1000 Dilution: 1 S.A.: 1

Mean RF for TEH: 31353.19

Rt range for Diesel Range Organics: 6.51 to 14.38

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*o-Terphenyl	12.222	.2	.187	93.65
*1-Chlorooctadecane	13.067	.2	.064	32.08

DRO Area: 2.060904E+08 DRO Amount: 6.573185
 TEH Area: 2.205345E+08 TEH Amount: 7.033879



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT


Sample Name: MB-162352 ;1221HP5 ,
 Raw File: G:\org\HP5\DAT\HP5122121_b\1221HP5.0008.RAW
 Date & Time Acquired: 12/21/2021 7:18:27 PM
 Method File: G:\Org\HP5\Methods\DR_8015-C24T-IK-L%.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO211102IK-24-Tri.CAL
 Sample Weight: 1000 Dilution: 1 S.A.: 1

Mean RF for TEH: 31353.19

Rt range for Diesel Range Organics: 6.51 to 14.38

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*o-Terphenyl	12.215	.2	.183	91.66	-
*1-Chlorooctadecane	13.059	.2	.04		-
*#Triacontane	16.293	.2	.095	47.54	-

DRO Area:334103.8 DRO Amount: 1.065613E-02
 TEH Area:580763.4 TEH Amount: 1.852326E-02



RESIDUAL RANGE ORGANICS CHROMATOGRAM

Sample Name: MB-162352 ;1221HP5 ,
 Raw File: G:\org\HP5\DAT\HP5122121_b\1221HP5.0008.RAW
 Date & Time Acquired: 12/21/2021 7:18:27 PM
 Method File: G:\Org\HP5\Methods\DR_OROS-AK-L%.MET
 Calibration File: G:\Org\HP5\Cals\SW8015C_ORO211017AK-SAMP.CAL
 Sample Weight: 1000 Dilution: 1 S.A.: 1

Mean RF for for Residual Range Organics Calculations: 28542.41
 Rt range for Residual Range Organics: 14.29 to 21.62


SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*#Triacontane	16.293	.5	.095	19.02

RRO Area:141137.4 RRO AMOUNT: 4.944833E-03

Batch ID: 162352

G:\org\HP5\DAT\HP5122121_b\1221HP5.0008.RAW

MB-162352 ;1221HP5 ,



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: MB-162352 ;1221HP5 ,
Raw File: G:\org\HP5\DAT\HP5122121_b\1221HP5.0008.RAW
Date & Time Acquired: 12/21/2021 7:18:27 PM
Method File: G:\Org\HP5\Methods\DS_8015-C24T-IK-L#.met
Calibration File: G:\Org\HP5\Cals\SW8015C_DRO211102IK-24-Tri.CAL
Sample Weight: 1000 Dilution: 1 S.A.: 1

Mean RF for TEH: 31353.19

Rt range for Diesel Range Organics: 6.51 to 14.38

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*o-Terphenyl	12.215	.2	.183	91.5
*1-Chlorooctadecane	13.059	.2	.02	-
*#Triacontane	16.293	.2	.094	47.18


DRO Area:312933.7 DRO Amount: 9.98092E-03
TEH Area:2935838 TEH Amount: 9.363762E-02

ERH2184 (RHMW05)

Batch ID: 162352

G:\org\HP5\DAT\HP5122121_b\1221HP5.0009.RAW

B21121613-001B ;1221HP5 , \$HC-8015-DRO-W,



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: B21121613-001B ;1221HP5 , \$HC-8015-DRO-W,
Raw File: G:\org\HP5\DAT\HP5122121_b\1221HP5.0009.RAW
Date & Time Acquired: 12/21/2021 8:01:38 PM
Method File: G:\Org\HP5\Methods\DR_8015-C24T-IK-L%.met
Calibration File: G:\Org\HP5\Cals\SW8015C_DRO211102IK-24-Tri.CAL
Sample Weight: 990 Dilution: 1 S.A.: 1

Mean RF for TEH: 31353.19

Rt range for Diesel Range Organics: 6.51 to 14.38

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*o-Terphenyl	12.214	.202	.172	84.99	-
*1-Chlorooctadecane	13.063	.202	.	.04	-
*#Triacontane	16.292	.202	.093	46.03	-


DRO Area:587625.2 DRO Amount: 1.893143E-02
TEH Area:822057.5 TEH Amount: 0.0264841

ERH2184 (RHMW05)

Batch ID: 162352

G:\org\HP5\DAT\HP5122121_b\1221HP5.0009.RAW

B21121613-001B ;1221HP5 , \$HC-8015-DRO-W,



RESIDUAL RANGE ORGANICS CHROMATOGRAM

Sample Name: B21121613-001B ;1221HP5 , \$HC-8015-DRO-W,
Raw File: G:\org\HP5\DAT\HP5122121_b\1221HP5.0009.RAW
Date & Time Acquired: 12/21/2021 8:01:38 PM
Method File: G:\Org\HP5\Methods\DR_OROS-AK-L%.MET
Calibration File: G:\Org\HP5\Cals\SW8015C_ORO211017AK-SAMP.CAL
Sample Weight: 990 Dilution: 1 S.A.: 1

Mean RF for for Residual Range Organics Calculations: 28542.41
Rt range for Residual Range Organics: 14.29 to 21.62

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*#Triacontane	16.292	.505	.093	18.41


RRO Area:144972.6 RRO AMOUNT: 5.130504E-03

ERH2184 (RHMW05)

Batch ID: 162352

G:\Org\HP5\DAT\HP5122121_b\1221HP5.0009.RAW

B21121613-001B ;1221HP5 , \$HC-8015-DRO-W,



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: B21121613-001B ;1221HP5 , \$HC-8015-DRO-W,
Raw File: G:\Org\HP5\DAT\HP5122121_b\1221HP5.0009.RAW
Date & Time Acquired: 12/21/2021 8:01:38 PM
Method File: G:\Org\HP5\Methods\DS_8015-C24T-IK-L#.met
Calibration File: G:\Org\HP5\Cals\SW8015C_DRO211102IK-24-Tri.CAL
Sample Weight: 990 Dilution: 1 S.A.: 1

Mean RF for TEH: 31353.19

Rt range for Diesel Range Organics: 6.51 to 14.38


SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*o-Terphenyl	12.214	.202	.172	84.89	-
*1-Chlorooctadecane	13.063	.202	.	.03	-
*#Triacontane	16.292	.202	.092	45.69	-

DRO Area:547150.2 DRO Amount: 1.762745E-02
TEH Area:2952943 TEH Amount: 9.513452E-02

Batch ID: 162352

B21121613-001BMS ;1221HP5 ,

G:\org\HP5\DAT\HP5122121_b\1221HP5.0010.RAW



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: B21121613-001BMS ;1221HP5 ,
 Raw File: G:\org\HP5\DAT\HP5122121_b\1221HP5.0010.RAW
 Date & Time Acquired: 12/21/2021 8:44:53 PM
 Method File: G:\Org\HP5\Methods\D3_8015-24-IK-L%.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO211102IK-24.CAL
 Sample Weight: 990 Dilution: 1 S.A.: 1

Mean RF for TEH: 31353.19

Rt range for Diesel Range Organics: 6.51 to 14.38


SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*o-Terphenyl	12.22	.202	.341	168.93	-
*1-Chlorooctadecane	13.067	.202	.152	75.3	-

DRO Area: 4.286411E+08 DRO Amount: 13.80946
 TEH Area: 4.605626E+08 TEH Amount: 14.83788

Batch ID: 162352

G:\org\HP5\DAT\HP5122121_b\1221HP5.0010.RAW

B21121613-001BMS ;1221HP5 ,



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: B21121613-001BMS ;1221HP5 ,
 Raw File: G:\org\HP5\DAT\HP5122121_b\1221HP5.0010.RAW
 Date & Time Acquired: 12/21/2021 8:44:53 PM
 Method File: G:\Org\HP5\Methods\DS_8015-24-IK-L#.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO211102Ik-24.CAL
 Sample Weight: 990 Dilution: 1 S.A.: 1

Mean RF for TEH: 31353.19

Rt range for Diesel Range Organics: 6.51 to 14.38


SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*o-Terphenyl	12.22	.202	.184	90.9
*1-Chlorooctadecane	13.067	.202	.066	32.87

DRO Area: 2.076382E+08 DRO Amount: 6.689449
 TEH Area: 2.217393E+08 TEH Amount: 7.143741

Batch ID: 162352

B21121613-001BMSD ;1221HP5 ,

G:\org\HP5\DAT\HP5122121_b\1221HP5.0011.RAW



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: B21121613-001BMSD ;1221HP5 ,
 Raw File: G:\org\HP5\DAT\HP5122121_b\1221HP5.0011.RAW
 Date & Time Acquired: 12/21/2021 9:28:03 PM
 Method File: G:\Org\HP5\Methods\D3_8015-24-IK-L%.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO211102IK-24.CAL
 Sample Weight: 1040 Dilution: 1 S.A.: 1

Mean RF for TEH: 31353.19

Rt range for Diesel Range Organics: 6.51 to 14.38


SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*o-Terphenyl	12.218	.192	.305	158.7	-
*1-Chlorooctadecane	13.066	.192	.134	69.84	-

DRO Area: 3.981509E+08 DRO Amount: 12.21048
 TEH Area: 4.272252E+08 TEH Amount: 13.10213

Batch ID: 162352

B21121613-001BMSD ;1221HP5 ,

G:\org\HP5\DAT\HP5122121_b\1221HP5.0011.RAW



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT


Sample Name: B21121613-001BMSD ;1221HP5 ,
 Raw File: G:\org\HP5\DAT\HP5122121_b\1221HP5.0011.RAW
 Date & Time Acquired: 12/21/2021 9:28:03 PM
 Method File: G:\Org\HP5\Methods\DS_8015-24-IK-L#.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO211102Ik-24.CAL
 Sample Weight: 1040 Dilution: 1 S.A.: 1

Mean RF for TEH: 31353.19

Rt range for Diesel Range Organics: 6.51 to 14.38

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*o-Terphenyl	12.218	.192	.169	87.93
*1-Chlorooctadecane	13.066	.192	.06	31.3

DRO Area:1.900481E+08 DRO Amount: 5.828388
 TEH Area:2.029181E+08 TEH Amount: 6.223084



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: DCM-Baseline Check-V12
 Raw File: G:\org\HP5\DAT\HP5122121_b\1221HP5.0012.RAW
 Date & Time Acquired: 12/21/2021 10:11:15 PM
 Method File: G:\Org\HP5\Methods\DR_8015-IBb-LEXP.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO211102IB.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for TEH: 31353.19
 Rt range for Diesel Range Organics: 6.54 to 15.86

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*o-Terphenyl	12.212	200.	.022	.01
*1-Chlorooctadecane	13.075	200.	.019	.01


DRO Area:311077.7 DRO Amount: 9.921722
 TEH Area:484036.6 TEH Amount: 15.43819

ERH2201 (RHMW13)

Batch ID: 162352

G:\org\HP5\DAT\HP5122121_b\1221HP5.0013.RAW

B21121613-002B ;1221HP5 , \$HC-8015-DRO-W,



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: B21121613-002B ;1221HP5 , \$HC-8015-DRO-W,
Raw File: G:\org\HP5\DAT\HP5122121_b\1221HP5.0013.RAW
Date & Time Acquired: 12/21/2021 10:54:18 PM
Method File: G:\Org\HP5\Methods\DR_8015-C24T-IK-L%.met
Calibration File: G:\Org\HP5\Cals\SW8015C_DRO211102IK-24-Tri.CAL
Sample Weight: 1020 Dilution: 1 S.A.: 1

Mean RF for TEH: 31353.19

Rt range for Diesel Range Organics: 6.51 to 14.38

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*o-Terphenyl	12.213	.196	.166	84.86	-
*1-Chlorooctadecane	13.065	.196	.	.04	-
*#Triacontane	16.291	.196	.09	46.15	-


DRO Area:868001 DRO Amount: 2.714178E-02
TEH Area:1171759 TEH Amount: 3.664007E-02

ERH2201 (RHMW13)

Batch ID: 162352

G:\org\HP5\DAT\HP5122121_b\1221HP5.0013.RAW

B21121613-002B ;1221HP5 , \$HC-8015-DRO-W,



RESIDUAL RANGE ORGANICS CHROMATOGRAM

Sample Name: B21121613-002B ;1221HP5 , \$HC-8015-DRO-W,
Raw File: G:\org\HP5\DAT\HP5122121_b\1221HP5.0013.RAW
Date & Time Acquired: 12/21/2021 10:54:18 PM
Method File: G:\Org\HP5\Methods\DR_OROS-AK-L%.MET
Calibration File: G:\Org\HP5\Cals\SW8015C_ORO211017AK-SAMP.CAL
Sample Weight: 1020 Dilution: 1 S.A.: 1

Mean RF for for Residual Range Organics Calculations: 28542.41
Rt range for Residual Range Organics: 14.29 to 21.62

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*#Triacontane	16.291	.49	.09	18.46


RRO Area:213874.8 RRO AMOUNT: 7.346302E-03

ERH2201 (RHMW13)

Batch ID: 162352

G:\org\HP5\DAT\HP5122121_b\1221HP5.0013.RAW

B21121613-002B ;1221HP5 , \$HC-8015-DRO-W,



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT


Sample Name: B21121613-002B ;1221HP5 , \$HC-8015-DRO-W,
Raw File: G:\org\HP5\DAT\HP5122121_b\1221HP5.0013.RAW
Date & Time Acquired: 12/21/2021 10:54:18 PM
Method File: G:\Org\HP5\Methods\DS_8015-C24T-IK-L#.met
Calibration File: G:\Org\HP5\Cals\SW8015C_DRO211102IK-24-Tri.CAL
Sample Weight: 1020 Dilution: 1 S.A.: 1

Mean RF for TEH: 31353.19

Rt range for Diesel Range Organics: 6.51 to 14.38

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*o-Terphenyl	12.213	.196	.166	84.76	-
*1-Chlorooctadecane	13.065	.196	.	.04	-
*#Triacontane	16.291	.196	.09	45.82	-

DRO Area:790954 DRO Amount: 2.473257E-02
TEH Area:5707875 TEH Amount: 0.1784812



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: DCM-Baseline Check-V14
 Raw File: G:\org\HP5\DAT\HP5122121_b\1221HP5.0014.RAW
 Date & Time Acquired: 12/21/2021 11:37:27 PM
 Method File: G:\Org\HP5\Methods\DR_8015-IBb-LEXP.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO211102IB.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for TEH: 31353.19

Rt range for Diesel Range Organics: 6.54 to 15.86

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*o-Terphenyl	12.216	200.	.021	.01
*1-Chlorooctadecane	29.98	200.	.	.


DRO Area:259222 DRO Amount: 8.267801
 TEH Area:322606.3 TEH Amount: 10.28943

ERH2230 (Sump Adit 3)

G:\org\HP5\DAT\HP5122121_b\1221HP5.0015.RAW

Batch ID: 162352

B21121609-001C ; 1221HP5 , \$HC-8015-DRO-W,



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: B21121609-001C ; 1221HP5 , \$HC-8015-DRO-W,
Raw File: G:\org\HP5\DAT\HP5122121_b\1221HP5.0015.RAW
Date & Time Acquired: 12/22/2021 12:20:36 AM
Method File: G:\Org\HP5\Methods\D3_8015-C24T-IK-L%.met
Calibration File: G:\Org\HP5\Cals\SW8015C_DRO211102IK-24-Tri.CAL
Sample Weight: 1060 Dilution: 1 S.A.: 1

Mean RF for TEH: 31353.19

Rt range for Diesel Range Organics: 6.51 to 14.38

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*o-Terphenyl	12.213	.189	.17	90.31	-
*1-Chlorooctadecane	13.081	.189	.01	5.47	-
*#Triacontane	16.29	.189	.095	50.37	-

DRO Area: 1.547657E+08 DRO Amount: 4.656794


TEH Area: 1.707978E+08 TEH Amount: 5.139189

ERH2230 (Sump Adit 3)

Batch ID: 162352

G:\org\HP5\DAT\HP5122121_b\1221HP5.0015.RAW

B21121609-001C ; 1221HP5 , \$HC-8015-DRO-W,



RESIDUAL RANGE ORGANICS CHROMATOGRAM

Sample Name: B21121609-001C ; 1221HP5 , \$HC-8015-DRO-W,
Raw File: G:\org\HP5\DAT\HP5122121_b\1221HP5.0015.RAW
Date & Time Acquired: 12/22/2021 12:20:36 AM
Method File: G:\Org\HP5\Methods\D3_OROS-AK-L%.MET
Calibration File: G:\Org\HP5\Cals\SW8015C_ORO211017AK-SAMP.CAL
Sample Weight: 1060 Dilution: 1 S.A.: 1

Mean RF for for Residual Range Organics Calculations: 28542.41
Rt range for Residual Range Organics: 14.29 to 21.62

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*#Triacontane	16.29	.472	.095	20.15	-


RRO Area: 1.13758E+07 RRO AMOUNT: 0.3759978

ERH2230 (Sump Adit 3)

Batch ID: 162352

G:\org\HP5\DAT\HP5122121_b\1221HP5.0015.RAW

B21121609-001C ;1221HP5 , \$HC-8015-DRO-W,




DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: B21121609-001C ;1221HP5 , \$HC-8015-DRO-W,
Raw File: G:\org\HP5\DAT\HP5122121_b\1221HP5.0015.RAW
Date & Time Acquired: 12/22/2021 12:20:36 AM
Method File: G:\Org\HP5\Methods\DS_8015-122115-IK-L#.met
Calibration File: G:\Org\HP5\Cals\SW8015C_DRO211102IK-24-Tri.CAL
Sample Weight: 1060 Dilution: 1 S.A.: 1

Mean RF for TEH: 31353.19
Rt range for Diesel Range Organics: 6.51 to 14.38

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*o-Terphenyl	12.213	.189	.152	80.43	-
*1-Chlorooctadecane	13.081	.189	.	.22	-
*#Triacontane	16.29	.189	.084	44.67	-

DRO Area:1.245673E+08 DRO Amount: 3.748147
TEH Area:1.272026E+08 TEH Amount: 3.827441



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: DCM-Baseline Check-V16
 Raw File: G:\org\HP5\DAT\HP5122121_b\1221HP5.0016.RAW
 Date & Time Acquired: 12/22/2021 1:03:48 AM
 Method File: G:\Org\HP5\Methods\DR_8015-IBb-LEXP.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO211102IB.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for TEH: 31353.19
 Rt range for Diesel Range Organics: 6.54 to 15.86

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*o-Terphenyl	12.214	200.	.028	.01 -
*1-Chlorooctadecane	29.972	200.	.	. -


DRO Area:240423.9 DRO Amount: 7.668243
 TEH Area:282585.8 TEH Amount: 9.012982

ERH2182 (RHMW03)

G:\org\HP5\DAT\HP5122121_b\1221HP5.0017.RAW

Batch ID: 162352

B21121611-001B ;1221HP5 , \$HC-8015-DRO-W,



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: B21121611-001B ;1221HP5 , \$HC-8015-DRO-W,
Raw File: G:\org\HP5\DAT\HP5122121_b\1221HP5.0017.RAW
Date & Time Acquired: 12/22/2021 1:46:53 AM
Method File: G:\Org\HP5\Methods\D3_8015-C24T-IK-L%.met
Calibration File: G:\Org\HP5\Cals\SW8015C_DRO211102IK-24-Tri.CAL
Sample Weight: 1030 Dilution: 1 S.A.: 1

Mean RF for TEH: 31353.19

Rt range for Diesel Range Organics: 6.51 to 14.38

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*o-Terphenyl	12.213	.194	.176	90.51	-
*1-Chlorooctadecane	13.028	.194	.001	.76	-
*#Triacontane	16.289	.194	.109	56.12	-


DRO Area:5308837 DRO Amount: 0.1643919
TEH Area:1.946388E+07 TEH Amount: 0.6027129

ERH2182 (RHMW03)

Batch ID: 162352

G:\org\HP5\DAT\HP5122121_b\1221HP5.0017.RAW

B21121611-001B ;1221HP5 , \$HC-8015-DRO-W,



RESIDUAL RANGE ORGANICS CHROMATOGRAM

Sample Name: B21121611-001B ;1221HP5 , \$HC-8015-DRO-W,
Raw File: G:\org\HP5\DAT\HP5122121_b\1221HP5.0017.RAW
Date & Time Acquired: 12/22/2021 1:46:53 AM
Method File: G:\Org\HP5\Methods\D3_OROS-AK-L%.MET
Calibration File: G:\Org\HP5\Cals\SW8015C_ORO211017AK-SAMP.CAL
Sample Weight: 1030 Dilution: 1 S.A.: 1

Mean RF for for Residual Range Organics Calculations: 28542.41
Rt range for Residual Range Organics: 14.29 to 21.62

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*#Triacontane	16.289	.485	.109	22.45	-


RRO Area:8578406 RRO AMOUNT: 0.2917956

ERH2182 (RHMW03)

Batch ID: 162352

G:\org\HP5\DAT\HP5122121_b\1221HP5.0017.RAW

B21121611-001B ;1221HP5 , \$HC-8015-DRO-W,



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: B21121611-001B ;1221HP5 , \$HC-8015-DRO-W,
Raw File: G:\org\HP5\DAT\HP5122121_b\1221HP5.0017.RAW
Date & Time Acquired: 12/22/2021 1:46:53 AM
Method File: G:\Org\HP5\Methods\DS_8015-122117-IK-L#.met
Calibration File: G:\Org\HP5\Cals\SW8015C_DRO211102IK-24-Tri.CAL
Sample Weight: 1030 Dilution: 1 S.A.: 1

Mean RF for TEH: 31353.19

Rt range for Diesel Range Organics: 6.51 to 14.38

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*o-Terphenyl	12.213	.194	.173	89.1	-
*1-Chlorooctadecane	13.028	.194	.	.09	-
*#Triacontane	16.289	.194	.094	48.33	-


DRO Area:2406410 DRO Amount: 0.0745162
TEH Area:6452488 TEH Amount: 0.1998059

ERH2180 (RHMW02)

Batch ID: 162352

G:\org\HP5\DAT\HP5122121_b\1221HP5.0018.RAW

B21121616-001C ;1221HP5 , \$HC-8015-DRO-W,



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: B21121616-001C ;1221HP5 , \$HC-8015-DRO-W,
Raw File: G:\org\HP5\DAT\HP5122121_b\1221HP5.0018.RAW
Date & Time Acquired: 12/22/2021 2:30:01 AM
Method File: G:\Org\HP5\Methods\D3_8015-C24T-IK-L%.met
Calibration File: G:\Org\HP5\Cals\SW8015C_DRO211102IK-24-Tri.CAL
Sample Weight: 990 Dilution: 1 S.A.: 1

Mean RF for TEH: 31353.19

Rt range for Diesel Range Organics: 6.51 to 14.38

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*o-Terphenyl	12.214	.202	.178	88.08	-
*1-Chlorooctadecane	29.984	.202	.	.	-
*#Triacontane	16.289	.202	.111	55.1	-


DRO Area:3.772203E+07 DRO Amount: 1.215285
TEH Area:5.138606E+07 TEH Amount: 1.655497

ERH2180 (RHMW02)

Batch ID: 162352

G:\org\HP5\DAT\HP5122121_b\1221HP5.0018.RAW

B21121616-001C ;1221HP5 , \$HC-8015-DRO-W,



RESIDUAL RANGE ORGANICS CHROMATOGRAM

Sample Name: B21121616-001C ;1221HP5 , \$HC-8015-DRO-W,
Raw File: G:\org\HP5\DAT\HP5122121_b\1221HP5.0018.RAW
Date & Time Acquired: 12/22/2021 2:30:01 AM
Method File: G:\Org\HP5\Methods\D3_OROS-AK-L%.MET
Calibration File: G:\Org\HP5\Cals\SW8015C_ORO211017AK-SAMP.CAL
Sample Weight: 990 Dilution: 1 S.A.: 1

Mean RF for for Residual Range Organics Calculations: 28542.41
Rt range for Residual Range Organics: 14.29 to 21.62

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*#Triacontane	16.289	.505	.111	22.04


RRO Area:7434693 RRO AMOUNT: 0.2631099

ERH2180 (RHMW02)

Batch ID: 162352

G:\org\HP5\DAT\HP5122121_b\1221HP5.0018.RAW

B21121616-001C ;1221HP5 , \$HC-8015-DRO-W,



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: B21121616-001C ;1221HP5 , \$HC-8015-DRO-W,
Raw File: G:\org\HP5\DAT\HP5122121_b\1221HP5.0018.RAW
Date & Time Acquired: 12/22/2021 2:30:01 AM
Method File: G:\Org\HP5\Methods\DS_8015-122115-IK-L#.met
Calibration File: G:\Org\HP5\Cals\SW8015C_DRO211102IK-24-Tri.CAL
Sample Weight: 990 Dilution: 1 S.A.: 1


Mean RF for TEH: 31353.19


Rt range for Diesel Range Organics: 6.51 to 14.38

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*o-Terphenyl	12.214	.202	.152	75.42	-
*1-Chlorooctadecane	29.984	.202	.	.	-
*#Triacontane	16.289	.202	.1	49.51	-

DRO Area:1.264122E+07 DRO Amount: 0.4072601

TEH Area:1.548429E+07 TEH Amount: 0.4988549





RESIDUAL RANGE ORGANICS CHROMATOGRAM

Sample Name: CCV_1221HP520r, RRO ;1221HP5 , DRO211220C
 Raw File: G:\org\HP5\DAT\HP5122121_b\1221HP5.0020.RAW
 Date & Time Acquired: 12/22/2021 3:56:16 AM
 Method File: G:\Org\HP5\Methods\DC_ORO-AK-L%.MET
 Calibration File: G:\Org\HP5\Cals\SW8015C_ORO211017AK.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for for ~~Residual~~ TEH (Oil Range) Organics Calculations: 28542.41
 Rt range for ~~Residual~~ TEH (Oil Range) Organics: 12.58 to 30.05

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*#Triacontane	16.296	500.	349.556	69.91	-


~~RRO~~ TEH (Oil Range)Area:1.37078E+08 ~~RRO~~ TEH (Oil Range)AMOUNT: 4802.609

CONTINUING CALIBRATION REPORT: G:\org\HP5\DAT\HP5122121_b\1221HP5.0020.RAW

COMPOUND	ACTUAL (NG)	MEASURED (NG)	%RECOVERY	LIMITS
*30-40 Motor Oil	5000.	.043	.	75-125

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	LIMITS
*#Triacontane	16.296	200.	349.556	174.78	75-125

AMN 01/11/2022



RESIDUAL RANGE ORGANICS CHROMATOGRAM

Sample Name: CCV_1221HP520r, RRO ;1221HP5 , DRO211220C
 Raw File: G:\org\HP5\DAT\HP5122121_b\1221HP5.0020.RAW
 Date & Time Acquired: 12/22/2021 3:56:16 AM
 Method File: G:\Org\HP5\Methods\DS_ORO-AK-L%.MET
 Calibration File: G:\Org\HP5\Cals\SW8015C_ORO211017AK.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for for Residual Range Organics Calculations: 28542.41
 Rt range for Residual Range Organics: 12.58 to 30.05


SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*#Triacontane	16.296	500.	211.851	42.37	-

RRO Area:6182796 RRO AMOUNT: 216.6179

CONTINUING CALIBRATION REPORT: G:\org\HP5\DAT\HP5122121_b\1221HP5.0020.RAW

COMPOUND	ACTUAL (NG)	MEASURED (NG)	%RECOVERY	LIMITS
*30-40 Motor Oil	5000.	.043	.	75-125

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	LIMITS
*#Triacontane	16.296	200.	211.851	105.93	75-125



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: CCV_1221HP521r, DRO ;1221HP5 , DRO211220A
 Raw File: G:\org\HP5\DAT\HP5122121_b\1221HP5.0021.RAW
 Date & Time Acquired: 12/22/2021 4:39:23 AM
 Method File: G:\Org\HP5\Methods\DC_8015-24-IK-L%.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO211102Ik-24.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for TEH: 31353.19
 Rt range for Diesel Range Organics: 6.51 to 14.38

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*o-Terphenyl	12.221	200.	327.331	163.67
*1-Chlorooctadecane	13.063	200.	158.207	79.1

DRO Area: 4.661037E+08 DRO Amount: 14866.23
 TEH Area: 4.828347E+08 TEH Amount: 15399.86


CONTINUING CALIBRATION REPORT: G:\org\HP5\DAT\HP5122121_b\1221HP5.0021.RAW

COMPOUND	ACTUAL (NG)	MEASURED (NG)	%RECOVERY	LIMITS
TOTAL DRO	15000.	15399.86	102.67	85-115

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	LIMITS
*o-Terphenyl	12.221	200.	327.331	163.67	85-115
*1-Chlorooctadecane	13.063	200.	158.207	79.1	85-115

G:\org\HP5\DAT\HP5122121_b\1221HP5.0021.RAW

CCV_1221HP521r, DRO ;1221HP5 , DRO211220A



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: CCV_1221HP521r, DRO ;1221HP5 , DRO211220A
 Raw File: G:\org\HP5\DAT\HP5122121_b\1221HP5.0021.RAW
 Date & Time Acquired: 12/22/2021 4:39:23 AM
 Method File: G:\Org\HP5\Methods\DS_8015-24-IK-L#.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO211102Ik-24.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for TEH: 31353.19

Rt range for Diesel Range Organics: 6.51 to 14.38


SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*o-Terphenyl	12.221	200.	198.526	99.26
*1-Chlorooctadecane	13.063	200.	37.526	18.76

DRO Area: 2.583275E+08 DRO Amount: 8239.272
 TEH Area: 2.689115E+08 TEH Amount: 8576.846

CONTINUING CALIBRATION REPORT: G:\org\HP5\DAT\HP5122121_b\1221HP5.0021.RAW

COMPOUND	ACTUAL (NG)	MEASURED (NG)	%RECOVERY	LIMITS
TOTAL DRO	15000.	8576.85	57.18	85-115

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	LIMITS
*o-Terphenyl	12.221	200.	198.526	99.26	85-115
*1-Chlorooctadecane	13.063	200.	37.526	18.76	85-115



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: DCM-Baseline Check-V22
 Raw File: G:\org\HP5\DAT\HP5122121_b\1221HP5.0022.RAW
 Date & Time Acquired: 12/22/2021 7:02:09 AM
 Method File: G:\Org\HP5\Methods\DR_8015-IBb-LEXP.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO211102IB.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for TEH: 31353.19
 Rt range for Diesel Range Organics: 6.54 to 15.86

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*o-Terphenyl	12.217	200.	.094	.05	-
*1-Chlorooctadecane	13.077	200.	.042	.02	-


DRO Area:425246.2 DRO Amount: 13.56309
 TEH Area:545316.9 TEH Amount: 17.39271

ERH2178 (RHMW01R)

G:\org\HP5\DAT\HP5122121_b\1221HP5.0023.RAW

Batch ID: 162352

B21121623-001C ;1221HP5 , \$HC-8015-DRO-W,



RESIDUAL RANGE ORGANICS CHROMATOGRAM

Sample Name: B21121623-001C ;1221HP5 , \$HC-8015-DRO-W,
Raw File: G:\org\HP5\DAT\HP5122121_b\1221HP5.0023.RAW
Date & Time Acquired: 12/22/2021 7:44:43 AM
Method File: G:\Org\HP5\Methods\D3_OROS-122123-AK-L%.MET
Calibration File: G:\Org\HP5\Cals\SW8015C_ORO211017AK-SAMP.CAL
Sample Weight: 1020 Dilution: 1 S.A.: 1

Mean RF for for Residual Range Organics Calculations: 28542.41
Rt range for Residual Range Organics: 14.29 to 21.62

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*#Triacontane	16.29	.49	.099	20.11	-


RRO Area:2271961 RRO AMOUNT: 7.803871E-02

ERH2178 (RHMW01R)

Batch ID: 162352

G:\org\HP5\DAT\HP5122121_b\1221HP5.0023.RAW

B21121623-001C ;1221HP5 , \$HC-8015-DRO-W,



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: B21121623-001C ;1221HP5 , \$HC-8015-DRO-W,
Raw File: G:\org\HP5\DAT\HP5122121_b\1221HP5.0023.RAW
Date & Time Acquired: 12/22/2021 7:44:43 AM
Method File: G:\Org\HP5\Methods\DR_8015-122123-IK-L%.met
Calibration File: G:\Org\HP5\Cals\SW8015C_DRO211102IK-24-Tri.CAL
Sample Weight: 1020 Dilution: 1 S.A.: 1

Mean RF for TEH: 31353.19

Rt range for Diesel Range Organics: 6.51 to 14.38

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*o-Terphenyl	12.214	.196	.172	87.6	-
*1-Chlorooctadecane	13.052	.196	.001	.33	-
*#Triacontane	16.29	.196	.099	50.26	-


DRO Area:6223481 DRO Amount: 0.1946039
TEH Area:1.165042E+07 TEH Amount: 0.3643004

ERH2178 (RHMW01R)

Batch ID: 162352

G:\org\HP5\DAT\HP5122121_b\1221HP5.0023.RAW

B21121623-001C ;1221HP5 , \$HC-8015-DRO-W,



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT


Sample Name: B21121623-001C ;1221HP5 , \$HC-8015-DRO-W,
Raw File: G:\org\HP5\DAT\HP5122121_b\1221HP5.0023.RAW
Date & Time Acquired: 12/22/2021 7:44:43 AM
Method File: G:\Org\HP5\Methods\DS_8015-C24T-IK-L#.met
Calibration File: G:\Org\HP5\Cals\SW8015C_DRO211102IK-24-Tri.CAL
Sample Weight: 1020 Dilution: 1 S.A.: 1

Mean RF for TEH: 31353.19

Rt range for Diesel Range Organics: 6.51 to 14.38

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*o-Terphenyl	12.214	.196	.168	85.78	-
*1-Chlorooctadecane	13.052	.196	.	.01	-
*#Triacontane	16.29	.196	.096	49.16	-

DRO Area:3023054 DRO Amount: 9.452877E-02
TEH Area:6626082 TEH Amount: 0.2071929



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: DCM-Baseline Check-V24
 Raw File: G:\org\HP5\DAT\HP5122121_b\1221HP5.0024.RAW
 Date & Time Acquired: 12/22/2021 8:27:37 AM
 Method File: G:\Org\HP5\Methods\DR_8015-IBb-LEXP.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO211102IB.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for TEH: 31353.19
 Rt range for Diesel Range Organics: 6.54 to 15.86

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*o-Terphenyl	12.217	200.	.076	.04	-
*1-Chlorooctadecane	29.958	200.	.	.	-


DRO Area: 355771 DRO Amount: 11.3472
 TEH Area: 479438.5 TEH Amount: 15.29154

ERH2175 (RHMW2254-001)

Batch ID: 162352

G:\org\HP5\DAT\HP5122121_b\1221HP5.0025.RAW

B21121605-001A ;1221HP5 , \$HC-8015-DRO-W,



RESIDUAL RANGE ORGANICS CHROMATOGRAM

Sample Name: B21121605-001A ;1221HP5 , \$HC-8015-DRO-W,
Raw File: G:\org\HP5\DAT\HP5122121_b\1221HP5.0025.RAW
Date & Time Acquired: 12/22/2021 9:10:20 AM
Method File: G:\Org\HP5\Methods\D3_OROS-122123-AK-L%.MET
Calibration File: G:\Org\HP5\Cals\SW8015C_ORO211017AK-SAMP.CAL
Sample Weight: 1050 Dilution: 1 S.A.: 1

Mean RF for for Residual Range Organics Calculations: 28542.41
Rt range for Residual Range Organics: 14.29 to 21.62

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*#Triacontane	16.288	.476	.089	18.75


RRO Area:3106085 RRO AMOUNT: 0.1036414

ERH2175 (RHMW2254-001)

Batch ID: 162352

G:\org\HP5\DAT\HP5122121_b\1221HP5.0025.RAW

B21121605-001A ;1221HP5 , \$HC-8015-DRO-W,



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: B21121605-001A ;1221HP5 , \$HC-8015-DRO-W,
Raw File: G:\org\HP5\DAT\HP5122121_b\1221HP5.0025.RAW
Date & Time Acquired: 12/22/2021 9:10:20 AM
Method File: G:\Org\HP5\Methods\DR_8015-122123-IK-L%.met
Calibration File: G:\Org\HP5\Cals\SW8015C_DRO211102IK-24-Tri.CAL
Sample Weight: 1050 Dilution: 1 S.A.: 1

Mean RF for TEH: 31353.19

Rt range for Diesel Range Organics: 6.51 to 14.38

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*o-Terphenyl	12.212	.19	.167	87.79	-
*1-Chlorooctadecane	13.06	.19	.	.13	-
*#Triacontane	16.288	.19	.089	46.88	-


DRO Area:4817939 DRO Amount: 0.1463492
TEH Area:9000337 TEH Amount: 0.2733932

ERH2175 (RHMW2254-001)

Batch ID: 162352

G:\org\HP5\DAT\HP5122121_b\1221HP5.0025.RAW

B21121605-001A ;1221HP5 , \$HC-8015-DRO-W,



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT


Sample Name: B21121605-001A ;1221HP5 , \$HC-8015-DRO-W,
 Raw File: G:\org\HP5\DAT\HP5122121_b\1221HP5.0025.RAW
 Date & Time Acquired: 12/22/2021 9:10:20 AM
 Method File: G:\Org\HP5\Methods\DS_8015-C24T-IK-L#.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO211102IK-24-Tri.CAL
 Sample Weight: 1050 Dilution: 1 S.A.: 1

Mean RF for TEH: 31353.19

Rt range for Diesel Range Organics: 6.51 to 14.38

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*o-Terphenyl	12.212	.19	.165	86.55	-
*1-Chlorooctadecane	13.06	.19	.	.03	-
*#Triacontane	16.288	.19	.086	44.92	-

DRO Area:3396285 DRO Amount: 0.1031651
 TEH Area:6202406 TEH Amount: 0.1884036



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: DCM-Baseline Check-V26
 Raw File: G:\org\HP5\DAT\HP5122121_b\1221HP5.0026.RAW
 Date & Time Acquired: 12/22/2021 9:53:10 AM
 Method File: G:\Org\HP5\Methods\DR_8015-IBb-LEXP.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO211102IB.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for TEH: 31353.19
 Rt range for Diesel Range Organics: 6.54 to 15.86

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*o-Terphenyl	12.215	200.	.096	.05	-
*1-Chlorooctadecane	29.94	200.	.	.	-


DRO Area:377188.1 DRO Amount: 12.03029
 TEH Area:465741 TEH Amount: 14.85466

ERH2206 (RHMW2254-01)

Batch ID: 162352

G:\org\HP5\DAT\HP5122121_b\1221HP5.0027.RAW

B21121605-002A ;1221HP5 , \$HC-8015-DRO-W,



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: B21121605-002A ;1221HP5 , \$HC-8015-DRO-W,
 Raw File: G:\org\HP5\DAT\HP5122121_b\1221HP5.0027.RAW
 Date & Time Acquired: 12/22/2021 10:36:02 AM
 Method File: G:\Org\HP5\Methods\D3_8015-C24T-IK-L%.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO211102IK-24-Tri.CAL
 Sample Weight: 1050 Dilution: 1 S.A.: 1

Mean RF for TEH: 31353.19

Rt range for Diesel Range Organics: 6.51 to 14.38

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*o-Terphenyl	12.214	.19	.181	95.06	-
*1-Chlorooctadecane	13.06	.19	.001	.56	-
*#Triacontane	16.288	.19	.104	54.41	-


DRO Area: 5.062064E+07 DRO Amount: 1.537647
 TEH Area: 5.954589E+07 TEH Amount: 1.808759

ERH2206 (RHMW2254-01)

Batch ID: 162352

G:\org\HP5\DAT\HP5122121_b\1221HP5.0027.RAW

B21121605-002A ;1221HP5 , \$HC-8015-DRO-W,



RESIDUAL RANGE ORGANICS CHROMATOGRAM

Sample Name: B21121605-002A ;1221HP5 , \$HC-8015-DRO-W,
Raw File: G:\org\HP5\DAT\HP5122121_b\1221HP5.0027.RAW
Date & Time Acquired: 12/22/2021 10:36:02 AM
Method File: G:\Org\HP5\Methods\D3_OROS-AK-L%.MET
Calibration File: G:\Org\HP5\Cals\SW8015C_ORO211017AK-SAMP.CAL
Sample Weight: 1050 Dilution: 1 S.A.: 1

Mean RF for for Residual Range Organics Calculations: 28542.41
Rt range for Residual Range Organics: 14.29 to 21.62

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*#Triacontane	16.288	.476	.104	21.78	-


RRO Area:5258103 RRO AMOUNT: 0.1754483

ERH2206 (RHMW2254-01)

Batch ID: 162352

G:\org\HP5\DAT\HP5122121_b\1221HP5.0027.RAW

B21121605-002A ;1221HP5 , \$HC-8015-DRO-W,



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: B21121605-002A ;1221HP5 , \$HC-8015-DRO-W,
Raw File: G:\org\HP5\DAT\HP5122121_b\1221HP5.0027.RAW
Date & Time Acquired: 12/22/2021 10:36:02 AM
Method File: G:\Org\HP5\Methods\DS_8015-122127-IK-L#.met
Calibration File: G:\Org\HP5\Cals\SW8015C_DRO211102IK-24-Tri.CAL
Sample Weight: 1050 Dilution: 1 S.A.: 1

Mean RF for TEH: 31353.19

Rt range for Diesel Range Organics: 6.51 to 14.38

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*o-Terphenyl	12.214	.19	.177	92.78	-
*1-Chlorooctadecane	13.06	.19	.	.04	-
*#Triacontane	16.288	.19	.094	49.56	-


DRO Area: 4.603533E+07 DRO Amount: 1.398364
TEH Area: 4.740095E+07 TEH Amount: 1.439846

ERH2176 (RHMW2254-01)

Batch ID: 162352

G:\org\HP5\DAT\HP5122121_b\1221HP5.0028.RAW

B21121605-003A ;1221HP5 , \$HC-8015-DRO-W,



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: B21121605-003A ;1221HP5 , \$HC-8015-DRO-W,
Raw File: G:\org\HP5\DAT\HP5122121_b\1221HP5.0028.RAW
Date & Time Acquired: 12/22/2021 11:19:14 AM
Method File: G:\Org\HP5\Methods\D3_8015-C24T-IK-L%.met
Calibration File: G:\Org\HP5\Cals\SW8015C_DRO211102IK-24-Tri.CAL
Sample Weight: 1010 Dilution: 1 S.A.: 1

Mean RF for TEH: 31353.19

Rt range for Diesel Range Organics: 6.51 to 14.38

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*o-Terphenyl	12.213	.198	.188	94.92	-
*1-Chlorooctadecane	13.062	.198	.001	.56	-
*#Triacontane	16.289	.198	.107	54.11	-


DRO Area:5.57105E+07 DRO Amount: 1.759276
TEH Area:6.364356E+07 TEH Amount: 2.009793

ERH2176 (RHMW2254-01)

Batch ID: 162352

G:\org\HP5\DAT\HP5122121_b\1221HP5.0028.RAW

B21121605-003A ;1221HP5 , \$HC-8015-DRO-W,



RESIDUAL RANGE ORGANICS CHROMATOGRAM

Sample Name: B21121605-003A ;1221HP5 , \$HC-8015-DRO-W,
Raw File: G:\org\HP5\DAT\HP5122121_b\1221HP5.0028.RAW
Date & Time Acquired: 12/22/2021 11:19:14 AM
Method File: G:\Org\HP5\Methods\D3_OROS-AK-L%.MET
Calibration File: G:\Org\HP5\Cals\SW8015C_ORO211017AK-SAMP.CAL
Sample Weight: 1010 Dilution: 1 S.A.: 1

Mean RF for for Residual Range Organics Calculations: 28542.41
Rt range for Residual Range Organics: 14.29 to 21.62

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*#Triacontane	16.289	.495	.107	21.65


RRO Area:5145561 RRO AMOUNT: 0.1784928

ERH2176 (RHMW2254-01)

Batch ID: 162352

G:\org\HP5\DAT\HP5122121_b\1221HP5.0028.RAW

B21121605-003A ;1221HP5 , \$HC-8015-DRO-W,



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT


Sample Name: B21121605-003A ;1221HP5 , \$HC-8015-DRO-W,
Raw File: G:\org\HP5\DAT\HP5122121_b\1221HP5.0028.RAW
Date & Time Acquired: 12/22/2021 11:19:14 AM
Method File: G:\Org\HP5\Methods\DS_8015-122127-IK-L#.met
Calibration File: G:\Org\HP5\Cals\SW8015C_DRO211102IK-24-Tri.CAL
Sample Weight: 1010 Dilution: 1 S.A.: 1

Mean RF for TEH: 31353.19

Rt range for Diesel Range Organics: 6.51 to 14.38

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*o-Terphenyl	12.213	.198	.184	93.11	-
*1-Chlorooctadecane	13.062	.198	.04		-
*#Triacontane	16.289	.198	.096	48.71	-

DRO Area: 5.194986E+07 DRO Amount: 1.640519
TEH Area: 5.338243E+07 TEH Amount: 1.685758



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: DCM-Baseline Check-V29
 Raw File: G:\Org\HP5\DAT\HP5122121_b\1221HP5.0029.RAW
 Date & Time Acquired: 12/22/2021 12:01:53 PM
 Method File: G:\Org\HP5\Methods\DR_8015-IBb-Lexp.MET
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO211102IB.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for TEH: 31353.19
 Rt range for Diesel Range Organics: 6.54 to 15.86

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*o-Terphenyl	12.217	200.	.086	.04
*1-Chlorooctadecane	29.97	200.	.	.


DRO Area:334877.2 DRO Amount: 10.6808
 TEH Area:493060.2 TEH Amount: 15.726

ERH2193 (RHMW12A)

Batch ID: 162352

G:\Org\HP5\DAT\HP5122121_b\1221HP5.0030.RAW

B21121622-001B ;1221HP5 , \$HC-8015-DRO-W,



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: B21121622-001B ;1221HP5 , \$HC-8015-DRO-W,
Raw File: G:\Org\HP5\DAT\HP5122121_b\1221HP5.0030.RAW
Date & Time Acquired: 12/22/2021 12:44:34 PM
Method File: G:\Org\HP5\Methods\DR_8015-IBb-Lexp.MET
Calibration File: G:\Org\HP5\Cals\SW8015C_DRO211102IB.CAL
Sample Weight: 960 Dilution: 1 S.A.: 1

Mean RF for TEH: 31353.19

Rt range for Diesel Range Organics: 6.54 to 15.86

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*o-Terphenyl	12.212	.208	.187	89.87	-
*1-Chlorooctadecane	13.019	.208	.	.1	-


DRO Area:1038049 DRO Amount: 3.448777E-02
TEH Area:5803344 TEH Amount: 0.1928081

ERH2195 (RHMW16)

Batch ID: 162352

G:\org\HP5\DAT\HP5122121_b\1221HP5.0031.RAW

B21121622-002B ;1221HP5 , \$HC-8015-DRO-W,



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: B21121622-002B ;1221HP5 , \$HC-8015-DRO-W,
Raw File: G:\org\HP5\DAT\HP5122121_b\1221HP5.0031.RAW
Date & Time Acquired: 12/22/2021 1:27:19 PM
Method File: G:\Org\HP5\Methods\D3_8015-122131-IK-L%.met
Calibration File: G:\Org\HP5\Cals\SW8015C_DRO211102IK-24-Tri.CAL
Sample Weight: 1050 Dilution: 1 S.A.: 1

Mean RF for TEH: 31353.19

Rt range for Diesel Range Organics: 6.51 to 14.38

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*o-Terphenyl	12.212	.19	.184	96.4	-
*1-Chlorooctadecane	13.047	.19	.008	4.38	-
*#Triacontane	16.287	.19	.107	56.15	-


DRO Area:1.007537E+07 DRO Amount: 0.3060482
TEH Area:2.61747E+07 TEH Amount: 0.7950797

ERH2195 (RHMW16)

Batch ID: 162352

G:\org\HP5\DAT\HP5122121_b\1221HP5.0031.RAW

B21121622-002B ;1221HP5 , \$HC-8015-DRO-W,



RESIDUAL RANGE ORGANICS CHROMATOGRAM

Sample Name: B21121622-002B ;1221HP5 , \$HC-8015-DRO-W,
Raw File: G:\org\HP5\DAT\HP5122121_b\1221HP5.0031.RAW
Date & Time Acquired: 12/22/2021 1:27:19 PM
Method File: G:\Org\HP5\Methods\D3_OROS-122131-AK-L%.MET
Calibration File: G:\Org\HP5\Cals\SW8015C_ORO211017AK-SAMP.CAL
Sample Weight: 1050 Dilution: 1 S.A.: 1

Mean RF for for Residual Range Organics Calculations: 28542.41
Rt range for Residual Range Organics: 14.29 to 21.62

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*#Triacontane	16.287	.476	.107	22.47


RRO Area:1.16533E+07 RRO AMOUNT: 0.3888381

ERH2195 (RHMW16)

Batch ID: 162352

G:\org\HP5\DAT\HP5122121_b\1221HP5.0031.RAW

B21121622-002B ;1221HP5 , \$HC-8015-DRO-W,



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: B21121622-002B ;1221HP5 , \$HC-8015-DRO-W,
Raw File: G:\org\HP5\DAT\HP5122121_b\1221HP5.0031.RAW
Date & Time Acquired: 12/22/2021 1:27:19 PM
Method File: G:\Org\HP5\Methods\DS_8015-C24T-IK-L#.met
Calibration File: G:\Org\HP5\Cals\SW8015C_DRO211102IK-24-Tri.CAL
Sample Weight: 1050 Dilution: 1 S.A.: 1

Mean RF for TEH: 31353.19

Rt range for Diesel Range Organics: 6.51 to 14.38

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*o-Terphenyl	12.212	.19	.179	93.93	-
*1-Chlorooctadecane	13.047	.19	.002	1.21	-
*#Triacontane	16.287	.19	.095	50.12	-


DRO Area:5942135 DRO Amount: 0.1804976
TEH Area:2.337826E+07 TEH Amount: 0.7101353

ERH2205 (RHMW15-05)

Batch ID: 162352

G:\Org\HP5\DAT\HP5122121_b\1221HP5.0032.RAW

B21121622-003B ;1221HP5 , \$HC-8015-DRO-W,



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT


Sample Name: B21121622-003B ;1221HP5 , \$HC-8015-DRO-W,
Raw File: G:\Org\HP5\DAT\HP5122121_b\1221HP5.0032.RAW
Date & Time Acquired: 12/22/2021 2:10:15 PM
Method File: G:\Org\HP5\Methods\DR_8015-IBb-Lexp.MET
Calibration File: G:\Org\HP5\Cals\SW8015C_DRO211102IB.CAL
Sample Weight: 900 Dilution: 1 S.A.: 1


Mean RF for TEH: 31353.19

Rt range for Diesel Range Organics: 6.54 to 15.86

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*o-Terphenyl	12.212	.222	.191	85.82	-
*1-Chlorooctadecane	13.023	.222	.	.04	-

DRO Area:734614.3 DRO Amount: 2.603365E-02
TEH Area:3911347 TEH Amount: 0.1386124





RESIDUAL RANGE ORGANICS CHROMATOGRAM

Sample Name: CCV_1221HP536r, RRO ;1221HP5 , DRO211220C
 Raw File: G:\Org\HP5\DAT\HP5122121_b\1221HP5.0034.RAW
 Date & Time Acquired: 12/22/2021 3:35:47 PM
 Method File: G:\Org\HP5\Methods\DC_ORO-AK-L%.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_ORO211017AK.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for for ~~Residual~~ TEH (Oil Range) Organics Calculations: 28542.41
 Rt range for ~~Residual~~ TEH (Oil Range) Organics: 12.58 to 30.05

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*#Triacontane	16.291	500.	348.211	69.64	-


RRO TEH (Oil Range) Area:1.336989E+08 RRO TEH (Oil Range) AMOUNT: 4684.221

CONTINUING CALIBRATION REPORT: G:\Org\HP5\DAT\HP5122121_b\1221HP5.0034.RAW

COMPOUND	ACTUAL (NG)	MEASURED (NG)	%RECOVERY	LIMITS
*30-40 Motor Oil	5000.	.056	.	75-125

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	LIMITS
*#Triacontane	16.291	200.	348.211	174.11	75-125

AMN 01/11/2022



RESIDUAL RANGE ORGANICS CHROMATOGRAM

Sample Name: CCV_1221HP536r, RRO ;1221HP5 , DRO211220C
 Raw File: G:\Org\HP5\DAT\HP5122121_b\1221HP5.0034.RAW
 Date & Time Acquired: 12/22/2021 3:35:47 PM
 Method File: G:\Org\HP5\Methods\DS_ORO-AK-L%.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_ORO211017AK.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for for Residual Range Organics Calculations: 28542.41
 Rt range for Residual Range Organics: 12.58 to 30.05

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*#Triacontane	16.291	500.	209.356	41.87

RRO Area:6197081 RRO AMOUNT: 217.1183


CONTINUING CALIBRATION REPORT: G:\Org\HP5\DAT\HP5122121_b\1221HP5.0034.RAW

COMPOUND	ACTUAL (NG)	MEASURED (NG)	%RECOVERY	LIMITS
*30-40 Motor Oil	5000.	.056	.	75-125

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	LIMITS
*#Triacontane	16.291	200.	209.356	104.68	75-125

G:\Org\HP5\DAT\HP5122121_b\1221HP5.0035.RAW

CCV_1221HP537r, DRO ;1221HP5 , DRO211220A



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: CCV_1221HP537r, DRO ;1221HP5 , DRO211220A
 Raw File: G:\Org\HP5\DAT\HP5122121_b\1221HP5.0035.RAW
 Date & Time Acquired: 12/22/2021 4:18:49 PM
 Method File: G:\Org\HP5\Methods\DC_8015-24-IK-L%.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO211102Ik-24.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for TEH: 31353.19

Rt range for Diesel Range Organics: 6.51 to 14.38

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*o-Terphenyl	12.219	200.	307.877	153.94
*1-Chlorooctadecane	13.061	200.	150.6	75.3

DRO Area: 4.394882E+08 DRO Amount: 14017.34
 TEH Area: 4.55703E+08 TEH Amount: 14534.5


CONTINUING CALIBRATION REPORT: G:\Org\HP5\DAT\HP5122121_b\1221HP5.0035.RAW

COMPOUND	ACTUAL (NG)	MEASURED (NG)	%RECOVERY	LIMITS
TOTAL DRO	15000.	14534.5	96.9	85-115

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	LIMITS
*o-Terphenyl	12.219	200.	307.877	153.94	85-115
*1-Chlorooctadecane	13.061	200.	150.6	75.3	85-115

G:\Org\HP5\DAT\HP5122121_b\1221HP5.0035.RAW

CCV_1221HP537r, DRO ;1221HP5 , DRO211220A



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: CCV_1221HP537r, DRO ;1221HP5 , DRO211220A
 Raw File: G:\Org\HP5\DAT\HP5122121_b\1221HP5.0035.RAW
 Date & Time Acquired: 12/22/2021 4:18:49 PM
 Method File: G:\Org\HP5\Methods\DS_8015-24-IK-L#.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO211102Ik-24.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for TEH: 31353.19

Rt range for Diesel Range Organics: 6.51 to 14.38


SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*o-Terphenyl	12.219	200.	187.134	93.57
*1-Chlorooctadecane	13.061	200.	35.511	17.76

DRO Area: 2.42773E+08 DRO Amount: 7743.167
 TEH Area: 2.529474E+08 TEH Amount: 8067.676

CONTINUING CALIBRATION REPORT: G:\Org\HP5\DAT\HP5122121_b\1221HP5.0035.RAW

COMPOUND	ACTUAL (NG)	MEASURED (NG)	%RECOVERY	LIMITS
TOTAL DRO	15000.	8067.68	53.78	85-115

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	LIMITS
*o-Terphenyl	12.219	200.	187.134	93.57	85-115
*1-Chlorooctadecane	13.061	200.	35.511	17.76	85-115




DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: DCM-Baseline Check-V36
 Raw File: G:\org\HP5\DAT\HP5122121_b\1221HP5.0036.RAW
 Date & Time Acquired: 12/22/2021 5:02:18 PM
 Method File: G:\Org\HP5\Methods\DR_8015-IBb-LEXP.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO211102IB.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for TEH: 31353.19
 Rt range for Diesel Range Organics: 6.54 to 15.86

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*o-Terphenyl	12.22	200.	.182	.09
*1-Chlorooctadecane	29.978	200.	.	.

DRO Area:266088.9 DRO Amount: 8.48682
 TEH Area:358649.6 TEH Amount: 11.43902



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: DCM-Baseline Check-V37
 Raw File: G:\org\HP5\DAT\HP5122121_b\1221HP5.0037.RAW
 Date & Time Acquired: 12/22/2021 5:46:06 PM
 Method File: G:\Org\HP5\Methods\DR_8015-IBb-LEXP.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO211102IB.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for TEH: 31353.19
 Rt range for Diesel Range Organics: 6.54 to 15.86

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*o-Terphenyl	12.219	200.	.312	.16	-
*1-Chlorooctadecane	13.008	200.	.019	.01	-


DRO Area: 582158 DRO Amount: 18.56774
 TEH Area: 803733.5 TEH Amount: 25.63482

ERH2193 (RHMW12A)

Batch ID: 162352

G:\org\HP5\DAT\HP5122121_b\1221HP5.0038.RAW

B21121622-001B ;1221HP5 , \$HC-8015-DRO-W, rr



RESIDUAL RANGE ORGANICS CHROMATOGRAM

Sample Name: B21121622-001B ;1221HP5 , \$HC-8015-DRO-W, rr
 Raw File: G:\org\HP5\DAT\HP5122121_b\1221HP5.0038.RAW
 Date & Time Acquired: 12/22/2021 6:29:38 PM
 Method File: G:\Org\HP5\Methods\D3_OROS-122138-AK-L%.MET
 Calibration File: G:\Org\HP5\Cals\SW8015C_ORO211017AK-SAMP.CAL
 Sample Weight: 960 Dilution: 1 S.A.: 1

Mean RF for for Residual Range Organics Calculations: 28542.41
 Rt range for Residual Range Organics: 14.29 to 21.62

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*#Triacontane	16.286	.521	.109	21.01


RRO Area:2858089 RRO AMOUNT: 0.1043071

ERH2193 (RHMW12A)

Batch ID: 162352

G:\org\HP5\DAT\HP5122121_b\1221HP5.0038.RAW

B21121622-001B ;1221HP5 , \$HC-8015-DRO-W, rr



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: B21121622-001B ;1221HP5 , \$HC-8015-DRO-W, rr
 Raw File: G:\org\HP5\DAT\HP5122121_b\1221HP5.0038.RAW
 Date & Time Acquired: 12/22/2021 6:29:38 PM
 Method File: G:\Org\HP5\Methods\DR_8015-122138-IK-L%.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO211102IK-24-Tri.CAL
 Sample Weight: 960 Dilution: 1 S.A.: 1

Mean RF for TEH: 31353.19

Rt range for Diesel Range Organics: 6.51 to 14.38

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*o-Terphenyl	12.212	.208	.195	93.41	-
*1-Chlorooctadecane	13.06	.208	.	.04	-
*#Triacontane	16.286	.208	.109	52.51	-


DRO Area:861031.2 DRO Amount: 2.860658E-02
 TEH Area:4050678 TEH Amount: 0.1345782

ERH2193 (RHMW12A)

Batch ID: 162352

G:\org\HP5\DAT\HP5122121_b\1221HP5.0038.RAW

B21121622-001B ;1221HP5 , \$HC-8015-DRO-W, rr



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: B21121622-001B ;1221HP5 , \$HC-8015-DRO-W, rr
 Raw File: G:\org\HP5\DAT\HP5122121_b\1221HP5.0038.RAW
 Date & Time Acquired: 12/22/2021 6:29:38 PM
 Method File: G:\Org\HP5\Methods\DS_8015-C24T-IK-L#.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO211102IK-24-Tri.CAL
 Sample Weight: 960 Dilution: 1 S.A.: 1

Mean RF for TEH: 31353.19
 Rt range for Diesel Range Organics: 6.51 to 14.38

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*o-Terphenyl	12.212	.208	.194	93.29	-
*1-Chlorooctadecane	13.06	.208	.	.03	-
*#Triacontane	16.286	.208	.102	48.75	-


DRO Area:734578.2 DRO Amount: 2.440535E-02
 TEH Area:1963439 TEH Amount: 6.523256E-02

ERH2205 (RHMW15-05)

Batch ID: 162352

G:\org\HP5\DAT\HP5122121_b\1221HP5.0039.RAW

B21121622-003B ;1221HP5 , \$HC-8015-DRO-W, rr



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: B21121622-003B ;1221HP5 , \$HC-8015-DRO-W, rr
 Raw File: G:\org\HP5\DAT\HP5122121_b\1221HP5.0039.RAW
 Date & Time Acquired: 12/22/2021 7:13:02 PM
 Method File: G:\Org\HP5\Methods\DR_8015-C24T-IK-L%.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO211102IK-24-Tri.CAL
 Sample Weight: 900 Dilution: 1 S.A.: 1

Mean RF for TEH: 31353.19

Rt range for Diesel Range Organics: 6.51 to 14.38

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*o-Terphenyl	12.21	.222	.201	90.25	-
*1-Chlorooctadecane	13.057	.222	.	.07	-
*#Triacontane	16.287	.222	.108	48.53	-


DRO Area:660509.8 DRO Amount: 0.0234075
 TEH Area:1004612 TEH Amount: 3.560196E-02

ERH2205 (RHMW15-05)

Batch ID: 162352

G:\org\HP5\DAT\HP5122121_b\1221HP5.0039.RAW

B21121622-003B ;1221HP5 , \$HC-8015-DRO-W, rr



RESIDUAL RANGE ORGANICS CHROMATOGRAM

Sample Name: B21121622-003B ;1221HP5 , \$HC-8015-DRO-W, rr
 Raw File: G:\org\HP5\DAT\HP5122121_b\1221HP5.0039.RAW
 Date & Time Acquired: 12/22/2021 7:13:02 PM
 Method File: G:\Org\HP5\Methods\DR_OROS-AK-L%.MET
 Calibration File: G:\Org\HP5\Cals\SW8015C_ORO211017AK-SAMP.CAL
 Sample Weight: 900 Dilution: 1 S.A.: 1

Mean RF for for Residual Range Organics Calculations: 28542.41
 Rt range for Residual Range Organics: 14.29 to 21.62

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*#Triacontane	16.287	.556	.108	19.41


RRO Area:193839.2 RRO AMOUNT: 7.545855E-03

ERH2205 (RHMW15-05)

Batch ID: 162352

G:\org\HP5\DAT\HP5122121_b\1221HP5.0039.RAW

B21121622-003B ;1221HP5 , \$HC-8015-DRO-W, rr



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT


Sample Name: B21121622-003B ;1221HP5 , \$HC-8015-DRO-W, rr
 Raw File: G:\org\HP5\DAT\HP5122121_b\1221HP5.0039.RAW
 Date & Time Acquired: 12/22/2021 7:13:02 PM
 Method File: G:\Org\HP5\Methods\DS_8015-C24T-IK-L#.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO211102IK-24-Tri.CAL
 Sample Weight: 900 Dilution: 1 S.A.: 1

Mean RF for TEH: 31353.19

Rt range for Diesel Range Organics: 6.51 to 14.38

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*o-Terphenyl	12.21	.222	.2	89.92	-
*1-Chlorooctadecane	13.057	.222	.	.04	-
*#Triacontane	16.287	.222	.106	47.49	-

DRO Area: 610374.4 DRO Amount: 2.163078E-02
 TEH Area: 1074221 TEH Amount: 0.0380688




DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: DCM-Baseline Check-V40
 Raw File: G:\org\HP5\DAT\HP5122121_b\1221HP5.0040.RAW
 Date & Time Acquired: 12/22/2021 7:56:20 PM
 Method File: G:\Org\HP5\Methods\DR_8015-IBb-LEXP.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO211102IB.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for TEH: 31353.19
 Rt range for Diesel Range Organics: 6.54 to 15.86

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*o-Terphenyl	12.221	200.	.223	.11
*1-Chlorooctadecane	29.928	200.	.	.

DRO Area:323035.9 DRO Amount: 10.30313
 TEH Area:410978.2 TEH Amount: 13.10802



RESIDUAL RANGE ORGANICS CHROMATOGRAM

Sample Name: B21121613-001BMS-RRO ;1221HP5 ,
 Raw File: G:\org\HP5\DAT\HP5122121_b\1221HP5.0041.RAW
 Date & Time Acquired: 12/22/2021 8:39:40 PM
 Method File: G:\Org\HP5\Methods\D3_ORO-AK-L%.MET
 Calibration File: G:\Org\HP5\Cals\SW8015C_ORO211017AK.CAL
 Sample Weight: 1045 Dilution: 1 S.A.: 1

Mean RF for for ~~Residual~~ TEH (Oil Range) Organics Calculations: 28542.41
 Rt range for ~~Residual~~ TEH (Oil Range) Organics: 12.58 to 30.05


SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*#Triacontane	16.288	.478	.18	37.6	-

RRO TEH (Oil Range) Area:1.338794E+08 RRO TEH (Oil Range) AMOUNT: 4.488557

AMN 01/11/2022

G:\org\HP5\DAT\HP5122121_b\1221HP5.0041.RAW

B21121613-001BMS-RRO ;1221HP5 ,




RESIDUAL RANGE ORGANICS CHROMATOGRAM

Sample Name: B21121613-001BMS-RRO ;1221HP5 ,
 Raw File: G:\org\HP5\DAT\HP5122121_b\1221HP5.0041.RAW
 Date & Time Acquired: 12/22/2021 8:39:40 PM
 Method File: G:\Org\HP5\Methods\DS_ORO-AK-L%.MET
 Calibration File: G:\Org\HP5\Cals\SW8015C_ORO211017AK.CAL
 Sample Weight: 1045 Dilution: 1 S.A.: 1

Mean RF for for Residual Range Organics Calculations: 28542.41
 Rt range for Residual Range Organics: 12.58 to 30.05

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*#Triacontane	16.288	.478	.091	19.09

RRO Area:4832993 RRO AMOUNT: 0.1620351




DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: DCM-Baseline Check-V42
 Raw File: G:\org\HP5\DAT\HP5122121_b\1221HP5.0042.RAW
 Date & Time Acquired: 12/22/2021 9:22:56 PM
 Method File: G:\Org\HP5\Methods\DR_8015-IBb-LEXP.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO211102IB.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for TEH: 31353.19
 Rt range for Diesel Range Organics: 6.54 to 15.86

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*o-Terphenyl	12.22	200.	.228	.11
*1-Chlorooctadecane	29.943	200.	.	.

DRO Area:634204.5 DRO Amount: 20.22775
 TEH Area:1122381 TEH Amount: 35.79798



RESIDUAL RANGE ORGANICS CHROMATOGRAM


Sample Name: B21121613-001BMSD-RRO ;1221HP5 ,
 Raw File: G:\org\HP5\DAT\HP5122121_b\1221HP5.0043.RAW
 Date & Time Acquired: 12/22/2021 10:06:14 PM
 Method File: G:\Org\HP5\Methods\D3_ORO-AK-L%.MET
 Calibration File: G:\Org\HP5\Cals\SW8015C_ORO211017AK.CAL
 Sample Weight: 1055 Dilution: 1 S.A.: 1

Mean RF for for ~~Residual~~ TEH (Oil Range) Organics Calculations: 28542.41
 Rt range for ~~Residual~~ TEH (Oil Range) Organics: 12.58 to 30.05

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*#Triacontane	16.289	.474	.174	36.77	-

~~RRO~~ TEH (Oil Range) Area:1.35403E+08 ~~RRO~~ TEH (Oil Range) AMOUNT: 4.496609

AMN 01/11/2022




RESIDUAL RANGE ORGANICS CHROMATOGRAM

Sample Name: B21121613-001BMSD-RRO ;1221HP5 ,
 Raw File: G:\org\HP5\DAT\HP5122121_b\1221HP5.0043.RAW
 Date & Time Acquired: 12/22/2021 10:06:14 PM
 Method File: G:\Org\HP5\Methods\DS_ORO-AK-L%.MET
 Calibration File: G:\Org\HP5\Cals\SW8015C_ORO211017AK.CAL
 Sample Weight: 1055 Dilution: 1 S.A.: 1

Mean RF for for Residual Range Organics Calculations: 28542.41
 Rt range for Residual Range Organics: 12.58 to 30.05

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*#Triacontane	16.289	.474	.09	18.89	-

RRO Area:5022230 RRO AMOUNT: 0.1667837



RESIDUAL RANGE ORGANICS CHROMATOGRAM


Sample Name: LCS-162352-RRO ;1221HP5 ,
 Raw File: G:\org\HP5\DAT\HP5122121_b\1221HP5.0044.RAW
 Date & Time Acquired: 12/22/2021 10:49:23 PM
 Method File: G:\Org\HP5\Methods\D3_ORO-AK-L%.MET
 Calibration File: G:\Org\HP5\Cals\SW8015C_ORO211017AK.CAL
 Sample Weight: 1000 Dilution: 1 S.A.: 1


Mean RF for for ~~Residual~~ TEH (Oil Range) Organics Calculations: 28542.41
 Rt range for ~~Residual~~ TEH (Oil Range) Organics: 12.58 to 30.05

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*#Triacontane	16.288	.5	.171	34.22	-

RRO TEH (Oil Range) Area:1.325306E+08 RRO TEH (Oil Range) AMOUNT: 4.643287

AMN 01/11/2022





RESIDUAL RANGE ORGANICS CHROMATOGRAM

Sample Name: CCV_1221HP547r, RRO ;1221HP5 , DRO211220C
 Raw File: G:\org\HP5\DAT\HP5122121_b\1221HP5.0046.RAW
 Date & Time Acquired: 12/23/2021 12:15:45 AM
 Method File: G:\Org\HP5\Methods\DC_ORO-AK-L%.MET
 Calibration File: G:\Org\HP5\Cals\SW8015C_ORO211017AK.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for for ~~Residual~~ TEH (Oil Range) Organics Calculations: 28542.41
 Rt range for ~~Residual~~ TEH (Oil Range) Organics: 12.58 to 30.05

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*#Triacontane	16.292	500.	337.08	67.42	-


~~RRO~~ TEH (Oil Range) Area:1.333428E+08 ~~RRO~~ TEH (Oil Range) AMOUNT: 4671.743

CONTINUING CALIBRATION REPORT: G:\org\HP5\DAT\HP5122121_b\1221HP5.0046.RAW

COMPOUND	ACTUAL (NG)	MEASURED (NG)	%RECOVERY	LIMITS
*30-40 Motor Oil	5000.	.03	.	75-125

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	LIMITS
*#Triacontane	16.292	200.	337.08	168.54	75-125

AMN 01/11/2022



RESIDUAL RANGE ORGANICS CHROMATOGRAM

Sample Name: CCV_1221HP547r, RRO ;1221HP5 , DRO211220C
 Raw File: G:\org\HP5\DAT\HP5122121_b\1221HP5.0046.RAW
 Date & Time Acquired: 12/23/2021 12:15:45 AM
 Method File: G:\Org\HP5\Methods\DS_ORO-AK-L%.MET
 Calibration File: G:\Org\HP5\Cals\SW8015C_ORO211017AK.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for for Residual Range Organics Calculations: 28542.41
 Rt range for Residual Range Organics: 12.58 to 30.05


SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*#Triacontane	16.292	500.	207.587	41.52	-

RRO Area:5995935 RRO AMOUNT: 210.0711

CONTINUING CALIBRATION REPORT: G:\org\HP5\DAT\HP5122121_b\1221HP5.0046.RAW

COMPOUND	ACTUAL (NG)	MEASURED (NG)	%RECOVERY	LIMITS
*30-40 Motor Oil	5000.	.03	.	75-125

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	LIMITS
*#Triacontane	16.292	200.	207.587	103.79	75-125



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: CCV_1221HP548r, DRO ;1221HP5 , DRO211220A
 Raw File: G:\org\HP5\DAT\HP5122121_b\1221HP5.0047.RAW
 Date & Time Acquired: 12/23/2021 12:59:02 AM
 Method File: G:\Org\HP5\Methods\DC_8015-24-IK-L%.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO211102Ik-24.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for TEH: 31353.19
 Rt range for Diesel Range Organics: 6.51 to 14.38

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*o-Terphenyl	12.216	200.	315.942	157.97
*1-Chlorooctadecane	13.059	200.	154.909	77.45

DRO Area: 4.495668E+08 DRO Amount: 14338.79
 TEH Area: 4.651225E+08 TEH Amount: 14834.93


CONTINUING CALIBRATION REPORT: G:\org\HP5\DAT\HP5122121_b\1221HP5.0047.RAW

COMPOUND	ACTUAL (NG)	MEASURED (NG)	%RECOVERY	LIMITS
TOTAL DRO	15000.	14834.93	98.9	85-115

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	LIMITS
*o-Terphenyl	12.216	200.	315.942	157.97	85-115
*1-Chlorooctadecane	13.059	200.	154.909	77.45	85-115

G:\org\HP5\DAT\HP5122121_b\1221HP5.0047.RAW

CCV_1221HP548r, DRO ;1221HP5 , DRO211220A



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: CCV_1221HP548r, DRO ;1221HP5 , DRO211220A
 Raw File: G:\org\HP5\DAT\HP5122121_b\1221HP5.0047.RAW
 Date & Time Acquired: 12/23/2021 12:59:02 AM
 Method File: G:\Org\HP5\Methods\DS_8015-24-IK-L#.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO211102Ik-24.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for TEH: 31353.19

Rt range for Diesel Range Organics: 6.51 to 14.38


SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*o-Terphenyl	12.216	200.	191.297	95.65	-
*1-Chlorooctadecane	13.059	200.	36.175	18.09	-

DRO Area: 2.494623E+08 DRO Amount: 7956.519
 TEH Area: 2.596854E+08 TEH Amount: 8282.583

CONTINUING CALIBRATION REPORT: G:\org\HP5\DAT\HP5122121_b\1221HP5.0047.RAW

COMPOUND	ACTUAL (NG)	MEASURED (NG)	%RECOVERY	LIMITS
TOTAL DRO	15000.	8282.58	55.22	85-115

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	LIMITS
*o-Terphenyl	12.216	200.	191.297	95.65	85-115
*1-Chlorooctadecane	13.059	200.	36.175	18.09	85-115




DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: DCM-Baseline Check-V01
 Raw File: G:\org\HP5\DAT\HP5122621_b\1226HP5.0001.RAW
 Date & Time Acquired: 12/26/2021 11:16:04 AM
 Method File: G:\Org\HP5\Methods\DR_8015-IBb-LEXP.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO211102IB.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for TEH: 31353.19
 Rt range for Diesel Range Organics: 6.54 to 15.86

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*o-Terphenyl	12.211	200.	2.352	1.18	-
*1-Chlorooctadecane	13.066	200.	3.042	1.52	-

DRO Area: 4547303 DRO Amount: 145.0348
 TEH Area: 4689133 TEH Amount: 149.5584




DIESEL RANGE ORGANICS CHROMATOGRAM REPORT


Sample Name: DCM-Baseline Check-V02
 Raw File: G:\org\HP5\DAT\HP5122621_b\1226HP5.0002.RAW
 Date & Time Acquired: 12/26/2021 11:58:52 AM
 Method File: G:\Org\HP5\Methods\DR_8015-IBb-LEXP.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO211102IB.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for TEH: 31353.19
 Rt range for Diesel Range Organics: 6.54 to 15.86

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*o-Terphenyl	12.219	200.	.233	.12	-
*1-Chlorooctadecane	13.066	200.	.063	.03	-

DRO Area:264991.9 DRO Amount: 8.451831
 TEH Area:313974.9 TEH Amount: 10.01413





RESIDUAL RANGE ORGANICS CHROMATOGRAM

Sample Name: CCV_1226HP504r, RRO ;1226HP5 , DRO211220C
 Raw File: G:\org\HP5\DAT\HP5122621_b\1226HP5.0004.RAW
 Date & Time Acquired: 12/26/2021 1:24:41 PM
 Method File: G:\Org\HP5\Methods\DC_ORO-AL-L%.MET
 Calibration File: G:\Org\HP5\Cals\SW8015C_ORO211017AL.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for for ~~Residual~~ TEH (Oil Range) Organics Calculations: 28542.41
 Rt range for ~~Residual~~ TEH (Oil Range) Organics: 12.56 to 30.05

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*#Triacontane	16.29	500.	379.474	75.89	-


RRO TEH (Oil Range) Area:1.44784E+08 RRO TEH (Oil Range) AMOUNT: 5072.594

CONTINUING CALIBRATION REPORT: G:\org\HP5\DAT\HP5122621_b\1226HP5.0004.RAW

COMPOUND	ACTUAL (NG)	MEASURED (NG)	%RECOVERY	LIMITS
*30-40 Motor Oil	5000.	.048	.	75-125

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	LIMITS
*#Triacontane	16.29	200.	379.474	189.74	75-125

AMN 01/11/2022



RESIDUAL RANGE ORGANICS CHROMATOGRAM

Sample Name: CCV_1226HP504r, RRO ;1226HP5 , DRO211220C
 Raw File: G:\org\HP5\DAT\HP5122621_b\1226HP5.0004.RAW
 Date & Time Acquired: 12/26/2021 1:24:41 PM
 Method File: G:\Org\HP5\Methods\DS_ORO-AL-L%.MET
 Calibration File: G:\Org\HP5\Cals\SW8015C_ORO211017AL.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for for Residual Range Organics Calculations: 28542.41
 Rt range for Residual Range Organics: 12.56 to 30.05

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*#Triacontane	16.29	500.	231.352	46.27	-

RRO Area:7208120 RRO AMOUNT: 252.5407


CONTINUING CALIBRATION REPORT: G:\org\HP5\DAT\HP5122621_b\1226HP5.0004.RAW

COMPOUND	ACTUAL (NG)	MEASURED (NG)	%RECOVERY	LIMITS
*30-40 Motor Oil	5000.	.048	.	75-125

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	LIMITS
*#Triacontane	16.29	200.	231.352	115.68	75-125

G:\org\HP5\DAT\HP5122621_b\1226HP5.0005.RAW

CCV_1226HP505r, DRO ;1226HP5 , DRO211220A



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: CCV_1226HP505r, DRO ;1226HP5 , DRO211220A
 Raw File: G:\org\HP5\DAT\HP5122621_b\1226HP5.0005.RAW
 Date & Time Acquired: 12/26/2021 2:07:41 PM
 Method File: G:\Org\HP5\Methods\DC_8015-24-IL-L%.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO211102IL-24.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for TEH: 31353.19

Rt range for Diesel Range Organics: 6.51 to 14.37

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*o-Terphenyl	12.221	200.	351.824	175.91
*1-Chlorooctadecane	13.064	200.	170.308	85.15

DRO Area: 4.966392E+08 DRO Amount: 15840.15
 TEH Area: 5.150671E+08 TEH Amount: 16427.9


CONTINUING CALIBRATION REPORT: G:\org\HP5\DAT\HP5122621_b\1226HP5.0005.RAW

COMPOUND	ACTUAL (NG)	MEASURED (NG)	%RECOVERY	LIMITS
TOTAL DRO	15000.	16427.9	109.52	85-115

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	LIMITS
*o-Terphenyl	12.221	200.	351.824	175.91	85-115
*1-Chlorooctadecane	13.064	200.	170.308	85.15	85-115

G:\org\HP5\DAT\HP5122621_b\1226HP5.0005.RAW

CCV_1226HP505r, DRO ;1226HP5 , DRO211220A



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: CCV_1226HP505r, DRO ;1226HP5 , DRO211220A
 Raw File: G:\org\HP5\DAT\HP5122621_b\1226HP5.0005.RAW
 Date & Time Acquired: 12/26/2021 2:07:41 PM
 Method File: G:\Org\HP5\Methods\DS_8015-24-IL-L#.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO211102IL-24.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for TEH: 31353.19

Rt range for Diesel Range Organics: 6.51 to 14.37


SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*o-Terphenyl	12.221	200.	213.3	106.65
*1-Chlorooctadecane	13.064	200.	40.032	20.02

DRO Area: 2.765112E+08 DRO Amount: 8819.237
 TEH Area: 2.879809E+08 TEH Amount: 9185.059

CONTINUING CALIBRATION REPORT: G:\org\HP5\DAT\HP5122621_b\1226HP5.0005.RAW

COMPOUND	ACTUAL (NG)	MEASURED (NG)	%RECOVERY	LIMITS
TOTAL DRO	15000.	9185.06	61.23	85-115

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	LIMITS
*o-Terphenyl	12.221	200.	213.3	106.65	85-115
*1-Chlorooctadecane	13.064	200.	40.032	20.02	85-115



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: DCM-Baseline Check-V06
 Raw File: G:\org\HP5\DAT\HP5122621_b\1226HP5.0006.RAW
 Date & Time Acquired: 12/26/2021 2:50:40 PM
 Method File: G:\Org\HP5\Methods\DR_8015-IBb-LEXP.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO211102IB.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for TEH: 31353.19
 Rt range for Diesel Range Organics: 6.54 to 15.86


SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*o-Terphenyl	12.218	200.	.313	.16	-
*1-Chlorooctadecane	13.066	200.	.038	.02	-


DRO Area:355889.2 DRO Amount: 11.35097
 TEH Area:464188.4 TEH Amount: 14.80514

Batch ID: 162352

G:\org\HP5\DAT\HP5122621_b\1226HP5.0007.RAW

LCS-162352 ;1226HP5 ,





DIESEL RANGE ORGANICS CHROMATOGRAM REPORT


Sample Name: MB-162352 ;1226HP5 , SGT
 Raw File: G:\org\HP5\DAT\HP5122621_b\1226HP5.0008.RAW
 Date & Time Acquired: 12/26/2021 4:16:38 PM
 Method File: G:\Org\HP5\Methods\DR_8015-C24T-IL-L%.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO211102IL-24-Tri.CAL
 Sample Weight: 1000 Dilution: 1 S.A.: 1

Mean RF for TEH: 31353.19

Rt range for Diesel Range Organics: 6.51 to 14.37

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*o-Terphenyl	12.215	.2	.218	109.05	-
*1-Chlorooctadecane	13.106	.2	.	.03	-
*#Triacontane	16.285	.2	.118	59.03	-

DRO Area:493325.4 DRO Amount: 1.573446E-02
 TEH Area:730344.6 TEH Amount: 2.329411E-02



RESIDUAL RANGE ORGANICS CHROMATOGRAM


Sample Name: MB-162352 ;1226HP5 , SGT
 Raw File: G:\org\HP5\DAT\HP5122621_b\1226HP5.0008.RAW
 Date & Time Acquired: 12/26/2021 4:16:38 PM
 Method File: G:\Org\HP5\Methods\DR_OROS-AL-L%.MET
 Calibration File: G:\Org\HP5\Cals\SW8015C_ORO211017AL-SAMP.CAL
 Sample Weight: 1000 Dilution: 1 S.A.: 1

Mean RF for for Residual Range Organics Calculations: 28542.41
 Rt range for Residual Range Organics: 14.27 to 21.54

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*#Triacontane	16.285	.5	.118	23.61

RRO Area:189282.5 RRO AMOUNT: 6.631623E-03

Batch ID: 162352
G:\org\HP5\DAT\HP5122621_b\1226HP5.0008.RAW MB-162352 ;1226HP5 , SGT



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: MB-162352 ;1226HP5 , SGT
 Raw File: G:\org\HP5\DAT\HP5122621_b\1226HP5.0008.RAW
 Date & Time Acquired: 12/26/2021 4:16:38 PM
 Method File: G:\Org\HP5\Methods\DS_8015-C24T-IL-L#.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO211102IL-24-Tri.CAL
 Sample Weight: 1000 Dilution: 1 S.A.: 1

Mean RF for TEH: 31353.19
 Rt range for Diesel Range Organics: 6.51 to 14.37

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*o-Terphenyl	12.215	.2	.218	108.83
*1-Chlorooctadecane	13.022	.2	.1	-
*#Triacontane	16.285	.2	.118	58.78


DRO Area:206228.3 DRO Amount: 6.577585E-03
 TEH Area:471035.9 TEH Amount: 1.502354E-02

ERH2184 (RHMW05)

Batch ID: 162352

G:\org\HP5\DAT\HP5122621_b\1226HP5.0009.RAW

B21121613-001B ;1226HP5 , \$HC-8015-DRO-W, SGT



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: B21121613-001B ;1226HP5 , \$HC-8015-DRO-W, SGT
 Raw File: G:\org\HP5\DAT\HP5122621_b\1226HP5.0009.RAW
 Date & Time Acquired: 12/26/2021 4:59:41 PM
 Method File: G:\Org\HP5\Methods\DR_8015-C24T-IL-L%.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO211102IL-24-Tri.CAL
 Sample Weight: 990 Dilution: 1 S.A.: 1

Mean RF for TEH: 31353.19

Rt range for Diesel Range Organics: 6.51 to 14.37

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*o-Terphenyl	12.213	.202	.154	76.36	-
*1-Chlorooctadecane	13.106	.202	.	.02	-
*#Triacontane	16.28	.202	.083	41.2	-


DRO Area:471021.8 DRO Amount: 1.517484E-02
 TEH Area:634403.6 TEH Amount: 2.043848E-02

ERH2184 (RHMW05)

Batch ID: 162352

G:\org\HP5\DAT\HP5122621_b\1226HP5.0009.RAW

B21121613-001B ;1226HP5 , \$HC-8015-DRO-W, SGT



RESIDUAL RANGE ORGANICS CHROMATOGRAM

Sample Name: B21121613-001B ;1226HP5 , \$HC-8015-DRO-W, SGT
 Raw File: G:\org\HP5\DAT\HP5122621_b\1226HP5.0009.RAW
 Date & Time Acquired: 12/26/2021 4:59:41 PM
 Method File: G:\Org\HP5\Methods\DR_OROS-AL-L%.MET
 Calibration File: G:\Org\HP5\Cals\SW8015C_ORO211017AL-SAMP.CAL
 Sample Weight: 990 Dilution: 1 S.A.: 1

Mean RF for for Residual Range Organics Calculations: 28542.41
 Rt range for Residual Range Organics: 14.27 to 21.54

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*#Triacontane	16.28	.505	.083	16.48


RRO Area:129915.1 RRO AMOUNT: 4.597627E-03

ERH2184 (RHMW05)

Batch ID: 162352

G:\org\HP5\DAT\HP5122621_b\1226HP5.0009.RAW

B21121613-001B ;1226HP5 , \$HC-8015-DRO-W, SGT



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: B21121613-001B ;1226HP5 , \$HC-8015-DRO-W, SGT
 Raw File: G:\org\HP5\DAT\HP5122621_b\1226HP5.0009.RAW
 Date & Time Acquired: 12/26/2021 4:59:41 PM
 Method File: G:\Org\HP5\Methods\DS_8015-C24T-IL-L#.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO211102IL-24-Tri.CAL
 Sample Weight: 990 Dilution: 1 S.A.: 1

Mean RF for TEH: 31353.19

Rt range for Diesel Range Organics: 6.51 to 14.37


SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*o-Terphenyl	12.213	.202	.154	76.15	-
*1-Chlorooctadecane	13.022	.202	.	.06	-
*#Triacontane	16.28	.202	.083	41.03	-

DRO Area:180099.7 DRO Amount: 5.802245E-03
 TEH Area:369090 TEH Amount: 1.189092E-02

Batch ID: 162352

B21121613-001BMS ;1226HP5 , SGT

G:\org\HP5\DAT\HP5122621_b\1226HP5.0010.RAW



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: B21121613-001BMS ;1226HP5 , SGT
 Raw File: G:\org\HP5\DAT\HP5122621_b\1226HP5.0010.RAW
 Date & Time Acquired: 12/26/2021 5:42:41 PM
 Method File: G:\Org\HP5\Methods\D3_8015-24-IL-L%.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO211102IL-24.CAL
 Sample Weight: 990 Dilution: 1 S.A.: 1


Mean RF for TEH: 31353.19

Rt range for Diesel Range Organics: 6.51 to 14.37

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*o-Terphenyl	12.219	.202	.376	186.21	-
*1-Chlorooctadecane	13.064	.202	.178	88.2	-

DRO Area: 4.916715E+08 DRO Amount: 15.84011
 TEH Area: 5.225287E+08 TEH Amount: 16.83423

Batch ID: 162352
G:\org\HP5\DAT\HP5122621_b\1226HP5.0010.RAW B21121613-001BMS ;1226HP5 , SGT



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: B21121613-001BMS ;1226HP5 , SGT
 Raw File: G:\org\HP5\DAT\HP5122621_b\1226HP5.0010.RAW
 Date & Time Acquired: 12/26/2021 5:42:41 PM
 Method File: G:\Org\HP5\Methods\DS_8015-24-IL-L#.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO211102IL-24.CAL
 Sample Weight: 990 Dilution: 1 S.A.: 1

Mean RF for TEH: 31353.19


Rt range for Diesel Range Organics: 6.51 to 14.37

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*o-Terphenyl	12.219	.202	.216	106.73
*1-Chlorooctadecane	13.064	.202	.072	35.76

DRO Area: 2.379687E+08 DRO Amount: 7.666602
 TEH Area: 2.525714E+08 TEH Amount: 8.137053

Batch ID: 162352
B21121613-001BMSD ;1226HP5 , SGT

G:\org\HP5\DAT\HP5122621_b\1226HP5.0011.RAW



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: B21121613-001BMSD ;1226HP5 , SGT
 Raw File: G:\org\HP5\DAT\HP5122621_b\1226HP5.0011.RAW
 Date & Time Acquired: 12/26/2021 6:25:39 PM
 Method File: G:\Org\HP5\Methods\D3_8015-24-IL-L%.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO211102IL-24.CAL
 Sample Weight: 1040 Dilution: 1 S.A.: 1


Mean RF for TEH: 31353.19

Rt range for Diesel Range Organics: 6.51 to 14.37

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*o-Terphenyl	12.22	.192	.37	192.2	-
*1-Chlorooctadecane	13.064	.192	.172	89.36	-

DRO Area: 4.760504E+08 DRO Amount: 14.5995
 TEH Area: 5.059904E+08 TEH Amount: 15.5177

Batch ID: 162352
G:\org\HP5\DAT\HP5122621_b\1226HP5.0011.RAW B21121613-001BMSD ;1226HP5 , SGT



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT


Sample Name: B21121613-001BMSD ;1226HP5 , SGT
 Raw File: G:\org\HP5\DAT\HP5122621_b\1226HP5.0011.RAW
 Date & Time Acquired: 12/26/2021 6:25:39 PM
 Method File: G:\Org\HP5\Methods\DS_8015-24-IL-L#.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO211102IL-24.CAL
 Sample Weight: 1040 Dilution: 1 S.A.: 1

Mean RF for TEH: 31353.19

Rt range for Diesel Range Organics: 6.51 to 14.37

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*o-Terphenyl	12.22	.192	.21	109.01	-
*1-Chlorooctadecane	13.064	.192	.07	36.28	-

DRO Area: 2.214392E+08 DRO Amount: 6.791088
 TEH Area: 2.351863E+08 TEH Amount: 7.212684



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: DCM-Baseline Check-V12
 Raw File: G:\org\HP5\DAT\HP5122621_b\1226HP5.0012.RAW
 Date & Time Acquired: 12/26/2021 7:08:39 PM
 Method File: G:\Org\HP5\Methods\DR_8015-IBb-LEXP.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO211102IB.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for TEH: 31353.19
 Rt range for Diesel Range Organics: 6.54 to 15.86

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*o-Terphenyl	12.215	200.	.298	.15	-
*1-Chlorooctadecane	13.069	200.	.042	.02	-


DRO Area:426971.7 DRO Amount: 13.61812
 TEH Area:551501.4 TEH Amount: 17.58996

ERH2193 (RHMW12A)

Batch ID: 162352

G:\org\HP5\DAT\HP5122621_b\1226HP5.0013.RAW

B21121622-001B ;1226HP5 , \$HC-8015-DRO-W, SGT



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: B21121622-001B ;1226HP5 , \$HC-8015-DRO-W, SGT
 Raw File: G:\org\HP5\DAT\HP5122621_b\1226HP5.0013.RAW
 Date & Time Acquired: 12/26/2021 7:51:32 PM
 Method File: G:\Org\HP5\Methods\DR_8015-C24T-IL-L%.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO211102IL-24-Tri.CAL
 Sample Weight: 960 Dilution: 1 S.A.: 1

Mean RF for TEH: 31353.19

Rt range for Diesel Range Organics: 6.51 to 14.37

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*o-Terphenyl	12.211	.208	.19	91.34	-
*1-Chlorooctadecane	13.117	.208	.	.01	-
*#Triacontane	16.276	.208	.101	48.65	-


DRO Area: 680582.4 DRO Amount: 2.261142E-02
 TEH Area: 995765.7 TEH Amount: 3.308295E-02

ERH2193 (RHMW12A)

Batch ID: 162352

G:\org\HP5\DAT\HP5122621_b\1226HP5.0013.RAW

B21121622-001B ;1226HP5 , \$HC-8015-DRO-W, SGT



RESIDUAL RANGE ORGANICS CHROMATOGRAM

Sample Name: B21121622-001B ;1226HP5 , \$HC-8015-DRO-W, SGT
 Raw File: G:\org\HP5\DAT\HP5122621_b\1226HP5.0013.RAW
 Date & Time Acquired: 12/26/2021 7:51:32 PM
 Method File: G:\Org\HP5\Methods\DR_OROS-AL-L%.MET
 Calibration File: G:\Org\HP5\Cals\SW8015C_ORO211017AL-SAMP.CAL
 Sample Weight: 960 Dilution: 1 S.A.: 1

Mean RF for for Residual Range Organics Calculations: 28542.41
 Rt range for Residual Range Organics: 14.27 to 21.54

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*#Triacontane	16.276	.521	.101	19.46

RRO Area:218408

RRO AMOUNT: 7.970887E-03




ERH2193 (RHMW12A)

Batch ID: 162352

G:\org\HP5\DAT\HP5122621_b\1226HP5.0013.RAW

B21121622-001B ;1226HP5 , \$HC-8015-DRO-W, SGT



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: B21121622-001B ;1226HP5 , \$HC-8015-DRO-W, SGT
Raw File: G:\org\HP5\DAT\HP5122621_b\1226HP5.0013.RAW
Date & Time Acquired: 12/26/2021 7:51:32 PM
Method File: G:\Org\HP5\Methods\DS_8015-C24T-IL-L#.met
Calibration File: G:\Org\HP5\Cals\SW8015C_DRO211102IL-24-Tri.CAL
Sample Weight: 960 Dilution: 1 S.A.: 1

Mean RF for TEH: 31353.19

Rt range for Diesel Range Organics: 6.51 to 14.37

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*o-Terphenyl	12.211	.208	.19	91.12	-
*1-Chlorooctadecane	29.978	.208	.		-
*#Triacontane	16.276	.208	.101	48.4	-


DRO Area:287369.7 DRO Amount: 9.547462E-03
TEH Area:615266.6 TEH Amount: 2.044139E-02

ERH2175 (RHMW2254-001)

Batch ID: 162352

G:\org\HP5\DAT\HP5122621_b\1226HP5.0014.RAW

B21121605-001A ;1226HP5 , \$HC-8015-DRO-W, SGT



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: B21121605-001A ;1226HP5 , \$HC-8015-DRO-W, SGT
 Raw File: G:\org\HP5\DAT\HP5122621_b\1226HP5.0014.RAW
 Date & Time Acquired: 12/26/2021 8:34:21 PM
 Method File: G:\Org\HP5\Methods\DR_8015-C24T-IL-L%.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO211102IL-24-Tri.CAL
 Sample Weight: 1050 Dilution: 1 S.A.: 1

Mean RF for TEH: 31353.19

Rt range for Diesel Range Organics: 6.51 to 14.37

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*o-Terphenyl	12.212	.19	.158	83.06	-
*1-Chlorooctadecane	13.022	.19	.	.11	-
*#Triacontane	16.277	.19	.082	43.03	-

DRO Area:1966356
 TEH Area:2154859


DRO Amount: 0.0597298
 TEH Amount: 6.545574E-02

ERH2175 (RHMW2254-001)

Batch ID: 162352

G:\org\HP5\DAT\HP5122621_b\1226HP5.0014.RAW

B21121605-001A ;1226HP5 , \$HC-8015-DRO-W, SGT



RESIDUAL RANGE ORGANICS CHROMATOGRAM

Sample Name: B21121605-001A ;1226HP5 , \$HC-8015-DRO-W, SGT
 Raw File: G:\org\HP5\DAT\HP5122621_b\1226HP5.0014.RAW
 Date & Time Acquired: 12/26/2021 8:34:21 PM
 Method File: G:\Org\HP5\Methods\DR_OROS-AL-L%.MET
 Calibration File: G:\Org\HP5\Cals\SW8015C_ORO211017AL-SAMP.CAL
 Sample Weight: 1050 Dilution: 1 S.A.: 1

Mean RF for for Residual Range Organics Calculations: 28542.41
 Rt range for Residual Range Organics: 14.27 to 21.54

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*#Triacontane	16.277	.476	.082	17.21

RRO Area:146590.7 RRO AMOUNT: 4.891324E-03




ERH2175 (RHMW2254-001)

Batch ID: 162352

G:\Org\HP5\DAT\HP5122621_b\1226HP5.0014.RAW

B21121605-001A ; 1226HP5 , \$HC-8015-DRO-W, SGT



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: B21121605-001A ; 1226HP5 , \$HC-8015-DRO-W, SGT
Raw File: G:\Org\HP5\DAT\HP5122621_b\1226HP5.0014.RAW
Date & Time Acquired: 12/26/2021 8:34:21 PM
Method File: G:\Org\HP5\Methods\DS_8015-C24T-IL-L#.met
Calibration File: G:\Org\HP5\Cals\SW8015C_DRO211102IL-24-Tri.CAL
Sample Weight: 1050 Dilution: 1 S.A.: 1

Mean RF for TEH: 31353.19

Rt range for Diesel Range Organics: 6.51 to 14.37

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*o-Terphenyl	12.212	.19	.158	82.81	-
*1-Chlorooctadecane	13.022	.19	.	.05	-
*#Triacontane	16.277	.19	.081	42.79	-


DRO Area:1280950 DRO Amount: 0.03891
TEH Area:1452287 TEH Amount: 4.411451E-02

ERH2178 (RHMW01R)

Batch ID: 162352

G:\org\HP5\DAT\HP5122621_b\1226HP5.0015.RAW

B21121623-001C ;1226HP5 , \$HC-8015-DRO-W, SGT



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: B21121623-001C ;1226HP5 , \$HC-8015-DRO-W, SGT
 Raw File: G:\org\HP5\DAT\HP5122621_b\1226HP5.0015.RAW
 Date & Time Acquired: 12/26/2021 9:17:18 PM
 Method File: G:\Org\HP5\Methods\DR_8015-C24T-IL-L%.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO211102IL-24-Tri.CAL
 Sample Weight: 1020 Dilution: 1 S.A.: 1

Mean RF for TEH: 31353.19

Rt range for Diesel Range Organics: 6.51 to 14.37

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*o-Terphenyl	12.214	.196	.193	98.2	-
*1-Chlorooctadecane	13.056	.196	.	.04	-
*#Triacontane	16.279	.196	.112	57.05	-


DRO Area:1858886 DRO Amount: 5.812604E-02
 TEH Area:2119170 TEH Amount: 6.626495E-02

ERH2178 (RHMW01R)

Batch ID: 162352

G:\org\HP5\DAT\HP5122621_b\1226HP5.0015.RAW

B21121623-001C ;1226HP5 , \$HC-8015-DRO-W, SGT



RESIDUAL RANGE ORGANICS CHROMATOGRAM

Sample Name: B21121623-001C ;1226HP5 , \$HC-8015-DRO-W, SGT
 Raw File: G:\org\HP5\DAT\HP5122621_b\1226HP5.0015.RAW
 Date & Time Acquired: 12/26/2021 9:17:18 PM
 Method File: G:\Org\HP5\Methods\DR_OROS-AL-L%.MET
 Calibration File: G:\Org\HP5\Cals\SW8015C_ORO211017AL-SAMP.CAL
 Sample Weight: 1020 Dilution: 1 S.A.: 1

Mean RF for for Residual Range Organics Calculations: 28542.41
 Rt range for Residual Range Organics: 14.27 to 21.54

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*#Triacontane_____	16.279	.49	.112	22.82	-


RRO Area:179215.4 RRO AMOUNT: 0.0061558

ERH2178 (RHMW01R)

Batch ID: 162352

G:\org\HP5\DAT\HP5122621_b\1226HP5.0015.RAW

B21121623-001C ; 1226HP5 , \$HC-8015-DRO-W, SGT



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT


Sample Name: B21121623-001C ; 1226HP5 , \$HC-8015-DRO-W, SGT
 Raw File: G:\org\HP5\DAT\HP5122621_b\1226HP5.0015.RAW
 Date & Time Acquired: 12/26/2021 9:17:18 PM
 Method File: G:\Org\HP5\Methods\DS_8015-C24T-IL-L#.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO211102IL-24-Tri.CAL
 Sample Weight: 1020 Dilution: 1 S.A.: 1

Mean RF for TEH: 31353.19

Rt range for Diesel Range Organics: 6.51 to 14.37

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*o-Terphenyl	12.214	.196	.192	97.95	-
*1-Chlorooctadecane	13.022	.196	.	.02	-
*Triacontane	16.279	.196	.111	56.79	-

DRO Area: 1364879 DRO Amount: 4.267881E-02
 TEH Area: 1628100 TEH Amount: 5.090954E-02



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: DCM-Baseline Check-V16
 Raw File: G:\org\HP5\DAT\HP5122621_b\1226HP5.0016.RAW
 Date & Time Acquired: 12/26/2021 10:00:07 PM
 Method File: G:\Org\HP5\Methods\DR_8015-IBb-LEXP.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO211102IB.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for TEH: 31353.19

Rt range for Diesel Range Organics: 6.54 to 15.86

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*o-Terphenyl	12.215	200.	.365	.18	-
*1-Chlorooctadecane	13.074	200.	.022	.01	-


DRO Area: 576114.7 DRO Amount: 18.37499
 TEH Area: 754904.2 TEH Amount: 24.07743

ERH2182 (RHMW03)

Batch ID: 162352

G:\org\HP5\DAT\HP5122621_b\1226HP5.0017.RAW

B21121611-001B ;1226HP5 , \$HC-8015-DRO-W, SGT



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: B21121611-001B ;1226HP5 , \$HC-8015-DRO-W, SGT
 Raw File: G:\org\HP5\DAT\HP5122621_b\1226HP5.0017.RAW
 Date & Time Acquired: 12/26/2021 10:43:04 PM
 Method File: G:\Org\HP5\Methods\DR_8015-C24T-IL-L%.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO211102IL-24-Tri.CAL
 Sample Weight: 1030 Dilution: 1 S.A.: 1

Mean RF for TEH: 31353.19

Rt range for Diesel Range Organics: 6.51 to 14.37

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*o-Terphenyl	12.214	.194	.2	103.14	-
*1-Chlorooctadecane	13.059	.194	.	.08	-
*#Triacontane	16.279	.194	.106	54.63	-


DRO Area:815550.4 DRO Amount: 2.525409E-02
 TEH Area:1293165 TEH Amount: 4.004378E-02

ERH2182 (RHMW03)

Batch ID: 162352

G:\org\HP5\DAT\HP5122621_b\1226HP5.0017.RAW

B21121611-001B ;1226HP5 , \$HC-8015-DRO-W, SGT



RESIDUAL RANGE ORGANICS CHROMATOGRAM

Sample Name: B21121611-001B ;1226HP5 , \$HC-8015-DRO-W, SGT
 Raw File: G:\org\HP5\DAT\HP5122621_b\1226HP5.0017.RAW
 Date & Time Acquired: 12/26/2021 10:43:04 PM
 Method File: G:\Org\HP5\Methods\DR_OROS-AL-L%.MET
 Calibration File: G:\Org\HP5\Cals\SW8015C_ORO211017AL-SAMP.CAL
 Sample Weight: 1030 Dilution: 1 S.A.: 1

Mean RF for for Residual Range Organics Calculations: 28542.41
 Rt range for Residual Range Organics: 14.27 to 21.54

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*#Triacontane	16.279	.485	.106	21.85


RRO Area:377598.7 RRO AMOUNT: 1.284407E-02

ERH2182 (RHMW03)

Batch ID: 162352

G:\org\HP5\DAT\HP5122621_b\1226HP5.0017.RAW

B21121611-001B ;1226HP5 , \$HC-8015-DRO-W, SGT



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: B21121611-001B ;1226HP5 , \$HC-8015-DRO-W, SGT
 Raw File: G:\org\HP5\DAT\HP5122621_b\1226HP5.0017.RAW
 Date & Time Acquired: 12/26/2021 10:43:04 PM
 Method File: G:\Org\HP5\Methods\DS_8015-C24T-IL-L#.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO211102IL-24-Tri.CAL
 Sample Weight: 1030 Dilution: 1 S.A.: 1

Mean RF for TEH: 31353.19

Rt range for Diesel Range Organics: 6.51 to 14.37

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*o-Terphenyl	12.214	.194	.2	102.84	-
*1-Chlorooctadecane	13.059	.194	.	.02	-
*#Triacontane	16.279	.194	.106	54.35	-


DRO Area:347881.4 DRO Amount: 1.077239E-02
 TEH Area:819693.3 TEH Amount: 2.538238E-02

ERH2195 (RHMW16)

Batch ID: 162352

G:\org\HP5\DAT\HP5122621_b\1226HP5.0018.RAW

B21121622-002B ;1226HP5 , \$HC-8015-DRO-W, SGT



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: B21121622-002B ;1226HP5 , \$HC-8015-DRO-W, SGT
 Raw File: G:\org\HP5\DAT\HP5122621_b\1226HP5.0018.RAW
 Date & Time Acquired: 12/26/2021 11:25:55 PM
 Method File: G:\Org\HP5\Methods\DR_8015-C24T-IL-L%.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO211102IL-24-Tri.CAL
 Sample Weight: 1050 Dilution: 1 S.A.: 1

Mean RF for TEH: 31353.19

Rt range for Diesel Range Organics: 6.51 to 14.37

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*o-Terphenyl	12.211	.19	.156	81.7	-
*1-Chlorooctadecane	13.058	.19	.	.05	-
*#Triacontane	16.279	.19	.083	43.8	-


DRO Area: 646287.3 DRO Amount: 1.963155E-02
 TEH Area: 1010775 TEH Amount: 3.070318E-02

ERH2195 (RHMW16)

Batch ID: 162352

G:\org\HP5\DAT\HP5122621_b\1226HP5.0018.RAW

B21121622-002B ;1226HP5 , \$HC-8015-DRO-W, SGT



RESIDUAL RANGE ORGANICS CHROMATOGRAM

Sample Name: B21121622-002B ;1226HP5 , \$HC-8015-DRO-W, SGT
 Raw File: G:\org\HP5\DAT\HP5122621_b\1226HP5.0018.RAW
 Date & Time Acquired: 12/26/2021 11:25:55 PM
 Method File: G:\Org\HP5\Methods\DR_OROS-AL-L%.MET
 Calibration File: G:\Org\HP5\Cals\SW8015C_ORO211017AL-SAMP.CAL
 Sample Weight: 1050 Dilution: 1 S.A.: 1

Mean RF for for Residual Range Organics Calculations: 28542.41
 Rt range for Residual Range Organics: 14.27 to 21.54

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*#Triacontane	16.279	.476	.083	17.52


RRO Area:129004.1 RRO AMOUNT: 4.304509E-03

ERH2195 (RHMW16)

Batch ID: 162352

G:\Org\HP5\DAT\HP5122621_b\1226HP5.0018.RAW

B21121622-002B ;1226HP5 , \$HC-8015-DRO-W, SGT



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT


Sample Name: B21121622-002B ;1226HP5 , \$HC-8015-DRO-W, SGT
 Raw File: G:\Org\HP5\DAT\HP5122621_b\1226HP5.0018.RAW
 Date & Time Acquired: 12/26/2021 11:25:55 PM
 Method File: G:\Org\HP5\Methods\DS_8015-C24T-IL-L#.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO211102IL-24-Tri.CAL
 Sample Weight: 1050 Dilution: 1 S.A.: 1


Mean RF for TEH: 31353.19

Rt range for Diesel Range Organics: 6.51 to 14.37

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*o-Terphenyl	12.211	.19	.155	81.46	-
*1-Chlorooctadecane	13.022	.19	.	.02	-
*#Triacontane	16.279	.19	.083	43.55	-

DRO Area:248425.1 DRO Amount: 7.546134E-03
 TEH Area:620805.1 TEH Amount: 0.0188575





RESIDUAL RANGE ORGANICS CHROMATOGRAM

Sample Name: CCV_1226HP504r, RRO ;1226HP5 , DRO211220C
 Raw File: G:\org\HP5\DAT\HP5122621_b\1226HP5.0020.RAW
 Date & Time Acquired: 12/27/2021 12:51:35 AM
 Method File: G:\Org\HP5\Methods\DC_ORO-AL-L%.MET
 Calibration File: G:\Org\HP5\Cals\SW8015C_ORO211017AL.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for for ~~Residual~~ TEH (Oil Range) Organics Calculations: 28542.41
 Rt range for ~~Residual~~ TEH (Oil Range) Organics: 12.56 to 30.05

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*#Triacontane	16.287	500.	362.188	72.44	-

RRO TEH (Oil Range) Area:1.392909E+08 RRO TEH (Oil Range) AMOUNT: 4880.14

CONTINUING CALIBRATION REPORT: G:\org\HP5\DAT\HP5122621_b\1226HP5.0020.RAW


COMPOUND	ACTUAL (NG)	MEASURED (NG)	%RECOVERY	LIMITS
*30-40 Motor Oil	5000.	.063	.	75-125

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	LIMITS
*#Triacontane	16.287	200.	362.188	181.09	75-125

AMN 01/11/2022

G:\org\HP5\DAT\HP5122621_b\1226HP5.0020.RAW

CCV_1226HP504r, RRO ;1226HP5 , DRO211220C



RESIDUAL RANGE ORGANICS CHROMATOGRAM

Sample Name: CCV_1226HP504r, RRO ;1226HP5 , DRO211220C
 Raw File: G:\org\HP5\DAT\HP5122621_b\1226HP5.0020.RAW
 Date & Time Acquired: 12/27/2021 12:51:35 AM
 Method File: G:\Org\HP5\Methods\DS_ORO-AL-L%.MET
 Calibration File: G:\Org\HP5\Cals\SW8015C_ORO211017AL.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for for Residual Range Organics Calculations: 28542.41
 Rt range for Residual Range Organics: 12.56 to 30.05

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*#Triacontane	16.287	500.	222.05	44.41	-

RRO Area:6503688 RRO AMOUNT: 227.8605


CONTINUING CALIBRATION REPORT: G:\org\HP5\DAT\HP5122621_b\1226HP5.0020.RAW

COMPOUND	ACTUAL (NG)	MEASURED (NG)	%RECOVERY	LIMITS
*30-40 Motor Oil	5000.	.063	.	75-125

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	LIMITS
*#Triacontane	16.287	200.	222.05	111.03	75-125

G:\org\HP5\DAT\HP5122621_b\1226HP5.0021.RAW

CCV_1226HP505r, DRO ;1226HP5 , DRO211220A



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: CCV_1226HP505r, DRO ;1226HP5 , DRO211220A
 Raw File: G:\org\HP5\DAT\HP5122621_b\1226HP5.0021.RAW
 Date & Time Acquired: 12/27/2021 1:34:33 AM
 Method File: G:\Org\HP5\Methods\DC_8015-24-IL-L%.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO211102IL-24.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for TEH: 31353.19

Rt range for Diesel Range Organics: 6.51 to 14.37

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*o-Terphenyl	12.218	200.	350.91	175.45
*1-Chlorooctadecane	13.058	200.	170.747	85.37

DRO Area: 4.973627E+08 DRO Amount: 15863.22
 TEH Area: 5.159366E+08 TEH Amount: 16455.63


CONTINUING CALIBRATION REPORT: G:\org\HP5\DAT\HP5122621_b\1226HP5.0021.RAW

COMPOUND	ACTUAL (NG)	MEASURED (NG)	%RECOVERY	LIMITS
TOTAL DRO	15000.	16455.63	109.7	85-115

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	LIMITS
*o-Terphenyl	12.218	200.	350.91	175.45	85-115
*1-Chlorooctadecane	13.058	200.	170.747	85.37	85-115

G:\org\HP5\DAT\HP5122621_b\1226HP5.0021.RAW

CCV_1226HP505r, DRO ;1226HP5 , DRO211220A



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: CCV_1226HP505r, DRO ;1226HP5 , DRO211220A
 Raw File: G:\org\HP5\DAT\HP5122621_b\1226HP5.0021.RAW
 Date & Time Acquired: 12/27/2021 1:34:33 AM
 Method File: G:\Org\HP5\Methods\DS_8015-24-IL-L#.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO211102IL-24.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for TEH: 31353.19

Rt range for Diesel Range Organics: 6.51 to 14.37


SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*o-Terphenyl	12.218	200.	212.438	106.22
*1-Chlorooctadecane	13.058	200.	39.892	19.95

DRO Area: 2.749491E+08 DRO Amount: 8769.412
 TEH Area: 2.863747E+08 TEH Amount: 9133.829

CONTINUING CALIBRATION REPORT: G:\org\HP5\DAT\HP5122621_b\1226HP5.0021.RAW

COMPOUND	ACTUAL (NG)	MEASURED (NG)	%RECOVERY	LIMITS
TOTAL DRO	15000.	9133.83	60.89	85-115

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	LIMITS
*o-Terphenyl	12.218	200.	212.438	106.22	85-115
*1-Chlorooctadecane	13.058	200.	39.892	19.95	85-115




DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: DCM-Baseline Check-V22
 Raw File: G:\org\HP5\DAT\HP5122621_b\1226HP5.0022.RAW
 Date & Time Acquired: 12/27/2021 2:17:25 AM
 Method File: G:\Org\HP5\Methods\DR_8015-IBb-LEXP.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO211102IB.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for TEH: 31353.19
 Rt range for Diesel Range Organics: 6.54 to 15.86

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*o-Terphenyl	12.214	200.	.317	.16	-
*1-Chlorooctadecane	13.067	200.	.024	.01	-

DRO Area: 468725.9 DRO Amount: 14.94986
 TEH Area: 629374.6 TEH Amount: 20.0737



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: DCM-Baseline Check-V23
 Raw File: G:\org\HP5\DAT\HP5122621_b\1226HP5.0023.RAW
 Date & Time Acquired: 12/27/2021 3:00:22 AM
 Method File: G:\Org\HP5\Methods\DR_8015-IBb-LEXP.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO211102IB.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for TEH: 31353.19
 Rt range for Diesel Range Organics: 6.54 to 15.86

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*o-Terphenyl	12.216	200.	.296	.15	-
*1-Chlorooctadecane	13.021	200.	.027	.01	-


DRO Area:446954.4 DRO Amount: 14.25547
 TEH Area:585899.8 TEH Amount: 18.68709

ERH2180 (RHMW02)

Batch ID: 162352

G:\org\HP5\DAT\HP5122621_b\1226HP5.0024.RAW

B21121616-001C ;1226HP5 , \$HC-8015-DRO-W, SGT



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: B21121616-001C ;1226HP5 , \$HC-8015-DRO-W, SGT
 Raw File: G:\org\HP5\DAT\HP5122621_b\1226HP5.0024.RAW
 Date & Time Acquired: 12/27/2021 3:43:15 AM
 Method File: G:\Org\HP5\Methods\DR_8015-C24T-IL-L%.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO211102IL-24-Tri.CAL
 Sample Weight: 990 Dilution: 1 S.A.: 1

Mean RF for TEH: 31353.19

Rt range for Diesel Range Organics: 6.51 to 14.37

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*o-Terphenyl	12.21	.202	.146	72.04	-
*1-Chlorooctadecane	13.02	.202	.001	.28	-
*#Triacontane	16.278	.202	.097	47.94	-


DRO Area: 7969666 DRO Amount: 0.2567575
 TEH Area: 8381835 TEH Amount: 0.2700363

ERH2180 (RHMW02)

Batch ID: 162352

G:\org\HP5\DAT\HP5122621_b\1226HP5.0024.RAW

B21121616-001C ;1226HP5 , \$HC-8015-DRO-W,



RESIDUAL RANGE ORGANICS CHROMATOGRAM

Sample Name: B21121616-001C ;1226HP5 , \$HC-8015-DRO-W,
Raw File: G:\org\HP5\DAT\HP5122621_b\1226HP5.0024.RAW
Date & Time Acquired: 12/27/2021 3:43:15 AM
Method File: G:\Org\HP5\Methods\DR_OROS-AL-L%.MET
Calibration File: G:\Org\HP5\Cals\SW8015C_ORO211017AL-SAMP.CAL
Sample Weight: 990 Dilution: 1 S.A.: 1

Mean RF for for Residual Range Organics Calculations: 28542.41
Rt range for Residual Range Organics: 14.27 to 21.54

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*#Triacontane	16.278	.505	.097	19.18


RRO Area:209556 RRO AMOUNT: 7.416076E-03

ERH2180 (RHMW02)

Batch ID: 162352

G:\Org\HP5\DAT\HP5122621_b\1226HP5.0024.RAW

B21121616-001C ;1226HP5 , \$HC-8015-DRO-W,



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: B21121616-001C ;1226HP5 , \$HC-8015-DRO-W,
Raw File: G:\Org\HP5\DAT\HP5122621_b\1226HP5.0024.RAW
Date & Time Acquired: 12/27/2021 3:43:15 AM
Method File: G:\Org\HP5\Methods\DS_8015-C24T-IL-L#.met
Calibration File: G:\Org\HP5\Cals\SW8015C_DRO211102IL-24-Tri.CAL
Sample Weight: 990 Dilution: 1 S.A.: 1

Mean RF for TEH: 31353.19

Rt range for Diesel Range Organics: 6.51 to 14.37

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*o-Terphenyl	12.21	.202	.145	71.84	-
*1-Chlorooctadecane	13.02	.202	.	.21	-
*#Triacontane	16.278	.202	.096	47.61	-


DRO Area:7265083 DRO Amount: 0.2340581
TEH Area:7652590 TEH Amount: 0.2465423

ERH2206 (RHMW2254-01)

Batch ID: 162352

G:\org\HP5\DAT\HP5122621_b\1226HP5.0025.RAW

B21121605-002A ;1226HP5 , \$HC-8015-DRO-W, SGT



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: B21121605-002A ;1226HP5 , \$HC-8015-DRO-W, SGT
 Raw File: G:\org\HP5\DAT\HP5122621_b\1226HP5.0025.RAW
 Date & Time Acquired: 12/27/2021 4:26:07 AM
 Method File: G:\Org\HP5\Methods\D3_8015-C24T-IL-L%.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO211102IL-24-Tri.CAL
 Sample Weight: 1050 Dilution: 1 S.A.: 1

Mean RF for TEH: 31353.19

Rt range for Diesel Range Organics: 6.51 to 14.37

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*o-Terphenyl	12.211	.19	.189	99.3	-
*1-Chlorooctadecane	13.057	.19	.	.05	-
*#Triacontane	16.276	.19	.103	54.06	-


DRO Area: 4.140924E+07 DRO Amount: 1.257842
 TEH Area: 4.56042E+07 TEH Amount: 1.385268

ERH2206 (RHMW2254-01)

Batch ID: 162352

G:\org\HP5\DAT\HP5122621_b\1226HP5.0025.RAW

B21121605-002A ;1226HP5 , \$HC-8015-DRO-W, SGT



RESIDUAL RANGE ORGANICS CHROMATOGRAM

Sample Name: B21121605-002A ;1226HP5 , \$HC-8015-DRO-W, SGT
 Raw File: G:\org\HP5\DAT\HP5122621_b\1226HP5.0025.RAW
 Date & Time Acquired: 12/27/2021 4:26:07 AM
 Method File: G:\Org\HP5\Methods\D3_OROS-AL-L%.MET
 Calibration File: G:\Org\HP5\Cals\SW8015C_ORO211017AL-SAMP.CAL
 Sample Weight: 1050 Dilution: 1 S.A.: 1

Mean RF for for Residual Range Organics Calculations: 28542.41
 Rt range for Residual Range Organics: 14.27 to 21.54

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*#Triacontane	16.276	.476	.103	21.62


RRO Area:3100391 RRO AMOUNT: 0.1034514

ERH2206 (RHMW2254-01)

Batch ID: 162352

G:\org\HP5\DAT\HP5122621_b\1226HP5.0025.RAW

B21121605-002A ; 1226HP5 , \$HC-8015-DRO-W, SGT



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: B21121605-002A ; 1226HP5 , \$HC-8015-DRO-W, SGT
 Raw File: G:\org\HP5\DAT\HP5122621_b\1226HP5.0025.RAW
 Date & Time Acquired: 12/27/2021 4:26:07 AM
 Method File: G:\Org\HP5\Methods\DS_8015-C24T-IL-L#.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO211102IL-24-Tri.CAL
 Sample Weight: 1050 Dilution: 1 S.A.: 1

Mean RF for TEH: 31353.19

Rt range for Diesel Range Organics: 6.51 to 14.37

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*o-Terphenyl	12.211	.19	.189	99.	-
*1-Chlorooctadecane	13.057	.19	.	.02	-
*#Triacontane	16.276	.19	.099	51.79	-


DRO Area: 3.710659E+07 DRO Amount: 1.127145
 TEH Area: 3.79839E+07 TEH Amount: 1.153795

ERH2176 (RHMW2254-01)

Batch ID: 162352

G:\org\HP5\DAT\HP5122621_b\1226HP5.0026.RAW

B21121605-003A ;1226HP5 , \$HC-8015-DRO-W, SGT



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: B21121605-003A ;1226HP5 , \$HC-8015-DRO-W, SGT
 Raw File: G:\org\HP5\DAT\HP5122621_b\1226HP5.0026.RAW
 Date & Time Acquired: 12/27/2021 5:09:00 AM
 Method File: G:\Org\HP5\Methods\D3_8015-122626-IL-L%.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO211102IL-24-Tri.CAL
 Sample Weight: 1010 Dilution: 1 S.A.: 1

Mean RF for TEH: 31353.19

Rt range for Diesel Range Organics: 6.51 to 14.37

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*o-Terphenyl	12.213	.198	.182	91.98	-
*1-Chlorooctadecane	13.058	.198	.	.03	-
*#Triacontane	16.278	.198	.097	48.87	-


DRO Area: 4.426259E+07 DRO Amount: 1.397764
 TEH Area: 4.758381E+07 TEH Amount: 1.502644

ERH2176 (RHMW2254-01)

Batch ID: 162352

G:\org\HP5\DAT\HP5122621_b\1226HP5.0026.RAW

B21121605-003A ;1226HP5 , \$HC-8015-DRO-W, SGT



RESIDUAL RANGE ORGANICS CHROMATOGRAM

Sample Name: B21121605-003A ;1226HP5 , \$HC-8015-DRO-W, SGT
 Raw File: G:\org\HP5\DAT\HP5122621_b\1226HP5.0026.RAW
 Date & Time Acquired: 12/27/2021 5:09:00 AM
 Method File: G:\Org\HP5\Methods\D3_OROS-122626-AL-L%.MET
 Calibration File: G:\Org\HP5\Cals\SW8015C_ORO211017AL-SAMP.CAL
 Sample Weight: 1010 Dilution: 1 S.A.: 1

Mean RF for for Residual Range Organics Calculations: 28542.41
 Rt range for Residual Range Organics: 14.27 to 21.54

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*#Triacontane	16.278	.495	.097	19.55


RRO Area:2488046 RRO AMOUNT: 8.630706E-02

ERH2176 (RHMW2254-01)

Batch ID: 162352

G:\org\HP5\DAT\HP5122621_b\1226HP5.0026.RAW

B21121605-003A ; 1226HP5 , \$HC-8015-DRO-W, SGT



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: B21121605-003A ; 1226HP5 , \$HC-8015-DRO-W, SGT
 Raw File: G:\org\HP5\DAT\HP5122621_b\1226HP5.0026.RAW
 Date & Time Acquired: 12/27/2021 5:09:00 AM
 Method File: G:\Org\HP5\Methods\DS_8015-C24T-IL-L#.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO211102IL-24-Tri.CAL
 Sample Weight: 1010 Dilution: 1 S.A.: 1

Mean RF for TEH: 31353.19

Rt range for Diesel Range Organics: 6.51 to 14.37

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*o-Terphenyl	12.213	.198	.182	91.75	-
*1-Chlorooctadecane	13.022	.198	.	.03	-
*#Triacontane	16.278	.198	.093	46.92	-


DRO Area: 4.033741E+07 DRO Amount: 1.273811
 TEH Area: 4.118676E+07 TEH Amount: 1.300632

ERH2230 (Sump Adit 3)

Batch ID: 162352

G:\org\HP5\DAT\HP5122621_b\1226HP5.0027.RAW

B21121609-001C ; 1226HP5 , \$HC-8015-DRO-W, SGT



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: B21121609-001C ; 1226HP5 , \$HC-8015-DRO-W, SGT
 Raw File: G:\org\HP5\DAT\HP5122621_b\1226HP5.0027.RAW
 Date & Time Acquired: 12/27/2021 5:51:47 AM
 Method File: G:\Org\HP5\Methods\D3_8015-122627-IL-L%.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO211102IL-24-Tri.CAL
 Sample Weight: 1060 Dilution: 1 S.A.: 1

Mean RF for TEH: 31353.19

Rt range for Diesel Range Organics: 6.51 to 14.37

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*o-Terphenyl	12.213	.189	.194	102.82	-
*1-Chlorooctadecane	13.084	.189	.009	4.53	-
*#Triacontane	16.28	.189	.102	54.15	-


DRO Area: 1.321865E+08 DRO Amount: 3.977403
 TEH Area: 1.39549E+08 TEH Amount: 4.198934

ERH2230 (Sump Adit 3)

Batch ID: 162352

G:\org\HP5\DAT\HP5122621_b\1226HP5.0027.RAW

B21121609-001C ; 1226HP5 , \$HC-8015-DRO-W, SGT



RESIDUAL RANGE ORGANICS CHROMATOGRAM

Sample Name: B21121609-001C ; 1226HP5 , \$HC-8015-DRO-W, SGT
 Raw File: G:\org\HP5\DAT\HP5122621_b\1226HP5.0027.RAW
 Date & Time Acquired: 12/27/2021 5:51:47 AM
 Method File: G:\Org\HP5\Methods\D3_OROS-122627-AL-L%.MET
 Calibration File: G:\Org\HP5\Cals\SW8015C_ORO211017AL-SAMP.CAL
 Sample Weight: 1060 Dilution: 1 S.A.: 1

Mean RF for for Residual Range Organics Calculations: 28542.41
 Rt range for Residual Range Organics: 14.27 to 21.54

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*#Triacontane	16.28	.472	.102	21.66


RRO Area: 5354811 RRO AMOUNT: 0.1769896

ERH2230 (Sump Adit 3)

G:\Org\HP5\DAT\HP5122621_b\1226HP5.0027.RAW

Batch ID: 162352

B21121609-001C ; 1226HP5 , \$HC-8015-DRO-W, SGT



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT


Sample Name: B21121609-001C ; 1226HP5 , \$HC-8015-DRO-W, SGT
 Raw File: G:\Org\HP5\DAT\HP5122621_b\1226HP5.0027.RAW
 Date & Time Acquired: 12/27/2021 5:51:47 AM
 Method File: G:\Org\HP5\Methods\DS_8015-C24T-IL-L#.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO211102IL-24-Tri.CAL
 Sample Weight: 1060 Dilution: 1 S.A.: 1

Mean RF for TEH: 31353.19

Rt range for Diesel Range Organics: 6.51 to 14.37

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*o-Terphenyl	12.213	.189	.178	94.57	-
*1-Chlorooctadecane	13.084	.189	.	.08	-
*#Triacontane	16.28	.189	.095	50.6	-

DRO Area: 9.555618E+07 DRO Amount: 2.875221
 TEH Area: 9.704482E+07 TEH Amount: 2.920012



RESIDUAL RANGE ORGANICS CHROMATOGRAM

Sample Name: B21121613-001BMS-RRO ;1226HP5 , SGT
 Raw File: G:\org\HP5\DAT\HP5122621_b\1226HP5.0028.RAW
 Date & Time Acquired: 12/27/2021 6:34:35 AM
 Method File: G:\Org\HP5\Methods\D3_ORO-AL-L%.MET
 Calibration File: G:\Org\HP5\Cals\SW8015C_ORO211017AL.CAL
 Sample Weight: 1045 Dilution: 1 S.A.: 1

Mean RF for for ~~Residual~~ TEH (Oil Range) Organics Calculations: 28542.41
 Rt range for ~~Residual~~ TEH (Oil Range) Organics: 12.56 to 30.05


SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*#Triacontane	16.283	.478	.185	38.61	-

RRO TEH (Oil Range) Area:1.425497E+08 RRO TEH (Oil Range) AMOUNT: 4.779245

AMN 01/11/2022

G:\org\HP5\DAT\HP5122621_b\1226HP5.0028.RAW

B21121613-001BMS-RRO ;1226HP5 , SGT




RESIDUAL RANGE ORGANICS CHROMATOGRAM

Sample Name: B21121613-001BMS-RRO ;1226HP5 , SGT
 Raw File: G:\org\HP5\DAT\HP5122621_b\1226HP5.0028.RAW
 Date & Time Acquired: 12/27/2021 6:34:35 AM
 Method File: G:\Org\HP5\Methods\DS_ORO-AL-L%.MET
 Calibration File: G:\Org\HP5\Cals\SW8015C_ORO211017AL.CAL
 Sample Weight: 1045 Dilution: 1 S.A.: 1

Mean RF for for Residual Range Organics Calculations: 28542.41
 Rt range for Residual Range Organics: 12.56 to 30.05

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*#Triacontane	16.283	.478	.092	19.31

RRO Area:5297075 RRO AMOUNT: 0.1775943



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: DCM-Baseline Check-V29
 Raw File: G:\org\HP5\DAT\HP5122621_b\1226HP5.0029.RAW
 Date & Time Acquired: 12/27/2021 7:17:33 AM
 Method File: G:\Org\HP5\Methods\DR_8015-IBb-LEXP.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO211102IB.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for TEH: 31353.19
 Rt range for Diesel Range Organics: 6.54 to 15.86


SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*o-Terphenyl	12.217	200.	.303	.15
*1-Chlorooctadecane	29.925	200.	.	-

DRO Area: 587752.8 DRO Amount: 18.74619
 TEH Area: 1014402 TEH Amount: 32.35403

Batch ID: 162352

LCS-162352 ;1226HP5 , RR-SGT

G:\org\HP5\DAT\HP5122621_b\1226HP5.0030.RAW



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT


Sample Name: LCS-162352 ;1226HP5 , RR-SGT
 Raw File: G:\org\HP5\DAT\HP5122621_b\1226HP5.0030.RAW
 Date & Time Acquired: 12/27/2021 8:00:26 AM
 Method File: G:\Org\HP5\Methods\D3_8015-122630-IL-L%.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO211102IL-24.CAL
 Sample Weight: 1000 Dilution: 1 S.A.: 1

Mean RF for TEH: 31353.19

Rt range for Diesel Range Organics: 6.51 to 14.37

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*o-Terphenyl	12.217	.2	.365	182.68	-
*1-Chlorooctadecane	13.063	.2	.171	85.29	-

DRO Area: 4.67076E+08 DRO Amount: 14.89724
 TEH Area: 4.96762E+08 TEH Amount: 15.84407



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT


Sample Name: LCS-162352 ;1226HP5 , RR-SGT
 Raw File: G:\org\HP5\DAT\HP5122621_b\1226HP5.0030.RAW
 Date & Time Acquired: 12/27/2021 8:00:26 AM
 Method File: G:\Org\HP5\Methods\DS_8015-24-IL-L#.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO211102IL-24.CAL
 Sample Weight: 1000 Dilution: 1 S.A.: 1

Mean RF for TEH: 31353.19

Rt range for Diesel Range Organics: 6.51 to 14.37

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*o-Terphenyl	12.217	.2	.202	100.96
*1-Chlorooctadecane	13.063	.2	.073	36.39

DRO Area: 2.254774E+08 DRO Amount: 7.191529
 TEH Area: 2.392939E+08 TEH Amount: 7.632204



RESIDUAL RANGE ORGANICS CHROMATOGRAM


Sample Name: B21121613-001BMSD-RRO ;1226HP5 , SGT
 Raw File: G:\org\HP5\DAT\HP5122621_b\1226HP5.0031.RAW
 Date & Time Acquired: 12/27/2021 8:43:15 AM
 Method File: G:\Org\HP5\Methods\D3_ORO-AL-L%.MET
 Calibration File: G:\Org\HP5\Cals\SW8015C_ORO211017AL.CAL
 Sample Weight: 1055 Dilution: 1 S.A.: 1

Mean RF for for ~~Residual~~ TEH (Oil Range) Organics Calculations: 28542.41
 Rt range for ~~Residual~~ TEH (Oil Range) Organics: 12.56 to 30.05

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*#Triacontane	16.286	.474	.189	39.84	-

~~RRO~~ TEH (Oil Range) Area:1.468009E+08 ~~RRO~~ TEH (Oil Range) AMOUNT: 4.875122

AMN 01/11/2022




RESIDUAL RANGE ORGANICS CHROMATOGRAM

Sample Name: B21121613-001BMSD-RRO ;1226HP5 , SGT
 Raw File: G:\org\HP5\DAT\HP5122621_b\1226HP5.0031.RAW
 Date & Time Acquired: 12/27/2021 8:43:15 AM
 Method File: G:\Org\HP5\Methods\DS_ORO-AL-L%.MET
 Calibration File: G:\Org\HP5\Cals\SW8015C_ORO211017AL.CAL
 Sample Weight: 1055 Dilution: 1 S.A.: 1

Mean RF for for Residual Range Organics Calculations: 28542.41
 Rt range for Residual Range Organics: 12.56 to 30.05

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*#Triacontane	16.286	.474	.094	19.8

RRO Area:5556087 RRO AMOUNT: 0.1845125




RESIDUAL RANGE ORGANICS CHROMATOGRAM


Sample Name: LCS-162352-RRO ;1226HP5 ,
 Raw File: G:\org\HP5\DAT\HP5122621_b\1226HP5.0032.RAW
 Date & Time Acquired: 12/27/2021 9:26:27 AM
 Method File: G:\Org\HP5\Methods\D3_ORO-AK-L0.MET
 Calibration File: G:\Org\HP5\Cals\SW8015C_ORO211017AK.CAL
 Sample Weight: 1000 Dilution: 1 S.A.: 1

Mean RF for for Residual Range Organics Calculations: 28542.41
 Rt range for Residual Range Organics: 12.58 to 30.05

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*#Triacontane	16.284	.5	.231	46.14	-

RRO Area:1.63749E+08 RRO AMOUNT: 5.737041





RESIDUAL RANGE ORGANICS CHROMATOGRAM

Sample Name: CCV_1226HP534r, RRO ;1226HP5 , DRO211220C
 Raw File: G:\org\HP5\DAT\HP5122621_b\1226HP5.0034.RAW
 Date & Time Acquired: 12/27/2021 10:52:16 AM
 Method File: G:\Org\HP5\Methods\DC_ORO-AL-L%.MET
 Calibration File: G:\Org\HP5\Cals\SW8015C_ORO211017AL.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for for ~~Residual~~ TEH (Oil Range) Organics Calculations: 28542.41
 Rt range for ~~Residual~~ TEH (Oil Range) Organics: 12.56 to 30.05

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*#Triacontane	16.284	500.	351.055	70.21	-

RRO TEH (Oil Range) Area:1.362173E+08 RRO TEH (Oil Range) AMOUNT: 4772.454

CONTINUING CALIBRATION REPORT: G:\org\HP5\DAT\HP5122621_b\1226HP5.0034.RAW


COMPOUND	ACTUAL (NG)	MEASURED (NG)	%RECOVERY	LIMITS
*30-40 Motor Oil	5000.	.029	.	75-125

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	LIMITS
*#Triacontane	16.284	200.	351.055	175.53	75-125

AMN 01/11/2022

G:\org\HP5\DAT\HP5122621_b\1226HP5.0034.RAW

CCV_1226HP534r, RRO ;1226HP5 , DRO211220C



RESIDUAL RANGE ORGANICS CHROMATOGRAM

Sample Name: CCV_1226HP534r, RRO ;1226HP5 , DRO211220C
 Raw File: G:\org\HP5\DAT\HP5122621_b\1226HP5.0034.RAW
 Date & Time Acquired: 12/27/2021 10:52:16 AM
 Method File: G:\Org\HP5\Methods\DS_ORO-AL-L%.MET
 Calibration File: G:\Org\HP5\Cals\SW8015C_ORO211017AL.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for for Residual Range Organics Calculations: 28542.41
 Rt range for Residual Range Organics: 12.56 to 30.05


SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*#Triacontane	16.284	500.	215.283	43.06	-

RRO Area:6614671 RRO AMOUNT: 231.7488

CONTINUING CALIBRATION REPORT: G:\org\HP5\DAT\HP5122621_b\1226HP5.0034.RAW

COMPOUND	ACTUAL (NG)	MEASURED (NG)	%RECOVERY	LIMITS
*30-40 Motor Oil	5000.	.029	.	75-125

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	LIMITS
*#Triacontane	16.284	200.	215.283	107.64	75-125



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: CCV_1226HP535r, DRO ;1226HP5 , DRO211220A
 Raw File: G:\org\HP5\DAT\HP5122621_b\1226HP5.0035.RAW
 Date & Time Acquired: 12/27/2021 11:35:09 AM
 Method File: G:\Org\HP5\Methods\DC_8015-24-IL-L%.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO211102IL-24.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for TEH: 31353.19
 Rt range for Diesel Range Organics: 6.51 to 14.37


SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*o-Terphenyl	12.218	200.	326.58	163.29
*1-Chlorooctadecane	13.06	200.	159.338	79.67

DRO Area: 4.668803E+08 DRO Amount: 14891
 TEH Area: 4.842929E+08 TEH Amount: 15446.37

CONTINUING CALIBRATION REPORT: G:\org\HP5\DAT\HP5122621_b\1226HP5.0035.RAW

COMPOUND	ACTUAL (NG)	MEASURED (NG)	%RECOVERY	LIMITS
TOTAL DRO	15000.	15446.37	102.98	85-115

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	LIMITS
*o-Terphenyl	12.218	200.	326.58	163.29	85-115
*1-Chlorooctadecane	13.06	200.	159.338	79.67	85-115



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: CCV_1226HP535r, DRO ;1226HP5 , DRO211220A
 Raw File: G:\org\HP5\DAT\HP5122621_b\1226HP5.0035.RAW
 Date & Time Acquired: 12/27/2021 11:35:09 AM
 Method File: G:\Org\HP5\Methods\DS_8015-24-IL-L#.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO211102IL-24.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for TEH: 31353.19

Rt range for Diesel Range Organics: 6.51 to 14.37


SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*o-Terphenyl	12.218	200.	199.502	99.75
*1-Chlorooctadecane	13.06	200.	38.038	19.02

DRO Area: 2.589707E+08 DRO Amount: 8259.787
 TEH Area: 2.696968E+08 TEH Amount: 8601.894

CONTINUING CALIBRATION REPORT: G:\org\HP5\DAT\HP5122621_b\1226HP5.0035.RAW

COMPOUND	ACTUAL (NG)	MEASURED (NG)	%RECOVERY	LIMITS
TOTAL DRO	15000.	8601.89	57.35	85-115

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	LIMITS
*o-Terphenyl	12.218	200.	199.502	99.75	85-115
*1-Chlorooctadecane	13.06	200.	38.038	19.02	85-115




DIESEL RANGE ORGANICS CHROMATOGRAM REPORT


Sample Name: DCM-Baseline Check-V36
 Raw File: G:\org\HP5\DAT\HP5122621_b\1226HP5.0036.RAW
 Date & Time Acquired: 12/27/2021 12:18:05 PM
 Method File: G:\Org\HP5\Methods\DR_8015-IBb-LEXP.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO211102IB.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for TEH: 31353.19
 Rt range for Diesel Range Organics: 6.54 to 15.86

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*o-Terphenyl	12.216	200.	.41	.2	-
*1-Chlorooctadecane	13.073	200.	.033	.02	-

DRO Area:505658.4 DRO Amount: 16.12781
 TEH Area:627625.6 TEH Amount: 20.01792





RESIDUAL RANGE ORGANICS CHROMATOGRAM

Sample Name: CCV_1226HP550r, RRO ;1226HP5 , DRO211220C
 Raw File: G:\org\HP5\DAT\HP5122621_b\1226HP5.0050.RAW
 Date & Time Acquired: 12/27/2021 10:24:33 PM
 Method File: G:\Org\HP5\Methods\DC_ORO-AL-L%.MET
 Calibration File: G:\Org\HP5\Cals\SW8015C_ORO211017AL.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for for ~~Residual~~ TEH (Oil Range) Organics Calculations: 28542.41
 Rt range for ~~Residual~~ TEH (Oil Range) Organics: 12.56 to 30.05

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*#Triacontane	16.294	500.	353.668	70.73	-

~~RRO~~ TEH (Oil Range) Area:1.464909E+08 ~~RRO~~ TEH (Oil Range) AMOUNT: 5132.395

CONTINUING CALIBRATION REPORT: G:\org\HP5\DAT\HP5122621_b\1226HP5.0050.RAW


COMPOUND	ACTUAL (NG)	MEASURED (NG)	%RECOVERY	LIMITS
*30-40 Motor Oil	5000.	.065	.	75-125

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	LIMITS
*#Triacontane	16.294	200.	353.668	176.83	75-125

AMN 01/11/2022

G:\org\HP5\DAT\HP5122621_b\1226HP5.0050.RAW

CCV_1226HP550r, RRO ;1226HP5 , DRO211220C



RESIDUAL RANGE ORGANICS CHROMATOGRAM

Sample Name: CCV_1226HP550r, RRO ;1226HP5 , DRO211220C
Raw File: G:\org\HP5\DAT\HP5122621_b\1226HP5.0050.RAW
Date & Time Acquired: 12/27/2021 10:24:33 PM
Method File: G:\Org\HP5\Methods\DS_ORO-AL-L%.MET
Calibration File: G:\Org\HP5\Cals\SW8015C_ORO211017AL.CAL
Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for for Residual Range Organics Calculations: 28542.41
Rt range for Residual Range Organics: 12.56 to 30.05

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*#Triacontane	16.294	500.	217.331	43.47	-

RRO Area:6985090 RRO AMOUNT: 244.7267


CONTINUING CALIBRATION REPORT: G:\org\HP5\DAT\HP5122621_b\1226HP5.0050.RAW

COMPOUND	ACTUAL (NG)	MEASURED (NG)	%RECOVERY	LIMITS
*30-40 Motor Oil	5000.	.065	.	75-125

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	LIMITS
*#Triacontane	16.294	200.	217.331	108.67	75-125

G:\org\HP5\DAT\HP5122621_b\1226HP5.0051.RAW

CCV_1226HP551r, DRO ;1226HP5 , DRO211220A



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: CCV_1226HP551r, DRO ;1226HP5 , DRO211220A
 Raw File: G:\org\HP5\DAT\HP5122621_b\1226HP5.0051.RAW
 Date & Time Acquired: 12/27/2021 11:07:32 PM
 Method File: G:\Org\HP5\Methods\DC_8015-24-IL-L%.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO211102IL-24.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for TEH: 31353.19

Rt range for Diesel Range Organics: 6.51 to 14.37

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*o-Terphenyl	12.227	200.	347.264	173.63
*1-Chlorooctadecane	13.067	200.	162.652	81.33

DRO Area: 4.842769E+08 DRO Amount: 15445.86
 TEH Area: 5.001953E+08 TEH Amount: 15953.57


CONTINUING CALIBRATION REPORT: G:\org\HP5\DAT\HP5122621_b\1226HP5.0051.RAW

COMPOUND	ACTUAL (NG)	MEASURED (NG)	%RECOVERY	LIMITS
TOTAL DRO	15000.	15953.57	106.36	85-115

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	LIMITS
*o-Terphenyl	12.227	200.	347.264	173.63	85-115
*1-Chlorooctadecane	13.067	200.	162.652	81.33	85-115

G:\org\HP5\DAT\HP5122621_b\1226HP5.0051.RAW

CCV_1226HP551r, DRO ;1226HP5 , DRO211220A



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: CCV_1226HP551r, DRO ;1226HP5 , DRO211220A
 Raw File: G:\org\HP5\DAT\HP5122621_b\1226HP5.0051.RAW
 Date & Time Acquired: 12/27/2021 11:07:32 PM
 Method File: G:\Org\HP5\Methods\DS_8015-24-IL-L#.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO211102IL-24.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for TEH: 31353.19

Rt range for Diesel Range Organics: 6.51 to 14.37


SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*o-Terphenyl	12.227	200.	212.485	106.24
*1-Chlorooctadecane	13.067	200.	40.035	20.02

DRO Area: 2.73378E+08 DRO Amount: 8719.302
 TEH Area: 2.844651E+08 TEH Amount: 9072.922

CONTINUING CALIBRATION REPORT: G:\org\HP5\DAT\HP5122621_b\1226HP5.0051.RAW

COMPOUND	ACTUAL (NG)	MEASURED (NG)	%RECOVERY	LIMITS
TOTAL DRO	15000.	9072.92	60.49	85-115

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	LIMITS
*o-Terphenyl	12.227	200.	212.485	106.24	85-115
*1-Chlorooctadecane	13.067	200.	40.035	20.02	85-115




DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: DCM-Baseline Check-V52
 Raw File: G:\org\HP5\DAT\HP5122621_b\1226HP5.0052.RAW
 Date & Time Acquired: 12/27/2021 11:50:28 PM
 Method File: G:\Org\HP5\Methods\DR_8015-IBb-LEXP.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO211102IB.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for TEH: 31353.19
 Rt range for Diesel Range Organics: 6.54 to 15.86

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*o-Terphenyl	12.217	200.	.251	.13
*1-Chlorooctadecane	13.001	200.	.017	.01

DRO Area: 429180.7 DRO Amount: 13.68858
 TEH Area: 747937.4 TEH Amount: 23.85522




DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: DCM-Baseline Check-V56
 Raw File: G:\org\HP5\DAT\HP5122621_b\1226HP5.0056.RAW
 Date & Time Acquired: 12/28/2021 2:42:24 AM
 Method File: G:\Org\HP5\Methods\DR_8015-IBb-LEXP.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO211102IB.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for TEH: 31353.19
 Rt range for Diesel Range Organics: 6.54 to 15.86

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*o-Terphenyl	12.218	200.	.265	.13
*1-Chlorooctadecane	29.99	200.	.	.

DRO Area: 415776.2 DRO Amount: 13.26105
 TEH Area: 960770.3 TEH Amount: 30.64346



RESIDUAL RANGE ORGANICS CHROMATOGRAM

Sample Name: LCS-162352-RRO ;1226HP5 , RR-SGT
 Raw File: G:\Org\HP5\DAT\HP5122621_b\1226HP5.0057.RAW
 Date & Time Acquired: 12/28/2021 3:25:24 AM
 Method File: G:\Org\HP5\Methods\D3_ORO- AL-L%.MET
 Calibration File: G:\Org\HP5\Cals\SW8015C_ORO211017AL.CAL
 Sample Weight: 1000 Dilution: 1 S.A.: 1

Mean RF for for ~~Residual~~ TEH (Oil Range) Organics Calculations: 28542.41
 Rt range for ~~Residual~~ TEH (Oil Range) Organics: 12.56 to 30.05


SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*#Triacontane	16.288	.5	.217	43.41

RRO TEH (Oil Range) Area:1.61902E+08 RRO TEH (Oil Range) AMOUNT: 5.672332

AMN 01/11/2022

G:\Org\HP5\DAT\HP5122621_b\1226HP5.0057.RAW

LCS-162352-RRO ;1226HP5 , RR-SGT




RESIDUAL RANGE ORGANICS CHROMATOGRAM


Sample Name: LCS-162352-RRO ;1226HP5 , RR-SGT
 Raw File: G:\Org\HP5\DAT\HP5122621_b\1226HP5.0057.RAW
 Date & Time Acquired: 12/28/2021 3:25:24 AM
 Method File: G:\Org\HP5\Methods\DS_ORO-AL-L%.MET
 Calibration File: G:\Org\HP5\Cals\SW8015C_ORO211017AL.CAL
 Sample Weight: 1000 Dilution: 1 S.A.: 1

Mean RF for for Residual Range Organics Calculations: 28542.41
 Rt range for Residual Range Organics: 12.56 to 30.05

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*#Triacontane	16.288	.5	.113	22.66

RRO Area:5707735 RRO AMOUNT: 0.1999738





RESIDUAL RANGE ORGANICS CHROMATOGRAM

Sample Name: CCV_1226HP561r, RRO ;1226HP5 , DRO211220C
 Raw File: G:\org\HP5\DAT\HP5122621_b\1226HP5.0061.RAW
 Date & Time Acquired: 12/28/2021 6:17:22 AM
 Method File: G:\Org\HP5\Methods\DC_ORO-AL-L%.MET
 Calibration File: G:\Org\HP5\Cals\SW8015C_ORO211017AL.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for for ~~Residual~~ TEH (Oil Range) Organics Calculations: 28542.41
 Rt range for ~~Residual~~ TEH (Oil Range) Organics: 12.56 to 30.05

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*#Triacontane	16.296	500.	373.382	74.68	-

~~RRO~~ TEH (Oil Range) Area:1.585933E+08 ~~RRO~~ TEH (Oil Range) AMOUNT: 5556.41

CONTINUING CALIBRATION REPORT: G:\org\HP5\DAT\HP5122621_b\1226HP5.0061.RAW


COMPOUND	ACTUAL (NG)	MEASURED (NG)	%RECOVERY	LIMITS
*30-40 Motor Oil	5000.	.054	.	75-125

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	LIMITS
*#Triacontane	16.296	200.	373.382	186.69	75-125

AMN 01/11/2022

G:\org\HP5\DAT\HP5122621_b\1226HP5.0061.RAW

CCV_1226HP561r, RRO ;1226HP5 , DRO211220C



RESIDUAL RANGE ORGANICS CHROMATOGRAM

Sample Name: CCV_1226HP561r, RRO ;1226HP5 , DRO211220C
Raw File: G:\org\HP5\DAT\HP5122621_b\1226HP5.0061.RAW
Date & Time Acquired: 12/28/2021 6:17:22 AM
Method File: G:\Org\HP5\Methods\DS_ORO-AL-L%.MET
Calibration File: G:\Org\HP5\Cals\SW8015C_ORO211017AL.CAL
Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for for Residual Range Organics Calculations: 28542.41
Rt range for Residual Range Organics: 12.56 to 30.05

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*#Triacontane	16.296	500.	223.119	44.62	-

RRO Area:6751890 RRO AMOUNT: 236.5564

CONTINUING CALIBRATION REPORT: G:\org\HP5\DAT\HP5122621_b\1226HP5.0061.RAW

COMPOUND	ACTUAL (NG)	MEASURED (NG)	%RECOVERY	LIMITS
*30-40 Motor Oil	5000.	.054	.	75-125

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	LIMITS
*#Triacontane	16.296	200.	223.119	111.56	75-125

Write Sequence	Data File	Sample Name	Method	Weight	Dil Factor	Amnt Inj.	IS	Cal D	Manual Integrations
	G:\org\HP5\DAT\HP5122121_b1221HP5.01r	DCM-Baseline Check-V01	G:\org\HP5\Methods\DR_8015-IBb-LEXP.met	1	1	1	1	1	0 No integrations
	G:\org\HP5\DAT\HP5122121_b1221HP5.02r	DCM-Baseline Check-V02	G:\org\HP5\Methods\DR_8015-IBb-LEXP.met	1	1	1	1	1	0 No integrations
	G:\org\HP5\DAT\HP5122121_b1221HP5.03r	MARKER_1220HP503r_C40_1220HP5_DRO211207A	G:\org\HP5\Methods\GSCS21221.met	1	1	1	1	1	0 No integrations
	G:\org\HP5\DAT\HP5122121_b1221HP5.04r	CCV_1221HP504r_RRO_1221HP5_DRO211220C	G:\org\HP5\Methods\DC_ORO-AK-L%.MET G:\org\HP5\Methods\DS_ORO-AK-L%.MET	1	1	1	1	1	0 The integration of Oil Range hydrocarbon is the hydrocarbon response with reference to the baseline. Surrogates are integrated using a valley to valley integration using Set Baseline Now placed slight before at 16.18 minutes slightly after the surrogate peak at 16.4 minutes and X-axis scaling showing surrogate peak from 14-18 minutes.
	G:\org\HP5\DAT\HP5122121_b1221HP5.05r	CCV_1221HP505r_DRO_1221HP5_DRO211220A	G:\org\HP5\Methods\DC_8015-24-ik-L%.met G:\org\HP5\Methods\DS_8015-24-ik-L%.met	1	1	1	1	1	0 The integration of Diesel Range Organics (C10-C24) and Total Extractable Hydrocarbons is the hydrocarbon response with reference to the baseline. Assigned Set Baseline All Valley on at 16.83 minutes. Surrogates are integrated using a valley to valley integration using Set Baseline Now placed slight before at 12.14 minutes and slightly after the surrogate peak at 12.31 minutes and X-axis scaling showing surrogate peak from 11.5-14 minutes.
	G:\org\HP5\DAT\HP5122121_b1221HP5.06r	DCM-Baseline Check-V06	G:\org\HP5\Methods\DR_8015-IBb-LEXP.met	1	1	1	1	1	0 No integrations
	G:\org\HP5\DAT\HP5122121_b1221HP5.07r	LCS-162352_1221HP5_	G:\org\HP5\Methods\D3_8015-24-ik-L%.met G:\org\HP5\Methods\DS_8015-24-ik-L%.met	1000	1	1	1	1	0 The integration of Diesel Range Organics (C10-C24) and Total Extractable Hydrocarbons is the hydrocarbon response with reference to the baseline. Assigned Set Baseline All Valley on at 16.83 minutes. Surrogates are integrated using a valley to valley integration using Set Baseline Now placed slight before at 12.14 minutes and slightly after the surrogate peak at 12.31 minutes and X-axis scaling showing surrogate peak from 11.5-14 minutes.
	G:\org\HP5\DAT\HP5122121_b1221HP5.08r	MB-162352_1221HP5_	G:\org\HP5\Methods\DR_8015-C24T-ik-L%.met G:\org\HP5\Methods\DR_OROS-AK-L%.MET G:\org\HP5\Methods\DS_8015-C24T-ik-L%.met	1000	1	1	1	1	0 The integration of Diesel Range Organics (C10-C24), C24-C40, and Total Extractable Hydrocarbons (TEH) is the hydrocarbon response with reference to the baseline. Assigned Set Baseline All Valley on at 17.12 minutes. Surrogates are integrated using a valley to valley integration using Set Baseline Now placed slight before at 12.11 minutes and 16.21 minutes and slightly after the surrogate peaks at 12.46 and 16.41 minutes. X-axis scaling showing surrogate peak from 11-18 minutes.
	G:\org\HP5\DAT\HP5122121_b1221HP5.09r	B21121613-001B_1221HP5_ \$HC-8015-DRO-W,	G:\org\HP5\Methods\DR_8015-C24T-ik-L%.met G:\org\HP5\Methods\DR_OROS-AK-L%.MET G:\org\HP5\Methods\DS_8015-C24T-ik-L%.met	990	1	1	1	1	0 The integration of Diesel Range Organics (C10-C24), C24-C40, and Total Extractable Hydrocarbons (TEH) is the hydrocarbon response with reference to the baseline. Assigned Set Baseline All Valley on at 17.12 minutes. Surrogates are integrated using a valley to valley integration using Set Baseline Now placed slight before at 12.11 minutes and 16.21 minutes and slightly after the surrogate peaks at 12.46 and 16.41 minutes. X-axis scaling showing surrogate peak from 11-18 minutes.
	G:\org\HP5\DAT\HP5122121_b1221HP5.10r	B21121613-001BMS_1221HP5_	G:\org\HP5\Methods\D3_8015-24-ik-L%.met G:\org\HP5\Methods\DS_8015-24-ik-L%.met	990	1	1	1	1	0 The integration of Diesel Range Organics (C10-C24) and Total Extractable Hydrocarbons is the hydrocarbon response with reference to the baseline. Assigned Set Baseline All Valley on at 16.83 minutes. Surrogates are integrated using a valley to valley integration using Set Baseline Now placed slight before at 12.14 minutes and slightly after the surrogate peak at 12.31 minutes and X-axis scaling showing surrogate peak from 11.5-14 minutes.
	G:\org\HP5\DAT\HP5122121_b1221HP5.11r	B21121613-001BMSD_1221HP5_	G:\org\HP5\Methods\D3_8015-24-ik-L%.met G:\org\HP5\Methods\DS_8015-24-ik-L%.met	1040	1	1	1	1	0 The integration of Diesel Range Organics (C10-C24) and Total Extractable Hydrocarbons is the hydrocarbon response with reference to the baseline. Assigned Set Baseline All Valley on at 16.83 minutes. Surrogates are integrated using a valley to valley integration using Set Baseline Now placed slight before at 12.14 minutes and slightly after the surrogate peak at 12.31 minutes and X-axis scaling showing surrogate peak from 11.5-14 minutes.
	G:\org\HP5\DAT\HP5122121_b1221HP5.12r	DCM-Baseline Check-V12	G:\org\HP5\Methods\DR_8015-IBb-LEXP.met	1	1	1	1	1	0 No integrations
	G:\org\HP5\DAT\HP5122121_b1221HP5.13r	B21121613-002B_1221HP5_ \$HC-8015-DRO-W,	G:\org\HP5\Methods\DR_8015-C24T-ik-L%.met G:\org\HP5\Methods\DR_OROS-AK-L%.MET G:\org\HP5\Methods\DS_8015-C24T-ik-L%.met	1020	1	1	1	1	0 The integration of Diesel Range Organics (C10-C24), C24-C40, and Total Extractable Hydrocarbons (TEH) is the hydrocarbon response with reference to the baseline. Assigned Set Baseline All Valley on at 17.12 minutes. Surrogates are integrated using a valley to valley integration using Set Baseline Now placed slight before at 12.11 minutes and 16.21 minutes and slightly after the surrogate peaks at 12.46 and 16.41 minutes. X-axis scaling showing surrogate peak from 11-18 minutes.
	G:\org\HP5\DAT\HP5122121_b1221HP5.14r	DCM-Baseline Check-V14	G:\org\HP5\Methods\DR_8015-IBb-LEXP.met	1	1	1	1	1	0 No integrations
	G:\org\HP5\DAT\HP5122121_b1221HP5.15r	B21121609-001C_1221HP5_ \$HC-8015-DRO-W,	G:\org\HP5\Methods\D3_8015-C24T-ik-L%.met G:\org\HP5\Methods\DR_OROS-AK-L%.MET G:\org\HP5\Methods\DS_8015-22115-ik-L%.met	1060	1	1	1	1	0 The integration of Diesel Range Organics (C10-C24), C24-C40, and Total Extractable Hydrocarbons (TEH) is the hydrocarbon response with reference to the baseline. Surrogates are integrated using a valley to valley integration using Set Baseline Now placed slight before at 12.11 minutes and 16.21 minutes and slightly after the surrogate peaks at 12.29 and 16.41 minutes. X-axis scaling showing surrogate peak from 11-18 minutes.
	G:\org\HP5\DAT\HP5122121_b1221HP5.16r	DCM-Baseline Check-V16	G:\org\HP5\Methods\DR_8015-IBb-LEXP.met	1	1	1	1	1	0 No integrations
	G:\org\HP5\DAT\HP5122121_b1221HP5.17r	B21121611-001B_1221HP5_ \$HC-8015-DRO-W,	G:\org\HP5\Methods\DR_8015-C24T-ik-L%.met G:\org\HP5\Methods\DR_OROS-AK-L%.MET G:\org\HP5\Methods\DS_8015-122115-ik-L%.met	1030	1	1	1	1	0 The integration of Diesel Range Organics (C10-C24), C24-C40, and Total Extractable Hydrocarbons (TEH) is the hydrocarbon response with reference to the baseline. Surrogates are integrated using a valley to valley integration using Set Baseline Now placed slight before at 12.11 minutes and 16.21 minutes and slightly after the surrogate peaks at 12.29 and 16.47 minutes. X-axis scaling showing surrogate peak from 11-18 minutes.

G:\org\HP5\DAT\HP5122121_b\1221HP5.18r	B21121616-001C ;1221HP5 , \$HC-8015-DRO-W,	G:\Org\HP5\Methods\D3_8015-C24T-IK-L%.met G:\Org\HP5\Methods\D3_OROS-AK-L%.MET G:\Org\HP5\Methods\DS_8015-122115-IK-L%.met	990	1	1	1	0	The integration of Diesel Range Organics (C10-C24), C24-C40, and Total Extractable Hydrocarbons (TEH) is the hydrocarbon response with reference to the baseline. Surrogates are integrated using a valley to valley integration using Set Baseline Now placed slight before at 12.11 minutes and 16.21 minutes and slightly after the surrogate peaks at 12.29 and 16.41 minutes. X-axis scaling showing surrogate peak from 11-18 minutes.
G:\org\HP5\DAT\HP5122121_b\1221HP5.19r	MARKER_1220HP519r_C40 ;1220HP5 , DRO211207A	G:\org\HP5\Methods\CS211221.met	1	1	1	1	0	No integrations
G:\org\HP5\DAT\HP5122121_b\1221HP5.20r	CCV_1221HP520r ,RRO ;1221HP5 , DRO211220C	G:\Org\HP5\Methods\DC_ORO-AK-L%.MET G:\Org\HP5\Methods\DS_ORO-AK-L%.MET	1	1	1	1	0	The integration of Oil Range hydrocarbon is the hydrocarbon response with reference to the baseline. Surrogates are integrated using a valley to valley integration using Set Baseline Now placed slight before at 16.18 minutes slightly after the surrogate peak at 16.4 minutes and X-axis scaling showing surrogate peak from 14-18 minutes.
G:\org\HP5\DAT\HP5122121_b\1221HP5.21r	CCV_1221HP521r ,DRO ;1221HP5 , DRO211220A	G:\Org\HP5\Methods\DC_8015-24-IK-L%.met G:\Org\HP5\Methods\DS_8015-24-IK-L%.met	1	1	1	1	0	The integration of Diesel Range Organics (C10-C24) and Total Extractable Hydrocarbons is the hydrocarbon response with reference to the baseline. Assigned Set Baseline All Valley on at 16.83 minutes. Surrogates are integrated using a valley to valley integration using Set Baseline Now placed slight before at 12.14 minutes and slightly after the surrogate peak at 12.31 minutes and X-axis scaling showing surrogate peak from 11.5-14 minutes.
G:\org\HP5\DAT\HP5122121_b\1221HP5.22r	DCM-Baseline Check-V22	G:\Org\HP5\Methods\DR_8015-IBB-LEXP.met	1	1	1	1	0	No integrations
G:\org\HP5\DAT\HP5122121_b\1221HP5.23r	B21121623-001C ;1221HP5 , \$HC-8015-DRO-W,	G:\Org\HP5\Methods\DR_8015-122123-IK-L%.met G:\Org\HP5\Methods\D3_OROS-122123-AK-L%.MET G:\Org\HP5\Methods\DS_8015-C24T-IK-L%.met	1020	1	1	1	0	The integration of Diesel Range Organics (C10-C24), C24-C40, and Total Extractable Hydrocarbons (TEH) is the hydrocarbon response with reference to the baseline with Set Baseline Now at 25.75 minutes. Surrogates are integrated using a valley to valley integration using Set Baseline Now placed slight before at 12.11 minutes and 16.21 minutes and slightly after the surrogate peaks at 12.46 and 16.41 minutes. X-axis scaling showing surrogate peak from 11-18 minutes.
G:\org\HP5\DAT\HP5122121_b\1221HP5.24r	DCM-Baseline Check-V24	G:\Org\HP5\Methods\DR_8015-IBB-LEXP.met	1	1	1	1	0	No integrations
G:\org\HP5\DAT\HP5122121_b\1221HP5.25r	B21121605-001A ;1221HP5 , \$HC-8015-DRO-W,	G:\Org\HP5\Methods\DR_8015-122123-IK-L%.met G:\Org\HP5\Methods\D3_OROS-122123-AK-L%.MET G:\Org\HP5\Methods\DS_8015-C24T-IK-L%.met	1050	1	1	1	0	The integration of Diesel Range Organics (C10-C24), C24-C40, and Total Extractable Hydrocarbons (TEH) is the hydrocarbon response with reference to the baseline with Set Baseline Now at 25.75 minutes. Surrogates are integrated using a valley to valley integration using Set Baseline Now placed slight before at 12.11 minutes and 16.21 minutes and slightly after the surrogate peaks at 12.46 and 16.41 minutes. X-axis scaling showing surrogate peak from 11-18 minutes.
G:\org\HP5\DAT\HP5122121_b\1221HP5.26r	DCM-Baseline Check-V26	G:\Org\HP5\Methods\DR_8015-IBB-LEXP.met	1	1	1	1	0	No integrations
G:\org\HP5\DAT\HP5122121_b\1221HP5.27r	B21121605-002A ;1221HP5 , \$HC-8015-DRO-W,	G:\Org\HP5\Methods\DR_8015-C24T-IK-L%.met G:\Org\HP5\Methods\D3_OROS-AK-L%.MET G:\Org\HP5\Methods\DS_8015-122127-IK-L%.met	1050	1	1	1	0	The integration of Diesel Range Organics (C10-C24), C24-C40, and Total Extractable Hydrocarbons (TEH) is the hydrocarbon response with reference to the baseline. Surrogates are integrated using a valley to valley integration using Set Baseline Now placed slight before at 12.11 minutes and 16.21 minutes and slightly after the surrogate peaks at 12.34 and 16.54 minutes. X-axis scaling showing surrogate peak from 11-18 minutes.
G:\org\HP5\DAT\HP5122121_b\1221HP5.28r	B21121605-003A ;1221HP5 , \$HC-8015-DRO-W,	G:\Org\HP5\Methods\D3_8015-C24T-IK-L%.met G:\Org\HP5\Methods\D3_OROS-AK-L%.MET G:\Org\HP5\Methods\DS_8015-122127-IK-L%.met	1010	1	1	1	0	The integration of Diesel Range Organics (C10-C24), C24-C40, and Total Extractable Hydrocarbons (TEH) is the hydrocarbon response with reference to the baseline. Surrogates are integrated using a valley to valley integration using Set Baseline Now placed slight before at 12.11 minutes and 16.21 minutes and slightly after the surrogate peaks at 12.34 and 16.54 minutes. X-axis scaling showing surrogate peak from 11-18 minutes.
G:\org\HP5\DAT\HP5122121_b\1221HP5.29r	DCM-Baseline Check-V29	G:\Org\HP5\Methods\DR_8015-IBB-LEXP.met	1	1	1	1	0	No integrations
G:\org\HP5\DAT\HP5122121_b\1221HP5.30r	B21121622-001B ;1221HP5 , \$HC-8015-DRO-W, due to baseline	G:\Org\HP5\Methods\DR_8015-IBB-LEXP.met	960	1	1	1	0	No integrations
G:\org\HP5\DAT\HP5122121_b\1221HP5.31r	B21121622-002B ;1221HP5 , \$HC-8015-DRO-W,	G:\Org\HP5\Methods\D3_8015-122131-IK-L%.met G:\Org\HP5\Methods\D3_OROS-122131-AK-L%.MET G:\Org\HP5\Methods\DS_8015-C24T-IK-L%.met	1050	1	1	1	0	The integration of Diesel Range Organics (C10-C24), C24-C40, and Total Extractable Hydrocarbons (TEH) is the hydrocarbon response with reference to the baseline with peak peak adjusted. Surrogates are integrated using a valley to valley integration using Set Baseline Now placed slight before at 12.11 minutes and 16.21 minutes and slightly after the surrogate peaks at 12.46 and 16.41 minutes. X-axis scaling showing surrogate peak from 11-18 minutes.
G:\org\HP5\DAT\HP5122121_b\1221HP5.32r	B21121622-003B ;1221HP5 , \$HC-8015-DRO-W, need rerun to verify no carryover	G:\Org\HP5\Methods\DR_8015-C24T-IJ-L0.met	900	1	1	1	0	No integrations
G:\org\HP5\DAT\HP5122121_b\1221HP5.33r	MARKER_1220HP535r_C40 ;1220HP5 , DRO211207A	G:\org\HP5\Methods\CS211221.met	1	1	1	1	0	No integrations
G:\org\HP5\DAT\HP5122121_b\1221HP5.34r	CCV_1221HP536r ,RRO ;1221HP5 , DRO211220C	G:\Org\HP5\Methods\DC_ORO-AK-L%.MET G:\Org\HP5\Methods\DS_ORO-AK-L%.MET	1	1	1	1	0	The integration of Oil Range hydrocarbon is the hydrocarbon response with reference to the baseline. Surrogates are integrated using a valley to valley integration using Set Baseline Now placed slight before at 16.18 minutes slightly after the surrogate peak at 16.4 minutes and X-axis scaling showing surrogate peak from 14-18 minutes.
G:\org\HP5\DAT\HP5122121_b\1221HP5.35r	CCV_1221HP537r ,DRO ;1221HP5 , DRO211220A	G:\Org\HP5\Methods\DC_8015-24-IK-L%.met G:\Org\HP5\Methods\DS_8015-24-IK-L%.met	1	1	1	1	0	The integration of Diesel Range Organics (C10-C24) and Total Extractable Hydrocarbons is the hydrocarbon response with reference to the baseline. Assigned Set Baseline All Valley on at 16.83 minutes. Surrogates are integrated using a valley to valley integration using Set Baseline Now placed slight before at 12.14 minutes and slightly after the surrogate peak at 12.31 minutes and X-axis scaling showing surrogate peak from 11.5-14 minutes.
G:\org\HP5\DAT\HP5122121_b\1221HP5.36r	DCM-Baseline Check-V36	G:\Org\HP5\Methods\DR_8015-IBB-LEXP.met	1	1	1	1	0	No integrations
G:\org\HP5\DAT\HP5122121_b\1221HP5.37r	DCM-Baseline Check-V37	G:\Org\HP5\Methods\DR_8015-IBB-LEXP.met	1	1	1	1	0	No integrations

G:\org\HP5\DAT\HP5122121_b\1221HP5.38r	B21121622-001B ;1221HP5 , \$HC-8015-DRO-W, rr	G:\Org\HP5\Methods\DR_8015-122138-IK-L%.met G:\Org\HP5\Methods\D3_OROS-122138-AK-L%.MET G:\Org\HP5\Methods\DS_8015-C24T-IK-L%.met	960	1	1	1	0	The integration of Diesel Range Organics (C10-C24), C24-C40, and Total Extractable Hydrocarbons (TEH) is the hydrocarbon response with reference to the baseline with Set Baseline Now at 23.94 minutes. Surrogates are integrated using a valley to valley integration using Set Baseline Now placed slight before at 12.11 minutes and 16.21 minutes and slightly after the surrogate peaks at 12.46 and 16.41 minutes. X-axis scaling showing surrogate peak from 11-18 minutes.
G:\org\HP5\DAT\HP5122121_b\1221HP5.39r	B21121622-003B ;1221HP5 , \$HC-8015-DRO-W, rr	G:\Org\HP5\Methods\DR_8015-C24T-IK-L%.met G:\Org\HP5\Methods\DR_OROS-AK-L%.MET G:\Org\HP5\Methods\DS_8015-C24T-IK-L%.met	900	1	1	1	0	The integration of Diesel Range Organics (C10-C24), C24-C40, and Total Extractable Hydrocarbons (TEH) is the hydrocarbon response with reference to the baseline. Assigned Set Baseline All Valley on at 17.12 minutes. Surrogates are integrated using a valley to valley integration using Set Baseline Now placed slight before at 12.11 minutes and 16.21 minutes and slightly after the surrogate peaks at 12.46 and 16.41 minutes. X-axis scaling showing surrogate peak from 11-18 minutes.
G:\org\HP5\DAT\HP5122121_b\1221HP5.40r G:\org\HP5\DAT\HP5122121_b\1221HP5.41r	DCM-Baseline Check-V40 B21121613-001BMS-RRO ;1221HP5 ,	G:\Org\HP5\Methods\DR_8015-IBb-LEXP.met G:\Org\HP5\Methods\D3_ORO-AK-L%.MET G:\Org\HP5\Methods\DS_ORO-AK-L%.MET	1 1045	1 1	1 1	1 1	0 0	No integrations The integration of Oil Range hydrocarbon is the hydrocarbon response with reference to the baseline. Surrogates are integrated using a valley to valley integration using Set Baseline Now placed slight before at 16.18 minutes slightly after the surrogate peak at 16.4 minutes and X-axis scaling showing surrogate peak from 14-18 minutes.
G:\org\HP5\DAT\HP5122121_b\1221HP5.42r G:\org\HP5\DAT\HP5122121_b\1221HP5.43r	DCM-Baseline Check-V42 B21121613-001BMSD-RRO ;1221HP5 ,	G:\Org\HP5\Methods\DR_8015-IBb-LEXP.met G:\Org\HP5\Methods\D3_ORO-AK-L%.MET G:\Org\HP5\Methods\DS_ORO-AK-L%.MET	1 1055	1 1	1 1	1 1	0 0	No integrations The integration of Oil Range hydrocarbon is the hydrocarbon response with reference to the baseline. Surrogates are integrated using a valley to valley integration using Set Baseline Now placed slight before at 16.18 minutes slightly after the surrogate peak at 16.4 minutes and X-axis scaling showing surrogate peak from 14-18 minutes.
G:\org\HP5\DAT\HP5122121_b\1221HP5.44r	LCS-162352-RRO ;1221HP5 ,	G:\Org\HP5\Methods\D3_ORO-AK-L%.MET G:\Org\HP5\Methods\DS_ORO-AK-L%.MET	1000	1	1	1	0	The integration of Oil Range hydrocarbon is the hydrocarbon response with reference to the baseline. Surrogates are integrated using a valley to valley integration using Set Baseline Now placed slight before at 16.18 minutes slightly after the surrogate peak at 16.4 minutes and X-axis scaling showing surrogate peak from 14-18 minutes.
G:\org\HP5\DAT\HP5122121_b\1221HP5.45r G:\org\HP5\DAT\HP5122121_b\1221HP5.46r	MARKER_1220HP546r_C40 ;1220HP5 , DRO211207A CCV_1221HP547r_RRO ;1221HP5 , DRO211220C	G:\org\HP5\Methods\CSC211221.met G:\Org\HP5\Methods\DC_ORO-AK-L%.MET G:\Org\HP5\Methods\DS_ORO-AK-L%.MET	1 1	1 1	1 1	1 1	0 0	No integrations The integration of Oil Range hydrocarbon is the hydrocarbon response with reference to the baseline. Surrogates are integrated using a valley to valley integration using Set Baseline Now placed slight before at 16.18 minutes slightly after the surrogate peak at 16.4 minutes and X-axis scaling showing surrogate peak from 14-18 minutes.
G:\org\HP5\DAT\HP5122121_b\1221HP5.47r	CCV_1221HP548r_DRO ;1221HP5 , DRO211220A	G:\Org\HP5\Methods\DC_8015-24-IK-L%.met G:\Org\HP5\Methods\DS_8015-24-IK-L%.met	1	1	1	1	0	The integration of Diesel Range Organics (C10-C24) and Total Extractable Hydrocarbons is the hydrocarbon response with reference to the baseline. Assigned Set Baseline All Valley on at 16.83 minutes. Surrogates are integrated using a valley to valley integration using Set Baseline Now placed slight before at 12.14 minutes and slightly after the surrogate peak at 12.31 minutes and X-axis scaling showing surrogate peak from 11.5-14 minutes.

Ann Nebel

Digitally signed by
Ann Nebel
Date: 2022.01.11 13:09:01 -07:00

G:\org\HP5\DAT\HP5122621_bi1226HP5.29	B21121605-002A ;1226HP5 , \$HC-8015-DRO-W, SGT	G:\Org\HP5\Methods\D3_8015-C24T-IL-L%.met G:\Org\HP5\Methods\D3_OROS-AL-L%.MET G:\Org\HP5\Methods\DS_8015-C24T-IL-L#.met	1050	1	1	1	0	The integration of Diesel Range Organics (C10-C24), C24-C40, and Total Extractable Hydrocarbons (TEH) is the hydrocarbon response with reference to the baseline. Surrogates are integrated using a valley to valley integration using Set Baseline All Valleys on at 10.78 minutes and X-axis scaling showing surrogate peak from 11-18 minutes.
G:\org\HP5\DAT\HP5122621_bi1226HP5.26	B21121605-003A ;1226HP5 , \$HC-8015-DRO-W, SGT	G:\Org\HP5\Methods\D3_8015-122626-IL-L%.met G:\Org\HP5\Methods\D3_OROS-122626-AL-L%.MET G:\Org\HP5\Methods\DS_8015-C24T-IL-L#.met	1010	1	1	1	0	The integration of Diesel Range Organics (C10-C24), C24-C40, and Total Extractable Hydrocarbons (TEH) is the hydrocarbon response with reference to the baseline. Assigned Set Now at 22.97 minutes. Surrogates are integrated using a valley to valley integration using Set Baseline All Valleys on at 10.78 minutes and X-axis scaling showing surrogate peak from 11-18 minutes.
G:\org\HP5\DAT\HP5122621_bi1226HP5.27	B21121609-001C ;1226HP5 , \$HC-8015-DRO-W, SGT	G:\Org\HP5\Methods\D3_8015-122627-IL-L%.met G:\Org\HP5\Methods\D3_OROS-122627-AL-L%.MET G:\Org\HP5\Methods\DS_8015-C24T-IL-L#.met	1060	1	1	1	0	The integration of Diesel Range Organics (C10-C24), C24-C40, and Total Extractable Hydrocarbons (TEH) is the hydrocarbon response with reference to the baseline. Assigned Set Now at 26.45 minutes. Surrogates are integrated using a valley to valley integration using Set Baseline All Valleys on at 10.78 minutes and X-axis scaling showing surrogate peak from 11-18 minutes.
G:\org\HP5\DAT\HP5122621_bi1226HP5.28	B21121613-001BMS-RRO ;1226HP5 , SGT	G:\Org\HP5\Methods\D3_ORO-AL-L%.MET G:\Org\HP5\Methods\DS_ORO-AL-L%.MET	1045	1	1	1	0	The integration of Oil Range hydrocarbon is the hydrocarbon response with reference to the baseline. Surrogates are integrated using a valley to valley integration using Set Baseline Now placed slight before at 16.18 minutes slightly after the surrogate peak at 16.4 minutes and X-axis scaling showing surrogate peak from 14-18 minutes.
G:\org\HP5\DAT\HP5122621_bi1226HP5.29	DCM-Baseline Check-V29	G:\Org\HP5\Methods\DR_8015-18b-LEXP.met	1	1	1	1	0	No integration
G:\org\HP5\DAT\HP5122621_bi1226HP5.30	LCS-162352 ;1226HP5 , RR-SGT	G:\Org\HP5\Methods\D3_8015-122630-IL-L%.met G:\Org\HP5\Methods\DS_8015-24-IL-L#.met	1000	1	1	1	0	The integration of Diesel Range Organics (C10-C24) and Total Extractable Hydrocarbons is the hydrocarbon response with reference to the baseline with peak width adjusted. Surrogates are integrated using a valley to valley integration using Set Baseline Now placed slight before at 12.14 minutes and slightly after the surrogate peak at 12.31 minutes and X-axis scaling showing surrogate peak from 11.5-14 minutes.
G:\org\HP5\DAT\HP5122621_bi1226HP5.31	B21121613-001BMSD-RRO ;1226HP5 , SGT	G:\Org\HP5\Methods\D3_ORO-AL-L%.MET G:\Org\HP5\Methods\DS_ORO-AL-L%.MET	1055	1	1	1	0	The integration of Oil Range hydrocarbon is the hydrocarbon response with reference to the baseline. Surrogates are integrated using a valley to valley integration using Set Baseline Now placed slight before at 16.18 minutes slightly after the surrogate peak at 16.4 minutes and X-axis scaling showing surrogate peak from 14-18 minutes.
G:\org\HP5\DAT\HP5122621_bi1226HP5.32	LCS-162352-RRO ;1226HP5 , NEEDS RR	G:\Org\HP5\Methods\DR_8015-18b-LEXP.met	1000	1	1	1	0	No integration
G:\org\HP5\DAT\HP5122621_bi1226HP5.33	MARKER_1226HP533r_C40 ;1226HP5 , DRO211207A	g:\org\HP5\Methods\CSC211226.met	1	1	1	1	0	No integration
G:\org\HP5\DAT\HP5122621_bi1226HP5.34	CCV_1226HP534r_RRO ;1226HP5 , DRO211220C	G:\Org\HP5\Methods\DC_ORO-AL-L%.MET G:\Org\HP5\Methods\DS_ORO-AL-L%.MET	1	1	1	1	0	The integration of Oil Range hydrocarbon is the hydrocarbon response with reference to the baseline. Surrogates are integrated using a valley to valley integration using Set Baseline Now placed slight before at 16.18 minutes slightly after the surrogate peak at 16.4 minutes and X-axis scaling showing surrogate peak from 14-18 minutes.
G:\org\HP5\DAT\HP5122621_bi1226HP5.35	CCV_1226HP535r_DRO ;1226HP5 , DRO211220A	G:\Org\HP5\Methods\DC_8015-24-IL-L%.met G:\Org\HP5\Methods\DS_8015-24-IL-L#.met	1	1	1	1	0	The integration of Diesel Range Organics (C10-C24) and Total Extractable Hydrocarbons is the hydrocarbon response with reference to the baseline. Assigned Set Baseline All Valley on at 16.83 minutes. Surrogates are integrated using a valley to valley integration using Set Baseline Now placed slight before at 12.14 minutes and slightly after the surrogate peak at 12.31 minutes and X-axis scaling showing surrogate peak from 11.5-14 minutes.
G:\org\HP5\DAT\HP5122621_bi1226HP5.36	DCM-Baseline Check-V36	G:\Org\HP5\Methods\DR_8015-18b-LEXP.met	1	1	1	1	0	No integration
G:\org\HP5\DAT\HP5122621_bi1226HP5.49	MARKER_1226HP549r_C40 ;1226HP5 , DRO211207A	g:\org\HP5\Methods\CSC211226.met	1	1	1	1	0	No integration
G:\org\HP5\DAT\HP5122621_bi1226HP5.50	CCV_1226HP550r_RRO ;1226HP5 , DRO211220C	G:\Org\HP5\Methods\DC_ORO-AL-L%.MET G:\Org\HP5\Methods\DS_ORO-AL-L%.MET	1	1	1	1	0	The integration of Oil Range hydrocarbon is the hydrocarbon response with reference to the baseline. Surrogates are integrated using a valley to valley integration using Set Baseline Now placed slight before at 16.18 minutes slightly after the surrogate peak at 16.4 minutes and X-axis scaling showing surrogate peak from 14-18 minutes.
G:\org\HP5\DAT\HP5122621_bi1226HP5.51	CCV_1226HP551r_DRO ;1226HP5 , DRO211220A	G:\Org\HP5\Methods\DC_8015-24-IL-L%.met G:\Org\HP5\Methods\DS_8015-24-IL-L#.met	1	1	1	1	0	The integration of Diesel Range Organics (C10-C24) and Total Extractable Hydrocarbons is the hydrocarbon response with reference to the baseline. Assigned Set Baseline All Valley on at 16.83 minutes. Surrogates are integrated using a valley to valley integration using Set Baseline Now placed slight before at 12.14 minutes and slightly after the surrogate peak at 12.31 minutes and X-axis scaling showing surrogate peak from 11.5-14 minutes.
G:\org\HP5\DAT\HP5122621_bi1226HP5.52	DCM-Baseline Check-V52	G:\Org\HP5\Methods\DR_8015-18b-LEXP.met	1	1	1	1	0	No integration
G:\org\HP5\DAT\HP5122621_bi1226HP5.56	DCM-Baseline Check-V56	G:\Org\HP5\Methods\DR_8015-18b-LEXP.met	1	1	1	1	0	No integration
G:\org\HP5\DAT\HP5122621_bi1226HP5.57	LCS-162352-RRO ;1226HP5 , RR-SGT	G:\Org\HP5\Methods\D3_8015-122630-IL-L%.met G:\Org\HP5\Methods\DS_ORO-AL-L%.MET	1000	1	1	1	0	The integration of Oil Range hydrocarbon is the hydrocarbon response with reference to the baseline. Surrogates are integrated using a valley to valley integration using Set Baseline Now placed slight before at 16.18 minutes slightly after the surrogate peak at 16.4 minutes and X-axis scaling showing surrogate peak from 14-18 minutes.
G:\org\HP5\DAT\HP5122621_bi1226HP5.60	MARKER_1226HP560r_C40 ;1226HP5 , DRO211207A	g:\org\HP5\Methods\CSC211226.met	1	1	1	1	0	No integration
G:\org\HP5\DAT\HP5122621_bi1226HP5.61	CCV_1226HP561r_RRO ;1226HP5 , DRO211220C	G:\Org\HP5\Methods\DC_ORO-AL-L%.MET G:\Org\HP5\Methods\DS_ORO-AL-L%.MET	1	1	1	1	0	The integration of Oil Range hydrocarbon is the hydrocarbon response with reference to the baseline. Surrogates are integrated using a valley to valley integration using Set Baseline Now placed slight before at 16.18 minutes slightly after the surrogate peak at 16.4 minutes and X-axis scaling showing surrogate peak from 14-18 minutes.

Ann Nebel

Digitally signed by
Ann Nebel
Date: 2022.01.11 13:09:31 -07:00

Energy Laboratories Inc

Spike LOG

Standard ID: DRO211101A
Standard Name: OTP-4000 ug/mL DCM
Date Prepared: 11/1/2021
Date Expires: 9/30/2024
Department: dropr
Vendor:
Lot Number:
Balance ID: BAL-DRO
Comments: Used to Prep DRO-8015 ICAL and CCV Solutions

Type: Secondary
BY: Ann Nebel
Status: Open

Chemical / Solvent Used	BottleNo	Amt	Units	Exp
Dichloromethane EC328	14408	25	mL	8/19/

Final Volume: 25 mL

Stock Source

DRO200430B O-Terphenyl

Base Units

ug/mL

Amount Added

0.1012 g

Analtes

A O-Terphenyl

CAS

84-15-1

Conc:

ug/mL

4000

Energy Laboratories Inc

Standard LOG

Standard ID: DRO200430B
Standard Name: O-Terphenyl
Date Prepared: 4/30/2020
Date Expires: 9/30/2024
Department: dropr
Vendor: Chemservice
Lot Number: 9972100
Balance ID:
Comments: ID#: 6271

Type: Neat
BY: Ann Nebel
Status: New

Chemical / Solvent Used	BottleNo	Amt	Units	Exp
o-Terphenyl	12650	500	mg	9/30/

Final Volume: mL

Stock Source

Base Units

Amount Added

Analtes

CAS

Conc: ug/mL

A O-Terphenyl

84-15-1

1

660 Tower Lane • P.O. Box 599 • West Chester, PA 19381-0599
1-800-452-9994 • 1-610-692-3026 • Fax 1-610-692-8729
info@chemservice.com • www.chemservice.com

CERTIFICATE OF ANALYSIS

o-Terphenyl

CATALOG NUMBER N-12693-500MG
LOT NUMBER 9972100
DATE CERTIFIED 09/23/19
EXPIRATION DATE 09/30/24
CAS NUMBER 84-15-1
MOLECULAR FORMULA C18H14
MOLECULAR WEIGHT 230.32
STORAGE Store in a cool dry place.
HANDLING See Safety Data Sheet
INTENDED USE For laboratory use only.

Analytical Test	Value
FT-IR SPECTROSCOPY	CONFORMS TO STRUCTURE
GC/MS SPECTRA ID	MATCHES NIST DATABASE
MELTING POINT (°C)	57.1
% PURITY (GC/FID)	99.5

Chem Service, Inc. guarantees the purity to be +/- 0.5% deviation prior to the expiration date shown on the label and exclusive of any customer contamination.

Certified By:

Mary Beth O'Donnell

Mary Beth O'Donnell
CSM/TC

ID #: 12650

Opened: _____

o-Terphenyl

Expires: 9/30/2024

Rec'd: 4/30/2020

Energyl Laboratories Inc 1120 So. 27th Street
Billings MT 59107


Chem Service is accredited to ISO 17034:2016, ISO/IEC 17025:2017 and certified to ISO 9001:2015



CERTIFICATE OF ANALYSIS

Analysis Method:

Catalog Number: N-12693-500MG
Description: o-Terphenyl
Lot Number: 9972100
Expiration Date: 09/30/24



Chem Service is accredited to ISO 17034:2016, ISO/IEC 17025:2017 and certified to ISO 9001:2015



CERTIFICATE OF ANALYSIS

Analysis Method:

Catalog Number: N-12693-500MG
Description: o-Terphenyl
Lot Number: 9972100
Expiration Date: 09/30/24

Chem Service Inc Area Percent Report

Data File: D:\msdchem\2019 DATA\0919\0923-01.D
Acq On : 23 Sep 2019 10:40
Operator :
Sample : n-12693
Misc :
ALS Vial : 95

Integration Parameters: autoint1.e
Integrator: ChemStation

DataAcq Meth: SCREEN.M
Method : D:\msdchem\2019 DATA\0919\0903-09.D\ERIN.M

Signal : TIC: 0923-01.D\data.ms

peak #	R.T. min	first scan	max scan	last scan	PK TY	peak height	corr. area	corr. % max.	% of total
1	11.844	1597	1606	1613	BB	32038221	432253484	100.00%	100.000%

Sum of corrected areas: 432253484

ERIN.M Mon Sep 23 10:55:51 2019


660 Tower Lane • P.O. Box 599 • West Chester, PA 19381-0599
1-800-452-9994 • 1-610-692-3026 • Fax 1-610-692-8729
info@chemservice.com • www.chemservice.com

CERTIFICATE OF ANALYSIS


Analysis Method:

Catalog Number: N-12693-500MG
Description: o-Terphenyl
Lot Number: 9972100
Expiration Date: 09/30/24

Abundance



Chem Service is accredited to ISO 17034:2016, ISO/IEC 17025:2017 and certified to ISO 9001:2015




660 Tower Lane • P.O. Box 599 • West Chester, PA 19381-0599
1-800-452-9994 • 1-610-692-3026 • Fax 1-610-692-8729
info@chemservice.com • www.chemservice.com

CERTIFICATE OF ANALYSIS

Analysis Method:

Catalog Number: N-12693-500MG
Description: o-Terphenyl
Lot Number: 9972100
Expiration Date: 09/30/24

Abundance



Chem Service is accredited to ISO 17034:2016, ISO/IEC 17025:2017 and certified to ISO 9001:2015.



660 Tower Lane • P.O. Box 599 • West Chester, PA 19381-0599
1-800-452-9994 • 1-610-692-3026 • Fax 1-610-692-8729
info@chemservice.com • www.chemservice.com

CERTIFICATE OF ANALYSIS

Analysis Method:

Catalog Number:	N-12693-500MG
Description:	o-Terphenyl
Lot Number:	9972100
Expiration Date:	09/30/24

Chem Service is accredited to ISO 17034:2016, ISO/IEC 17025:2017 and certified to ISO 9001:2015




660 Tower Lane • P.O. Box 599 • West Chester, PA 19381-0599
 1-800-452-9994 • 1-610-692-3026 • Fax 1-610-692-8729
info@chemservice.com • www.chemservice.com

Gas

Data file: C:\CHEM3\
 Sample name: N-12893
 Instrument: GC 2
 Injection date: 8/23/2019 9:58:34 AM
 Acq. method: SCREEN.M
 Column name: HP-5

CERTIFICATE OF ANALYSIS

Location: Vial 141
 Injection volume: 1.0uL



Signal: FID1 A, Front Signal

RT [min]	Type	Width [min]	Area	Height	Area%
4.565	BB	0.0305	1.2408	0.5122	0.11
20.152	BB	0.0391	1171.9556	439.4599	99.89
		Sum	1173.1963		

Chem Service is accredited to ISO 17034:2016, ISO/IEC 17025:2017 and certified to ISO 9001:2015



Energy Laboratories Inc

Standard LOG

Standard ID: DRO211012A
Standard Name Diesel Fuel #2 50,000 ug/mL in DCM
Date Prepared 10/12/2021
Date Expires: 4/30/2023
Department dropr
Vendor: Sigma-Aldrich
Lot Number: LRAC6316
Balance ID:
Comments: Diesel Fuel #2 For CCVs.

Type: Primary
BY: Ann Nebel
Status: New

Chemical / Solvent Used	BottleNo	Amt	Units	Exp
Diesel Fuel No. 2	14376	1	mL	4/30/

Final Volume: mL

Stock Source

Base Units

Amount Added

Analtes

CAS

Conc: ug/mL

Diesel Fuel #2

0

Certificate of Analysis

Certified
Reference
Material

Diesel Fuel No. 2

Description

Product ID UST148
Lot LRAC6316
Expiration Date April 2023
Manufacturing Date April 2020
Storage Conditions Room Temperature
Solvent/Matrix DICHLOROMETHANE

ID #: 14376

Opened: _____

Diesel Fuel No. 2

Expires: 4/30/2023


Rec'd: 10/12/2021

Energy Laboratories Inc 1120 So. 27th Street
Billings MT 59107

Certified Values

Analyte	Certified Value ^{1,4}	Units	Raw Material Purity,%	Raw Material Lot	CAS
NO.2 FUEL OIL	50001 ± 2770	µg/mL	100.0	LA80505	68476-34-6

Informational Values



Additional Information:

Analytical Method Parameters:

Column: SPB-5, 30 m × 0.53 mm I.D., 1.5 µm film thickness (Column #214)

Carrier Gas: H₂, Flow: 4.0 mL/min

Inlet Temperature: 250 °C, Injection Volume: 1.0 µL

Injection Mode: Split, Split Ratio: 10:1

Temperature Program: 40 °C (Hold 2 min) @ 15 °C/min to 300 °C (Hold 5 min)

Detector: FID

Detector Temperature: 300 °C



SIGMA-ALDRICH

2931 Soldier Springs Rd. Laramie, Wyoming 82070 USA
800-325-5832
TechService@milliporesigma.com www.sigma-aldrich.com

Description

Lot **LRAC6316**
Expiration Date April 2023
Manufacturing Date April 2020
Storage Conditions Room Temperature
Solvent/Matrix DICHLOROMETHANE

1 Metrological traceability: Traceable to the SI and higher order standards from NIST through an unbroken chain of comparisons. The balance used to weigh raw materials is accurate to +/-0.0001 g and calibrated regularly using mass standards traceable to NIST. All dilutions were performed gravimetrically. Additionally, individual analytes are traceable to NIST SRMs where available and specified above.
4 Ucrm - Uncertainty values in this document are expressed as Expanded Uncertainty (Ucrm) corresponding to the 95% confidence interval. Ucrm is derived from the combined standard uncertainty multiplied by the coverage factor k, which is obtained from a t-distribution and degrees of freedom. The components of combined standard uncertainty include the uncertainties due to characterization, homogeneity, long term stability, and short term stability (transport). The components due to stability are generally considered to be negligible unless otherwise indicated by stability studies. The mathematical representation of the Ucrm calculation is as follows:

$$u_{CRM} = \sqrt{u_{char}^2 + u_{homogeneity}^2 + u_{stability}^2}$$

k: Coverage factor derived from a t-distribution table, based on the degrees of freedom of the data set. Assume 2.0 for a **Confidence interval = 95%**

6 Analytical Value- For QC verification of the certified value only- not to be used in calculations. Represents the analytical data obtained by comparison to a standard as analyzed by the method described in the CoA or another acceptable method. The result may differ from the certified value and UCRM based on method uncertainty as well as the uncertainty associated with the standard used for comparison.


Traceability: The standard was manufactured under an ISO/IEC 17025:2017 certified quality system. The balance used to weigh raw materials is accurate to +/- 0.0001g and calibrated regularly using mass standards traceable to NIST. All dilutions were performed gravimetrically. Additionally, individual analytes are traceable to NIST SRMs where available and specified above.

Homogeneity: Homogeneity was assessed in accordance with ISO 17034:2016. Completed units were sampled using a random stratified sampling protocol. The results of chemical analysis were then compared using a one-way analysis of variance approach as described by TNI EL-V3-2009 Appendix A.2. See Instructions for minimum sub-sample size.


Expiration is at end of month given on certificate and label.

MSDS reports for components comprising greater than 1.0% of the solution or 0.1% for components known to be carcinogens are available upon request.

THIS PRODUCT WAS DESIGNED, PRODUCED AND VERIFIED FOR ACCURACY AND STABILITY IN ACCORDANCE WITH ISO/IEC 17025:2017 (ANAB Cert AT-1467) and ISO 17034:2016 (ANAB Cert AR-1470).



Andy Ommen - QC Manager



Mark Pooler - QA Supervisor

Certification Date April 30, 2020
Version 0-4302020



Energy Laboratories Inc

Standard LOG

Standard ID: DRO180918C
Standard Name: 50,000 ug/mL Oil Std For AK103 RRO-In DC
Date Prepared: 9/18/2018
Date Expires: 8/31/2025
Department: dropr
Vendor: Restek
Lot Number: A0140080
Balance ID: Sartorius 4 place balance

Type: Primary
BY: Ann Nebel
Status: Open

Comments:

Chemical / Solvent Used	BottleNo	Amt	Units	Exp
Residual Range Calibration Standard	10787	1	mL	8/31/

Final Volume: 1 mL

Stock Source

Base Units

Amount Added

Analtes

CAS

Conc: **ug/mL**



CERTIFIED REFERENCE MATERIAL

110 Benner Circle
Bellefonte, PA 16823-8812
Tel: (800)356-1688
Fax: (814)353-1309

www.restek.com

Certificate of Analysis



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31817 Lot No.: A0140080

Description : Residual Range Calibration Standard (RCS)
Residual Range Calib Std (RCS) 50,000µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL Pkg Amt: > 1 mL

Expiration Date : August 31, 2025 Storage: 25°C nominal

CERTIFIED VALUES

Elution Order	Compound	Grav. Conc. (weight/volume)	Expanded Uncertainty (95% C.L.; K=2)		
1	Motor Oil SAE30 & SAE40 Blend (Pennzoil) CAS # 64742-65-0.F (Lot A0126386) Purity ----%	50,113.0 µg/mL	+/- 293.4226	µg/mL	Gravimetric
			+/- 1,492.4284	µg/mL	Unstressed
			+/- 1,591.6738	µg/mL	Stressed

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

ID #: 10787
Opened: _____
Residual Range Calibration Standard
Expires: **8/31/2025**
Rec'd: 9/18/2018
Eneray Laboratories Inc 1120 So. 27th Street
Billings MT 59107

Column:
30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)


Carrier Gas:
hydrogen-constant pressure 10 psi.

Temp. Program:
40°C (hold 2 min.) to 330°C
@ 10°C/min. (hold 10 min.)

Inj. Temp:
250°C

Det. Temp:
330°C

Det. Type:
FID



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Brandon Reish
Brandon Reish - Mix Technician

Date Mixed: 28-Jul-2018

Balance: B345965662

Diane Shaffer
Diane Shaffer - Operations Tech-ARM QC

Date Passed: 30-Jul-2018

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

Energy Laboratories Inc

Standard LOG

Standard ID: DRO211006A
Standard Name: Triacontane SURR 2000 ug/mL
Date Prepared: 10/6/2021
Date Expires: 4/6/2026
Department: dropr
Vendor:
Lot Number:
Balance ID: BAL-DRO
Comments: Triacontane SURR 2000 ug/mL

Type: Secondary
BY: Jillian L Bostwick
Status: New

Chemical / Solvent Used	BottleNo	Amt	Units	Exp
Acetone DZ509	13553	50	mL	7/22/

Final Volume: 50 mL

Stock Source
DRO210406A Triacontane-d62 Surr For AK103 RRO

Base Units
ug/mL

Amount Added
0.1001 g

Analtes
A Triacontane-d62

CAS

Conc: **ug/mL**
2000

Energy Laboratories Inc

Standard LOG

Standard ID: DRO210406A
Standard Name: Triacontane-d62 Surr For AK103 RRO
Date Prepared: 4/6/2021
Date Expires: 4/6/2026
Department: dropr
Vendor: Sigma-Aldrich
Lot Number: MBBC4347
Balance ID:
Comments: Alaska surr [for AK103 RRO]

Type: Neat
BY: Ann Nebel
Status: New

Chemical / Solvent Used	BottleNo	Amt	Units	Exp
Triacontane-d62-98 atom % D	13736		mL	4/6/2026

Final Volume: mL

Stock Source

Base Units

Amount Added

Analtes

CAS

Conc: ug/mL

A Triacontane-d62

1

3050 Spruce Street, Saint Louis, MO 63103, USA
 Website: www.sigmaaldrich.com
 Email USA: techserv@sial.com
 Outside USA: eurtechserv@sial.com

Certificate of Analysis

Product Name:
 Triacontane-d62 - 98 atom % D

Product Number: 451789
 Batch Number: MBBC4347
 Brand: ALDRICH
 CAS Number: 93952-07-9
 MDL Number: MFCD00209794
 Formula: C30D62
 Formula Weight: 485.20 g/mol
 Quality Release Date: 27 APR 2018



ID #: 13736

Opened: _____


Triacontane-d62-98 atom % D

Expires: 4/6/2026

Rec'd: 4/6/2021

Energx Laboratories Inc 1120 So. 27th Street
 Billings MT 59107

Test	Specification	Result
Purity (HPLC)	≥ 99.0 %	99.0 %
Proton NMR Spectrum	Conforms to Structure	Conforms
D Enrichment	≥ 98.0 %	99.0 %
Initial Melting Point		60.0 °C
Final Melting Point		62.0 °C



Greg Abernathy, Supervisor
 Quality Control
 Miamisburg, Ohio US

Sigma-Aldrich warrants, that at the time of the quality release or subsequent retest date this product conformed to the information contained in this publication. The current Specification sheet may be available at Sigma-Aldrich.com. For further inquiries, please contact Technical Service. Purchaser must determine the suitability of the product for its particular use. See reverse side of invoice or packing slip for additional terms and conditions of sale.

Energy Laboratories Inc

Standard LOG

Standard ID: DRO211012B
Standard Name #2 Diesel in Acetone 150,000 ug/mL Type: Secondary
Date Prepared 10/12/2021 BY: Ann Nebel
Date Expires: 11/5/2023
Department dropr Status: New
Vendor:
Lot Number:
Balance ID: BAL-DRO
Comments: #2 Diesel in Acetone 150,000 ug/mL.

Chemical / Solvent Used	BottleNo	Amt	Units	Exp
Acetone EA662	14050	25	mL	1/7/2

Final Volume: 25 mL

Stock Source

DRO181105A #2 Diesel (NEAT)

Base Units

ug/mL

Amount Added

3.7507 g

Analtes

A #2 Diesel

CAS

68476-34-6

Conc:

ug/mL

150000

Energy Laboratories Inc

Standard LOG

Standard ID: DRO181105A
Standard Name: #2 Diesel (NEAT) Type: Neat
Date Prepared: 11/5/2018 BY: Ann Nebel
Date Expires: 11/5/2023
Department: dropr Status: New
Vendor: conoco
Lot Number:
Balance ID:
Comments: -18 Cloud peak. (Conoco Gas Station 1240 S. 27th Billings, MT) 2nd Source

<u>Stock Source</u>	<u>Base Units</u>	<u>Final Volume:</u>	<u>Amount Added</u>
<u>Analvtes</u>	<u>CAS</u>	250 mL	
A #2 Diesel	68476-34-6	Conc:	ug/mL 1

Energy Laboratories Inc

Standard LOG

Standard ID: DRO210217A
 Standard Name: 20,000 ug/mL Oil Std For AK103 RRO-In DC
 Date Prepared: 2/17/2021
 Date Expires: 8/23/2021
 Department: dropr
 Vendor:
 Lot Number:
 Balance ID: Sartorius 4 place balance
 Type: Secondary
 BY: Ann Nebel
 Status: Expired
 Comments:

Chemical / Solvent Used	BottleNo	Amt	Units	Exp
Dichloromethane EA342	13510	25	mL	11/17

Final Volume: 25 mL

<u>Stock Source</u>	<u>Base Units</u>	<u>Amount Added</u>
DRO160823C 30W Motor Oil-Valvoline	ug/mL	0.2501 g
DRO160823D 40W Motor Oil-Valvoline	ug/mL	0.2527 g

<u>Analtes</u>	<u>CAS</u>	<u>Conc:</u>	<u>ug/mL</u>
A 30W Motor Oil			10000
A 30W-Motor oil			0
A 40W Motor Oil			10000
A 40W-Motor oil			0

Energy Laboratories Inc

Standard LOG

Standard ID: DRO160823C
Standard Name: 30W Motor Oil-Valvoline
Date Prepared: 8/23/2016
Date Expires: 8/23/2021
Department: dropr
Vendor:
Lot Number:
Balance ID:

Type: Primary
BY: Todd C Cooper
Status: Expired

Comments: Used to make 2nd Source Standard for AK103 method.

Chemical / Solvent Used	BottleNo	Amt	Units	Exp
Valvoline SAE 30 Motor Oil	8637		mL	8/23/

Final Volume: mL

Stock Source

Base Units

Amount Added

Analtes

CAS

Conc: ug/mL

A 30W-Motor oil

1

Energy Laboratories Inc

Standard LOG

Standard ID: DRO160823D
Standard Name: 40W Motor Oil-Valvoline
Date Prepared: 8/23/2016
Date Expires: 8/23/2021
Department: dropr
Vendor:
Lot Number:
Balance ID:
Type: Primary
BY: Todd C Cooper
Status: Expired
Comments: Used to Make 2nd Source Standards For Alaska AK103 RRO Method

Chemical / Solvent Used	BottleNo	Amt	Units	Exp
Valvoline SAE 40 Motor Oil	8638		mL	8/23/

Final Volume: mL

Stock Source

Base Units

Amount Added

Analtes

CAS

Conc: ug/mL

A 40W-Motor oil

1

Energy Laboratories Inc

Standard LOG

Standard ID: DRO211220A
 Standard Name: 8015 CCV-15,000ug/mL + 200 OTP
 Date Prepared: 12/20/2021
 Date Expires: 4/30/2023
 Department: dropr
 Vendor:
 Lot Number:
 Balance ID:
 Comments: 8015DRO CCV MIX-15,000ug/mL +200 OTP #2 Diesel

Type: Secondary
 BY: Ann Nebel
 Status: New

Chemical / Solvent Used	BottleNo	Amt	Units	Exp
Dichloromethane EC757	14596	2.6	mL	10/20

Final Volume: 4 mL

<u>Stock Source</u>	<u>Base Units</u>	<u>Amount Added</u>
DRO211101A OTP-4000 ug/mL DCM	ug/mL	0.2 mL
DRO211102B Diesel Fuel #2 50,000 ug/mL in DCM	ug/mL	1.2 mL

<u>Analtes</u>	<u>CAS</u>	<u>Conc:</u>	<u>ug/mL</u>
A #2 Diesel			15000
Diesel Fuel #2			0
A O-Terphenyl	84-15-1		200

Energy Laboratories Inc

Standard LOG

Standard ID: DRO211102B
Standard Name Diesel Fuel #2 50,000 ug/mL in DCM Type: Primary
Date Prepared 11/2/2021 BY: Ann Nebel
Date Expires: 4/30/2023
Department dropr Status: New
Vendor: Sigma-Aldrich
Lot Number: LRAC6316
Balance ID:
Comments: Diesel Fuel #2 For CCVs.

Chemical / Solvent Used	BottleNo	Amt	Units	Exp
Diesel Fuel No. 2	14478	1	mL	4/30/

Final Volume: mL

Stock Source

Base Units

Amount Added

Analtes

CAS

Conc: ug/mL

Diesel Fuel #2

0

Certificate of Analysis

Diesel Fuel No. 2

Certified
Reference
Material

Description

Product ID UST148
Lot LRAC6316
Expiration Date April 2023
Manufacturing Date April 2020
Storage Conditions Room Temperature
Solvent/Matrix DICHLOROMETHANE

ID #: 14478

Opened: _____

Diesel Fuel No. 2

Expires: 4/30/2023


Rec'd: 11/2/2021

Energex Laboratories Inc 1120 So. 27th Street
Billings MT 59107

Certified Values

Analyte	Certified Value ^{1,4}	Units	Raw Material Purity, %	Raw Material Lot	CAS
NO.2 FUEL OIL	50001 ± 2770	µg/mL	100.0	LA80505	68476-34-6

Informational Values



Additional Information:

Analytical Method Parameters:

Column: SPB-5, 30 m × 0.53 mm I.D., 1.5 µm film thickness (Column #214)

Carrier Gas: H₂, Flow: 4.0 mL/min

Inlet Temperature: 250 °C, Injection Volume: 1.0 µL

Injection Mode: Split, Split Ratio: 10: 1

Temperature Program: 40 °C (Hold 2 min) @ 15 °C/min to 300 °C (Hold 5 min)

Detector: FID

Detector Temperature: 300 °C



SIGMA-ALDRICH

2931 Soldier Springs Rd. Laramie, Wyoming 82070 USA
800-325-5832
TechService@milliporesigma.com www.sigma-aldrich.com

Description

Lot **LRAC6316**
Expiration Date April 2023
Manufacturing Date April 2020
Storage Conditions Room Temperature
Solvent/Matrix DICHLOROMETHANE

1 Metrological traceability: Traceable to the SI and higher order standards from NIST through an unbroken chain of comparisons. The balance used to weigh raw materials is accurate to +/-0.0001 g and calibrated regularly using mass standards traceable to NIST. All dilutions were performed gravimetrically. Additionally, individual analytes are traceable to NIST SRMs where available and specified above.

4 Ucrm - Uncertainty values in this document are expressed as Expanded Uncertainty (Ucrm) corresponding to the 95% confidence interval. Ucrm is derived from the combined standard uncertainty multiplied by the coverage factor k, which is obtained from a t-distribution and degrees of freedom. The components of combined standard uncertainty include the uncertainties due to characterization, homogeneity, long term stability, and short term stability (transport). The components due to stability are generally considered to be negligible unless otherwise indicated by stability studies. The mathematical representation of the Ucrm calculation is as follows:

$$u_{CRM} = \sqrt{u_{char}^2 + u_{homogeneity}^2 + u_{stability}^2}$$

k: Coverage factor derived from a t-distribution table, based on the degrees of freedom of the data set. Assume 2.0 for a **Confidence interval = 95%**

6 Analytical Value- For QC verification of the certified value only- not to be used in calculations. Represents the analytical data obtained by comparison to a standard as analyzed by the method described in the CoA or another acceptable method. The result may differ from the certified value and UCRM based on method uncertainty as well as the uncertainty associated with the standard used for comparison.


Traceability: The standard was manufactured under an ISO/IEC 17025:2017 certified quality system. The balance used to weigh raw materials is accurate to +/- 0.0001g and calibrated regularly using mass standards traceable to NIST. All dilutions were performed gravimetrically. Additionally, individual analytes are traceable to NIST SRMs where available and specified above.

Homogeneity: Homogeneity was assessed in accordance with ISO 17034:2016. Completed units were sampled using a random stratified sampling protocol. The results of chemical analysis were then compared using a one-way analysis of variance approach as described by TNI EL-V3-2009 Appendix A.2. See Instructions for minimum sub-sample size.

Expiration is at end of month given on certificate and label.

MSDS reports for components comprising greater than 1.0% of the solution or 0.1% for components known to be carcinogens are available upon request.

THIS PRODUCT WAS DESIGNED, PRODUCED AND VERIFIED FOR ACCURACY AND STABILITY IN ACCORDANCE WITH **ISO/IEC 17025:2017 (ANAB Cert AT-1467)** and **ISO 17034:2016 (ANAB Cert AR-1470)**.



Andy Ommen - QC Manager



Mark Pooler - QA Supervisor

Certification Date April 30, 2020
Version 0-4302020



Energy Laboratories Inc

Spike LOG

Standard ID: DRO211101A
Standard Name: OTP-4000 ug/mL DCM
Date Prepared: 11/1/2021
Date Expires: 9/30/2024
Department: dropr
Vendor:
Lot Number:
Balance ID: BAL-DRO
Comments: Used to Prep DRO-8015 ICAL and CCV Solutions

Type: Secondary
BY: Ann Nebel
Status: Open

Chemical / Solvent Used	BottleNo	Amt	Units	Exp
Dichloromethane EC328	14408	25	mL	8/19/

Final Volume: 25 mL

Stock Source

DRO200430B O-Terphenyl

Base Units

ug/mL

Amount Added

0.1012 g

Analtes

A O-Terphenyl

CAS

84-15-1

Conc:

ug/mL

4000

Energy Laboratories Inc

Standard LOG

Standard ID: DRO200430B
Standard Name: O-Terphenyl
Date Prepared: 4/30/2020
Date Expires: 9/30/2024
Department: dropr
Vendor: Chemservice
Lot Number: 9972100
Balance ID:
Comments: ID#: 6271

Type: Neat
BY: Ann Nebel
Status: New

Chemical / Solvent Used	BottleNo	Amt	Units	Exp
o-Terphenyl	12650	500	mg	9/30/

Final Volume: mL

Stock Source

Base Units

Amount Added

Analtes

CAS

Conc: ug/mL

A O-Terphenyl

84-15-1

1

660 Tower Lane • P.O. Box 599 • West Chester, PA 19381-0599
1-800-452-9994 • 1-610-692-3026 • Fax 1-610-692-8729
info@chemservice.com • www.chemservice.com

CERTIFICATE OF ANALYSIS

o-Terphenyl

CATALOG NUMBER N-12693-500MG
LOT NUMBER 9972100
DATE CERTIFIED 09/23/19
EXPIRATION DATE 09/30/24
CAS NUMBER 84-15-1
MOLECULAR FORMULA C₁₈H₁₄
MOLECULAR WEIGHT 230.32
STORAGE Store in a cool dry place.
HANDLING See Safety Data Sheet
INTENDED USE For laboratory use only.

Analytical Test	Value
FT-IR SPECTROSCOPY	CONFORMS TO STRUCTURE
GC/MS SPECTRA ID	MATCHES NIST DATABASE
MELTING POINT (°C)	57.1
% PURITY (GC/FID)	99.5

Chem Service, Inc. guarantees the purity to be +/- 0.5% deviation prior to the expiration date shown on the label and exclusive of any customer contamination.

Certified By:

Mary Beth O'Donnell

Mary Beth O'Donnell
CSM/TC

ID #: 12650

Opened: _____

o-Terphenyl

Expires: 9/30/2024


Rec'd: 4/30/2020

Energy Laboratories Inc 1120 So. 27th Street
Billings MT 59107

CERTIFICATE OF ANALYSIS

Analysis Method:

Catalog Number: N-12693-500MG
Description: o-Terphenyl
Lot Number: 9972100
Expiration Date: 09/30/24



Chem Service is accredited to ISO 17034:2016, ISO/IEC 17025:2017 and certified to ISO 9001:2015



660 Tower Lane • P.O. Box 599 • West Chester, PA 19381-0599
1-800-452-9994 • 1-610-692-3026 • Fax 1-610-692-8729
info@chemservice.com • www.chemservice.com

CERTIFICATE OF ANALYSIS

Analysis Method:

Catalog Number: N-12693-500MG
Description: o-Terphenyl
Lot Number: 9972100
Expiration Date: 09/30/24
Chem Service Inc Area Percent Report

Data File: D:\msdchem\2019 DATA\0919\0923-01.D
Acq On : 23 Sep 2019 10:40
Operator :
Sample : n-12693
Misc :
ALS Vial : 95

Integration Parameters: autoint1.e
Integrator: ChemStation

DataAcq Meth: SCREEN.M
Method : D:\msdchem\2019 DATA\0919\0903-09.D\ERIN.M

Signal : TIC: 0923-01.D\data.ms

peak #	R.T. min	first scan	max scan	last scan	PK TY	peak height	corr. area	corr. % max.	% of total
1	11.844	1597	1606	1613	BB	32038221	432253484	100.00%	100.000%

Sum of corrected areas: 432253484

ERIN.M Mon Sep 23 10:55:51 2019


660 Tower Lane • P.O. Box 599 • West Chester, PA 19381-0599
1-800-452-9994 • 1-610-692-3026 • Fax 1-610-692-8729
info@chemservice.com • www.chemservice.com

CERTIFICATE OF ANALYSIS

Analysis Method:

Catalog Number: N-12693-500MG
Description: o-Terphenyl
Lot Number: 9972100
Expiration Date: 09/30/24

Abundance



Chem. Service is accredited to ISO 17034:2016, ISO/IEC 17025:2017 and certified to ISO 9001:2015




660 Tower Lane • P.O. Box 599 • West Chester, PA 19381-0599
1-800-452-9994 • 1-610-692-3026 • Fax 1-610-692-8729
info@chemservice.com • www.chemservice.com

CERTIFICATE OF ANALYSIS

Analysis Method:

Catalog Number: N-12693-500MG
Description: o-Terphenyl
Lot Number: 9972100
Expiration Date: 09/30/24

Abundance



Chem Service is accredited to ISO 17034:2016, ISO/IEC 17025:2017 and certified to ISO 9001:2015



660 Tower Lane • P.O. Box 599 • West Chester, PA 19381-0599
1-800-452-9994 • 1-610-692-3026 • Fax 1-610-692-8729
info@chemservice.com • www.chemservice.com

CERTIFICATE OF ANALYSIS

Analysis Method:

Catalog Number:	N-12693-500MG
Description:	o-Terphenyl
Lot Number:	9972100
Expiration Date:	09/30/24

Chem Service is accredited to ISO 17034:2016, ISO/IEC 17025:2017 and certified to ISO 9001:2015




660 Tower Lane • P.O. Box 599 • West Chester, PA 19381-0599
 1-800-452-9994 • 1-610-692-3026 • Fax 1-610-692-8729
info@chemservice.com • www.chemservice.com

Gas

Data file: C:\CHEM3\
 Sample name: N-12683
 Instrument: GC 2
 Injection date: 9/23/2019 9:58:34 AM
 Acq. method: SCREEN.M
 Column name: HP-5

CERTIFICATE OF ANALYSIS

Sample type:
 Location: Vial 141
 Injection volume: 1.0uL



Signal: FID1 A, Front Signal

RT [min]	Type	Width [min]	Area	Height	Area%
4.565	BB	0.0305	1.2408	0.5122	0.11
20.152	BB	0.0391	1171.9556	439.4599	99.89
		Sum	1173.1963		

Chem Service is accredited to ISO 17034:2016, ISO/IEC 17025:2017 and certified to ISO 9001:2015



Energy Laboratories Inc

Standard LOG

Standard ID: DRO211220C
Standard Name 5,000 ug/mL RRO CCV 200 ug/mL Triaconta Type: Secondary
Date Prepared 12/20/2021 BY: Ann Nebel
Date Expires: 4/6/2026
Department dropr Status: New
Vendor:
Lot Number:
Balance ID: Sartorius 4 place balance
Comments: CCV for AK102 and 8015C RRO.

Chemical / Solvent Used	BottleNo	Amt	Units	Exp
Dichloromethane EC757	14596	2.8	mL	10/20

Final Volume: 4 mL

<u>Stock Source</u>		Base Units	Amount Added
DRO210401B	50,000 ug/mL Oil Std For AK103 RRO-I	ug/mL	400 µL
DRO211129A	Triacontane SURR 1000 ug/mL	ug/mL	800 µL

<u>Analtes</u>		CAS	Conc: ug/mL
A	30/40W Motor Oil		5000
A	Triacontane-d62		200

Energy Laboratories Inc

Standard LOG

Standard ID: DRO210401B
Standard Name: 50,000 ug/mL Oil Std For AK103 RRO-In DC
Date Prepared: 4/1/2021
Date Expires: 1/31/2028
Department: dropr
Vendor: Restek
Lot Number: A0166827
Balance ID: Sartorius 4 place balance

Type: Primary
BY: Ann Nebel
Status: Open

Comments:

Chemical / Solvent Used	BottleNo	Amt	Units	Exp
Residual Range Calibration Standard (13714	1	mL	1/31/

Final Volume: 1 mL

Stock Source

Base Units

Amount Added

Analtes

CAS

Conc: **ug/mL**



CERTIFIED REFERENCE MATERIAL

110 Benner Circle
Bellefonte, PA 16823-8812
Tel: (800)356-1688
Fax: (814)353-1309

www.restek.com

Certificate of Analysis



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31817 **Lot No.:** A0166827

Description : Residual Range Calibration Standard (RCS)
Residual Range Calib Std (RCS) 50,000µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : January 31, 2028 **Storage:** 25°C nominal

Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	Grav. Conc. (weight/volume)	Expanded Uncertainty (95% C.L.; K=2)		
1	Motor Oil SAE30 & SAE40 Blend (Pennzoil) CAS # 64742-65-0.F (Lot A0126386) Purity ----%	50,056.0 µg/mL	+/- 293.0889 µg/mL	+/- 1,490.7309 µg/mL	+/- 1,589.8634 µg/mL
			Gravimetric	Unstressed	Stressed

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

ID #: 13714
Opened: _____
Residual Range Calibration Standard (RCS)
Expires: 1/31/2028
Rec'd: 4/1/2021
Energy Laboratories Inc 1120 So. 27th Street
Billings MT 59107

Column:

30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)

Carrier Gas:

hydrogen-constant pressure 10 psi.

Temp. Program:

40°C (hold 2 min.) to 330°C
@ 10°C/min. (hold 10 min.)

Inj. Temp:


250°C

Det. Temp:

330°C

Det. Type:

FID



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Kylie Struble
Kylie Struble - Operations Technician I

Date Mixed: 02-Dec-2020

Balance: 1128353505

Justin Albertson
Justin Albertson - Operations Tech-ARM QC

Date Passed: 07-Dec-2020

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/ μ ECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified combined stressed uncertainty value (includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ stressed} = k \sqrt{U_{gravimetric}^2 + U_{homogeneity}^2 + U_{storage\ stability}^2 + U_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- It is important to note that the shipping stability uncertainty was obtained under temperature extremes for specific time intervals; therefore, the certified combined stressed uncertainty value should only be applied to the product if it was stored at non-standard temperature conditions up to and including 7 days. Contact Restek Technical Service at www.restek.com/Contact-Us for use recommendations if your shipment was in-transit for more than 7 days at non-standard temperature conditions.
- Apply the certified combined unstressed uncertainty value if the product was received under standard shipping conditions. Apply the certified combined stressed uncertainty value if the product was received under non-standard conditions as specified below.

Label Conditions	Standard Conditions	Non-Standard Conditions
25°C Nominal (Room Temperature)	< 60°C	≥ 60°C up to 7 days
10°C or colder (Refrigerate)	< 40°C	≥ 40°C up to 7 days
0°C or colder (Freezer) -20°C or colder (Deep Freezer)	< 25°C	≥ 25°C up to 7 days

- Separate (not combined) uncertainty values for gravimetric uncertainty are also displayed on the certificate, if needed, separate homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty values are available by contacting Restek Technical Service at www.restek.com/Contact-Us.
- The packaged amount is the minimum sample size for which uncertainty is valid. The ampules are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampules. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.

Energy Laboratories Inc

Standard LOG

Standard ID: DRO211129A
Standard Name: Triacotane SURR 1000 ug/mL
Date Prepared: 11/29/2021
Date Expires: 4/6/2026
Department: dropr
Vendor:
Lot Number:
Balance ID: BAL-DRO
Comments: 2X dilution of Triacotane SURR 2000 ug/mL

Type: Secondary
BY: Jillian L Bostwick
Status: New

Chemical / Solvent Used	BottleNo	Amt	Units	Exp
Dichloromethane EC735	14518	5	mL	10/14

Final Volume: 10 mL

Stock Source
DRO211006A Triacotane SURR 2000 ug/mL

Base Units
ug/mL

Amount Added
5 mL

Analvtes
A Triacotane-d62

CAS

Conc: **ug/mL**
1000

Energy Laboratories Inc

Standard LOG

Standard ID: DRO211006A
Standard Name: Triacontane SURR 2000 ug/mL
Date Prepared: 10/6/2021
Date Expires: 4/6/2026
Department: dropr
Vendor:
Lot Number:
Balance ID: BAL-DRO
Comments: Triacontane SURR 2000 ug/mL

Type: Secondary
BY: Jillian L Bostwick
Status: New

Chemical / Solvent Used	BottleNo	Amt	Units	Exp
Acetone DZ509	13553	50	mL	7/22/

Final Volume: 50 mL

Stock Source
DRO210406A Triacontane-d62 Surr For AK103 RRO

Base Units
ug/mL

Amount Added
0.1001 g

Analtes
A Triacontane-d62

CAS

Conc: **ug/mL**
2000

Energy Laboratories Inc

Standard LOG

Standard ID: DRO210406A
Standard Name: Triacontane-d62 Surr For AK103 RRO
Date Prepared: 4/6/2021
Date Expires: 4/6/2026
Department: dropr
Vendor: Sigma-Aldrich
Lot Number: MBBC4347
Balance ID:
Comments: Alaska surr [for AK103 RRO]

Type: Neat
BY: Ann Nebel
Status: New

Chemical / Solvent Used	BottleNo	Amt	Units	Exp
Triacontane-d62-98 atom % D	13736		mL	4/6/2026

Final Volume: mL

Stock Source

Base Units

Amount Added

Analtes

CAS

Conc: ug/mL

A Triacontane-d62

3050 Spruce Street, Saint Louis, MO 63103, USA

Website: www.sigmaaldrich.com


Email USA: techserv@sial.com

Outside USA: eurtechserv@sial.com

Certificate of Analysis

Product Name:
Triacontane-d62 - 98 atom % D

Product Number: 451789
 Batch Number: MBBC4347
 Brand: ALDRICH
 CAS Number: 93952-07-9
 MDL Number: MFCD00209794
 Formula: C30D62
 Formula Weight: 485.20 g/mol
 Quality Release Date: 27 APR 2018



ID #: 13736

Opened: _____


Triacontane-d62-98 atom % D

Expires: 4/6/2026

Rec'd: 4/6/2021

Energy Laboratories Inc 1120 So. 27th Street
 Billings MT 59107

Test	Specification	Result
Purity (HPLC)	≥ 99.0 %	99.0 %
Proton NMR Spectrum	Conforms to Structure	Conforms
D Enrichment	≥ 98.0 %	99.0 %
Initial Melting Point		60.0 °C
Final Melting Point		62.0 °C



Greg Abernathy, Supervisor
 Quality Control
 Miamisburg, Ohio US

Sigma-Aldrich warrants, that at the time of the quality release or subsequent retest date this product conformed to the information contained in this publication. The current Specification sheet may be available at Sigma-Aldrich.com. For further inquiries, please contact Technical Service. Purchaser must determine the suitability of the product for its particular use. See reverse side of invoice or packing slip for additional terms and conditions of sale.

Energy Laboratories Inc

Standard LOG

Standard ID: DRO211213A
Standard Name: OTP only SURR 2000 ug/mL
Date Prepared: 12/13/2021
Date Expires: 9/30/2024
Department: dropr
Vendor:
Lot Number:
Balance ID: BAL-DRO
Comments: OTP SURR 2000 ug/mL

Type: Secondary
BY: Jillian L Bostwick
Status: New

Chemical / Solvent Used	BottleNo	Amt	Units	Exp
Acetone DZ509	13553	100	mL	7/22/

Final Volume: 100 mL

Stock Source

DRO200430B O-Terphenyl

Base Units

ug/mL

Amount Added

0.2015 g

Analtes

A 1-Chlorooctadecane

CAS

3386-33-2

Conc:

ug/mL

2000

Energy Laboratories Inc

Standard LOG

Standard ID: DRO200430B
Standard Name: O-Terphenyl
Date Prepared: 4/30/2020
Date Expires: 9/30/2024
Department: dropr
Vendor: Chemservice
Lot Number: 9972100
Balance ID:
Comments: ID#: 6271

Type: Neat
BY: Ann Nebel
Status: New

Chemical / Solvent Used	BottleNo	Amt	Units	Exp
o-Terphenyl	12650	500	mg	9/30/

Final Volume: mL

Stock Source

Base Units

Amount Added

Analtes

CAS

Conc: ug/mL

A O-Terphenyl

84-15-1

1

660 Tower Lane • P.O. Box 599 • West Chester, PA 19381-0599
1-800-452-9994 • 1-610-692-3026 • Fax 1-610-692-8729
info@chemservice.com • www.chemservice.com

CERTIFICATE OF ANALYSIS

o-Terphenyl

CATALOG NUMBER N-12693-500MG
LOT NUMBER 9972100
DATE CERTIFIED 09/23/19
EXPIRATION DATE 09/30/24
CAS NUMBER 84-15-1
MOLECULAR FORMULA C₁₈H₁₄
MOLECULAR WEIGHT 230.32
STORAGE Store in a cool dry place.
HANDLING See Safety Data Sheet
INTENDED USE For laboratory use only.

Analytical Test	Value
FT-IR SPECTROSCOPY	CONFORMS TO STRUCTURE
GC/MS SPECTRA ID	MATCHES NIST DATABASE
MELTING POINT (°C)	57.1
% PURITY (GC/FID)	99.5

Chem Service, Inc. guarantees the purity to be +/- 0.5% deviation prior to the expiration date shown on the label and exclusive of any customer contamination.

Certified By:

Mary Beth O'Donnell

Mary Beth O'Donnell
CSM/TC

ID #: 12650

Opened: _____

o-Terphenyl

Expires: 9/30/2024


Rec'd: 4/30/2020

Energy Laboratories Inc 1120 So. 27th Street
Billings MT 59107

CERTIFICATE OF ANALYSIS

Analysis Method:

Catalog Number: N-12693-500MG
Description: o-Terphenyl
Lot Number: 9972100
Expiration Date: 09/30/24



Chem Service is accredited to ISO 17034:2016, ISO/IEC 17025:2017 and certified to ISO 9001:2015



660 Tower Lane • P.O. Box 599 • West Chester, PA 19381-0599
1-800-452-9994 • 1-610-692-3026 • Fax 1-610-692-8729
info@chemservice.com • www.chemservice.com

CERTIFICATE OF ANALYSIS

Analysis Method:

Catalog Number: N-12693-500MG
Description: o-Terphenyl
Lot Number: 9972100
Expiration Date: 09/30/24
Chem Service Inc Area Percent Report

Data File: D:\msdchem\2019 DATA\0919\0923-01.D
Acq On : 23 Sep 2019 10:40
Operator :
Sample : n-12693
Misc :
ALS Vial : 95

Integration Parameters: autoint1.e
Integrator: ChemStation

DataAcq Meth: SCREEN.M
Method : D:\msdchem\2019 DATA\0919\0903-09.D\ERIN.M

Signal : TIC: 0923-01.D\data.ms

peak #	R.T. min	first scan	max scan	last scan	PK TY	peak height	corr. area	corr. % max.	% of total
1	11.844	1597	1606	1613	BB	32038221	432253484	100.00%	100.000%

Sum of corrected areas: 432253484

ERIN.M Mon Sep 23 10:55:51 2019


660 Tower Lane • P.O. Box 599 • West Chester, PA 19381-0599
1-800-452-9994 • 1-610-692-3026 • Fax 1-610-692-8729
info@chemservice.com • www.chemservice.com

CERTIFICATE OF ANALYSIS

Analysis Method:

Catalog Number: N-12693-500MG
Description: o-Terphenyl
Lot Number: 9972100
Expiration Date: 09/30/24

Abundance



Time-->

Chem. Service is accredited to ISO 17034:2016, ISO/IEC 17025:2017 and certified to ISO 9001:2015




660 Tower Lane • P.O. Box 599 • West Chester, PA 19381-0599
1-800-452-9994 • 1-610-692-3026 • Fax 1-610-692-8729
info@chemservice.com • www.chemservice.com

CERTIFICATE OF ANALYSIS

Analysis Method:

Catalog Number: N-12693-500MG
Description: o-Terphenyl
Lot Number: 9972100
Expiration Date: 09/30/24

Abundance



Chem Service is accredited to ISO 17034:2016, ISO/IEC 17025:2017 and certified to ISO 9001:2015



660 Tower Lane • P.O. Box 599 • West Chester, PA 19381-0599
1-800-452-9994 • 1-610-692-3026 • Fax 1-610-692-8729
info@chemservice.com • www.chemservice.com

CERTIFICATE OF ANALYSIS

Analysis Method:

Catalog Number:	N-12693-500MG
Description:	o-Terphenyl
Lot Number:	9972100
Expiration Date:	09/30/24

Chem Service is accredited to ISO 17034:2016, ISO/IEC 17025:2017 and certified to ISO 9001:2015




660 Tower Lane • P.O. Box 599 • West Chester, PA 19381-0599
 1-800-452-9994 • 1-610-692-3026 • Fax 1-610-692-8729
info@chemservice.com • www.chemservice.com

Gas

Data file: C:\CHEM3\
 Sample name: N-12683
 Instrument: GC 2
 Injection date: 9/23/2019 9:58:34 AM
 Acq. method: SCREEN.M
 Column name: HP-5

CERTIFICATE OF ANALYSIS

Sample type:
 Location: Vial 141
 Injection volume: 1.0uL



Signal: FID1 A, Front Signal

RT [min]	Type	Width [min]	Area	Height	Area%
4.565	BB	0.0305	1.2408	0.5122	0.11
20.152	BB	0.0391	1171.9556	439.4599	99.89
		Sum	1173.1963		

Chem Service is accredited to ISO 17034:2016, ISO/IEC 17025:2017 and certified to ISO 9001:2015



Energy Laboratories Inc

Standard LOG

Standard ID: DRO210902A
 Standard Name: 50,000 ug/mL Oil Std for RRO-In DCM
 Date Prepared: 9/2/2021
 Date Expires: 9/1/2026
 Department: dropr
 Vendor:
 Lot Number:
 Balance ID: BAL-DRO
 Comments: .625 g of 30W and 40 W each LCS for Oil range

Type: Secondary
 BY: Jillian L Bostwick
 Status: New

Chemical / Solvent Used	BottleNo	Amt	Units	Exp
Dichloromethane EB867	14196	25	mL	6/18/

Final Volume: 25 mL

<u>Stock Source</u>	<u>Base Units</u>	<u>Amount Added</u>
DRO210901B 40W Motor Oil-Valvoline	ug/mL	0.6261 g
DRO210901A 30W Motor Oil-Valvoline	ug/mL	0.6254 g

<u>Analtes</u>	<u>CAS</u>	<u>Conc:</u>	<u>ug/mL</u>
A 30W Motor Oil			10000
A 30W-Motor oil			0
A 40W Motor Oil			10000
A 40W-Motor oil			0

Energy Laboratories Inc

Standard LOG

Standard ID: DRO210901B
Standard Name: 40W Motor Oil-Valvoline
Date Prepared: 9/1/2021
Date Expires: 9/1/2026
Department: dropr
Vendor:
Lot Number:
Balance ID:
Type: Primary
BY: Jillian L Bostwick
Status: New
Comments: Used to Make 2nd Source Standards For Alaska AK103 RRO Method and Oil

Chemical / Solvent Used	BottleNo	Amt	Units	Exp
Valvoline SAE 40 Motor Oil	14231		mL	9/1/2026

Final Volume: mL

Stock Source

Base Units

Amount Added

Analtes

CAS

Conc: ug/mL

A 40W-Motor oil

1

Energy Laboratories Inc

Standard LOG

Standard ID: DRO210901A
Standard Name: 30W Motor Oil-Valvoline
Date Prepared: 9/1/2021
Date Expires: 9/1/2026
Department: dropr
Vendor:
Lot Number:
Balance ID:
Type: Primary
BY: Jillian L Bostwick
Status: New
Comments: Used to make 2nd Source Standard for AK103 method.

Chemical / Solvent Used	BottleNo	Amt	Units	Exp
Valvoline SAE 30 Motor Oil	14232		mL	9/1/26

Final Volume: mL

Stock Source

Base Units

Amount Added

Analtes

CAS

Conc: ug/mL

A 30W-Motor oil

1

Energy Laboratories Inc

Standard LOG

Standard ID: DRO211129A
Standard Name: Triacotane SURR 1000 ug/mL
Date Prepared: 11/29/2021
Date Expires: 4/6/2026
Department: dropr
Vendor:
Lot Number:
Balance ID: BAL-DRO
Comments: 2X dilution of Triacotane SURR 2000 ug/mL

Type: Secondary
BY: Jillian L Bostwick
Status: New

Chemical / Solvent Used	BottleNo	Amt	Units	Exp
Dichloromethane EC735	14518	5	mL	10/14

Final Volume: 10 mL

Stock Source
DRO211006A Triacotane SURR 2000 ug/mL

Base Units
ug/mL

Amount Added
5 mL

Analtes
A Triacotane-d62

CAS

Conc: ug/mL
1000

Energy Laboratories Inc

Standard LOG

Standard ID: DRO211006A
Standard Name: Triacontane SURR 2000 ug/mL
Date Prepared: 10/6/2021
Date Expires: 4/6/2026
Department: dropr
Vendor:
Lot Number:
Balance ID: BAL-DRO
Comments: Triacontane SURR 2000 ug/mL

Type: Secondary
BY: Jillian L Bostwick
Status: New

Chemical / Solvent Used	BottleNo	Amt	Units	Exp
Acetone DZ509	13553	50	mL	7/22/

Final Volume: 50 mL

Stock Source
DRO210406A Triacontane-d62 Surr For AK103 RRO

Base Units
ug/mL

Amount Added
0.1001 g

Analtes
A Triacontane-d62

CAS

Conc: **ug/mL**
2000

Energy Laboratories Inc

Standard LOG

Standard ID: DRO210406A
Standard Name: Triacontane-d62 Surr For AK103 RRO
Date Prepared: 4/6/2021
Date Expires: 4/6/2026
Department: dropr
Vendor: Sigma-Aldrich
Lot Number: MBBC4347
Balance ID:
Comments: Alaska surr [for AK103 RRO]

Type: Neat
BY: Ann Nebel
Status: New

Chemical / Solvent Used	BottleNo	Amt	Units	Exp
Triacontane-d62-98 atom % D	13736		mL	4/6/2026

Final Volume: mL

Stock Source

Base Units

Amount Added

Analtes

CAS

Conc: ug/mL

A Triacontane-d62

3050 Spruce Street, Saint Louis, MO 63103, USA

Website: www.sigmaaldrich.com

Email USA: techserv@sial.com

Outside USA: eurtechserv@sial.com

Certificate of Analysis

Product Name:
Triacontane-d62 - 98 atom % D

Product Number: 451789
 Batch Number: MBBC4347
 Brand: ALDRICH
 CAS Number: 93952-07-9
 MDL Number: MFCD00209794
 Formula: C30D62
 Formula Weight: 485.20 g/mol
 Quality Release Date: 27 APR 2018



ID #: 13736

Opened: _____


Triacontane-d62-98 atom % D

Expires: 4/6/2026

Rec'd: 4/6/2021

Energy Laboratories Inc 1120 So. 27th Street
 Billings MT 59107

Test	Specification	Result
Purity (HPLC)	≥ 99.0 %	99.0 %
Proton NMR Spectrum	Conforms to Structure	Conforms
D Enrichment	≥ 98.0 %	99.0 %
Initial Melting Point		60.0 °C
Final Melting Point		62.0 °C



Greg Abernathy, Supervisor
 Quality Control
 Miamisburg, Ohio US

Sigma-Aldrich warrants, that at the time of the quality release or subsequent retest date this product conformed to the information contained in this publication. The current Specification sheet may be available at Sigma-Aldrich.com. For further inquiries, please contact Technical Service. Purchaser must determine the suitability of the product for its particular use. See reverse side of invoice or packing slip for additional terms and conditions of sale.