

Energy Laboratories Inc

ANALYTICAL RUN Summary

13-Jan-22

Run ID GCFID-HP5-B_220111A

Run Start Date: 1/11/2022
Analyst: Ann Nebel
Ical:
Column ID:
Comments: ICAL- SW8015C_DRO220111JA.CAL

Std ID	Std Name	Std Amount	Std Units	Samp Amount	Samp Units	SampType	Expiration Date
DRO211012B	#2 Diesel in Acetone 150,000 ug/mL					ICV	11/5/2023
DRO211101A	OTP-4000 ug/mL DCM					OTP-CAL	9/30/2024
DRO211214C	Diesel Fuel #2 50,000 ug/mL in DCM					CCV-CAL	4/30/2023
DRO220102D	ALASKA MARKER-200ug/mL					MARKER	5/31/2022

Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist				
14976981	CCV_0111HP50	HC-8015-DRO-	CCV		1/11/2022 8:59:2	1	R373149		0	0					
Analyte		T Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD Q
Total Extractable Hydrocarbons		A mg/L		3.205893		15	0	0	0.0749	0.3	50	21%	80	120	0% S
o-Terphenyl		S mg/L		0.1968894		0.2	0	0	0.000429	0.002	0	98%	80	120	0%

Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist				
14976982	CCV_0111HP50	HC-8015-DRO-	CAL1		1/11/2022 10:25:	1	R373149		0	0					
Analyte		T Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD Q
o-Terphenyl		S mg/L		0.00201677		0.002	0	0	0.000429	0.002	0	101%	80	120	0%

Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist				
14976983	CCV_0111HP50	HC-8015-DRO-	CAL2		1/11/2022 11:08:	1	R373149		0	0					
Analyte		T Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD Q
o-Terphenyl		S mg/L		0.0489019		0.05	0	0	0.000429	0.002	0	98%	80	120	0%

Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist					
14976984	CCV_0111HP50	HC-8015-DRO-	CAL3		1/11/2022 11:51:	1	R373149		0	0						
Analyte	T	Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD	Q
o-Terphenyl	S	mg/L		0.2047389		0.2	0	0	0.000429	0.002	0	102%	80	120	0%	
Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist					
14976985	CCV_0111HP50	HC-8015-DRO-	CAL4		1/11/2022 12:34:	1	R373149		0	0						
Analyte	T	Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD	Q
o-Terphenyl	S	mg/L		0.4884362		0.5	0	0	0.000429	0.002	0	98%	80	120	0%	
Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist					
14976986	CCV_0111HP50	HC-8015-DRO-	CAL5		1/11/2022 1:17:0	1	R373149		0	0						
Analyte	T	Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD	Q
o-Terphenyl	S	mg/L		1.013008		1	0	0	0.000429	0.002	0	101%	80	120	0%	
Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist					
14976987	CCV_0111HP50	HC-8015-DRO-	CAL1		1/11/2022 1:59:5	1	R373149		0	0						
Analyte	T	Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD	Q
Total Extractable Hydrocarbons	A	mg/L		0.1635249		0.15	0	0	0.0749	0.3	50	109%	80	120	0%	
Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist					
14976989	CCV_0111HP51	HC-8015-DRO-	CAL2		1/11/2022 2:42:3	1	R373149		0	0						
Analyte	T	Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD	Q
Total Extractable Hydrocarbons	A	mg/L		3.698293		3.75	0	0	0.0749	0.3	50	99%	80	120	0%	
Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist					
14976990	CCV_0111HP51	HC-8015-DRO-	CAL3		1/11/2022 3:25:2	1	R373149		0	0						
Analyte	T	Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD	Q
Total Extractable Hydrocarbons	A	mg/L		14.75864		15	0	0	0.0749	0.3	50	98%	80	120	0%	

Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist					
14976991	CCV_0111HP51	HC-8015-DRO-	CAL4		1/11/2022 4:08:0	1	R373149		0	0						
Analyte	T	Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD	Q
Total Extractable Hydrocarbons	A	mg/L		36.29137		37.5	0	0	0.0749	0.3	50	97%	80	120	0%	
Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist					
14976992	CCV_0111HP51	HC-8015-DRO-	CAL5		1/11/2022 4:51:0	1	R373149		0	0						
Analyte	T	Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD	Q
Total Extractable Hydrocarbons	A	mg/L		48.59718		50	0	0	0.0749	0.3	50	97%	80	120	0%	
Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist					
14976993	CCV_0111HP51	HC-8015-DRO-	ICV		1/11/2022 5:34:2	1	R373149		0	0						
Analyte	T	Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD	Q
Total Extractable Hydrocarbons	A	mg/L		14.05379		15	0	0	0.0749	0.3	50	94%	80	120	0%	

Write Sequence	Data File	Sample Name	Method	Weight	Dil Factor	Amt Inj.	IS	Cal ID
	G:\org\HP5\DAT\HP5011122_b\0111HP5.01r	DCM-Baseline Check-V01	G:\Org\HP5\Methods\DR_8015-IC-LEXP.met	1	1	1	1	0
	G:\org\HP5\DAT\HP5011122_b\0111HP5.02r	CCV_0111HP502r, DRO ;0111HP5 , DRO220102D	G:\Org\HP5\Methods\DC_8015-JA-L%.met	1	1	1	1	0
	G:\org\HP5\DAT\HP5011122_b\0111HP5.03r	DCM-Baseline Check-V03	G:\Org\HP5\Methods\DR_8015-IC-LEXP.met	1	1	1	1	0
	G:\org\HP5\DAT\HP5011122_b\0111HP5.04r	CCV_0111HP504r, CAL1 ;0111HP5 , 2 ug per mL OTP (10 uL of Cal3 + 990 uL DCM(14647)	G:\Org\HP5\Methods\DS_8015-JA-L#.met	1	1	1	1	0
	G:\org\HP5\DAT\HP5011122_b\0111HP5.05r	CCV_0111HP505r, CAL2 ;0111HP5 , 50 ug per mL OTP (100 uL Cal4 + 900 uL of DCM(14647)	G:\Org\HP5\Methods\DS_8015-JA-L#.met	1	1	1	1	0
	G:\org\HP5\DAT\HP5011122_b\0111HP5.06r	CCV_0111HP506r, CAL3 ;0111HP5 , 200 ug per mL OTP (100uL of Cal5 + 400 uL DCM(14647)	G:\Org\HP5\Methods\DS_8015-JA-L#.met	1	1	1	1	0
	G:\org\HP5\DAT\HP5011122_b\0111HP5.07r	CCV_0111HP507r, CAL4 ;0111HP5 , 500 ug per mL OTP (250uL of Cal5 + 250 uL DCM(14647)	G:\Org\HP5\Methods\DS_8015-JA-L#.met	1	1	1	1	0
	G:\org\HP5\DAT\HP5011122_b\0111HP5.08r	CCV_0111HP508r, CAL5 ;0111HP5 , 1000 ug per mL OTP (250 uL 4000 ug/mL OTP DRO211101A + 750 DCM(14647)	G:\Org\HP5\Methods\DS_8015-JA-L#.met	1	1	1	1	0
	G:\org\HP5\DAT\HP5011122_b\0111HP5.09r	CCV_0111HP509r, CAL1 ;0111HP5 , 150 ug per mL Diesel (20 uL of Cal3 + 980 uL DCM(14647), then 100 uL of that + 100 uL of DCM (14647)	G:\Org\HP5\Methods\DC_8015-JA-L%.met	1	1	1	1	0
	G:\org\HP5\DAT\HP5011122_b\0111HP5.10r	CCV_0111HP510r, CAL2 ;0111HP5 , 3750 ug per mL Diesel (100 uL Cal4 + 900 uL of DCM(14647)	G:\Org\HP5\Methods\DC_8015-JA-L%.met	1	1	1	1	0
	G:\org\HP5\DAT\HP5011122_b\0111HP5.11r	CCV_0111HP511r, CAL3 ;0111HP5 , 15000 ug per mL Diesel (300 uL of DRO211214C + 700 uL DCM(14647)	G:\Org\HP5\Methods\DC_8015-JA-L%.met	1	1	1	1	0
	G:\org\HP5\DAT\HP5011122_b\0111HP5.12r	CCV_0111HP512r, CAL4 ;0111HP5 , 37500ug per mL Diesel (750 uL of DRO211214C + 250 uL DCM(14647)	G:\Org\HP5\Methods\DC_8015-JA-L%.met	1	1	1	1	0
	G:\org\HP5\DAT\HP5011122_b\0111HP5.13r	CCV_0111HP513r, CAL5 ;0111HP5 , 50000 ug per mL Diesel (200 uL of DRO211214C)	G:\Org\HP5\Methods\DC_8015-JA-L%.met	1	1	1	1	0
	G:\org\HP5\DAT\HP5011122_b\0111HP5.14r	CCV_0111HP514r, Second Source ;0111HP5 , 15000 ug per mL (100uL of DRO211012B + 900uL DCM(14647)	G:\Org\HP5\Methods\DC_8015-JA-L%.met	1	1	1	1	0

File Name: G:\Org\HP5\Cals\SW8015C_DRO220111JA.CAL

Version: 12

Creator: AMN 01/13/2022

Description: 8015C-DRO. New ICal Per 0111HP5 (2022)-2 uL Inj.; COD added using OTP RFs

Reason for change:

External standard calibration

Standard injection volume: 1

Standard sample weight: 1

Area reject threshold: 500

Reference peak area reject threshold: 500

Amount units: nanograms


No default component

Method of calculating data point averages: Equal weight for all updates

No calibration update report

All levels are normal data points.

1 DRO Range Start



Expected retention time: 6.68 minutes
 Search window: 0.05 minutes
 No retention time reference component
 Group number: 0
 High alarm limit: 0
 Low alarm limit: 0
 Component constant: 0


Single peak quantification by area

$Y = 32675.36 X + 0$

Average CF fit with equal weighting, forced to origin
 Coefficient of determination: 0.9980255
 Average error: 3.607%
 Average CF: 32675.36
 RSD: 5.100%

Level	Amount	Response	Cal Factor	Error, %	Source	Date and time
1	150	5343235	35621.57	9.017	Manual	1/13/2022 12:28:36 PM
2	3750	1.20843E+08	32224.8	-1.379	Manual	1/13/2022 12:29:11 PM
3	15000	4.82244E+08	32149.6	-1.609	Manual	1/13/2022 12:29:24 PM
4	37500	1.185834E+09	31622.24	-3.223	Manual	1/13/2022 12:29:37 PM
5	50000	1.58793E+09	31758.6	-2.806	Manual	1/13/2022 12:28:57 PM


2 *o-Terphenyl



Expected retention time: 12.35 minutes
 Search window: 0.05 minutes
 No retention time reference component
 Group number: 0
 High alarm limit: 0
 Low alarm limit: 0
 Component constant: 0
 Single peak quantification by area
 $Y = 36857.86 X + 0$
 Average CF fit with equal weighting, forced to origin
 Coefficient of determination: 0.9995278
 Average error: 1.804%
 Average CF: 36857.86
 RSD: 2.132%

Level	Amount	Response	Cal Factor	Error, %	Source	Date and time
1	2	74333.97	37166.98	0.839	G:\Org\HP5\DAT\HP5011122_b\0111HP5.0004.BND	1/13/2022 12:27:15 PM
2	50	1802420	36048.4	-2.196	G:\Org\HP5\DAT\HP5011122_b\0111HP5.0005.BND	1/13/2022 12:27:23 PM
3	200	7546240	37731.2	2.369	G:\Org\HP5\DAT\HP5011122_b\0111HP5.0006.BND	1/13/2022 12:27:28 PM
4	500	1.800271E+07	36005.42	-2.313	G:\Org\HP5\DAT\HP5011122_b\0111HP5.0007.BND	1/13/2022 12:27:34 PM
5	1000	3.733731E+07	37337.31	1.301	G:\Org\HP5\DAT\HP5011122_b\0111HP5.0008.BND	1/13/2022 12:27:40 PM

3 *1-Chlorooctadecane



Expected retention time: 13.16 minutes
 Search window: 0.05 minutes
 No retention time reference component
 Group number: 0
 High alarm limit: 0
 Low alarm limit: 0
 Component constant: 0

Single peak quantification by area

$Y = 36857.86 X + 0$

Average CF fit with equal weighting, forced to origin
 Coefficient of determination: 0.9995278
 Average error: 1.804%
 Average CF: 36857.86
 RSD: 2.132%

Level	Amount	Response	Cal Factor	Error, %	Source	Date and time
1	2	74333.97	37166.98	0.839	Manual	1/13/2022 12:27:45 PM
2	50	1802420	36048.4	-2.196	Manual	1/13/2022 12:27:47 PM
3	200	7546240	37731.2	2.369	Manual	1/13/2022 12:27:49 PM
4	500	1.800271E+07	36005.42	-2.313	Manual	1/13/2022 12:27:51 PM
5	1000	3.733731E+07	37337.31	1.301	Manual	1/13/2022 12:27:53 PM

Write Sequence	Data File	Sample Name	Method	Weight	Dil Factor	Amt Inj.	IS	Cal ID	Manual Integrations
		DCM-Baseline Check-V01	G:\Org\HP5\Methods\DR_8015-IC-LEXP.met	1	1	1	1	0	No Integration
		CCV_0111HP502r, DRO ;0111HP5 , DRO220102D	G:\Org\HP5\Methods\DC_8015-JA-L%.met	1	1	1	1	0	No Integration
		DCM-Baseline Check-V03	G:\Org\HP5\Methods\DR_8015-IC-LEXP.met	1	1	1	1	0	No Integration
		CCV_0111HP504r, CAL1 ;0111HP5 , 2 ug per mL OTP (10 uL of Cal3 + 990 uL DCM(14647)	G:\Org\HP5\Methods\DS_8015-JA-L#.met	1	1	1	1	0	Surrogates are integrated using a valley to valley integration Set Baseline All Valley on at 12.01 minutes.
		CCV_0111HP505r, CAL2 ;0111HP5 , 50 ug per mL OTP (100 uL Cal4 + 900 uL of DCM(14647)	G:\Org\HP5\Methods\DS_8015-JA-L#.met	1	1	1	1	0	Surrogates are integrated using a valley to valley integration Set Baseline All Valley on at 12.01 minutes.
		CCV_0111HP506r, CAL3 ;0111HP5 , 200 ug per mL OTP (100uL of Cal5 + 400 uL DCM(14647)	G:\Org\HP5\Methods\DS_8015-JA-L#.met	1	1	1	1	0	Surrogates are integrated using a valley to valley integration Set Baseline All Valley on at 12.01 minutes.
		CCV_0111HP507r, CAL4 ;0111HP5 , 500 ug per mL OTP (250uL of Cal5 + 250 uL DCM(14647)	G:\Org\HP5\Methods\DS_8015-JA-L#.met	1	1	1	1	0	Surrogates are integrated using a valley to valley integration Set Baseline All Valley on at 12.01 minutes.
		CCV_0111HP508r, CAL5 ;0111HP5 , 1000 ug per mL OTP (250 uL 4000 ug/mL OTP DRO211101A + 750 DCM(14647)	G:\Org\HP5\Methods\DS_8015-JA-L#.met	1	1	1	1	0	Surrogates are integrated using a valley to valley integration Set Baseline All Valley on at 12.01 minutes.
		CCV_0111HP509r, CAL1 ;0111HP5 , 150 ug per mL Diesel (20 uL of Cal3 + 980 uL DCM(14647), then 100 uL of that + 100 uL of DCM (14647))	G:\Org\HP5\Methods\DC_8015-JA-L%.met	1	1	1	1	0	The integration of Diesel Range Organics and Total Extractable Hydrocarbons is the hydrocarbon response with reference to the baseline. Assigned Set Baseline on All Valley on at 16.36 minutes.
		CCV_0111HP510r, CAL2 ;0111HP5 , 3750 ug per mL Diesel (100 uL Cal4 + 900 uL of DCM(14647)	G:\Org\HP5\Methods\DC_8015-JA-L%.met	1	1	1	1	0	The integration of Diesel Range Organics and Total Extractable Hydrocarbons is the hydrocarbon response with reference to the baseline. Assigned Set Baseline on All Valley on at 16.36 minutes.
		CCV_0111HP511r, CAL3 ;0111HP5 , 15000 ug per mL Diesel (300 uL of DRO211214C + 700 uL DCM(14647)	G:\Org\HP5\Methods\DC_8015-JA-L%.met	1	1	1	1	0	The integration of Diesel Range Organics and Total Extractable Hydrocarbons is the hydrocarbon response with reference to the baseline. Assigned Set Baseline on All Valley on at 16.36 minutes.
		CCV_0111HP512r, CAL4 ;0111HP5 , 37500ug per mL Diesel (750 uL of DRO211214C + 250 uL DCM(14647)	G:\Org\HP5\Methods\DC_8015-JA-L%.met	1	1	1	1	0	The integration of Diesel Range Organics and Total Extractable Hydrocarbons is the hydrocarbon response with reference to the baseline. Assigned Set Baseline on All Valley on at 16.36 minutes.
		CCV_0111HP513r, CAL5 ;0111HP5 , 50000 ug per mL Diesel (200 uL of DRO211214C)	G:\Org\HP5\Methods\DC_8015-JA-L%.met	1	1	1	1	0	The integration of Diesel Range Organics and Total Extractable Hydrocarbons is the hydrocarbon response with reference to the baseline. Assigned Set Baseline on All Valley on at 16.36 minutes.
		CCV_0111HP514r, Second Source ;0111HP5 , 15000 ug per mL (100uL of DRO211012B + 900uL DCM(14647)	G:\Org\HP5\Methods\DC_8015-JA-L%.met	1	1	1	1	0	The integration of Diesel Range Organics and Total Extractable Hydrocarbons is the hydrocarbon response with reference to the baseline. Assigned Set Baseline on All Valley on at 16.36 minutes.

Ann Nebel

Digitally signed by
Ann Nebel
Date: 2022.02.11 10:29:19 -07:00

Write Sequence	Data File	Sample Name	Method	Weight	Dil Factor	Amt Inj.	IS	Cal ID	Manual Integrations
		DCM-Baseline Check-V01	G:\Org\HP5-Methods\DR_8015-IC-LEXP.met	1	1	1	1	0	No Integration
		CCV_0111HP502r, DRO_0111HP5, DRO220102D	G:\Org\HP5-Methods\DC_8015-JA-L%.met	1	1	1	1	0	No Integration
		DCM-Baseline Check-V03	G:\Org\HP5-Methods\DR_8015-IC-LEXP.met	1	1	1	1	0	No Integration
		CCV_0111HP504r, CAL1_0111HP5, 2 ug per mL OTP (10 uL of Cal3 + 990 uL DCM(14647))	G:\Org\HP5-Methods\DS_8015-JA-L#.met	1	1	1	1	0	Surrogates are integrated using a valley to valley integration Set Baseline All Valley on at 12.01 minutes.
		CCV_0111HP505r, CAL2_0111HP5, 50 ug per mL OTP (100 uL Cal4 + 900 uL DCM(14647))	G:\Org\HP5-Methods\DS_8015-JA-L#.met	1	1	1	1	0	Surrogates are integrated using a valley to valley integration Set Baseline All Valley on at 12.01 minutes.
		CCV_0111HP506r, CAL3_0111HP5, 200 ug per mL OTP (100uL of Cal5 + 400 uL DCM(14647))	G:\Org\HP5-Methods\DS_8015-JA-L#.met	1	1	1	1	0	Surrogates are integrated using a valley to valley integration Set Baseline All Valley on at 12.01 minutes.
		CCV_0111HP507r, CAL4_0111HP5, 500 ug per mL OTP (250uL of Cal5 + 250 uL DCM(14647))	G:\Org\HP5-Methods\DS_8015-JA-L#.met	1	1	1	1	0	Surrogates are integrated using a valley to valley integration Set Baseline All Valley on at 12.01 minutes.
		CCV_0111HP508r, CAL5_0111HP5, 1000 ug per mL OTP (250 uL 4000 ug/mL OTP DRO211101A + 750 DCM(14647))	G:\Org\HP5-Methods\DS_8015-JA-L#.met	1	1	1	1	0	Surrogates are integrated using a valley to valley integration Set Baseline All Valley on at 12.01 minutes.
		CCV_0111HP509r, CAL1_0111HP5, 150 ug per mL Diesel (20 uL of Cal3 + 980 uL DCM(14647)), then 100 uL of that + 100 uL of DCM (14647))	G:\Org\HP5-Methods\DC_8015-JA-L%.met	1	1	1	1	0	The integration of Diesel Range Organics and Total Extractable Hydrocarbons is the hydrocarbon response with reference to the baseline. Assigned Set Baseline on All Valley on at 16.36 minutes.
		CCV_0111HP510r, CAL2_0111HP5, 3750 ug per mL Diesel (100 uL Cal4 + 900 uL of DCM(14647))	G:\Org\HP5-Methods\DC_8015-JA-L%.met	1	1	1	1	0	The integration of Diesel Range Organics and Total Extractable Hydrocarbons is the hydrocarbon response with reference to the baseline. Assigned Set Baseline on All Valley on at 16.36 minutes.
		CCV_0111HP511r, CAL3_0111HP5, 15000 ug per mL Diesel (300 uL of DRO211214C + 700 uL DCM(14647))	G:\Org\HP5-Methods\DC_8015-JA-L%.met	1	1	1	1	0	The integration of Diesel Range Organics and Total Extractable Hydrocarbons is the hydrocarbon response with reference to the baseline. Assigned Set Baseline on All Valley on at 16.36 minutes.
		CCV_0111HP512r, CAL4_0111HP5, 37500ug per mL Diesel (750 uL of DRO211214C + 250 uL DCM(14647))	G:\Org\HP5-Methods\DC_8015-JA-L%.met	1	1	1	1	0	The integration of Diesel Range Organics and Total Extractable Hydrocarbons is the hydrocarbon response with reference to the baseline. Assigned Set Baseline on All Valley on at 16.36 minutes.
		CCV_0111HP513r, CAL5_0111HP5, 50000 ug per mL Diesel (200 uL of DRO211214C)	G:\Org\HP5-Methods\DC_8015-JA-L%.met	1	1	1	1	0	The integration of Diesel Range Organics and Total Extractable Hydrocarbons is the hydrocarbon response with reference to the baseline. Assigned Set Baseline on All Valley on at 16.36 minutes.
		CCV_0111HP514r, Second Source_0111HP5, 15000 ug per mL (100uL of DRO211012B + 900uL DCM(14647))	G:\Org\HP5-Methods\DC_8015-JA-L%.met	1	1	1	1	0	The integration of Diesel Range Organics and Total Extractable Hydrocarbons is the hydrocarbon response with reference to the baseline. Assigned Set Baseline on All Valley on at 16.36 minutes.

Ann Nebel

Digitally signed by
Ann Nebel
Date: 2022.02.11 10:29:19 -07:00

Energy Laboratories Inc

ANALYTICAL RUN Summary

14-Jan-22

Run ID GCFID-HP5-B_220111C

Run Start Date: 1/11/2022
Analyst: Ann Nebel
Ical:
Column ID:
Comments: ICAL- SW8015C_ORO220111BA.CAL with Triacontane

Std ID	Std Name	Std Amount	Std Units	Samp Amount	Samp Units	SampType	Expiration Date
DRO210902A	50,000 ug/mL Oil Std for RRO-In DCM					ICV	9/1/2026
DRO211006A	Triacontane SURR 2000 ug/mL					CAL-SURR	4/6/2026
DRO211118A	50,000 ug/mL Oil Std For AK103 RRO-In DCM					CAL-ORO	10/31/2028

Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist					
14977288	CCV_0111HP52	HC-8015-DRO-	CAL1		1/12/2022 3:39:1	1	R373160		0	0						
Analyte	T	Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD	Q
n-Triacontane	S	mg/L		0.00190245		0.002	0	0	0.000336	0.002	0	95%	80	120	0%	

Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist					
14977289	CCV_0111HP52	HC-8015-DRO-	CAL2		1/12/2022 4:22:1	1	R373160		0	0						
Analyte	T	Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD	Q
n-Triacontane	S	mg/L		0.04984459		0.05	0	0	0.000336	0.002	0	100%	80	120	0%	

Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist					
14977290	CCV_0111HP53	HC-8015-DRO-	CAL3		1/12/2022 5:05:2	1	R373160		0	0						
Analyte	T	Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD	Q
n-Triacontane	S	mg/L		0.2024053		0.2	0	0	0.000336	0.002	0	101%	80	120	0%	

Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist					
14977291	CCV_0111HP53	HC-8015-DRO-	CAL4		1/12/2022 5:48:3	1	R373160		0	0						
Analyte	T	Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD	Q
n-Triacontane	S	mg/L		0.5035697		0.5	0	0	0.000336	0.002	0	101%	80	120	0%	
Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist					
14977292	CCV_0111HP55	HC-8015-DRO-	CAL5		1/12/2022 8:49:5	1	R373160		0	0						
Analyte	T	Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD	Q
n-Triacontane	S	mg/L		1.032718		1	0	0	0.000336	0.002	0	103%	80	120	0%	
Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist					
14977293	CCV_0111HP55	HC-8015-DRO-	CAL1		1/13/2022 3:06:1	1	R373160		0	0						
Analyte	T	Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD	Q
TEH(Oil Range)	A	mg/L		0.15954587		0.15	0	0	0.0879	0.3	0	106%	80	120	0%	
Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist					
14977294	CCV_0111HP55	HC-8015-DRO-	CAL2		1/13/2022 4:31:3	1	R373160		0	0						
Analyte	T	Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD	Q
TEH(Oil Range)	A	mg/L		1.03294141		1	0	0	0.0879	0.3	0	103%	80	120	0%	
Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist					
14977295	CCV_0111HP55	HC-8015-DRO-	CAL3		1/13/2022 5:57:4	1	R373160		0	0						
Analyte	T	Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD	Q
TEH(Oil Range)	A	mg/L		4.9326875		5	0	0	0.0879	0.3	0	99%	80	120	0%	
Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist					
14977296	CCV_0111HP56	HC-8015-DRO-	CAL4		1/13/2022 7:24:1	1	R373160		0	0						
Analyte	T	Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD	Q
TEH(Oil Range)	A	mg/L		14.328667		15	0	0	0.0879	0.3	0	96%	80	120	0%	

Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist					
14977297	CCV_0111HP56	HC-8015-DRO-	CAL5		1/13/2022 8:50:3	1	R373160		0	0						
Analyte	T	Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD	Q
TEH(Oil Range)	A	mg/L		28.7914395		30	0	0	0.0879	0.3	0	96%	80	120	0%	

Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist					
14977298	CCV_0111HP56	HC-8015-DRO-	ICV		1/14/2022 8:18:1	0	R373160		0	0						
Analyte	T	Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD	Q
TEH(Oil Range)	A	mg/L		5.07699902		5	0	0	0	0.3	0	102%	80	120	0%	

Write Sequence	Data File	Sample Name	Method	Weight	Dil Factor	Amt Inj.	IS	Cal ID
	G:\org\HP5\DAT\HP5011122 b\0111HP5.25r	DCM-Baseline Check-V25	G:\Org\HP5\Methods\DR_8015-IC-LEXP.met	1	1	1	1	0
	G:\org\HP5\DAT\HP5011122 b\0111HP5.26r	Marker_0111HP526r, DRO :0111HP5 , DRO220111A	G:\org\HP5\Methods\CSC210212.met	1	1	1	1	0
	G:\org\HP5\DAT\HP5011122 b\0111HP5.27r	DCM-Baseline Check-V27	G:\Org\HP5\Methods\DR_8015-HS-LEXP.met	1	1	1	1	0
	G:\org\HP5\DAT\HP5011122 b\0111HP5.28r	CCV_0111HP528r, CAL1 :0111HP5 , 2 ug per mL Triacotane (10 uL of Cal3 + 990 uL DCM(14647)	G:\Org\HP5\Methods\DS_ORO-BA-L#.MET	1	1	1	1	0
	G:\org\HP5\DAT\HP5011122 b\0111HP5.29r	CCV_0111HP529r, CAL2 :0111HP5 , 50 ug per mL Triacotane (100 uL Cal4 + 900 uL of DCM(14647)	G:\Org\HP5\Methods\DS_ORO-BA-L#.MET	1	1	1	1	0
	G:\org\HP5\DAT\HP5011122 b\0111HP5.30r	CCV_0111HP530r, CAL3 :0111HP5 , 200 ug per mL Triacotane (100uL of Cal5 + 400 uL DCM(14647)	G:\Org\HP5\Methods\DS_ORO-BA-L#.MET	1	1	1	1	0
	G:\org\HP5\DAT\HP5011122 b\0111HP5.31r	CCV_0111HP531r, CAL4 :0111HP5 , 500 ug per mL Triacotane (250uL of Cal5 + 250 uL DCM(14647)	G:\Org\HP5\Methods\DS_ORO-BA-L#.MET	1	1	1	1	0
	G:\org\HP5\DAT\HP5011122 b\0111HP5.32r	DCM-Baseline Check-V32	G:\Org\HP5\Methods\DR_8015-HE-LEXP.met	1	1	1	1	0
	G:\org\HP5\DAT\HP5011122 b\0111HP5.50r	CCV_0111HP550r, CAL5 :0111HP5 , 1000 ug per mL Triacotane (DRO211006A)	G:\Org\HP5\Methods\DS_ORO-BA-L#.MET	1	1	1	1	0
	G:\org\HP5\DAT\HP5011122 b\0111HP5.51r	DCM-Baseline Check-V51	G:\Org\HP5\Methods\DR_8015-IC-LEXP.met	1	1	1	1	0
	G:\org\HP5\DAT\HP5011122 b\0111HP5.52r	DCM-Baseline Check-V52	G:\Org\HP5\Methods\DR_8015-IC-LEXP.met	1	1	1	1	0
	G:\org\HP5\DAT\HP5011122 b\0111HP5.53r	Marker_0111HP553r, DRO :0111HP5 , DRO220111A	G:\org\HP5\Methods\CSC210212.met	1	1	1	1	0
	G:\org\HP5\DAT\HP5011122 b\0111HP5.54r	DCM-Baseline Check-V54	G:\Org\HP5\Methods\DR_8015-IC-LEXP.met	1	1	1	1	0
	G:\org\HP5\DAT\HP5011122 b\0111HP5.55r	CCV_0111HP555r, CAL1 :0111HP5 , 150 ug per mL Oil (10 uL of Cal4 + 990 uL DCM(14647)	G:\Org\HP5\Methods\DC_ORO-55-BA-L%.xls	1	1	1	1	0
	G:\org\HP5\DAT\HP5011122 b\0111HP5.56r	DCM-Baseline Check-V56	G:\Org\HP5\Methods\DR_8015-HE-LEXP.met	1	1	1	1	0
	G:\org\HP5\DAT\HP5011122 b\0111HP5.57r	CCV_0111HP557r, CAL2 :0111HP5 , 1000 ug per mL Oil (200 uL of Cal 3 +800 uL DCM(14647)	G:\Org\HP5\Methods\DC_ORO-57-BA-L%.xls	1	1	1	1	0
	G:\org\HP5\DAT\HP5011122 b\0111HP5.58r	DCM-Baseline Check-V58	G:\Org\HP5\Methods\DR_8015-HE-LEXP.met	1	1	1	1	0
	G:\org\HP5\DAT\HP5011122 b\0111HP5.59r	CCV_0111HP559r, CAL3 :0111HP5 , 5000 ug per mL Oil (100 uL of DRO211118A + 900 uL DCM(14647)	G:\Org\HP5\Methods\DC_ORO-59-BA-L%.xls	1	1	1	1	0
	G:\org\HP5\DAT\HP5011122 b\0111HP5.60r	DCM-Baseline Check-V60	G:\Org\HP5\Methods\DR_8015-HE-LEXP.met	1	1	1	1	0
	G:\org\HP5\DAT\HP5011122 b\0111HP5.61r	CCV_0111HP561r, CAL4 :0111HP5 , 15000 ug per mL Oil (200 uL of CAL5 + 200 uL DCM(14647)	G:\Org\HP5\Methods\DC_ORO-61-BA-L%.xls	1	1	1	1	0
	G:\org\HP5\DAT\HP5011122 b\0111HP5.62r	DCM-Baseline Check-V62	G:\Org\HP5\Methods\DR_8015-HE-LEXP.met	1	1	1	1	0
	G:\org\HP5\DAT\HP5011122 b\0111HP5.63r	CCV_0111HP563r, CAL5 :0111HP5 , 30000 ug per mL Oil (600 uL of DRO211118A + 400 uL of DCM)	G:\Org\HP5\Methods\DC_ORO-BA-L%.xls	1	1	1	1	0
	G:\org\HP5\DAT\HP5011122 b\0111HP5.64r	DCM-Baseline Check-V64	G:\Org\HP5\Methods\DR_8015-HE-LEXP.met	1	1	1	1	0
	G:\org\HP5\DAT\HP5011122 b\0111HP5.65r	DCM-Baseline Check-V65	G:\Org\HP5\Methods\DR_8015-HE-LEXP.met	1	1	1	1	0
	G:\org\HP5\DAT\HP5011122 b\0111HP5.66r	DCM-Baseline Check-V66	G:\Org\HP5\Methods\DR_8015-HE-LEXP.met	1	1	1	1	0
	G:\org\HP5\DAT\HP5011122 b\0111HP5.68r	DCM-Baseline Check-V68	G:\Org\HP5\Methods\DR_8015-IC-LEXP.met	1	1	1	1	0
	G:\org\HP5\DAT\HP5011122 b\0111HP5.69r	CCV_0111HP567r, Second Source :0111HP5 , 5000 ug per mL (100uL of DRO210902A + 900uL DCM(14647)	G:\Org\HP5\Methods\DC_ORO-59-BA-L%.xls	1	1	1	1	0

File Name: G:\Org\HP5\Cals\SW8015C_ORO220111BA.CAL
Version: 11

Creator: AMN
Description: 8015C-Oil Range with Triacontane. New ICal Per 0111HP5,(2022)-2 uL Inj.;
Reason for change:


External standard calibration

Standard injection volume: 1
Standard sample weight: 1
Area reject threshold: 500
Reference peak area reject threshold: 500
Amount units: nanograms
No default component

Method of calculating data point averages: Equal weight for all updates
No calibration update report

All levels are normal data points.

1 *30-40 Motor Oil



Expected retention time: 6.4 minutes
 Search window: 0.05 minutes
 No retention time reference component
 Group number: 0
 High alarm limit: 0
 Low alarm limit: 0
 Component constant: 0


Single peak quantification by area

$Y = 26424.55 X + 0$

Average CF fit with equal weighting, forced to origin
 Coefficient of determination: 0.9969108
 Average error: 3.495%
 Average CF: 26424.55
 RSD: 4.293%

Level	Amount	Response	Cal Factor	Error, %	Source	Date and time
1	150	4177025	27846.83	5.382	Manual	1/14/2022 7:51:42 AM
2	1000	2.73111E+07	27311.1	3.355	Manual	1/14/2022 8:05:40 AM
3	5000	1.313247E+08	26264.94	-0.604	Manual	1/14/2022 8:05:24 AM
4	15000	3.796282E+08	25308.55	-4.223	Manual	1/14/2022 8:05:07 AM
5	30000	7.617404E+08	25391.35	-3.910	Manual	1/14/2022 8:04:35 AM


2 #C20



Expected retention time: 12.56 minutes
 Search window: 0.05 minutes
 No retention time reference component
 Group number: 0
 High alarm limit: 0
 Low alarm limit: 0
 Component constant: 0
 Single peak quantification by area
 $Y = 26424.55 X + 0$
 Average CF fit with equal weighting, forced to origin
 Coefficient of determination: 0.9969108
 Average error: 3.495%
 Average CF: 26424.55
 RSD: 4.293%


Level	Amount	Response	Cal Factor	Error, %	Source	Date and time
1	150	4177025	27846.83	5.382	Manual	1/14/2022 8:06:03 AM
2	1000	2.73111E+07	27311.1	3.355	Manual	1/14/2022 8:06:05 AM
3	5000	1.313247E+08	26264.94	-0.604	Manual	1/14/2022 8:06:06 AM
4	15000	3.796282E+08	25308.55	-4.223	Manual	1/14/2022 8:06:11 AM
5	30000	7.617404E+08	25391.35	-3.910	Manual	1/14/2022 8:06:13 AM

3 *#Triacontane



Expected retention time: 16.44 minutes
 Search window: 0.05 minutes
 No retention time reference component
 Group number: 0
 High alarm limit: 0
 Low alarm limit: 0
 Component constant: 0
 Single peak quantification by area
 $Y = 29636.1 X + 0$
 Average CF fit with equal weighting, forced to origin
 Coefficient of determination: 0.9984925
 Average error: 2.075%
 Average CF: 29636.1
 RSD: 3.023%

Level	Amount	Response	Cal Factor	Error, %	Source	Date and time
1	2	56381.2	28190.6	-4.878	Manual	1/13/2022 12:38:47 PM
2	50	1477199	29543.98	-0.311	Manual	1/13/2022 12:38:50 PM
3	200	5998503	29992.52	1.203	Manual	1/13/2022 12:38:53 PM
4	500	1.492384E+07	29847.68	0.714	Manual	1/13/2022 12:38:56 PM
5	1000	3.060573E+07	30605.73	3.272	Manual	1/13/2022 12:39:03 PM




DIESEL RANGE ORGANICS CHROMATOGRAM REPORT


Sample Name: DCM-Baseline Check-V25
 Raw File: G:\org\HP5\DAT\HP5011122_b\0111HP5.0025.RAW
 Date & Time Acquired: 1/12/2022 1:29:46 AM
 Method File: G:\Org\HP5\Methods\DR_8015-IC-LEXP.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO211102IC.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for TEH: 31353.19
 Rt range for Diesel Range Organics: 6.5 to 15.82

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*o-Terphenyl	29.899	200.	.	-
*1-Chlorooctadecane	12.975	200.	.017	.01 -

DRO Area:132028.6 DRO Amount: 4.211011
 TEH Area:186308.4 TEH Amount: 5.942247





DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: DCM-Baseline Check-V27
 Raw File: G:\org\HP5\DAT\HP5011122_b\0111HP5.0027.RAW
 Date & Time Acquired: 1/12/2022 2:56:04 AM
 Method File: G:\Org\HP5\Methods\DR_8015-HS-LEXP.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO210108Hs.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1


Mean RF for TEH: 29457.33
 Rt range for Diesel Range Organics: 6.51 to 15.85

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*o-Terphenyl	12.261	200.	.017	.01
*1-Chlorooctadecane	29.983	200.	.	.

DRO Area:193795.7 DRO Amount: 6.578862
 TEH Area:272770 TEH Amount: 9.259835

G:\Org\HP5\DAT\HP5011122_b\0111HP5.0028.RAW

CCV_0111HP528r, CAL1 ;0111HP5 , 2 ug per mL Triacontane



RESIDUAL RANGE ORGANICS CHROMATOGRAM

Sample Name: CCV_0111HP528r, CAL1 ;0111HP5 , 2 ug per mL Triacontane
 Raw File: G:\Org\HP5\DAT\HP5011122_b\0111HP5.0028.RAW
 Date & Time Acquired: 1/12/2022 3:39:11 AM
 Method File: G:\Org\HP5\Methods\DS_ORO-BA-L%.MET
 Calibration File: G:\Org\HP5\Cals\SW8015C_ORO220111ba.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for for Residual Range Organics Calculations: 28542.41
 Rt range for Residual Range Organics: 12.51 to 30.05

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*#Triacontane	16.421	500.	1.902	.38

RRO Area:11465.21 RRO AMOUNT: 0.4016902


CONTINUING CALIBRATION REPORT: G:\Org\HP5\DAT\HP5011122_b\0111HP5.0028.RAW

COMPOUND	ACTUAL (NG)	MEASURED (NG)	%RECOVERY	LIMITS
*30-40 Motor Oil	5000.	.056	.	75-125

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	LIMITS
*#Triacontane	16.421	200.	1.902	.95	75-125

G:\org\HP5\DAT\HP5011122_b\0111HP5.0029.RAW

CCV_0111HP529r, CAL2 ;0111HP5 , 50 ug per mL Triacontane



RESIDUAL RANGE ORGANICS CHROMATOGRAM

Sample Name: CCV_0111HP529r, CAL2 ;0111HP5 , 50 ug per mL Triacontane
 Raw File: G:\org\HP5\DAT\HP5011122_b\0111HP5.0029.RAW
 Date & Time Acquired: 1/12/2022 4:22:15 AM
 Method File: G:\Org\HP5\Methods\DS_ORO-BA-L%.MET
 Calibration File: G:\Org\HP5\Cals\SW8015C_ORO220111ba.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for for Residual Range Organics Calculations: 28542.41
 Rt range for Residual Range Organics: 12.51 to 30.05


SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*#Triacontane	16.423	500.	49.845	9.97

RRO Area:60154.51 RRO AMOUNT: 2.107548

CONTINUING CALIBRATION REPORT: G:\org\HP5\DAT\HP5011122_b\0111HP5.0029.RAW

COMPOUND	ACTUAL (NG)	MEASURED (NG)	%RECOVERY	LIMITS
*30-40 Motor Oil	5000.	.023	.	75-125

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	LIMITS
*#Triacontane	16.423	200.	49.845	24.92	75-125



RESIDUAL RANGE ORGANICS CHROMATOGRAM

Sample Name: CCV_0111HP530r, CAL3 ;0111HP5 , 200 ug per mL Triacontane
 Raw File: G:\org\HP5\DAT\HP5011122_b\0111HP5.0030.RAW
 Date & Time Acquired: 1/12/2022 5:05:25 AM
 Method File: G:\Org\HP5\Methods\DS_ORO-BA-L%.MET
 Calibration File: G:\Org\HP5\Cals\SW8015C_ORO220111ba.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for for Residual Range Organics Calculations: 28542.41
 Rt range for Residual Range Organics: 12.51 to 30.05

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*#Triacontane	16.427	500.	202.405	40.48	-

RRO Area:200104.8 RRO AMOUNT: 7.01079


CONTINUING CALIBRATION REPORT: G:\org\HP5\DAT\HP5011122_b\0111HP5.0030.RAW

COMPOUND	ACTUAL (NG)	MEASURED (NG)	%RECOVERY	LIMITS
*30-40 Motor Oil	5000.	.	75-125	

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	LIMITS
*#Triacontane	16.427	200.	202.405	101.2	75-125

G:\org\HP5\DAT\HP5011122_b\0111HP5.0031.RAW

CCV_0111HP531r, CAL4 ;0111HP5 , 500 ug per mL Triacontane



RESIDUAL RANGE ORGANICS CHROMATOGRAM

Sample Name: CCV_0111HP531r, CAL4 ;0111HP5 , 500 ug per mL Triacontane
 Raw File: G:\org\HP5\DAT\HP5011122_b\0111HP5.0031.RAW
 Date & Time Acquired: 1/12/2022 5:48:34 AM
 Method File: G:\Org\HP5\Methods\DS_ORO-BA-L%.MET
 Calibration File: G:\Org\HP5\Cals\SW8015C_ORO220111ba.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for for Residual Range Organics Calculations: 28542.41
 Rt range for Residual Range Organics: 12.51 to 30.05


SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*#Triacontane	16.434	500.	503.57	100.71	-

RRO Area:497882.9 RRO AMOUNT: 17.44362

CONTINUING CALIBRATION REPORT: G:\org\HP5\DAT\HP5011122_b\0111HP5.0031.RAW

COMPOUND	ACTUAL (NG)	MEASURED (NG)	%RECOVERY	LIMITS
*30-40 Motor Oil	5000.	.	75-125	

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	LIMITS
*#Triacontane	16.434	200.	503.57	251.78	75-125



RESIDUAL RANGE ORGANICS CHROMATOGRAM

Sample Name: CCV_0111HP550r, CAL5 ;0111HP5 , 1000 ug per mL Triacontane
 Raw File: G:\org\HP5\DAT\HP5011122_b\0111HP5.0050.RAW
 Date & Time Acquired: 1/12/2022 8:49:58 PM
 Method File: G:\Org\HP5\Methods\DS_ORO-BA-L%.MET
 Calibration File: G:\Org\HP5\Cals\SW8015C_ORO220111ba.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1


Mean RF for for Residual Range Organics Calculations: 28542.41
 Rt range for Residual Range Organics: 12.51 to 30.05

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*#Triacontane	16.444	500.	1032.718	206.54

RRO Area:993904.8 RRO AMOUNT: 34.82203

CONTINUING CALIBRATION REPORT: G:\org\HP5\DAT\HP5011122_b\0111HP5.0050.RAW
 COMPOUND ACTUAL (NG) MEASURED (NG) %RECOVERY LIMITS
 *30-40 Motor Oil 5000. . . 75-125

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	LIMITS
*#Triacontane	16.444	200.	1032.718	516.36	75-125




DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: DCM-Baseline Check-V51
 Raw File: G:\org\HP5\DAT\HP5011122_b\0111HP5.0051.RAW
 Date & Time Acquired: 1/13/2022 12:15:29 PM
 Method File: G:\Org\HP5\Methods\DR_8015-IC-LEXP.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO211102IC.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for TEH: 31353.19
 Rt range for Diesel Range Organics: 6.5 to 15.82

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*o-Terphenyl	12.166	200.	.369	.18	-
*1-Chlorooctadecane	12.994	200.	.464	.23	-

DRO Area:587062.5 DRO Amount: 18.72417
 TEH Area:891448.4 TEH Amount: 28.43246




DIESEL RANGE ORGANICS CHROMATOGRAM REPORT


Sample Name: DCM-Baseline Check-V52
 Raw File: G:\org\HP5\DAT\HP5011122_b\0111HP5.0052.RAW
 Date & Time Acquired: 1/13/2022 12:58:31 PM
 Method File: G:\Org\HP5\Methods\DR_8015-IC-LEXP.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO211102IC.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for TEH: 31353.19
 Rt range for Diesel Range Organics: 6.5 to 15.82

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*o-Terphenyl	12.192	200.	.855	.43	-
*1-Chlorooctadecane	13.007	200.	.955	.48	-

DRO Area:2710300 DRO Amount: 86.44414
 TEH Area:2842315 TEH Amount: 90.65472






DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: DCM-Baseline Check-V54
 Raw File: G:\org\HP5\DAT\HP5011122_b\0111HP5.0054.RAW
 Date & Time Acquired: 1/13/2022 2:23:42 PM
 Method File: G:\Org\HP5\Methods\DR_8015-IC-LEXP.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO211102IC.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for TEH: 31353.19
 Rt range for Diesel Range Organics: 6.5 to 15.82

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*o-Terphenyl	29.882	200.	.	-
*1-Chlorooctadecane	29.882	200.	.	-

DRO Area:44798.44 DRO Amount: 1.428832
 TEH Area:97771.24 TEH Amount: 3.118382



RESIDUAL RANGE ORGANICS CHROMATOGRAM

Sample Name: CCV_0111HP555r, CAL1 ;0111HP5 , 150 ug per mL Oil
 Raw File: G:\org\HP5\DAT\HP5011122_b\0111HP5.0055.RAW
 Date & Time Acquired: 1/13/2022 3:06:11 PM
 Method File: G:\Org\HP5\Methods\DC_ORO-55-BA-L%.MET
 Calibration File: G:\Org\HP5\Cals\SW8015C_ORO220111BA.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for for Residual Range Organics Calculations: 26424.55
 Rt range for Residual Range Organics: 12.51 to 30.05


SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*#Triacontane	16.447	500.	.47	.09	-

RRO Area: 4215928 RRO AMOUNT: 159.5459

CONTINUING CALIBRATION REPORT: G:\org\HP5\DAT\HP5011122_b\0111HP5.0055.RAW

COMPOUND	ACTUAL (NG)	MEASURED (NG)	%RECOVERY	LIMITS
*30-40 Motor Oil	5000.	.	.	75-125

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	LIMITS
*#Triacontane	16.447	200.	.47	.23	75-125



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT


Sample Name: DCM-Baseline Check-V56
 Raw File: G:\org\HP5\DAT\HP5011122_b\0111HP5.0056.RAW
 Date & Time Acquired: 1/13/2022 3:48:53 PM
 Method File: G:\Org\HP5\Methods\DR_8015-HE-LEXP.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO210108HE.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for TEH: 29457.33

Rt range for Diesel Range Organics: 6.49 to 15.75

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*o-Terphenyl	12.162	200.	.406	.2
*1-Chlorooctadecane	29.946	200.	.	.

DRO Area:2125703 DRO Amount: 72.16209
 TEH Area:2146824 TEH Amount: 72.8791



RESIDUAL RANGE ORGANICS CHROMATOGRAM

Sample Name: CCV_0111HP557r, CAL2 ;0111HP5 , 1000 ug per mL Oil
 Raw File: G:\org\HP5\DAT\HP5011122_b\0111HP5.0057.RAW
 Date & Time Acquired: 1/13/2022 4:31:31 PM
 Method File: G:\Org\HP5\Methods\DC_ORO-57-BA-L\MET
 Calibration File: G:\Org\HP5\Cals\SW8015C_ORO220111BA.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for for Residual Range Organics Calculations: 26424.55
 Rt range for Residual Range Organics: 12.51 to 30.05


SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*#Triacontane	16.454	500.	3.058	.61	-

RRO Area: 2.729502E+07 RRO AMOUNT: 1032.941

CONTINUING CALIBRATION REPORT: G:\org\HP5\DAT\HP5011122_b\0111HP5.0057.RAW

COMPOUND	ACTUAL (NG)	MEASURED (NG)	%RECOVERY	LIMITS
*30-40 Motor Oil	5000.	.	75-125	

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	LIMITS
*#Triacontane	16.454	200.	3.058	1.53	75-125



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT


Sample Name: DCM-Baseline Check-V58
 Raw File: G:\org\HP5\DAT\HP5011122_b\0111HP5.0058.RAW
 Date & Time Acquired: 1/13/2022 5:14:45 PM
 Method File: G:\Org\HP5\Methods\DR_8015-HE-LEXP.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO210108HE.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for TEH: 29457.33

Rt range for Diesel Range Organics: 6.49 to 15.75

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*o-Terphenyl	29.94	200.	.	-
*1-Chlorooctadecane	29.94	200.	.	-

DRO Area:48306.73 DRO Amount: 1.639888
 TEH Area:141285.5 TEH Amount: 4.796276



RESIDUAL RANGE ORGANICS CHROMATOGRAM

Sample Name: CCV_0111HP559r, CAL3 ;0111HP5 , 5000 ug per mL Oil
 Raw File: G:\org\HP5\DAT\HP5011122_b\0111HP5.0059.RAW
 Date & Time Acquired: 1/13/2022 5:57:48 PM
 Method File: G:\Org\HP5\Methods\DC_ORO-59-BA-L%.MET
 Calibration File: G:\Org\HP5\Cals\SW8015C_ORO220111BA.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for for Residual Range Organics Calculations: 26424.55
 Rt range for Residual Range Organics: 12.51 to 30.05


SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*#Triacontane	16.442	500.	33.09	6.62	-

RRO Area:1.303441E+08 RRO AMOUNT: 4932.688

CONTINUING CALIBRATION REPORT: G:\org\HP5\DAT\HP5011122_b\0111HP5.0059.RAW

COMPOUND	ACTUAL (NG)	MEASURED (NG)	%RECOVERY	LIMITS
*30-40 Motor Oil	5000.	.037	.	75-125

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	LIMITS
*#Triacontane	16.442	200.	33.09	16.54	75-125



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT


Sample Name: DCM-Baseline Check-V60
 Raw File: G:\org\HP5\DAT\HP5011122_b\0111HP5.0060.RAW
 Date & Time Acquired: 1/13/2022 6:41:03 PM
 Method File: G:\Org\HP5\Methods\DR_8015-HE-LEXP.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO210108HE.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for TEH: 29457.33

Rt range for Diesel Range Organics: 6.49 to 15.75

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*o-Terphenyl	29.911	200.	.	-
*1-Chlorooctadecane	29.911	200.	.	-

DRO Area:316779.5 DRO Amount: 10.75384
 TEH Area:980005.5 TEH Amount: 33.26864



RESIDUAL RANGE ORGANICS CHROMATOGRAM

Sample Name: CCV_0111HP561r, CAL4 ;0111HP5 , 15000 ug per mL Oil
 Raw File: G:\org\HP5\DAT\HP5011122_b\0111HP5.0061.RAW
 Date & Time Acquired: 1/13/2022 7:24:16 PM
 Method File: G:\Org\HP5\Methods\DC_ORO-61-BA-L%.MET
 Calibration File: G:\Org\HP5\Cals\SW8015C_ORO220111BA.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for for Residual Range Organics Calculations: 26424.55
 Rt range for Residual Range Organics: 12.51 to 30.05


SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*#Triacontane	16.429	500.	33.728	6.75	-

RRO Area:3.786286E+08 RRO AMOUNT: 14328.67

CONTINUING CALIBRATION REPORT: G:\org\HP5\DAT\HP5011122_b\0111HP5.0061.RAW

COMPOUND	ACTUAL (NG)	MEASURED (NG)	%RECOVERY	LIMITS
*30-40 Motor Oil	5000.	.086	.	75-125

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	LIMITS
*#Triacontane	16.429	200.	33.728	16.86	75-125



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT


Sample Name: DCM-Baseline Check-V62
 Raw File: G:\org\HP5\DAT\HP5011122_b\0111HP5.0062.RAW
 Date & Time Acquired: 1/13/2022 8:07:28 PM
 Method File: G:\Org\HP5\Methods\DR_8015-HE-LEXP.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO210108HE.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for TEH: 29457.33

Rt range for Diesel Range Organics: 6.49 to 15.75

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*o-Terphenyl	29.883	200.	.	.
*1-Chlorooctadecane	29.883	200.	.	.

DRO Area:289041.4 DRO Amount: 9.812207
 TEH Area:1408450 TEH Amount: 47.81323



RESIDUAL RANGE ORGANICS CHROMATOGRAM

Sample Name: CCV_0111HP563r, CAL5 ;0111HP5 , 30000 ug per mL Oil
 Raw File: G:\org\HP5\DAT\HP5011122_b\0111HP5.0063.RAW
 Date & Time Acquired: 1/13/2022 8:50:32 PM
 Method File: G:\Org\HP5\Methods\DC_ORO-BA-L%.MET
 Calibration File: G:\Org\HP5\Cals\SW8015C_ORO220111BA.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for for Residual Range Organics Calculations: 26424.55
 Rt range for Residual Range Organics: 12.51 to 30.05


SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*#Triacontane	16.44	500.	102.625	20.52	-

RRO Area: 7.608009E+08 RRO AMOUNT: 28791.44

CONTINUING CALIBRATION REPORT: G:\org\HP5\DAT\HP5011122_b\0111HP5.0063.RAW

COMPOUND	ACTUAL (NG)	MEASURED (NG)	%RECOVERY	LIMITS
*30-40 Motor Oil	5000.	.102	.	75-125

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	LIMITS
*#Triacontane	16.44	200.	102.625	51.31	75-125




DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: DCM-Baseline Check-V64
 Raw File: G:\org\HP5\DAT\HP5011122_b\0111HP5.0064.RAW
 Date & Time Acquired: 1/13/2022 9:33:32 PM
 Method File: G:\Org\HP5\Methods\DR_8015-HE-LEXP.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO210108HE.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for TEH: 29457.33
 Rt range for Diesel Range Organics: 6.49 to 15.75

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*o-Terphenyl	29.741	200.	.	-
*1-Chlorooctadecane	29.741	200.	.	-

DRO Area:282237.4 DRO Amount: 9.581227
 TEH Area:2669631 TEH Amount: 90.62704



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT


Sample Name: DCM-Baseline Check-V65
 Raw File: G:\org\HP5\DAT\HP5011122_b\0111HP5.0065.RAW
 Date & Time Acquired: 1/13/2022 10:16:33 PM
 Method File: G:\Org\HP5\Methods\DR_8015-HE-LEXP.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO210108HE.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for TEH: 29457.33

Rt range for Diesel Range Organics: 6.49 to 15.75

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*o-Terphenyl	12.129	200.	.021	.01
*1-Chlorooctadecane	29.884	200.	.	.

DRO Area:397141.5 DRO Amount: 13.48192
 TEH Area:1310457 TEH Amount: 44.48662




DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: DCM-Baseline Check-V66
 Raw File: G:\org\HP5\DAT\HP5011122_b\0111HP5.0066.RAW
 Date & Time Acquired: 1/13/2022 10:59:39 PM
 Method File: G:\Org\HP5\Methods\DR_8015-HE-LEXP.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO210108HE.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for TEH: 29457.33
 Rt range for Diesel Range Organics: 6.49 to 15.75

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*o-Terphenyl	12.12	200.	.025	.01
*1-Chlorooctadecane	12.922	200.	.037	.02

DRO Area: 278500.4 DRO Amount: 9.454367
 TEH Area: 757930.2 TEH Amount: 25.72976




DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: DCM-Baseline Check-V68
 Raw File: G:\org\HP5\DAT\HP5011122_b\0111HP5.0068.RAW
 Date & Time Acquired: 1/14/2022 7:35:26 AM
 Method File: G:\Org\HP5\Methods\DR_8015-IC-LEXP.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO211102IC.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for TEH: 31353.19
 Rt range for Diesel Range Organics: 6.5 to 15.82

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*o-Terphenyl	29.833	200.	.	-
*1-Chlorooctadecane	29.833	200.	.	-

DRO Area:178261.1 DRO Amount: 5.685582
 TEH Area:1513925 TEH Amount: 48.28614



RESIDUAL RANGE ORGANICS CHROMATOGRAM

Sample Name: CCV_0111HP567r, Second Source ;0111HP5 , 5000 ug per mL (100
 Raw File: G:\Org\HP5\DAT\HP5011122_b\0111HP5.0069.RAW
 Date & Time Acquired: 1/14/2022 8:18:14 AM
 Method File: G:\Org\HP5\Methods\DC_ORO-59-BA-L%.MET
 Calibration File: G:\Org\HP5\Cals\SW8015C_ORO220111BA.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for for Residual Range Organics Calculations: 26424.55
 Rt range for Residual Range Organics: 12.51 to 30.05

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*#Triacontane_____	16.408	500.	23.958	4.79	-

RRO Area:1.341574E+08 RRO AMOUNT: 5076.999

CONTINUING CALIBRATION REPORT: G:\Org\HP5\DAT\HP5011122_b\0111HP5.0069.RAW

COMPOUND	ACTUAL (NG)	MEASURED (NG)	%RECOVERY	LIMITS
*30-40 Motor Oil_____	5000.	.033	.	75-125

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	LIMITS
*#Triacontane_____	16.408	200.	23.958	11.98	75-125

Write Sequence	Data File	Sample Name	Method	Weight	Dil Factor	Amt Inj.	IS	Cal ID	Manual Integrations
	G:\org\HP5\DAT\HP5011122_b\0111HP5.25r	DCM-Baseline Check-V25	G:\Org\HP5\Methods\DR_8015-IC-LEXP.met	1	1	1	1	0	No Integrations
	G:\org\HP5\DAT\HP5011122_b\0111HP5.26r	Marker_0111HP526r_DRO :0111HP5 , DRO220111A	G:\org\HP5\Methods\CSC210212.met	1	1	1	1	0	No Integrations
	G:\org\HP5\DAT\HP5011122_b\0111HP5.27r	DCM-Baseline Check-V27	G:\Org\HP5\Methods\DR_8015-HS-LEXP.met	1	1	1	1	0	No Integrations
	G:\org\HP5\DAT\HP5011122_b\0111HP5.28r	CCV_0111HP528r, CAL1 :0111HP5 , 2 ug per mL Triacotane (10 uL of Cal3 + 990 uL DCM(14647)	G:\Org\HP5\Methods\DS_ORO-BA-L#.MET	1	1	1	1	0	Surrogates are integrated using a valley to valley integration Set Baseline All Valley on at 16.04 minutes.
	G:\org\HP5\DAT\HP5011122_b\0111HP5.29r	CCV_0111HP529r, CAL2 :0111HP5 , 50 ug per mL Triacotane (100 uL Cal4 + 900 uL of DCM(14647)	G:\Org\HP5\Methods\DS_ORO-BA-L#.MET	1	1	1	1	0	Surrogates are integrated using a valley to valley integration Set Baseline All Valley on at 16.04 minutes.
	G:\org\HP5\DAT\HP5011122_b\0111HP5.30r	CCV_0111HP530r, CAL3 :0111HP5 , 200 ug per mL Triacotane (100uL of Cal5 + 400 uL DCM(14647)	G:\Org\HP5\Methods\DS_ORO-BA-L#.MET	1	1	1	1	0	Surrogates are integrated using a valley to valley integration Set Baseline All Valley on at 16.04 minutes.
	G:\org\HP5\DAT\HP5011122_b\0111HP5.31r	CCV_0111HP531r, CAL4 :0111HP5 , 500 ug per mL Triacotane (250uL of Cal5 + 250 uL DCM(14647)	G:\Org\HP5\Methods\DS_ORO-BA-L#.MET	1	1	1	1	0	Surrogates are integrated using a valley to valley integration Set Baseline All Valley on at 16.04 minutes.
	G:\org\HP5\DAT\HP5011122_b\0111HP5.32r	DCM-Baseline Check-V33	G:\Org\HP5\Methods\DR_8015-HE-LEXP.met	1	1	1	1	0	No Integrations
	G:\org\HP5\DAT\HP5011122_b\0111HP5.50r	CCV_0111HP550r, CAL5 :0111HP5 , 1000 ug per mL Triacotane (DRO211006A)	G:\Org\HP5\Methods\DS_ORO-BA-L#.MET	1	1	1	1	0	Surrogates are integrated using a valley to valley integration Set Baseline All Valley on at 16.04 minutes.
	G:\org\HP5\DAT\HP5011122_b\0111HP5.51r	DCM-Baseline Check-V51	G:\Org\HP5\Methods\DR_8015-IC-LEXP.met	1	1	1	1	0	No Integrations
	G:\org\HP5\DAT\HP5011122_b\0111HP5.52r	DCM-Baseline Check-V52	G:\Org\HP5\Methods\DR_8015-IC-LEXP.met	1	1	1	1	0	No Integrations
	G:\org\HP5\DAT\HP5011122_b\0111HP5.53r	Marker_0111HP553r_DRO :0111HP5 , DRO220111A	G:\org\HP5\Methods\CSC210212.met	1	1	1	1	0	No Integrations
	G:\org\HP5\DAT\HP5011122_b\0111HP5.54r	DCM-Baseline Check-V54	G:\Org\HP5\Methods\DR_8015-IC-LEXP.met	1	1	1	1	0	No Integrations
	G:\org\HP5\DAT\HP5011122_b\0111HP5.55r	CCV_0111HP555r, CAL1 :0111HP5 , 150 ug per mL Oil (10 uL of Cal4 + 990 uL DCM(14647)	G:\Org\HP5\Methods\DC_ORO-55-BA-L%.xls	1	1	1	1	0	The integration of TEH(Oil Range)is the hydrocarbon response with reference to the baseline. Assigned Set Baseline Now at 25 minutes. Y-Scale adjusted.
	G:\org\HP5\DAT\HP5011122_b\0111HP5.56r	DCM-Baseline Check-V56	G:\Org\HP5\Methods\DR_8015-HE-LEXP.met	1	1	1	1	0	No Integrations
	G:\org\HP5\DAT\HP5011122_b\0111HP5.57r	CCV_0111HP557r, CAL2 :0111HP5 , 1000 ug per mL Oil (200 uL of Cal 3 +800 uL DCM(14647)	G:\Org\HP5\Methods\DC_ORO-57-BA-L%.xls	1	1	1	1	0	The integration of TEH(Oil Range)is the hydrocarbon response with reference to the baseline. Y-Scale adjusted.
	G:\org\HP5\DAT\HP5011122_b\0111HP5.58r	DCM-Baseline Check-V58	G:\Org\HP5\Methods\DR_8015-HE-LEXP.met	1	1	1	1	0	No Integrations
	G:\org\HP5\DAT\HP5011122_b\0111HP5.59r	CCV_0111HP559r, CAL3 :0111HP5 , 5000 ug per mL Oil (100 uL of DRO211118A + 900 uL DCM(14647)	G:\Org\HP5\Methods\DC_ORO-59-BA-L%.xls	1	1	1	1	0	The integration of TEH(Oil Range)is the hydrocarbon response with reference to the baseline. Y-Scale adjusted.
	G:\org\HP5\DAT\HP5011122_b\0111HP5.60r	DCM-Baseline Check-V60	G:\Org\HP5\Methods\DR_8015-HE-LEXP.met	1	1	1	1	0	No Integrations
	G:\org\HP5\DAT\HP5011122_b\0111HP5.61r	CCV_0111HP561r, CAL4 :0111HP5 , 15000 ug per mL Oil (200 uL of CAL5 + 200 uL DCM(14647)	G:\Org\HP5\Methods\DC_ORO-61-BA-L%.xls	1	1	1	1	0	The integration of TEH(Oil Range)is the hydrocarbon response with reference to the baseline. Y-Scale adjusted.
	G:\org\HP5\DAT\HP5011122_b\0111HP5.62r	DCM-Baseline Check-V62	G:\Org\HP5\Methods\DR_8015-HE-LEXP.met	1	1	1	1	0	No Integrations
	G:\org\HP5\DAT\HP5011122_b\0111HP5.63r	CCV_0111HP563r, CAL5 :0111HP5 , 30000 ug per mL Oil (600 uL of DRO211118A + 400 uL of DCM)	G:\Org\HP5\Methods\DC_ORO-BA-L%.xls	1	1	1	1	0	The integration of TEH(Oil Range)is the hydrocarbon response with reference to the baseline. Y-Scale adjusted.
	G:\org\HP5\DAT\HP5011122_b\0111HP5.64r	DCM-Baseline Check-V64	G:\Org\HP5\Methods\DR_8015-HE-LEXP.met	1	1	1	1	0	No Integrations
	G:\org\HP5\DAT\HP5011122_b\0111HP5.65r	DCM-Baseline Check-V65	G:\Org\HP5\Methods\DR_8015-HE-LEXP.met	1	1	1	1	0	No Integrations
	G:\org\HP5\DAT\HP5011122_b\0111HP5.66r	DCM-Baseline Check-V66	G:\Org\HP5\Methods\DR_8015-HE-LEXP.met	1	1	1	1	0	No Integrations
	G:\org\HP5\DAT\HP5011122_b\0111HP5.68r	DCM-Baseline Check-V68	G:\Org\HP5\Methods\DR_8015-IC-LEXP.met	1	1	1	1	0	No Integrations
	G:\org\HP5\DAT\HP5011122_b\0111HP5.69r	CCV_0111HP567r, Second Source :0111HP5 , 5000 ug per mL (100uL of DRO210902A + 900uL DCM(14647)	G:\Org\HP5\Methods\DC_ORO-59-BA-L%.xls	1	1	1	1	0	The integration of TEH(Oil Range)is the hydrocarbon response with reference to the baseline. Y-Scale adjusted.

Ann Nebel

Digitally signed by
Ann Nebel
Date: 2022.02.11 10:29:31 -07:00

PREP BATCH REPORT

Prep Code: **HC-3520-DRO**
 Prep Batch **163748** Prep Temp **NA °C**

Technician: **Jillian L Bostwick**
 Batch Units: **ML**

Prep Start Date: **2/14/2022 1:41:09 PM**
 Prep End Date: **2/15/2022 12:42:00 P**

Sample ID	Matrix	pH	Initial Samp Amt	Sol Added	Sol Recovered	Final Vol (mL)	Factor	Balance	Prep Start Date	Prep End Date
MB-163748			1000	0	0	1.00	0.001		2/14/2022	2/15/2022
Start time: 1:33 PM, 2/14/2022. End time: 02/15/2022 at 8:53 AM. SGT by ALN on remaining sample on 2/16/2022.										
LCS-163748			1000	0	0	1.00	0.001		2/14/2022	2/15/2022
All bottles were completely used, defaced and disposed of on 2/14/2022. SGT by ALN on remaining sample on 2/16/2022.										
LCSD-163748			1000	0	0	1.00	0.001		2/14/2022	2/15/2022
ALN took down, transferred, blew down and bottled samples. SGT by ALN on remaining sample on 2/16/2022.										
LCS-163748-RRO			1000	0	0	1.00	0.001		2/14/2022	2/15/2022
SGT by ALN on remaining sample on 2/16/2022.										
LCSD-163748-RRO			1000	0	0	1.00	0.001		2/14/2022	2/15/2022
SGT by ALN on remaining sample on 2/16/2022.										
B22020962-001D	Ground Water	2	1050	0	0	1.00	0.000952		2/14/2022	2/15/2022
Bottle 1/2. Clear. SGT by ALN on remaining sample on 2/16/2022.										
B22020962-001DMS	Ground Water	2	1050	0	0	1.00	0.000952		2/14/2022	2/15/2022
Bottle 2/2. Clear. SGT by ALN on remaining sample on 2/16/2022.										
B22020962-006D	Ground Water	2	1050	0	0	1.00	0.000952		2/14/2022	2/15/2022
Bottle 1/2. Clear. SGT by ALN on remaining sample on 2/16/2022.										
B22020962-006DMS-RRO	Ground Water	2	1050	0	0	1.00	0.000952		2/14/2022	2/15/2022
Bottle 2/2. Clear. SGT by ALN on remaining sample on 2/16/2022.										
B22020962-011D	Ground Water	2	1040	0	0	1.00	0.000957		2/14/2022	2/15/2022
Bottle 1/2. Clear, light sediment. SGT by ALN on remaining sample on 2/16/2022.										
B22020962-016D	Ground Water	2	1010	0	0	1.00	0.00099		2/14/2022	2/15/2022
Bottle 1/2. Clear.										
B22020962-021D	Ground Water	2	1040	0	0	1.00	0.000957		2/14/2022	2/15/2022
Bottle 1/2. Clear. SGT by ALN on remaining sample on 2/16/2022.										
B22020962-026D	Ground Water	2	1000	0	0	1.00	0.001		2/14/2022	2/15/2022
Bottle 1/2. Turbid, sediment. SGT by ALN on remaining sample on 2/16/2022.										
B22020962-031D	Ground Water	2	1060	0	0	1.00	0.000943		2/14/2022	2/15/2022
Bottle 1/2. Clear.										

Number	Reagent Name	Exp Date
11	Carbon Filter Water	1/1/2023
13379	PTFE Boiling Stones 27463755	12/30/2025
14206	pH-indicator Strips 0-14 HC160347	8/26/2026
14719	4ML, Amber Vial, 20220104	1/4/2027
14828	Dichloromethane ED092	12/12/2023

Spk ID	Spike Name	SampType	AmtAdd	Exp Date
FP220126 14244	DCM RINSED FILTER PAPER	all	1	4/6/2026
Sulfate 01/25/22 (Baked Sodium Sulfate	all	Varies	11/29/2026
DRO220119A	Triacontane SURR 1000 ug/mL	All except LCS, L	100 uL	4/6/2026
DRO211213A	OTP only SURR 2000 ug/mL	All except RRO-L	100 uL	9/30/2024
DRO220106C	#2 Diesel in Acetone 150,000 ug/mL	LCS, LCSD, MS,	100 uL	11/5/2023
DRO220112A	50,000 ug/mL Oil Std for RRO-In D	LCS-RRO, LCSD	100 uL	9/1/2026
SG220126(13376)	Baked Silica Gel	SGT	5g	2/28/2030

PREP BATCH REPORT

Prep Code: **HC-3520-DRO**
 Prep Batch **163748** Prep Temp **NA °C**

Technician: **Jillian L Bostwick**
 Batch Units: **ML**

Prep Start Date: **2/14/2022 1:41:09 PM**
 Prep End Date: **2/15/2022 12:42:00 P**

Sample ID	Matrix	pH	Initial Samp Amt	Sol Added	Sol Recovered	Final Vol (mL)	Factor	Balance	Prep Start Date	Prep End Date
B22020962-032B	Ground Water	2	1060	0	0	1.00	0.000943		2/14/2022	2/15/2022
Bottle 1/2. Clear. SGT by ALN on remaining sample on 2/16/2022.										

Number	Reagent Name	Exp Date
11	Carbon Filter Water	1/1/2023
13379	PTFE Boiling Stones 27463755	12/30/2025
14206	pH-indicator Strips 0-14 HC160347	8/26/2026
14719	4ML, Amber Vial, 20220104	1/4/2027
14828	Dichloromethane ED092	12/12/2023

Spk ID	Spike Name	SampType	AmtAdd	Exp Date
FP220126 14244	DCM RINSED FILTER PAPER	all	1	4/6/2026
Sulfate 01/25/22 (Baked Sodium Sulfate	all	Varies	11/29/2026
DRO220119A	Triacontane SURR 1000 ug/mL	All except LCS, L	100 uL	4/6/2026
DRO211213A	OTP only SURR 2000 ug/mL	All except RRO-L	100 uL	9/30/2024
DRO220106C	#2 Diesel in Acetone 150,000 ug/mL	LCS, LCSD, MS,	100 uL	11/5/2023
DRO220112A	50,000 ug/mL Oil Std for RRO-In D	LCS-RRO, LCSD	100 uL	9/1/2026
SG220126(13376)	Baked Silica Gel	SGT	5g	2/28/2030

Energy Laboratories Inc

ANALYTICAL RUN Summary

17-Feb-22

Run ID GCFID-HP5-B_220215A

Run Start Date: 2/15/2022
Analyst: Ann Nebel
Ical:
Column ID:
Comments: DRO-8015-ICAL information is in Index GCFID-HP5-B_220111A 8015C OIL range calibration GCFID-HP5-B_220111C

Std ID	Std Name	Std Amount	Std Units	Samp Amount	Samp Units	SampType	Expiration Date
DRO220128A	8015 CCV-15,000ug/mL + 200 OTP					CCV-DRO	4/30/2023
DRO220128B	Carbon Scan STD-Marker					MARKER	7/13/2026
DRO220201A	5,000 ug/mL RRO CCV 200 ug/mL Triacontane					CCV-RRO	4/6/2026

Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist					
15041578	CCV_0215HP50	HC-8015-DRO-	CCV		2/15/2022 12:08:	1	R374849		0	0						
Analyte	T	Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD	Q
TEH(Oil Range)	A	mg/L		5.1684375		5	0	0	0.0879	0.3	0	103%	80	120	0%	
n-Triacontane	S	mg/L		0.1922915		0.2	0	0	0.000336	0.002	0	96%	80	120	0%	

Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist					
15041579	CCV_0215HP50	HC-8015-DRO-	CCV		2/15/2022 1:09:2	1	R374849		0	0						
Analyte	T	Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD	Q
Diesel Range Organics (C10 to C24)	A	mg/L		14.96687		15	0	0	0.0389	0.3	0	100%	80	120	0%	
Total Extractable Hydrocarbons	A	mg/L		14.96687		15	0	0	0.0749	0.3	50	100%	80	120	0%	
o-Terphenyl	S	mg/L		0.183568		0.2	0	0	0.000429	0.002	0	92%	80	120	0%	

Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist					
15041580	LCS-163748	HC-8015-DRO-	LCS-DOD		2/15/2022 3:17:5	1	163748	2/14/2022 1:	0	0						
Analyte	T	Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD	Q

Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist					
15041580	LCS-163748	HC-8015-DRO-	LCS-DOD		2/15/2022 3:17:5	1	163748	2/14/2022 1:	0	0						
Analyte	T	Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD	Q
Diesel Range Organics (C10 to C24)	A	mg/L		14.47345		15	0	0	0.0389	0.3	0	96%	36	132	0%	
Total Extractable Hydrocarbons	A	mg/L		14.09594		15	0	0	0.0749	0.3	50	94%	60	132	0%	
o-Terphenyl	S	mg/L		0.2007885		0.2	0	0	0.000429	0.002	0	100%	56	125	0%	
Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist					
15041581	LCSD-163748	HC-8015-DRO-	LCSD-DOD		2/15/2022 4:00:5	1	163748	2/14/2022 1:	0	2E+07						
Analyte	T	Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD	Q
Diesel Range Organics (C10 to C24)	A	mg/L		12.87807		15	0	14.47345	0.0389	0.3	0	86%	36	132	12%	
Total Extractable Hydrocarbons	A	mg/L		13.7668		15	0	14.09594	0.0749	0.3	50	92%	60	132	2%	
o-Terphenyl	S	mg/L		0.2013523		0.2	0	0	0.000429	0.002	0	101%	56	125	0%	
Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist					
15041582	MB-163748	HC-8015-DRO-	MBLK		2/15/2022 4:43:4	1	163748	2/14/2022 1:	0	0						
Analyte	T	Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD	Q
Diesel Range Organics (C10 to C24)	A	mg/L		0		0	0	0	0.0389	0.15	0	0%	0	0	0%	
Oil Range Hydrocarbons (C24 to C40)	A	mg/L		0		0	0	0	0.0879	0.15	0	0%	0	0	0%	
Total Extractable Hydrocarbons	A	mg/L		0		0	0	0	0.0749	0.15	50	0%	0	0	0%	
n-Triacontane	S	mg/L		0.115		0.1	0	0	0.000336	0.002	0	115%	50	150	0%	
o-Terphenyl	S	mg/L		0.2150964		0.2	0	0	0.000429	0.002	0	108%	56	125	0%	
Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist					
15041583	B22020962-001	HC-8015-DRO-	SAMP		2/15/2022 5:26:5	1	163748	2/14/2022 1:	0	0						
Analyte	T	Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD	Q
Diesel Range Organics (C10 to C24)	A	mg/L		0		0	0	0	0.0370328	0.3	0	0%	0	0	0%	U
Oil Range Hydrocarbons (C24 to C40)	A	mg/L		0		0	0	0	0.0836808	0.3	0	0%	0	0	0%	U
Total Extractable Hydrocarbons	A	mg/L		0		0	0	0	0.0713048	0.3	50	0%	0	0	0%	U
n-Triacontane	S	mg/L		0.104		0.0952	0	0	0.0003199	0.001904	0	109%	50	150	0%	
o-Terphenyl	S	mg/L		0.1903713		0.1904	0	0	0.0004084	0.002	0	100%	56	125	0%	

Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist					
15041584	B22020962-001	HC-8015-DRO-	MS-DOD		2/15/2022 6:10:0	1	163748	2/14/2022 1:	2E+07	0						
Analyte	T	Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD	Q
Diesel Range Organics (C10 to C24)	A	mg/L		13.67493		14.28	0	0	0.0370328	0.3	0	96%	36	132	0%	
Total Extractable Hydrocarbons	A	mg/L		14.55865		14.28	0	0	0.0713048	0.3	50	102%	60	132	0%	
o-Terphenyl	S	mg/L		0.2079687		0.1904	0	0	0.0004084	0.002	0	109%	56	125	0%	
Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist					
15041585	B22020962-031	HC-8015-DRO-	SAMP		2/15/2022 7:36:0	1	163748	2/14/2022 1:		0	0					
Analyte	T	Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD	Q
Diesel Range Organics (C10 to C24)	A	mg/L		0		0	0	0	0.0366827	0.3	0	0%	0	0	0%	U
Oil Range Hydrocarbons (C24 to C40)	A	mg/L		0		0	0	0	0.0828897	0.3	0	0%	0	0	0%	U
Total Extractable Hydrocarbons	A	mg/L		0		0	0	0	0.0706307	0.3	50	0%	0	0	0%	U
n-Triacontane	S	mg/L		0.103		0.0943	0	0	0.0003168	0.001886	0	109%	50	150	0%	
o-Terphenyl	S	mg/L		0.1898616		0.1886	0	0	0.0004045	0.002	0	101%	56	125	0%	
Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist					
15041586	B22020962-032	HC-8015-DRO-	SAMP		2/15/2022 8:19:0	1	163748	2/14/2022 1:		0	0					
Analyte	T	Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD	Q
Diesel Range Organics (C10 to C24)	A	mg/L		0.1867703		0	0	0	0.0366827	0.3	0	0%	0	0	0%	J
Oil Range Hydrocarbons (C24 to C40)	A	mg/L		0.25140500		0	0	0	0.0828897	0.3	0	0%	0	0	0%	J
Total Extractable Hydrocarbons	A	mg/L		0.4578004		0	0	0	0.0706307	0.3	50	0%	0	0	0%	
n-Triacontane	S	mg/L		0.101		0.0943	0	0	0.0003168	0.001886	0	107%	50	150	0%	
o-Terphenyl	S	mg/L		0.1831251		0.1886	0	0	0.0004045	0.002	0	97%	56	125	0%	
Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist					
15041587	B22020962-021	HC-8015-DRO-	SAMP		2/15/2022 9:01:5	1	163748	2/14/2022 1:		0	0					
Analyte	T	Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD	Q
Diesel Range Organics (C10 to C24)	A	mg/L		0.1293294		0	0	0	0.0372273	0.3	0	0%	0	0	0%	J
Oil Range Hydrocarbons (C24 to C40)	A	mg/L		0.27971750		0	0	0	0.0841203	0.3	0	0%	0	0	0%	J
Total Extractable Hydrocarbons	A	mg/L		0.4272106		0	0	0	0.0716793	0.3	50	0%	0	0	0%	
n-Triacontane	S	mg/L		0.102		0.0957	0	0	0.0003216	0.001914	0	107%	50	150	0%	
o-Terphenyl	S	mg/L		0.1867143		0.1914	0	0	0.0004106	0.002	0	98%	56	125	0%	

Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist					
15041588	B22020962-026	HC-8015-DRO-	SAMP		2/15/2022 10:28:	1	163748	2/14/2022 1:	0	0						
Analyte	T	Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD	Q
Diesel Range Organics (C10 to C24)	A	mg/L		3.523833		0	0	0	0.0389	0.3	0	0%	0	0	0%	
Oil Range Hydrocarbons (C24 to C40)	A	mg/L		0.87978595		0	0	0	0.0879	0.3	0	0%	0	0	0%	
Total Extractable Hydrocarbons	A	mg/L		4.347314		0	0	0	0.0749	0.3	50	0%	0	0	0%	
n-Triacontane	S	mg/L		0.116		0.1	0	0	0.000336	0.002	0	116%	50	150	0%	
o-Terphenyl	S	mg/L		0.1962887		0.2	0	0	0.000429	0.002	0	98%	56	125	0%	
Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist					
15041589	CCV_0215HP52	HC-8015-DRO-	CCV		2/15/2022 11:53:	1	R374849				0	0				
Analyte	T	Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD	Q
TEH(Oil Range)	A	mg/L		5.20969434		5	0	0	0.0879	0.3	0	104%	80	120	0%	
n-Triacontane	S	mg/L		0.1934945		0.2	0	0	0.000336	0.002	0	97%	80	120	0%	
Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist					
15041590	CCV_0215HP52	HC-8015-DRO-	CCV		2/16/2022 12:36:	1	R374849				0	0				
Analyte	T	Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD	Q
Diesel Range Organics (C10 to C24)	A	mg/L		14.72236		15	0	0	0.0389	0.3	0	98%	80	120	0%	
Total Extractable Hydrocarbons	A	mg/L		15.23634		15	0	0	0.0749	0.3	50	102%	80	120	0%	
o-Terphenyl	S	mg/L		0.1880856		0.2	0	0	0.000429	0.002	0	94%	80	120	0%	
Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist					
15041591	B22020962-011	HC-8015-DRO-	SAMP		2/16/2022 2:02:5	1	163748	2/14/2022 1:	0	0						
Analyte	T	Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD	Q
Diesel Range Organics (C10 to C24)	A	mg/L		0.1697735		0	0	0	0.0372273	0.3	0	0%	0	0	0%	J
Oil Range Hydrocarbons (C24 to C40)	A	mg/L		0.49424407		0	0	0	0.0841203	0.3	0	0%	0	0	0%	
Total Extractable Hydrocarbons	A	mg/L		0.670882		0	0	0	0.0716793	0.3	50	0%	0	0	0%	
n-Triacontane	S	mg/L		0.106		0.0957	0	0	0.0003216	0.001914	0	111%	50	150	0%	
o-Terphenyl	S	mg/L		0.1480728		0.1914	0	0	0.0004106	0.002	0	77%	56	125	0%	

Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist					
15041592	B22020962-006	HC-8015-DRO-	SAMP		2/16/2022 3:28:5	1	163748	2/14/2022 1:	0	0						
Analyte	T	Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD	Q
Diesel Range Organics (C10 to C24)	A	mg/L		0.9079867		0	0	0	0.0370328	0.3	0	0%	0	0	0%	
Oil Range Hydrocarbons (C24 to C40)	A	mg/L		0.72268534		0	0	0	0.0836808	0.3	0	0%	0	0	0%	
Total Extractable Hydrocarbons	A	mg/L		1.637198		0	0	0	0.0713048	0.3	50	0%	0	0	0%	
n-Triacontane	S	mg/L		0.103		0.0952	0	0	0.0003199	0.001904	0	108%	50	150	0%	
o-Terphenyl	S	mg/L		0.1653225		0.1904	0	0	0.0004084	0.002	0	87%	56	125	0%	
TEH(Oil Range)	X	mg/L		1.08583021		0	0	0	0.0836808	0.3	0	0%	0	0	0%	

Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist					
15041593	B22020962-006	HC-8015-DRO-	MS-DOD		2/16/2022 4:11:5	1	163748	2/14/2022 1:	2E+07	0						
Analyte	T	Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD	Q
TEH(Oil Range)	A	mg/L		5.22538567		4.76	1.0858302	0	0.0836808	0.3	0	87%	41	113	0%	
n-Triacontane	S	mg/L		0.1		0.0952	0	0	0.0003199	0.002	0	105%	50	150	0%	

Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist					
15041594	B22020962-016	HC-8015-DRO-	SAMP		2/16/2022 6:20:4	1	163748	2/14/2022 1:	0	0						
Analyte	T	Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD	Q
Diesel Range Organics (C10 to C24)	A	mg/L		0		0	0	0	0.038511	0.3	0	0%	0	0	0%	U
Oil Range Hydrocarbons (C24 to C40)	A	mg/L		0		0	0	0	0.087021	0.3	0	0%	0	0	0%	U
Total Extractable Hydrocarbons	A	mg/L		0		0	0	0	0.074151	0.3	50	0%	0	0	0%	U
n-Triacontane	S	mg/L		0.115		0.099	0	0	0.0003326	0.00198	0	116%	50	150	0%	
o-Terphenyl	S	mg/L		0.2137025		0.198	0	0	0.0004247	0.002	0	108%	56	125	0%	

Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist					
15041595	LCS-163748-RR	HC-8015-DRO-	LCS-DOD		2/16/2022 8:29:2	1	163748	2/14/2022 1:	0	0						
Analyte	T	Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD	Q
TEH(Oil Range)	A	mg/L		5.50884819		5	0	0	0.0879	0.3	0	110%	41	113	0%	
n-Triacontane	S	mg/L		0.112		0.1	0	0	0.000336	0.002	0	112%	50	150	0%	

Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist					
15041596	LCSD-163748-R	HC-8015-DRO-	LCSD-DOD		2/16/2022 9:54:5	1	163748	2/14/2022 1:	0	2E+07						
Analyte	T	Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD	Q
TEH(Oil Range)	A	mg/L		5.23484612		5	0	5.5088482	0.0879	0.3	0	105%	41	113	5%	
n-Triacontane	S	mg/L		0.108		0.1	0	0	0.000336	0.002	0	108%	50	150	0%	
Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist					
15041597	CCV_0215HP53	HC-8015-DRO-	CCV		2/16/2022 11:20:	1	R374849		0	0						
Analyte	T	Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD	Q
TEH(Oil Range)	A	mg/L		5.31640527		5	0	0	0.0879	0.3	0	106%	80	120	0%	
n-Triacontane	S	mg/L		0.1953403		0.2	0	0	0.000336	0.002	0	98%	80	120	0%	
Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist					
15041598	CCV_0215HP53	HC-8015-DRO-	CCV		2/16/2022 12:03:	1	R374849		0	0						
Analyte	T	Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD	Q
Diesel Range Organics (C10 to C24)	A	mg/L		14.46479		15	0	0	0.0389	0.3	0	96%	80	120	0%	
Total Extractable Hydrocarbons	A	mg/L		14.96941		15	0	0	0.0749	0.3	50	100%	80	120	0%	
o-Terphenyl	S	mg/L		0.1842026		0.2	0	0	0.000429	0.002	0	92%	80	120	0%	

Energy Laboratories Inc

ANALYTICAL RUN Summary

18-Feb-22

Run ID GCFID-HP5-B_220215B

Run Start Date: 2/15/2022
Analyst: Ann Nebel
Ical:
Column ID:
Comments: IDOC_JLB supervised by AMN. DRO-8015-ICAL information is in Index GCFID-HP5-B_220111A 8015C OIL range calibration GCFID-HP5-B_220111C.

Std ID	Std Name	Std Amount	Std Units	Samp Amount	Samp Units	SampType	Expiration Date
DRO220128A	8015 CCV-15,000ug/mL + 200 OTP					CCV-DRO	4/30/2023
DRO220128B	Carbon Scan STD-Marker					MARKER	7/13/2026
DRO220201A	5,000 ug/mL RRO CCV 200 ug/mL Triacontane					CCV-RRO	4/6/2026

Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist
--------	--------	-----------	------------	---------	---------------	----	----------	-----------	--------	--------	--------

15042679	CCV_0215HP53	HC-8015-DRO-	CCV		2/16/2022 11:20:	1	R374888			0	0					
Analyte	T	Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD	Q
TEH(Oil Range)	A	mg/L		5.31640527		5	0	0	0.0879	0.3	0	106%	80	120	0%	
n-Triacontane	S	mg/L		0.1953403		0.2	0	0	0.000336	0.002	0	98%	80	120	0%	

Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist
--------	--------	-----------	------------	---------	---------------	----	----------	-----------	--------	--------	--------

15042680	CCV_0215HP53	HC-8015-DRO-	CCV		2/16/2022 12:03:	1	R374888			0	0					
Analyte	T	Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD	Q
Diesel Range Organics (C10 to C24)	A	mg/L		14.46479		15	0	0	0.0389	0.3	0	96%	80	120	0%	
Total Extractable Hydrocarbons	A	mg/L		14.96941		15	0	0	0.0749	0.3	50	100%	80	120	0%	
o-Terphenyl	S	mg/L		0.1842026		0.2	0	0	0.000429	0.002	0	92%	80	120	0%	

Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist
--------	--------	-----------	------------	---------	---------------	----	----------	-----------	--------	--------	--------

15042681	LCS-163748	HC-8015-DRO-	LCS-DOD		2/16/2022 2:12:1	1	163748	2/14/2022 1:		0	0					
Analyte	T	Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD	Q

Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist					
15042681	LCS-163748	HC-8015-DRO-	LCS-DOD		2/16/2022 2:12:1	1	163748	2/14/2022 1:	0	0						
Analyte	T	Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD	Q
Diesel Range Organics (SGT-C10 to	A	mg/L		12.75872		15	0	0	0.0281	0.3	0	85%	36	132	0%	
Total Extractable Hydrocarbons (SGT	A	mg/L		13.51961		15	0	0	0.0357	0.3	0	90%	60	132	0%	
o-Terphenyl (SGT)	S	mg/L		0.2035284		0.2	0	0	0.000429	0.002	0	102%	56	125	0%	
Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist					
15042682	LCSD-163748	HC-8015-DRO-	LCSD-DOD		2/16/2022 2:54:4	1	163748	2/14/2022 1:	0	2E+07						
Analyte	T	Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD	Q
Diesel Range Organics (SGT-C10 to	A	mg/L		11.6549		15	0	12.75872	0.0281	0.3	0	78%	36	132	9%	
Total Extractable Hydrocarbons (SGT	A	mg/L		12.3811		15	0	13.51961	0.0357	0.3	0	83%	60	132	9%	
o-Terphenyl (SGT)	S	mg/L		0.18874		0.2	0	0	0.000429	0.002	0	94%	56	125	0%	
Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist					
15042683	MB-163748	HC-8015-DRO-	MBLK		2/16/2022 3:37:3	1	163748	2/14/2022 1:	0	0						
Analyte	T	Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD	Q
Diesel Range Organics (SGT-C10 to	A	mg/L		0		0	0	0	0.0281	0.15	0	0%	0	0	0%	
Oil Range Hydrocarbons (SGT-C24 t	A	mg/L		0		0	0	0	0.0879	0.15	0	0%	0	0	0%	
Total Extractable Hydrocarbons (SGT	A	mg/L		0		0	0	0	0.0357	0.15	0	0%	0	0	0%	
n-Triacontane (SGT)	S	mg/L		0.1		0.1	0	0	0.000336	0.002	0	100%	50	150	0%	
o-Terphenyl (SGT)	S	mg/L		0.1969335		0.2	0	0	0.000429	0.002	0	98%	56	125	0%	
Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist					
15042684	B22020962-001	HC-8015-DRO-	SAMP		2/16/2022 4:20:2	1	163748	2/14/2022 1:	0	0						
Analyte	T	Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD	Q
Diesel Range Organics (SGT-C10 to	A	mg/L		0		0	0	0	0.0267512	0.3	0	0%	0	0	0%	U
Oil Range Hydrocarbons (SGT-C24 t	A	mg/L		0		0	0	0	0.0836808	0.3	0	0%	0	0	0%	U
Total Extractable Hydrocarbons (SGT	A	mg/L		0		0	0	0	0.0339864	0.3	0	0%	0	0	0%	U
n-Triacontane (SGT)	S	mg/L		0.094		0.0952	0	0	0.0003199	0.001904	0	99%	50	150	0%	
o-Terphenyl (SGT)	S	mg/L		0.1839699		0.1904	0	0	0.0004084	0.001904	0	97%	56	125	0%	

Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist					
15042685	B22020962-001	HC-8015-DRO-	MS-DOD		2/16/2022 5:03:1	1	163748	2/14/2022 1:	2E+07	0						
Analyte	T	Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD	Q
Diesel Range Organics (SGT-C10 to	A	mg/L		11.8361		14.28	0	0	0.0267512	0.3	0	83%	36	132	0%	
Total Extractable Hydrocarbons (SGT	A	mg/L		12.53939		14.28	0	0	0.0339864	0.3	0	88%	60	132	0%	
o-Terphenyl (SGT)	S	mg/L		0.1872221		0.1904	0	0	0.0004084	0.002	0	98%	56	125	0%	

Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist					
15042686	B22020962-021	HC-8015-DRO-	SAMP		2/16/2022 6:28:5	1	163748	2/14/2022 1:		0	0					
Analyte	T	Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD	Q
Diesel Range Organics (SGT-C10 to	A	mg/L		0.02979903		0	0	0	0.0268917	0.3	0	0%	0	0	0%	J
Oil Range Hydrocarbons (SGT-C24 t	A	mg/L		0		0	0	0	0.0841203	0.3	0	0%	0	0	0%	U
Total Extractable Hydrocarbons (SGT	A	mg/L		0.03910249		0	0	0	0.0341649	0.3	0	0%	0	0	0%	J
n-Triacontane (SGT)	S	mg/L		0.09		0.0957	0	0	0.0003216	0.001914	0	94%	50	150	0%	
o-Terphenyl (SGT)	S	mg/L		0.1799082		0.1914	0	0	0.0004106	0.001914	0	94%	56	125	0%	

Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist					
15042687	B22020962-032	HC-8015-DRO-	SAMP		2/16/2022 7:12:0	1	163748	2/14/2022 1:		0	0					
Analyte	T	Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD	Q
Diesel Range Organics (SGT-C10 to	A	mg/L		0		0	0	0	0.0264983	0.3	0	0%	0	0	0%	U
Oil Range Hydrocarbons (SGT-C24 t	A	mg/L		0		0	0	0	0.0828897	0.3	0	0%	0	0	0%	U
Total Extractable Hydrocarbons (SGT	A	mg/L		0		0	0	0	0.0336651	0.3	0	0%	0	0	0%	U
n-Triacontane (SGT)	S	mg/L		0.087		0.0943	0	0	0.0003168	0.001886	0	92%	50	150	0%	
o-Terphenyl (SGT)	S	mg/L		0.1722048		0.1886	0	0	0.0004045	0.001886	0	91%	56	125	0%	

Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist					
15042688	B22020962-011	HC-8015-DRO-	SAMP		2/16/2022 8:37:5	1	163748	2/14/2022 1:		0	0					
Analyte	T	Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD	Q
Diesel Range Organics (SGT-C10 to	A	mg/L		0		0	0	0	0.0268917	0.3	0	0%	0	0	0%	U
Oil Range Hydrocarbons (SGT-C24 t	A	mg/L		0.18008812		0	0	0	0.0841203	0.3	0	0%	0	0	0%	J
Total Extractable Hydrocarbons (SGT	A	mg/L		0.2671859		0	0	0	0.0341649	0.3	0	0%	0	0	0%	J
n-Triacontane (SGT)	S	mg/L		0.09		0.0957	0	0	0.0003216	0.001914	0	94%	50	150	0%	
o-Terphenyl (SGT)	S	mg/L		0.1338975		0.1914	0	0	0.0004106	0.001914	0	70%	56	125	0%	


Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist					
15042689	B22020962-026	HC-8015-DRO-	SAMP		2/16/2022 9:20:5	1	163748	2/14/2022 1:	0	0						
Analyte	T	Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD	Q
Diesel Range Organics (SGT-C10 to	A	mg/L		2.253301		0	0	0	0.0281	0.3	0	0%	0	0	0%	
Oil Range Hydrocarbons (SGT-C24 t	A	mg/L		0.47409186		0	0	0	0.0879	0.3	0	0%	0	0	0%	
Total Extractable Hydrocarbons (SGT	A	mg/L		2.685523		0	0	0	0.0357	0.3	0	0%	0	0	0%	
n-Triacontane (SGT)	S	mg/L		0.079		0.1	0	0	0.000336	0.002	0	79%	50	150	0%	
o-Terphenyl (SGT)	S	mg/L		0.151495		0.2	0	0	0.000429	0.002	0	76%	56	125	0%	
Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist					
15042690	CCV_0215HP55	HC-8015-DRO-	CCV		2/16/2022 10:46:	1	R374888				0	0				
Analyte	T	Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD	Q
TEH(Oil Range)	A	mg/L		5.21005078		5	0	0	0.0879	0.3	0	104%	80	120	0%	
n-Triacontane	S	mg/L		0.1927553		0.2	0	0	0.000336	0.002	0	96%	80	120	0%	
Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist					
15042691	CCV_0215HP55	HC-8015-DRO-	CCV		2/16/2022 11:29:	1	R374888				0	0				
Analyte	T	Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD	Q
Diesel Range Organics (C10 to C24)	A	mg/L		14.07582		15	0	0	0.0389	0.3	0	94%	80	120	0%	
Total Extractable Hydrocarbons	A	mg/L		14.56759		15	0	0	0.0749	0.3	50	97%	80	120	0%	
o-Terphenyl	S	mg/L		0.1791171		0.2	0	0	0.000429	0.002	0	90%	80	120	0%	
Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist					
15042692	CCV_0215HP55	HC-8015-DRO-	CCV		2/17/2022 9:51:1	1	R374888				0	0				
Analyte	T	Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD	Q
TEH(Oil Range)	A	mg/L		5.24299192		5	0	0	0.0879	0.3	0	105%	80	120	0%	
n-Triacontane	S	mg/L		0.1916585		0.2	0	0	0.000336	0.002	0	96%	80	120	0%	
Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist					
15042693	CCV_0215HP55	HC-8015-DRO-	CCV		2/17/2022 10:34:	1	R374888				0	0				
Analyte	T	Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD	Q

Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist					
15042693	CCV_0215HP55	HC-8015-DRO-	CCV		2/17/2022 10:34:	1	R374888		0	0						
Analyte	T	Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD	Q
Diesel Range Organics (C10 to C24)	A	mg/L		14.43044		15	0	0	0.0389	0.3	0	96%	80	120	0%	
Total Extractable Hydrocarbons	A	mg/L		14.93495		15	0	0	0.0749	0.3	50	100%	80	120	0%	
o-Terphenyl	S	mg/L		0.1845303		0.2	0	0	0.000429	0.002	0	92%	80	120	0%	
Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist					
15042694	B22020962-006	HC-8015-DRO-	SAMP		2/17/2022 12:42:	1	163748	2/14/2022 1:	0	0						
Analyte	T	Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD	Q
Diesel Range Organics (SGT-C10 to	A	mg/L		0.04571747		0	0	0	0.0267512	0.3	0	0%	0	0	0%	J
Oil Range Hydrocarbons (SGT-C24 t	A	mg/L		0		0	0	0	0.0836808	0.3	0	0%	0	0	0%	U
Total Extractable Hydrocarbons (SGT	A	mg/L		0.0521687		0	0	0	0.0339864	0.3	0	0%	0	0	0%	J
n-Triacontane (SGT)	S	mg/L		0.095		0.0952	0	0	0.0003199	0.001904	0	100%	50	150	0%	
o-Terphenyl (SGT)	S	mg/L		0.1632467		0.1904	0	0	0.0004084	0.001904	0	86%	56	125	0%	
TEH (SGT-Oil Range)	X	mg/L		0		0	0	0	0.0836808	0.2856	0	0%	0	0	0%	U
Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist					
15042695	B22020962-006	HC-8015-DRO-	MS-DOD		2/17/2022 1:25:3	1	163748	2/14/2022 1:	2E+07	0						
Analyte	T	Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD	Q
TEH (SGT-Oil Range)	A	mg/L		4.82638168		4.76	0	0	0.0836808	0.3	0	101%	41	113	0%	
n-Triacontane (SGT)	S	mg/L		0.091		0.0952	0	0	0.0003199	0.002	0	96%	50	150	0%	
Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist					
15042696	LCS-163748-RR	HC-8015-DRO-	LCS-DOD		2/17/2022 2:51:0	1	163748	2/14/2022 1:	0	0						
Analyte	T	Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD	Q
TEH (SGT-Oil Range)	A	mg/L		5.19994164		5	0	0	0.0879	0.3	0	104%	41	113	0%	
n-Triacontane (SGT)	S	mg/L		0.099		0.1	0	0	0.000336	0.002	0	99%	50	150	0%	
Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist					
15042697	LCSD-163748-R	HC-8015-DRO-	LCSD-DOD		2/17/2022 4:16:1	1	163748	2/14/2022 1:	0	2E+07						
Analyte	T	Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD	Q

Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist					
15042697	LCSD-163748-R	HC-8015-DRO-	LCSD-DOD		2/17/2022 4:16:1	1	163748	2/14/2022 1:	0	2E+07						
Analyte	T	Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD	Q
TEH (SGT-Oil Range)	A	mg/L		4.98869514		5	0	5.1999416	0.0879	0.3	0	100%	41	113	4%	
n-Triacontane (SGT)	S	mg/L		0.094		0.1	0	0	0.000336	0.002	0	94%	50	150	0%	
Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist					
15042698	CCV_0215HP56	HC-8015-DRO-	CCV		2/17/2022 5:42:5	1	R374888		0	0						
Analyte	T	Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD	Q
TEH(Oil Range)	A	mg/L		5.21258545		5	0	0	0.0879	0.3	0	104%	80	120	0%	
n-Triacontane	S	mg/L		0.1896352		0.2	0	0	0.000336	0.002	0	95%	80	120	0%	
Seq No	Lab ID	Test Code	Sample Typ	File ID	Analysis Date	DF	Batch ID	Prep Date	SPKref	RPDref	pmoist					
15042699	CCV_0215HP56	HC-8015-DRO-	CCV		2/17/2022 6:26:1	1	R374888		0	0						
Analyte	T	Units	RAW	Final	Text	Spike	SPKref	RPDref	MDL	PQL	UQL	%REC	LOW	HIGH	%RPD	Q
Diesel Range Organics (C10 to C24)	A	mg/L		14.50951		15	0	0	0.0389	0.3	0	97%	80	120	0%	
Total Extractable Hydrocarbons	A	mg/L		15.01328		15	0	0	0.0749	0.3	50	100%	80	120	0%	
o-Terphenyl	S	mg/L		0.184661		0.2	0	0	0.000429	0.002	0	92%	80	120	0%	

Write Sequence Data File	Insert Entries(Have the first call for entries select) Sample Name	Method	Weight	Dil Factor	Amt Inj.	IS	Cal ID
G:\org\HP5\DAT\HP5021522_b\0215HP5.01r	DCM-Baseline Check-V01	G:\Org\HP5\Methods\DR_8015-JA-LEXP.met	1	1	1	1	0
G:\org\HP5\DAT\HP5021522_b\0215HP5.02r	DCM-Baseline Check-V02	G:\Org\HP5\Methods\DR_8015-JA-LEXP.met	1	1	1	1	0
G:\org\HP5\DAT\HP5021522_b\0215HP5.03r	MARKER_0215HP503r_DRO_0215HP5 , DRO220128B	G:\org\HP5\Methods\CSC220215.met	1	1	1	1	0
G:\org\HP5\DAT\HP5021522_b\0215HP5.04r	CCV_0215HP504r_RRO_0215HP5 , DRO220201A	G:\Org\HP5\Methods\DC_ORO-BE-L%.MET G:\Org\HP5\Methods\DS_ORO-BE-L%.MET	1	1	1	1	0
G:\org\HP5\DAT\HP5021522_b\0215HP5.05r	CCV_0215HP505r_DRO_0215HP5 , DRO220128A	G:\Org\HP5\Methods\DC_8015-C24-JEB-L%.met G:\Org\HP5\Methods\DS_8015-C24-JEB-L#.met	1	1	1	1	0
G:\org\HP5\DAT\HP5021522_b\0215HP5.06r	DCM-Baseline Check-V06	G:\Org\HP5\Methods\DR_8015-JA-LEXP.met	1	1	1	1	0
G:\org\HP5\DAT\HP5021522_b\0215HP5.07r	DCM-Baseline Check-V07	G:\Org\HP5\Methods\DR_8015-JA-LEXP.met	1	1	1	1	0
G:\org\HP5\DAT\HP5021522_b\0215HP5.08r	LCS-163748_0215HP5 ,	G:\Org\HP5\Methods\I3_8015-C24-JEB-L%.met G:\Org\HP5\Methods\DS_8015-C24-JEB-L#.met	1000	1	1	1	0
G:\org\HP5\DAT\HP5021522_b\0215HP5.09r	LCSD-163748_0215HP5 ,	G:\Org\HP5\Methods\I3_8015-C24-JEB-L%.met G:\Org\HP5\Methods\DS_8015-C24-JEB-L#.met	1000	1	1	1	0
G:\org\HP5\DAT\HP5021522_b\0215HP5.10r	MB-163748_0215HP5 ,	G:\Org\HP5\Methods\DR_8015-C24Tb-JEB-L%.met G:\Org\HP5\Methods\DR_OROS-BEB-L%.MET G:\Org\HP5\Methods\DS_8015-C24T-JEB-L#.met	1000	1	1	1	0
G:\org\HP5\DAT\HP5021522_b\0215HP5.11r	B22020962-001D_0215HP5 , \$HC-8015-DRO-W,	G:\Org\HP5\Methods\DR_8015-C24Tb-JEB-L%.met G:\Org\HP5\Methods\DR_OROS-BEB-L%.MET G:\Org\HP5\Methods\DS_8015-C24T-JEB-L#.met	1050	1	1	1	0
G:\org\HP5\DAT\HP5021522_b\0215HP5.12r	B22020962-001DMS_0215HP5 ,	G:\Org\HP5\Methods\I3_8015-C24-JEB-L%.met G:\Org\HP5\Methods\DS_8015-C24-JEB-L#.met	1050	1	1	1	0
G:\org\HP5\DAT\HP5021522_b\0215HP5.13r	DCM-Baseline Check-V13	G:\Org\HP5\Methods\DR_8015-JA-LEXP.met	1	1	1	1	0
G:\org\HP5\DAT\HP5021522_b\0215HP5.14r	B22020962-031D_0215HP5 , \$HC-8015-DRO-W,	G:\Org\HP5\Methods\DR_8015-C24Tb-JEB-L%.met G:\Org\HP5\Methods\DR_OROS-BEB-L%.MET G:\Org\HP5\Methods\DS_8015-C24T-JEB-L#.met	1060	1	1	1	0
G:\org\HP5\DAT\HP5021522_b\0215HP5.15r	B22020962-032B_0215HP5 , \$HC-8015-DRO-W,	G:\Org\HP5\Methods\I3_8015-C24Tb-JEB-L%.met G:\Org\HP5\Methods\I3_OROS-BEB-L%.MET G:\Org\HP5\Methods\DS_8015-C24T-JEB-L#.met	1060	1	1	1	0
G:\org\HP5\DAT\HP5021522_b\0215HP5.16r	B22020962-021D_0215HP5 , \$HC-8015-DRO-W,	G:\Org\HP5\Methods\I3_8015-021516-JEB-L%.met G:\Org\HP5\Methods\I3_OROS-021516-BEB-L%.MET G:\Org\HP5\Methods\DS_8015-C24T-JEB-L#.met	1045	1	1	1	0
G:\org\HP5\DAT\HP5021522_b\0215HP5.17r	DCM-Baseline Check-V17	G:\Org\HP5\Methods\DR_8015-JA-LEXP.met	1	1	1	1	0
G:\org\HP5\DAT\HP5021522_b\0215HP5.18r	B22020962-026D_0215HP5 , \$HC-8015-DRO-W,	G:\Org\HP5\Methods\I3_8015-C24Tb-JEB-L%.met G:\Org\HP5\Methods\I3_OROS-BEB-L%.MET G:\Org\HP5\Methods\DS_8015-C24T-JEB-L#.met	1000	1	1	1	0
G:\org\HP5\DAT\HP5021522_b\0215HP5.19r	MARKER_0215HP519r_DRO_0215HP5 , DRO220128B	G:\org\HP5\Methods\CSC220215.met	1	1	1	1	0
G:\org\HP5\DAT\HP5021522_b\0215HP5.20r	CCV_0215HP520r_RRO_0215HP5 , DRO220201A	G:\Org\HP5\Methods\DC_ORO-BE-L%.MET G:\Org\HP5\Methods\DS_ORO-BE-L%.MET	1	1	1	1	0
G:\org\HP5\DAT\HP5021522_b\0215HP5.21r	CCV_0215HP521r_DRO_0215HP5 , DRO220128A	G:\Org\HP5\Methods\DC_8015-C24-JEB-L%.met G:\Org\HP5\Methods\DS_8015-C24-JEB-L#.met	1	1	1	1	0
G:\org\HP5\DAT\HP5021522_b\0215HP5.22r	DCM-Baseline Check-V22	G:\Org\HP5\Methods\DR_8015-JA-LEXP.met	1	1	1	1	0
G:\org\HP5\DAT\HP5021522_b\0215HP5.23r	B22020962-011D_0215HP5 , \$HC-8015-DRO-W,	G:\Org\HP5\Methods\I3_8015-C24Tb-JEB-L%.met G:\Org\HP5\Methods\I3_OROS-BEB-L%.MET G:\Org\HP5\Methods\DS_8015-C24T-JEB-L#.met	1045	1	1	1	0
G:\org\HP5\DAT\HP5021522_b\0215HP5.24r	DCM-Baseline Check-V24	G:\Org\HP5\Methods\DR_8015-JA-LEXP.met	1	1	1	1	0
G:\org\HP5\DAT\HP5021522_b\0215HP5.25r	B22020962-006D_0215HP5 , \$HC-8015-DRO-W,	G:\Org\HP5\Methods\I3_8015-021516-JEB-L%.met G:\Org\HP5\Methods\I3_OROS-021516-BEB-L%.MET G:\Org\HP5\Methods\DS_8015-C24T-JEB-L#.met	1050	1	1	1	0
G:\org\HP5\DAT\HP5021522_b\0215HP5.26r	B22020962-006DMS-RRO_0215HP5 ,	G:\Org\HP5\Methods\I3_ORO-BE-L%.MET G:\Org\HP5\Methods\DS_ORO-BE-L%.MET	1050	1	1	1	0
G:\org\HP5\DAT\HP5021522_b\0215HP5.27r	DCM-Baseline Check-V27	G:\Org\HP5\Methods\DR_8015-JA-LEXP.met	1	1	1	1	0
G:\org\HP5\DAT\HP5021522_b\0215HP5.28r	DCM-Baseline Check-V28	G:\Org\HP5\Methods\DR_8015-JA-LEXP.met	1	1	1	1	0
G:\org\HP5\DAT\HP5021522_b\0215HP5.29r	B22020962-016D_0215HP5 , \$HC-8015-DRO-W,	G:\Org\HP5\Methods\DR_8015-C24Tb-JEB-L%.met G:\Org\HP5\Methods\DR_OROS-BEB-L%.MET G:\Org\HP5\Methods\DS_8015-C24T-JEB-L#.met	1010	1	1	1	0
G:\org\HP5\DAT\HP5021522_b\0215HP5.30r	DCM-Baseline Check-V30	G:\Org\HP5\Methods\DR_8015-JA-LEXP.met	1	1	1	1	0
G:\org\HP5\DAT\HP5021522_b\0215HP5.31r	DCM-Baseline Check-V31	G:\Org\HP5\Methods\DR_8015-JA-LEXP.met	1	1	1	1	0
G:\org\HP5\DAT\HP5021522_b\0215HP5.32r	LCS-163748-RRO_0215HP5 ,	G:\Org\HP5\Methods\I3_ORO-BE-L%.MET G:\Org\HP5\Methods\DS_ORO-BE-L%.MET	1000	1	1	1	0
G:\org\HP5\DAT\HP5021522_b\0215HP5.33r	DCM-Baseline Check-V33	G:\Org\HP5\Methods\DR_8015-JA-LEXP.met	1	1	1	1	0
G:\org\HP5\DAT\HP5021522_b\0215HP5.34r	LCSD-163748-RRO_0215HP5 ,	G:\Org\HP5\Methods\I3_ORO-BE-L%.MET G:\Org\HP5\Methods\DS_ORO-BE-L%.MET	1000	1	1	1	0
G:\org\HP5\DAT\HP5021522_b\0215HP5.35r	MARKER_0215HP535r_DRO_0215HP5 , DRO220128B	G:\org\HP5\Methods\CSC220215.met	1	1	1	1	0
G:\org\HP5\DAT\HP5021522_b\0215HP5.36r	CCV_0215HP536r_RRO_0215HP5 , DRO220201A	G:\Org\HP5\Methods\DC_ORO-BE-L%.MET G:\Org\HP5\Methods\DS_ORO-BE-L%.MET	1	1	1	1	0
G:\org\HP5\DAT\HP5021522_b\0215HP5.37r	CCV_0215HP537r_DRO_0215HP5 , DRO220128A	G:\Org\HP5\Methods\DC_8015-C24-JEB-L%.met G:\Org\HP5\Methods\DS_8015-C24-JEB-L#.met	1	1	1	1	0

Write Sequence Data File	Insert Entries(Have the first cell for entries select) Sample Name	Method	Weight	Dil Factor	Amt Inj.	IS	Cal ID
G:\org\HP5\DAT\HP5021522_b\0215HP5.35r	MARKER_0215HP535r, DRO_0215HP5, DRO220128B	G:\org\HP5\Methods\CSC220215.met	1	1	1	1	0
G:\org\HP5\DAT\HP5021522_b\0215HP5.36r	CCV_0215HP536r, RRO_0215HP5, DRO220201A	G:\Org\HP5\Methods\DC_ORO-BE-L%.MET G:\Org\HP5\Methods\DS_ORO-BE-L%.MET	1	1	1	1	0
G:\org\HP5\DAT\HP5021522_b\0215HP5.37r	CCV_0215HP537r, DRO_0215HP5, DRO220128A	G:\Org\HP5\Methods\DC_8015-C24-JEB-L%.met G:\Org\HP5\Methods\DS_8015-C24-JEB-L#.met	1	1	1	1	0
G:\org\HP5\DAT\HP5021522_b\0215HP5.38r	DCM-Baseline Check-V38	G:\Org\HP5\Methods\DR_8015-JA-LEXP.met	1	1	1	1	0
G:\org\HP5\DAT\HP5021522_b\0215HP5.39r	DCM-Baseline Check-V39	G:\Org\HP5\Methods\DR_8015-JA-LEXP.met	1	1	1	1	0
G:\org\HP5\DAT\HP5021522_b\0215HP5.40r	LCS-163748_0215HP5, SGT	G:\Org\HP5\Methods\D3_8015-021540-JEB-L%.met G:\Org\HP5\Methods\DS_8015-C24-JEB-L#.met	1000	1	1	1	0
G:\org\HP5\DAT\HP5021522_b\0215HP5.41r	LCSD-163748_0215HP5, SGT	G:\Org\HP5\Methods\D3_8015-C24-JEB-L%.met G:\Org\HP5\Methods\DS_8015-C24-JEB-L#.met	1000	1	1	1	0
G:\org\HP5\DAT\HP5021522_b\0215HP5.42r	MB-163748_0215HP5, SGT	G:\Org\HP5\Methods\DR_8015-C24Tb-JEB-L%.met G:\Org\HP5\Methods\DR_OROS-BEB-L%.MET G:\Org\HP5\Methods\DS_8015-C24T-JEB-L#.met	1000	1	1	1	0
G:\org\HP5\DAT\HP5021522_b\0215HP5.43r	B22020962-001D_0215HP5, \$HC-8015-DRO-W, SGT	G:\Org\HP5\Methods\DR_8015-C24Tb-JEB-L%.met G:\Org\HP5\Methods\DR_OROS-BEB-L%.MET G:\Org\HP5\Methods\DS_8015-C24T-JEB-L#.met	1050	1	1	1	0
G:\org\HP5\DAT\HP5021522_b\0215HP5.44r	B22020962-001DMS_0215HP5, SGT	G:\Org\HP5\Methods\D3_8015-021540-JEB-L%.met G:\Org\HP5\Methods\DS_8015-C24-JEB-L#.met	1050	1	1	1	0
G:\org\HP5\DAT\HP5021522_b\0215HP5.45r	DCM-Baseline Check-V45	G:\Org\HP5\Methods\DR_8015-JA-LEXP.met	1	1	1	1	0
G:\org\HP5\DAT\HP5021522_b\0215HP5.46r	B22020962-021D_0215HP5, \$HC-8015-DRO-W, SGT	G:\Org\HP5\Methods\DR_8015-C24Tb-JEB-L%.met G:\Org\HP5\Methods\DR_OROS-BEB-L%.MET G:\Org\HP5\Methods\DS_8015-C24T-JEB-L#.met	1045	1	1	1	0
G:\org\HP5\DAT\HP5021522_b\0215HP5.47r	B22020962-032B_0215HP5, \$HC-8015-DRO-W, SGT	G:\Org\HP5\Methods\DR_8015-C24Tb-JEB-L%.met G:\Org\HP5\Methods\DR_OROS-BEB-L%.MET G:\Org\HP5\Methods\DS_8015-C24T-JEB-L#.met	1060	1	1	1	0
G:\org\HP5\DAT\HP5021522_b\0215HP5.48r	DCM-Baseline Check-V48	G:\Org\HP5\Methods\DR_8015-JA-LEXP.met	1	1	1	1	0
G:\org\HP5\DAT\HP5021522_b\0215HP5.49r	B22020962-011D_0215HP5, \$HC-8015-DRO-W, SGT	G:\Org\HP5\Methods\D3_8015-021549-JEB-L%.met G:\Org\HP5\Methods\D3_OROS-021549-BEB-L%.MET G:\Org\HP5\Methods\DS_8015-C24T-JEB-L#.met	1045	1	1	1	0
G:\org\HP5\DAT\HP5021522_b\0215HP5.50r	B22020962-026D_0215HP5, \$HC-8015-DRO-W, SGT	G:\Org\HP5\Methods\D3_8015-C24Tb-JEB-L%.met G:\Org\HP5\Methods\D3_OROS-BEB-L%.MET G:\Org\HP5\Methods\DS_8015-C24T-JEB-L#.met	1000	1	1	1	0
G:\org\HP5\DAT\HP5021522_b\0215HP5.51r	MARKER_0215HP551r, DRO_0215HP5, DRO220128B	G:\org\HP5\Methods\CSC220215.met	1	1	1	1	0
G:\org\HP5\DAT\HP5021522_b\0215HP5.52r	CCV_0215HP552r, RRO_0215HP5, DRO220201A	G:\Org\HP5\Methods\DC_ORO-BE-L%.MET G:\Org\HP5\Methods\DS_ORO-BE-L%.MET	1	1	1	1	0
G:\org\HP5\DAT\HP5021522_b\0215HP5.53r	CCV_0215HP553r, DRO_0215HP5, DRO220128A	G:\Org\HP5\Methods\DC_8015-C24-JEB-L%.met G:\Org\HP5\Methods\DS_8015-C24-JEB-L#.met	1	1	1	1	0
G:\org\HP5\DAT\HP5021522_b\0215HP5.54r	DCM-Baseline Check-V54	G:\Org\HP5\Methods\DR_8015-JA-LEXP.met	1	1	1	1	0
G:\org\HP5\DAT\HP5021522_b\0215HP5.55r	MARKER_0215HP555r, DRO_0215HP5, DRO220128B	G:\org\HP5\Methods\CSC220215.met	1	1	1	1	0
G:\org\HP5\DAT\HP5021522_b\0215HP5.56r	CCV_0215HP556r, RRO_0215HP5, DRO220201A	G:\Org\HP5\Methods\DC_ORO-BE-L%.MET G:\Org\HP5\Methods\DS_ORO-BE-L%.MET	1050	1	1	1	0
G:\org\HP5\DAT\HP5021522_b\0215HP5.57r	CCV_0215HP557r, DRO_0215HP5, DRO220128A	G:\Org\HP5\Methods\DC_8015-C24-JEB-L%.met G:\Org\HP5\Methods\DS_8015-C24-JEB-L#.met	1050	1	1	1	0
G:\org\HP5\DAT\HP5021522_b\0215HP5.58r	DCM-Baseline Check-V58	G:\Org\HP5\Methods\DR_8015-JA-LEXP.met	1	1	1	1	0
G:\org\HP5\DAT\HP5021522_b\0215HP5.59r	DCM-Baseline Check-V59	G:\Org\HP5\Methods\DR_8015-JA-LEXP.met	1	1	1	1	0
G:\org\HP5\DAT\HP5021522_b\0215HP5.60r	B22020962-006D_0215HP5, \$HC-8015-DRO-W, SGT	G:\Org\HP5\Methods\DR_8015-C24Tb-JEB-L%.met G:\Org\HP5\Methods\DR_OROS-BEB-L%.MET G:\Org\HP5\Methods\DS_8015-C24T-JEB-L#.met	1050	1	1	1	0
G:\org\HP5\DAT\HP5021522_b\0215HP5.61r	B22020962-006DMS-RRO_0215HP5, SGT	G:\Org\HP5\Methods\D3_ORO-BE-L%.MET G:\Org\HP5\Methods\DS_ORO-BE-L%.MET	1050	1	1	1	0
G:\org\HP5\DAT\HP5021522_b\0215HP5.62r	DCM-Baseline Check-V62	G:\Org\HP5\Methods\DR_8015-JA-LEXP.met	1	1	1	1	0
G:\org\HP5\DAT\HP5021522_b\0215HP5.63r	LCS-163748-RRO_0215HP5, SGT	G:\Org\HP5\Methods\D3_ORO-BE-L%.MET G:\Org\HP5\Methods\DS_ORO-BE-L%.MET	1000	1	1	1	0
G:\org\HP5\DAT\HP5021522_b\0215HP5.64r	DCM-Baseline Check-V64	G:\Org\HP5\Methods\DR_8015-JA-LEXP.met	1	1	1	1	0
G:\org\HP5\DAT\HP5021522_b\0215HP5.65r	LCSD-163748-RRO_0215HP5, SGT	G:\Org\HP5\Methods\D3_ORO-BE-L%.MET G:\Org\HP5\Methods\DS_ORO-BE-L%.MET	1000	1	1	1	0
G:\org\HP5\DAT\HP5021522_b\0215HP5.66r	MARKER_0215HP566r, DRO_0215HP5, DRO220128B	G:\org\HP5\Methods\CSC220215.met	1	1	1	1	0
G:\org\HP5\DAT\HP5021522_b\0215HP5.67r	CCV_0215HP567r, RRO_0215HP5, DRO220201A	G:\Org\HP5\Methods\DC_ORO-BE-L%.MET G:\Org\HP5\Methods\DS_ORO-BE-L%.MET	1	1	1	1	0
G:\org\HP5\DAT\HP5021522_b\0215HP5.68r	CCV_0215HP568r, DRO_0215HP5, DRO220128A	G:\Org\HP5\Methods\DC_8015-C24-JEB-L%.met G:\Org\HP5\Methods\DS_8015-C24-JEB-L#.met	1	1	1	1	0



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT


Sample Name: DCM-Baseline Check-V01
 Raw File: G:\org\HP5\DAT\HP5021522_b\0215HP5.0001.RAW
 Date & Time Acquired: 2/15/2022 9:59:30 AM
 Method File: G:\Org\HP5\Methods\DR_8015-JA-LEXP.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO220111JA.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for TEH: 32675.36

Rt range for Diesel Range Organics: 6.63 to 15.99

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*o-Terphenyl	12.395	200.	1.509	.75	-
*1-Chlorooctadecane	13.116	200.	1.966	.98	-

DRO Area:2208238 DRO Amount: 67.58113
 TEH Area:2314252 TEH Amount: 70.82561




DIESEL RANGE ORGANICS CHROMATOGRAM REPORT


Sample Name: DCM-Baseline Check-V02
 Raw File: G:\org\HP5\DAT\HP5021522_b\0215HP5.0002.RAW
 Date & Time Acquired: 2/15/2022 10:42:27 AM
 Method File: G:\Org\HP5\Methods\DR_8015-JA-LEXP.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO220111JA.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for TEH: 32675.36
 Rt range for Diesel Range Organics: 6.63 to 15.99

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*o-Terphenyl	29.984	200.	.	-
*1-Chlorooctadecane	13.162	200.	.074	.04

DRO Area:114416.5 DRO Amount: 3.501614
 TEH Area:178990.9 TEH Amount: 5.477857





RESIDUAL RANGE ORGANICS CHROMATOGRAM

Sample Name: CCV_0215HP504r, RRO ;0215HP5 , DRO220201A
 Raw File: G:\org\HP5\DAT\HP5021522_b\0215HP5.0004.RAW
 Date & Time Acquired: 2/15/2022 12:08:46 PM
 Method File: G:\Org\HP5\Methods\DC_ORO-BE-L%.MET
 Calibration File: G:\Org\HP5\Cals\SW8015C_ORO220111BE.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for for ~~Residual~~ TEH(Oil Range) Organics Calculations: 26424.55
 Rt range for ~~Residual~~ TEH(Oil Range) Organics: 12.61 to 30.05

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*#Triacontane	16.34	500.	333.272	66.65	-


~~RRO~~ TEH(Oil Range) Area:1.365736E+08 ~~RRO~~ TEH(Oil Range) AMOUNT: 5168.438

CONTINUING CALIBRATION REPORT: G:\org\HP5\DAT\HP5021522_b\0215HP5.0004.RAW

COMPOUND	ACTUAL (NG)	MEASURED (NG)	%RECOVERY	LIMITS
*30-40 Motor Oil	5000.	.033	.	75-125

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	LIMITS
*#Triacontane	16.34	200.	333.272	166.64	75-125

AMN 02/18/2022



RESIDUAL RANGE ORGANICS CHROMATOGRAM

Sample Name: CCV_0215HP504r, RRO ;0215HP5 , DRO220201A
 Raw File: G:\org\HP5\DAT\HP5021522_b\0215HP5.0004.RAW
 Date & Time Acquired: 2/15/2022 12:08:46 PM
 Method File: G:\Org\HP5\Methods\DS_ORO-BE-L%.MET
 Calibration File: G:\Org\HP5\Cals\SW8015C_ORO220111BE.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for for Residual Range Organics Calculations: 26424.55
 Rt range for Residual Range Organics: 12.61 to 30.05


SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*#Triacontane	16.34	500.	192.292	38.46	-

RRO Area:3412283 RRO AMOUNT: 129.133

CONTINUING CALIBRATION REPORT: G:\org\HP5\DAT\HP5021522_b\0215HP5.0004.RAW

COMPOUND	ACTUAL (NG)	MEASURED (NG)	%RECOVERY	LIMITS
*30-40 Motor Oil	5000.	.033	.	75-125

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	LIMITS
*#Triacontane	16.34	200.	192.292	96.15	75-125



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: CCV_0215HP505r, DRO ;0215HP5 , DRO220128A
 Raw File: G:\org\HP5\DAT\HP5021522_b\0215HP5.0005.RAW
 Date & Time Acquired: 2/15/2022 1:09:22 PM
 Method File: G:\Org\HP5\Methods\DC_8015-C24-JEb-L%.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO220111JEB-C24.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for TEH: 32675.36

Rt range for Diesel Range Organics: 6.52 to 14.41

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*o-Terphenyl	12.256	200.	306.319	153.16	-
*1-Chlorooctadecane	13.106	200.	151.347	75.67	-

DRO Area: 4.729253E+08 DRO Amount: 14473.45
 TEH Area: 4.890477E+08 TEH Amount: 14966.87


CONTINUING CALIBRATION REPORT: G:\org\HP5\DAT\HP5021522_b\0215HP5.0005.RAW

COMPOUND	ACTUAL (NG)	MEASURED (NG)	%RECOVERY	LIMITS
TOTAL DRO	15000.	14966.87	99.78	85-115

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	LIMITS
*o-Terphenyl	12.256	200.	306.319	153.16	85-115
*1-Chlorooctadecane	13.106	200.	151.347	75.67	85-115

G:\org\HP5\DAT\HP5021522_b\0215HP5.0005.RAW

CCV_0215HP505r, DRO ;0215HP5 , DRO220128A



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: CCV_0215HP505r, DRO ;0215HP5 , DRO220128A
 Raw File: G:\org\HP5\DAT\HP5021522_b\0215HP5.0005.RAW
 Date & Time Acquired: 2/15/2022 1:09:22 PM
 Method File: G:\Org\HP5\Methods\DS_8015-C24-JEb-L#.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO220111JEB-C24.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for TEH: 32675.36

Rt range for Diesel Range Organics: 6.52 to 14.41


SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*o-Terphenyl	12.256	200.	183.568	91.78
*1-Chlorooctadecane	13.106	200.	35.59	17.8

DRO Area: 2.43595E+08 DRO Amount: 7455.007
 TEH Area: 2.54425E+08 TEH Amount: 7786.448

CONTINUING CALIBRATION REPORT: G:\org\HP5\DAT\HP5021522_b\0215HP5.0005.RAW

COMPOUND	ACTUAL (NG)	MEASURED (NG)	%RECOVERY	LIMITS
TOTAL DRO	15000.	7786.45	51.91	85-115

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	LIMITS
*o-Terphenyl	12.256	200.	183.568	91.78	85-115
*1-Chlorooctadecane	13.106	200.	35.59	17.8	85-115




DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: DCM-Baseline Check-V06
 Raw File: G:\org\HP5\DAT\HP5021522_b\0215HP5.0006.RAW
 Date & Time Acquired: 2/15/2022 1:52:09 PM
 Method File: G:\Org\HP5\Methods\DR_8015-JA-LEXP.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO220111JA.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for TEH: 32675.36
 Rt range for Diesel Range Organics: 6.63 to 15.99

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*o-Terphenyl	29.911	200.	.034	.02
*1-Chlorooctadecane	13.144	200.		

DRO Area: 89566.79 DRO Amount: 2.741111
 TEH Area: 213378.2 TEH Amount: 6.530248



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: DCM-Baseline Check-V07
 Raw File: G:\org\HP5\DAT\HP5021522_b\0215HP5.0007.RAW
 Date & Time Acquired: 2/15/2022 2:34:56 PM
 Method File: G:\Org\HP5\Methods\DR_8015-JA-LEXP.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO220111JA.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for TEH: 32675.36
 Rt range for Diesel Range Organics: 6.63 to 15.99


SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*o-Terphenyl	29.922	200.	.	-
*1-Chlorooctadecane	13.155	200.	.031	.02

DRO Area:81352.14 DRO Amount: 2.489709
 TEH Area:158437.8 TEH Amount: 4.848848

Batch ID: 163748

LCS-163748 ;0215HP5 ,

G:\org\HP5\DAT\HP5021522_b\0215HP5.0008.RAW



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: LCS-163748 ;0215HP5 ,
 Raw File: G:\org\HP5\DAT\HP5021522_b\0215HP5.0008.RAW
 Date & Time Acquired: 2/15/2022 3:17:56 PM
 Method File: G:\Org\HP5\Methods\D3_8015-C24-JEb-L%.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO220111JEB-C24.CAL
 Sample Weight: 1000 Dilution: 1 S.A.: 1


Mean RF for TEH: 32675.36

Rt range for Diesel Range Organics: 6.52 to 14.41

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*o-Terphenyl	12.254	.2	.342	170.82	-
*1-Chlorooctadecane	13.106	.2	.212	105.93	-

DRO Area: 4.317244E+08 DRO Amount: 13.21254
 TEH Area: 4.6059E+08 TEH Amount: 14.09594

G:\Org\HP5\DAT\HP5021522_b\0215HP5.0008.RAW Batch ID: 163748
LCS-163748 ;0215HP5 ,



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: LCS-163748 ;0215HP5 ,
Raw File: G:\Org\HP5\DAT\HP5021522_b\0215HP5.0008.RAW
Date & Time Acquired: 2/15/2022 3:17:56 PM
Method File: G:\Org\HP5\Methods\DS_8015-C24-JEb-L#.met
Calibration File: G:\Org\HP5\Cals\SW8015C_DRO220111JEb-C24.CAL
Sample Weight: 1000 Dilution: 1 S.A.: 1

Mean RF for TEH: 32675.36
Rt range for Diesel Range Organics: 6.52 to 14.41


SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*o-Terphenyl	12.254	.2	.201	100.39
*1-Chlorooctadecane	13.106	.2	.058	28.95

DRO Area: 2.02994E+08 DRO Amount: 6.212448
TEH Area: 2.175475E+08 TEH Amount: 6.657846

Batch ID: 163748

LCSD-163748 ;0215HP5 ,

G:\Org\HP5\DAT\HP5021522_b\0215HP5.0009.RAW



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: LCSD-163748 ;0215HP5 ,
 Raw File: G:\Org\HP5\DAT\HP5021522_b\0215HP5.0009.RAW
 Date & Time Acquired: 2/15/2022 4:00:54 PM
 Method File: G:\Org\HP5\Methods\D3_8015-C24-JEb-L%.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO220111JEB-C24.CAL
 Sample Weight: 1000 Dilution: 1 S.A.: 1

Mean RF for TEH: 32675.36

Rt range for Diesel Range Organics: 6.52 to 14.41


SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*o-Terphenyl	12.254	.2	.365	182.54	-
*1-Chlorooctadecane	13.107	.2	.202	101.1	-

DRO Area: 4.207957E+08 DRO Amount: 12.87807
 TEH Area: 4.49835E+08 TEH Amount: 13.7668

Batch ID: 163748

G:\Org\HP5\DAT\HP5021522_b\0215HP5.0009.RAW

LCSD-163748 ;0215HP5 ,



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT


Sample Name: LCSD-163748 ;0215HP5 ,
 Raw File: G:\Org\HP5\DAT\HP5021522_b\0215HP5.0009.RAW
 Date & Time Acquired: 2/15/2022 4:00:54 PM
 Method File: G:\Org\HP5\Methods\DS_8015-C24-JEb-L#.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO220111JEB-C24.CAL
 Sample Weight: 1000 Dilution: 1 S.A.: 1

Mean RF for TEH: 32675.36

Rt range for Diesel Range Organics: 6.52 to 14.41

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*o-Terphenyl	12.254	.2	.201	100.68
*1-Chlorooctadecane	13.107	.2	.056	28.13

DRO Area:1.991289E+08 DRO Amount: 6.094161
 TEH Area:2.141581E+08 TEH Amount: 6.554115



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: MB-163748 ;0215HP5 ,
 Raw File: G:\Org\HP5\DAT\HP5021522_b\0215HP5.0010.RAW
 Date & Time Acquired: 2/15/2022 4:43:40 PM
 Method File: G:\Org\HP5\Methods\DR_8015-C24Tb-JEb-L%.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO220111JEB-C24-T.CAL
 Sample Weight: 1000 Dilution: 1 S.A.: 1

Mean RF for TEH: 32675.36


Rt range for Diesel Range Organics: 6.52 to 14.41

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*o-Terphenyl	12.249	.2	.215	107.74 -
*1-Chlorooctadecane	29.979	.2	.	-
*#Triacontane	16.331	.2	.116	57.91 -

DRO Area: 552632.2 DRO Amount: 1.691281E-02
 TEH Area: 871436.4 TEH Amount: 2.666953E-02

G:\org\HP5\DAT\HP5021522_b\0215HP5.0010.RAW

MB-163748 ;0215HP5 ,



RESIDUAL RANGE ORGANICS CHROMATOGRAM

Sample Name: MB-163748 ;0215HP5 ,
 Raw File: G:\org\HP5\DAT\HP5021522_b\0215HP5.0010.RAW
 Date & Time Acquired: 2/15/2022 4:43:40 PM
 Method File: G:\Org\HP5\Methods\DR_OROS-BEb-L%.MET
 Calibration File: G:\Org\HP5\Cals\SW8015C_ORO220111BEb_SAMP.CAL
 Sample Weight: 1000 Dilution: 1 S.A.: 1

Mean RF for for Residual Range Organics Calculations: 26424.55
 Rt range for Residual Range Organics: 14.32 to 21.77


SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*#Triacontane_____	16.331	.5	.116	23.16

RRO Area:200168 RRO AMOUNT: 7.575076E-03

Batch ID: 163748

G:\Org\HP5\DAT\HP5021522_b\0215HP5.0010.RAW

MB-163748 ;0215HP5 ,



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: MB-163748 ;0215HP5 ,
Raw File: G:\Org\HP5\DAT\HP5021522_b\0215HP5.0010.RAW
Date & Time Acquired: 2/15/2022 4:43:40 PM
Method File: G:\Org\HP5\Methods\DS_8015-C24T-JEb-L#.met
Calibration File: G:\Org\HP5\Cals\SW8015C_DRO220111JEb-C24-T.CAL
Sample Weight: 1000 Dilution: 1 S.A.: 1

Mean RF for TEH: 32675.36

Rt range for Diesel Range Organics: 6.52 to 14.41

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*o-Terphenyl	12.249	.2	.215	107.55
*1-Chlorooctadecane	29.979	.2	.	-
*#Triacontane	16.331	.2	.115	57.63


DRO Area:395245.4 DRO Amount: 1.209613E-02
TEH Area:927861.2 TEH Amount: 2.839636E-02

ERH2524 (OWDFMW07A)

Batch ID: 163748

G:\Org\HP5\DAT\HP5021522_b\0215HP5.0011.RAW

B22020962-001D ;0215HP5 , \$HC-8015-DRO-W,



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: B22020962-001D ;0215HP5 , \$HC-8015-DRO-W,
Raw File: G:\org\HP5\DAT\HP5021522_b\0215HP5.0011.RAW
Date & Time Acquired: 2/15/2022 5:26:50 PM
Method File: G:\Org\HP5\Methods\DR_8015-C24Tb-JEb-L%.met
Calibration File: G:\Org\HP5\Cals\SW8015C_DRO220111JEB-C24-T.CAL
Sample Weight: 1050 Dilution: 1 S.A.: 1

Mean RF for TEH: 32675.36

Rt range for Diesel Range Organics: 6.52 to 14.41

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*o-Terphenyl	12.248	.19	.191	100.3	-
*1-Chlorooctadecane	13.103	.19	.	.14	-
*#Triacontane	16.331	.19	.105	54.97	-


DRO Area:995916.4 DRO Amount: 2.902774E-02
TEH Area:1289321 TEH Amount: 3.757952E-02

ERH2524 (OWDFMW07A)

Batch ID: 163748

G:\Org\HP5\DAT\HP5021522_b\0215HP5.0011.RAW

B22020962-001D ;0215HP5 , \$HC-8015-DRO-W,



RESIDUAL RANGE ORGANICS CHROMATOGRAM

Sample Name: B22020962-001D ;0215HP5 , \$HC-8015-DRO-W,
Raw File: G:\org\HP5\DAT\HP5021522_b\0215HP5.0011.RAW
Date & Time Acquired: 2/15/2022 5:26:50 PM
Method File: G:\Org\HP5\Methods\DR_OROS-BEb-L%.MET
Calibration File: G:\Org\HP5\Cals\SW8015C_ORO220111BEb_SAMP.CAL
Sample Weight: 1050 Dilution: 1 S.A.: 1

Mean RF for for Residual Range Organics Calculations: 26424.55
Rt range for Residual Range Organics: 14.32 to 21.77

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*#Triacontane_____	16.331	.476	.105	21.99


RRO Area:192831.7 RRO AMOUNT: 6.949947E-03

ERH2524 (OWDFMW07A)

Batch ID: 163748

G:\Org\HP5\DAT\HP5021522_b\0215HP5.0011.RAW

B22020962-001D ;0215HP5 , \$HC-8015-DRO-W,



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: B22020962-001D ;0215HP5 , \$HC-8015-DRO-W,
Raw File: G:\Org\HP5\DAT\HP5021522_b\0215HP5.0011.RAW
Date & Time Acquired: 2/15/2022 5:26:50 PM
Method File: G:\Org\HP5\Methods\DS_8015-C24T-JEb-L#.met
Calibration File: G:\Org\HP5\Cals\SW8015C_DRO220111JEB-C24-T.CAL
Sample Weight: 1050 Dilution: 1 S.A.: 1

Mean RF for TEH: 32675.36

Rt range for Diesel Range Organics: 6.52 to 14.41


SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*o-Terphenyl	12.248	.19	.19	99.94
*1-Chlorooctadecane	13.103	.19	.	.03
*#Triacontane	16.331	.19	.104	54.73

DRO Area:592515.9 DRO Amount: 1.726992E-02
TEH Area:1102999 TEH Amount: 3.214884E-02

Batch ID: 163748

B22020962-001DMS ;0215HP5 ,

G:\org\HP5\DAT\HP5021522_b\0215HP5.0012.RAW



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: B22020962-001DMS ;0215HP5 ,
 Raw File: G:\org\HP5\DAT\HP5021522_b\0215HP5.0012.RAW
 Date & Time Acquired: 2/15/2022 6:10:00 PM
 Method File: G:\Org\HP5\Methods\D3_8015-C24-JEb-L%.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO220111JEB-C24.CAL
 Sample Weight: 1050 Dilution: 1 S.A.: 1

Mean RF for TEH: 32675.36

Rt range for Diesel Range Organics: 6.52 to 14.41


SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*o-Terphenyl	12.252	.19	.372	195.34	-
*1-Chlorooctadecane	13.104	.19	.154	80.72	-

DRO Area: 4.691749E+08 DRO Amount: 13.67493
 TEH Area: 4.994944E+08 TEH Amount: 14.55865

Batch ID: 163748

G:\Org\HP5\DAT\HP5021522_b\0215HP5.0012.RAW

B22020962-001DMS ;0215HP5 ,



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT


Sample Name: B22020962-001DMS ;0215HP5 ,
 Raw File: G:\Org\HP5\DAT\HP5021522_b\0215HP5.0012.RAW
 Date & Time Acquired: 2/15/2022 6:10:00 PM
 Method File: G:\Org\HP5\Methods\DS_8015-C24-JEb-L#.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO220111JEB-C24.CAL
 Sample Weight: 1050 Dilution: 1 S.A.: 1

Mean RF for TEH: 32675.36

Rt range for Diesel Range Organics: 6.52 to 14.41

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*o-Terphenyl	12.252	.19	.208	109.18	-
*1-Chlorooctadecane	13.104	.19	.06	31.57	-

DRO Area: 2.180153E+08 DRO Amount: 6.354441
 TEH Area: 2.335988E+08 TEH Amount: 6.808648



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: DCM-Baseline Check-V13
 Raw File: G:\org\HP5\DAT\HP5021522_b\0215HP5.0013.RAW
 Date & Time Acquired: 2/15/2022 6:53:03 PM
 Method File: G:\Org\HP5\Methods\DR_8015-JA-LEXP.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO220111JA.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for TEH: 32675.36
 Rt range for Diesel Range Organics: 6.63 to 15.99

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*o-Terphenyl	29.963	200.	.	-
*1-Chlorooctadecane	13.169	200.	.033	.02


DRO Area:62458.48 DRO Amount: 1.911486
 TEH Area:109943 TEH Amount: 3.364706

ERH2526 (OWDFMW08A)

G:\Org\HP5\DAT\HP5021522_b\0215HP5.0014.RAW

Batch ID: 163748

B22020962-031D ;0215HP5 , \$HC-8015-DRO-W,



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: B22020962-031D ;0215HP5 , \$HC-8015-DRO-W,
Raw File: G:\Org\HP5\DAT\HP5021522_b\0215HP5.0014.RAW
Date & Time Acquired: 2/15/2022 7:36:03 PM
Method File: G:\Org\HP5\Methods\DR_8015-C24Tb-JEb-L%.met
Calibration File: G:\Org\HP5\Cals\SW8015C_DRO220111JEb-C24-T.CAL
Sample Weight: 1060 Dilution: 1 S.A.: 1

Mean RF for TEH: 32675.36

Rt range for Diesel Range Organics: 6.52 to 14.41

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*o-Terphenyl	12.247	.189	.19	100.78	-
*1-Chlorooctadecane	13.131	.189	.	.11	-
*#Triacontane	16.33	.189	.103	54.66	-


DRO Area:965439.4 DRO Amount: 2.787397E-02
TEH Area:1357710 TEH Amount: 3.919953E-02

ERH2526 (OWDFMW08A)

Batch ID: 163748

G:\Org\HP5\DAT\HP5021522_b\0215HP5.0014.RAW

B22020962-031D ;0215HP5 , \$HC-8015-DRO-W,



RESIDUAL RANGE ORGANICS CHROMATOGRAM

Sample Name: B22020962-031D ;0215HP5 , \$HC-8015-DRO-W,
Raw File: G:\org\HP5\DAT\HP5021522_b\0215HP5.0014.RAW
Date & Time Acquired: 2/15/2022 7:36:03 PM
Method File: G:\Org\HP5\Methods\DR_OROS-BEb-L%.MET
Calibration File: G:\Org\HP5\Cals\SW8015C_ORO220111BEb_SAMP.CAL
Sample Weight: 1060 Dilution: 1 S.A.: 1

Mean RF for for Residual Range Organics Calculations: 26424.55
Rt range for Residual Range Organics: 14.32 to 21.77

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*#Triacontane_____	16.33	.472	.103	21.87


RRO Area:184615.2 RRO AMOUNT: 6.59104E-03

ERH2526 (OWDFMW08A)

Batch ID: 163748

G:\Org\HP5\DAT\HP5021522_b\0215HP5.0014.RAW

B22020962-031D ;0215HP5 , \$HC-8015-DRO-W,



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: B22020962-031D ;0215HP5 , \$HC-8015-DRO-W,
Raw File: G:\Org\HP5\DAT\HP5021522_b\0215HP5.0014.RAW
Date & Time Acquired: 2/15/2022 7:36:03 PM
Method File: G:\Org\HP5\Methods\DS_8015-C24T-JEb-L#.met
Calibration File: G:\Org\HP5\Cals\SW8015C_DRO220111JEB-C24-T.CAL
Sample Weight: 1060 Dilution: 1 S.A.: 1

Mean RF for TEH: 32675.36

Rt range for Diesel Range Organics: 6.52 to 14.41

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*o-Terphenyl	12.247	.189	.19	100.63	-
*1-Chlorooctadecane	13.131	.189	.	.04	-
*#Triacontane	16.33	.189	.103	54.43	-


DRO Area:751452.7 DRO Amount: 2.169579E-02
TEH Area:1211146 TEH Amount: 3.496796E-02

ERH2527 (OWDFMW08A-FD)

Batch ID: 163748

G:\Org\HP5\DAT\HP5021522_b\0215HP5.0015.RAW

B22020962-032B ;0215HP5 , \$HC-8015-DRO-W,



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: B22020962-032B ;0215HP5 , \$HC-8015-DRO-W,
Raw File: G:\org\HP5\DAT\HP5021522_b\0215HP5.0015.RAW
Date & Time Acquired: 2/15/2022 8:19:02 PM
Method File: G:\Org\HP5\Methods\D3_8015-C24Tb-JEb-L%.met
Calibration File: G:\Org\HP5\Cals\SW8015C_DRO220111JEb-C24-T.CAL
Sample Weight: 1060 Dilution: 1 S.A.: 1

Mean RF for TEH: 32675.36

Rt range for Diesel Range Organics: 6.52 to 14.41

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*o-Terphenyl	12.247	.189	.186	98.82	-
*1-Chlorooctadecane	13.099	.189	.002	.81	-
*#Triacontane	16.331	.189	.106	56.26	-


DRO Area:6468954 DRO Amount: 0.1867703
TEH Area:1.585632E+07 TEH Amount: 0.4578004

ERH2527 (OWDFMW08A-FD)

Batch ID: 163748

G:\Org\HP5\DAT\HP5021522_b\0215HP5.0015.RAW

B22020962-032B ;0215HP5 , \$HC-8015-DRO-W,



RESIDUAL RANGE ORGANICS CHROMATOGRAM

Sample Name: B22020962-032B ;0215HP5 , \$HC-8015-DRO-W,
Raw File: G:\org\HP5\DAT\HP5021522_b\0215HP5.0015.RAW
Date & Time Acquired: 2/15/2022 8:19:02 PM
Method File: G:\Org\HP5\Methods\D3_OROS-BEb-L%.MET
Calibration File: G:\Org\HP5\Cals\SW8015C_ORO220111BEb_SAMP.CAL
Sample Weight: 1060 Dilution: 1 S.A.: 1

Mean RF for for Residual Range Organics Calculations: 26424.55
Rt range for Residual Range Organics: 14.32 to 21.77

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*#Triacontane_____	16.331	.472	.106	22.5


RRO Area:7041861 RRO AMOUNT: 0.251405

ERH2527 (OWDFMW08A-FD)

Batch ID: 163748

G:\Org\HP5\DAT\HP5021522_b\0215HP5.0015.RAW

B22020962-032B ;0215HP5 , \$HC-8015-DRO-W,



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: B22020962-032B ;0215HP5 , \$HC-8015-DRO-W,
Raw File: G:\Org\HP5\DAT\HP5021522_b\0215HP5.0015.RAW
Date & Time Acquired: 2/15/2022 8:19:02 PM
Method File: G:\Org\HP5\Methods\DS_8015-C24T-JEb-L#.met
Calibration File: G:\Org\HP5\Cals\SW8015C_DRO220111JEB-C24-T.CAL
Sample Weight: 1060 Dilution: 1 S.A.: 1

Mean RF for TEH: 32675.36

Rt range for Diesel Range Organics: 6.52 to 14.41

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*o-Terphenyl	12.247	.189	.183	97.06
*1-Chlorooctadecane	13.099	.189	.	.02
*#Triacontane	16.331	.189	.101	53.52


DRO Area:3218569 DRO Amount: 9.292586E-02
TEH Area:9788461 TEH Amount: 0.2826104

ERH2533 (RHMW2254-01-Bailer)

Batch ID: 163748

G:\Org\HP5\DAT\HP5021522_b\0215HP5.0016.RAW

B22020962-021D ;0215HP5 , \$HC-8015-DRO-W,



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: B22020962-021D ;0215HP5 , \$HC-8015-DRO-W,
Raw File: G:\Org\HP5\DAT\HP5021522_b\0215HP5.0016.RAW
Date & Time Acquired: 2/15/2022 9:01:59 PM
Method File: G:\Org\HP5\Methods\D3_8015-021516-JEb-L%.met
Calibration File: G:\Org\HP5\Cals\SW8015C_DRO220111JEb-C24-T.CAL
Sample Weight: 1045 Dilution: 1 S.A.: 1

Mean RF for TEH: 32675.36

Rt range for Diesel Range Organics: 6.52 to 14.41

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*o-Terphenyl	12.249	.191	.189	98.85	-
*1-Chlorooctadecane	13.098	.191	.001	.48	-
*#Triacontane	16.331	.191	.108	56.5	-


DRO Area:4416048 DRO Amount: 0.1293294
TEH Area:1.458743E+07 TEH Amount: 0.4272106

ERH2533 (RHMW2254-01-Bailer)

Batch ID: 163748

G:\org\HP5\DAT\HP5021522_b\0215HP5.0016.RAW

B22020962-021D ;0215HP5 , \$HC-8015-DRO-W,



RESIDUAL RANGE ORGANICS CHROMATOGRAM

Sample Name: B22020962-021D ;0215HP5 , \$HC-8015-DRO-W,
Raw File: G:\org\HP5\DAT\HP5021522_b\0215HP5.0016.RAW
Date & Time Acquired: 2/15/2022 9:01:59 PM
Method File: G:\Org\HP5\Methods\D3_OROS-021516-BEb-L%.MET
Calibration File: G:\Org\HP5\Cals\SW8015C_ORO220111BEb_SAMP.CAL
Sample Weight: 1045 Dilution: 1 S.A.: 1

Mean RF for for Residual Range Organics Calculations: 26424.55
Rt range for Residual Range Organics: 14.32 to 21.77

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*#Triacontane_____	16.331	.478	.108	22.6


RRO Area:7724023 RRO AMOUNT: 0.2797175

ERH2533 (RHMW2254-01-Bailer)

Batch ID: 163748

G:\Org\HP5\DAT\HP5021522_b\0215HP5.0016.RAW

B22020962-021D ;0215HP5 , \$HC-8015-DRO-W,



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT


Sample Name: B22020962-021D ;0215HP5 , \$HC-8015-DRO-W,
 Raw File: G:\Org\HP5\DAT\HP5021522_b\0215HP5.0016.RAW
 Date & Time Acquired: 2/15/2022 9:01:59 PM
 Method File: G:\Org\HP5\Methods\DS_8015-C24T-JEb-L#.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO220111JEB-C24-T.CAL
 Sample Weight: 1045 Dilution: 1 S.A.: 1

Mean RF for TEH: 32675.36

Rt range for Diesel Range Organics: 6.52 to 14.41

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*o-Terphenyl	12.249	.191	.187	97.56	-
*1-Chlorooctadecane	29.994	.191	.		-
*#Triacontane	16.331	.191	.102	53.44	-

DRO Area:1740724 DRO Amount: 5.097923E-02
 TEH Area:3092168 TEH Amount: 9.055791E-02



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: DCM-Baseline Check-V17
 Raw File: G:\org\HP5\DAT\HP5021522_b\0215HP5.0017.RAW
 Date & Time Acquired: 2/15/2022 9:45:00 PM
 Method File: G:\Org\HP5\Methods\DR_8015-JA-LEXP.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO220111JA.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for TEH: 32675.36

Rt range for Diesel Range Organics: 6.63 to 15.99

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*o-Terphenyl	29.906	200.	.	-
*1-Chlorooctadecane	13.165	200.	.015	.01 -


DRO Area: 79769.91 DRO Amount: 2.441287
 TEH Area: 166497.7 TEH Amount: 5.095512

ERH2537 (Sump Adit 3)

Batch ID: 163748

G:\Org\HP5\DAT\HP5021522_b\0215HP5.0018.RAW

B22020962-026D ;0215HP5 , \$HC-8015-DRO-W,



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: B22020962-026D ;0215HP5 , \$HC-8015-DRO-W,
Raw File: G:\Org\HP5\DAT\HP5021522_b\0215HP5.0018.RAW
Date & Time Acquired: 2/15/2022 10:28:02 PM
Method File: G:\Org\HP5\Methods\D3_8015-C24Tb-JEb-L%.met
Calibration File: G:\Org\HP5\Cals\SW8015C_DRO220111JEB-C24-T.CAL
Sample Weight: 1000 Dilution: 1 S.A.: 1

Mean RF for TEH: 32675.36

Rt range for Diesel Range Organics: 6.52 to 14.41

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*o-Terphenyl	12.248	.2	.212	105.87	-
*1-Chlorooctadecane	13.111	.2	.018	8.94	-
*#Triacontane	16.331	.2	.146	72.98	-


DRO Area:1.151425E+08 DRO Amount: 3.523833
TEH Area:1.420501E+08 TEH Amount: 4.347314

ERH2537 (Sump Adit 3)

Batch ID: 163748

G:\Org\HP5\DAT\HP5021522_b\0215HP5.0018.RAW

B22020962-026D ;0215HP5 , \$HC-8015-DRO-W,



RESIDUAL RANGE ORGANICS CHROMATOGRAM

Sample Name: B22020962-026D ;0215HP5 , \$HC-8015-DRO-W,
Raw File: G:\org\HP5\DAT\HP5021522_b\0215HP5.0018.RAW
Date & Time Acquired: 2/15/2022 10:28:02 PM
Method File: G:\Org\HP5\Methods\D3_OROS-BEb-L%.MET
Calibration File: G:\Org\HP5\Cals\SW8015C_ORO220111BEb_SAMP.CAL
Sample Weight: 1000 Dilution: 1 S.A.: 1

Mean RF for for Residual Range Organics Calculations: 26424.55
Rt range for Residual Range Organics: 14.32 to 21.77

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*#Triacontane_____	16.331	.5	.146	29.19


RRO Area:2.324795E+07 RRO AMOUNT: 0.879786

ERH2537 (Sump Adit 3)

Batch ID: 163748

G:\Org\HP5\DAT\HP5021522_b\0215HP5.0018.RAW

B22020962-026D ;0215HP5 , \$HC-8015-DRO-W,



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT


Sample Name: B22020962-026D ;0215HP5 , \$HC-8015-DRO-W,
 Raw File: G:\Org\HP5\DAT\HP5021522_b\0215HP5.0018.RAW
 Date & Time Acquired: 2/15/2022 10:28:02 PM
 Method File: G:\Org\HP5\Methods\DS_8015-C24T-JEb-L#.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO220111JEB-C24-T.CAL
 Sample Weight: 1000 Dilution: 1 S.A.: 1


Mean RF for TEH: 32675.36

Rt range for Diesel Range Organics: 6.52 to 14.41

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*o-Terphenyl	12.248	.2	.196	98.14
*1-Chlorooctadecane	13.111	.2	.003	1.43
*#Triacontane	16.331	.2	.116	58.11

DRO Area: 7.851872E+07 DRO Amount: 2.402995
 TEH Area: 8.013168E+07 TEH Amount: 2.452358





RESIDUAL RANGE ORGANICS CHROMATOGRAM

Sample Name: CCV_0215HP520r, RRO ;0215HP5 , DRO220201A
 Raw File: G:\org\HP5\DAT\HP5021522_b\0215HP5.0020.RAW
 Date & Time Acquired: 2/15/2022 11:53:56 PM
 Method File: G:\Org\HP5\Methods\DC_ORO-BE-L%.MET
 Calibration File: G:\Org\HP5\Cals\SW8015C_ORO220111BE.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for for ~~Residual~~ TEH(Oil Range) Organics Calculations: 26424.55
 Rt range for ~~Residual~~ TEH(Oil Range) Organics: 12.61 to 30.05

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*#Triacontane	16.338	500.	334.512	66.9	-


~~RRO~~ TEH(Oil Range) Area:1.376638E+08 ~~RRO~~ TEH(Oil Range) AMOUNT: 5209.694

CONTINUING CALIBRATION REPORT: G:\org\HP5\DAT\HP5021522_b\0215HP5.0020.RAW

COMPOUND	ACTUAL (NG)	MEASURED (NG)	%RECOVERY	LIMITS
*30-40 Motor Oil	5000.	.043	.	75-125

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	LIMITS
*#Triacontane	16.338	200.	334.512	167.26	75-125

AMN 02/18/2022



RESIDUAL RANGE ORGANICS CHROMATOGRAM

Sample Name: CCV_0215HP520r, RRO ;0215HP5 , DRO220201A
 Raw File: G:\org\HP5\DAT\HP5021522_b\0215HP5.0020.RAW
 Date & Time Acquired: 2/15/2022 11:53:56 PM
 Method File: G:\Org\HP5\Methods\DS_ORO-BE-L%.MET
 Calibration File: G:\Org\HP5\Cals\SW8015C_ORO220111BE.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for for Residual Range Organics Calculations: 26424.55
 Rt range for Residual Range Organics: 12.61 to 30.05


SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*#Triacontane	16.338	500.	193.495	38.7	-

RRO Area:3563541 RRO AMOUNT: 134.8572

CONTINUING CALIBRATION REPORT: G:\org\HP5\DAT\HP5021522_b\0215HP5.0020.RAW

COMPOUND	ACTUAL (NG)	MEASURED (NG)	%RECOVERY	LIMITS
*30-40 Motor Oil	5000.	.043	.	75-125

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	LIMITS
*#Triacontane	16.338	200.	193.495	96.75	75-125



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: CCV_0215HP521r, DRO ;0215HP5 , DRO220128A
 Raw File: G:\org\HP5\DAT\HP5021522_b\0215HP5.0021.RAW
 Date & Time Acquired: 2/16/2022 12:36:55 AM
 Method File: G:\Org\HP5\Methods\DC_8015-C24-JEb-L%.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO220111JEB-C24.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for TEH: 32675.36
 Rt range for Diesel Range Organics: 6.52 to 14.41


SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*o-Terphenyl	12.254	200.	312.292	156.15
*1-Chlorooctadecane	13.104	200.	158.623	79.31

DRO Area: 4.810584E+08 DRO Amount: 14722.36
 TEH Area: 4.978528E+08 TEH Amount: 15236.34

CONTINUING CALIBRATION REPORT: G:\org\HP5\DAT\HP5021522_b\0215HP5.0021.RAW

COMPOUND	ACTUAL (NG)	MEASURED (NG)	%RECOVERY	LIMITS
TOTAL DRO	15000.	15236.34	101.58	85-115

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	LIMITS
*o-Terphenyl	12.254	200.	312.292	156.15	85-115
*1-Chlorooctadecane	13.104	200.	158.623	79.31	85-115



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: CCV_0215HP521r, DRO ;0215HP5 , DRO220128A
 Raw File: G:\org\HP5\DAT\HP5021522_b\0215HP5.0021.RAW
 Date & Time Acquired: 2/16/2022 12:36:55 AM
 Method File: G:\Org\HP5\Methods\DS_8015-C24-JEb-L#.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO220111JEB-C24.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for TEH: 32675.36
 Rt range for Diesel Range Organics: 6.52 to 14.41


SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*o-Terphenyl	12.254	200.	188.086	94.04
*1-Chlorooctadecane	13.104	200.	36.938	18.47

DRO Area: 2.49074E+08 DRO Amount: 7622.685
 TEH Area: 2.600504E+08 TEH Amount: 7958.609

CONTINUING CALIBRATION REPORT: G:\org\HP5\DAT\HP5021522_b\0215HP5.0021.RAW

COMPOUND	ACTUAL (NG)	MEASURED (NG)	%RECOVERY	LIMITS
TOTAL DRO	15000.	7958.61	53.06	85-115

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	LIMITS
*o-Terphenyl	12.254	200.	188.086	94.04	85-115
*1-Chlorooctadecane	13.104	200.	36.938	18.47	85-115



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: DCM-Baseline Check-V22
 Raw File: G:\org\HP5\DAT\HP5021522_b\0215HP5.0022.RAW
 Date & Time Acquired: 2/16/2022 1:19:56 AM
 Method File: G:\Org\HP5\Methods\DR_8015-JA-LEXP.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO220111JA.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for TEH: 32675.36

Rt range for Diesel Range Organics: 6.63 to 15.99

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*o-Terphenyl	29.967	200.	.	-
*1-Chlorooctadecane	13.162	200.	.016	.01


DRO Area:207963 DRO Amount: 6.364522
 TEH Area:442068.5 TEH Amount: 13.52911

ERH2529 (RHMW19)

Batch ID: 163748

G:\Org\HP5\DAT\HP5021522_b\0215HP5.0023.RAW

B22020962-011D ;0215HP5 , \$HC-8015-DRO-W,



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: B22020962-011D ;0215HP5 , \$HC-8015-DRO-W,
Raw File: G:\Org\HP5\DAT\HP5021522_b\0215HP5.0023.RAW
Date & Time Acquired: 2/16/2022 2:02:54 AM
Method File: G:\Org\HP5\Methods\D3_8015-C24Tb-JEb-L%.met
Calibration File: G:\Org\HP5\Cals\SW8015C_DRO220111JEb-C24-T.CAL
Sample Weight: 1045 Dilution: 1 S.A.: 1

Mean RF for TEH: 32675.36

Rt range for Diesel Range Organics: 6.52 to 14.41

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*o-Terphenyl	12.248	.191	.151	78.89	-
*1-Chlorooctadecane	13.101	.191	.001	.69	-
*#Triacontane	16.333	.191	.118	61.57	-


DRO Area:5797046 DRO Amount: 0.1697735
TEH Area:2.290777E+07 TEH Amount: 0.670882

ERH2529 (RHMW19)

Batch ID: 163748

G:\org\HP5\DAT\HP5021522_b\0215HP5.0023.RAW

B22020962-011D ;0215HP5 , \$HC-8015-DRO-W,



RESIDUAL RANGE ORGANICS CHROMATOGRAM

Sample Name: B22020962-011D ;0215HP5 , \$HC-8015-DRO-W,
Raw File: G:\org\HP5\DAT\HP5021522_b\0215HP5.0023.RAW
Date & Time Acquired: 2/16/2022 2:02:54 AM
Method File: G:\Org\HP5\Methods\D3_OROS-BEb-L%.MET
Calibration File: G:\Org\HP5\Cals\SW8015C_ORO220111BEb_SAMP.CAL
Sample Weight: 1045 Dilution: 1 S.A.: 1

Mean RF for for Residual Range Organics Calculations: 26424.55
Rt range for Residual Range Organics: 14.32 to 21.77

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*#Triacontane	16.333	.478	.118	24.63

RRO Area:1.364789E+07 RRO AMOUNT: 0.4942441

ERH2529 (RHMW19)

Batch ID: 163748

G:\Org\HP5\DAT\HP5021522_b\0215HP5.0023.RAW

B22020962-011D ;0215HP5 , \$HC-8015-DRO-W,




DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: B22020962-011D ;0215HP5 , \$HC-8015-DRO-W,
 Raw File: G:\Org\HP5\DAT\HP5021522_b\0215HP5.0023.RAW
 Date & Time Acquired: 2/16/2022 2:02:54 AM
 Method File: G:\Org\HP5\Methods\DS_8015-C24T-JEb-L#.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO220111JEb-C24-T.CAL
 Sample Weight: 1045 Dilution: 1 S.A.: 1

Mean RF for TEH: 32675.36
 Rt range for Diesel Range Organics: 6.52 to 14.41

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*o-Terphenyl	12.248	.191	.148	77.37
*1-Chlorooctadecane	13.101	.191	.	.02
*#Triacontane	16.333	.191	.106	55.15

DRO Area:2870218 DRO Amount: 8.405783E-02
 TEH Area:5793790 TEH Amount: 0.1696782



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: DCM-Baseline Check-V24
 Raw File: G:\org\HP5\DAT\HP5021522_b\0215HP5.0024.RAW
 Date & Time Acquired: 2/16/2022 2:45:53 AM
 Method File: G:\Org\HP5\Methods\DR_8015-JA-LEXP.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO220111JA.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for TEH: 32675.36
 Rt range for Diesel Range Organics: 6.63 to 15.99

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*o-Terphenyl	29.945	200.	.	-
*1-Chlorooctadecane	29.945	200.	.	-


DRO Area:149604 DRO Amount: 4.578497
 TEH Area:281437.3 TEH Amount: 8.613134

ERH2531 (RHMW01R)

Batch ID: 163748

G:\Org\HP5\DAT\HP5021522_b\0215HP5.0025.RAW

B22020962-006D ;0215HP5 , \$HC-8015-DRO-W,



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: B22020962-006D ;0215HP5 , \$HC-8015-DRO-W,
Raw File: G:\Org\HP5\DAT\HP5021522_b\0215HP5.0025.RAW
Date & Time Acquired: 2/16/2022 3:28:52 AM
Method File: G:\Org\HP5\Methods\D3_8015-021516-JEb-L%.met
Calibration File: G:\Org\HP5\Cals\SW8015C_DRO220111JEB-C24-T.CAL
Sample Weight: 1050 Dilution: 1 S.A.: 1

Mean RF for TEH: 32675.36

Rt range for Diesel Range Organics: 6.52 to 14.41

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*o-Terphenyl	12.247	.19	.181	95.13	-
*1-Chlorooctadecane	29.979	.19	.	.	-
*#Triacontane	16.331	.19	.12	63.24	-

DRO Area:3.115223E+07 DRO Amount: 0.9079867


TEH Area:5.617084E+07 TEH Amount: 1.637198

ERH2531 (RHMW01R)

Batch ID: 163748

G:\Org\HP5\DAT\HP5021522_b\0215HP5.0025.RAW

B22020962-006D ;0215HP5 , \$HC-8015-DRO-W,



RESIDUAL RANGE ORGANICS CHROMATOGRAM

Sample Name: B22020962-006D ;0215HP5 , \$HC-8015-DRO-W,
Raw File: G:\Org\HP5\DAT\HP5021522_b\0215HP5.0025.RAW
Date & Time Acquired: 2/16/2022 3:28:52 AM
Method File: G:\Org\HP5\Methods\D3_OROS-021516-BEb-L%.MET
Calibration File: G:\Org\HP5\Cals\SW8015C_ORO220111BEb_SAMP.CAL
Sample Weight: 1050 Dilution: 1 S.A.: 1

Mean RF for for Residual Range Organics Calculations: 26424.55
Rt range for Residual Range Organics: 14.32 to 21.77

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*#Triacontane	16.331	.476	.12	25.3


RRO Area: 2.005147E+07 RRO AMOUNT: 0.7226853

ERH2531 (RHMW01R)

Batch ID: 163748

G:\Org\HP5\DAT\HP5021522_b\0215HP5.0025.RAW

B22020962-006D ;0215HP5 , \$HC-8015-DRO-W,



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT


Sample Name: B22020962-006D ;0215HP5 , \$HC-8015-DRO-W,
Raw File: G:\Org\HP5\DAT\HP5021522_b\0215HP5.0025.RAW
Date & Time Acquired: 2/16/2022 3:28:52 AM
Method File: G:\Org\HP5\Methods\DS_8015-C24T-JEb-L#.met
Calibration File: G:\Org\HP5\Cals\SW8015C_DRO220111JEB-C24-T.CAL
Sample Weight: 1050 Dilution: 1 S.A.: 1

Mean RF for TEH: 32675.36

Rt range for Diesel Range Organics: 6.52 to 14.41

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*o-Terphenyl	12.247	.19	.165	86.79	-
*1-Chlorooctadecane	29.979	.19	.	.	-
*#Triacontane	16.331	.19	.103	54.18	-

DRO Area:1.24974E+07 DRO Amount: 0.3642586
TEH Area:2.131612E+07 TEH Amount: 0.6212959



RESIDUAL RANGE ORGANICS CHROMATOGRAM


Sample Name: B22020962-006DMS-RRO ;0215HP5 ,
 Raw File: G:\org\HP5\DAT\HP5021522_b\0215HP5.0026.RAW
 Date & Time Acquired: 2/16/2022 4:11:53 AM
 Method File: G:\Org\HP5\Methods\D3_ORO-BE-L%.MET
 Calibration File: G:\Org\HP5\Cals\SW8015C_ORO220111BE.CAL
 Sample Weight: 1050 Dilution: 1 S.A.: 1

Mean RF for for ~~Residual~~ TEH(Oil Range) Organics Calculations: 26424.55
 Rt range for ~~Residual~~ TEH(Oil Range) Organics: 12.61 to 30.05

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*#Triacontane	16.336	.476	.188	39.53	-

~~RRO~~ TEH(Oil Range) Area:1.449824E+08 ~~RRO~~ TEH(Oil Range) AMOUNT: 5.225386

AMN 02/18/2022




RESIDUAL RANGE ORGANICS CHROMATOGRAM

Sample Name: B22020962-006DMS-RRO ;0215HP5 ,
 Raw File: G:\org\HP5\DAT\HP5021522_b\0215HP5.0026.RAW
 Date & Time Acquired: 2/16/2022 4:11:53 AM
 Method File: G:\Org\HP5\Methods\DS_ORO-BE-L%.MET
 Calibration File: G:\Org\HP5\Cals\SW8015C_ORO220111BE.CAL
 Sample Weight: 1050 Dilution: 1 S.A.: 1

Mean RF for for Residual Range Organics Calculations: 26424.55
 Rt range for Residual Range Organics: 12.61 to 30.05

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*#Triacontane_____	16.336	.476	.1	21.01 -

RRO Area:3530088 RRO AMOUNT: 0.1272297




DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: DCM-Baseline Check-V27
 Raw File: G:\org\HP5\DAT\HP5021522_b\0215HP5.0027.RAW
 Date & Time Acquired: 2/16/2022 4:54:47 AM
 Method File: G:\Org\HP5\Methods\DR_8015-JA-LEXP.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO220111JA.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for TEH: 32675.36
 Rt range for Diesel Range Organics: 6.63 to 15.99

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*o-Terphenyl	29.879	200.	.	-
*1-Chlorooctadecane	29.879	200.	.	-

DRO Area:146151.7 DRO Amount: 4.472841
 TEH Area:708222.8 TEH Amount: 21.67452



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: DCM-Baseline Check-V28
 Raw File: G:\org\HP5\DAT\HP5021522_b\0215HP5.0028.RAW
 Date & Time Acquired: 2/16/2022 5:37:44 AM
 Method File: G:\Org\HP5\Methods\DR_8015-JA-LEXP.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO220111JA.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for TEH: 32675.36
 Rt range for Diesel Range Organics: 6.63 to 15.99

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*o-Terphenyl	29.919	200.	.	-
*1-Chlorooctadecane	29.919	200.	.	-


DRO Area:79279.07 DRO Amount: 2.426265
 TEH Area:181465.6 TEH Amount: 5.55359

ERH2535 (RHMW2254-01) Low Flow

Batch ID: 163748

G:\org\HP5\DAT\HP5021522_b\0215HP5.0029.RAW

B22020962-016D ;0215HP5 , \$HC-8015-DRO-W,



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: B22020962-016D ;0215HP5 , \$HC-8015-DRO-W,
Raw File: G:\org\HP5\DAT\HP5021522_b\0215HP5.0029.RAW
Date & Time Acquired: 2/16/2022 6:20:41 AM
Method File: G:\Org\HP5\Methods\DR_8015-C24Tb-JE-L%.met
Calibration File: G:\Org\HP5\Cals\SW8015C_DRO220111JE-C24-T.CAL
Sample Weight: 1010 Dilution: 1 S.A.: 1

Mean RF for TEH: 32675.36

Rt range for Diesel Range Organics: 6.56 to 14.42

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*o-Terphenyl	12.248	.198	.214	108.13	-
*1-Chlorooctadecane	13.106	.198	.	.03	-
*#Triacontane	16.331	.198	.115	58.2	-


DRO Area:568843.7 DRO Amount: 1.723658E-02
TEH Area:840378.4 TEH Amount: 2.546438E-02

ERH2535 (RHMW2254-01) Low Flow

Batch ID: 163748

G:\Org\HP5\DAT\HP5021522_b\0215HP5.0029.RAW

B22020962-016D ;0215HP5 , \$HC-8015-DRO-W,



RESIDUAL RANGE ORGANICS CHROMATOGRAM

Sample Name: B22020962-016D ;0215HP5 , \$HC-8015-DRO-W,
Raw File: G:\Org\HP5\DAT\HP5021522_b\0215HP5.0029.RAW
Date & Time Acquired: 2/16/2022 6:20:41 AM
Method File: G:\Org\HP5\Methods\DR_OROS-BEb-L%.MET
Calibration File: G:\Org\HP5\Cals\SW8015C_ORO220111BEb_SAMP.CAL
Sample Weight: 1010 Dilution: 1 S.A.: 1

Mean RF for for Residual Range Organics Calculations: 26424.55
Rt range for Residual Range Organics: 14.32 to 21.77

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*#Triacontane_____	16.331	.495	.115	23.28


RRO Area:177785.6 RRO AMOUNT: 6.661431E-03

ERH2535 (RHMW2254-01) Low Flow

Batch ID: 163748

G:\Org\HP5\DAT\HP5021522_b\0215HP5.0029.RAW

B22020962-016D ;0215HP5 , \$HC-8015-DRO-W,



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT


Sample Name: B22020962-016D ;0215HP5 , \$HC-8015-DRO-W,
Raw File: G:\Org\HP5\DAT\HP5021522_b\0215HP5.0029.RAW
Date & Time Acquired: 2/16/2022 6:20:41 AM
Method File: G:\Org\HP5\Methods\DS_8015-C24T-JEb-L#.met
Calibration File: G:\Org\HP5\Cals\SW8015C_DRO220111JEB-C24-T.CAL
Sample Weight: 1010 Dilution: 1 S.A.: 1

Mean RF for TEH: 32675.36

Rt range for Diesel Range Organics: 6.52 to 14.41

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*o-Terphenyl	12.248	.198	.214	107.92
*1-Chlorooctadecane	13.106	.198	.01	-
*#Triacontane	16.331	.198	.115	57.9

DRO Area:410844.8 DRO Amount: 1.244904E-02
TEH Area:871540.6 TEH Amount: 2.640863E-02




DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: DCM-Baseline Check-V30
 Raw File: G:\org\HP5\DAT\HP5021522_b\0215HP5.0030.RAW
 Date & Time Acquired: 2/16/2022 7:03:40 AM
 Method File: G:\Org\HP5\Methods\DR_8015-JA-LEXP.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO220111JA.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for TEH: 32675.36
 Rt range for Diesel Range Organics: 6.63 to 15.99

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*o-Terphenyl	29.98	200.	.	-
*1-Chlorooctadecane	29.98	200.	.	-

DRO Area:228848.1 DRO Amount: 7.00369
 TEH Area:409650.3 TEH Amount: 12.53698




DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: DCM-Baseline Check-V31
 Raw File: G:\org\HP5\DAT\HP5021522_b\0215HP5.0031.RAW
 Date & Time Acquired: 2/16/2022 7:46:38 AM
 Method File: G:\Org\HP5\Methods\DR_8015-JA-LEXP.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO220111JA.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for TEH: 32675.36
 Rt range for Diesel Range Organics: 6.63 to 15.99

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*o-Terphenyl	29.984	200.	.	-
*1-Chlorooctadecane	29.984	200.	.	-

DRO Area:180408.3 DRO Amount: 5.521234
 TEH Area:291673.2 TEH Amount: 8.926396



RESIDUAL RANGE ORGANICS CHROMATOGRAM


Sample Name: LCS-163748-RRO ;0215HP5 ,
 Raw File: G:\org\HP5\DAT\HP5021522_b\0215HP5.0032.RAW
 Date & Time Acquired: 2/16/2022 8:29:23 AM
 Method File: G:\Org\HP5\Methods\D3_ORO-BE-L%.MET
 Calibration File: G:\Org\HP5\Cals\SW8015C_ORO220111BE.CAL
 Sample Weight: 1000 Dilution: 1 S.A.: 1

Mean RF for for ~~Residual~~ TEH(Oil Range) Organics Calculations: 26424.55
 Rt range for ~~Residual~~ TEH(Oil Range) Organics: 12.61 to 30.05

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*#Triacontane	16.333	.5	.204	40.89	-

~~RRO~~ TEH(Oil Range) Area:1.455688E+08 ~~RRO~~ TEH(Oil Range) AMOUNT: 5.508848

AMN 02/18/2022




RESIDUAL RANGE ORGANICS CHROMATOGRAM

Sample Name: LCS-163748-RRO ;0215HP5 ,
 Raw File: G:\org\HP5\DAT\HP5021522_b\0215HP5.0032.RAW
 Date & Time Acquired: 2/16/2022 8:29:23 AM
 Method File: G:\Org\HP5\Methods\DS_ORO-BE-L%.MET
 Calibration File: G:\Org\HP5\Cals\SW8015C_ORO220111BE.CAL
 Sample Weight: 1000 Dilution: 1 S.A.: 1

Mean RF for for Residual Range Organics Calculations: 26424.55
 Rt range for Residual Range Organics: 12.61 to 30.05

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*#Triacontane	16.333	.5	.112	22.49

RRO Area:3495212 RRO AMOUNT: 0.1322714




DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: DCM-Baseline Check-V33
 Raw File: G:\org\HP5\DAT\HP5021522_b\0215HP5.0033.RAW
 Date & Time Acquired: 2/16/2022 9:12:08 AM
 Method File: G:\Org\HP5\Methods\DR_8015-JA-LEXP.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO220111JA.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for TEH: 32675.36
 Rt range for Diesel Range Organics: 6.63 to 15.99

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*o-Terphenyl	29.886	200.	.	-
*1-Chlorooctadecane	29.886	200.	.	-

DRO Area:143342.9 DRO Amount: 4.38688
 TEH Area:713584.4 TEH Amount: 21.83861



RESIDUAL RANGE ORGANICS CHROMATOGRAM


Sample Name: LCSD-163748-RRO ;0215HP5 ,
 Raw File: G:\org\HP5\DAT\HP5021522_b\0215HP5.0034.RAW
 Date & Time Acquired: 2/16/2022 9:54:56 AM
 Method File: G:\Org\HP5\Methods\D3_ORO-BE-L%.MET
 Calibration File: G:\Org\HP5\Cals\SW8015C_ORO220111BE.CAL
 Sample Weight: 1000 Dilution: 1 S.A.: 1

Mean RF for for ~~Residual~~ TEH(Oil Range) Organics Calculations: 26424.55
 Rt range for ~~Residual~~ TEH(Oil Range) Organics: 12.61 to 30.05

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*#Triacontane	16.331	.5	.209	41.71

RRO TEH(Oil Range) Area:1.383285E+08 RRO TEH(Oil Range) AMOUNT: 5.234846

AMN 02/18/2022




RESIDUAL RANGE ORGANICS CHROMATOGRAM


Sample Name: LCSD-163748-RRO ;0215HP5 ,
 Raw File: G:\org\HP5\DAT\HP5021522_b\0215HP5.0034.RAW
 Date & Time Acquired: 2/16/2022 9:54:56 AM
 Method File: G:\Org\HP5\Methods\DS_ORO-BE-L%.MET
 Calibration File: G:\Org\HP5\Cals\SW8015C_ORO220111BE.CAL
 Sample Weight: 1000 Dilution: 1 S.A.: 1

Mean RF for for Residual Range Organics Calculations: 26424.55
 Rt range for Residual Range Organics: 12.61 to 30.05

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*#Triacontane_____	16.331	.5	.108	21.52

RRO Area:3196335 RRO AMOUNT: 0.1209608





RESIDUAL RANGE ORGANICS CHROMATOGRAM

Sample Name: CCV_0215HP536r, RRO ;0215HP5 , DRO220201A
 Raw File: G:\org\HP5\DAT\HP5021522_b\0215HP5.0036.RAW
 Date & Time Acquired: 2/16/2022 11:20:37 AM
 Method File: G:\Org\HP5\Methods\DC_ORO-BE-L%.MET
 Calibration File: G:\Org\HP5\Cals\SW8015C_ORO220111BE.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for for ~~Residual~~ TEH(Oil Range) Organics Calculations: 26424.55
 Rt range for ~~Residual~~ TEH(Oil Range) Organics: 12.61 to 30.05

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*#Triacontane	16.333	500.	333.835	66.77	-


~~RRO~~ TEH(Oil Range) Area:1.404836E+08 ~~RRO~~ TEH(Oil Range) AMOUNT: 5316.405

CONTINUING CALIBRATION REPORT: G:\org\HP5\DAT\HP5021522_b\0215HP5.0036.RAW

COMPOUND	ACTUAL (NG)	MEASURED (NG)	%RECOVERY	LIMITS
*30-40 Motor Oil	5000.	.035	.	75-125

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	LIMITS
*#Triacontane	16.333	200.	333.835	166.92	75-125

AMN 02/18/2022



RESIDUAL RANGE ORGANICS CHROMATOGRAM

Sample Name: CCV_0215HP536r, RRO ;0215HP5 , DRO220201A
 Raw File: G:\org\HP5\DAT\HP5021522_b\0215HP5.0036.RAW
 Date & Time Acquired: 2/16/2022 11:20:37 AM
 Method File: G:\Org\HP5\Methods\DS_ORO-BE-L%.MET
 Calibration File: G:\Org\HP5\Cals\SW8015C_ORO220111BE.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for for Residual Range Organics Calculations: 26424.55
 Rt range for Residual Range Organics: 12.61 to 30.05


SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*#Triacontane	16.333	500.	195.34	39.07	-

RRO Area:3682178 RRO AMOUNT: 139.3468

CONTINUING CALIBRATION REPORT: G:\org\HP5\DAT\HP5021522_b\0215HP5.0036.RAW

COMPOUND	ACTUAL (NG)	MEASURED (NG)	%RECOVERY	LIMITS
*30-40 Motor Oil	5000.	.035	.	75-125

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	LIMITS
*#Triacontane	16.333	200.	195.34	97.67	75-125



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: CCV_0215HP537r, DRO ;0215HP5 , DRO220128A
 Raw File: G:\org\HP5\DAT\HP5021522_b\0215HP5.0037.RAW
 Date & Time Acquired: 2/16/2022 12:03:21 PM
 Method File: G:\Org\HP5\Methods\DC_8015-C24-JEb-L%.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO220111JEB-C24.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for TEH: 32675.36
 Rt range for Diesel Range Organics: 6.52 to 14.41

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*o-Terphenyl	12.251	200.	308.255	154.13
*1-Chlorooctadecane	13.1	200.	150.52	75.26

DRO Area: 4.726423E+08 DRO Amount: 14464.79
 TEH Area: 4.89131E+08 TEH Amount: 14969.41


CONTINUING CALIBRATION REPORT: G:\org\HP5\DAT\HP5021522_b\0215HP5.0037.RAW

COMPOUND	ACTUAL (NG)	MEASURED (NG)	%RECOVERY	LIMITS
TOTAL DRO	15000.	14969.41	99.8	85-115

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	LIMITS
*o-Terphenyl	12.251	200.	308.255	154.13	85-115
*1-Chlorooctadecane	13.1	200.	150.52	75.26	85-115

G:\org\HP5\DAT\HP5021522_b\0215HP5.0037.RAW

CCV_0215HP537r, DRO ;0215HP5 , DRO220128A



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: CCV_0215HP537r, DRO ;0215HP5 , DRO220128A
 Raw File: G:\org\HP5\DAT\HP5021522_b\0215HP5.0037.RAW
 Date & Time Acquired: 2/16/2022 12:03:21 PM
 Method File: G:\Org\HP5\Methods\DS_8015-C24-JEb-L#.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO220111JEB-C24.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for TEH: 32675.36

Rt range for Diesel Range Organics: 6.52 to 14.41


SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*o-Terphenyl	12.251	200.	184.203	92.1
*1-Chlorooctadecane	13.1	200.	35.818	17.91


DRO Area: 2.441834E+08 DRO Amount: 7473.015
 TEH Area: 2.549449E+08 TEH Amount: 7802.36

CONTINUING CALIBRATION REPORT: G:\org\HP5\DAT\HP5021522_b\0215HP5.0037.RAW

COMPOUND	ACTUAL (NG)	MEASURED (NG)	%RECOVERY	LIMITS
TOTAL DRO	15000.	7802.36	52.02	85-115

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	LIMITS
*o-Terphenyl	12.251	200.	184.203	92.1	85-115
*1-Chlorooctadecane	13.1	200.	35.818	17.91	85-115





RESIDUAL RANGE ORGANICS CHROMATOGRAM

Sample Name: CCV_0215HP536r, RRO ;0215HP5 , DRO220201A
 Raw File: G:\org\HP5\DAT\HP5021522_b\0215HP5.0036.RAW
 Date & Time Acquired: 2/16/2022 11:20:37 AM
 Method File: G:\Org\HP5\Methods\DC_ORO-BE-L%.MET
 Calibration File: G:\Org\HP5\Cals\SW8015C_ORO220111BE.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for for ~~Residual~~ TEH(Oil Range) Organics Calculations: 26424.55
 Rt range for ~~Residual~~ TEH(Oil Range) Organics: 12.61 to 30.05

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*#Triacontane	16.333	500.	333.835	66.77	-


RRO TEH(Oil Range) Area:1.404836E+08 RRO TEH(Oil Range) AMOUNT: 5316.405

CONTINUING CALIBRATION REPORT: G:\org\HP5\DAT\HP5021522_b\0215HP5.0036.RAW

COMPOUND	ACTUAL (NG)	MEASURED (NG)	%RECOVERY	LIMITS
*30-40 Motor Oil	5000.	.035	.	75-125

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	LIMITS
*#Triacontane	16.333	200.	333.835	166.92	75-125

AMN 02/18/2022



RESIDUAL RANGE ORGANICS CHROMATOGRAM

Sample Name: CCV_0215HP536r, RRO ;0215HP5 , DRO220201A
 Raw File: G:\org\HP5\DAT\HP5021522_b\0215HP5.0036.RAW
 Date & Time Acquired: 2/16/2022 11:20:37 AM
 Method File: G:\Org\HP5\Methods\DS_ORO-BE-L%.MET
 Calibration File: G:\Org\HP5\Cals\SW8015C_ORO220111BE.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for for Residual Range Organics Calculations: 26424.55
 Rt range for Residual Range Organics: 12.61 to 30.05


SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*#Triacontane	16.333	500.	195.34	39.07	-

RRO Area:3682178 RRO AMOUNT: 139.3468

CONTINUING CALIBRATION REPORT: G:\org\HP5\DAT\HP5021522_b\0215HP5.0036.RAW

COMPOUND	ACTUAL (NG)	MEASURED (NG)	%RECOVERY	LIMITS
*30-40 Motor Oil	5000.	.035	.	75-125

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	LIMITS
*#Triacontane	16.333	200.	195.34	97.67	75-125



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: CCV_0215HP537r, DRO ;0215HP5 , DRO220128A
 Raw File: G:\org\HP5\DAT\HP5021522_b\0215HP5.0037.RAW
 Date & Time Acquired: 2/16/2022 12:03:21 PM
 Method File: G:\Org\HP5\Methods\DC_8015-C24-JEb-L%.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO220111JEB-C24.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for TEH: 32675.36
 Rt range for Diesel Range Organics: 6.52 to 14.41

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*o-Terphenyl	12.251	200.	308.255	154.13	-
*1-Chlorooctadecane	13.1	200.	150.52	75.26	-

DRO Area: 4.726423E+08 DRO Amount: 14464.79
 TEH Area: 4.89131E+08 TEH Amount: 14969.41


CONTINUING CALIBRATION REPORT: G:\org\HP5\DAT\HP5021522_b\0215HP5.0037.RAW

COMPOUND	ACTUAL (NG)	MEASURED (NG)	%RECOVERY	LIMITS
TOTAL DRO	15000.	14969.41	99.8	85-115

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	LIMITS
*o-Terphenyl	12.251	200.	308.255	154.13	85-115
*1-Chlorooctadecane	13.1	200.	150.52	75.26	85-115

G:\org\HP5\DAT\HP5021522_b\0215HP5.0037.RAW

CCV_0215HP537r, DRO ;0215HP5 , DRO220128A



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: CCV_0215HP537r, DRO ;0215HP5 , DRO220128A
 Raw File: G:\org\HP5\DAT\HP5021522_b\0215HP5.0037.RAW
 Date & Time Acquired: 2/16/2022 12:03:21 PM
 Method File: G:\Org\HP5\Methods\DS_8015-C24-JEb-L#.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO220111JEB-C24.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for TEH: 32675.36

Rt range for Diesel Range Organics: 6.52 to 14.41


SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*o-Terphenyl	12.251	200.	184.203	92.1
*1-Chlorooctadecane	13.1	200.	35.818	17.91

DRO Area: 2.441834E+08 DRO Amount: 7473.015
 TEH Area: 2.549449E+08 TEH Amount: 7802.36

CONTINUING CALIBRATION REPORT: G:\org\HP5\DAT\HP5021522_b\0215HP5.0037.RAW

COMPOUND	ACTUAL (NG)	MEASURED (NG)	%RECOVERY	LIMITS
TOTAL DRO	15000.	7802.36	52.02	85-115

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	LIMITS
*o-Terphenyl	12.251	200.	184.203	92.1	85-115
*1-Chlorooctadecane	13.1	200.	35.818	17.91	85-115




DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: DCM-Baseline Check-V38
 Raw File: G:\org\HP5\DAT\HP5021522_b\0215HP5.0038.RAW
 Date & Time Acquired: 2/16/2022 12:46:01 PM
 Method File: G:\Org\HP5\Methods\DR_8015-JA-LEXP.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO220111JA.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for TEH: 32675.36
 Rt range for Diesel Range Organics: 6.63 to 15.99

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*o-Terphenyl	29.93	200.	.	-
*1-Chlorooctadecane	29.93	200.	.	-

DRO Area:87362.3 DRO Amount: 2.673645
 TEH Area:216466 TEH Amount: 6.624748




DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: DCM-Baseline Check-V39
 Raw File: G:\org\HP5\DAT\HP5021522_b\0215HP5.0039.RAW
 Date & Time Acquired: 2/16/2022 1:29:02 PM
 Method File: G:\Org\HP5\Methods\DR_8015-JA-LEXP.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO220111JA.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for TEH: 32675.36
 Rt range for Diesel Range Organics: 6.63 to 15.99

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*o-Terphenyl	12.315	200.	.036	.02 -
*1-Chlorooctadecane	29.868	200.	.	. -

DRO Area:79034.05 DRO Amount: 2.418766
 TEH Area:136178.5 TEH Amount: 4.167621



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: LCS-163748 ;0215HP5 , SGT
 Raw File: G:\org\HP5\DAT\HP5021522_b\0215HP5.0040.RAW
 Date & Time Acquired: 2/16/2022 2:12:10 PM
 Method File: G:\Org\HP5\Methods\D3_8015-021540-JEb-L%.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO220111JEb-C24.CAL
 Sample Weight: 1000 Dilution: 1 S.A.: 1

Mean RF for TEH: 32675.36
 Rt range for Diesel Range Organics: 6.52 to 14.41


SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*o-Terphenyl	12.25	.2	.338	168.78	-
*1-Chlorooctadecane	13.101	.2	.149	74.7	-

DRO Area: 4.168957E+08 DRO Amount: 12.75872
 TEH Area: 4.417582E+08 TEH Amount: 13.51961

Batch ID: 163748

G:\Org\HP5\DAT\HP5021522_b\0215HP5.0040.RAW

LCS-163748 ;0215HP5 , SGT



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: LCS-163748 ;0215HP5 , SGT
 Raw File: G:\Org\HP5\DAT\HP5021522_b\0215HP5.0040.RAW
 Date & Time Acquired: 2/16/2022 2:12:10 PM
 Method File: G:\Org\HP5\Methods\DS_8015-C24-JEb-L#.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO220111JEB-C24.CAL
 Sample Weight: 1000 Dilution: 1 S.A.: 1

Mean RF for TEH: 32675.36
 Rt range for Diesel Range Organics: 6.52 to 14.41


SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*o-Terphenyl	12.25	.2	.204	101.76
*1-Chlorooctadecane	13.101	.2	.055	27.66

DRO Area: 1.894617E+08 DRO Amount: 5.798304
 TEH Area: 1.999649E+08 TEH Amount: 6.119746

Batch ID: 163748

LCSD-163748 ;0215HP5 , SGT

G:\Org\HP5\DAT\HP5021522_b\0215HP5.0041.RAW



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT


Sample Name: LCSD-163748 ;0215HP5 , SGT
 Raw File: G:\Org\HP5\DAT\HP5021522_b\0215HP5.0041.RAW
 Date & Time Acquired: 2/16/2022 2:54:46 PM
 Method File: G:\Org\HP5\Methods\D3_8015-C24-JEb-L%.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO220111JEB-C24.CAL
 Sample Weight: 1000 Dilution: 1 S.A.: 1

Mean RF for TEH: 32675.36

Rt range for Diesel Range Organics: 6.52 to 14.41

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*o-Terphenyl	12.247	.2	.312	156.16	-
*1-Chlorooctadecane	13.099	.2	.186	92.9	-

DRO Area: 3.808281E+08 DRO Amount: 11.6549
 TEH Area: 4.045568E+08 TEH Amount: 12.3811




DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: LCSD-163748 ;0215HP5 , SGT
 Raw File: G:\org\HP5\DAT\HP5021522_b\0215HP5.0041.RAW
 Date & Time Acquired: 2/16/2022 2:54:46 PM
 Method File: G:\Org\HP5\Methods\DS_8015-C24-JEb-L#.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO220111JEb-C24.CAL
 Sample Weight: 1000 Dilution: 1 S.A.: 1

Mean RF for TEH: 32675.36
 Rt range for Diesel Range Organics: 6.52 to 14.41

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*o-Terphenyl	12.247	.2	.189	94.37
*1-Chlorooctadecane	13.099	.2	.051	25.31

DRO Area: 1.75291E+08 DRO Amount: 5.364624
 TEH Area: 1.858709E+08 TEH Amount: 5.688413



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: MB-163748 ;0215HP5 , SGT
 Raw File: G:\Org\HP5\DAT\HP5021522_b\0215HP5.0042.RAW
 Date & Time Acquired: 2/16/2022 3:37:30 PM
 Method File: G:\Org\HP5\Methods\DR_8015-C24Tb-JEb-L%.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO220111JEb-C24-T.CAL
 Sample Weight: 1000 Dilution: 1 S.A.: 1


Mean RF for TEH: 32675.36
 Rt range for Diesel Range Organics: 6.52 to 14.41

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*o-Terphenyl	12.243	.2	.197	98.74
*1-Chlorooctadecane	13.074	.2	.2	-
*#Triacontane	16.329	.2	.101	50.5

DRO Area:354511.5 DRO Amount: 0.0108495
 TEH Area:611334.3 TEH Amount: 1.870934E-02

G:\Org\HP5\DAT\HP5021522_b\0215HP5.0042.RAW

MB-163748 ;0215HP5 , SGT



RESIDUAL RANGE ORGANICS CHROMATOGRAM

Sample Name: MB-163748 ;0215HP5 , SGT
 Raw File: G:\org\HP5\DAT\HP5021522_b\0215HP5.0042.RAW
 Date & Time Acquired: 2/16/2022 3:37:30 PM
 Method File: G:\Org\HP5\Methods\DR_OROS-BEb-L%.MET
 Calibration File: G:\Org\HP5\Cals\SW8015C_ORO220111BEb_SAMP.CAL
 Sample Weight: 1000 Dilution: 1 S.A.: 1

Mean RF for for Residual Range Organics Calculations: 26424.55
 Rt range for Residual Range Organics: 14.32 to 21.77


SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*#Triacontane_____	16.329	.5	.101	20.21

RRO Area:178422.5 RRO AMOUNT: 6.752147E-03

Batch ID: 163748

MB-163748 ;0215HP5 , SGT

G:\Org\HP5\DAT\HP5021522_b\0215HP5.0042.RAW



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: MB-163748 ;0215HP5 , SGT
 Raw File: G:\Org\HP5\DAT\HP5021522_b\0215HP5.0042.RAW
 Date & Time Acquired: 2/16/2022 3:37:30 PM
 Method File: G:\Org\HP5\Methods\DS_8015-C24T-JEb-L#.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO220111JEB-C24-T.CAL
 Sample Weight: 1000 Dilution: 1 S.A.: 1

Mean RF for TEH: 32675.36

Rt range for Diesel Range Organics: 6.52 to 14.41

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*o-Terphenyl	12.243	.2	.197	98.47
*1-Chlorooctadecane	13.074	.2	.	.12
*#Triacontane	16.329	.2	.1	50.18


DRO Area:191846.4 DRO Amount: 5.871288E-03
 TEH Area:454867.8 TEH Amount: 1.392082E-02

ERH2524 (OWDFMW07A)

G:\org\HP5\DAT\HP5021522_b\0215HP5.0043.RAW

Batch ID: 163748

B22020962-001D ;0215HP5 , \$HC-8015-DRO-W, SGT



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: B22020962-001D ;0215HP5 , \$HC-8015-DRO-W, SGT
 Raw File: G:\org\HP5\DAT\HP5021522_b\0215HP5.0043.RAW
 Date & Time Acquired: 2/16/2022 4:20:20 PM
 Method File: G:\Org\HP5\Methods\DR_8015-C24Tb-JEb-L%.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO220111JEb-C24-T.CAL
 Sample Weight: 1050 Dilution: 1 S.A.: 1

Mean RF for TEH: 32675.36

Rt range for Diesel Range Organics: 6.52 to 14.41

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*o-Terphenyl	12.246	.19	.185	96.93	-
*1-Chlorooctadecane	13.074	.19	.	.25	-
*#Triacontane	16.328	.19	.095	49.62	-


DRO Area:329790.6 DRO Amount: 9.612327E-03
 TEH Area:527998.8 TEH Amount: 1.538946E-02

ERH2524 (OWDFMW07A)

Batch ID: 163748

G:\Org\HP5\DAT\HP5021522_b\0215HP5.0043.RAW

B22020962-001D ;0215HP5 , \$HC-8015-DRO-W, SGT



RESIDUAL RANGE ORGANICS CHROMATOGRAM

Sample Name: B22020962-001D ;0215HP5 , \$HC-8015-DRO-W, SGT
 Raw File: G:\Org\HP5\DAT\HP5021522_b\0215HP5.0043.RAW
 Date & Time Acquired: 2/16/2022 4:20:20 PM
 Method File: G:\Org\HP5\Methods\DR_OROS-BEb-L%.MET
 Calibration File: G:\Org\HP5\Cals\SW8015C_ORO220111BEb_SAMP.CAL
 Sample Weight: 1050 Dilution: 1 S.A.: 1

Mean RF for for Residual Range Organics Calculations: 26424.55
 Rt range for Residual Range Organics: 14.32 to 21.77

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*#Triacontane_____	16.328	.476	.095	19.87


RRO Area:148512.9 RRO AMOUNT: 5.352631E-03

ERH2524 (OWDFMW07A)

Batch ID: 163748

G:\Org\HP5\DAT\HP5021522_b\0215HP5.0043.RAW

B22020962-001D ;0215HP5 , \$HC-8015-DRO-W, SGT



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: B22020962-001D ;0215HP5 , \$HC-8015-DRO-W, SGT
 Raw File: G:\Org\HP5\DAT\HP5021522_b\0215HP5.0043.RAW
 Date & Time Acquired: 2/16/2022 4:20:20 PM
 Method File: G:\Org\HP5\Methods\DS_8015-C24T-JEb-L#.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO220111JEB-C24-T.CAL
 Sample Weight: 1050 Dilution: 1 S.A.: 1

Mean RF for TEH: 32675.36

Rt range for Diesel Range Organics: 6.52 to 14.41


SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*o-Terphenyl	12.246	.19	.184	96.58	-
*1-Chlorooctadecane	13.074	.19	.	.13	-
*Triacontane	16.328	.19	.094	49.3	-

DRO Area:157078.4 DRO Amount: 4.578327E-03
 TEH Area:355845.6 TEH Amount: 1.037175E-02

Batch ID: 163748

B22020962-001DMS ;0215HP5 , SGT

G:\org\HP5\DAT\HP5021522_b\0215HP5.0044.RAW



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: B22020962-001DMS ;0215HP5 , SGT
 Raw File: G:\org\HP5\DAT\HP5021522_b\0215HP5.0044.RAW
 Date & Time Acquired: 2/16/2022 5:03:15 PM
 Method File: G:\Org\HP5\Methods\D3_8015-021540-JEb-L%.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO220111JEb-C24.CAL
 Sample Weight: 1050 Dilution: 1 S.A.: 1

Mean RF for TEH: 32675.36

Rt range for Diesel Range Organics: 6.52 to 14.41


SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*o-Terphenyl	12.247	.19	.314	164.87	-
*1-Chlorooctadecane	13.098	.19	.138	72.21	-

DRO Area: 4.060861E+08 DRO Amount: 11.8361
 TEH Area: 4.302156E+08 TEH Amount: 12.53939

Batch ID: 163748

G:\Org\HP5\DAT\HP5021522_b\0215HP5.0044.RAW

B22020962-001DMS ;0215HP5 , SGT



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: B22020962-001DMS ;0215HP5 , SGT
 Raw File: G:\Org\HP5\DAT\HP5021522_b\0215HP5.0044.RAW
 Date & Time Acquired: 2/16/2022 5:03:15 PM
 Method File: G:\Org\HP5\Methods\DS_8015-C24-JEb-L#.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO220111JEb-C24.CAL
 Sample Weight: 1050 Dilution: 1 S.A.: 1

Mean RF for TEH: 32675.36

Rt range for Diesel Range Organics: 6.52 to 14.41


SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*o-Terphenyl	12.247	.19	.187	98.29	-
*1-Chlorooctadecane	13.098	.19	.051	26.62	-

DRO Area:1.854669E+08

DRO Amount: 5.40576

TEH Area:1.960038E+08

TEH Amount: 5.712876



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: DCM-Baseline Check-V45
 Raw File: G:\org\HP5\DAT\HP5021522_b\0215HP5.0045.RAW
 Date & Time Acquired: 2/16/2022 5:46:07 PM
 Method File: G:\Org\HP5\Methods\DR_8015-JA-LEXP.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO220111JA.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for TEH: 32675.36
 Rt range for Diesel Range Organics: 6.63 to 15.99

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*o-Terphenyl	12.354	200.	.048	.02
*1-Chlorooctadecane	29.915	200.	.	.


DRO Area: 81007.41 DRO Amount: 2.479159
 TEH Area: 123701.9 TEH Amount: 3.785787

ERH2533 (RHMW2254-01-Bailer)

Batch ID: 163748

G:\Org\HP5\DAT\HP5021522_b\0215HP5.0046.RAW

B22020962-021D ;0215HP5 , \$HC-8015-DRO-W, SGT



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: B22020962-021D ;0215HP5 , \$HC-8015-DRO-W, SGT
 Raw File: G:\Org\HP5\DAT\HP5021522_b\0215HP5.0046.RAW
 Date & Time Acquired: 2/16/2022 6:28:59 PM
 Method File: G:\Org\HP5\Methods\DR_8015-C24Tb-JEb-L%.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO220111JEb-C24-T.CAL
 Sample Weight: 1045 Dilution: 1 S.A.: 1

Mean RF for TEH: 32675.36
 Rt range for Diesel Range Organics: 6.52 to 14.41

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*o-Terphenyl	12.244	.191	.18	94.21	-
*1-Chlorooctadecane	13.081	.191	.	.07	-
*#Triacontane	16.326	.191	.09	47.19	-


DRO Area:1017510 DRO Amount: 2.979903E-02
 TEH Area:1335184 TEH Amount: 3.910249E-02

ERH2533 (RHMW2254-01-Bailer)

Batch ID: 163748

G:\org\HP5\DAT\HP5021522_b\0215HP5.0046.RAW

B22020962-021D ;0215HP5 , \$HC-8015-DRO-W, SGT



RESIDUAL RANGE ORGANICS CHROMATOGRAM

Sample Name: B22020962-021D ;0215HP5 , \$HC-8015-DRO-W, SGT
 Raw File: G:\org\HP5\DAT\HP5021522_b\0215HP5.0046.RAW
 Date & Time Acquired: 2/16/2022 6:28:59 PM
 Method File: G:\Org\HP5\Methods\DR_OROS-BEb-L%.MET
 Calibration File: G:\Org\HP5\Cals\SW8015C_ORO220111BEb_SAMP.CAL
 Sample Weight: 1045 Dilution: 1 S.A.: 1

Mean RF for for Residual Range Organics Calculations: 26424.55
 Rt range for Residual Range Organics: 14.32 to 21.77

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*#Triacontane_____	16.326	.478	.09	18.91	-

RRO Area:140164 RRO AMOUNT: 5.075896E-03

ERH2533 (RHMW2254-01-Bailer)

Batch ID: 163748

G:\Org\HP5\DAT\HP5021522_b\0215HP5.0046.RAW

B22020962-021D ;0215HP5 , \$HC-8015-DRO-W, SGT



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: B22020962-021D ;0215HP5 , \$HC-8015-DRO-W, SGT
 Raw File: G:\Org\HP5\DAT\HP5021522_b\0215HP5.0046.RAW
 Date & Time Acquired: 2/16/2022 6:28:59 PM
 Method File: G:\Org\HP5\Methods\DS_8015-C24T-JEb-L#.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO220111JEB-C24-T.CAL
 Sample Weight: 1045 Dilution: 1 S.A.: 1

Mean RF for TEH: 32675.36

Rt range for Diesel Range Organics: 6.52 to 14.41

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*o-Terphenyl	12.244	.191	.18	94.	-
*1-Chlorooctadecane	13.081	.191	.	.04	-
*#Triacontane	16.326	.191	.09	46.9	-

DRO Area:742216.1
TEH Area:1050881


DRO Amount: 0.0217367
TEH Amount: 3.077633E-02

ERH2527 (OWDFMW08A-FD)

Batch ID: 163748

G:\Org\HP5\DAT\HP5021522_b\0215HP5.0047.RAW

B22020962-032B ;0215HP5 , \$HC-8015-DRO-W, SGT



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: B22020962-032B ;0215HP5 , \$HC-8015-DRO-W, SGT
 Raw File: G:\org\HP5\DAT\HP5021522_b\0215HP5.0047.RAW
 Date & Time Acquired: 2/16/2022 7:12:01 PM
 Method File: G:\Org\HP5\Methods\DR_8015-C24Tb-JEb-L%.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO220111JEb-C24-T.CAL
 Sample Weight: 1060 Dilution: 1 S.A.: 1

Mean RF for TEH: 32675.36

Rt range for Diesel Range Organics: 6.52 to 14.41

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*o-Terphenyl	12.243	.189	.173	91.61	-
*1-Chlorooctadecane	13.111	.189	.	.15	-
*#Triacontane	16.328	.189	.088	46.63	-


DRO Area:333775.1 DRO Amount: 9.636686E-03
 TEH Area:751659.8 TEH Amount: 2.170176E-02

ERH2527 (OWDFMW08A-FD)

Batch ID: 163748

G:\org\HP5\DAT\HP5021522_b\0215HP5.0047.RAW

B22020962-032B ;0215HP5 , \$HC-8015-DRO-W, SGT



RESIDUAL RANGE ORGANICS CHROMATOGRAM

Sample Name: B22020962-032B ;0215HP5 , \$HC-8015-DRO-W, SGT
 Raw File: G:\org\HP5\DAT\HP5021522_b\0215HP5.0047.RAW
 Date & Time Acquired: 2/16/2022 7:12:01 PM
 Method File: G:\Org\HP5\Methods\DR_OROS-BEb-L%.MET
 Calibration File: G:\Org\HP5\Cals\SW8015C_ORO220111BEb_SAMP.CAL
 Sample Weight: 1060 Dilution: 1 S.A.: 1

Mean RF for for Residual Range Organics Calculations: 26424.55
 Rt range for Residual Range Organics: 14.32 to 21.77

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*#Triacontane_____	16.328	.472	.088	18.68


RRO Area:133513 RRO AMOUNT: 4.766613E-03

ERH2527 (OWDFMW08A-FD)

Batch ID: 163748

G:\Org\HP5\DAT\HP5021522_b\0215HP5.0047.RAW

B22020962-032B ;0215HP5 , \$HC-8015-DRO-W, SGT



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: B22020962-032B ;0215HP5 , \$HC-8015-DRO-W, SGT
 Raw File: G:\Org\HP5\DAT\HP5021522_b\0215HP5.0047.RAW
 Date & Time Acquired: 2/16/2022 7:12:01 PM
 Method File: G:\Org\HP5\Methods\DS_8015-C24T-JEb-L#.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO220111JEB-C24-T.CAL
 Sample Weight: 1060 Dilution: 1 S.A.: 1


Mean RF for TEH: 32675.36

Rt range for Diesel Range Organics: 6.52 to 14.41

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*o-Terphenyl	12.243	.189	.172	91.27	-
*1-Chlorooctadecane	13.111	.189	.	.06	-
*#Triacontane	16.328	.189	.087	46.37	-

DRO Area:202319.3
 TEH Area:629508.5

DRO Amount: 5.84132E-03
 TEH Amount: 1.817504E-02



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: DCM-Baseline Check-V48
 Raw File: G:\org\HP5\DAT\HP5021522_b\0215HP5.0048.RAW
 Date & Time Acquired: 2/16/2022 7:54:57 PM
 Method File: G:\Org\HP5\Methods\DR_8015-JA-LEXP.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO220111JA.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for TEH: 32675.36
 Rt range for Diesel Range Organics: 6.63 to 15.99

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*o-Terphenyl	12.351	200.	.077	.04	-
*1-Chlorooctadecane	13.134	200.	.019	.01	-


DRO Area:83686.39 DRO Amount: 2.561147
 TEH Area:122317.8 TEH Amount: 3.743425

ERH2529 (RHMW19)

Batch ID: 163748

G:\Org\HP5\DAT\HP5021522_b\0215HP5.0049.RAW

B22020962-011D ;0215HP5 , \$HC-8015-DRO-W, SGT



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: B22020962-011D ;0215HP5 , \$HC-8015-DRO-W, SGT
 Raw File: G:\org\HP5\DAT\HP5021522_b\0215HP5.0049.RAW
 Date & Time Acquired: 2/16/2022 8:37:56 PM
 Method File: G:\Org\HP5\Methods\D3_8015-021549-JEb-L%.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO220111JEb-C24.CAL
 Sample Weight: 1045 Dilution: 1 S.A.: 1

Mean RF for TEH: 32675.36

Rt range for Diesel Range Organics: 6.52 to 14.41

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*o-Terphenyl	12.241	.191	.134	70.14	-
*1-Chlorooctadecane	13.103	.191	.	.03	-


DRO Area:297080.8 DRO Amount: 8.700373E-03
 TEH Area:9123262 TEH Amount: 0.2671859

ERH2529 (RHMW19)

Batch ID: 163748

G:\org\HP5\DAT\HP5021522_b\0215HP5.0049.RAW

B22020962-011D ;0215HP5 , \$HC-8015-DRO-W, SGT



RESIDUAL RANGE ORGANICS CHROMATOGRAM

Sample Name: B22020962-011D ;0215HP5 , \$HC-8015-DRO-W, SGT
 Raw File: G:\org\HP5\DAT\HP5021522_b\0215HP5.0049.RAW
 Date & Time Acquired: 2/16/2022 8:37:56 PM
 Method File: G:\Org\HP5\Methods\D3_OROS-021549-BEb-L%.MET
 Calibration File: G:\Org\HP5\Cals\SW8015C_ORO220111BEb_SAMP.CAL
 Sample Weight: 1045 Dilution: 1 S.A.: 1

Mean RF for for Residual Range Organics Calculations: 26424.55
 Rt range for Residual Range Organics: 14.32 to 21.77

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*#Triacontane	16.324	.478	.095	19.91


RRO Area:4972892 RRO AMOUNT: 0.1800881

ERH2529 (RHMW19)

Batch ID: 163748

G:\Org\HP5\DAT\HP5021522_b\0215HP5.0049.RAW

B22020962-011D ;0215HP5 , \$HC-8015-DRO-W, SGT



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: B22020962-011D ;0215HP5 , \$HC-8015-DRO-W, SGT
 Raw File: G:\Org\HP5\DAT\HP5021522_b\0215HP5.0049.RAW
 Date & Time Acquired: 2/16/2022 8:37:56 PM
 Method File: G:\Org\HP5\Methods\DS_8015-C24T-JEb-L#.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO220111JEB-C24-T.CAL
 Sample Weight: 1045 Dilution: 1 S.A.: 1

Mean RF for TEH: 32675.36

Rt range for Diesel Range Organics: 6.52 to 14.41

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*o-Terphenyl	12.241	.191	.134	69.96	-
*1-Chlorooctadecane	13.103	.191	.	.01	-
*#Triacontane	16.324	.191	.09	46.82	-


DRO Area:151309.4 DRO Amount: 4.431281E-03
 TEH Area:566896.4 TEH Amount: 1.660225E-02

ERH2537 (Sump Adit 3)

G:\Org\HP5\DAT\HP5021522_b\0215HP5.0050.RAW

Batch ID: 163748

B22020962-026D ;0215HP5 , \$HC-8015-DRO-W, SGT



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: B22020962-026D ;0215HP5 , \$HC-8015-DRO-W, SGT
 Raw File: G:\Org\HP5\DAT\HP5021522_b\0215HP5.0050.RAW
 Date & Time Acquired: 2/16/2022 9:20:56 PM
 Method File: G:\Org\HP5\Methods\D3_8015-C24Tb-JEb-L%.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO220111JEb-C24-T.CAL
 Sample Weight: 1000 Dilution: 1 S.A.: 1

Mean RF for TEH: 32675.36

Rt range for Diesel Range Organics: 6.52 to 14.41

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*o-Terphenyl	12.244	.2	.163	81.56	-
*1-Chlorooctadecane	13.111	.2	.012	6.13	-
*#Triacontane	16.324	.2	.1	50.16	-


DRO Area: 7.362742E+07 DRO Amount: 2.253301
 TEH Area: 8.775042E+07 TEH Amount: 2.685523

ERH2537 (Sump Adit 3)

Batch ID: 163748

G:\Org\HP5\DAT\HP5021522_b\0215HP5.0050.RAW

B22020962-026D ;0215HP5 , \$HC-8015-DRO-W, SGT



RESIDUAL RANGE ORGANICS CHROMATOGRAM

Sample Name: B22020962-026D ;0215HP5 , \$HC-8015-DRO-W, SGT
 Raw File: G:\org\HP5\DAT\HP5021522_b\0215HP5.0050.RAW
 Date & Time Acquired: 2/16/2022 9:20:56 PM
 Method File: G:\Org\HP5\Methods\D3_OROS-BEb-L%.MET
 Calibration File: G:\Org\HP5\Cals\SW8015C_ORO220111BEb_SAMP.CAL
 Sample Weight: 1000 Dilution: 1 S.A.: 1

Mean RF for for Residual Range Organics Calculations: 26424.55
 Rt range for Residual Range Organics: 14.32 to 21.77

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*#Triacontane_____	16.324	.5	.1	20.06


RRO Area:1.252767E+07 RRO AMOUNT: 0.4740919

ERH2537 (Sump Adit 3)

Batch ID: 163748

G:\Org\HP5\DAT\HP5021522_b\0215HP5.0050.RAW

B22020962-026D ;0215HP5 , \$HC-8015-DRO-W, SGT



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT


Sample Name: B22020962-026D ;0215HP5 , \$HC-8015-DRO-W, SGT
 Raw File: G:\Org\HP5\DAT\HP5021522_b\0215HP5.0050.RAW
 Date & Time Acquired: 2/16/2022 9:20:56 PM
 Method File: G:\Org\HP5\Methods\DS_8015-C24T-JEb-L#.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO220111JEB-C24-T.CAL
 Sample Weight: 1000 Dilution: 1 S.A.: 1


Mean RF for TEH: 32675.36

Rt range for Diesel Range Organics: 6.52 to 14.41

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*o-Terphenyl	12.244	.2	.151	75.75	-
*1-Chlorooctadecane	13.111	.2	.002	1.08	-
*#Triacontane	16.324	.2	.079	39.7	-

DRO Area: 5.111595E+07 DRO Amount: 1.564358
 TEH Area: 5.183577E+07 TEH Amount: 1.586387





RESIDUAL RANGE ORGANICS CHROMATOGRAM

Sample Name: CCV_0215HP552r, RRO ;0215HP5 , DRO220201A
 Raw File: G:\org\HP5\DAT\HP5021522_b\0215HP5.0052.RAW
 Date & Time Acquired: 2/16/2022 10:46:46 PM
 Method File: G:\Org\HP5\Methods\DC_ORO-BE-L%.MET
 Calibration File: G:\Org\HP5\Cals\SW8015C_ORO220111BE.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for for ~~Residual~~ TEH (Oil Range) Organics Calculations: 26424.55
 Rt range for ~~Residual~~ TEH (Oil Range) Organics: 12.61 to 30.05

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*#Triacontane	16.332	500.	341.007	68.2	-


~~RRO~~ TEH (Oil Range) Area:1.376733E+08 ~~RRO~~ TEH (Oil Range) AMOUNT: 5210.051

CONTINUING CALIBRATION REPORT: G:\org\HP5\DAT\HP5021522_b\0215HP5.0052.RAW

COMPOUND	ACTUAL (NG)	MEASURED (NG)	%RECOVERY	LIMITS
*30-40 Motor Oil	5000.	.023	.	75-125

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	LIMITS
*#Triacontane	16.332	200.	341.007	170.5	75-125

JLB 02/18/2022



RESIDUAL RANGE ORGANICS CHROMATOGRAM

Sample Name: CCV_0215HP552r, RRO ;0215HP5 , DRO220201A
 Raw File: G:\org\HP5\DAT\HP5021522_b\0215HP5.0052.RAW
 Date & Time Acquired: 2/16/2022 10:46:46 PM
 Method File: G:\Org\HP5\Methods\DS_ORO-BE-L%.MET
 Calibration File: G:\Org\HP5\Cals\SW8015C_ORO220111BE.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for for Residual Range Organics Calculations: 26424.55
 Rt range for Residual Range Organics: 12.61 to 30.05


SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*#Triacontane	16.332	500.	192.755	38.55	-

RRO Area:3658005 RRO AMOUNT: 138.432

CONTINUING CALIBRATION REPORT: G:\org\HP5\DAT\HP5021522_b\0215HP5.0052.RAW

COMPOUND	ACTUAL (NG)	MEASURED (NG)	%RECOVERY	LIMITS
*30-40 Motor Oil	5000.	.023	.	75-125

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	LIMITS
*#Triacontane	16.332	200.	192.755	96.38	75-125



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: CCV_0215HP553r, DRO ;0215HP5 , DRO220128A
 Raw File: G:\org\HP5\DAT\HP5021522_b\0215HP5.0053.RAW
 Date & Time Acquired: 2/16/2022 11:29:41 PM
 Method File: G:\Org\HP5\Methods\DC_8015-C24-JEb-L%.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO220111JEB-C24.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for TEH: 32675.36
 Rt range for Diesel Range Organics: 6.52 to 14.41

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*o-Terphenyl	12.25	200.	299.474	149.74
*1-Chlorooctadecane	13.099	200.	147.386	73.69

DRO Area: 4.599326E+08 DRO Amount: 14075.82
 TEH Area: 4.760012E+08 TEH Amount: 14567.59


CONTINUING CALIBRATION REPORT: G:\org\HP5\DAT\HP5021522_b\0215HP5.0053.RAW

COMPOUND	ACTUAL (NG)	MEASURED (NG)	%RECOVERY	LIMITS
TOTAL DRO	15000.	14567.59	97.12	85-115

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	LIMITS
*o-Terphenyl	12.25	200.	299.474	149.74	85-115
*1-Chlorooctadecane	13.099	200.	147.386	73.69	85-115

G:\Org\HP5\DAT\HP5021522_b\0215HP5.0053.RAW

CCV_0215HP553r, DRO ;0215HP5 , DRO220128A



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: CCV_0215HP553r, DRO ;0215HP5 , DRO220128A
 Raw File: G:\Org\HP5\DAT\HP5021522_b\0215HP5.0053.RAW
 Date & Time Acquired: 2/16/2022 11:29:41 PM
 Method File: G:\Org\HP5\Methods\DS_8015-C24-JEb-L#.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO220111JEB-C24.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for TEH: 32675.36

Rt range for Diesel Range Organics: 6.52 to 14.41


SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*o-Terphenyl	12.25	200.	179.117	89.56
*1-Chlorooctadecane	13.099	200.	34.224	17.11

DRO Area: 2.36808E+08 DRO Amount: 7247.297
 TEH Area: 2.47181E+08 TEH Amount: 7564.753

CONTINUING CALIBRATION REPORT: G:\Org\HP5\DAT\HP5021522_b\0215HP5.0053.RAW

COMPOUND	ACTUAL (NG)	MEASURED (NG)	%RECOVERY	LIMITS
TOTAL DRO	15000.	7564.75	50.43	85-115

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	LIMITS
*o-Terphenyl	12.25	200.	179.117	89.56	85-115
*1-Chlorooctadecane	13.099	200.	34.224	17.11	85-115




DIESEL RANGE ORGANICS CHROMATOGRAM REPORT


Sample Name: DCM-Baseline Check-V54
 Raw File: G:\org\HP5\DAT\HP5021522_b\0215HP5.0054.RAW
 Date & Time Acquired: 2/17/2022 8:25:35 AM
 Method File: G:\Org\HP5\Methods\DR_8015-JA-LEXP.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO220111JA.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for TEH: 32675.36
 Rt range for Diesel Range Organics: 6.63 to 15.99

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*o-Terphenyl	12.306	200.	.237	.12
*1-Chlorooctadecane	29.97	200.	.	.

DRO Area:604715.8 DRO Amount: 18.50678
 TEH Area:692572 TEH Amount: 21.19554





RESIDUAL RANGE ORGANICS CHROMATOGRAM

Sample Name: CCV_0215HP556r, RRO ;0215HP5 , DRO220201A
 Raw File: G:\org\HP5\DAT\HP5021522_b\0215HP5.0056.RAW
 Date & Time Acquired: 2/17/2022 9:51:17 AM
 Method File: G:\Org\HP5\Methods\DC_ORO-BE-L%.MET
 Calibration File: G:\Org\HP5\Cals\SW8015C_ORO220111BE.CAL
 Sample Weight: 1000 Dilution: 1 S.A.: 1

Mean RF for for ~~Residual~~ TEH (Oil Range) Organics Calculations: 26424.55
 Rt range for ~~Residual~~ TEH (Oil Range) Organics: 12.61 to 30.05


SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*#Triacontane	16.333	.5	.333	66.56

RRO TEH (Oil Range) Area:1.385437E+08 RRO TEH (Oil Range) AMOUNT: 5.242992

CONTINUING CALIBRATION REPORT: G:\org\HP5\DAT\HP5021522_b\0215HP5.0056.RAW

COMPOUND	ACTUAL (NG)	MEASURED (NG)	%RECOVERY	LIMITS	
*30-40 Motor Oil	5000.	.	75-125		
SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	LIMITS
*#Triacontane	16.333	.2	.333	166.4	75-125

JLB 02/18/2022



RESIDUAL RANGE ORGANICS CHROMATOGRAM

Sample Name: CCV_0215HP556r, RRO ;0215HP5 , DRO220201A
 Raw File: G:\org\HP5\DAT\HP5021522_b\0215HP5.0056.RAW
 Date & Time Acquired: 2/17/2022 9:51:17 AM
 Method File: G:\Org\HP5\Methods\DS_ORO-BE-L%.MET
 Calibration File: G:\Org\HP5\Cals\SW8015C_ORO220111BE.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1


Mean RF for for Residual Range Organics Calculations: 26424.55
 Rt range for Residual Range Organics: 12.61 to 30.05

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*#Triacontane	16.333	500.	191.659	38.33	-

RRO Area:3454260 RRO AMOUNT: 130.7216

CONTINUING CALIBRATION REPORT: G:\org\HP5\DAT\HP5021522_b\0215HP5.0056.RAW
 COMPOUND ACTUAL (NG) MEASURED (NG) %RECOVERY LIMITS
 *30-40 Motor Oil 5000. . . 75-125

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	LIMITS
*#Triacontane	16.333	200.	191.659	95.83	75-125



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: CCV_0215HP557r, DRO ;0215HP5 , DRO220128A
 Raw File: G:\org\HP5\DAT\HP5021522_b\0215HP5.0057.RAW
 Date & Time Acquired: 2/17/2022 10:34:03 AM
 Method File: G:\Org\HP5\Methods\DC_8015-C24-JEb-L%.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO220111JEB-C24.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for TEH: 32675.36
 Rt range for Diesel Range Organics: 6.52 to 14.41

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*o-Terphenyl	12.248	200.	308.664	154.33
*1-Chlorooctadecane	13.098	200.	149.97	74.98

DRO Area: 4.715199E+08 DRO Amount: 14430.44
 TEH Area: 4.880048E+08 TEH Amount: 14934.95


CONTINUING CALIBRATION REPORT: G:\org\HP5\DAT\HP5021522_b\0215HP5.0057.RAW

COMPOUND	ACTUAL (NG)	MEASURED (NG)	%RECOVERY	LIMITS
TOTAL DRO	15000.	14934.95	99.57	85-115

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	LIMITS
*o-Terphenyl	12.248	200.	308.664	154.33	85-115
*1-Chlorooctadecane	13.098	200.	149.97	74.98	85-115

G:\org\HP5\DAT\HP5021522_b\0215HP5.0057.RAW

CCV_0215HP557r, DRO ;0215HP5 , DRO220128A



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: CCV_0215HP557r, DRO ;0215HP5 , DRO220128A
 Raw File: G:\org\HP5\DAT\HP5021522_b\0215HP5.0057.RAW
 Date & Time Acquired: 2/17/2022 10:34:03 AM
 Method File: G:\Org\HP5\Methods\DS_8015-C24-JEb-L#.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO220111JEB-C24.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for TEH: 32675.36

Rt range for Diesel Range Organics: 6.52 to 14.41


SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*o-Terphenyl	12.248	200.	184.53	92.27
*1-Chlorooctadecane	13.098	200.	35.514	17.76

DRO Area: 2.422591E+08 DRO Amount: 7414.122
 TEH Area: 2.528782E+08 TEH Amount: 7739.11

CONTINUING CALIBRATION REPORT: G:\org\HP5\DAT\HP5021522_b\0215HP5.0057.RAW

COMPOUND	ACTUAL (NG)	MEASURED (NG)	%RECOVERY	LIMITS
TOTAL DRO	15000.	7739.11	51.59	85-115

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	LIMITS
*o-Terphenyl	12.248	200.	184.53	92.27	85-115
*1-Chlorooctadecane	13.098	200.	35.514	17.76	85-115




DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: DCM-Baseline Check-V58
 Raw File: G:\org\HP5\DAT\HP5021522_b\0215HP5.0058.RAW
 Date & Time Acquired: 2/17/2022 11:16:56 AM
 Method File: G:\Org\HP5\Methods\DR_8015-JA-LEXP.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO220111JA.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for TEH: 32675.36
 Rt range for Diesel Range Organics: 6.63 to 15.99

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*o-Terphenyl	12.335	200.	.026	.01
*1-Chlorooctadecane	29.965	200.	.	.

DRO Area:153488.6 DRO Amount: 4.69738
 TEH Area:310934.1 TEH Amount: 9.51586



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: DCM-Baseline Check-V59
 Raw File: G:\org\HP5\DAT\HP5021522_b\0215HP5.0059.RAW
 Date & Time Acquired: 2/17/2022 11:59:46 AM
 Method File: G:\Org\HP5\Methods\DR_8015-JA-LEXP.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO220111JA.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for TEH: 32675.36
 Rt range for Diesel Range Organics: 6.63 to 15.99

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*o-Terphenyl	12.322	200.	.039	.02 -
*1-Chlorooctadecane	29.98	200.	.	. -


DRO Area:105162.9 DRO Amount: 3.218416
 TEH Area:178633.2 TEH Amount: 5.466907

ERH2531 (RHMW01R)

Batch ID: 163748

G:\Org\HP5\DAT\HP5021522_b\0215HP5.0060.RAW

B22020962-006D ;0215HP5 , \$HC-8015-DRO-W, SGT



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: B22020962-006D ;0215HP5 , \$HC-8015-DRO-W, SGT
 Raw File: G:\Org\HP5\DAT\HP5021522_b\0215HP5.0060.RAW
 Date & Time Acquired: 2/17/2022 12:42:40 PM
 Method File: G:\Org\HP5\Methods\DR_8015-C24Tb-JEb-L%.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO220111JEb-C24-T.CAL
 Sample Weight: 1050 Dilution: 1 S.A.: 1

Mean RF for TEH: 32675.36

Rt range for Diesel Range Organics: 6.52 to 14.41

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*o-Terphenyl	12.246	.19	.164	86.05	-
*1-Chlorooctadecane	13.081	.19	.	.16	-
*#Triacontane	16.329	.19	.096	50.26	-


DRO Area:1568527 DRO Amount: 4.571747E-02
 TEH Area:1789863 TEH Amount: 0.0521687

ERH2531 (RHMW01R)

Batch ID: 163748

G:\org\HP5\DAT\HP5021522_b\0215HP5.0060.RAW

B22020962-006D ;0215HP5 , \$HC-8015-DRO-W, SGT



RESIDUAL RANGE ORGANICS CHROMATOGRAM

Sample Name: B22020962-006D ;0215HP5 , \$HC-8015-DRO-W, SGT
 Raw File: G:\org\HP5\DAT\HP5021522_b\0215HP5.0060.RAW
 Date & Time Acquired: 2/17/2022 12:42:40 PM
 Method File: G:\Org\HP5\Methods\DR_OROS-BEb-L%.MET
 Calibration File: G:\Org\HP5\Cals\SW8015C_ORO220111BEb_SAMP.CAL
 Sample Weight: 1050 Dilution: 1 S.A.: 1

Mean RF for for Residual Range Organics Calculations: 26424.55
 Rt range for Residual Range Organics: 14.32 to 21.77

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*#Triacontane	16.329	.476	.096	20.2

RRO Area:155611.4 RRO AMOUNT: 5.608471E-03

ERH2531 (RHMW01R)

Batch ID: 163748

G:\Org\HP5\DAT\HP5021522_b\0215HP5.0060.RAW

B22020962-006D ;0215HP5 , \$HC-8015-DRO-W, SGT



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT


Sample Name: B22020962-006D ;0215HP5 , \$HC-8015-DRO-W, SGT
 Raw File: G:\Org\HP5\DAT\HP5021522_b\0215HP5.0060.RAW
 Date & Time Acquired: 2/17/2022 12:42:40 PM
 Method File: G:\Org\HP5\Methods\DS_8015-C24T-JEb-L#.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO220111JEB-C24-T.CAL
 Sample Weight: 1050 Dilution: 1 S.A.: 1

Mean RF for TEH: 32675.36

Rt range for Diesel Range Organics: 6.52 to 14.41

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*o-Terphenyl	12.246	.19	.163	85.7	-
*1-Chlorooctadecane	13.081	.19	.	.07	-
*#Triacontane	16.329	.19	.095	50.06	-

DRO Area:1186501 DRO Amount: 3.458267E-02
 TEH Area:1395105 TEH Amount: 4.066279E-02



RESIDUAL RANGE ORGANICS CHROMATOGRAM


Sample Name: B22020962-006DMS-RRO ;0215HP5 , SGT
 Raw File: G:\org\HP5\DAT\HP5021522_b\0215HP5.0061.RAW
 Date & Time Acquired: 2/17/2022 1:25:38 PM
 Method File: G:\Org\HP5\Methods\D3_ORO-BE-L%.MET
 Calibration File: G:\Org\HP5\Cals\SW8015C_ORO220111BE.CAL
 Sample Weight: 1050 Dilution: 1 S.A.: 1

Mean RF for for ~~Residual~~ TEH (Oil Range) Organics Calculations: 26424.55
 Rt range for ~~Residual~~ TEH (Oil Range) Organics: 12.61 to 30.05

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*#Triacontane	16.336	.476	.176	36.94	-

~~RRO~~ TEH (Oil Range) Area:1.339117E+08 ~~RRO~~ TEH (Oil Range) AMOUNT: 4.826382

JLB 02/18/2022




RESIDUAL RANGE ORGANICS CHROMATOGRAM

Sample Name: B22020962-006DMS-RRO ;0215HP5 , SGT
 Raw File: G:\org\HP5\DAT\HP5021522_b\0215HP5.0061.RAW
 Date & Time Acquired: 2/17/2022 1:25:38 PM
 Method File: G:\Org\HP5\Methods\DS_ORO-BE-L%.MET
 Calibration File: G:\Org\HP5\Cals\SW8015C_ORO220111BE.CAL
 Sample Weight: 1050 Dilution: 1 S.A.: 1

Mean RF for for Residual Range Organics Calculations: 26424.55
 Rt range for Residual Range Organics: 12.61 to 30.05

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*#Triacontane	16.336	.476	.091	19.09

RRO Area:2727465 RRO AMOUNT: 9.830197E-02



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT


Sample Name: DCM-Baseline Check-V62
 Raw File: G:\org\HP5\DAT\HP5021522_b\0215HP5.0062.RAW
 Date & Time Acquired: 2/17/2022 2:08:23 PM
 Method File: G:\Org\HP5\Methods\DR_8015-JA-LEXP.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO220111JA.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for TEH: 32675.36

Rt range for Diesel Range Organics: 6.63 to 15.99

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*o-Terphenyl	29.823	200.	.	-
*1-Chlorooctadecane	13.168	200.	1.594	.8

DRO Area: 5972160 DRO Amount: 182.7726
 TEH Area: 6589083 TEH Amount: 201.653



RESIDUAL RANGE ORGANICS CHROMATOGRAM


Sample Name: LCS-163748-RRO ;0215HP5 , SGT
 Raw File: G:\org\HP5\DAT\HP5021522_b\0215HP5.0063.RAW
 Date & Time Acquired: 2/17/2022 2:51:01 PM
 Method File: G:\Org\HP5\Methods\D3_ORO-BE-L%.MET
 Calibration File: G:\Org\HP5\Cals\SW8015C_ORO220111BE.CAL
 Sample Weight: 1000 Dilution: 1 S.A.: 1

Mean RF for for ~~Residual~~ TEH (Oil Range) Organics Calculations: 26424.55
 Rt range for ~~Residual~~ TEH (Oil Range) Organics: 12.61 to 30.05

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*#Triacontane	16.338	.5	.216	43.27

RRO TEH (Oil Range) Area:1.374061E+08 RRO TEH (Oil Range) AMOUNT: 5.199942

JLB 02/18/2022




RESIDUAL RANGE ORGANICS CHROMATOGRAM

Sample Name: LCS-163748-RRO ;0215HP5 , SGT
 Raw File: G:\Org\HP5\DAT\HP5021522_b\0215HP5.0063.RAW
 Date & Time Acquired: 2/17/2022 2:51:01 PM
 Method File: G:\Org\HP5\Methods\DS_ORO-BE-L%.MET
 Calibration File: G:\Org\HP5\Cals\SW8015C_ORO220111BE.CAL
 Sample Weight: 1000 Dilution: 1 S.A.: 1

Mean RF for for Residual Range Organics Calculations: 26424.55
 Rt range for Residual Range Organics: 12.61 to 30.05

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*#Triacontane_____	16.338	.5	.099	19.84

RRO Area:3217107 RRO AMOUNT: 0.1217469



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT


Sample Name: DCM-Baseline Check-V64
 Raw File: G:\org\HP5\DAT\HP5021522_b\0215HP5.0064.RAW
 Date & Time Acquired: 2/17/2022 3:33:37 PM
 Method File: G:\Org\HP5\Methods\DR_8015-JA-LEXP.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO220111JA.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for TEH: 32675.36

Rt range for Diesel Range Organics: 6.63 to 15.99

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*o-Terphenyl	12.364	200.	.191	.1
*1-Chlorooctadecane	29.796	200.	.	.

DRO Area:414697.6 DRO Amount: 12.69145
 TEH Area:1052639 TEH Amount: 32.21505



RESIDUAL RANGE ORGANICS CHROMATOGRAM


Sample Name: LCSD-163748-RRO ;0215HP5 , SGT
 Raw File: G:\org\HP5\DAT\HP5021522_b\0215HP5.0065.RAW
 Date & Time Acquired: 2/17/2022 4:16:18 PM
 Method File: G:\Org\HP5\Methods\D3_ORO-BE-L%.MET
 Calibration File: G:\Org\HP5\Cals\SW8015C_ORO220111BE.CAL
 Sample Weight: 1000 Dilution: 1 S.A.: 1

Mean RF for for ~~Residual~~ TEH (Oil Range) Organics Calculations: 26424.55
 Rt range for ~~Residual~~ TEH (Oil Range) Organics: 12.61 to 30.05

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*#Triacontane	16.335	.5	.194	38.88

RRO TEH (Oil Range) Area:1.31824E+08 RRO TEH (Oil Range) AMOUNT: 4.988695

JLB 02/18/2022




RESIDUAL RANGE ORGANICS CHROMATOGRAM


Sample Name: LCSD-163748-RRO ;0215HP5 , SGT
 Raw File: G:\org\HP5\DAT\HP5021522_b\0215HP5.0065.RAW
 Date & Time Acquired: 2/17/2022 4:16:18 PM
 Method File: G:\Org\HP5\Methods\DS_ORO-BE-L%.MET
 Calibration File: G:\Org\HP5\Cals\SW8015C_ORO220111BE.CAL
 Sample Weight: 1000 Dilution: 1 S.A.: 1

Mean RF for for Residual Range Organics Calculations: 26424.55
 Rt range for Residual Range Organics: 12.61 to 30.05

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*#Triacontane_____	16.335	.5	.094	18.86 -

RRO Area:2586011 RRO AMOUNT: 9.786393E-02





RESIDUAL RANGE ORGANICS CHROMATOGRAM

Sample Name: CCV_0215HP567r, RRO ;0215HP5 , DRO220201A
 Raw File: G:\org\HP5\DAT\HP5021522_b\0215HP5.0067.RAW
 Date & Time Acquired: 2/17/2022 5:42:51 PM
 Method File: G:\Org\HP5\Methods\DC_ORO-BE-L%.MET
 Calibration File: G:\Org\HP5\Cals\SW8015C_ORO220111BE.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for for ~~Residual~~ TEH (Oil Range) Organics Calculations: 26424.55
 Rt range for ~~Residual~~ TEH (Oil Range) Organics: 12.61 to 30.05

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*#Triacontane	16.338	500.	324.741	64.95	-


~~RRO TEH (Oil Range)~~ Area:1.377402E+08 ~~RRO TEH (Oil Range)~~ AMOUNT: 5212.585

CONTINUING CALIBRATION REPORT: G:\org\HP5\DAT\HP5021522_b\0215HP5.0067.RAW

COMPOUND	ACTUAL (NG)	MEASURED (NG)	%RECOVERY	LIMITS
*30-40 Motor Oil	5000.	.	75-125	

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	LIMITS
*#Triacontane	16.338	200.	324.741	162.37	75-125

JLB 02/18/2022



RESIDUAL RANGE ORGANICS CHROMATOGRAM

Sample Name: CCV_0215HP567r, RRO ;0215HP5 , DRO220201A
 Raw File: G:\org\HP5\DAT\HP5021522_b\0215HP5.0067.RAW
 Date & Time Acquired: 2/17/2022 5:42:51 PM
 Method File: G:\Org\HP5\Methods\DS_ORO-BE-L%.MET
 Calibration File: G:\Org\HP5\Cals\SW8015C_ORO220111BE.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for for Residual Range Organics Calculations: 26424.55
 Rt range for Residual Range Organics: 12.61 to 30.05


SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	
*#Triacontane	16.338	500.	189.635	37.93	-

RRO Area:3517665 RRO AMOUNT: 133.1211

CONTINUING CALIBRATION REPORT: G:\org\HP5\DAT\HP5021522_b\0215HP5.0067.RAW

COMPOUND	ACTUAL (NG)	MEASURED (NG)	%RECOVERY	LIMITS
*30-40 Motor Oil	5000.	.	75-125	

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	LIMITS
*#Triacontane	16.338	200.	189.635	94.82	75-125



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: CCV_0215HP568r, DRO ;0215HP5 , DRO220128A
 Raw File: G:\org\HP5\DAT\HP5021522_b\0215HP5.0068.RAW
 Date & Time Acquired: 2/17/2022 6:26:16 PM
 Method File: G:\Org\HP5\Methods\DC_8015-C24-JEb-L%.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO220111JEb-C24.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for TEH: 32675.36
 Rt range for Diesel Range Organics: 6.52 to 14.41

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*o-Terphenyl	12.256	200.	309.656	154.83
*1-Chlorooctadecane	13.106	200.	157.561	78.78

DRO Area: 4.741033E+08 DRO Amount: 14509.51
 TEH Area: 4.905642E+08 TEH Amount: 15013.28


CONTINUING CALIBRATION REPORT: G:\org\HP5\DAT\HP5021522_b\0215HP5.0068.RAW

COMPOUND	ACTUAL (NG)	MEASURED (NG)	%RECOVERY	LIMITS
TOTAL DRO	15000.	15013.28	100.09	85-115

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	LIMITS
*o-Terphenyl	12.256	200.	309.656	154.83	85-115
*1-Chlorooctadecane	13.106	200.	157.561	78.78	85-115

G:\org\HP5\DAT\HP5021522_b\0215HP5.0068.RAW

CCV_0215HP568r, DRO ;0215HP5 , DRO220128A



DIESEL RANGE ORGANICS CHROMATOGRAM REPORT

Sample Name: CCV_0215HP568r, DRO ;0215HP5 , DRO220128A
 Raw File: G:\org\HP5\DAT\HP5021522_b\0215HP5.0068.RAW
 Date & Time Acquired: 2/17/2022 6:26:16 PM
 Method File: G:\Org\HP5\Methods\DS_8015-C24-JEb-L#.met
 Calibration File: G:\Org\HP5\Cals\SW8015C_DRO220111JEB-C24.CAL
 Sample Weight: 1 Dilution: 1 S.A.: 1

Mean RF for TEH: 32675.36

Rt range for Diesel Range Organics: 6.52 to 14.41

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC
*o-Terphenyl	12.256	200.	184.661	92.33
*1-Chlorooctadecane	13.106	200.	35.941	17.97

DRO Area: 2.443996E+08 DRO Amount: 7479.629
 TEH Area: 2.551085E+08 TEH Amount: 7807.365

CONTINUING CALIBRATION REPORT: G:\org\HP5\DAT\HP5021522_b\0215HP5.0068.RAW

COMPOUND	ACTUAL (NG)	MEASURED (NG)	%RECOVERY	LIMITS
TOTAL DRO	15000.	7807.37	52.05	85-115

SURROGATE COMPOUND	RT	ACTUAL	MEASURED	%REC	LIMITS
*o-Terphenyl	12.256	200.	184.661	92.33	85-115
*1-Chlorooctadecane	13.106	200.	35.941	17.97	85-115

G:\org\HP5\DAT\HP5021522_b\0215HP5.27r	DCM-Baseline Check-V27	G:\Org\HP5\Methods\DR_8015-JA-LEXP.met	1	1	1	1	0	No integration
G:\org\HP5\DAT\HP5021522_b\0215HP5.28r	DCM-Baseline Check-V28	G:\Org\HP5\Methods\DR_8015-JA-LEXP.met	1	1	1	1	0	No integration
G:\org\HP5\DAT\HP5021522_b\0215HP5.29r	B22020962-016D ;0215HP5 , \$HC-8015-DRO-W,	G:\Org\HP5\Methods\DR_8015-C24T-JEb-L%.met G:\Org\HP5\Methods\DR_OROS-BEb-L%.MET G:\Org\HP5\Methods\DS_8015-C24T-JEb-L%.met	1010	1	1	1	0	The integration of Diesel Range Organics (C10-C24), C24-C40, and Total Extractable Hydrocarbons (TEH) is the hydrocarbon response with reference to the baseline. Assigned Set Baseline All Valleys on at 17.12 minutes. Surrogates are integrated using a valley to valley integration using Set baseline All Valleys on at 10.78 minutes and X-axis scaling showing surrogate peak from 11-18 minutes.
G:\org\HP5\DAT\HP5021522_b\0215HP5.30r	DCM-Baseline Check-V30	G:\Org\HP5\Methods\DR_8015-JA-LEXP.met	1	1	1	1	0	No integration
G:\org\HP5\DAT\HP5021522_b\0215HP5.31r	DCM-Baseline Check-V31	G:\Org\HP5\Methods\DR_8015-JA-LEXP.met	1	1	1	1	0	No integration
G:\org\HP5\DAT\HP5021522_b\0215HP5.32r	LCS-163748-RRO ;0215HP5 ,	G:\Org\HP5\Methods\DS_ORO-BE-L%.MET G:\Org\HP5\Methods\DS_ORO-BE-L%.MET	1000	1	1	1	0	The integration of Oil Range hydrocarbon is the hydrocarbon response with reference to the baseline. Surrogates are integrated using a valley to valley integration using Set baseline All Valleys on placed at 16.14 minutes and X-axis scaling showing surrogate peak from 14-18 minutes.
G:\org\HP5\DAT\HP5021522_b\0215HP5.33r	DCM-Baseline Check-V33	G:\Org\HP5\Methods\DR_8015-JA-LEXP.met	1	1	1	1	0	No integration
G:\org\HP5\DAT\HP5021522_b\0215HP5.34r	LCSD-163748-RRO ;0215HP5 ,	G:\Org\HP5\Methods\DS_ORO-BE-L%.MET G:\Org\HP5\Methods\DS_ORO-BE-L%.MET	1000	1	1	1	0	The integration of Oil Range hydrocarbon is the hydrocarbon response with reference to the baseline. Surrogates are integrated using a valley to valley integration using Set baseline All Valleys on placed at 16.14 minutes and X-axis scaling showing surrogate peak from 14-18 minutes.
G:\org\HP5\DAT\HP5021522_b\0215HP5.35r	MARKER ;0215HP535r, DRO ;0215HP5 , DRO220128B	G:\org\HP5\Methods\CSC220215.met	1	1	1	1	0	No integration
G:\org\HP5\DAT\HP5021522_b\0215HP5.31r	CCV_0215HP536r, RRO ;0215HP5 , DRO220201A	G:\Org\HP5\Methods\DC_ORO-BE-L%.MET G:\Org\HP5\Methods\DS_ORO-BE-L%.MET	1	1	1	1	0	The integration of Oil Range hydrocarbon is the hydrocarbon response with reference to the baseline. Surrogates are integrated using a valley to valley integration using Set baseline All Valleys on placed at 16.14 minutes and X-axis scaling showing surrogate peak from 14-18 minutes.
G:\org\HP5\DAT\HP5021522_b\0215HP5.37r	CCV_0215HP537r, DRO ;0215HP5 , DRO220128A	G:\Org\HP5\Methods\DC_8015-C24-JEb-L%.met G:\Org\HP5\Methods\DS_8015-C24-JEb-L%.met	1	1	1	1	0	The integration of Diesel Range Organics (C10-C24) and Total Extractable Hydrocarbons is the hydrocarbon response with reference to the baseline. Assigned Set Baseline All Valley on at 16.35 minutes. Surrogates are integrated using a valley to valley integration using Set baseline All Valleys on placed at 12.01 minutes and X-axis scaling showing surrogate peak from 11.5-14 minutes.

Ann Nebel

Digitally signed by
Ann Nebel
Date: 2022.02.18 12:36:24 -07:00

Write Sequence Data File	Insert Entries(Have the first cell for entries selected) Sample Name	Method	Weight	Dil Factor	Amt Inj.	IS	Cal ID	Manual Integration
G:\org\HP5\DAT\HP5021522_b\0215HP5.35r G:\org\HP5\DAT\HP5021522_b\0215HP5.36r	MARKER_0215HP535r, DRO_0215HP5_DRO220128B CCV_0215HP536r, RRO_0215HP5_DRO220201A	G:\org\HP5\Methods\CSC220215.met G:\org\HP5\Methods\DC_ORO-BE-L%.MET G:\org\HP5\Methods\DS_ORO-BE-L%.MET	1	1	1	1	0	No integration The integration of Oil Range hydrocarbon is the hydrocarbon response with reference to the baseline. Surrogates are integrated using a valley to valley integration using Set baseline All Valleys on placed at 16.14 minutes and X-axis scaling showing surrogate peak from 14-18 minutes.
G:\org\HP5\DAT\HP5021522_b\0215HP5.37r	CCV_0215HP537r, DRO_0215HP5_DRO220128A	G:\org\HP5\Methods\DC_8015-C24-JEB-L%.met G:\org\HP5\Methods\DS_8015-C24-JEB-L%.met	1	1	1	1	0	The integration of Diesel Range Organics (C10-C24) and Total Extractable Hydrocarbons is the hydrocarbon response with reference to the baseline. Assigned Set Baseline All Valley on at 16.35 minutes. Surrogates are integrated using a valley to valley integration using Set baseline All Valleys on placed at 12.01 minutes and X-axis scaling showing surrogate peak from 11.5-14 minutes.
G:\org\HP5\DAT\HP5021522_b\0215HP5.38r G:\org\HP5\DAT\HP5021522_b\0215HP5.39r G:\org\HP5\DAT\HP5021522_b\0215HP5.40r	DCM-Baseline Check-V38 DCM-Baseline Check-V39 LCS-163748_0215HP5_SGT	G:\org\HP5\Methods\DR_8015-JA-LEXP.met G:\org\HP5\Methods\DR_8015-JA-LEXP.met G:\org\HP5\Methods\D3_8015-021540-JEB-L%.met G:\org\HP5\Methods\DS_8015-C24-JEB-L%.met	1 1 1000	1 1 1	1 1 1	1 1 1	0	No integration No integration The integration of Diesel Range Organics (C10-C24) and Total Extractable Hydrocarbons is the hydrocarbon response with reference to the baseline. Adjusted peak width and placed split at 12.3 minutes. Surrogates are integrated using a valley to valley integration using Set baseline All Valleys on placed at 12.01 minutes and X-axis scaling showing surrogate peak from 11.5-14 minutes.
G:\org\HP5\DAT\HP5021522_b\0215HP5.41r	LCSD-163748_0215HP5_SGT	G:\org\HP5\Methods\D3_8015-C24-JEB-L%.met G:\org\HP5\Methods\DS_8015-C24-JEB-L%.met	1000	1	1	1	0	The integration of Diesel Range Organics (C10-C24) and Total Extractable Hydrocarbons is the hydrocarbon response with reference to the baseline. Surrogates are integrated using a valley to valley integration using Set baseline All Valleys on placed at 12.01 minutes and X-axis scaling showing surrogate peak from 11.5-14 minutes.
G:\org\HP5\DAT\HP5021522_b\0215HP5.42r	MB-163748_0215HP5_SGT	G:\org\HP5\Methods\DR_8015-C24Tb-JEB-L%.met G:\org\HP5\Methods\DR_OROS-BEB-L%.MET G:\org\HP5\Methods\DS_8015-C24T-JEB-L%.met	1000	1	1	1	0	The integration of Diesel Range Organics (C10-C24), C24-C40, and Total Extractable Hydrocarbons (TEH) is the hydrocarbon response with reference to the baseline. Assigned Set Baseline All Valley on at 17.12 minutes. Surrogates are integrated using a valley to valley integration using Set baseline All Valleys on at 10.78 minutes and X-axis scaling showing surrogate peak from 11-18 minutes.
G:\org\HP5\DAT\HP5021522_b\0215HP5.43r	B22020962-001D_0215HP5_SHC-8015-DRO-W, SGT	G:\org\HP5\Methods\DR_8015-C24Tb-JEB-L%.met G:\org\HP5\Methods\DR_OROS-BEB-L%.MET G:\org\HP5\Methods\DS_8015-C24T-JEB-L%.met	1050	1	1	1	0	The integration of Diesel Range Organics (C10-C24), C24-C40, and Total Extractable Hydrocarbons (TEH) is the hydrocarbon response with reference to the baseline. Assigned Set Baseline All Valley on at 17.12 minutes. Surrogates are integrated using a valley to valley integration using Set baseline All Valleys on at 10.78 minutes and X-axis scaling showing surrogate peak from 11-18 minutes.
G:\org\HP5\DAT\HP5021522_b\0215HP5.44r	B22020962-001DMS_0215HP5_SGT	G:\org\HP5\Methods\D3_8015-021540-JEB-L%.met G:\org\HP5\Methods\DS_8015-C24-JEB-L%.met	1050	1	1	1	0	The integration of Diesel Range Organics (C10-C24) and Total Extractable Hydrocarbons is the hydrocarbon response with reference to the baseline. Adjusted peak width and placed split at 12.3 minutes. Surrogates are integrated using a valley to valley integration using Set baseline All Valleys on placed at 12.01 minutes and X-axis scaling showing surrogate peak from 11.5-14 minutes.
G:\org\HP5\DAT\HP5021522_b\0215HP5.45r G:\org\HP5\DAT\HP5021522_b\0215HP5.46r	DCM-Baseline Check-V45 B22020962-021D_0215HP5_SHC-8015-DRO-W, SGT	G:\org\HP5\Methods\DR_8015-JA-LEXP.met G:\org\HP5\Methods\DR_8015-C24Tb-JEB-L%.met G:\org\HP5\Methods\DR_OROS-BEB-L%.MET G:\org\HP5\Methods\DS_8015-C24T-JEB-L%.met	1 1045	1 1	1 1	1 1	0	No integration The integration of Diesel Range Organics (C10-C24), C24-C40, and Total Extractable Hydrocarbons (TEH) is the hydrocarbon response with reference to the baseline. Assigned Set Baseline All Valley on at 17.12 minutes. Surrogates are integrated using a valley to valley integration using Set baseline All Valleys on at 10.78 minutes and X-axis scaling showing surrogate peak from 11-18 minutes.
G:\org\HP5\DAT\HP5021522_b\0215HP5.47r	B22020962-032B_0215HP5_SHC-8015-DRO-W, SGT	G:\org\HP5\Methods\DR_8015-C24Tb-JEB-L%.met G:\org\HP5\Methods\DR_OROS-BEB-L%.MET G:\org\HP5\Methods\DS_8015-C24T-JEB-L%.met	1060	1	1	1	0	The integration of Diesel Range Organics (C10-C24), C24-C40, and Total Extractable Hydrocarbons (TEH) is the hydrocarbon response with reference to the baseline. Assigned Set Baseline All Valley on at 17.12 minutes. Surrogates are integrated using a valley to valley integration using Set baseline All Valleys on at 10.78 minutes and X-axis scaling showing surrogate peak from 11-18 minutes.
G:\org\HP5\DAT\HP5021522_b\0215HP5.48r G:\org\HP5\DAT\HP5021522_b\0215HP5.49r	DCM-Baseline Check-V48 B22020962-011D_0215HP5_SHC-8015-DRO-W, SGT	G:\org\HP5\Methods\DR_8015-JA-LEXP.met G:\org\HP5\Methods\D3_8015-021549-JEB-L%.met G:\org\HP5\Methods\D3_OROS-021549-BEB-L%.MET G:\org\HP5\Methods\DS_8015-C24T-JEB-L%.met	1 1045	1 1	1 1	1 1	0	No integration The integration of Diesel Range Organics (C10-C24), C24-C40, and Total Extractable Hydrocarbons (TEH) is the hydrocarbon response with reference to the baseline. Set Baseline now placed at 27.85 minutes. Surrogates are integrated using a valley to valley integration using Set baseline All Valleys on at 10.78 minutes and X-axis scaling showing surrogate peak from 11-18 minutes.
G:\org\HP5\DAT\HP5021522_b\0215HP5.50r	B22020962-026D_0215HP5_SHC-8015-DRO-W, SGT	G:\org\HP5\Methods\D3_8015-C24Tb-JEB-L%.met G:\org\HP5\Methods\DS_8015-C24T-JEB-L%.met	1000	1	1	1	0	The integration of Diesel Range Organics (C10-C24), C24-C40, and Total Extractable Hydrocarbons (TEH) is the hydrocarbon response with reference to the baseline. Surrogates are integrated using a valley to valley integration using Set baseline All Valleys on at 10.78 minutes and X-axis scaling showing surrogate peak from 11-18 minutes.
G:\org\HP5\DAT\HP5021522_b\0215HP5.51r G:\org\HP5\DAT\HP5021522_b\0215HP5.52r	MARKER_0215HP551r, DRO_0215HP5_DRO220128B CCV_0215HP552r, RRO_0215HP5_DRO220201A	G:\org\HP5\Methods\CSC220215.met G:\org\HP5\Methods\DC_ORO-BE-L%.MET G:\org\HP5\Methods\DS_ORO-BE-L%.MET	1 1	1 1	1 1	1 1	0	No integration The integration of Oil Range hydrocarbon is the hydrocarbon response with reference to the baseline. Surrogates are integrated using a valley to valley integration using Set baseline All Valleys on placed at 16.14 minutes and X-axis scaling showing surrogate peak from 14-18 minutes.
G:\org\HP5\DAT\HP5021522_b\0215HP5.53r	CCV_0215HP553r, DRO_0215HP5_DRO220128A	G:\org\HP5\Methods\DC_8015-C24-JEB-L%.met G:\org\HP5\Methods\DS_8015-C24-JEB-L%.met	1	1	1	1	0	The integration of Diesel Range Organics (C10-C24) and Total Extractable Hydrocarbons is the hydrocarbon response with reference to the baseline. Assigned Set Baseline All Valley on at 16.35 minutes. Surrogates are integrated using a valley to valley integration using Set baseline All Valleys on placed at 12.01 minutes and X-axis scaling showing surrogate peak from 11.5-14 minutes.
G:\org\HP5\DAT\HP5021522_b\0215HP5.54r G:\org\HP5\DAT\HP5021522_b\0215HP5.55r G:\org\HP5\DAT\HP5021522_b\0215HP5.56r	DCM-Baseline Check-V54 MARKER_0215HP555r, DRO_0215HP5_DRO220128B CCV_0215HP556r, RRO_0215HP5_DRO220201A	G:\org\HP5\Methods\DR_8015-JA-LEXP.met G:\org\HP5\Methods\CSC220215.met G:\org\HP5\Methods\DC_ORO-BE-L%.MET G:\org\HP5\Methods\DS_ORO-BE-L%.MET	1 1 1050	1 1 1	1 1 1	1 1 1	0	No integration No integration The integration of Oil Range hydrocarbon is the hydrocarbon response with reference to the baseline. Surrogates are integrated using a valley to valley integration using Set baseline All Valleys on placed at 16.14 minutes and X-axis scaling showing surrogate peak from 14-18 minutes.
G:\org\HP5\DAT\HP5021522_b\0215HP5.57r	CCV_0215HP557r, DRO_0215HP5_DRO220128A	G:\org\HP5\Methods\DC_8015-C24-JEB-L%.met G:\org\HP5\Methods\DS_8015-C24-JEB-L%.met	1050	1	1	1	0	The integration of Diesel Range Organics (C10-C24) and Total Extractable Hydrocarbons is the hydrocarbon response with reference to the baseline. Assigned Set Baseline All Valley on at 16.35 minutes. Surrogates are integrated using a valley to valley integration using Set baseline All Valleys on placed at 12.01 minutes and X-axis scaling showing surrogate peak from 11.5-14 minutes.
G:\org\HP5\DAT\HP5021522_b\0215HP5.58r G:\org\HP5\DAT\HP5021522_b\0215HP5.59r	DCM-Baseline Check-V58 DCM-Baseline Check-V59	G:\org\HP5\Methods\DR_8015-JA-LEXP.met G:\org\HP5\Methods\DR_8015-JA-LEXP.met	1 1	1 1	1 1	1 1	0	No integration No integration

G:\org\HP5\DAT\HP5021522_b\0215HP5.60r	B22020962-006D_0215HP5_ \$HC-8015-DRO-W, SGT	G:\Org\HP5\Methods\DR_8015-C24Tb-JEb-L%.met G:\Org\HP5\Methods\DR_OROS-BEb-L%.MET G:\Org\HP5\Methods\DS_8015-C24T-JEb-L#.met	1050	1	1	1	0	The integration of Diesel Range Organics (C10-C24), C24-C40, and Total Extractable Hydrocarbons (TEH) is the hydrocarbon response with reference to the baseline. Assigned Set Baseline All Valley on at 17.12 minutes. Surrogates are integrated using a valley to valley integration using Set baseline All Valleys on at 10.78 minutes and X-axis scaling showing surrogate peak from 11-18 minutes.
G:\org\HP5\DAT\HP5021522_b\0215HP5.61r	B22020962-006DMS-RRO_0215HP5_ SGT	G:\Org\HP5\Methods\D3_ORO-BE-L%.MET G:\Org\HP5\Methods\DS_ORO-BE-L%.MET	1050	1	1	1	0	The integration of Oil Range hydrocarbon is the hydrocarbon response with reference to the baseline. Surrogates are integrated using a valley to valley integration using Set baseline All Valleys on placed at 16.14 minutes and X-axis scaling showing surrogate peak from 14-18 minutes.
G:\org\HP5\DAT\HP5021522_b\0215HP5.62r	DCM-Baseline Check-V62	G:\Org\HP5\Methods\DR_8015-JA-LEXP.met	1	1	1	1	0	No integration
G:\org\HP5\DAT\HP5021522_b\0215HP5.63r	LCS-163748-RRO_0215HP5_ SGT	G:\Org\HP5\Methods\D3_ORO-BE-L%.MET G:\Org\HP5\Methods\DS_ORO-BE-L%.MET	1000	1	1	1	0	The integration of Oil Range hydrocarbon is the hydrocarbon response with reference to the baseline. Surrogates are integrated using a valley to valley integration using Set baseline All Valleys on placed at 16.14 minutes and X-axis scaling showing surrogate peak from 14-18 minutes.
G:\org\HP5\DAT\HP5021522_b\0215HP5.64r	DCM-Baseline Check-V64	G:\Org\HP5\Methods\DR_8015-JA-LEXP.met	1	1	1	1	0	No integration
G:\org\HP5\DAT\HP5021522_b\0215HP5.65r	LCS-163748-RRO_0215HP5_ SGT	G:\Org\HP5\Methods\D3_ORO-BE-L%.MET G:\Org\HP5\Methods\DS_ORO-BE-L%.MET	1000	1	1	1	0	The integration of Oil Range hydrocarbon is the hydrocarbon response with reference to the baseline. Surrogates are integrated using a valley to valley integration using Set baseline All Valleys on placed at 16.14 minutes and X-axis scaling showing surrogate peak from 14-18 minutes.
G:\org\HP5\DAT\HP5021522_b\0215HP5.66r	MARKER_0215HP566r_DRO_0215HP5_DRO220128B	G:\org\HP5\Methods\CSC220215.met	1	1	1	1	0	No integration
G:\org\HP5\DAT\HP5021522_b\0215HP5.67r	CCV_0215HP567r_RRO_0215HP5_DRO220201A	G:\Org\HP5\Methods\DC_ORO-BE-L%.MET G:\Org\HP5\Methods\DS_ORO-BE-L%.MET	1	1	1	1	0	The integration of Oil Range hydrocarbon is the hydrocarbon response with reference to the baseline. Surrogates are integrated using a valley to valley integration using Set baseline All Valleys on placed at 16.14 minutes and X-axis scaling showing surrogate peak from 14-18 minutes.
G:\org\HP5\DAT\HP5021522_b\0215HP5.68r	CCV_0215HP568r_DRO_0215HP5_DRO220128A	G:\Org\HP5\Methods\DC_8015-C24-JEb-L%.met G:\Org\HP5\Methods\DS_8015-C24-JEb-L#.met	1	1	1	1	0	The integration of Diesel Range Organics (C10-C24) and Total Extractable Hydrocarbons is the hydrocarbon response with reference to the baseline. Assigned Set Baseline All Valley on at 16.35 minutes. Surrogates are integrated using a valley to valley integration using Set baseline All Valleys on placed at 12.01 minutes and X-axis scaling showing surrogate peak from 11.5-14 minutes.

Jillian Bostwick
Analyst

Digitally signed by
Jillian L. Bostwick
Date: 2022.02.18 12:05:11 -07:00



Analytical RunID GCFID-HP5-B_220111A Standards Traceability Report

Standard ID: DRO180126C

Standard Name: 2-Fluorobiphenyl

Prep Date: 1/26/2018

Exp Date: 10/31/2024

Department: dropr

Vendor: Chemservice

Lot Number: 5599700

Balance ID:

Comments:

Type: Neat

Prep By: Todd C Cooper

Status: New

Final Volume: mL

Chemical/Solvent Used	Bottle No	Amt	Units	Expires
2-Fluorobiphenyl	10069		mL	10/31/2024
Stock Source	Base Units	Amount Added		



Analytical RunID GCFID-HP5-B_220111A Standards Traceability Report

Standard ID: DRO180823A

Standard Name: 2-Bromonaphthalene

Prep Date: 8/22/2016

Exp Date: 5/31/2022

Department: dropr

Vendor: Chemservice

Lot Number: 3150700

Balance ID:

Comments:

Type: Neat

Prep By: Ann Nebel

Status: New

Final Volume: mL

Chemical/Solvent Used	Bottle No	Amt	Units	Expires
2-Bromonaphthalene	10701		mL	5/31/2022
Stock Source	Base Units	Amount Added		



Analytical RunID GCFID-HP5-B_220111A Standards Traceability Report

Standard ID: DRO181105A

Standard Name: #2 Diesel (NEAT)

Prep Date: 11/5/2018

Exp Date: 11/5/2023

Department: dropr

Vendor: conoco

Lot Number:

Balance ID:

Comments: -18 Cloud peak. (Conoco Gas Sation 1240 S. 27th Billings, MT) 2nd Source

Type: Neat

Prep By: Ann Nebel

Status: New

Final Volume: 250 mL

Chemical/Solvent Used	Bottle No	Amt	Units	Expires
				11/5/2023
Stock Source	Base Units	Amount Added		



Analytical RunID GCFID-HP5-B_220111A Standards Traceability Report

Standard ID: DRO200430B
Standard Name: O-Terphenyl
Prep Date: 4/30/2020
Exp Date: 9/30/2024
Department: dropr
Vendor: Chemservice
Lot Number: 9972100
Balance ID:
Comments: ID#: 6271

Type: Neat
Prep By: Ann Nebel
Status: New

Final Volume: mL

Chemical/Solvent Used	Bottle No	Amt	Units	Expires
o-Terphenyl	12650	500	mg	9/30/2024
Stock Source	Base Units	Amount Added		



Analytical RunID GCFID-HP5-B_220111A Standards Traceability Report

Standard ID: DRO201014C

Standard Name: 1-Chlorooctadecane

Prep Date: 10/14/2019

Exp Date: 12/31/2024

Department: dropr

Vendor: CSI1

Lot Number: 10809500

Balance ID:

Comments: Date Certified: 12/9/16 ; N-10042-1G; 99.5% purity

Type: Neat

Prep By: Ann Nebel

Status: Open

Final Volume: mL

Chemical/Solvent Used	Bottle No	Amt	Units	Expires
1-Chlorooctadecane	13192	1	g	12/31/2024

Stock Source	Base Units	Amount Added
--------------	------------	--------------



Analytical RunID GCFID-HP5-B_220111A Standards Traceability Report

Standard ID: DRO201014D

Standard Name: n-Pentacosane

Prep Date: 10/14/2020

Exp Date: 2/28/2025

Department: dropr

Vendor: Chem Service

Lot Number: 9642200

Balance ID:

Comments: C-25; Used in AKDRO Marker

Type: Neat

Prep By: Ann Nebel

Status: New

Final Volume: mL

Chemical/Solvent Used	Bottle No	Amt	Units	Expires
n-Pentacosane	13193	100	mg	2/28/2025
Stock Source	Base Units	Amount Added		



Analytical RunID GCFID-HP5-B_220111A Standards Traceability Report

Standard ID: DRO211012B

Standard Name: #2 Diesel in Acetone 150,000 ug/mL

Prep Date: 10/12/2021

Exp Date: 11/5/2023

Department: dropr

Vendor:

Lot Number:

Balance ID: BAL-DRO

Comments: #2 Diesel in Acetone 150,000 ug/mL.

Type: Secondary

Prep By: Ann Nebel

Status: New

Final Volume: 25 mL

Chemical/Solvent Used	Bottle No	Amt	Units	Expires
Acetone EA662	14050	25	mL	11/5/2023
Stock Source	Base Units	Amount Added		
DRO181105A	ug/mL	3.7507 g		



Analytical RunID GCFID-HP5-B_220111A Standards Traceability Report

Standard ID: DRO211025B

Standard Name: Ali Hydro Std 1000ug/mL

Prep Date: 10/25/2021

Exp Date: 11/30/2024

Department: dropr

Vendor: Agilent

Lot Number: 0006643302

Balance ID:

Comments: Ali Hydro Std 1000ug/mL For CCVs.

Type: Primary

Prep By: Ann Nebel

Status: New

Final Volume: mL

Chemical/Solvent Used	Bottle No	Amt	Units	Expires
Aliphatic Hydrocarbon Standard	14434	1	mL	11/30/2024
Stock Source	Base Units	Amount Added		



Analytical RunID GCFID-HP5-B_220111A Standards Traceability Report

Spike ID: DRO211101A

Spike Name: OTP-4000 ug/mL DCM

Prep Date: 11/1/2021

Exp Date: 9/30/2024

Department: dropr

Vendor:

Lot Number:

Balance ID: BAL-DRO

Comments: Used to Prep DRO-8015 ICAL and CCV Solutions

Type: Secondary

Prep By: Ann Nebel

Status: Open

Final Volume: 25 mL

Chemical/Solvent Used	Bottle No	Amt	Units	Expires
Dichloromethane EC328	14408	25	mL	9/30/2024

Stock Source	Base Units	Amount Added
DRO200430B	ug/mL	0.1012 g



Analytical RunID GCFID-HP5-B_220111A Standards Traceability Report

Standard ID: DRO211214C

Standard Name: Diesel Fuel #2 50,000 ug/mL in DCM

Prep Date: 12/14/2021

Exp Date: 4/30/2023

Department: dropr

Vendor: Sigma-Aldrich

Lot Number: LRAC6316

Balance ID:

Comments: Diesel Fuel #2 For CCVs.

Type: Primary

Prep By: Ann Nebel

Status: New

Final Volume: mL

Chemical/Solvent Used	Bottle No	Amt	Units	Expires
Diesel Fuel No. 2	14623	1	mL	4/30/2023
Stock Source	Base Units	Amount Added		
DRO211214C	ug/mL			



Analytical RunID GCFID-HP5-B_220111A Standards Traceability Report

Standard ID: DRO211222B

Standard Name: EPH (4) SURR-1000 ug/mL ea. in Hexane

Prep Date: 12/22/2021

Exp Date: 5/31/2022

Department: dropr

Vendor:

Lot Number:

Balance ID: BAL-DRO

Comments: EPH (4) SURR-1000 ug/mL ea. in Hexane

Type: Secondary

Prep By: Jillian L Bostwick

Status: Open

Final Volume: 50 mL

Chemical/Solvent Used	Bottle No	Amt	Units	Expires
Hexane EB754	14543	50	mL	5/31/2022

Stock Source	Base Units	Amount Added
DRO180823A	ug/mL	0.0507 g
DRO200430B	ug/mL	0.0504 g
DRO180126C	ug/mL	0.0496 g
DRO201014C	ug/mL	0.0504 g



Analytical RunID GCFID-HP5-B_220111A Standards Traceability Report

Standard ID: DRO220102D

Standard Name: ALASKA MARKER-200ug/mL

Prep Date: 1/2/2022

Exp Date: 5/31/2022

Department: dropr

Vendor:

Lot Number:

Balance ID:

Comments: ALASKA MARKER w/ C-10, C-25, and OTP/COD. Optimal C-25 is 0.0012g.

Type: Secondary

Prep By: Ann Nebel

Status: New

Final Volume: 5.5 mL

Chemical/Solvent Used	Bottle No	Amt	Units	Expires
Hexane EB754	14543	3.3	mL	5/31/2022

Stock Source	Base Units	Amount Added
DRO201014D	ug/mL	0.0016 g
DRO211222B	ug/mL	1.1 mL
DRO211025B	ug/mL	1.1 mL

Certificate of Analysis

Diesel Fuel No. 2

Certified
Reference
Material

Description

Product ID UST148
Lot LRAC6316
Expiration Date April 2023
Manufacturing Date April 2020
Storage Conditions Room Temperature
Solvent/Matrix DICHLOROMETHANE

ID #: 14623

Opened: _____

Diesel Fuel No. 2

Expires: 4/30/2023


Rec'd: 12/14/2021

Energy Laboratories Inc 1120 So. 27th Street
Billings MT 59107

Certified Values

Analyte	Certified Value ^{1,4}	Units	Raw Material Purity, %	Raw Material Lot	CAS
NO.2 FUEL OIL	50001 ± 2770	µg/mL	100.0	LA80505	68476-34-6

Informational Values



Additional Information:

Analytical Method Parameters:

Column: SPB-5, 30 m × 0.53 mm I.D., 1.5 µm film thickness (Column #214)

Carrier Gas: H₂, Flow: 4.0 mL/min

Inlet Temperature: 250 °C, Injection Volume: 1.0 µL

Injection Mode: Split, Split Ratio: 10: 1

Temperature Program: 40 °C (Hold 2 min) @ 15 °C/min to 300 °C (Hold 5 min)

Detector: FID

Detector Temperature: 300 °C



SIGMA-ALDRICH

2931 Soldier Springs Rd. Laramie, Wyoming 82070 USA
800-325-5832
TechService@milliporesigma.com www.sigma-aldrich.com

Description

Lot **LRAC6316**
Expiration Date April 2023
Manufacturing Date April 2020
Storage Conditions Room Temperature
Solvent/Matrix DICHLOROMETHANE

1 Metrological traceability: Traceable to the SI and higher order standards from NIST through an unbroken chain of comparisons. The balance used to weigh raw materials is accurate to +/-0.0001 g and calibrated regularly using mass standards traceable to NIST. All dilutions were performed gravimetrically. Additionally, individual analytes are traceable to NIST SRMs where available and specified above.
4 Ucrm - Uncertainty values in this document are expressed as Expanded Uncertainty (Ucrm) corresponding to the 95% confidence interval. Ucrm is derived from the combined standard uncertainty multiplied by the coverage factor k, which is obtained from a t-distribution and degrees of freedom. The components of combined standard uncertainty include the uncertainties due to characterization, homogeneity, long term stability, and short term stability (transport). The components due to stability are generally considered to be negligible unless otherwise indicated by stability studies. The mathematical representation of the Ucrm calculation is as follows:

$$U_{CRM} = \sqrt{u_{char}^2 + u_{homogeneity}^2 + u_{stability}^2}$$

k: Coverage factor derived from a t-distribution table, based on the degrees of freedom of the data set. Assume 2.0 for a **Confidence interval = 95%**

6 Analytical Value- For QC verification of the certified value only- not to be used in calculations. Represents the analytical data obtained by comparison to a standard as analyzed by the method described in the CoA or another acceptable method. The result may differ from the certified value and UCRM based on method uncertainty as well as the uncertainty associated with the standard used for comparison.


Traceability: The standard was manufactured under an ISO/IEC 17025:2017 certified quality system. The balance used to weigh raw materials is accurate to +/- 0.0001g and calibrated regularly using mass standards traceable to NIST. All dilutions were performed gravimetrically. Additionally, individual analytes are traceable to NIST SRMs where available and specified above.

Homogeneity: Homogeneity was assessed in accordance with ISO 17034:2016. Completed units were sampled using a random stratified sampling protocol. The results of chemical analysis were then compared using a one-way analysis of variance approach as described by TNI EL-V3-2009 Appendix A.2. See Instructions for minimum sub-sample size.

Expiration is at end of month given on certificate and label.


MSDS reports for components comprising greater than 1.0% of the solution or 0.1% for components known to be carcinogens are available upon request.

THIS PRODUCT WAS DESIGNED, PRODUCED AND VERIFIED FOR ACCURACY AND STABILITY IN ACCORDANCE WITH ISO/IEC 17025:2017 (ANAB Cert AT-1467) and ISO 17034:2016 (ANAB Cert AR-1470).



Andy Ommen - QC Manager

Certification Date April 30, 2020
Version 0-4302020



Mark Pooler - QA Supervisor



660 Tower Lane • P.O. Box 599 • West Chester, PA 19381-0599
1-800-452-9994 • 1-610-692-3026 • Fax 1-610-692-8729
info@chemservice.com • www.chemservice.com

CERTIFICATE OF ANALYSIS

o-Terphenyl

CATALOG NUMBER N-12693-500MG
LOT NUMBER 9972100
DATE CERTIFIED 09/23/19
EXPIRATION DATE 09/30/24
CAS NUMBER 84-15-1
MOLECULAR FORMULA C18H14
MOLECULAR WEIGHT 230.32
STORAGE Store in a cool dry place.
HANDLING See Safety Data Sheet
INTENDED USE For laboratory use only.

Analytical Test	Value
FT-IR SPECTROSCOPY	CONFORMS TO STRUCTURE
GC/MS SPECTRA ID	MATCHES NIST DATABASE
MELTING POINT (°C)	57.1
% PURITY (GC/FID)	99.5

Chem Service, Inc. guarantees the purity to be +/- 0.5% deviation prior to the expiration date shown on the label and exclusive of any customer contamination.

Certified By:

Mary Beth O'Donnell

Mary Beth O'Donnell
CSM/TC

ID #: 12650

Opened: _____

o-Terphenyl

Expires: 9/30/2024

Rec'd: 4/30/2020

Energx Laboratories Inc 1120 So. 27th Street

Billings MT 59107


Chem Service is accredited to ISO 17034:2016, ISO/IEC 17025:2017 and certified to ISO 9001:2015



CERTIFICATE OF ANALYSIS

Analysis Method:

Catalog Number: N-12693-500MG
Description: o-Terphenyl
Lot Number: 9972100
Expiration Date: 09/30/24



Chem Service is accredited to ISO 17034:2016, ISO/IEC 17025:2017 and certified to ISO 9001:2015



660 Tower Lane • P.O. Box 599 • West Chester, PA 19381-0599
1-800-452-9994 • 1-610-692-3026 • Fax 1-610-692-8729
info@chemservice.com • www.chemservice.com

CERTIFICATE OF ANALYSIS

Analysis Method:

Catalog Number: N-12693-500MG
Description: o-Terphenyl
Lot Number: 9972100
Expiration Date: 09/30/24

Chem Service Inc Area Percent Report

Data File: D:\msdchem\2019 DATA\0919\0923-01.D
Acq On : 23 Sep 2019 10:40
Operator :
Sample : n-12693
Misc :
ALS Vial : 95

Integration Parameters: autoint1.e
Integrator: ChemStation

DataAcq Meth: SCREEN.M
Method : D:\msdchem\2019 DATA\0919\0903-09.D\ERIN.M

Signal : TIC: 0923-01.D\data.ms

peak #	R.T. min	first scan	max scan	last scan	PK TY	peak height	corr. area	corr. % max.	% of total
1	11.844	1597	1606	1613	BB	32038221	432253484	100.00%	100.000%

Sum of corrected areas: 432253484

ERIN.M Mon Sep 23 10:55:51 2019

Chem Service is accredited to ISO 17034:2016, ISO/IEC 17025:2017 and certified to ISO 9001:2015




660 Tower Lane • P.O. Box 599 • West Chester, PA 19381-0599
1-800-452-9994 • 1-610-692-3026 • Fax 1-610-692-8729
info@chemservice.com • www.chemservice.com

CERTIFICATE OF ANALYSIS

Analysis Method:

Catalog Number: N-12693-500MG
Description: o-Terphenyl
Lot Number: 9972100
Expiration Date: 09/30/24

Abundance



Time-->

Chem Service is accredited to ISO 17034:2016, ISO/IEC 17025:2017 and certified to ISO 9001:2015




CERTIFICATE OF ANALYSIS

Analysis Method:

Catalog Number: N-12693-500MG
Description: o-Terphenyl
Lot Number: 9972100
Expiration Date: 09/30/24

Abundance



Chem Service is accredited to ISO 17034:2016, ISO/IEC 17025:2017 and certified to ISO 9001:2015.



660 Tower Lane • P.O. Box 599 • West Chester, PA 19381-0599
1-800-452-9994 • 1-610-692-3026 • Fax 1-610-692-8729
info@chemservice.com • www.chemservice.com

CERTIFICATE OF ANALYSIS

Analysis Method:

Catalog Number:	N-12693-500MG
Description:	o-Terphenyl
Lot Number:	9972100
Expiration Date:	09/30/24

Chem Service is accredited to ISO 17034:2016, ISO/IEC 17025:2017 and certified to ISO 9001:2015




660 Tower Lane • P.O. Box 599 • West Chester, PA 19381-0599
 1-800-452-9994 • 1-610-692-3026 • Fax 1-610-692-8729
info@chemservice.com • www.chemservice.com

Gas

Data file: C:\CHEM3\
Sample name: N-12893
Instrument: GC 2
Injection date: 8/23/2019 9:58:34 AM
Acq. method: SCREEN.M
Column name: HP-5

CERTIFICATE OF ANALYSIS

Location: Vial 141
Injection volume: 1.0uL



Signal: FID1 A, Front Signal

RT [min]	Type	Width [min]	Area	Height	Area%
4.565	BB	0.0305	1.2408	0.5122	0.11
20.152	BB	0.0391	1171.9556	439.4599	99.89
		Sum	1173.1963		

Chem Service is accredited to ISO 17034:2016, ISO/IEC 17025:2017 and certified to ISO 9001:2015





Analytical RunID GCFID-HP5-B_220111c Standards Traceability Report

Standard ID: DRO210406A

Standard Name: Triacontane-d62 Surr For AK103 RRO

Prep Date: 4/6/2021

Exp Date: 4/6/2026

Department: dropr

Vendor: Sigma-Aldrich

Lot Number: MBBC4347

Balance ID:

Comments: Alaska surr [for AK103 RRO]

Type: Neat

Prep By: Ann Nebel

Status: New

Final Volume: mL

Chemical/Solvent Used	Bottle No	Amt	Units	Expires
Triacontane-d62-98 atom % D	13736		mL	4/6/2026
Stock Source	Base Units	Amount Added		



Analytical RunID GCFID-HP5-B_220111c Standards Traceability Report

Standard ID: DRO210901A

Standard Name: 30W Motor Oil-Valvoline

Prep Date: 9/1/2021

Exp Date: 9/1/2026

Department: dropr

Vendor:

Lot Number: F1620C1

Balance ID:

Comments: Used to make 2nd Source Standard for AK103 method.

Type: Primary

Prep By: Jillian L Bostwick

Status: New

Final Volume: mL

Chemical/Solvent Used	Bottle No	Amt	Units	Expires
Valvoline SAE 30 Motor Oil	14232		mL	9/1/2026
Stock Source	Base Units	Amount Added		



Analytical RunID GCFID-HP5-B_220111c Standards Traceability Report

Standard ID: DRO210901B

Standard Name: 40W Motor Oil-Valvoline

Prep Date: 9/1/2021

Exp Date: 9/1/2026

Department: dropr

Vendor:

Lot Number: L0717H2

Balance ID:

Comments: Used to Make 2nd Source Standards For Alaska AK103 RRO Method and Oil

Type: Primary

Prep By: Jillian L Bostwick

Status: New

Final Volume: mL

Chemical/Solvent Used	Bottle No	Amt	Units	Expires
Valvoline SAE 40 Motor Oil	14231		mL	9/1/2026
Stock Source	Base Units	Amount Added		



Analytical RunID GCFID-HP5-B_220111c Standards Traceability Report

Standard ID: DRO210902A

Standard Name: 50,000 ug/mL Oil Std for RRO-In DCM

Prep Date: 9/2/2021

Exp Date: 9/1/2026

Department: dropr

Vendor:

Lot Number:

Balance ID: BAL-DRO

Comments: .625 g of 30W and 40 W each LCS for Oil range

Type: Secondary

Prep By: Jillian L Bostwick

Status: New

Final Volume: 25 mL

Chemical/Solvent Used	Bottle No	Amt	Units	Expires
Dichloromethane EB867	14196	25	mL	9/1/2026

Stock Source	Base Units	Amount Added
DRO210901A	ug/mL	0.6254 g
DRO210901B	ug/mL	0.6261 g



Analytical RunID GCFID-HP5-B_220111c Standards Traceability Report

Standard ID: DRO211006A

Standard Name: Triacontane SURR 2000 ug/mL

Prep Date: 10/6/2021

Exp Date: 4/6/2026

Department: dropr

Vendor:

Lot Number:

Balance ID: BAL-DRO

Comments: Triacontane SURR 2000 ug/mL

Type: Secondary

Prep By: Jillian L Bostwick

Status: New

Final Volume: 50 mL

Chemical/Solvent Used	Bottle No	Amt	Units	Expires
Acetone DZ509	13553	50	mL	4/6/2026

Stock Source	Base Units	Amount Added
DRO210406A	ug/mL	0.1001 g



Analytical RunID GCFID-HP5-B_220111c Standards Traceability Report

Standard ID: DRO211118A

Standard Name: 50,000 ug/mL Oil Std For AK103 RRO-In DCM

Prep Date: 11/18/2021

Exp Date: 10/31/2028

Department: dropr

Vendor: Restek

Lot Number: A0176667

Balance ID: Sartorius 4 place balance

Comments:

Type: Primary

Prep By: Ann Nebel

Status: Open

Final Volume: 1 mL

Chemical/Solvent Used	Bottle No	Amt	Units	Expires
Residual Range Calibration Standard	14531	1	mL	10/31/2028

Stock Source	Base Units	Amount Added
DRO211118A	ug/mL	



CERTIFIED REFERENCE MATERIAL

110 Benner Circle
Bellefonte, PA 16823-8812
Tel: (800)356-1688
Fax: (814)353-1309

www.restek.com

Certificate of Analysis



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31817 **Lot No.:** A0176667

Description : Residual Range Calibration Standard (RCS)

Residual Range Calib Std (RCS) 50,000µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : October 31, 2028 **Storage:** 25°C nominal

Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	Grav. Conc. (weight/volume)	Expanded Uncertainty (95% C.L.; K=2)		
1	Motor Oil SAE30 & SAE40 Blend (Pennzoil) CAS # 64742-65-0.F Purity ----%	50,102.0 µg/mL	+/- 293.3582	µg/mL	Gravimetric
	(Lot A0126386)		+/- 1,492.1008	µg/mL	Unstressed
			+/- 1,591.3244	µg/mL	Stressed

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

ID #: 14531

Opened: _____

Residual Range Calibration Standard

Expires: 10/31/2028

Rec'd: 11/18/2021

Energ Laboratories Inc 1120 So. 27th Street

Billings MT 59107

Column:

30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)

Carrier Gas:

hydrogen-constant pressure 10 psi.

Temp. Program:

40°C (hold 2 min.) to 330°C
@ 10°C/min. (hold 10 min.)

Inj. Temp:


250°C

Det. Temp:

330°C

Det. Type:

FID



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Sam Moodler

Sam Moodler - Operations Tech I

Date Mixed: 22-Sep-2021

Balance: 1128360905

Alexis Shelow

Alexis Shelow - Operations Tech I

Date Passed: 23-Sep-2021

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/ μ ECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified combined stressed uncertainty value (includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ stressed} = k \sqrt{U_{gravimetric}^2 + U_{homogeneity}^2 + U_{storage\ stability}^2 + U_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- It is important to note that the shipping stability uncertainty was obtained under temperature extremes for specific time intervals; therefore, the certified combined stressed uncertainty value should only be applied to the product if it was stored at non-standard temperature conditions up to and including 7 days. Contact Restek Technical Service at www.restek.com/Contact-Us for use recommendations if your shipment was in-transit for more than 7 days at non-standard temperature conditions.
- Apply the certified combined unstressed uncertainty value if the product was received under standard shipping conditions. Apply the certified combined stressed uncertainty value if the product was received under non-standard conditions as specified below.

Label Conditions	Standard Conditions	Non-Standard Conditions
25°C Nominal (Room Temperature)	< 60°C	≥ 60°C up to 7 days
10°C or colder (Refrigerate)	< 40°C	≥ 40°C up to 7 days
0°C or colder (Freezer) -20°C or colder (Deep Freezer)	< 25°C	≥ 25°C up to 7 days

- Separate (not combined) uncertainty values for gravimetric uncertainty are also displayed on the certificate, if needed, separate homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty values are available by contacting Restek Technical Service at www.restek.com/Contact-Us.
- The packaged amount is the minimum sample size for which uncertainty is valid. The ampules are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampules. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.

3050 Spruce Street, Saint Louis, MO 63103, USA

Website: www.sigmaaldrich.com


Email USA: techserv@sial.com

Outside USA: eurtechserv@sial.com

Certificate of Analysis

Product Name:
Triacontane-d62 - 98 atom % D

Product Number: 451789
 Batch Number: MBBC4347
 Brand: ALDRICH
 CAS Number: 93952-07-9
 MDL Number: MFCD00209794
 Formula: C30D62
 Formula Weight: 485.20 g/mol
 Quality Release Date: 27 APR 2018



ID #: 13736

Opened: _____


Triacontane-d62-98 atom % D

Expires: 4/6/2026

Rec'd: 4/6/2021

Energx Laboratories Inc 1120 So. 27th Street
 Billings MT 59107

Test	Specification	Result
Purity (HPLC)	≥ 99.0 %	99.0 %
Proton NMR Spectrum	Conforms to Structure	Conforms
D Enrichment	≥ 98.0 %	99.0 %
Initial Melting Point		60.0 °C
Final Melting Point		62.0 °C



Greg Abernathy, Supervisor
 Quality Control
 Miamisburg, Ohio US

Sigma-Aldrich warrants, that at the time of the quality release or subsequent retest date this product conformed to the information contained in this publication. The current Specification sheet may be available at Sigma-Aldrich.com. For further inquiries, please contact Technical Service. Purchaser must determine the suitability of the product for its particular use. See reverse side of invoice or packing slip for additional terms and conditions of sale.



Prep Batch 163748 Standards Traceability Report

Spike ID: DRO181105A

Spike Name: #2 Diesel (NEAT)

Prep Date: 11/5/2018

Exp Date: 11/5/2023

Department: dropr

Vendor: conoco

Lot Number:

Balance ID:

Comments: -18 Cloud peak. (Conoco Gas Sation 1240 S. 27th Billings, MT) 2nd Source

Type: Neat

Prep By: Ann Nebel

Status: New

Final Volume: 250 mL

Chemical/Solvent Used	Bottle No	Amt	Units	Expires
				11/5/2023
Stock Source	Base Units	Amount Added		



Prep Batch 163748 Standards Traceability Report

Spike ID: DRO200430B

Spike Name: O-Terphenyl

Prep Date: 4/30/2020

Exp Date: 9/30/2024

Department: dropr

Vendor: Chemservice

Lot Number: 9972100

Balance ID:

Comments: ID#: 6271

Type: Neat

Prep By: Ann Nebel

Status: New

Final Volume: mL

Chemical/Solvent Used	Bottle No	Amt	Units	Expires
o-Terphenyl	12650	500	mg	9/30/2024
Stock Source	Base Units	Amount Added		



Prep Batch 163748 Standards Traceability Report

Spike ID: DRO210406A

Spike Name: Triacontane-d62 Surr For AK103 RRO

Type: Neat

Prep Date: 4/6/2021

Prep By: Ann Nebel

Exp Date: 4/6/2026

Status: New

Department: dropr

Vendor: Sigma-Aldrich

Final Volume: mL

Lot Number: MBBC4347

Balance ID:

Comments: Alaska surr [for AK103 RRO]

Chemical/Solvent Used	Bottle No	Amt	Units	Expires
Triacontane-d62-98 atom % D	13736	500	mg	4/6/2026

Stock Source	Base Units	Amount Added
--------------	------------	--------------



Prep Batch 163748 Standards Traceability Report

Spike ID: DRO210901A

Spike Name: 30W Motor Oil-Valvoline

Prep Date: 9/1/2021

Exp Date: 9/1/2026

Department: dropr

Vendor:

Lot Number: F1620C1

Balance ID:

Comments: Used to make 2nd Source Standard for AK103 method.

Type: Primary

Prep By: Jillian L Bostwick

Status: New

Final Volume: mL

Chemical/Solvent Used	Bottle No	Amt	Units	Expires
Valvoline SAE 30 Motor Oil	14232		mL	9/1/2026
Stock Source	Base Units	Amount Added		



Prep Batch 163748 Standards Traceability Report

Spike ID: DRO210901B

Spike Name: 40W Motor Oil-Valvoline

Prep Date: 9/1/2021

Exp Date: 9/1/2026

Department: dropr

Vendor:

Lot Number: L0717H2

Balance ID:

Comments: Used to Make 2nd Source Standards For Alaska AK103 RRO Method and Oil

Type: Primary

Prep By: Jillian L Bostwick

Status: New

Final Volume: mL

Chemical/Solvent Used	Bottle No	Amt	Units	Expires
Valvoline SAE 40 Motor Oil	14231		mL	9/1/2026
Stock Source	Base Units	Amount Added		



Prep Batch 163748 Standards Traceability Report

Spike ID: DRO211006A

Spike Name: Triacontane SURR 2000 ug/mL

Prep Date: 10/6/2021

Exp Date: 4/6/2026

Department: dropr

Vendor:

Lot Number:

Balance ID: BAL-DRO

Comments: Triacontane SURR 2000 ug/mL

Type: Primary

Prep By: Jillian L Bostwick

Status: New

Final Volume: 50 mL

Chemical/Solvent Used	Bottle No	Amt	Units	Expires
Acetone DZ509	13553	50	mL	4/6/2026

Stock Source	Base Units	Amount Added
DRO210406A	ug/mL	0.1001 g



Prep Batch 163748 Standards Traceability Report

Spike ID: DRO211213A

Spike Name: OTP only SURR 2000 ug/mL

Prep Date: 12/13/2021

Exp Date: 9/30/2024

Department: dropr

Vendor:

Lot Number:

Balance ID: BAL-DRO

Comments: OTP SURR 2000 ug/mL

Type: Primary

Prep By: Jillian L Bostwick

Status: New

Final Volume: 100 mL

Chemical/Solvent Used	Bottle No	Amt	Units	Expires
Acetone DZ509	13553	100	mL	9/30/2024

Stock Source	Base Units	Amount Added
DRO200430B	ug/mL	0.2015 g



Prep Batch 163748 Standards Traceability Report

Spike ID: DRO220106C

Spike Name: #2 Diesel in Acetone 150,000 ug/mL

Type: Secondary

Prep Date: 1/6/2022

Prep By: Ann Nebel

Exp Date: 11/5/2023

Status: New

Department: dropr

Vendor:

Final Volume: 25 mL

Lot Number:

Balance ID: BAL-DRO

Comments:

Chemical/Solvent Used	Bottle No	Amt	Units	Expires
Acetone DZ509	13553	25	mL	11/5/2023

Stock Source	Base Units	Amount Added
DRO181105A	ug/mL	3.7506 g



Prep Batch 163748 Standards Traceability Report

Spike ID: DRO220112A

Spike Name: 50,000 ug/mL Oil Std for RRO-In DCM

Prep Date: 1/12/2022

Exp Date: 9/1/2026

Department: dropr

Vendor:

Lot Number:

Balance ID: BAL-DRO

Comments: .625 g of 30W and 40 W each LCS for Oil range

Type: Secondary

Prep By: Jillian L Bostwick

Status: New

Final Volume: 25 mL

Chemical/Solvent Used	Bottle No	Amt	Units	Expires
Dichloromethane EC832	14647	25	mL	9/1/2026

Stock Source	Base Units	Amount Added
DRO210901A	ug/mL	0.6225 g
DRO210901B	ug/mL	0.6273 g



Prep Batch 163748 Standards Traceability Report

Spike ID: DRO220119A

Spike Name: Triacontane SURR 1000 ug/mL

Prep Date: 1/19/2022

Exp Date: 4/6/2026

Department: dropr

Vendor:

Lot Number:

Balance ID: BAL-DRO

Comments: 2X dilution of Triacontane SURR 2000 ug/mL

Type: Secondary

Prep By: Jillian L Bostwick

Status: New

Final Volume: 10 mL

Chemical/Solvent Used	Bottle No	Amt	Units	Expires
Dichloromethane EC849	14747	5	mL	4/6/2026

Stock Source	Base Units	Amount Added
DRO211006A	ug/mL	5 mL

Anna

660 Tower Lane • P.O. Box 599 • West Chester, PA 19381-0599
1-800-452-9994 • 1-610-692-3026 • Fax 1-610-692-8729
info@chemservice.com • www.chemservice.com

CERTIFICATE OF ANALYSIS

o-Terphenyl

CATALOG NUMBER N-12693-500MG
LOT NUMBER 9972100
DATE CERTIFIED 09/23/19
EXPIRATION DATE 09/30/24
CAS NUMBER 84-15-1
MOLECULAR FORMULA C18H14
MOLECULAR WEIGHT 230.32
STORAGE Store in a cool dry place.
HANDLING See Safety Data Sheet
INTENDED USE For laboratory use only.

Analytical Test	Value
FT-IR SPECTROSCOPY	CONFORMS TO STRUCTURE
GC/MS SPECTRA ID	MATCHES NIST DATABASE
MELTING POINT (°C)	57.1
% PURITY (GC/FID)	99.5

Chem Service, Inc. guarantees the purity to be +/- 0.5% deviation prior to the expiration date shown on the label and exclusive of any customer contamination.

Certified By:

Mary Beth O'Donnell

Mary Beth O'Donnell
CSM/TC

ID #: 12650

Opened: _____

o-Terphenyl

Expires: 9/30/2024

Rec'd: 4/30/2020

Energyl Laboratories Inc 1120 So. 27th Street

Billings MT 59107

COA Form
Revision 3 (3/2015)


Chem Service is accredited to ISO 17034:2016, ISO/IEC 17025:2017 and certified to ISO 9001:2015



CERTIFICATE OF ANALYSIS

Analysis Method:

Catalog Number: N-12693-500MG
Description: o-Terphenyl
Lot Number: 9972100
Expiration Date: 09/30/24



Chem Service is accredited to ISO 17034:2016, ISO/IEC 17025:2017 and certified to ISO 9001:2015



CERTIFICATE OF ANALYSIS

Analysis Method:

Catalog Number: N-12693-500MG
Description: o-Terphenyl
Lot Number: 9972100
Expiration Date: 09/30/24

Chem Service Inc Area Percent Report

Data File: D:\msdchem\2019 DATA\0919\0923-01.D
Acq On : 23 Sep 2019 10:40
Operator :
Sample : n-12693
Misc :
ALS Vial : 95

Integration Parameters: autoint1.e
Integrator: ChemStation

DataAcq Meth: SCREEN.M
Method : D:\msdchem\2019 DATA\0919\0903-09.D\ERIN.M

Signal : TIC: 0923-01.D\data.ms

peak #	R.T. min	first scan	max scan	last scan	PK TY	peak height	corr. area	corr. % max.	% of total
1	11.844	1597	1606	1613	BB	32038221	432253484	100.00%	100.000%

Sum of corrected areas: 432253484

ERIN.M Mon Sep 23 10:55:51 2019


660 Tower Lane • P.O. Box 599 • West Chester, PA 19381-0599
1-800-452-9994 • 1-610-692-3026 • Fax 1-610-692-8729
info@chemservice.com • www.chemservice.com

CERTIFICATE OF ANALYSIS

Analysis Method:

Catalog Number: N-12693-500MG
Description: o-Terphenyl
Lot Number: 9972100
Expiration Date: 09/30/24

Abundance



Chem Service is accredited to ISO 17034:2016, ISO/IEC 17025:2017 and certified to ISO 9001:2015




660 Tower Lane • P.O. Box 599 • West Chester, PA 19381-0599
1-800-452-9994 • 1-610-692-3026 • Fax 1-610-692-8729
info@chemservice.com • www.chemservice.com

CERTIFICATE OF ANALYSIS

Analysis Method:

Catalog Number: N-12693-500MG
Description: o-Terphenyl
Lot Number: 9972100
Expiration Date: 09/30/24

Abundance



Chem Service is accredited to ISO 17034:2016, ISO/IEC 17025:2017 and certified to ISO 9001:2015.



660 Tower Lane • P.O. Box 599 • West Chester, PA 19381-0599
1-800-452-9994 • 1-610-692-3026 • Fax 1-610-692-8729
info@chemservice.com • www.chemservice.com

CERTIFICATE OF ANALYSIS

Analysis Method:

Catalog Number:	N-12693-500MG
Description:	o-Terphenyl
Lot Number:	9972100
Expiration Date:	09/30/24

Chem Service is accredited to ISO 17034:2016, ISO/IEC 17025:2017 and certified to ISO 9001:2015




660 Tower Lane • P.O. Box 599 • West Chester, PA 19381-0599
 1-800-452-9994 • 1-610-692-3026 • Fax 1-610-692-8729
info@chemservice.com • www.chemservice.com

Gas

Data file: C:\CHEM3\
 Sample name: N-12893
 Instrument: GC 2
 Injection date: 8/23/2019 9:58:34 AM
 Acq. method: SCREEN.M
 Column name: HP-5

CERTIFICATE OF ANALYSIS

Location: Vial 141
 Injection volume: 1.0uL



Signal: FID1 A, Front Signal

RT [min]	Type	Width [min]	Area	Height	Area%
4.565	BB	0.0305	1.2408	0.5122	0.11
20.152	BB	0.0391	1171.9556	439.4599	99.89
		Sum	1173.1963		

Chem Service is accredited to ISO 17034:2016, ISO/IEC 17025:2017 and certified to ISO 9001:2015



3050 Spruce Street, Saint Louis, MO 63103, USA

Website: www.sigmaaldrich.com

Email USA: techserv@sial.com

Outside USA: eurtechserv@sial.com

Certificate of Analysis

Product Name:
Triacontane-d62 - 98 atom % D

Product Number: 451789
 Batch Number: MBBC4347
 Brand: ALDRICH
 CAS Number: 93952-07-9
 MDL Number: MFCD00209794
 Formula: C30D62
 Formula Weight: 485.20 g/mol
 Quality Release Date: 27 APR 2018



ID #: 13736

Opened: _____


Triacontane-d62-98 atom % D

Expires: 4/6/2026

Rec'd: 4/6/2021

Energx Laboratories Inc 1120 So. 27th Street
 Billings MT 59107

Test	Specification	Result
Purity (HPLC)	≥ 99.0 %	99.0 %
Proton NMR Spectrum	Conforms to Structure	Conforms
D Enrichment	≥ 98.0 %	99.0 %
Initial Melting Point		60.0 °C
Final Melting Point		62.0 °C



Greg Abernathy, Supervisor
 Quality Control
 Miamisburg, Ohio US

Sigma-Aldrich warrants, that at the time of the quality release or subsequent retest date this product conformed to the information contained in this publication. The current Specification sheet may be available at Sigma-Aldrich.com. For further inquiries, please contact Technical Service. Purchaser must determine the suitability of the product for its particular use. See reverse side of invoice or packing slip for additional terms and conditions of sale.



Analytical RunID GCFID-HP5-B_220215A Standards Traceability Report

Standard ID: DRO200430B
Standard Name: O-Terphenyl
Prep Date: 4/30/2020
Exp Date: 9/30/2024
Department: dropr
Vendor: Chemservice
Lot Number: 9972100
Balance ID:
Comments: ID#: 6271

Type: Neat
Prep By: Ann Nebel
Status: New

Final Volume: mL

Chemical/Solvent Used	Bottle No	Amt	Units	Expires
o-Terphenyl	12650	500	mg	9/30/2024
Stock Source	Base Units	Amount Added		



Analytical RunID GCFID-HP5-B_220215A Standards Traceability Report

Standard ID: DRO210406A

Standard Name: Triacontane-d62 Surr For AK103 RRO

Prep Date: 4/6/2021

Exp Date: 4/6/2026

Department: dropr

Vendor: Sigma-Aldrich

Lot Number: MBBC4347

Balance ID:

Comments: Alaska surr [for AK103 RRO]

Type: Neat

Prep By: Ann Nebel

Status: New

Final Volume: mL

Chemical/Solvent Used	Bottle No	Amt	Units	Expires
Triacontane-d62-98 atom % D	13736	500	mg	4/6/2026
Stock Source	Base Units	Amount Added		



Analytical RunID GCFID-HP5-B_220215A Standards Traceability Report

Standard ID: DRO211006A

Standard Name: Triacontane SURR 2000 ug/mL

Prep Date: 10/6/2021

Exp Date: 4/6/2026

Department: dropr

Vendor:

Lot Number:

Balance ID: BAL-DRO

Comments: Triacontane SURR 2000 ug/mL

Type: Primary

Prep By: Jillian L Bostwick

Status: New

Final Volume: 50 mL

Chemical/Solvent Used	Bottle No	Amt	Units	Expires
Acetone DZ509	13553	50	mL	4/6/2026
Stock Source	Base Units	Amount Added		
DRO210406A	ug/mL	0.1001 g		



Analytical RunID GCFID-HP5-B_220215A Standards Traceability Report

Spike ID: DRO211101A

Spike Name: OTP-4000 ug/mL DCM

Prep Date: 11/1/2021

Exp Date: 9/30/2024

Department: dropr

Vendor:

Lot Number:

Balance ID: BAL-DRO

Comments: Used to Prep DRO-8015 ICAL and CCV Solutions

Type: Secondary

Prep By: Ann Nebel

Status: Open

Final Volume: 25 mL

Chemical/Solvent Used	Bottle No	Amt	Units	Expires
Dichloromethane EC328	14408	25	mL	9/30/2024

Stock Source	Base Units	Amount Added
DRO200430B	ug/mL	0.1012 g



Analytical RunID GCFID-HP5-B_220215A Standards Traceability Report

Standard ID: DRO211118A

Standard Name: 50,000 ug/mL Oil Std For AK103 RRO-In DCM

Prep Date: 11/18/2021

Exp Date: 10/31/2028

Department: dropr

Vendor: Restek

Lot Number: A0176667

Balance ID: Sartorius 4 place balance

Comments:

Type: Primary

Prep By: Ann Nebel

Status: Open

Final Volume: 1 mL

Chemical/Solvent Used	Bottle No	Amt	Units	Expires
Residual Range Calibration Standard	14531	1	mL	10/31/2028
Stock Source	Base Units	Amount Added		
DRO211118A	ug/mL			



Analytical RunID GCFID-HP5-B_220215A Standards Traceability Report

Standard ID: DRO211214C

Standard Name: Diesel Fuel #2 50,000 ug/mL in DCM

Prep Date: 12/14/2021

Exp Date: 4/30/2023

Department: dropr

Vendor: Sigma-Aldrich

Lot Number: LRAC6316

Balance ID:

Comments: Diesel Fuel #2 For CCVs.

Type: Primary

Prep By: Ann Nebel

Status: New

Final Volume: mL

Chemical/Solvent Used	Bottle No	Amt	Units	Expires
Diesel Fuel No. 2	14623	1	mL	4/30/2023
Stock Source	Base Units	Amount Added		
DRO211214C	ug/mL			



Analytical RunID GCFID-HP5-B_220215A Standards Traceability Report

Standard ID: DRO220110A

Standard Name: Carbon Scan STD-Marker

Prep Date: 1/11/2022

Exp Date: 7/13/2026

Department: dropr

Vendor: ASI2

Lot Number: 55064

Balance ID:

Comments: FOR Qualitative analyst only.31 compounds-C5 to C30,32,34,36,38,40.

Type: Neat

Prep By: Ann Nebel

Status: Open

Final Volume: 1.2 mL

Chemical/Solvent Used	Bottle No	Amt	Units	Expires
n-Hydrocarbons- C5 to C30, C32, C34, C36, C38, C40	14737	1.2	mL	7/13/2026

Stock Source	Base Units	Amount Added
DRO220110A	ug/mL	



Analytical RunID GCFID-HP5-B_220215A Standards Traceability Report

Standard ID: DRO220119A

Standard Name: Triacontane SURR 1000 ug/mL

Prep Date: 1/19/2022

Exp Date: 4/6/2026

Department: dropr

Vendor:

Lot Number:

Balance ID: BAL-DRO

Comments: 2X dilution of Triacontane SURR 2000 ug/mL

Type: Secondary

Prep By: Jillian L Bostwick

Status: New

Final Volume: 10 mL

Chemical/Solvent Used	Bottle No	Amt	Units	Expires
Dichloromethane EC849	14747	5	mL	4/6/2026

Stock Source	Base Units	Amount Added
DRO211006A	ug/mL	5 mL



Analytical RunID GCFID-HP5-B_220215A Standards Traceability Report

Standard ID: DRO220128A

Standard Name: 8015 CCV-15,000ug/mL + 200 OTP

Prep Date: 1/28/2022

Exp Date: 4/30/2023

Department: dropr

Vendor:

Lot Number:

Balance ID:

Comments: 8015DRO CCV MIX-15,000ug/mL +200 OTP #2 Diesel

Type: Secondary

Prep By: Jillian L Bostwick

Status: New

Final Volume: 4 mL

Chemical/Solvent Used	Bottle No	Amt	Units	Expires
Dichloromethane EC 978	14777	2.6	mL	4/30/2023

Stock Source	Base Units	Amount Added
DRO211214C	ug/mL	1.2 mL
DRO211101A	ug/mL	0.2 mL



Analytical RunID GCFID-HP5-B_220215A Standards Traceability Report

Standard ID: DRO220128B

Standard Name: Carbon Scan STD-Marker

Prep Date: 1/28/2022

Exp Date: 7/13/2026

Department: dropr

Vendor: ASI2

Lot Number: 071306

Balance ID:

Comments: FOR Qualitative analyst only.31 compounds-C5 to C30,32,34,36,38,40.

Type: Primary

Prep By: Jillian L Bostwick

Status: Open

Final Volume: 2.4 mL

Chemical/Solvent Used	Bottle No	Amt	Units	Expires
Carbon Disulfide 55064	7477	1.2	mL	7/13/2026

Stock Source	Base Units	Amount Added
DRO220110A	ug/mL	1.2 mL



Analytical RunID GCFID-HP5-B_220215A Standards Traceability Report

Standard ID: DRO220201A

Standard Name: 5,000 ug/mL RRO CCV 200 ug/mL Triacontane

Type: Secondary

Prep Date: 2/1/2022

Prep By: Ann Nebel

Exp Date: 4/6/2026

Status: New

Department: dropr

Vendor:

Final Volume: 4 mL

Lot Number:

Balance ID:

Comments: CCV for AK102 and 8015C RRO.

Chemical/Solvent Used	Bottle No	Amt	Units	Expires
Dichloromethane EC 978	14777	2.8	mL	4/6/2026

Stock Source	Base Units	Amount Added
DRO220119A	ug/mL	800 µL
DRO211118A	ug/mL	400 µL

Anna

660 Tower Lane • P.O. Box 599 • West Chester, PA 19381-0599
1-800-452-9994 • 1-610-692-3026 • Fax 1-610-692-8729
info@chemservice.com • www.chemservice.com

CERTIFICATE OF ANALYSIS

o-Terphenyl

CATALOG NUMBER N-12693-500MG
LOT NUMBER 9972100
DATE CERTIFIED 09/23/19
EXPIRATION DATE 09/30/24
CAS NUMBER 84-15-1
MOLECULAR FORMULA C18H14
MOLECULAR WEIGHT 230.32
STORAGE Store in a cool dry place.
HANDLING See Safety Data Sheet
INTENDED USE For laboratory use only.

Analytical Test	Value
FT-IR SPECTROSCOPY	CONFORMS TO STRUCTURE
GC/MS SPECTRA ID	MATCHES NIST DATABASE
MELTING POINT (°C)	57.1
% PURITY (GC/FID)	99.5

Chem Service, Inc. guarantees the purity to be +/- 0.5% deviation prior to the expiration date shown on the label and exclusive of any customer contamination.

Certified By:

Mary Beth O'Donnell

Mary Beth O'Donnell
CSM/TC

ID #: 12650

Opened: _____

o-Terphenyl

Expires: 9/30/2024

Rec'd: 4/30/2020

Energyl Laboratories Inc 1120 So. 27th Street
Billings MT 59107


Chem Service is accredited to ISO 17034:2016, ISO/IEC 17025:2017 and certified to ISO 9001:2015



CERTIFICATE OF ANALYSIS

Analysis Method:

Catalog Number: N-12693-500MG
Description: o-Terphenyl
Lot Number: 9972100
Expiration Date: 09/30/24



Chem Service is accredited to ISO 17034:2016, ISO/IEC 17025:2017 and certified to ISO 9001:2015



CERTIFICATE OF ANALYSIS

Analysis Method:

Catalog Number: N-12693-500MG
Description: o-Terphenyl
Lot Number: 9972100
Expiration Date: 09/30/24

Chem Service Inc Area Percent Report

Data File: D:\msdchem\2019 DATA\0919\0923-01.D
Acq On : 23 Sep 2019 10:40
Operator :
Sample : n-12693
Misc :
ALS Vial : 95

Integration Parameters: autoint1.e
Integrator: ChemStation

DataAcq Meth: SCREEN.M
Method : D:\msdchem\2019 DATA\0919\0903-09.D\ERIN.M

Signal : TIC: 0923-01.D\data.ms

peak #	R.T. min	first scan	max scan	last scan	PK TY	peak height	corr. area	corr. % max.	% of total
1	11.844	1597	1606	1613	BB	32038221	432253484	100.00%	100.000%

Sum of corrected areas: 432253484

ERIN.M Mon Sep 23 10:55:51 2019


660 Tower Lane • P.O. Box 599 • West Chester, PA 19381-0599
1-800-452-9994 • 1-610-692-3026 • Fax 1-610-692-8729
info@chemservice.com • www.chemservice.com

CERTIFICATE OF ANALYSIS


Analysis Method:

Catalog Number: N-12693-500MG
Description: o-Terphenyl
Lot Number: 9972100
Expiration Date: 09/30/24

Abundance



Chem Service is accredited to ISO 17034:2016, ISO/IEC 17025:2017 and certified to ISO 9001:2015




660 Tower Lane • P. O. Box 599 • West Chester, PA 19381-0599
1-800-452-9994 • 1-610-692-3026 • Fax 1-610-692-8729
info@chemservice.com • www.chemservice.com

CERTIFICATE OF ANALYSIS

Analysis Method:

Catalog Number: N-12693-500MG
Description: o-Terphenyl
Lot Number: 9972100
Expiration Date: 09/30/24

Abundance



Chem Service is accredited to ISO 17034:2016, ISO/IEC 17025:2017 and certified to ISO 9001:2015.



660 Tower Lane • P.O. Box 599 • West Chester, PA 19381-0599
1-800-452-9994 • 1-610-692-3026 • Fax 1-610-692-8729
info@chemservice.com • www.chemservice.com

CERTIFICATE OF ANALYSIS

Analysis Method:

Catalog Number:	N-12693-500MG
Description:	o-Terphenyl
Lot Number:	9972100
Expiration Date:	09/30/24

Chem Service is accredited to ISO 17034:2016, ISO/IEC 17025:2017 and certified to ISO 9001:2015




660 Tower Lane • P.O. Box 599 • West Chester, PA 19381-0599
 1-800-452-9994 • 1-610-692-3026 • Fax 1-610-692-8729
info@chemservice.com • www.chemservice.com

Gas

Data file: C:\CHEM3\
 Sample name: N-12893
 Instrument: GC 2
 Injection date: 8/23/2019 9:58:34 AM
 Acq. method: SCREEN.M
 Column name: HP-5

CERTIFICATE OF ANALYSIS

Location: Vial 141
 Injection volume: 1.0uL



Signal: FID1 A, Front Signal

RT [min]	Type	Width [min]	Area	Height	Area%
4.565	BB	0.0305	1.2408	0.5122	0.11
20.152	BB	0.0391	1171.9556	439.4599	99.89
		Sum	1173.1963		

Chem Service is accredited to ISO 17034:2016, ISO/IEC 17025:2017 and certified to ISO 9001:2015



3050 Spruce Street, Saint Louis, MO 63103, USA

Website: www.sigmaaldrich.com

Email USA: techserv@sial.com

Outside USA: eurtechserv@sial.com

Certificate of Analysis

Product Name:
Triacontane-d62 - 98 atom % D

Product Number: 451789
 Batch Number: MBBC4347
 Brand: ALDRICH
 CAS Number: 93952-07-9
 MDL Number: MFCD00209794
 Formula: C30D62
 Formula Weight: 485.20 g/mol
 Quality Release Date: 27 APR 2018



ID #: 13736

Opened: _____


Triacontane-d62-98 atom % D

Expires: 4/6/2026

Rec'd: 4/6/2021

Energx Laboratories Inc 1120 So. 27th Street
 Billings MT 59107

Test	Specification	Result
Purity (HPLC)	≥ 99.0 %	99.0 %
Proton NMR Spectrum	Conforms to Structure	Conforms
D Enrichment	≥ 98.0 %	99.0 %
Initial Melting Point		60.0 °C
Final Melting Point		62.0 °C



Greg Abernathy, Supervisor
 Quality Control
 Miamisburg, Ohio US

Sigma-Aldrich warrants, that at the time of the quality release or subsequent retest date this product conformed to the information contained in this publication. The current Specification sheet may be available at Sigma-Aldrich.com. For further inquiries, please contact Technical Service. Purchaser must determine the suitability of the product for its particular use. See reverse side of invoice or packing slip for additional terms and conditions of sale.



CERTIFIED REFERENCE MATERIAL

110 Benner Circle
Bellefonte, PA 16823-8812
Tel: (800)356-1688
Fax: (814)353-1309

www.restek.com

Certificate of Analysis



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31817

Lot No.: A0176667

Description : Residual Range Calibration Standard (RCS)

Residual Range Calib Std (RCS) 50,000µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL

Pkg Amt: > 1 mL

Expiration Date : October 31, 2028

Storage: 25°C nominal

Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	Grav. Conc. (weight/volume)	Expanded Uncertainty (95% C.L.; K=2)	
1	Motor Oil SAE30 & SAE40 Blend (Pennzoil) CAS # 64742-65-0.F Purity ----%	50,102.0 µg/mL	+/- 293.3582 µg/mL +/- 1,492.1008 µg/mL +/- 1,591.3244 µg/mL	Gravimetric Unstressed Stressed

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

ID #: 14531

Opened: _____

Residual Range Calibration Standard

Expires: 10/31/2028

Rec'd: 11/18/2021

Energ Laboratories Inc 1120 So. 27th Street
Billings MT 59107

Column:

30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)

Carrier Gas:

hydrogen-constant pressure 10 psi.

Temp. Program:

40°C (hold 2 min.) to 330°C
@ 10°C/min. (hold 10 min.)

Inj. Temp:


250°C

Det. Temp:

330°C

Det. Type:

FID



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Sam Moodler

Sam Moodler - Operations Tech I

Date Mixed: 22-Sep-2021

Balance: 1128360905

Alexis Shelow

Alexis Shelow - Operations Tech I

Date Passed: 23-Sep-2021

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/ μ ECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified combined stressed uncertainty value (includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ stressed} = k \sqrt{U_{gravimetric}^2 + U_{homogeneity}^2 + U_{storage\ stability}^2 + U_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- It is important to note that the shipping stability uncertainty was obtained under temperature extremes for specific time intervals; therefore, the certified combined stressed uncertainty value should only be applied to the product if it was stored at non-standard temperature conditions up to and including 7 days. Contact Restek Technical Service at www.restek.com/Contact-Us for use recommendations if your shipment was in-transit for more than 7 days at non-standard temperature conditions.
- Apply the certified combined unstressed uncertainty value if the product was received under standard shipping conditions. Apply the certified combined stressed uncertainty value if the product was received under non-standard conditions as specified below.

Label Conditions	Standard Conditions	Non-Standard Conditions
25°C Nominal (Room Temperature)	< 60°C	≥ 60°C up to 7 days
10°C or colder (Refrigerate)	< 40°C	≥ 40°C up to 7 days
0°C or colder (Freezer) -20°C or colder (Deep Freezer)	< 25°C	≥ 25°C up to 7 days

- Separate (not combined) uncertainty values for gravimetric uncertainty are also displayed on the certificate, if needed, separate homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty values are available by contacting Restek Technical Service at www.restek.com/Contact-Us.
- The packaged amount is the minimum sample size for which uncertainty is valid. The ampules are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampules. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.

Certificate of Analysis

Diesel Fuel No. 2

*Certified
Reference
Material*

Description

Product ID UST148
Lot LRAC6316
Expiration Date April 2023
Manufacturing Date April 2020
Storage Conditions Room Temperature
Solvent/Matrix DICHLOROMETHANE

ID #: 14623

Opened: _____

Diesel Fuel No. 2

Expires: 4/30/2023


Rec'd: 12/14/2021

Energy Laboratories Inc 1120 So. 27th Street
Billings MT 59107

Certified Values

Analyte	Certified Value ^{1,4}	Units	Raw Material Purity,%	Raw Material Lot	CAS
NO.2 FUEL OIL	50001 ± 2770	µg/mL	100.0	LA80505	68476-34-6

Informational Values



Additional Information:

Analytical Method Parameters:

Column: SPB-5, 30 m × 0.53 mm I.D., 1.5 µm film thickness (Column #214)

Carrier Gas: H₂, Flow: 4.0 mL/min

Inlet Temperature: 250 °C, Injection Volume: 1.0 µL

Injection Mode: Split, Split Ratio: 10: 1

Temperature Program: 40 °C (Hold 2 min) @ 15 °C/min to 300 °C (Hold 5 min)

Detector: FID

Detector Temperature: 300 °C



SIGMA-ALDRICH

2931 Soldier Springs Rd. Laramie, Wyoming 82070 USA
800-325-5832
TechService@milliporesigma.com www.sigma-aldrich.com

Description

Lot **LRAC6316**
Expiration Date April 2023
Manufacturing Date April 2020
Storage Conditions Room Temperature
Solvent/Matrix DICHLOROMETHANE

1 Metrological traceability: Traceable to the SI and higher order standards from NIST through an unbroken chain of comparisons. The balance used to weigh raw materials is accurate to +/-0.0001 g and calibrated regularly using mass standards traceable to NIST. All dilutions were performed gravimetrically. Additionally, individual analytes are traceable to NIST SRMs where available and specified above.
4 Ucrm - Uncertainty values in this document are expressed as Expanded Uncertainty (Ucrm) corresponding to the 95% confidence interval. Ucrm is derived from the combined standard uncertainty multiplied by the coverage factor k, which is obtained from a t-distribution and degrees of freedom. The components of combined standard uncertainty include the uncertainties due to characterization, homogeneity, long term stability, and short term stability (transport). The components due to stability are generally considered to be negligible unless otherwise indicated by stability studies. The mathematical representation of the Ucrm calculation is as follows:

$$u_{CRM} = \sqrt{u_{char}^2 + u_{homogeneity}^2 + u_{stability}^2}$$

k: Coverage factor derived from a t-distribution table, based on the degrees of freedom of the data set. Assume 2.0 for a **Confidence interval = 95%**

6 Analytical Value- For QC verification of the certified value only- not to be used in calculations. Represents the analytical data obtained by comparison to a standard as analyzed by the method described in the CoA or another acceptable method. The result may differ from the certified value and UCRM based on method uncertainty as well as the uncertainty associated with the standard used for comparison.


Traceability: The standard was manufactured under an ISO/IEC 17025:2017 certified quality system. The balance used to weigh raw materials is accurate to +/- 0.0001g and calibrated regularly using mass standards traceable to NIST. All dilutions were performed gravimetrically. Additionally, individual analytes are traceable to NIST SRMs where available and specified above.

Homogeneity: Homogeneity was assessed in accordance with ISO 17034:2016. Completed units were sampled using a random stratified sampling protocol. The results of chemical analysis were then compared using a one-way analysis of variance approach as described by TNI EL-V3-2009 Appendix A.2. See Instructions for minimum sub-sample size.

Expiration is at end of month given on certificate and label.

MSDS reports for components comprising greater than 1.0% of the solution or 0.1% for components known to be carcinogens are available upon request.

THIS PRODUCT WAS DESIGNED, PRODUCED AND VERIFIED FOR ACCURACY AND STABILITY IN ACCORDANCE WITH ISO/IEC 17025:2017 (ANAB Cert AT-1467) and ISO 17034:2016 (ANAB Cert AR-1470).



Andy Ommen - QC Manager

Certification Date April 30, 2020
Version 0-4302020



Mark Pooler - QA Supervisor





Analytical RunID GCFID-HP5-B_220215B Standards Traceability Report

Standard ID: DRO200430B
Standard Name: O-Terphenyl
Prep Date: 4/30/2020
Exp Date: 9/30/2024
Department: dropr
Vendor: Chemservice
Lot Number: 9972100
Balance ID:
Comments: ID#: 6271

Type: Neat
Prep By: Ann Nebel
Status: New

Final Volume: mL

Chemical/Solvent Used	Bottle No	Amt	Units	Expires
o-Terphenyl	12650	500	mg	9/30/2024
Stock Source	Base Units	Amount Added		



Analytical RunID GCFID-HP5-B_220215B Standards Traceability Report

Standard ID: DRO210406A

Standard Name: Triacontane-d62 Surr For AK103 RRO

Prep Date: 4/6/2021

Exp Date: 4/6/2026

Department: dropr

Vendor: Sigma-Aldrich

Lot Number: MBBC4347

Balance ID:

Comments: Alaska surr [for AK103 RRO]

Type: Neat

Prep By: Ann Nebel

Status: New

Final Volume: mL

Chemical/Solvent Used	Bottle No	Amt	Units	Expires
Triacontane-d62-98 atom % D	13736	500	mg	4/6/2026
Stock Source	Base Units	Amount Added		



Analytical RunID GCFID-HP5-B_220215B Standards Traceability Report

Standard ID: DRO211006A

Standard Name: Triacontane SURR 2000 ug/mL

Prep Date: 10/6/2021

Exp Date: 4/6/2026

Department: dropr

Vendor:

Lot Number:

Balance ID: BAL-DRO

Comments: Triacontane SURR 2000 ug/mL

Type: Primary

Prep By: Jillian L Bostwick

Status: New

Final Volume: 50 mL

Chemical/Solvent Used	Bottle No	Amt	Units	Expires
Acetone DZ509	13553	50	mL	4/6/2026

Stock Source	Base Units	Amount Added
DRO210406A	ug/mL	0.1001 g



Analytical RunID GCFID-HP5-B_220215B Standards Traceability Report

Spike ID: DRO211101A

Spike Name: OTP-4000 ug/mL DCM

Prep Date: 11/1/2021

Exp Date: 9/30/2024

Department: dropr

Vendor:

Lot Number:

Balance ID: BAL-DRO

Comments: Used to Prep DRO-8015 ICAL and CCV Solutions

Type: Secondary

Prep By: Ann Nebel

Status: Open

Final Volume: 25 mL

Chemical/Solvent Used	Bottle No	Amt	Units	Expires
Dichloromethane EC328	14408	25	mL	9/30/2024

Stock Source	Base Units	Amount Added
DRO200430B	ug/mL	0.1012 g



Analytical RunID GCFID-HP5-B_220215B Standards Traceability Report

Standard ID: DRO211118A

Standard Name: 50,000 ug/mL Oil Std For AK103 RRO-In DCM

Prep Date: 11/18/2021

Exp Date: 10/31/2028

Department: dropr

Vendor: Restek

Lot Number: A0176667

Balance ID: Sartorius 4 place balance

Comments:

Type: Primary

Prep By: Ann Nebel

Status: Open

Final Volume: 1 mL

Chemical/Solvent Used	Bottle No	Amt	Units	Expires
Residual Range Calibration Standard	14531	1	mL	10/31/2028
Stock Source	Base Units	Amount Added		
DRO211118A	ug/mL			



Analytical RunID GCFID-HP5-B_220215B Standards Traceability Report

Standard ID: DRO211214C

Standard Name: Diesel Fuel #2 50,000 ug/mL in DCM

Prep Date: 12/14/2021

Exp Date: 4/30/2023

Department: dropr

Vendor: Sigma-Aldrich

Lot Number: LRAC6316

Balance ID:

Comments: Diesel Fuel #2 For CCVs.

Type: Primary

Prep By: Ann Nebel

Status: New

Final Volume: mL

Chemical/Solvent Used	Bottle No	Amt	Units	Expires
Diesel Fuel No. 2	14623	1	mL	4/30/2023
Stock Source	Base Units	Amount Added		
DRO211214C	ug/mL			



Analytical RunID GCFID-HP5-B_220215B Standards Traceability Report

Standard ID: DRO220110A

Standard Name: Carbon Scan STD-Marker

Prep Date: 1/11/2022

Exp Date: 7/13/2026

Department: dropr

Vendor: ASI2

Lot Number: 55064

Balance ID:

Comments: FOR Qualitative analyst only.31 compounds-C5 to C30,32,34,36,38,40.

Type: Neat

Prep By: Ann Nebel

Status: Open

Final Volume: 1.2 mL

Chemical/Solvent Used	Bottle No	Amt	Units	Expires
n-Hydrocarbons- C5 to C30, C32, C34, C36, C38, C40	14737	1.2	mL	7/13/2026

Stock Source	Base Units	Amount Added
DRO220110A	ug/mL	



Analytical RunID GCFID-HP5-B_220215B Standards Traceability Report

Standard ID: DRO220119A

Standard Name: Triacontane SURR 1000 ug/mL

Prep Date: 1/19/2022

Exp Date: 4/6/2026

Department: dropr

Vendor:

Lot Number:

Balance ID: BAL-DRO

Comments: 2X dilution of Triacontane SURR 2000 ug/mL

Type: Secondary

Prep By: Jillian L Bostwick

Status: New

Final Volume: 10 mL

Chemical/Solvent Used	Bottle No	Amt	Units	Expires
Dichloromethane EC849	14747	5	mL	4/6/2026

Stock Source	Base Units	Amount Added
DRO211006A	ug/mL	5 mL



Analytical RunID GCFID-HP5-B_220215B Standards Traceability Report

Standard ID: DRO220128A

Standard Name: 8015 CCV-15,000ug/mL + 200 OTP

Prep Date: 1/28/2022

Exp Date: 4/30/2023

Department: dropr

Vendor:

Lot Number:

Balance ID:

Comments: 8015DRO CCV MIX-15,000ug/mL +200 OTP #2 Diesel

Type: Secondary

Prep By: Jillian L Bostwick

Status: New

Final Volume: 4 mL

Chemical/Solvent Used	Bottle No	Amt	Units	Expires
Dichloromethane EC 978	14777	2.6	mL	4/30/2023

Stock Source	Base Units	Amount Added
DRO211214C	ug/mL	1.2 mL
DRO211101A	ug/mL	0.2 mL



Analytical RunID GCFID-HP5-B_220215B Standards Traceability Report

Standard ID: DRO220128B

Standard Name: Carbon Scan STD-Marker

Prep Date: 1/28/2022

Exp Date: 7/13/2026

Department: dropr

Vendor: ASI2

Lot Number: 071306

Balance ID:

Comments: FOR Qualitative analyst only.31 compounds-C5 to C30,32,34,36,38,40.

Type: Primary

Prep By: Jillian L Bostwick

Status: Open

Final Volume: 2.4 mL

Chemical/Solvent Used	Bottle No	Amt	Units	Expires
Carbon Disulfide 55064	7477	1.2	mL	7/13/2026

Stock Source	Base Units	Amount Added
DRO220110A	ug/mL	1.2 mL



Analytical RunID GCFID-HP5-B_220215B Standards Traceability Report

Standard ID: DRO220201A

Standard Name: 5,000 ug/mL RRO CCV 200 ug/mL Triacontane

Type: Secondary

Prep Date: 2/1/2022

Prep By: Ann Nebel

Exp Date: 4/6/2026

Status: New

Department: dropr

Vendor:

Final Volume: 4 mL

Lot Number:

Balance ID:

Comments: CCV for AK102 and 8015C RRO.

Chemical/Solvent Used	Bottle No	Amt	Units	Expires
Dichloromethane EC 978	14777	2.8	mL	4/6/2026

Stock Source	Base Units	Amount Added
DRO220119A	ug/mL	800 µL
DRO211118A	ug/mL	400 µL

660 Tower Lane • P.O. Box 599 • West Chester, PA 19381-0599
1-800-452-9994 • 1-610-692-3026 • Fax 1-610-692-8729
info@chemservice.com • www.chemservice.com

CERTIFICATE OF ANALYSIS

o-Terphenyl

CATALOG NUMBER N-12693-500MG
LOT NUMBER 9972100
DATE CERTIFIED 09/23/19
EXPIRATION DATE 09/30/24
CAS NUMBER 84-15-1
MOLECULAR FORMULA C18H14
MOLECULAR WEIGHT 230.32
STORAGE Store in a cool dry place.
HANDLING See Safety Data Sheet
INTENDED USE For laboratory use only.

Analytical Test	Value
FT-IR SPECTROSCOPY	CONFORMS TO STRUCTURE
GC/MS SPECTRA ID	MATCHES NIST DATABASE
MELTING POINT (°C)	57.1
% PURITY (GC/FID)	99.5

Chem Service, Inc. guarantees the purity to be +/- 0.5% deviation prior to the expiration date shown on the label and exclusive of any customer contamination.

Certified By:

Mary Beth O'Donnell

Mary Beth O'Donnell
CSM/TC

ID #: 12650

Opened: _____

o-Terphenyl

Expires: 9/30/2024

Rec'd: 4/30/2020


Energyl Laboratories Inc 1120 So. 27th Street

Billings MT 59107

CERTIFICATE OF ANALYSIS

Analysis Method:

Catalog Number: N-12693-500MG
Description: o-Terphenyl
Lot Number: 9972100
Expiration Date: 09/30/24



Chem Service is accredited to ISO 17034:2016, ISO/IEC 17025:2017 and certified to ISO 9001:2015



660 Tower Lane • P.O. Box 599 • West Chester, PA 19381-0599
1-800-452-9994 • 1-610-692-3026 • Fax 1-610-692-8729
info@chemservice.com • www.chemservice.com

CERTIFICATE OF ANALYSIS

Analysis Method:

Catalog Number: N-12693-500MG
Description: o-Terphenyl
Lot Number: 9972100
Expiration Date: 09/30/24

Chem Service Inc Area Percent Report

Data File: D:\msdchem\2019 DATA\0919\0923-01.D
Acq On : 23 Sep 2019 10:40
Operator :
Sample : n-12693
Misc :
ALS Vial : 95

Integration Parameters: autoint1.e
Integrator: ChemStation

DataAcq Meth: SCREEN.M
Method : D:\msdchem\2019 DATA\0919\0903-09.D\ERIN.M

Signal : TIC: 0923-01.D\data.ms

peak #	R.T. min	first scan	max scan	last scan	PK TY	peak height	corr. area	corr. % max.	% of total
1	11.844	1597	1606	1613	BB	32038221	432253484	100.00%	100.000%

Sum of corrected areas: 432253484

ERIN.M Mon Sep 23 10:55:51 2019

Chem Service is accredited to ISO 17034:2016, ISO/IEC 17025:2017 and certified to ISO 9001:2015




660 Tower Lane • P.O. Box 599 • West Chester, PA 19381-0599
1-800-452-9994 • 1-610-692-3026 • Fax 1-610-692-8729
info@chemservice.com • www.chemservice.com

CERTIFICATE OF ANALYSIS


Analysis Method:

Catalog Number: N-12693-500MG
Description: o-Terphenyl
Lot Number: 9972100
Expiration Date: 09/30/24

Abundance



Chem Service is accredited to ISO 17034:2016, ISO/IEC 17025:2017 and certified to ISO 9001:2015




CERTIFICATE OF ANALYSIS

Analysis Method:

Catalog Number: N-12693-500MG
Description: o-Terphenyl
Lot Number: 9972100
Expiration Date: 09/30/24

Abundance



Chem Service is accredited to ISO 17034:2016, ISO/IEC 17025:2017 and certified to ISO 9001:2015.



660 Tower Lane • P.O. Box 599 • West Chester, PA 19381-0599
1-800-452-9994 • 1-610-692-3026 • Fax 1-610-692-8729
info@chemservice.com • www.chemservice.com

CERTIFICATE OF ANALYSIS

Analysis Method:

Catalog Number:	N-12693-500MG
Description:	o-Terphenyl
Lot Number:	9972100
Expiration Date:	09/30/24

Chem Service is accredited to ISO 17034:2016, ISO/IEC 17025:2017 and certified to ISO 9001:2015




660 Tower Lane • P.O. Box 599 • West Chester, PA 19381-0599
 1-800-452-9994 • 1-610-692-3026 • Fax 1-610-692-8729
info@chemservice.com • www.chemservice.com

Gas

Data file: C:\CHEM3\
 Sample name: N-12893
 Instrument: GC 2
 Injection date: 8/23/2019 9:58:34 AM
 Acq. method: SCREEN.M
 Column name: HP-5

CERTIFICATE OF ANALYSIS

Location: Vial 141
 Injection volume: 1.0uL



Signal: FID1 A, Front Signal

RT [min]	Type	Width [min]	Area	Height	Area%
4.565	BB	0.0305	1.2408	0.5122	0.11
20.152	BB	0.0391	1171.9556	439.4599	99.89
		Sum	1173.1963		

Chem Service is accredited to ISO 17034:2016, ISO/IEC 17025:2017 and certified to ISO 9001:2015



3050 Spruce Street, Saint Louis, MO 63103, USA

Website: www.sigmaaldrich.com

Email USA: techserv@sial.com

Outside USA: eurtechserv@sial.com

Certificate of Analysis

Product Name:
Triacontane-d62 - 98 atom % D

Product Number: 451789
 Batch Number: MBBC4347
 Brand: ALDRICH
 CAS Number: 93952-07-9
 MDL Number: MFCD00209794
 Formula: C30D62
 Formula Weight: 485.20 g/mol
 Quality Release Date: 27 APR 2018



ID #: 13736

Opened: _____


Triacontane-d62-98 atom % D

Expires: 4/6/2026

Rec'd: 4/6/2021

Energx Laboratories Inc 1120 So. 27th Street
 Billings MT 59107

Test	Specification	Result
Purity (HPLC)	≥ 99.0 %	99.0 %
Proton NMR Spectrum	Conforms to Structure	Conforms
D Enrichment	≥ 98.0 %	99.0 %
Initial Melting Point		60.0 °C
Final Melting Point		62.0 °C



Greg Abernathy, Supervisor
 Quality Control
 Miamisburg, Ohio US

Sigma-Aldrich warrants, that at the time of the quality release or subsequent retest date this product conformed to the information contained in this publication. The current Specification sheet may be available at Sigma-Aldrich.com. For further inquiries, please contact Technical Service. Purchaser must determine the suitability of the product for its particular use. See reverse side of invoice or packing slip for additional terms and conditions of sale.



CERTIFIED REFERENCE MATERIAL

110 Benner Circle
Bellefonte, PA 16823-8812

Tel: (800)356-1688

Fax: (814)353-1309

www.restek.com

Certificate of Analysis



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31817

Lot No.: A0176667

Description : Residual Range Calibration Standard (RCS)

Residual Range Calib Std (RCS) 50,000µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL

Pkg Amt: > 1 mL

Expiration Date : October 31, 2028

Storage: 25°C nominal

Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	Grav. Conc. (weight/volume)	Expanded Uncertainty (95% C.L.; K=2)		
1	Motor Oil SAE30 & SAE40 Blend (Pennzoil) CAS # 64742-65-0.F Purity ----%	50,102.0 µg/mL	+/- 293.3582	µg/mL	Gravimetric
	(Lot A0126386)		+/- 1,492.1008	µg/mL	Unstressed
			+/- 1,591.3244	µg/mL	Stressed

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

ID #: 14531

Opened: _____

Residual Range Calibration Standard

Expires: 10/31/2028

Rec'd: 11/18/2021

Energv Laboratories Inc 1120 So. 27th Street

Billings MT 59107

Column:

30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)

Carrier Gas:

hydrogen-constant pressure 10 psi.

Temp. Program:

40°C (hold 2 min.) to 330°C
@ 10°C/min. (hold 10 min.)

Inj. Temp:


250°C

Det. Temp:

330°C

Det. Type:

FID



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Sam Moodler

Sam Moodler - Operations Tech I

Date Mixed: 22-Sep-2021

Balance: 1128360905

Alexis Shelow

Alexis Shelow - Operations Tech I

Date Passed: 23-Sep-2021

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/ μ ECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified combined stressed uncertainty value (includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ stressed} = k \sqrt{U_{gravimetric}^2 + U_{homogeneity}^2 + U_{storage\ stability}^2 + U_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- It is important to note that the shipping stability uncertainty was obtained under temperature extremes for specific time intervals; therefore, the certified combined stressed uncertainty value should only be applied to the product if it was stored at non-standard temperature conditions up to and including 7 days. Contact Restek Technical Service at www.restek.com/Contact-Us for use recommendations if your shipment was in-transit for more than 7 days at non-standard temperature conditions.
- Apply the certified combined unstressed uncertainty value if the product was received under standard shipping conditions. Apply the certified combined stressed uncertainty value if the product was received under non-standard conditions as specified below.

Label Conditions	Standard Conditions	Non-Standard Conditions
25°C Nominal (Room Temperature)	< 60°C	≥ 60°C up to 7 days
10°C or colder (Refrigerate)	< 40°C	≥ 40°C up to 7 days
0°C or colder (Freezer) -20°C or colder (Deep Freezer)	< 25°C	≥ 25°C up to 7 days

- Separate (not combined) uncertainty values for gravimetric uncertainty are also displayed on the certificate, if needed, separate homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty values are available by contacting Restek Technical Service at www.restek.com/Contact-Us.
- The packaged amount is the minimum sample size for which uncertainty is valid. The ampules are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampules. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.

Certificate of Analysis

Diesel Fuel No. 2

*Certified
Reference
Material*

Description

Product ID UST148
Lot LRAC6316
Expiration Date April 2023
Manufacturing Date April 2020
Storage Conditions Room Temperature
Solvent/Matrix DICHLOROMETHANE

ID #: 14623

Opened: _____

Diesel Fuel No. 2

Expires: 4/30/2023


Rec'd: 12/14/2021

Energy Laboratories Inc 1120 So. 27th Street
Billings MT 59107

Certified Values

Analyte	Certified Value ^{1,4}	Units	Raw Material Purity,%	Raw Material Lot	CAS
NO.2 FUEL OIL	50001 ± 2770	µg/mL	100.0	LA80505	68476-34-6

Informational Values



Additional Information:

Analytical Method Parameters:

Column: SPB-5, 30 m × 0.53 mm I.D., 1.5 µm film thickness (Column #214)

Carrier Gas: H₂, Flow: 4.0 mL/min

Inlet Temperature: 250 °C, Injection Volume: 1.0 µL

Injection Mode: Split, Split Ratio: 10: 1

Temperature Program: 40 °C (Hold 2 min) @ 15 °C/min to 300 °C (Hold 5 min)

Detector: FID

Detector Temperature: 300 °C



SIGMA-ALDRICH®

2931 Soldier Springs Rd. Laramie, Wyoming 82070 USA
800-325-5832
TechService@milliporesigma.com www.sigma-aldrich.com

Description

Lot **LRAC6316**
Expiration Date April 2023
Manufacturing Date April 2020
Storage Conditions Room Temperature
Solvent/Matrix DICHLOROMETHANE

1 Metrological traceability: Traceable to the SI and higher order standards from NIST through an unbroken chain of comparisons. The balance used to weigh raw materials is accurate to +/-0.0001 g and calibrated regularly using mass standards traceable to NIST. All dilutions were performed gravimetrically. Additionally, individual analytes are traceable to NIST SRMs where available and specified above.
4 Ucrm - Uncertainty values in this document are expressed as Expanded Uncertainty (Ucrm) corresponding to the 95% confidence interval. Ucrm is derived from the combined standard uncertainty multiplied by the coverage factor k, which is obtained from a t-distribution and degrees of freedom. The components of combined standard uncertainty include the uncertainties due to characterization, homogeneity, long term stability, and short term stability (transport). The components due to stability are generally considered to be negligible unless otherwise indicated by stability studies. The mathematical representation of the Ucrm calculation is as follows:

$$u_{CRM} = \sqrt{u_{char}^2 + u_{homogeneity}^2 + u_{stability}^2}$$

k: Coverage factor derived from a t-distribution table, based on the degrees of freedom of the data set. Assume 2.0 for a **Confidence interval = 95%**

6 Analytical Value- For QC verification of the certified value only- not to be used in calculations. Represents the analytical data obtained by comparison to a standard as analyzed by the method described in the CoA or another acceptable method. The result may differ from the certified value and UCRM based on method uncertainty as well as the uncertainty associated with the standard used for comparison.


Traceability: The standard was manufactured under an ISO/IEC 17025:2017 certified quality system. The balance used to weigh raw materials is accurate to +/- 0.0001g and calibrated regularly using mass standards traceable to NIST. All dilutions were performed gravimetrically. Additionally, individual analytes are traceable to NIST SRMs where available and specified above.

Homogeneity: Homogeneity was assessed in accordance with ISO 17034:2016. Completed units were sampled using a random stratified sampling protocol. The results of chemical analysis were then compared using a one-way analysis of variance approach as described by TNI EL-V3-2009 Appendix A.2. See Instructions for minimum sub-sample size.

Expiration is at end of month given on certificate and label.

MSDS reports for components comprising greater than 1.0% of the solution or 0.1% for components known to be carcinogens are available upon request.

THIS PRODUCT WAS DESIGNED, PRODUCED AND VERIFIED FOR ACCURACY AND STABILITY IN ACCORDANCE WITH ISO/IEC 17025:2017 (ANAB Cert AT-1467) and ISO 17034:2016 (ANAB Cert AR-1470).



Andy Ommen - QC Manager

Certification Date April 30, 2020
Version 0-4302020



Mark Pooler - QA Supervisor

